

IEC 60601-2-19
(Second edition – 2009)

Medical electrical equipment –

**Part 2-19: Particular requirements for the basic safety
and essential performance of infant incubators**

CORRIGENDUM 1

201.4.3.101 ESSENTIAL PERFORMANCE OF INFANT INCUBATORS

Replace the existing Table 201.101 by the following:

Table 201.101 – Additional ESSENTIAL PERFORMANCE requirements

Requirement	Subclause
ESSENTIAL PERFORMANCE requirement 1	201.12.1.104 or generation of a visual and audible alarm in compliance with 201.15.4.2.1 ee)
ESSENTIAL PERFORMANCE requirement 2	201.12.1.106 or generation of a visual and audible alarm in compliance with 201.15.4.2.1 dd)

201.9.6.2.1.101 * Sound level within the COMPARTMENT

In the first paragraph, replace the reference to "201.12.3.107" by "201.9.6.2.1.103".

201.12.1.107 * Warm-up time

In the first paragraph, replace "(see item subclause 201.7.9.2.2)" by "(see subclause 201.7.9.2.8)".