

## CORRIGENDUM 1

### 201.3.201

Replace the existing term " $B_1$ RMS" by " $B_{1+}$ RMS".

Replace, in the definition:

" $B_1$ , the radio frequency magnetic induction"  
by

" $B_{1+}$ , the MR relevant radiofrequency  
magnetic induction"

Replace, in the formula, " $B_1$ RMS" by " $B_{1+}$ RMS"  
and " $B_1$ " by " $B_{1+}$ ".

### 201.3.201

Remplacer le terme existant " $B_1$ RMS" par " $B_{1+}$ RMS".

Remplacer, dans la définition:

" $B_1$ , l'induction magnétique à radiofréquence"  
par

" $B_{1+}$ , l'induction magnétique à radiofréquence  
RM pertinente".

Remplacer, dans la formule, " $B_1$ RMS" par " $B_{1+}$ RMS" et " $B_1$ " par " $B_{1+}$ ".

### Table 201.101 – List of symbols

In the 2<sup>nd</sup> row, 1<sup>st</sup> column replace " $B_1$ " by " $B_{1+}$ ".

In the 3<sup>rd</sup> row, 1<sup>st</sup> column replace " $B_1$ RMS" by " $B_{1+}$ RMS".

In the 3<sup>rd</sup> row, 3<sup>rd</sup> column replace " $B_1$ " by " $B_{1+}$ ".

### Tableau 201.101 – Liste des symboles

Dans la 2<sup>ème</sup> ligne, 1<sup>ère</sup> colonne remplacer " $B_1$ " par " $B_{1+}$ ".

Dans la 3<sup>ème</sup> ligne, 1<sup>ère</sup> colonne remplacer " $B_1$ RMS" par " $B_{1+}$ RMS".

Dans la 3<sup>ème</sup> ligne, 3<sup>ème</sup> colonne remplacer " $B_1$ " par " $B_{1+}$ ".

### 201.7.9.2.101 Instructions for use for MR EQUIPMENT

j) Exposure of the PATIENT to radio  
frequency magnetic fields

In the final dashed item replace " $B_1$ RMS" by " $B_{1+}$ RMS"

### 201.7.9.2.101 Instructions d'utilisation des APPAREILS A RM

j) Exposition du PATIENT aux champs  
magnétiques à radiofréquence

Dans le dernier tiret, remplacer " $B_1$ RMS" par " $B_{1+}$ RMS".