

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Medical electrical equipment –

Part 2-52: Particular requirements for the basic safety and essential performance of medical beds

Appareils électromédicaux –

Partie 2-52: Exigences particulières de sécurité de base et de performances essentielles des lits médicaux

IECNORM.COM : Click to view the full PDF of IEC 60601-2-52:2009



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2009 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester.

If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de la CEI ou du Comité national de la CEI du pays du demandeur.

Si vous avez des questions sur le copyright de la CEI ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de la CEI de votre pays de résidence.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland
Email: inmail@iec.ch
Web: www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

- Catalogue of IEC publications: www.iec.ch/searchpub

The IEC on-line Catalogue enables you to search by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, withdrawn and replaced publications.

- IEC Just Published: www.iec.ch/online_news/justpub

Stay up to date on all new IEC publications. Just Published details twice a month all new publications released. Available on-line and also by email.

- Electropedia: www.electropedia.org

The world's leading online dictionary of electronic and electrical terms containing more than 20 000 terms and definitions in English and French, with equivalent terms in additional languages. Also known as the International Electrotechnical Vocabulary online.

- Customer Service Centre: www.iec.ch/webstore/custserv

If you wish to give us your feedback on this publication or need further assistance, please visit the Customer Service Centre FAQ or contact us:

Email: csc@iec.ch

Tel.: +41 22 919 02 11

Fax: +41 22 919 03 00

A propos de la CEI

La Commission Electrotechnique Internationale (CEI) est la première organisation mondiale qui élabore et publie des normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications CEI

Le contenu technique des publications de la CEI est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

- Catalogue des publications de la CEI: www.iec.ch/searchpub/cur_fut-f.htm

Le Catalogue en-ligne de la CEI vous permet d'effectuer des recherches en utilisant différents critères (numéro de référence, texte, comité d'études,...). Il donne aussi des informations sur les projets et les publications retirées ou remplacées.

- Just Published CEI: www.iec.ch/online_news/justpub

Restez informé sur les nouvelles publications de la CEI. Just Published détaille deux fois par mois les nouvelles publications parues. Disponible en-ligne et aussi par email.

- Electropedia: www.electropedia.org

Le premier dictionnaire en ligne au monde de termes électroniques et électriques. Il contient plus de 20 000 termes et définitions en anglais et en français, ainsi que les termes équivalents dans les langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International en ligne.

- Service Clients: www.iec.ch/webstore/custserv/custserv_entry-f.htm

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions, visitez le FAQ du Service clients ou contactez-nous:

Email: csc@iec.ch

Tél.: +41 22 919 02 11

Fax: +41 22 919 03 00



IEC 60601-2-52

Edition 1.0 2009-12

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Medical electrical equipment –

**Part 2-52: Particular requirements for the basic safety and essential performance
of medical beds**

Appareils électromédicaux –

**Partie 2-52: Exigences particulières de sécurité de base et de performances
essentielles des lits médicaux**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

PRICE CODE
CODE PRIX

XC

ICS 11.140

ISBN 2-8318-1068-1

CONTENTS

FOREWORD.....	5
INTRODUCTION.....	7
201.1 Scope, object and related standards	8
201.2 Normative references.....	9
201.3 Terms and definitions.....	10
201.4 General requirements	13
201.5 General requirements for testing of ME EQUIPMENT	13
201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....	15
201.7 ME EQUIPMENT identification, marking and documents	15
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	20
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	21
201.10 Protection against unwanted and excessive radiation HAZARDS	44
201.11 Protection against excessive temperatures and other HAZARDS	44
201.12 Accuracy of controls and instruments and protection against hazardous outputs	46
201.13 HAZARDOUS SITUATIONS and fault conditions	47
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS).....	47
201.15 Construction of ME EQUIPMENT.....	48
201.16 ME SYSTEMS	51
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	51
Annexes	51
Annex AA (informative) Particular guidance and rationale.....	52
Annex BB (normative) Design requirements and recommendations for MEDICAL BEDS.....	67
Annex CC (informative) Particular guidance for assessing risk of entrapment in v-shaped openings	75
Bibliography.....	81
Index of defined terms used in this particular standard.....	82
Figure 201.101 – APPLIED PART.....	10
Figure 201.102 – MEDICAL BED, general arrangement (example, schematic presentation only)	12
Figure 201.103a – Cone tool.....	14
Figure 201.103b – Cylinder tool	14
Figure 201.103 – Entrapment test tools.....	14
Figure 201.104 – Loading pad	15
Figure 201.105 – Graphic symbol for maximum PATIENT weight and SAFE WORKING LOAD	16
Figure 201.106 – MEDICAL BED function controls and/or actuators: guidelines for creating graphic symbols	18
Figure 201.107 – Example of MEDICAL BED with segmented or split SIDE RAIL	22
Figure 201.108 – Example of MEDICAL BED with single piece SIDE RAIL	23
Figure 201.109 – Allowable spacing for fingers in areas of normal reach around the perimeter of the MATTRESS SUPPORT PLATFORM	28

Figure 201.110 – Example using barriers for clearance measurement around the perimeter of the MATTRESS SUPPORT PLATFORM to mitigate PATIENT-finger entrapment	29
Figure 201.111a – Foot and toe clearance area between moving parts and the floor.....	29
Figure 201.111b – Toe clearance area between moving parts and the floor	30
Figure 201.111 – Clearance areas	30
Figure 201.112 – Lateral stability test along the side of the MEDICAL BED	32
Figure 201.113 – Longitudinal stability test with removable FOOT BOARD	32
Figure 201.114 – Longitudinal stability test with fixed HEAD/FOOT BOARDS.....	33
Figure 201.115 – Distribution of SAFE WORKING LOAD for tests.....	37
Figure 201.116 – Position of loading pad (see Figure 201.104).....	40
Figure 201.117 – Application of forces for test of SIDE RAIL.....	42
Figure 201.118 – Height of SIDE RAIL	43
Figure 201.119a – Angle γ between the back section and the leg section of the MATTRESS SUPPORT PLATFORM.....	49
Figure 201.119b – Angle γ between the back section and the upper leg section of the MATTRESS SUPPORT PLATFORM.....	49
Figure 201.119c – Angle γ between the angled back section and upper leg section of the MATTRESS SUPPORT PLATFORM	49
Figure 201.119d – Angle γ between the angled back section and the leg/upper leg section of the MATTRESS SUPPORT PLATFORM	50
Figure 201.119 – Configurations of the MATTRESS SUPPORT PLATFORM	50
Figure AA.1 – Marking to select recommended mattresses specified by the MANUFACTURER.....	54
Figure AA.2 – Marking for detachable SIDE RAILS specified by the MANUFACTURER	54
Figure AA.3 – Resultant forces without mattress	58
Figure AA.4 – Resultant forces with mattress.....	58
Figure AA.5 – Example of 60 mm gap measurement of B.....	58
Figure AA.6 – Angle measurement example of B	58
Figure AA.7 – Placement of measurement TOOL for measurement of D	59
Figure AA.8 – Example of area D measurement that passes	59
Figure AA.9 – Example of area D measurement that fails.....	59
Figure AA.10 – Example of area D measurement that fails (on limit)	60
Figure AA.11 – Example of potential PATIENT entrapment in area A within the SIDE RAIL	60
Figure AA.12 – Example of potential PATIENT entrapment in area A below the SIDE RAIL	60
Figure AA.13 – Example of potential PATIENT entrapment in area B.....	60
Figure AA.14 – Example of potential PATIENT entrapment in area C between split SIDE RAIL	60
Figure AA.15 – Example of potential PATIENT entrapment in area C between SIDE RAIL and HEAD BOARD	61
Figure AA.16 – Example of potential PATIENT entrapment in area D.....	61
Figure AA.17 – Example of potential PATIENT entrapment in area A below a single piece SIDE RAIL	61
Figure BB.1 – Other areas of possible impact testing.....	68
Figure BB.2 – Impactor	69
Figure BB.3 – Schematic presentation of under MEDICAL BED clearance.....	72

Figure BB.4 – Recommendations and requirements regarding angles for different sections of the MATTRESS SUPPORT PLATFORM	74
Figure CC.1 – Wedge tool.....	76
Figure CC.2 – V-shaped opening in relation to B.....	77
Figure CC.3 – Pass/fail in relation to area B	77
Figure CC.4 – Positioning of wedge tool	78
Figure CC.5 – Pass/fail in relation to area C between HEAD BOARD and FOOT BOARD.....	79
Figure CC.6 – Pass/fail in relation to area C between split SIDE RAILS	80
 Table 201.101 – Protection against PATIENT entrapment	 24
Table 201.102 – Protection against inadvertent PATIENT falls	44
Table 24 – Allowable maximum temperatures for skin contact with MEDICAL BED APPLIED PARTS.....	45
Table BB.1 – Normative and informative requirements for different APPLICATION ENVIRONMENTS 1 to 5	67

IECNORM.COM : Click to view the full PDF of IEC 60601-2-52:2009

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –**Part 2-52: Particular requirements for the basic safety
and essential performance of medical beds**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 60601-2-52 has been prepared by IEC subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice, and by ISO technical committee 173: Assistive products for persons with disability.

It is published as double logo standard.

This first edition cancels and replaces the first edition of IEC 60601-2-38, published in 1996, and its Amendment 1 (1999). This edition constitutes a technical revision.

The text of this particular standard is based on the following documents:

FDIS	Report on voting
62D/795/FDIS	62D/815/RVD

Full information on the voting for the approval of this particular standard can be found in the report on voting indicated in the above table. In ISO, the standard has been approved by 20 P-members out of 20 having cast a vote.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this collateral standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the maintenance result date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

INTRODUCTION

In 1996, the IEC published the first edition of the particular standard for electrically operated hospital beds, IEC 60601-2-38. The publication was in response to demand in the field for a universal standard addressing HAZARDS specific to the safety of the hospital bed. Used in conjunction with a MANUFACTURER'S RISK ASSESSMENT, the standard was felt to be the current thinking on establishing a basic safety benchmark for industry.

An amendment of IEC 60601-2-38 issued in 1999 recognized the need to mitigate against a RISK of PATIENT entrapment in the SIDE RAILS, again combined with the use of the MANUFACTURER'S RISK ASSESSMENT. Although this improved the particular standard, it still was centered upon electrically operated hospital beds, and failed to take into account manually operated hospital beds and products in other medical environments.

In 2000, the EN 1970 standard (*Adjustable beds for DISABLED PERSONS – Requirements and test methods*) was published, which addressed beds used by DISABLED PERSONS to alleviate or compensate for a disability or handicap. This standard offered a broadened scope in conjunction with IEC 60601-2-38, but after the edition of Amendment 1 to IEC 60601-2-38, the opportunity presented itself to combine the two standards to a common, international standard.

As work began on the integration, the IEC adjusted its stance on BASIC SAFETY and ESSENTIAL PERFORMANCE, integrating them into the third edition of IEC 60601-1. It therefore became necessary to align the new standard with the third edition. The particular standard was given a new number, IEC 60601-2-52, and work began on alignment to third edition.

This particular standard, therefore, is the realization of much work in alignment, and scope adjustment between IEC 60601-2-38, EN 1970, and the third edition of IEC 60601-1. It represents the current thinking in BASIC SAFETY and ESSENTIAL PERFORMANCE of the MEDICAL BED as used to alleviate illness of PATIENTS and disability of DISABLED PERSONS. This is the effort of a joint working group of the IEC and the ISO.

IECNORM.COM : Click to view the full PDF of IEC 60601-2-52:2009

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-52: Particular requirements for the basic safety and essential performance of medical beds

201.1 Scope, object and related standards

Clause 1 of the general standard¹⁾ applies, except as follows:

201.1.1 * Scope

Replacement:

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of MEDICAL BEDS intended for adults, hereafter referred to as MEDICAL BED as defined in 201.3.212.

If a clause or subclause is specifically intended to be applicable to a MEDICAL BED only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to MEDICAL BED and to ME SYSTEMS, as relevant.

HAZARDS inherent in the intended physiological function of MEDICAL BED or ME SYSTEMS within the scope of this standard are not covered by specific requirements in this standard except in 7.2.13 and 8.4.1 of the general standard.

NOTE See also 4.2 of the General Standard.

201.1.2 Object

Replacement:

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for MEDICAL BEDS as defined in 201.3.212.

201.1.3 Collateral standards

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 2 of this particular standard.

IEC 60601-1-3, IEC 60601-1-8 and IEC 60601-1-10²⁾ do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

¹⁾ The general standard is IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*.

²⁾ IEC 60601-1-10:2007, *Medical electrical equipment – Part 1-10: General requirements for basic safety and essential performance – Collateral Standard: Requirements for the development of physiologic closed-loop controllers*

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1 is referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this particular standard corresponds to that of the general standard with the prefix "201" (e.g. 201.1 in this standard addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix "20x" where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

"Replacement" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

"Addition" means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral standard.

"Amendment" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However, due to the fact that definitions in the general standard are numbered 3.1 through 3.139, additional definitions in this standard are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term "this standard" is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

201.2 Normative references

NOTE Informative references are listed in the bibliography on page 81.

Clause 2 of the general standard applies except as follows:

Addition:

IEC 60068-2-31:2008, *Environmental testing – Part 2-31: Tests – Test Ec: Rough handling shocks, primarily for equipment-type specimens*

201.3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 60601-1:2005 apply, except as follows:


NOTE An index of defined terms is found beginning on page 82.

201.3.8

APPLIED PART

Addition:

The APPLIED PART includes all parts of the MEDICAL BED that are within reach of the PATIENT, even if they are underneath the MATTRESS SUPPORT SURFACE (see Figure 201.101)



IEC 2115/09

Key

1 Region of APPLIED PART including ACCESSORIES

Figure 201.101 – APPLIED PART

201.3.76

PATIENT

Replacement:

person undergoing a medical procedure or DISABLED PERSON

201.3.131*** TRAPPING ZONE**

Addition:

locations where the body of a MEDICAL BED occupant can become entrapped, entangled, wedged, or stuck in or between parts of the MEDICAL BED, such as the SIDE RAILS, HEAD/FOOT BOARD, MATTRESS SUPPORT PLATFORM or mattress

Addition:

201.3.201*** APPLICATION ENVIRONMENT 1**

intensive/critical care provided in a hospital where 24 h medical supervision and constant monitoring is required and provision of life support system/equipment used in medical procedures is essential to maintain or improve the vital functions of the PATIENT

201.3.202*** APPLICATION ENVIRONMENT 2**

acute care provided in a hospital or other medical facility where medical supervision and monitoring is required and ME EQUIPMENT used in medical procedures is often provided to help maintain or improve the condition of the PATIENT

201.3.203*** APPLICATION ENVIRONMENT 3**

long-term care in a medical area where medical supervision is required and monitoring is provided if necessary and ME EQUIPMENT used in medical procedures may be provided to help maintain or improve the condition of the PATIENT

NOTE This includes use in nursing homes and in rehabilitation and geriatric facilities.

201.3.204*** APPLICATION ENVIRONMENT 4**

care provided in a domestic area where ME EQUIPMENT is used to alleviate or compensate for an injury, disability or disease

NOTE This excludes use in all other APPLICATION ENVIRONMENTS (e.g. nursing homes, rehabilitation and geriatric facilities) when a MEDICAL BED is purely designed for APPLICATION ENVIRONMENT 4.

201.3.205*** APPLICATION ENVIRONMENT 5**

outpatient (ambulatory) care, which is provided in a hospital or other medical facility, under medical supervision where ME EQUIPMENT, is provided for the need of persons with illness, injury or disability for treatment, diagnosis or monitoring

201.3.206**BED-LIFT**

height adjustable mechanism on which a MATTRESS SUPPORT PLATFORM can be mounted

NOTE The combination of a BED LIFT and a compatible non-MEDICAL BED as specified by the MANUFACTURER is considered to be a MEDICAL BED.

201.3.207**DISABLED PERSON**

person with one or more impairments, one or more activity limitations, one or more participation restrictions or a combination thereof

[ISO 9999:2007]

201.3.208

HEAD/FOOT BOARD

assembly/assemblies mounted to MEDICAL BED, which identifies for the PATIENT the edge of the head or foot end of the MEDICAL BED and/or MATTRESS SUPPORT PLATFORM

NOTE It may be used as handles to push a MEDICAL BED intended to transport PATIENTS.

201.3.209

LIFTING POLE

ACCESSORY attached to a MEDICAL BED and intended to assist support of a PATIENT when changing position by providing a gripping support above the PATIENT

201.3.210

MATTRESS OVERLAY


supplementary mattress (support surface), which is intended to be placed on an existing mattress, and generally used for prophylactic or therapeutic effect

201.3.211

MATTRESS SUPPORT PLATFORM

structure which supports a PATIENT surface (for example mattress)

NOTE It can articulate or change positions to facilitate various therapeutic, diagnostic and convenience positions (See Figures 201.102 and 201.119 a) to 201.119 d)).



IEC 2116/09

Key

- 1 HEAD BOARD
- 2 Back section
- 3 Seat section
- 4 Upper leg section
- 5 Lower leg section
- 6 FOOT BOARD

**Figure 201.102 – MEDICAL BED, general arrangement
(example, schematic presentation only)**

201.3.212*** MEDICAL BED**

device for which the INTENDED USE is sleeping/resting that contains a MATTRESS SUPPORT PLATFORM and intended to assist in diagnosis, monitoring, prevention, treatment, alleviation of disease or compensation for an injury or handicap

NOTE 1 A BED-LIFT and/or a detachable MATTRESS SUPPORT PLATFORM in combination with a compatible non-MEDICAL BED as specified by the MANUFACTURER is also considered a MEDICAL BED.

NOTE 2 Excluded are devices for which the INTENDED USE is mainly for examination or transportation under medical supervision (e.g. stretcher, examination table).

201.3.213**MOTION LOCKOUT CONTROL**

auxiliary subsystem that deactivates motion controls

201.3.214**PENDANT CONTROL**

handheld device, which has a FUNCTIONAL CONNECTION to the MEDICAL BED, controlling at least MEDICAL BED articulations and/or movements

NOTE PENDANT CONTROLS may be wired, or wireless, and may integrate other functions, (e.g. communications, radio/tv, etc.).

201.3.215**SIDE RAIL**

physical barrier, which may be a detachable ACCESSORY or integral to the overall construction of a MEDICAL BED and is mounted to the side(s) of the MEDICAL BED

NOTE When a SIDE RAIL is located in the “up” position it provides a physical barrier, which is intended to reduce the RISK of the PATIENT accidentally slipping or rolling off the mattress.

201.3.216**SPECIALTY MATTRESS**

mattress intended for prophylactic or therapeutic effect

201.3.217**TEST BED BOARD**

flat, rigid loading board of dimensions as specified by the MANUFACTURER representing the dimensions of the MEDICAL BED

201.3.218**UNDERCARRIAGE**

all components of the MEDICAL BED or BED-LIFT below the MATTRESS SUPPORT PLATFORM

201.4 General requirements

Clause 4 of the general standard applies.

201.5 General requirements for testing of ME EQUIPMENT

Clause 5 of the general standard applies, except as follows:

Additional subclauses:

201.5.101 * Entrapment test tools

Figures 201.103a and 201.103b illustrate entrapment test tools (a cone tool and a cylinder tool respectively).

Dimensions in millimetres




Figure 201.103a – Cone tool

Dimensions in millimetres




Figure 201.103b – Cylinder tool

Figure 201.103 – Entrapment test tools

201.5.102 Loading pad

The loading pad (see Figure 201.104) is a rigid circular object, (355+/-5) mm in diameter, the face of which has a convex spherical curvature of 800 mm radius with a 20 mm front edge radius.




Figure 201.104 – Loading pad

201.6 Classification of ME EQUIPMENT and ME SYSTEMS

Clause 6 of the general standard applies, except as follows:

201.6.2 * Protection against electrical shock

Addition:

For APPLICATION ENVIRONMENT 4, the MEDICAL BED shall be CLASS II.

201.7 ME EQUIPMENT identification, marking and documents

Clause 7 of the general standard applies, except as follows:

201.7.2.2 Identification

Replacement of the first paragraph:


The MEDICAL BED shall be marked with the name or trademark and address of the MANUFACTURER, MODEL OR TYPE REFERENCE and means to allow traceability.

The detachable components shall be marked with the name or trademark and address of the MANUFACTURER, MODEL OR TYPE REFERENCE and means to allow traceability unless mis-identification does not present an unacceptable RISK.

Additional subclauses:

201.7.2.2.101 * Marking of maximum PATIENT weight and SAFE WORKING LOAD

The MEDICAL BED shall be marked with the corresponding maximum PATIENT weight (see 201.9.8.3.1) and SAFE WORKING LOAD (for symbol see Figure 201.105).



IEC 2120/09

Figure 201.105 – Graphic symbol for maximum PATIENT weight and SAFE WORKING LOAD

Detachable parts of a MEDICAL BED of a mass of more than 20 kg shall be marked with symbol ISO 7000-1321 (2004-01): .

201.7.2.2.102 Marking for machine washable MEDICAL BEDS by an automatic washing system

MEDICAL BEDS intended for use with an automatic washing system shall be marked with the following text to distinguish them from MEDICAL BEDS which cannot tolerate such cleaning methods: “Caution, for cleaning purposes this bed can be used with automatic washing systems.”

201.7.2.2.103 Marking for MEDICAL BEDS intended for jet stream washing

MEDICAL BEDS intended for use with jet stream washing shall be marked with the following text: “Caution, for cleaning purposes this bed can be used with jet stream washing.”

201.7.2.2.104 Width of carriage of BED-LIFT

If a BED-LIFT has an adjustable width carriage, the range shall be marked, e.g. by linear measurement indicator fixed to the adjustable parts.

201.7.2.2.105 * Marking of replacement mattresses

MEDICAL BEDS designed to have replacement mattresses shall be marked with the following sentence as a warning: “Incompatible mattresses can create hazards. Read instructions for use” or a symbol as appropriate on a prominent place on the MATTRESS SUPPORT PLATFORM indicating the compatible mattresses (see example in AA.1).

201.7.2.2.106 * Marking of detachable SIDE RAILS

MEDICAL BEDS designed to have detachable SIDE RAILS shall be marked with the following sentence as a warning: “Incompatible SIDE RAILS can create HAZARDS. Read instructions for use” or use a symbol as appropriate on a prominent place near the attachment point of the SIDE RAIL, indicating the compatible SIDE RAILS (see example in AA.2).

201.7.2.4 ACCESSORIES

Addition:

Where an overload on an ACCESSORY that is intended to support loads can create an unacceptable RISK, the corresponding SAFE WORKING LOAD shall be marked on the ACCESSORY.

Compliance is checked by inspection.

201.7.4.2 Control devices

Addition:

Where reliance to prevent unintended movement of the MATTRESS SUPPORT PLATFORM is based on a MOTION LOCKOUT CONTROL which requires activation by the OPERATOR, this shall be disclosed by markings or symbols as appropriate on the outside of the MEDICAL BED and which are visible from a position of NORMAL USE (see Figure 201.106).

NOTE Example: "Engage the MOTION LOCKOUT CONTROL if a patient could be injured due to inadvertent motion of the MATTRESS SUPPORT PLATFORM."

Compliance is checked by inspection of the MEDICAL BED.


201.7.6.3 Symbols for controls and performance

Addition:

Controls and/or indicators, when possible, shall be marked using symbols that convey the intended function of those controls or indicators without the need for additional text.

NOTE Figure 201.106 is intended as a guideline when designing these symbols. In all cases where standard international symbols exist, they should be used.

IECNORM.COM : Click to view the full PDF of IEC 60601-2-52:2009



IEC 2121/09

Key

- 1 Back section down function
- 2 Back section up function
- 3 Leg down function
- 4 Leg up function
- 5 MATTRESS SUPPORT PLATFORM down
- 6 MATTRESS SUPPORT PLATFORM up
- 7 Trendelenburg
- 8 Anti-Trendelenburg/reverse Trendelenburg
- 9 Padlock-symbol (unlocked) for MOTION LOCKOUT CONTROL
- 10 Padlock-symbol (locked) for MOTION LOCKOUT CONTROL

**Figure 201.106 – MEDICAL BED function controls and/or actuators:
guidelines for creating graphic symbols**

201.7.9.2 Instructions for use**201.7.9.2.1 General**

Addition:

The instructions for use shall include:

- a) a description according to 201.3. of the intended APPLICATION ENVIRONMENT(S);
- b) the maximum PATIENT weight and SAFE WORKING LOAD. The SAFE WORKING LOAD is the sum of:
 - the PATIENT;
 - the mattress;
 - the ACCESSORIES of the MEDICAL BED (only if they are supported by the support system of the MEDICAL BED); and
 - the load supported by those ACCESSORIES (excluding PATIENT weight);
- c) an explanation of how to deactivate any MEDICAL BED function if movement caused by that function could cause injury to the PATIENT;
- d) for APPLICATION ENVIRONMENT 4 intended MEDICAL BEDS, the result of the measurement of auditable acoustic energy in accordance with ISO 3746.

201.7.9.2.2 Warning and safety notices

Addition:

- a) The instructions for use shall provide a warning that the MEDICAL BED should be left in its lowest position when the PATIENT is unattended in order to reduce RISK of injury due to falls.
- b) The instructions for use shall provide a warning on HAZARDS caused by inappropriate handling of the POWER SUPPLY CORD, e. g. by kinking, shearing or other mechanical damages.
- c) The instructions for use shall provide a warning, stating that when routing cables from other equipment in the MEDICAL BED, precautions shall be taken to avoid squeezing those between parts of the MEDICAL BED.
- d) The instructions for use shall provide a warning if a MEDICAL BED shall only be used with certain hoists, because of the limited space underneath the MEDICAL BED.

201.7.9.2.5 ME EQUIPMENT description

Addition:

201.7.9.2.5.101 Selection of mattress

The instructions for use shall contain information on the selection of mattress(es), including mattress dimensions and mattress characteristics (e.g. to reduce the RISK of entrapment and falls (see also 201.7.2.2.105)).

201.7.9.2.5.102 Selection of SIDE RAILS

The instructions for use shall contain information on the selection of SIDE RAILS, including SIDE RAIL dimensions and SIDE RAIL characteristics (e.g. to reduce the RISK of entrapment and falls (see also 201.7.2.2.106)).

201.7.9.2.5.103 * Angles and height of MEDICAL BED

The instructions for use shall identify the maximum angles, which can be achieved in NORMAL USE by each part of the MATTRESS SUPPORT PLATFORM with reference to horizontal. They shall

also identify the maximum and minimum heights from the floor which can be achieved by the MATTRESS SUPPORT PLATFORM in NORMAL USE. They shall also identify any emergency position(s) and the controls by which such position(s) are obtained.

201.7.9.2.5.104 * Maximum mass of MEDICAL BED

The instructions for use shall identify the maximum mass (in kg) of the MEDICAL BED. If the MEDICAL BED is intended to be disassembled into parts, the maximum mass (in kg) of any parts shall be stated in the instructions for use.

201.7.9.2.13 Maintenance

Replacement:

The instruction for use shall instruct the OPERATOR or RESPONSIBLE ORGANIZATION in sufficient detail concerning preventive inspection, maintenance and calibration to be performed by them, including the frequency of such maintenance.

The instruction for use shall provide information for the safe performance of such routine maintenance necessary to ensure the continued safe use of the MEDICAL BED.

Additionally, the instructions for use shall identify the parts on which preventive inspection and maintenance shall be performed by SERVICE PERSONNEL, including the periods to be applied and details about the actual performance of such maintenance.

For MEDICAL BEDS containing rechargeable batteries that are intended to be maintained by anyone other than SERVICE PERSONNEL, the instruction for use shall contain instructions to ensure adequate maintenance.

201.8 Protection against electrical HAZARDS from ME EQUIPMENT

Clause 8 of the general standard applies, except as follows:

201.8.11.3.2 Types

Replacement:

POWER SUPPLY CORDS shall be a minimum 2,5 m measured from the plug to the outside perimeter of the MEDICAL BED.

POWER SUPPLY CORDS and other external flexible mains cables and cords on MEDICAL BEDS shall be type HD22.10 H05-BQ-F [1]³⁾ or equivalent quality for mechanical robustness.

POWER SUPPLY CORDS shall be equipped with appropriate strain relief and bend protection.

POWER SUPPLY CORD sets shall have a moulded-on plug or other means to withstand the ingress of water during the cleaning PROCESS for which the MEDICAL BED is intended.

The MEDICAL BED shall be equipped with a means to keep the POWER SUPPLY CORD clear of any moving MEDICAL BED part or mechanism, when the MEDICAL BED is in use, transport or not in use to avoid a damage of the POWER SUPPLY CORD.

POWER SUPPLY CORDS shall be adequately protected against damages from contact with moving part(s) or from friction at sharp corners and edges within the MEDICAL BED.

³⁾ Figures in square brackets refer to the bibliography.

Compliance is checked by inspection.

201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS

Clause 9 of the general standard applies, except as follows:

201.9.1 * MECHANICAL HAZARDS of ME EQUIPMENT

Addition:

201.9.1.101 * Protection against PATIENT entrapment in non-moving parts

Any openings or areas (A_1 , A_2 , A_3 , A_4 , A_5 , A_6 , B, C and D) within the MEDICAL BED system and which are above the MATTRESS SUPPORT PLATFORM shall meet the dimensional and constructional requirements of Figures 201.107, 201.108 and Table 201.101. Where a RISK of PATIENT entrapment exists and is addressed in another way this shall be justified by the MANUFACTURER in the RISK MANAGEMENT FILE.

Compliance is checked before and after application of the SIDE RAIL strength and latch reliability tests (see 201.9.8.3.3.3).


Compliance is checked with the MATTRESS SUPPORT PLATFORM in the flat position unless otherwise noted as indicated in Table 201.101. The test shall be performed with the SIDE RAIL in all raised and locked positions.

All the tests are performed without the mattress except the test for Dimension D. Requirements involving the mattress are checked with the mattress(es) as specified by the MANUFACTURER. Requirements of Figures 201.107, 201.108 and Table 201.101 involving the mattress are excluded for SPECIALTY MATTRESSES.

A RISK ASSESSMENT shall be performed to evaluate:

- SPECIALTY MATTRESSES;
- MATTRESS OVERLAYS;
- ACCESSORIES;
- articulated MATTRESS SUPPORT PLATFORM positions.

Compliance is checked by the following tests and inspection of the RISK MANAGEMENT FILE.



IEC 2122/09


Only applies when the area C above is < 60 mm.

A_x represents the different areas A_1 , A_2 , A_3 , A_4 , A_5 and A_6

Key

- 1 Area of TOOL representing neck diameter (60 mm).
- 2 Area of TOOL representing chest breadth (318 mm).
- 3 Area of TOOL representing head breadth (120 mm).
- 4 MATTRESS SUPPORT PLATFORM
- 5 HEAD BOARD
- 6 FOOT BOARD
- 7 Mattress
- 8 SIDE RAIL

Figure 201.107 – Example of MEDICAL BED with segmented or split SIDE RAIL



Only applies when the area C is < 60 mm.

Key

- 1 Area of TOOL representing neck diameter (60 mm).
- 2 Area of TOOL representing chest breadth (318 mm).
- 3 Area of TOOL representing head breadth (120 mm).
- 4 MATTRESS SUPPORT PLATFORM
- 5 HEAD BOARD
- 6 FOOT BOARD
- 7 Mattress
- 8 SIDE RAIL

Figure 201.108 – Example of MEDICAL BED with single piece SIDE RAIL

Table 201.101 – Protection against PATIENT entrapment

Area	Description	Requirement/Compliance method
A ₁	Fully enclosed openings within a SIDE RAIL, HEAD or FOOT BOARD	Gap is specified to be less than 120 mm as defined by the following test. <i>Compliance is checked by the following test:</i> <i>Except for A₃ articulate the MEDICAL BED and find the largest opening. Insert 60 mm diameter part of cone tool (see Figure 201.103a, See Figures 201.107 and 201.108) through opening from inside of the MEDICAL BED system. Bring cone tool to bear on opening of interest. Exert 250 N force applied to 60 mm cylindrical end of cone tool in most disadvantageous direction.</i> <i>Pass/fail criterion: Opening shall not allow 120 mm diameter part of cone tool to enter and pass through.</i>
A ₂	Fully enclosed opening defined by the SIDE RAIL, its supports, and the MATTRESS SUPPORT PLATFORM	
A ₃	Partially enclosed opening defined by the HEAD BOARD, MATTRESS SUPPORT PLATFORM, and SIDE RAIL	
A ₄	Partially enclosed opening defined by the FOOT BOARD, MATTRESS SUPPORT PLATFORM, and SIDE RAIL (except where the gap between the SIDE RAIL and the FOOT BOARD > 318 mm)	
A ₅	Partially enclosed opening between segmented or split SIDE RAILS and the MATTRESS SUPPORT PLATFORM (except where the gap between the SIDE RAILS is > 318 mm)	
A ₆	Partially enclosed opening defined by the lowest point of a SIDE RAIL, the adjacent SIDE RAIL support, and MATTRESS SUPPORT PLATFORM, to the outside of the SIDE RAIL supports.	
A	Other opening(s) defined by ACCESSORIES (e.g., IV poles, fracture frames) and SIDE RAILS, HEAD/FOOT BOARDS, and/or MATTRESS SUPPORT PLATFORM. This is not in Figure 201.107 and Figure 201.108 as it depends on the position of ACCESSORIES and construction of the MEDICAL BED.	Gap < 60 mm. AND Angle between MATTRESS SUPPORT PLATFORM and the SIDE RAIL interface >60° over the entire range of mattress heights from the minimum recommended mattress height, minus 2 cm. to the maximum recommended mattress height, plus 2 cm. NOTE The RISK MANAGEMENT has to address the possibility of the use of a mattress not specified by the MANUFACTURER. RISK MANAGEMENT to address the entrapment condition of area B (as illustrated in Figure AA.13) should be performed, taking the following into consideration: i) The SIDE RAIL shape and geometry. ii) The distance between the lowest point of the SIDE RAIL and the MATTRESS SUPPORT PLATFORM. iii) The mattress material properties. iv) The mattress dimensions. v) The fit relationship between the SIDE RAIL, mattress and MATTRESS SUPPORT PLATFORM. <i>Compliance is checked by inspection of the RISK MANAGEMENT FILE. See also Annex CC for further information.</i>
B	Distance between the MATTRESS SUPPORT PLATFORM and the lowest point of the SIDE RAIL outside of the SIDE RAIL support. AND The angle between SIDE RAIL and the MATTRESS SUPPORT PLATFORM at the range of the mattress height defined by the MANUFACTURER ± 2 cm. NOTE ± 2 cm takes into account mattress compression and height of the neck above the mattress.	
C ₁	Gap between the HEAD BOARD and adjacent SIDE RAIL.	Gap between the HEAD BOARD and adjacent SIDE RAIL is required to be < 60 mm. <i>Compliance for a gap < 60 mm is checked by the following test:</i>

Area	Description	Requirement/Compliance method
		<i>The cylinder tool (see Figure 201.103b) shall be oriented parallel to floor, in the most disadvantageous angle in the horizontal plane above the gap. The 60 mm cylinder tool shall rest with the full weight on the gap where the cylinder tool intersects. Extra vertical force shall not be used. The cylinder tool shall not be used to pry apart parts of the MEDICAL BED.</i> <i>Pass/fail criterion:</i> <i>The 60 mm cylinder tool shall not slide into the opening. See Annex CC for further clarification.</i> <i>SEE ALSO ANNEX CC FOR FURTHER INFORMATION.</i>
C ₂	Gap between segmented or split SIDE RAILS with both SIDE RAILS raised.	Gap between segmented or split SIDE RAILS with both SIDE RAILS raised is required to be < 60 mm or > 318 mm. <i>Compliance for a gap < 60 mm is checked by the following test:</i> <i>The cylinder tool (see Figure 201.103b) shall be oriented parallel to floor, in the most disadvantageous angle in the horizontal plane above the gap. The 60 mm cylinder tool shall rest with the full weight on the gap where the cylinder tool intersects. Extra vertical force shall not be used. The cylinder tool shall not be used to pry apart parts of the MEDICAL BED.</i> <i>For MEDICAL BED with split SIDE RAILS, articulate the MATTRESS SUPPORT PLATFORM to identify the worst case opening between the SIDE RAILS and perform the test.</i> <i>Pass/fail criterion:</i> <i>The 60 mm cylinder tool shall not slide into the opening. See Annex CC for further clarification.</i> OR For a gap > 318 mm: The gap shall be > 318 mm for the entire vertical distance. <i>See Annex CC for further information.</i>

IECNORM.COM : Click to view the full PDF 60601-2-52:2009

Area	Description	Requirement/Compliance method
C ₃	Gap between SIDE RAIL and FOOTBOARD. Other openings(s) defined by ACCESSORIES (e.g. IV poles, fractures frames,.....) and SIDE RAILS, HEAD BOARD, FOOT BOARD, and / or MATTRESS SUPPORT PLATFORM.	Gap between SIDE RAIL and FOOTBOARD is required to be < 60 mm OR > 318 mm. <i>Compliance for a gap < 60 mm is checked by the following test:</i> <i>The cylinder tool (see Figure 201.103b) shall be oriented parallel to floor, in the most disadvantageous angle in the horizontal plane above the gap. The 60 mm cylinder tool shall rest with the full weight on gap where the cylinder intersects. Extra vertical force shall not be used. The cylinder tool shall not be used to pry apart parts of the MEDICAL BED.</i> <i>Pass/fail criterion:</i> <i>The 60 mm cylinder tool shall not slide into the opening. See Annex CC for further clarification.</i> OR For a gap > 318 mm: The gap shall be > 318 mm for the entire vertical distance. SEE ANNEX CC FOR FURTHER INFORMATION.
D	Region defined between the SIDE RAIL and the mattress.	<i>Compliance is checked by the following test:</i> <i>Push the mattress away from the SIDE RAIL being measured until the mattress retention system, or the opposing SIDE RAIL stops the mattress. Pull outward on the SIDE RAIL to remove any lateral play and during application of the force the cone tool (see Figure 201.103a) is placed with its longitudinal axis parallel to the SIDE RAIL, resting on the mattress in the horizontal gap between the SIDE RAIL and mattress. Turn the cone tool until the line on the face of the 120 mm diameter end is horizontal. Let the cone tool sink into the space by its own weight.</i> <i>If a mattress retention system, SIDE RAIL support or other structure keeps the cone tool from sinking in the gap, the cone tool shall be placed at a different location along the SIDE RAIL where there is no interference.</i> <i>Pass/fail criterion:</i> <i>The large end of the cone tool shall not sink below the mattress surface by 50 % or more of its 120 mm diameter.</i> NOTE The gap between HEADBOARD and MATTRESS SUPPORT PLATFORM is covered by the measurement of A (See designator A: other opening(s) defined by ACCESSORIES (e.g., IV poles, fracture frames) and SIDE RAILS, HEAD/FOOT BOARDS, and/or MATTRESS SUPPORT PLATFORM. This is not in Figure 201.107 or Figure 201.108 as it depends on the construction of the MEDICAL BED) Gap between HEADBOARD and top of mattress: The mattresses specified by the MANUFACTURER normally have no reasonable gaps between HEADBOARD and top end of mattresses for a possible head entrapment.

201.9.2.2 TRAPPING ZONE**201.9.2.2.1 General**

Addition:

The entire region in the UNDERCARRIAGE shall be considered in the RISK ANALYSIS with regard to trapping HAZARDS due to high/low motion.

Compliance is checked by inspection of the RISK MANAGEMENT FILE.


201.9.2.2.2 Gaps

Amendment:

The locations identified in Figure 201.109 and 201.110 shall be considered as TRAPPING ZONES for fingers.

Distances between moving parts shall always be either less than 8 mm or more than 25 mm (as dimensioned in Figure 201.109). The 200 mm cross-hatched area represents the areas of normal reach around the perimeter of the MATTRESS SUPPORT PLATFORM. The 200 mm distance can be measured taking into account any barrier preventing access to fingers (see Figure 201.110). The region within the APPLIED PART and above the MATTRESS SUPPORT PLATFORM shall be considered in the RISK ANALYSIS with regard to finger spacing between moving parts.

IECNORM.COM : Click to view the full PDF of IEC 60601-2-52:2009




IEC 2124/09

Key

- A Spacing between moving parts shall be less than 8 mm (non opening) or more than 25 mm.
- B Cross-hatched area representing entrapment reach zone for fingers 200 mm in from the outer edge.
- C HEAD BOARD
- D FOOT BOARD


Figure 201.109 – Allowable spacing for fingers in areas of normal reach around the perimeter of the MATTRESS SUPPORT PLATFORM

**Key**

- 1 Pinch point (TRAPPING ZONE)
- 2 MATTRESS SUPPORT PLATFORM

Figure 201.110 – Example using barriers for clearance measurement around the perimeter of the MATTRESS SUPPORT PLATFORM to mitigate PATIENT-finger entrapment


The locations identified in Figures 201.111 a) and 201.111 b) shall be considered as TRAPPING ZONES for feet.

**Key**

For area where $b \leq 130 \text{ mm}$;
shall always be $a \geq 120 \text{ mm}$

NOTE The dimension “a” is only measured from the floor. The dimension “b” is measured from the outer edge of the MEDICAL BED. Including any permanently fixed ACCESSORIES (e.g. SIDE RAIL), if applicable.

Figure 201.111 a) – Foot and toe clearance area between moving parts and the floor



Key

For area where b is between 130 mm to 180 mm;

shall always be $c \geq 50$ mm.

NOTE The dimension “ c ” is only measured from the floor. The dimension “ b ” is measured from the outer edge of the MEDICAL BED. Including any permanently fixed ACCESSORIES (e.g. SIDE RAIL), if applicable.

Figure 201.111 b) – Toe clearance area between moving parts and the floor

Figure 201.111 – Clearance areas

Measurements shall be done in the most disadvantageous condition for Figures 201.109, 201.110, 201.111a) and Figure 201.111b).

201.9.2.2.3 Safe distances

Addition:

For the reach of a hand the safe distance is 200 mm (see Figure 201.109 and Figure 201.110).

201.9.2.2.5 Continuous activation

Replacement of item b):

- b) all movement of the MEDICAL BED or its parts is possible only by the activation of control device(s) which initiate and maintain operation of the MEDICAL BED elements only as long as the manual control is actuated and where the manual control automatically returns to the 'Stop' or 'Off' position when released.

NOTE Manually and foot-operated movements are considered to comply with this clause, as long as mass and velocity allow adequate control of positioning without causing an unacceptable RISK.

Compliance is checked by inspection and functional tests.

201.9.2.3.1 * Unintended movement

Addition:

Except for emergency MEDICAL BED movements specified by the MANUFACTURER, the means to deactivate MEDICAL BED movement (such as those in PENDANT CONTROL and/or SIDE RAIL function controls intended to be activated to the PATIENT, OPERATOR, or visitor) shall be designed such that it cannot be accidentally re-activated by the PATIENT when the PATIENT is in the MEDICAL BED.

Compliance is checked by inspection.

Foot operated controls shall be designed to prevent accidental activation.

If the MEDICAL BED design cannot exclude trapping or crushing there shall be a means to deactivate any foot-operated controls used for any MEDICAL BED movement which is not manually operated. The control shall deactivate without OPERATOR action after use. Consideration shall be given to unintended activation by the PATIENT or other persons crawling under the MEDICAL BED or by objects used in close proximity.

If means for the foot-operated MEDICAL BED movement are provided, the means to deactivate foot-operated MEDICAL BED movement shall be located or designed such that a PATIENT cannot accidentally re-activate the functions, taking into account PATIENT mobility and medical supervision.

Compliance is checked by inspection.

201.9.4.2 Instability – overbalance

201.9.4.2.2 * Instability excluding transport

Addition:

When the following tests are performed, the MEDICAL BED shall not overbalance (tip over) with the height and length of the MATTRESS SUPPORT PLATFORM, castors, SIDE RAILS and other ACCESSORIES with their SAFE WORKING LOAD in their most adverse position of NORMAL USE, and the MEDICAL BED equipped with the lightest mattress as specified by the MANUFACTURER or a load representing the weight of the specified mattress that is centered on the MATTRESS SUPPORT PLATFORM.

Compliance is checked by the following tests conducted with the MATTRESS SUPPORT PLATFORM in the flat and horizontal position.

- *Lateral stability test:*

A load of 2 200 N is placed at the side edge of the MATTRESS SUPPORT PLATFORM and evenly distributed over an area 250 mm × 950 mm (see Figure 201.112).

If the maximum PATIENT load according to the MANUFACTURER exceeds 2 200 N, the maximum PATIENT load is to be used and is evenly distributed over an area 950 mm long and 250 mm wide (see Figure 201.112).

Perform the test at each corner of the MEDICAL BED.

- *Longitudinal stability test:*

aa) If the FOOT BOARDS are removable without the use of TOOLS:


Remove the FOOT BOARD; a load of 2 200 N for APPLICATION ENVIRONMENTS 1, 2, 3 and 5 and 1 850 N for APPLICATION ENVIRONMENT 4 is evenly distributed over an area of 250 mm across the full width of the MEDICAL BED (see Figure 201.113).

If the maximum PATIENT load according to the MANUFACTURER exceeds 2 200 N (or 1 850 N for APPLICATION ENVIRONMENT 4), the maximum PATIENT load is to be used.

bb) If the HEAD/FOOT BOARDS are permanently fixed or require the use of TOOLS to remove them:

Two loads each of 1 100 N for APPLICATION ENVIRONMENTS 1, 2, 3 and 5 and two loads each of 925 N for APPLICATION ENVIRONMENT 4 are at the same time evenly distributed over an area of 250 mm × 475 mm (Figure 201.114). Perform the test at both ends of the MEDICAL BED.

Dimensions in millimetres




Key

- 1 Load of 2 200 N evenly distributed or greater if maximum PATIENT load exceeds this.

Figure 201.112 – Lateral stability test along the side of the MEDICAL BED

Dimensions in millimetres




Key

- 1 Load of 2 200 N, (1 850 N for APPLICATION ENVIRONMENT 4), evenly distributed or greater if maximum PATIENT load exceeds this.

Figure 201.113 – Longitudinal stability test with removable FOOT BOARD

Dimensions in millimetres

**Key**

- 1 Two loads of 1 100 N, (two loads of 925 N for APPLICATION ENVIRONMENT 4), evenly distributed.

Figure 201.114 – Longitudinal stability test with fixed HEAD/FOOT BOARDS

201.9.4.2.3 Instability from horizontal and vertical forces

Replacement of item b):

- b) The MEDICAL BED shall not overbalance (tip over) due to sitting or stepping.

NOTE Requirements for PATIENT support surfaces are found in 201.9.8.3.

Compliance is checked by inspection and by the following test:

The MEDICAL BED is placed on a horizontal plane and a constant downward force of 1 100 N is applied at the point of maximum moment to any working surface, excluding the MATTRESS SUPPORT PLATFORM, offering an obvious foothold or sitting surface of a minimum 20 cm by 20 cm area, and at a height not exceeding 1 m from the floor. Prior to the test the MEDICAL BED is prepared as described in 201.9.4.2.2.

201.9.4.2.4.2 Force for propulsion

Addition:

For MEDICAL BEDS intended for PATIENT transport:

- perform the test with the MEDICAL BED with SAFE WORKING LOAD in place.

For MEDICAL BEDS not intended for PATIENT transport:

- perform the test with the MEDICAL BED without SAFE WORKING LOAD in place.

201.9.4.2.4.3 * Movement over a threshold

Replacement:

MOBILE MEDICAL BED intended to transport PATIENTS shall withstand the stresses caused by rough handling. The following threshold test shall be used.

This requirement does not apply to a MEDICAL BED specified by the MANUFACTURER only to be moved within the PATIENT room for cleaning or PATIENT access.

Compliance is checked by the following test:

The SIDE RAILS are raised and latched, with all other ACCESSORIES intended for NORMAL USE during transport attached to the MEDICAL BED and with the SAFE WORKING LOAD in place and the height in the worst case position.

The MEDICAL BED shall be moved at a speed of $0,8 \text{ m/s} \pm 0,1 \text{ m/s}$, or for motor-driven MEDICAL BEDS for transportation, the maximum speed shall be used, while all castors shall impact and pass over an obstruction which is fixed flat on the floor, with a rectangular cross-section, 20 mm high and 80 mm deep. The MEDICAL BED, with all castors, shall then be pulled back over the obstruction and back to the starting position of the test.

This is repeated 10 times.

Acceptance criteria for MOBILE and motor-driven MEDICAL BED:

At the end of the test, the MEDICAL BED, MEDICAL BEDS parts, ACCESSORIES shall present no loss of function, and without unlocking/unlatching of the SIDE RAILS, or physical deterioration, (e.g., deterioration of fixings, unlocking of ACCESSORIES if fixed on the MEDICAL BED (like SIDE RAILS), which can reduce the NORMAL USE or create a RISK like collapsing, or permanent deformation, modifying gap for entrapment or pinching etc.).

The MEDICAL BED shall go over the obstruction. The MEDICAL BED shall not overbalance (tip over). The MEDICAL BED or MEDICAL BED parts shall not present an unacceptable RISK. Unacceptable RISK is determined by inspection of the MEDICAL BED, its parts, and relevant information from the RISK MANAGEMENT FILE.

201.9.4.3.1 * Instability in transport

Replacement of item c):

- c) MOBILE MEDICAL BED shall not result in an unacceptable RISK due to unwanted horizontal movement.

Compliance is checked by the following test:

Prior to the test, the MEDICAL BED is prepared with the following in the most disadvantageous position of NORMAL USE:

- the height, articulation, and length of the MATTRESS SUPPORT PLATFORM;*
- castors;*
- SIDE RAILS;*
- ACCESSORIES with their SAFE WORKING LOAD in place, including in combination with other accessories (e.g. a platform with its SAFE WORKING LOAD; a lifting pole without its SAFE WORKING LOAD);*
- mattress (e.g. height and weight) as specified by the MANUFACTURER or a load representing the weight of the specified mattress that is uniformly distributed and centered on the MATTRESS SUPPORT PLATFORM.*

The MOBILE MEDICAL BED is placed with the SAFE WORKING LOAD in place, and the locking device (e.g. brakes) activated, on a plane covered with 2 mm to 4 mm thick vinyl flooring material and inclined at 6° from the horizontal plane on a concrete floor. Following initial elastic movement, initial creepage, and initial pivoting of castors, there shall be no movement of the MOBILE MEDICAL BED greater than 50 mm (in relation to the inclined plane). Any initial movement shall not result in an unacceptable RISK, taking into account the NORMAL USE of the MEDICAL BED.

201.9.4.3.2 Instability excluding transport

Replacement of item a):

See 201.9.4.3.1.

201.9.4.4 Grips and other handling devices

This subclause does not apply to APPLICATION ENVIRONMENTS 1, 2, 3, 5.

201.9.6.2.1 Audible acoustic energy

Addition:

The results of the measurement shall be given in the instruction for use

Amendment of compliance criteria:

Compliance is checked by measuring the maximum A-weighted sound pressure level at the minimum distances of PATIENT, OPERATOR and other persons from the source of acoustic energy (noise) in NORMAL USE, and if necessary, calculating the A-weighted sound pressure level produced by the ME EQUIPMENT in accordance with ISO 3746. The following conditions apply.

201.9.8 HAZARDS associated with support systems

201.9.8.1 General

Amendment:

Delete the first dashed item.

201.9.8.2 * TENSILE SAFETY FACTOR

This subclause does not apply (see 201.9.8.3.2).

201.9.8.3 Strength of PATIENT or OPERATOR support or suspension systems

201.9.8.3.1 General

Replacement:

MEDICAL BED parts serving for support or immobilization of the PATIENT shall be designed to minimize the RISK of physical injuries and of accidental loosening of fixings.

For APPLICATION ENVIRONMENTS 1 and 2, the SAFE WORKING LOAD of the MEDICAL BED shall be at least 2 000 N. It is considered to be the sum of the following minimum loads:

- 1 350 N, corresponding approximately to a mass of 135 kg for the PATIENT;
- 200 N, corresponding approximately to a mass of 20 kg for the mattress;
- 450 N, corresponding approximately to a mass of 45 kg for both the ACCESSORIES and the mass of the SAFE WORKING LOAD supported by those ACCESSORIES but excluding PATIENT weight.

For APPLICATION ENVIRONMENTS 3, 4 and 5, the SAFE WORKING LOAD of the MEDICAL BED shall be at least 1 700 N. It is considered to be the sum of the following minimum loads:

- 1 350 N, corresponding approximately to a mass of 135 kg for the PATIENT;
- 200 N, corresponding approximately to a mass of 20 kg for the mattress;
- 150 N, corresponding approximately to a mass of 15 kg for both the ACCESSORIES and the mass of the SAFE WORKING LOAD supported by those ACCESSORIES but excluding PATIENT weight.


The SAFE WORKING LOAD of a BED-LIFT shall be at least 2 200 N. It is considered to be the sum of the following minimum loads:

- 1 350 N, corresponding approximately to a mass of 135 kg for the PATIENT;
- 200 N, corresponding approximately to a mass of 20 kg for the mattress;
- 150 N, corresponding approximately to a mass of 15 kg for both the ACCESSORIES and the mass of the SAFE WORKING LOAD supported by those ACCESSORIES but excluding PATIENT weight;
- 500 N, corresponding approximately to a mass of 50 kg for those parts of the MEDICAL BED intended to be lifted by the BED-LIFT.

If the SAFE WORKING LOAD as specified by the MANUFACTURER is greater than 2 000 N for MEDICAL BEDS for APPLICATION ENVIRONMENTS 1 and 2, 1 700 N for MEDICAL BEDS FOR APPLICATION ENVIRONMENTS 3,4 and 5 or 2 200 N for BED-LIFTS, it shall be used as a basis for testing.

The SAFE WORKING LOAD shall be placed at the worst-case position permitted by the configuration or ACCESSORIES attachment on the support/suspension parts. The SAFE WORKING LOAD is distributed as shown in Figure 201.115.

IECNORM.COM : Click to view the full PDF of IEC 60601-2-52:2009



IEC 2131/09

Key

- 1 Back section
- 2 Seat section
- 3 Upper leg section
- 4 Lower leg section
- 5 SWL = SAFE WORKING LOAD

Figure 201.115 – Distribution of SAFE WORKING LOAD for tests

For a foot rest, the part of the SAFE WORKING LOAD representing the mass of PATIENTS is distributed over an area of 0,1 m² or whatever is available.

NOTE The position of the human body varies depending on the configuration of the support/suspension system and therefore the load acting on different sections will vary and should be taken into account. In analyzing loading forces and torques on support assemblies, the part of the SAFE WORKING LOAD represents the mass as defined.

The SAFE WORKING LOAD of the LIFTING POLE (included PATIENT handle) shall be at least 750 N

201.9.8.3.2 *Static forces due to loading from persons

Replacement:

The MEDICAL BED and BED-LIFT shall be capable of withstanding a uniformly distributed static load equal to two times the SAFE WORKING LOAD or 4 000 N, whichever is greater, in the most disadvantageous position on the MATTRESS SUPPORT PLATFORM in a horizontal position (see Figure 201.115).

When impairment by wear, corrosion, material fatigue or aging is expected, relevant supporting parts shall have a safety factor not less than 4 times the SAFE WORKING LOAD.

- a) For MEDICAL BEDS:

Compliance is checked by the following test:

Permanent deformation is acceptable only if the MEDICAL BED is in compliance with its intended function. The static load shall be applied for at least 1 min, unless material creep may be an issue, in which case the time shall be increased to at least 1 h.

- b) For BED-LIFTS:

Compliance is checked by the following test:

Mount a TEST BED BOARD to a BED-LIFT not supplied with a bed board.

Place the mattress, as specified by the MANUFACTURER, onto the bed board/TEST BED BOARD, in its flat position.

Apply a vertical load of two times the SAFE WORKING LOAD or 4 000 N, whichever is greater (excluding the mass of the mattress placed onto the MEDICAL BED or the mass of the TEST BED BOARD), equally distributed over the mattress.

Permanent deformation is acceptable only if the BED-LIFT is in compliance with its intended function. The static load shall be applied for at least 1 min, unless material creep may be an issue in which case the time shall be increased to at least 1 h.

All ACCESSORIES (including those not supporting PATIENT weight) shall be designed to support a load that is at least two times the SAFE WORKING LOAD specified for the ACCESSORY. This load is applied in the most disadvantageous direction and position on the ACCESSORY.

Compliance is checked by the following test:

With the ACCESSORY (others than LIFTING POLES) in its worst case NORMAL USE position, attach a static load equal to two times its SAFE WORKING LOAD for at least 1 min, unless material creep may be an issue in which case the time shall be increased to at least 1 h.

There shall be no HAZARD or loss of function.

The fastenings of LIFTING POLES shall still function normally and present no HAZARDS after the following tests. For the LIFTING POLE, permanent deformation is acceptable in the first test but a sudden movement of the LIFTING POLE shall be considered a HAZARD:

- 1) *Compliance is checked by the following test:*

Position the LIFTING POLE to the MEDICAL BED in its most adverse position intended for use.

Apply a downward load of two times the SAFE WORKING LOAD of LIFTING POLES (at least 1 500 N) to the outermost suspension point for the handle for at least 1 min, unless material creep may be an issue in which case the time shall be increased to at least 1 h.

Examine the LIFTING POLE and its fastenings during and after application of the load.

2) *Compliance is checked by the following test:*

Position the LIFTING POLE to the MEDICAL BED in its most adverse position intended for use. Fixate the MEDICAL BED during the test if necessary due to instability.

Apply perpendicular to the MEDICAL BED side a horizontal force of 350 N to the outermost suspension point for the handle. Examine the LIFTING POLE and its fastenings during and after application of the load.

201.9.8.3.3 * Dynamic forces due to loading from persons

Replacement:

201.9.8.3.3.1 General


Where dynamic forces (due to sitting down, standing up, PATIENT handling PROCESS or the like) can be exerted on equipment parts intended to support or suspend a PATIENT in NORMAL USE, they shall not result in an unacceptable RISK.

Durability shall be considered with regards to the most disadvantageous position of the MEDICAL BED parts intended to support or suspend a PATIENT in NORMAL USE.

Compliance is checked by the following tests.

Condition of test:

Place the worst-case mattress specified by MANUFACTURER onto the MEDICAL BED in its flat position. If height-adjustable, adjust the height to the worse case position. Utilize loading pad (Figure 201.104) at position A in Figure 201.116, at whichever side is regarded to be the weaker.



Key

- 1 Head-end of MEDICAL BED
- 2 Foot-end of MEDICAL BED

Figure 201.116 – Position of loading pad
(see Figure 201.104)

Test method:

Apply the loading pad (Figure 201.104) 10 000 times at the position A shown in Figure 201.116, with a load of 1 350 N or the maximum PATIENT load, whichever is greater.

Acceptance criteria/Compliance method:

When tested in accordance with the above, the MEDICAL BED shall still function normally and present no HAZARDS after removal of the loading pad.

Remove the mattress and perform the following:

- evaluate entrapment HAZARDS according to 201.9.1.101;
- evaluate TRAPPING ZONES according to 201.9.2.2;
- conduct test according to 201.9.8.3.2.

201.9.8.3.3.2 * Dynamic testing of the height adjustment mechanism

The height adjustment of MEDICAL BED or BED-LIFT shall function normally and present no unacceptable RISK after 3 000 cycles in NORMAL USE.

Compliance is checked by the following test:

Place the MEDICAL BED in its flat position. If the BED-LIFT is separate from a MATTRESS SUPPORT PLATFORM, mount a TEST BED BOARD to the BED-LIFT. Apply the SAFE WORKING LOAD distributed as indicated in 201.9.8.3.1 (Figure 201.115) on the MATTRESS SUPPORT PLATFORM. For BED-LIFTS apply the SAFE WORKING LOAD minus the weight of the TEST BED BOARD distributed as indicated in 201.9.8.3.1 (Figure 201.115) on the MATTRESS SUPPORT PLATFORM.

The MEDICAL BED, or BED-LIFT is raised and lowered completely 3 000 times in accordance with the procedure stated in the instructions for use. After the test remove the SAFE WORKING LOAD.

201.9.8.3.3.3 * SIDE RAIL strength and latch reliability

SIDE RAIL latches/locks shall remain secure when subjected to the forces of NORMAL USE.

Compliance is checked by the following test. Forces are applied at the worst-case accessible location for latching/locking:


Cycle the SIDE RAIL mechanism (from the latched, upper position, to the unlatched, lowered position, and back to the latched, upper position) for 30 000 cycles. A force A or B or C or D or E or F (as specified in Figure 201.117) shall then be applied to the worst-case position for locking/latching of the SIDE RAIL in the direction of unlatching/unlocking. The SIDE RAIL shall not become unlatched/unlocked or create any other unacceptable RISK.

SIDE RAILS shall be designed to withstand the forces applied during reasonably foreseeable misuse over the product life cycle without creating an unacceptable RISK.

Compliance is checked by the following test with the SIDE RAIL in upper position:

- a) Lateral force cycling test. Exert a force of 100 N that is perpendicular to the SIDE RAIL at the top middle section of the SIDE RAIL in the direction indicated by E or F in Figure 201.117. Reverse direction. Repeat for 3 000 cycles.*
- b) Longitudinal force cycling test. Exert a force of 100 N on the SIDE RAIL in the lengthwise direction of the SIDE RAIL as indicated by C or D in Figure 201.117. Reverse direction. Repeat for 3 000 cycles.*
- c) Vertical force cycling test. Exert a force of 100 N on the uppermost part of the SIDE RAIL in the vertical direction of the SIDE RAIL as indicated by B in Figure 201.117. Repeat for 3 000 cycles. No test is required in direction A.*
- d) Upon completion of a), b) and c) above, apply a static load in the directions of Figure 201.117 in worst case positions. The SIDE RAIL shall not become unlatched/unlocked or create any other unacceptable RISK.*

NOTE The MANUFACTURER shall take into account when these forces can be applied in the lowered position.



Direction of force	Force
A	500 N
B	750 N
C	500 N
D	500 N
E	500 N
F	500 N

Figure 201.117 – Application of forces for test of SIDE RAIL

201.9.8.3.3.4 Dynamic testing of the LIFTING POLE

The LIFTING POLE and its fastenings shall still function normally and present no HAZARDS after the following test. A sudden movement of the LIFTING POLE or its handle shall be considered a HAZARD.

Compliance is checked by the following test:

Position the LIFTING POLE to the MEDICAL BED in its most adverse position intended for use.

Apply a vertical downward load of the SAFE WORKING LOAD to the LIFTING POLE (at least 750 N) 1 000 times onto the handle of the LIFTING POLE. Examine the LIFTING POLE and its fastenings during and after application of the force and record deflection and deformation.

The deflection of the LIFTING POLE shall not be greater than 100 mm during application of the SAFE WORKING LOAD and the permanent deformation not more than 20 mm after the durability test measured in relation to the MATTRESS SUPPORT PLATFORM.

201.9.8.4 Systems with MECHANICAL PROTECTIVE DEVICES

201.9.8.4.1 General

This subclause does not apply.

201.9.8.5 Systems without MECHANICAL PROTECTIVE DEVICES

This subclause does not apply.


Additional subclause:

201.9.101 *Protection against inadvertent PATIENT falls

SIDE RAILS shall be designed with minimum height requirements as indicated by G in Figure 201.118 and Table 201.102.

Where a SPECIALTY MATTRESS or MATTRESS OVERLAY is used and the SIDE RAIL does not meet G as indicated in Table 201.102, a RISK ASSESSMENT shall be performed to assure equivalent safety.

Compliance is checked by measurement of G and inspection of the RISK MANAGEMENT FILE.



Key

- 1 MATTRESS SUPPORT PLATFORM
- 2 FOOT BOARD
- 3 HEAD BOARD
- 4 Mattress
- G Height of the top edge of the SIDE RAIL above the mattress

IEC 2134/09

Figure 201.118 – Height of SIDE RAIL

Table 201.102 – Protection against inadvertent PATIENT falls

Designator	Description	Requirement/Compliance method
G	Height of the top edge of the SIDE RAIL above the mattress without compression in at least 50 % of the length of the MATTRESS SUPPORT PLATFORM. Other raised and locked positions not meeting the height of ≥ 220 mm or length at least 50% for SIDE RAILS and specified for purposes other than SIDE RAIL application (e.g. mobilization bar or grab handles) shall be covered in the MANUFACTURER'S RISK MANAGEMENT FILE. When SIDE RAILS cover less than the full length of the MATTRESS SUPPORT PLATFORM, those parts that are above 220 mm shall be placed strategically to help prevent accidental falls from the MEDICAL BED. All Dimensions C of Table 201.101 shall be respected. NOTE Placement of the 220 mm portions of the rail may need to be placed next to those parts of the body that are highest (e.g. shoulder for a PATIENT lying on their side) and the parts which could lead to slipping out of the MEDICAL BED (legs) in order to prevent accidental rollout or sliding out. See Figure 201.116, for guidance on distribution of body mass.	≥ 220 mm <i>Compliance is checked by measurement</i>

201.10 Protection against unwanted and excessive radiation HAZARDS

Clause 10 of the general standard applies.

201.11 Protection against excessive temperatures and other HAZARDS

Clause 11 of the general standard applies, except as follows:

201.11.1.1 Maximum temperature during NORMAL USE

Amendment:

Replacement of Table 24:

**Table 24 – Allowable maximum temperatures for skin contact
with MEDICAL BED APPLIED PARTS**

APPLIED PARTS OF MEDICAL BEDS		Maximum temperature ^{a b} °C		
		Metal and liquids	Glass, porcelain, vitreous material	Moulded material, plastic, rubber, wood
APPLIED PART having contact with the PATIENT for a time “t”	$t < 1 \text{ min}$	51	56	60
	$1 \text{ min} \leq t < 10 \text{ min}$	48	48	48
	$10 \text{ min} \leq t$	41	41	41
^a These temperature limit values are applicable for the healthy skin of adults, but in the limitation for temperature for duration more than 10 min also DISABLED PERSONS with sensible skin or no sensation have been considered. They are not applicable when large areas of the skin (10 % of total body surface or more) can be in contact with a hot surface. They are not applicable in the case of skin contact with over 10 % of the head surface. Where this is the case, appropriate limits shall be determined and documented in the RISK MANAGEMENT FILE. ^b Where it is necessary for APPLIED PARTS to exceed the temperature limits of Table 201.103 in order to provide clinical benefit, the RISK MANAGEMENT FILE shall contain documentation showing that the resulting benefit exceeds any associated increase in RISK.				

201.11.6.5 Ingress of water or particulate matter into ME EQUIPMENT and ME SYSTEMS

Addition:

201.11.6.5.101 Ingress of water

ENCLOSURES of MEDICAL BED shall provide a minimum protection degree of IPX4 in accordance with the classification of IEC 60529.

Compliance is checked by the tests of IEC 60529 with the MEDICAL BED placed in the least favourable position of NORMAL USE (as defined in the instructions for use) and by inspection:

All parts that can be removed without a TOOL shall be removed (detached/opened) before the test. The test shall be executed without a mattress. After these procedures, the MEDICAL BED shall show no signs of bridging of insulation (or electrical components) that could result in a HAZARD in NORMAL CONDITION or in combination with a SINGLE FAULT CONDITION (based on a visual inspection) followed by the appropriate dielectric strength and LEAKAGE CURRENT tests.

The MEDICAL BED shall comply with its INTENDED USE.

201.11.6.6 Cleaning and disinfection of ME EQUIPMENT and ME SYSTEMS

Addition:

If the MEDICAL BED is intended for use with equipment for cleaning and disinfection, the MANUFACTURER shall specify the cleaning and/or disinfection PROCESS.

201.11.6.6.101 Machine washable MEDICAL BED

MEDICAL BED or BED-LIFT specified by the MANUFACTURER to be machine washable by an automatic washing system shall function normally after the test.

Variations to the test procedure regarding test-cycle, temperature, time and cleaning fluids shall be covered in the RISK MANAGEMENT FILE of the MANUFACTURER.

The present test method represents a basic procedure for disinfection of a MEDICAL BED in a washing machine.

Compliance is checked by the following test:

- a) *Parts and access covers which can be detached/opened without the use of a TOOL shall be detached/opened;*
 - *temperature preconditioning treatment of 10 days at 65 °C $\pm$ 2 °C or at maximum value of the rated storage temperature, if higher, is carried out;*
 - *the MEDICAL BED shall then be kept at room temperature for not less than 16 h;*
- b) *50 test cycles according to the procedure the MANUFACTURER described in the instructions for use or consisting of:*
 - *2 min wash with 70 °C water at a ph-value 5-8, 0,5 % cleaning and disinfectant solution as specified by the MANUFACTURER;*
 - *20 s rinse with 85 °C water at a ph-value 5-8, and 0,2 % clear rinsing solution according to the data of the MANUFACTURER.*
 - *10 min cooling at 20° C ambient temperature.*
 - *Acceptance criteria:*
 - 1) *Immediately after the test cycles, the MEDICAL BED is connected to mains. No unintentional movements shall arise.*
 - 2) *The MEDICAL BED shall function as specified by the INTENDED USE at the following intervals:*
 - *immediately after the test cycles;*
 - *5 min ($\pm$ 1 min) after the test cycles;*
 - *60 min ($\pm$ 5 min) after the test cycles;*
 - *24 h ($\pm$ 30 min) after the test cycles;*
 - 3) *Perform dielectric strength and LEAKAGE CURRENT tests according to 8.8.3 and 8.7 at the following intervals:*
 - *immediately after the test cycles*
 - *24 h ($\pm$ 30 min) after the test cycles*
 - 4) *Perform a visual inspection for ingress of water that may result in an unacceptable RISK (i.e., shorting of isolation barriers and violation of creepage distances).*

NOTE For some MEDICAL BED not all acceptance criteria apply (i.e. for manual MEDICAL BEDS, electrical acceptance criteria would not apply).

201.11.8 * Interruption of the power supply/SUPPLY MAINS to ME EQUIPMENT

Addition:

In an emergency situation when MAINS VOLTAGE has been interrupted, ESSENTIAL PERFORMANCE functions defined by the MANUFACTURER shall be able to be operated by other means and back section lowering or any other emergency position (specified by the MANUFACTURER) shall be achievable within 30 s.

Compliance is checked by functional test in the most adverse conditions.

201.12 Accuracy of controls and instruments and protection against hazardous outputs

Clause 12 of the general standard applies, except as follows:

201.12.2 USABILITY

Addition:

USABILITY analysis shall take into account MEDICAL BED design aspects such as the MATTRESS SUPPORT PLATFORM height, with regard to both OPERATOR and PATIENT, as relevant to the APPLICATION ENVIRONMENT.

NOTE Access to PATIENT for the purposes of medical treatment should not be delayed significantly by the SIDE RAILS and their respective control systems.

201.13 HAZARDOUS SITUATIONS and fault conditions

Clause 13 of the general standard applies, except as follows:

201.13.1.4 * Specific MECHANICAL HAZARDS

Addition:

Protection shall be provided in the motion controls of electrical MEDICAL BEDS such that a SINGLE FAULT CONDITION does not cause a movement without human intervention.

NOTE A MOTION LOCKOUT CONTROL may be used as a means of compliance. A MOTION LOCKOUT CONTROL may be activated automatically or as an OPERATOR option. (See also 201.7.4.2.).

If an auxiliary subsystem that deactivates motion controls (MOTION LOCKOUT CONTROL) is provided, it shall remain functional under SINGLE FAULT CONDITION.

For switches and other components that control motion in electrical MEDICAL BEDS, COMPONENTS WITH HIGH-INTEGRITY CHARACTERISTICS are required if their failure can cause a HAZARD.

Hydraulic, pneumatic, and mechanical subsystems of MEDICAL BEDS shall not cause unintended movement of the MATTRESS SUPPORT PLATFORM in SINGLE FAULT CONDITION, where such movement could cause a HAZARD.

Protection of SINGLE FAULT CONDITION for pneumatic, hydraulic, or mechanical MEDICAL BEDS, or for pneumatic, hydraulic, or mechanical subsystems of electrical MEDICAL BEDS, shall be met by compliance with 9.7.2 of the general standard, and other applicable clauses of the general standard and this standard pertaining to safety factors.

201.13.2.2 Electrical SINGLE FAULT CONDITION

Addition:

201.13.2.2.101 Power-driven MEDICAL BED for transportation

Power-driven MEDICAL BEDS for transportation shall have a means to allow transport by the OPERATOR under SINGLE FAULT CONDITION of the transport system.

NOTE A power-driven medical bed incorporates a propulsion system.

201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)

Clause 14 of the general standard applies.

201.15 Construction of ME EQUIPMENT

Clause 15 of the general standard applies, except as follows:

201.15.3 Mechanical strength

201.15.3.4.1 HAND-HELD ME EQUIPMENT

Replacement:

PENDANT CONTROL shall not result in an unacceptable RISK as a result of a free fall.

Compliance is checked by the following test:

The PENDANT CONTROL has to fall 1 000 times as specified in IEC 60068-2-31:2008.

After the test the PENDANT CONTROL is checked by inspection, functional test and relevant electrical safety tests.

201.15.3.5 * Rough handling test

Replacement of item c):

c) Door frame shock

The MEDICAL BED is moved three times in its normal direction of travel at a speed of 0,4 m/s $\pm$ 0,1 m/s with the MATTRESS SUPPORT PLATFORM in the worst case position, unless transport position is specified by the MANUFACTURER, against a hardwood vertical obstacle having a width and thickness of 40 mm affixed to a vertical rigid support (e.g. concrete). The height of the vertical obstacle shall be at the same level as the MEDICAL BED contact point(s). The direction of movement is perpendicular to the face of the obstacle.

Addition at the end of the paragraph:

This requirement does not apply to MEDICAL BED specified by the MANUFACTURER only to be moved within the PATIENT room for cleaning or PATIENT access.

201.15.4 ME EQUIPMENT components and general assembly

201.15.4.4 * Indicators

Amendment:

Delete the first two paragraphs.

201.15.4.6.2 * Limitation of movement

Addition:

The angle (designated " γ " in Figure 201.119) between the back section and leg/upper leg section for various configurations of the MATTRESS SUPPORT PLATFORM shall always be greater than 90° under NORMAL CONDITIONS.

In APPLICATION ENVIRONMENT 4, the angle α of the back section in relation to the horizontal shall always be greater than or equal to 0°.

If a MEDICAL BED in APPLICATION ENVIRONMENTS 1, 2, 3 or 5 is provided with Trendelenburg-position, the entire MATTRESS SUPPORT PLATFORM shall be able to achieve a minimum of 12° so that the PATIENT'S head is lower than the circulatory centre point of the body.

Compliance is checked by operating the controls to achieve the maximum angle γ and then measuring this angle.




Figure 201.119a – Angle γ between the back section and the leg section of the MATTRESS SUPPORT PLATFORM




Figure 201.119b – Angle γ between the back section and the upper leg section of the MATTRESS SUPPORT PLATFORM





Figure 201.119c – Angle γ between the angled back section and upper leg section of the MATTRESS SUPPORT PLATFORM



Key

- 1 Back section
- 2 Leg section
- 3 Upper leg section
- 4 Seat section
- 5 Lower leg section

Figure 201.119d – Angle γ between the angled back section and the leg/upper leg section of the MATTRESS SUPPORT PLATFORM

Figure 201.119 – Configurations of the MATTRESS SUPPORT PLATFORM

201.15.4.7.1 Mechanical strength

Amendment to item b):

This requirement does not apply if the foot-operated control device is so located that the operator's entire weight cannot be placed on it.

Compliance is checked by application to the foot-operated control device, in its position of NORMAL USE, of an actuating force of 1 350 N for 1 min. The force is applied over an area of 30 mm diameter. There shall be no damage to the device resulting in an unacceptable RISK.

Addition to item b):

Where the entire weight cannot be placed on the foot-operated control device, the RISK MANAGEMENT FILE shall specify the compliance criteria.

Compliance is checked by inspection of the RISK MANAGEMENT FILE.

201.15.4.7.3 Entry of liquids

Amendment:

Foot-operated controls shall be at least IPX4.

Additional subclauses:

201.15.4.101 HEAD/FOOT BOARD assembly

The design of the MEDICAL BED shall allow immediate and unimpeded access to the PATIENT from the head end of the MEDICAL BED in emergency situations. This applies for APPLICATION ENVIRONMENTS 1 and 2 only.

NOTE One means of accomplishing this is by a HEAD BOARD, which is removable without the use of a TOOL by one person.

Compliance is checked by the following test:

A trained OPERATOR shall be able to remove the HEAD BOARD within 15 s.

201.15.4.102 Mattress retention

The MEDICAL BED shall be designed to retain the mattress in position during NORMAL USE to prevent sliding off MATTRESS SUPPORT PLATFORM. A RISK ASSESSMENT shall be performed to evaluate the maximum movement of the retained mattress according to Area "D" of Table 201.101.

Compliance is checked by inspection of the RISK MANAGEMENT FILE.

201.16 ME SYSTEMS

Clause 16 of the general standard applies.

201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS

Clause 17 of the general standard applies.

Annexes

The annexes of the general standard apply.

IECNORM.COM : Click to view the full PDF of IEC 60601-2-52:2009

Annex AA (informative)

Particular guidance and rationale

AA.1 General guidance

In 1996, the IEC published the first edition of the particular standard for electrically operated hospital beds, IEC 60601-2-38. The publication was in response to the demand in the field for a universal standard addressing HAZARDS specific to the safety of the hospital bed. Used in conjunction with a MANUFACTURER'S RISK ASSESSMENT, the standard was felt to be the current thinking on establishing a BASIC SAFETY benchmark for industry.

An amendment of IEC 60601-2-38 issued in 1999 recognized the need to mitigate against a RISK of PATIENT entrapment in the SIDE RAILS, again combined with the use of the MANUFACTURER'S RISK ASSESSMENT. Although this improved the particular standard, it still was centered upon electrically operated hospital beds, and failed to take into account manually operated hospital beds and products in other medical environments.

In 2000, the EN 1970 standard (Adjustable beds for DISABLED PERSONS – Requirements and test methods) was published, which addressed beds used by DISABLED PERSONS to alleviate or compensate for a disability or handicap. This standard offered a broadened scope in conjunction with IEC 60601-2-38, but after the edition of Amendment 1 to IEC 60601-2-38, the opportunity presented itself to combine the two standards to a common, international standard.

As work began on the integration, the IEC adjusted its stance on BASIC SAFETY and ESSENTIAL PERFORMANCE, integrating them into the third edition of IEC 60601-1. It therefore became necessary to align the new standard with the third edition. The particular standard was given a new number, IEC 60601-2-52, and work began on alignment to third edition.

This particular standard, therefore, is the realization of much work in alignment, and scope adjustment between IEC 60601-2-38, EN 1970, and the third edition of IEC 60601-1. It represents the current thinking in BASIC SAFETY and ESSENTIAL PERFORMANCE of the MEDICAL BED as used to alleviate illness of PATIENTS and disability of DISABLED PERSONS. This is the effort of a joint working group of the IEC and the ISO.

AA.2 Rationale for particular clauses and subclauses

The following are rationales for specific clauses and subclauses in this particular standard, with clause and subclause numbers parallel to those in the body of the document.

Subclause 201.1.1 – Scope

The working group recognizes that the definitions of the terms “adult” and “child” are based on physical characteristics, which vary from one country to another. If the highest level of safety is to be achieved for PATIENTS and OPERATORS, it is necessary to rely on caregivers to use their professional judgement to differentiate the needs of children from those of adults in relation to the ME EQUIPMENT, taking into consideration not only the physical, psychological and medical needs of the individual, but also the PATIENT's preference. The dimensional requirements of this particular standard are based on anthropometric data based on PATIENTS ranging in physical size from a 146 cm tall female to a 185 cm tall male. For BEDS intended for use with PATIENTS outside this range, all dimensional characteristics in this particular standard should be adjusted accordingly.

Definition 201.3.131 – TRAPPING ZONE

The term “entrapment” describes an event in which a PATIENT is caught, trapped, or entangled in spaces in or about the SIDE RAIL, mattress, or MEDICAL BED frame.

Definitions 201.3.201 – 201.3.205 – APPLICATION ENVIRONMENTS 1, 2, 3, 4 and 5

The working group discussed many different APPLICATION ENVIRONMENTS in evaluating this particular standard. For example, a maternity bed designed to aid in the birthing process is primarily used with PATIENTS below the age of 50. These PATIENTS generally do not fit into a high entrapment RISK category, thus entrapment guidelines may not be considered, unless the product is designed to be used in various APPLICATION ENVIRONMENTS in NORMAL USE, or there is reasonable foreseeable misuse.

Definition 201.3.212 – MEDICAL BED

This particular standard merges two pre-existing standards: IEC 60601-2-38 (*Particular requirements for the safety of electrically-operated hospital beds*), and EN 1970 (*Adjustable beds for DISABLED PERSONS – requirements and test methods*). Since different products exist within different health-care markets, requirements are divided into different APPLICATION ENVIRONMENTS. Certain products may exist across APPLICATION ENVIRONMENTS, or different APPLICATION ENVIRONMENTS may develop in the future that are not specified in this particular standard. The environment of use has to be addressed through the MANUFACTURER'S RISK MANAGEMENT PROCESS.

Subclause 201.5. 101 – Entrapment test tools

The dimensions and weight of the cone tool and cylinder tool derived from the following observation:

a) Cone tool:

For a 95th percentile male head this weight is 5,13 Kg (see *Air Force Research Laboratory Design and Development of Anthropometrically Correct Head Forms for Joint Strike Fighter Ejection Seat Testing – February 2005*).

b) Cylinder tool:

60 mm is a conservative measure of 83 mm neck 5th percentile female.


Subclause 201.6.2 – Protection against electrical shock

In some countries a protective earthing system may not be available or reliable in the home environment (APPLICATION ENVIRONMENT 4).

Subclause 201.7.2.2.101 – Marking of maximum PATIENT weight and SAFE WORKING LOAD

It is important that the RESPONSIBLE ORGANIZATION and/or OPERATOR clearly understand the maximum PATIENT weight and SAFE WORKING LOAD of the MEDICAL BED, and of any ACCESSORY that is intended to support potentially hazardous masses, so that the MEDICAL BED and its ACCESSORIES are not used in ways, which could create a HAZARD.


Subclause 201.7.2.2.105 – Marking of replacement mattresses



IEC 2139/09

Figure AA.1 – Marking to select recommended mattresses specified by the MANUFACTURER

Subclause 201.7.2.2.106 – Marking of detachable SIDE RAILS



IEC 2140/09

Figure AA.2 – Marking for detachable SIDE RAILS specified by the MANUFACTURER

Subclause 201.7.9.2.5.103 – Angles and height of MEDICAL BED

When determining the adjustable height range of the MEDICAL BED, consideration should be given to regional anthropometrics for both caregiver and/or OPERATOR and/or PATIENT.

Attention should be paid to:

- PATIENT falling from the MEDICAL BED, where a low height minimizes injuries;
- ease of PATIENT entry and egress;
- transfers to and from a wheelchair;

- caregivers' and/or OPERATORS' safe working height in order to minimize back problems.

Subclause 201.7.9.2.5.104 – Maximum mass of MEDICAL BED

Where a MEDICAL BED may be disassembled for transport by hand, knowledge of the mass of the main parts may be useful in avoiding strain.

Subclause 201.9.1 – MECHANICAL HAZARDS of ME EQUIPMENT

General rationale for 201.9.1.101, and 201.9.101: There are various recognized RISKS that have been accepted on the basis of RISK/benefit or RISK/RISK tradeoffs. SPECIALTY MATTRESSES are excluded from testing because the clinical benefits derived from such mattresses can outweigh the RISKS involved with their use, and it might not be possible to satisfy the test requirements without compromising the clinical benefit of the SPECIALTY MATTRESS. However this standard requires a RISK ASSESSMENT to be performed.

The requirements are only applied with the MATTRESS SUPPORT PLATFORM in the flat position unless otherwise indicated. The reason is that there are various clinical benefits associated with the ability of MEDICAL BEDS to perform various articulations. It is not apparent how the functionality of MEDICAL BEDS, particularly the most advanced, most clinically beneficial beds, could be preserved if the requirements were extended to articulated positions. MANUFACTURERS are encouraged to understand the RISK involved and to strive to minimize this RISK by analysis and novel approaches to mitigation, while maintaining clinical utility of the bed.

The dimensional requirements form a baseline, but they are not relevant to every bed application, and they may occasionally interfere with intended usage without mitigating RISK. The MANUFACTURER is responsible for justifying deviations for specialty beds or surfaces in the RISK ANALYSIS.

IECNORM.COM : Click to view the full PDF of IEC 60601-2-52:2009

Subclause 201.9.1.101 – Protection against PATIENT entrapment in non-moving parts

The dimensional requirements listed in Figures 201.107 and 201.108 and in Table 201.101 are intended to restrict openings in and around the MEDICAL BED system so that parts of the anatomy cannot enter or can pass through easily. Each requirement is based on commonly available anthropometric data. PATIENTS most vulnerable to entrapment are those who have cognitive disability or have uncontrolled body movement. These PATIENTS most often are frail, elderly or confused. Few adverse event reports indicate the angular position of the MATTRESS SUPPORT PLATFORM in which entrapment occurs, or what position the SIDE RAILS are in. Therefore this particular standard requires that entrapment be evaluated by the MANUFACTURER'S RISK ASSESSMENT in flat and articulated bed positions, with the SIDE RAILS fully raised and in intermediate positions.

Three key body parts at RISK for life-threatening entrapment in MEDICAL BED systems are the head, neck, and chest.

- Head

The 120 mm dimension is based on the 5th percentile Sri Lanka female head breadth, the distance across the face from ear to ear. Gaps < 120 mm were chosen as this dimension represents a conservative measurement to reduce entrapment. A dimension of 120 mm includes all 5th percentile smallest female head breadth references and most 1st percentile international references.

In case of a permanently FIXED mattress specifically dedicated to the MEDICAL BED and SIDE RAIL and if sliding of the mattress on the SUPPORT MATTRESS PLATFORM is not possible, the test is performed by inserting the 60 mm cylindrical end of the cone tool with 120 mm cone between the permanently FIXED mattress and the SIDE RAIL (A₂ and A₃ areas).

- Neck

The 60 mm dimension is based on the 5th percentile width of a small adult female neck.

It is noted that people of small stature have a 1st percentile female neck diameter of 79 mm (5th percentile = 83 mm.). Several factors, such as neck compressibility, loss of muscle mass in the neck as PATIENTS age, and the asymmetrical shape of the neck, all support a conservative measure of less than 60 mm.

To reduce the RISK of neck entrapment, gaps in the MEDICAL BED system should be small enough to prevent passage of a small neck. Additionally to reduce neck entrapment with regards to wedging of the neck, angles for V-shaped openings (Reference EN 13451-1) should be greater than 60°. The 60° angle reflects an understanding that at some point, a narrow angle creates a v-shape space or geometry that could be conducive to entrapment. The 60° angle was chosen by reviewing the TOOLS to assess entrapment RISK for swimming pool equipment (EN 13451-1, *Swimming pool equipment – Part 1: General safety requirements and test Methods*), and a corresponding mathematical analysis which shows that when a cylinder-shaped weight is supported by two offsetting force vectors, each of the perpendicular forces becomes equal to the weight of the cylinder when the angles relative to the vertical are 60°. An assumption is made that 60° may be considered to be an inflection point for angles where wedging may become a prominent consideration (see Figures AA.5 and AA.6).

See also Annex CC Particular guidance for assessing v-shaped opening

- Chest

The 318 mm is based on the 95th percentile male chest depth. There have been reports of chest entrapment in MEDICAL BEDS in the space between split SIDE RAILS. To avoid chest entrapments, the spaces should be wide enough (greater than 318 mm) to allow a large chest to slip through without being entrapped.


Specification of cone tool and cylinder tool used to evaluate entrapment see Figure 201.103a and Figure 201.103b.

- Rationale for measurement of B

The measurement of area B includes two separate measurements: a measurement of the gap beneath the SIDE RAIL (see Figure AA.5), and a measurement of the interface angle at the mattress height, plus and minus 2 cm (Figure AA.6). The decision to include the mattress height in the measurement of the interface angle comes from an analysis of the forces on a neck with and without the mattress (Figures AA.3 and AA.4). It is demonstrated that the mattress plays a significant part in entrapment where there is a wedging effect. It is understood that the measurement is done at the mattress height as specified by the MANUFACTURER. The measurement should represent a most adverse condition of the MEDICAL BED (i.e., smallest thickness mattress –2 cm, and thickest mattress +2 cm).

The angle measurements are taken for the range defined by (4) in Figure AA.6. For complex geometries in this range, measure to the tangential(s) of the SIDE RAIL.


See also Annex CC Particular guidance for assessing v-shaped opening



IEC 2141/09

Key

- 1 Normal force from SIDE RAIL
- 2 Normal force from MATTRESS SUPPORT PLATFORM
- 3 Neck
- 4 Resultant (ejecting) force




IEC 2142/09

Key

- 1 Normal force from SIDE RAIL
- 2 Normal force to neck from mattress
- 3 Neck
- 4 Resultant force from mattress
- 5 Mattress

**Figure AA.3 –
Resultant forces without mattress**


**Figure AA.4 –
Resultant forces with mattress**



IEC 2143/09

Key

- 1 SIDE RAIL
- 2 MATTRESS SUPPORT PLATFORM
- 3 Gap measurement (<60 mm)



IEC 2144/09

Key

- 1 Mattress
- 2 SIDE RAIL
- 3 Mattress height(s) (specified by MANUFACTURER)
- 4 Angle measurement zone

**Figure AA.5 –
Example of 60 mm gap measurement of B**

**Figure AA.6 –
Angle measurement example of B**


- Rationale for measurement of D

Area D is the space between the inside surface of the SIDE RAIL and the mattress compressed by the weight of a PATIENT'S head. The space should be small enough to prevent head entrapment when taking into account the mattress compressibility, any lateral shift of the mattress or SIDE RAIL, and the degree of play from loosened SIDE RAILS. A dimension of less

than 120 mm is required by this particular standard because the head is presumed to enter the space before the neck.

The goal of testing Area D is to assess if a PATIENT could trap his head in the horizontal space between the mattress and inside of the SIDE RAIL.

Additional examples of measurement of D are given in Figures AA.7 to AA.10.




Key

- 1 SIDE RAILS
- 2 Mattress
- 3 Cone tool

IEC 2145/09

Figure AA.7 – Placement of measurement TOOL for measurement of D




IEC 2146/09

Key

- 1 SIDE RAIL
- 2 Mattress
- 3 Cone tool
- 4 MATTRESS SUPPORT PLATFORM

**Figure AA.8 –
Example of area D measurement
that passes**




IEC 2147/09

Key

- 1 SIDE RAIL
- 2 Mattress
- 3 Cone tool
- 4 MATTRESS SUPPORT PLATFORM

**Figure AA.9 –
Example of area D measurement that fails**



IEC 2148/09


Key

- 1 SIDE RAIL
- 2 Mattress
- 3 Cone tool
- 4 MATTRESS SUPPORT PLATFORM

Figure AA.10 – Example of area D measurement that fails (on limit)


- Rationale on side rail intermediate height entrapment determination

Some artist sketches are provided below in order to demonstrate how entrapment may occur in each of the areas:



IEC 2149/09

Figure AA.11 – Example of potential PATIENT entrapment in area A within the SIDE RAIL



IEC 2150/09

Figure AA.12 – Example of potential PATIENT entrapment in area A below the SIDE RAIL



IEC 2151/09

Figure AA.13 – Example of potential PATIENT entrapment in area B




IEC 2152/09

Figure AA.14 – Example of potential PATIENT entrapment in area C between split SIDE RAIL




IEC 2153/09

Figure AA.15 – Example of potential PATIENT entrapment in area C between SIDE RAIL and HEAD BOARD



IEC 2154/09

Figure AA.16 – Example of potential PATIENT entrapment in area D



IEC 2155/09

Figure AA.17 – Example of potential PATIENT entrapment in area A below a single piece SIDE RAIL

- Rationale for test force of 250 N instead of 50 N for side rail test:

The investigations of incidents with entrapment in full-length side rails that fulfil the standard EN 1970 or IEC 60601-2-38 regarding the measurement A, show that seven incidents were registered. Four of these incidents occurred between the lower bar and the MATTRESS SUPPORT PLATFORM, the other three occurred between the upper (or middle) bar and the lower bar of the SIDE RAIL.

In five of the seven incidents the PATIENTS died, one of seven led to a fracture and one had no serious consequences. The five PATIENTS died due to entrapment of the chest or the neck (body completely slipped under the bar (see Figure AA.17)).

These results lead to the conclusion that the current test force of 50 N in IEC 60601-2-38 and the 30 N in EN 1970 is not sufficient.

The detailed review shows that in some cases the whole body of the PATIENT was lying on the bar and even when the PATIENT was very thin, 40 kg to 45 kg (estimated weight, weight was not registered) the body would have been pressing onto the bar. The bar bent and the (thin) person slipped with the legs under the bar and got stuck with the chest or the neck.

Even when it is considered that not the whole weight might be applied on the bar, it is estimated that at least 25 kg are pressed against it. Based on this estimate, the test force has been raised to 250 N for the testing of dimension A of SIDE RAILS.

Subclause 201.9.101 – Protection against inadvertent PATIENT falls

The requirement for SIDE RAILS height to be at least 220 mm above the MATTRESS SUPPORT PLATFORM in at least 50 % of the length of the MATTRESS SUPPORT PLATFORM is intended to ensure that SIDE RAILS reduce the RISK of the PATIENT accidentally slipping or rolling off the mattress.

Due to the nature of the SIDE RAIL entrapment HAZARD, compliance with the listed dimensional requirements is not in itself considered adequate to address all possible dangers (including entrapment of legs and arms, or the RISK of a PATIENT falling from an attempt to climb over the SIDE RAILS). For this reason, an additional requirement to perform a RISK ASSESSMENT (evaluating all issues pertaining to SIDE RAILS) has been included.

The 220 mm dimension is the minimum height requirement for the top edge of a SIDE RAIL above the mattress and is based on the midpoint (center of gravity) of an adult male torso at the 95th percentile lying on the shoulder on a compressed mattress.

In Table 201.202 it says that the minimum height of the SIDE RAIL from the top of the SIDE RAIL to the surface of an uncompressed mattress (dimension G) should be a minimum of 220 mm, the rationale above states that the 220 mm dimension is taken from the midpoint of an adult male torso at the 95th percentile lying on a compressed mattress. If this were true then the 220 mm dimension would be much less due to the sink into the mattress. If this rationale is relating the application to anthropometrics then surely it should state uncompressed mattress.

The minimum height requirement for SIDE RAILS may increase the RISK of injury when a PATIENT climbs over and falls. The minimum length requirement on SIDE RAILS may increase entrapment RISK by reducing egress points for the PATIENT, and while it may decrease inadvertent fall RISK from rolling out of MEDICAL BED, it may increase RISK where PATIENTS may opt to climb over the SIDE RAILS. The dimensional requirements in this particular standard reflect an attempt to make the best RISK tradeoffs on the basis of available information. Future data may demonstrate that some of these requirements should be revised at a future date.

Subclause 201.9.2.3.1 – Unintended movement

The probability of inadvertent motion due to component failure is significantly reduced when the lockout is engaged.

Several accidents have occurred involving foot-operated control pedals where PATIENTS, OPERATOR or visitors were injured or killed in the UNDERCARRIAGE of MEDICAL BEDS. The majority of the reported accidents occurred in long-term care environments. The factors of primary significance are:

- the existence of exposed MECHANICAL HAZARDS in the moving UNDERCARRIAGE,
- foot-operated control design which allows inadvertent activation, especially by unintended users,
- intermittent caregiver supervision, PATIENTS or OPERATORS accidentally falling or crawling over the control, and
- PATIENT, resident, or visitor characteristics such that the OPERATOR of the control lacks normal adult judgement.

Some of these factors need to exist simultaneously to create a HAZARD of mechanical injury in the MEDICAL BED UNDERCARRIAGE, so elimination of some may eliminate the RISK. The RISK for this HAZARD varies with the care environment. In APPLICATION ENVIRONMENT 1 and APPLICATION ENVIRONMENT 5 (intensive care and ambulatory), PATIENTS are generally under nearly continuous supervision, which minimizes RISK without regard to PATIENT characteristics or the MEDICAL BED design. RISK is comparatively high in long-term care unless protections are implemented in the MEDICAL BED design, because of intermittent caregiver supervision and resident mobility and/or mental characteristics of the MEDICAL BED occupants.

Experience in some countries with foot control (some incidents over recent years, including several deaths) shows that:

- incidents are more frequent in long-term services and with weak, disoriented or agitated PATIENTS but some other incidents occurred in acute care services or intensive care services and with normal adult judgement PATIENT;
- foot controls can generate an unacceptable RISK regarding its utility;
- a manual disconnection of foot controls or MOTION LOCKOUT CONTROL, without an another safety system, may not be an adequate solution to prevent unacceptable RISK because this function is not used consistently by OPERATOR.

Subclause 201.9.4.2.2 – Instability excluding transport

- Use of 1 100 N for patient weight

1 100 N is used to simulate a PATIENT, resident or visitor weight. 110 kg is the 99th percentile male weight as per *Adultdata, the Handbook of Adult Anthropometric and Strength Measurements*, by Laura Peebles and Beverley Norris. 110 kg corresponds to 1 078 N, or approximately 1 100 N.

- Explanation of the area on which load is applied

For the lateral test, having 2 persons sitting on the side edge of the MEDICAL BED, as close as possible to the end of the MEDICAL BED is reasonable. Since a 95th percentile male has a hip breadth, while sitting, of 466 mm [2], the weight of $2 \times 1\,100\text{ N} = 2\,200\text{ N}$ is distributed over a length of 950 mm (466 mm rounded to 475 mm, $\times 2$) and a width of 250 mm as in IEC 60601-2-38 and EN 1970. For the single person load in the longitudinal test (with fixed HEAD/FOOT BOARDS), the weight would be distributed over an area of 475 mm $\times$ 250 mm and these loads are placed on each side of the MEDICAL BED since the HEAD/FOOT BOARDS prevents sitting on the end of the MEDICAL BED. For the longitudinal test where the FOOT BOARD is removable, the 2 200 N load is placed at the foot-end of the MEDICAL BED and distributed across the entire width of the MEDICAL BED, which is about 950 mm.

Testing is done without the SAFE WORKING LOAD, as it is common for visitors to sit on the side or end of a MEDICAL BED when the PATIENT is NOT in the MEDICAL BED; this is the worst case. The SAFE WORKING LOAD in place on the MEDICAL BED during the test would only help stabilise the MEDICAL BED and is therefore not used. The loads specified in the lateral and longitudinal stability tests should be placed on the side or end of the MEDICAL BED that is heaviest in order to maximise the potential for tipping, but since it is difficult to determine which side is heaviest, the test is done on both sides and both ends.

- Explanation for use of 1850 N load in application environment 4

There have not been reported any accidents with the loading of 1 350 N at the foot end of the bed as required in EN 1970. The MEDICAL BED in APPLICATION ENVIRONMENT 4 is still safe for the MEDICAL BED PATIENT to sit anywhere on the MEDICAL BED, as the maximum PATIENT load always shall be used both for this test and the lateral stability test, if it exceeds 1 850 N. Higher loads for the test will be in conflict with ergonomic requirement in APPLICATION ENVIRONMENT 4.

Subclause 201.9.4.2.4.3 – Movement over a threshold

The use of 0,8 m/s instead of 0,4 m/s in the general standard is necessary to achieve that the MEDICAL BED will pass over the threshold with all castors. The speed of 0,8 m/s represents the real situation, because the OPERATOR will use this speed to pass the threshold. Therefore the construction of MEDICAL BEDS in APPLICATION ENVIRONMENTS 1, 2, 3 and 5 has to withstand this situation.

Subclause 201.9.4.3.1 – Instability in transport

The 6° angle was determined from building code regulations in different countries. Such regulations set the maximum angle allowed for floors at 5,7° in Canada, 4,76° in the UK, 4,76° in the USA and 2,86° in Denmark.

Subclause 201.9.8.2 – TENSILE SAFETY FACTOR

The safety factor requirements from pre-existing standard IEC 60601-2-38 and EN 1970 were demonstrated to be sufficient by several years of field application. Consequently these requirements are carried over into this particular standard. Also, in the normative Annex BB, durability tests are given for specific APPLICATION ENVIRONMENTS with regard to the real situation over the lifetime of a MEDICAL BED.

Subclause 201.9.8.3.2 – Static forces due to loading from persons

- Static strength

EN 1970 was substantially similar to IEC 60601-2-38 in this regard with the exception of the MEDICAL BED being in the worst-case position. The general standard refers to the worst-case as the "most disadvantageous" position. The general standard text was therefore chosen, where "most disadvantageous height" refers to the up/down position of the MEDICAL BED. The test method requires also most disadvantageous length, as this could be worse case for MEDICAL BEDS, with built-in or added length adjustment. Additionally, a 1-hour time duration should be applied only when assemblies contain elements subject to material creep (plastics or other non-metals). Where all metal assemblies are being tested, timeframes above 1 min are not needed (see IEC 60601-1:2005, subclause 9.8.2, Note 2).

- ACCESSORIES strength

Terminology of the general standard was adopted. Where ACCESSORIES may be cycled to induce similar wear of components, use of the SAFE WORKING LOAD is acceptable to prove compliance with the requirement, provided expected wear is simulated on the ACCESSORY or attachment point prior to testing. The expected distribution of weight upon the ACCESSORY during reasonably foreseeable misuse should be considered when testing the ability of the ACCESSORY load support.

The 1 h requirement is only required in designs where material creep is a concern (plastic parts are used). For assemblies that are all metal, 1 min is enough to determine if the design is capable of withstanding the SAFE WORKING LOAD times the TENSILE SAFETY FACTOR (see IEC 60601-1:2005, subclause 9.8.2 Note 2).

Subclause 201.9.8.3.3 – Dynamic forces due to loading from persons

- Loading from persons

The above requirements were added due to issues seen in the field that were perceived to have been caused by improper specification of dynamic strength. The issues included the failure of the MEDICAL BED or undercarriage assemblies, resulting in the MEDICAL BED surface dropping to the lowest level. It was therefore felt that a minimum standard test should be specified. The dynamic strength tests are designed to be a minimum standard to assess the dynamic strength of the MEDICAL BED. Testing should be modified as per the MANUFACTURER'S RISK ASSESSMENT program based on such things as elimination of RISK by design, normal and abuse conditions by the OPERATOR or PATIENT, environmental issues and others. Components, which loosen, or malfunction during tests that do not contribute to additional RISK, or sudden collapse of product, may be replaced.

Subclause 201.9.8.3.3.2 – Dynamic testing of the height adjustment mechanism

The experience of the experts from test houses and MANUFACTURER with the market observation leads to the result that a dynamic test as described is necessary to provide a minimum of safety between reasonable service intervals.

The number of 3 000 cycles is calculated on the base of 1 to 2 cycles a day, which represents a product life time of approximately 4-8 years.

Subclause 201.9.8.3.3.3 – SIDE RAIL strength and latch reliability

This requirement combines the proposed cycle test with the load test in IEC 60601-2-38 to ensure proper latching of the SIDE RAIL after a simulated life cycle. "HAZARD" was changed to "RISK" to use terminology of IEC 60601-1:2005. 30 000 cycles is used considering an average product life cycle of 10 years, with 4 usages per day (14 600 cycles) multiplied by a safety factor of 2. The number of cycles should be representative of the actual product life cycle, and product estimated usage of the SIDE RAILS. Components, which loosen, or malfunction during tests that do not contribute to additional RISK or sudden collapse of SIDE RAILS, may be replaced.

- SIDE RAIL latch

Lateral and longitudinal cycling: The number of cycles chosen is arbitrary in an effort to reproduce the action of an agitated PATIENT exerting reasonably foreseeable misuse on the SIDE RAIL. Once again, testing should be tailored to the MANUFACTURER'S indicated life cycle, estimated SIDE RAIL usage and RISK ANALYSIS with regards to number of cycles, forces applied and direction of force.

The specified maximum PATIENT weight is used to simulate a PATIENT climbing over the SIDE RAIL, as an additional reasonably foreseeable misuse. No data has been presented on PATIENTS over 1 112 N climbing over SIDE RAILS, resulting in a fall.

100 N is selected for this test as an approximation of an agitated PATIENT shaking the SIDE RAIL from the MEDICAL BED.

For Figure 201.117 Force A represents someone pulling up on the SIDE RAIL, the application of force should be on top of the SIDE RAIL

For Figure 201.117 Force B represents a PATIENT climbing over the top of the SIDE RAIL, the force should be applied to top the SIDE RAIL also.

For Figure 201.117 Forces C, D, E and F represent a PATIENT/OPERATOR using the SIDE RAIL as a grab handle.

Subclause 201.11.8 – Interruption of power supply/SUPPLY MAINS to ME EQUIPMENT

Examples of emergency procedures (as specified by the MANUFACTURER) that are needed during an interruption of the power supply/SUPPLY MAINS to MEDICAL BEDS are:

- 1) positions for cardiopulmonary resuscitation (CPR), for example, emergency back flattening;
- 2) others as specified by the MANUFACTURER.

Subclause 201.13.1.4 – Specific MECHANICAL HAZARDS

The U.S. Food and Drug Administration incident database was reviewed in the preparation of this standard to relate field incidents of unintended movement to the concept of SINGLE FAULT

CONDITION and abnormal operation in IEC 60601-1:2005. This review indicates such incidents mainly manifest when an electrically driven actuation starts unexpectedly and drives to its limit in one direction. The apparent cause is a fault occurring in MEDICAL BED function controls.

Abnormal operation/fault condition requirements emphasize protection against faults in the motion control systems of electrically powered MEDICAL BEDS. An interpretation is also developed to extend the SINGLE FAULT SAFE concept for unintended motion to non-electrically powered MEDICAL BEDS and pneumatic and hydraulic subsystems of electrical MEDICAL BEDS.

The requirement to provide a means (not accessible by the PATIENT) for deactivating controls that are accessible by the PATIENT is retained from the previous standard, but regarded as a human factors aspect that is not inherently integral to SINGLE FAULT CONDITION and abnormal operation requirements.

Tradeoffs are required among human factors, clinical benefits available from positioning the MATTRESS SUPPORT PLATFORM, and technology. For instance, it is possible to reduce the probability of a SINGLE FAULT CONDITION causing a movement without human intervention by the use of redundant circuits and parallel switches, which have to be closed simultaneously. Such a requirement is out of proportion to the RISK for the vast majority of PATIENTS and clinical situations. It would degrade the USABILITY of the product for clinical users and be equivalent to removing PATIENT controls for those PATIENTS whose lack of dexterity or medical condition prevents them from operating two controls simultaneously.

This standard introduces the concept of the MOTION LOCKOUT CONTROL. The MOTION LOCKOUT CONTROL provides a means to reconcile varying PATIENT needs. Some PATIENTS need controls with a high degree of USABILITY and have little RISK exposure in the event of rare unintended motion. Other PATIENTS require exceptional protection against possible unintended motion, even if control ergonomics are entirely sacrificed. A MOTION LOCKOUT CONTROL deployable at OPERATOR option can meet both needs.

Subclause 201.15.3.5 – Rough handling test

The reason for the difference to the speed of 0,8 m/s used in 201.9.4.2.4.3 is because the OPERATOR normally did not intend to hit the door frame and will stay with the speed (0,4 m/s) according to the requirements of the general standard.

Subclause 201.15.4.4 – Indicators

MEDICAL BEDS connected to MAINS VOLTAGE are generally operational at all times. This is generally understood by the OPERATORS; therefore no indicator light is required.

Subclause 201.15.4.6.2 – Limitation of movement

Loss of blood flow to legs (from over-flexing of the torso) or build-up of blood or other fluids in the abdomen can cause severe damage to PATIENT.

Annex BB (normative)

Design requirements and recommendations for MEDICAL BEDS

BB.1 General

Although it is recognized that the responsibility for defining the INTENDED USE (and therefore the configuration of features) of the MEDICAL BED rests with the MANUFACTURER, certain factors shall be taken into consideration. The following is intended both as requirements for MEDICAL BEDS intended for certain APPLICATION ENVIRONMENTS and also as guidelines for MANUFACTURERS when designing MEDICAL BEDS for other APPLICATION ENVIRONMENTS.

Table BB.1 indicates normative and informative requirements for the different APPLICATION ENVIRONMENTS 1 to 5.

**Table BB.1 – Normative and informative requirements for different
APPLICATION ENVIRONMENTS 1 to 5**

Clause	APPLICATION ENVIRONMENT 1	APPLICATION ENVIRONMENT 2	APPLICATION ENVIRONMENT 3	APPLICATION ENVIRONMENT 4	APPLICATION ENVIRONMENT 5
BB.2.1	Normative	Normative	Normative	Normative	Normative
BB.2.2	Normative	Normative	Normative	Normative	Normative
BB.2.3	Normative	Normative	Normative	Normative	Normative
BB.2.4	Normative	Normative	Normative	Normative	Normative
BB.3.1	Informative	Informative	Informative	Normative	Informative
BB.3.2	Informative	Informative	Informative	Normative	Informative
BB.3.3.1	Normative	Normative	Normative	Normative	Normative
BB.3.3.2	Normative	Normative	Normative	Normative	Normative
BB.3.3.3	Informative	Informative	Informative	Normative	Informative
BB.3.3.4	Informative	Informative	Informative	Normative	Informative
BB.3.3.5	Informative	Informative	Informative	Normative	Informative
BB.3.3.6	Informative	Informative	Informative	Normative	Informative
BB.3.3.7	Informative	Informative	Informative	Normative	Informative
BB.3.4	Informative	Informative	Informative	Normative	Informative
BB.4.1	Normative	Normative	Normative	Normative	Normative
BB.4.2	Informative	Informative	Informative	Normative	Informative
BB.4.3	Normative	Normative	Normative	Normative	Normative
BB.4.4	Normative	Normative	Normative	Normative	Normative
BB.4.4/dash 1	Informative	Informative	Informative	Normative	Informative
BB.4.4dash 2	Informative	Informative	Informative	Normative	Informative
BB.4.4/dash 3	Normative	Normative	Normative	Normative	Normative
BB.4.4/dash 4	Informative	Informative	Informative	Normative	Informative
BB.4.4/dash 5	Normative	Normative	Normative	Normative	Normative

BB.2.1 General

Circle "C" is used for impact testing according to BB.2.3

Figure BB.1 – Other areas of possible impact testing

After application of the load, no elements of the MEDICAL BED shall become loose, fractured or present any HAZARD.




Figure BB.2 – Impactor

The impactor body is approximately 200 mm in diameter and is separated from the striking surface by helical compression springs. The body is free to move on a line perpendicular to the plane of the central area of the striking surface. The total mass of the assembly shall be $25 \text{ kg} \pm 0,1 \text{ kg}$ and the body and its associated parts (minus the springs) shall have a mass of $17 \text{ kg} \pm 0,1 \text{ kg}$.

The combined spring rate (for the system) shall be $6,9 \text{ N/mm} \pm 1 \text{ N/mm}$ and the total friction resistance of the moving parts is between 0,25 N and 0,45 N. When the spring system is compressed to a load of $1\,040 \text{ N} \pm 5 \text{ N}$, the remaining compression distance shall be 60 mm minimum.

The striking surface shall be a rigid circular object, $(200 \pm 5) \text{ mm}$ in diameter, the face of which has a convex spherical curvature of 300 mm radius with a 12 mm front edge radius.

BB.2.3 The effect of loading on the edge of the MATTRESS SUPPORT PLATFORM

The MEDICAL BED shall be able to withstand the effect of loading on the edge of the MATTRESS SUPPORT PLATFORM as applied during NORMAL USE.

Compliance is checked by the following test:

Place the MEDICAL BED in its flat position. If the BED-LIFT is separate from a MATTRESS SUPPORT PLATFORM, mount a TEST BED BOARD to the BED-LIFT.

Adjust the height to the most adverse position.

Load three loading pads with a load of 750 N each to the MATTRESS SUPPORT PLATFORM to the locations marked "C" in Figure BB.1.

Examine the specimen during and after application of the loads to determine if the deflections present any HAZARD.

After application of the loads, no elements of the BED shall become loose, fractured or present any HAZARD.

The deflection shall not be greater than 40 mm during application of the loads and not greater than 10 mm measured in relation to the floor after removal of the loads.

BB.2.4 Durability of the moveable sections (if provided) of the MATTRESS SUPPORT PLATFORM

The moveable sections of the MATTRESS SUPPORT PLATFORM shall comply with the following test:

Apply the SAFE WORKING LOAD distributed as indicated in Figure 201.115 on the MATTRESS SUPPORT PLATFORM.

Each of the moveable sections of the MATTRESS SUPPORT PLATFORM are operated through their full range of travel as in NORMAL USE.

The test is performed for 1 000 cycles.

Inspect the MEDICAL BED 5 min after cycling is completed.

After application of the load, no elements of the MEDICAL BED shall become loose, fractured or present any HAZARD and deformation of less than 10 mm (after removal of the load) from the corresponding measurements taken before the application of the load is acceptable.

BB.3 Human factors (ergonomic)

BB.3.1 General

It is recognized that for a MEDICAL BED to accommodate the medical practices and procedures for which it will be used, the interaction of the human body with the ME EQUIPMENT shall be considered as a method of assuring efficient and safe operation. The following is presented for consideration when addressing these issues for MEDICAL BEDS in APPLICATION ENVIRONMENTS 1, 2, 3 and 5. For MEDICAL BEDS intended for APPLICATION ENVIRONMENT 4, BB.2 is normative.

NOTE The following may not be the only way to address the issues and any dimensions should be based on statistical analyses of regional populations. When used, appropriate dimensions should always be determined based on anthropometric data for the population in the area in which the MEDICAL BED is intended to be used.

BB.3.2 Range of adjustment of the height of the MATTRESS SUPPORT PLATFORM

Required/recommended range of adjustment (if applicable) of the height of the MATTRESS SUPPORT PLATFORM above the floor:

The MATTRESS SUPPORT PLATFORM, in its flat position shall/should have a range of adjustment, which at least allows positioning between 400 mm and 800 mm above the floor, measured at the uppermost part of the MATTRESS SUPPORT PLATFORM.

MEDICAL BEDS without the possibility of adjustment of height are only suitable for persons not needing caregivers.

BB.3.3 Dimensions for handles and pedals

BB.3.3.1 All handles and pedals shall/should be reachable from normal working positions and be placed so they allow operations to be carried out in a safe and ergonomic manner.

BB.3.3.2 Location and configuration of all buttons / switches / actuators shall/should minimize the RISK of unintentional activation.

BB.3.3.3 The surface of buttons shall/should at least cover a circle of 15 mm in diameter. Any distance between buttons shall/should be more than 10 mm.

The buttons may have any shape, e.g. they need not to be circular.

BB.3.3.4 The free distance between any handle (those parts intended to be grabbed) requiring an operating force of more than 10 N and any construction part shall/should be more than 35 mm.

BB.3.3.5 The free distance between the upper side of any pedal (those parts intended to be stepped on) and any construction part shall/should be more than 75 mm.

BB.3.3.6 The diameter of operating handles requiring an operating force of more than 10 N shall/should be between 19 mm and 43 mm.

BB.3.3.7 Pedals shall/should not be more than 300 mm above the floor.

BB.3.4 Operating forces for handles and pedals

BB.3.4.1 In the case of electrically operated functions, the force to operate the control unit buttons shall/should be less than 5 N.

BB.3.4.2 In the case of non-electrically operated functions, the force to release operating handles shall/should be no more than 10 N and the force required to lift any part of the MATTRESS SUPPORT PLATFORM shall/should be no more than 200 N, when loads are fixed to the MATTRESS SUPPORT PLATFORM as shown in Figure 201.115, with their centre of gravity in the middle of the sections.

BB.3.4.3 The static force to operate pedals shall/should be no more than 300 N when loads are fixed to the MATTRESS SUPPORT PLATFORM, as shown in Figure 201.115, with their centre of gravity in the middle of the sections.

BB.3.4.4 For MOBILE MEDICAL BEDS intended to transport PATIENTS, the force required to initiate movement of the MEDICAL BED with the SAFE WORKING LOAD shall/should not exceed 160 N. The force required to maintain movement (after it has been initiated) shall/should not exceed 85 N.

BB.4 Functionality

BB.4.1 General

Frequently, MEDICAL BEDS are designed for use in combination with other ME EQUIPMENT. MANUFACTURERS are encouraged to evaluate not only those characteristics of the MEDICAL BED that could affect the safe operation of the system (the MEDICAL BED in combination with the ME EQUIPMENT) but also the characteristics of the MEDICAL BED, which facilitate the use of the system.

In BB.3 both requirements for MEDICAL BEDS intended for certain APPLICATION ENVIRONMENTS are presented as well as guidelines for MANUFACTURERS when designing MEDICAL BEDS for other APPLICATION ENVIRONMENTS.

BB.4.2 Combination of MEDICAL BED and MOBILE hoists

The following is presented for consideration when addressing this issue for MEDICAL BEDS in APPLICATION ENVIRONMENTS 1, 2, 3 and 5. For MEDICAL BEDS intended for APPLICATION ENVIRONMENT 4, BB.3.1 is normative.

With the MATTRESS SUPPORT PLATFORM lowered to any position higher than 400 mm, the height of the under MEDICAL BED clearance shall/should be at least 150 mm at a length of at least 500 mm positioned on each side of the symmetry plane of the MATTRESS SUPPORT PLATFORM (see Figure BB.2)




Figure BB.3 – Schematic presentation of under MEDICAL BED clearance

BB.4.3 Combination of MEDICAL BED and mattress

For all MEDICAL BEDS at the area of the MATTRESS SUPPORT PLATFORM intended for the ingress/egress of the BED, the height of the mattress used shall at least be 20 mm higher than any construction part e.g. MEDICAL BED frame, side structure or the lowered SIDE RAILS of the MATTRESS SUPPORT PLATFORM.

When SPECIALITY MATTRESSES are used, a RISK ASSESSMENT shall be performed if the requirement cannot be fulfilled.

BB.4.4 Range of adjustment of moveable sections of the MATTRESS SUPPORT PLATFORM


The required/recommended range of adjustment (if applicable) for back, upper leg, and lower leg sections of the MATTRESS SUPPORT PLATFORM is as follows.

- The angle *A* between the back rest section and the MEDICAL BED board frame (see Figure BB.4) shall/should at least be adjustable within the range of 0° (horizontal) and 70°. A maximum of 85° shall/should only be reachable when the MEDICAL BED is intended to raise a person to an upright sitting position for transfer. Recommendation for MEDICAL BEDS intended for APPLICATION ENVIRONMENTS 1, 2, 3 and 5, requirement for MEDICAL BEDS intended for APPLICATION ENVIRONMENT 4.
- The angle *B* between a line drawn between the turning point of the back rest section/seat section and the turning point of the upper/lower leg section and horizontal (see Figure BB.3) shall/should at least be adjustable between 0° and 12°. Recommendation for

MEDICAL BEDS intended for APPLICATION ENVIRONMENTS 1, 2, 3 and 5, requirement for MEDICAL BEDS intended for APPLICATION ENVIRONMENT 4.

- The angle *C* between the back rest section and a line drawn between the turning point of the back rest section/seat section and the turning point of the upper/lower leg section (see Figure BB.4) shall be greater than 90° for all MEDICAL BEDS (as in 201.15.4.6.2).
- The angle *D* between the lower leg section and horizontal (see Figure BB.4) shall/should at least be adjustable between 0° and 20° under horizontal. Recommendation for MEDICAL BEDS intended for APPLICATION ENVIRONMENTS 1, 2, 3 and 5, requirement for MEDICAL BEDS intended for APPLICATION ENVIRONMENT 4.
- The angle *E* between the upper side of the upper leg section and the upper side of the lower leg section (see Figure BB.4) shall be at least 180° for all MEDICAL BEDS.

IECNORM.COM : Click to view the full PDF of IEC 60601-2-52:2009



IEC 2159/09

Figure BB.4 – Recommendations and requirements regarding angles for different sections of the MATTRESS SUPPORT PLATFORM

Annex CC (informative)

Particular guidance for assessing risk of entrapment in v-shaped openings

An additional proposed test method of assessing RISK of entrapment in v-shaped openings in relation to areas B and C (see Table 201.101) between the elements of SIDE RAIL, HEAD/FOOT BOARDS and MATTRESS SUPPORT PLATFORM is offered. This test method is not validated at the moment and it is placed as an informative annex to receive comments from the IEC-community for future adoptions.

The reasons for offering this additional method are as follows:

For area B it offers an additional method of assessing the contribution to entrapment into the gap between the elements due to the geometry of the v-shaped openings. The test method at the same time reduces the influence of MATTRESS depth and compressibility.


For area C it offers a means of assessing the contribution to entrapment into the gap between the elements due to the geometry of the v-shaped openings, this is not covered in the normative part of the standard (see Table 201.101 area C).

Subclause 201.9.1.101 – Protection against PATIENT entrapment in non-moving parts

To reduce the risk of neck entrapment, openings between elements in the MEDICAL BED system above the MATTRESS SUPPORT PLATFORM should be small enough to prevent passage of a small neck. Additionally to reduce neck entrapment with regards to wedging of the neck, angles for V-shaped entry into the openings should be greater than 75°. The 75° angle reflects an understanding that at some point, a narrow angle creates a V-shaped space or geometry that could be conducive to entrapment. The 75° angle was chosen by reviewing the tools to assess entrapment risk for swimming pool equipment (see EN 13451-1) and making changes suitable to application on MEDICAL BEDS by increasing the wedging minimum angle of 60° to 75°. The 75° angle reduces the RISK of PATIENT entrapment through hooking of the neck and head in the V-shaped opening (See Figure AA.13 and Figure AA.15). The wider angle allows the PATIENT to pull his head free from a vertical opening with less effort. This is essential when considering a weak or confused PATIENT.

To measure V-shaped openings in relation to areas B and C a test method and wedge tool (see Figure CC.1) has been developed. The tool is made from aluminium and weighs 3,34 kg representing the mass of a small adults head and neck (the 3,34 kg was taken from Air Force Research laboratory Design and Development of Anthropometrically Correct Head Forms for Joint Strike Fighter Ejection Seat Testing – February 2005), the wedge angle is 75° and has a 60 mm diameter cylindrical end representing the small adult neck.

Dimension in millimetres



IEC 2160/09

Material: Aluminium

Weight: 3,34 kg

Figure CC.1 – Wedge tool

- Proposed rationale for measurement of area B


The measurement of area B includes the gap beneath the SIDE RAIL, the v-shaped opening between the underneath of the SIDE RAIL and the leading edge of the SIDE RAIL to the MATTRESS SUPPORT PLATFORM (see Figure CC.2). The 75° angle reduces the RISK of PATIENT entrapment through hooking of the neck and head in the v-shaped opening (see Figure AA.13).

- Proposed and unvalidated test method for V- shaped openings

The MATTRESS SUPPORT PLATFORM is set in the flat horizontal position (and at the most disadvantageous height depending on the design) with the SIDE RAILS raised and locked (with segmented SIDE RAILS test independently and together).

Rest the wedge tool against the edge of the MATTRESS SUPPORT PLATFORM and push it into the V-shaped opening between the lowest point of the SIDE RAIL outside of the SIDE RAIL support and the MATTRESS SUPPORT PLATFORM orientating the wedge tool into the narrowest opening. The wedge tool must contact against both sides of the 60 mm cylinder diameter part of the wedge tool (it should not pass through). The opening should not allow the wedge tool to come into contact with both sides of the 75° angle at the same time, see pass fail examples Figure CC.3.


The influence of any specified mattress fitted on the areas of potential entrapment should be assessed for acceptability and justified in the MANUFACTURERS RISK MANAGEMENT FILE.



Key

- 1 Leading edge of safety side
- 2 Gap (< 60 mm)
- 3 Mattress support platform
- 4 V-shaped opening (> 75°)

Figure CC.2 – V-shaped opening in relation to B



IEC 2162/09

Figure CC.3 – Pass/fail in relation to area B

- Proposed rationale for measurement of area C with the wedge tool

The measurement of area C includes the v-shaped openings between the SIDE RAIL and the HEAD/FOOTEND and the v-shaped opening between split SIDE RAILS. The 75° angle reduces the risk of PATIENT entrapment through hooking of the neck and head in the v-shaped opening (see Figures AA.14 and AA.15).

- Proposed and unvalidated test method area C for V- shaped openings

The mattress support platform is set in the flat horizontal position (and at the most disadvantageous height depending on the design) with the SIDE RAILS in any raised and locked positions. For beds with segmented SIDE RAILS the test has also to be performed in the worst case opening between SIDE RAILS in articulated positions.

Hold the wedge tool by the handle and place it between the narrowest distance between the gap in the V-shaped opening. Allow the tool to rest under its own weight in the opening, the 60 mm diameter cylindrical end should always contact both sides of the V-shaped opening and should not pass through into the gap. Rotate the tool until one of the 75° angled sides touches the edge of one side of the opening (it may be necessary to angle the tool to locate it against the narrowest part of the opening, see Figure CC.4).

The opening should not allow the wedge tool to come into contact with both sides of the 75° angle at the same time, see pass/fail examples Figures CC.5 and CC.6.




Figure CC.4 – Positioning of wedge tool





Figure CC.5 – Pass/fail in relation to area C between HEAD BOARD and FOOT BOARD



**Figure CC.6 – Pass/fail in relation to area C between split
SIDE RAILS**

Bibliography

- [1] HD22.10 H05-BQ-F (DIN VDE 0207-20, *Insulating and sheathing compounds for cables and flexible cords – Rubber insulating compounds*)
- [2] People Size, 1998
- [3] IEC 60601-2-38, *Medical electrical equipment – Part 2: Particular requirements for the safety of electrically operated hospital beds*
- [4] EN 1970, *Adjustable beds for disabled persons – Requirements and test methods*
- [5] EN 13451-1, *Swimming pool equipment – General safety requirements and test methods*
- [6] PEEBLES, Laura and NORRIS Beverley. *Adultdata: The handbook of Adult Anthropometric and Strength Measurements*. University of Nottingham, 1998
- [7] AFRL-HE-WP-TR-2005-0044, *Design and Development of Anthropometrically Correct Head Forms for Joint Strike Fighter Ejection Seat Testing*, Air Force Research Laboratory

IECNORM.COM : Click to view the full PDF of IEC 60601-2-52:2009

Index of defined terms used in this particular standard

ACCESSORY.....	IEC 60601-1:2005, 3.3
APPLICATION ENVIRONMENT 1	201.3.201
APPLICATION ENVIRONMENT 2	201.3.202
APPLICATION ENVIRONMENT 3	201.3.203
APPLICATION ENVIRONMENT 4	201.3.204
APPLICATION ENVIRONMENT 5	201.3.205
APPLIED PART.....	201.3.8
BASIC SAFETY	IEC 60601-1:2005, 3.10
BED-LIFT	201.3.206
CLASS II	IEC 60601-1:2005, 3.14
COMPONENTS WITH HIGH-INTEGRITY CHARACTERISTICS	IEC 60601-1:2005, 3-17
DISABLED PERSON.....	201.3.207
ENCLOSURE	IEC 60601-1:2005, 3.26
ESSENTIAL PERFORMANCE	IEC 60601-1:2005, 3.27
EXPECTED SERVICE LIFE	IEC 60601-1:2005, 3.28
FUNCTIONAL CONNECTION.....	IEC 60601-1:2005, 3.33
HAZARD	IEC 60601-1:2005, 3.39
HAZARDOUS SITUATION.....	IEC 60601-1:2005, 3.40
HEAD/FOOT BOARD.....	201.3.208
INTENDED USE	IEC 60601-1:2005, 3.44
LEAKAGE CURRENT.....	IEC 60601-1:2005, 3.47
LIFTING POLE	201.3.209
MAINS VOLTAGE	IEC 60601-1:2005, 3.54
MANUFACTURER.....	IEC 60601-1:2005, 3.55
MATTRESS OVERLAY	201.3.210
MATTRESS SUPPORT PLATFORM.....	201.3.211
MECHANICAL HAZARD.....	IEC 60601-1:2005, 3.61
MECHANICAL PROTECTIVE DEVICE	IEC 60601-1:2005, 3.62
MEDICAL BED	201.3.212
MEDICAL ELECTRICAL EQUIPMENT (ME EQUIPMENT).....	IEC 60601-1:2005, 3.63
MEDICAL ELECTRICAL SYSTEM (ME SYSTEM).....	IEC 60601-1:2005, 3.64
MOBILE	IEC 60601-1:2005, 3.65
MODEL OR TYPE REFERENCE	IEC 60601-1:2005, 3.66
MOTION LOCKOUT CONTROL	201.3.213
NORMAL USE.....	IEC 60601-1:2005, 3.71
OPERATOR	IEC 60601-1:2005, 3.73
PATIENT	201.3.76
PENDANT CONTROL	201.3.214
POWER SUPPLY CORD.....	IEC 60601-1:2005, 3.87
PROCESS	IEC 60601-1:2005, 3.89
PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	IEC 60601-1:2005, 3.90
RESPONSIBLE ORGANIZATION	IEC 60601-1:2005, 3.101

RISK	IEC 60601-1:2005, 3.102
RISK ANALYSIS	IEC 60601-1:2005, 3.103
RISK ASSESSMENT	IEC 60601-1:2005, 3.104
RISK MANAGEMENT FILE	IEC 60601-1:2005, 3.108
SAFE WORKING LOAD	IEC 60601-1:2005, 3.109
SIDE RAIL	201.3.215
SINGLE FAULT CONDITION	IEC 60601-1:2005, 3.116
SPECIALTY MATTRESS	201.3.216
SUPPLY MAINS	IEC 60601-1:2005, 3.120
TENSILE SAFETY FACTOR	IEC 60601-1:2005, 3.121
TEST BED BOARD	201.3.217
TOOL	IEC 60601-1:2005, 3.127
TRAPPING ZONE	201.3.131
UNDERCARRIAGE	201.3.218
USABILITY	IEC 60601-1:2005, 3.136

IECNORM.COM : Click to view the full PDF of IEC 60601-2-52:2009

SOMMAIRE

AVANT-PROPOS.....	87
INTRODUCTION.....	90
201.1 Domaine d'application, objet et normes connexes	91
201.2 Références normatives.....	92
201.3 Termes et définitions	93
201.4 Exigences générales	96
201.5 Exigences générales relatives aux essais des APPAREILS EM	96
201.6 Classification des APPAREILS EM et des SYSTÈMES EM	98
201.7 Identification, marquage et documentation des APPAREILS EM	98
201.8 Protection contre les DANGERS d'origine électrique provenant des APPAREILS EM	103
201.9 Protection contre les DANGERS MÉCANIQUES des APPAREILS EM ET SYSTÈMES EM	104
201.10 Protection contre les DANGERS dus aux rayonnements involontaires ou excessifs.....	127
201.11 Protection contre les températures excessives et les autres DANGERS	127
201.12 Précision des commandes et des instruments et protection contre les caractéristiques de sortie présentant des DANGERS	130
201.13 SITUATIONS DANGEREUSES et conditions de défaut	130
201.14 SYSTÈMES ÉLECTROMÉDICAUX PROGRAMMABLES (SEMP)	131
201.15 Construction de l'APPAREIL EM.....	131
201.16 SYSTÈMES EM	134
201.17 Compatibilité électromagnétique des APPAREILS EM et des SYSTÈMES EM	134
Annexes	134
Annexe AA (informative) Guide particulier et justifications	135
Annexe BB (normative) Exigences et recommandations de conception pour les LITS MÉDICAUX	151
Annexe CC (informative) Guide particulier pour l'évaluation du risque de piégeage dans les ouvertures en forme de V.....	159
Bibliographie.....	166
Index des termes définis utilisés dans la présente norme particulière	167
Figure 201.101 – PARTIE APPLIQUÉE	93
Figure 201.102 – LIT MÉDICAL, configuration générale (exemple, présentation schématique uniquement).....	95
Figure 201.103a – Outil conique	97
Figure 201.103b – Outil cylindrique.....	97
Figure 201.103 – Outils d'essai de piégeage.....	97
Figure 201.104 – Patin de charge	98
Figure 201.105 – Symbole graphique du poids maximal du PATIENT et de la CHARGE DE FONCTIONNEMENT EN SÉCURITÉ	99
Figure 201.106 – Commandes de fonctions et/ou actionneurs du LIT MÉDICAL: lignes directrices pour la création des symboles graphiques	101
Figure 201.107 – Exemple de LIT MÉDICAL avec une BARRIÈRE LATÉRALE segmentée ou scindée.....	105

Figure 201.108 – Exemple de LIT MÉDICAL avec une BARRIÈRE LATÉRALE d'un seul tenant	106
Figure 201.109 – Espacement admissible pour les doigts dans les zones de portée normale autour du périmètre du SOMMIER	111
Figure 201.110 – Exemple utilisant des barrières pour la mesure des distances d'isolement autour du périmètre du SOMMIER, pour réduire les risques de piégeage des doigts du PATIENT	112
Figure 201.111 a) – Zone de passage du pied et de l'orteil entre les parties mobiles et le sol	112
Figure 201.111b) – Zone de passage de l'orteil entre les parties mobiles et le sol	113
Figure 201.111 – Zones de passage	113
Figure 201.112 – Essai de stabilité latérale sur les côtés du LIT MÉDICAL	115
Figure 201.113 – Essai de stabilité longitudinale avec PANNEAU PIED DE LIT amovible	115
Figure 201.114 – Essai de stabilité longitudinale avec PANNEAUX TÊTE ET PIED DE LIT fixes	116
Figure 201.115 – Répartition de la CHARGE DE FONCTIONNEMENT EN SECURITÉ pour les essais	120
Figure 201.116 – Emplacement du patin de charge (voir la Figure 201.104)	123
Figure 201.117 – Application des forces pour l'essai de la BARRIÈRE LATÉRALE	125
Figure 201.118 – Hauteur de la BARRIÈRE LATÉRALE	126
Figure 201.119a – Angle γ entre la section relève-buste et la section relève-jambes du SOMMIER	132
Figure 201.119b – Angle γ entre la section relève-buste et la section plicature genoux supérieure du SOMMIER	132
Figure 201.119c – Angle γ entre la section relève-buste et la section plicature genoux supérieure du SOMMIER	132
Figure 201.119d – Angle γ entre la section relève-buste et la section relève-jambes du SOMMIER	133
Figure 201.119 – Configurations du SOMMIER	133
Figure AA.1 – Marquage pour le choix des matelas recommandés spécifiés par le FABRICANT	137
Figure AA.2 – Marquage des BARRIERES LATERALES amovibles spécifiées par le FABRICANT	137
Figure AA.3 – Forces résultantes sans matelas	141
Figure AA.4 – Forces résultantes avec matelas	141
Figure AA.5 – Exemple de mesure d'un espace de 60 mm de B	141
Figure AA.6 – Exemple de mesure d'angle de B	141
Figure AA.7 – Emplacement de l'OUTIL de mesure pour la mesure de D	142
Figure AA.8 – Exemple d'une mesure de la zone D réalisée avec succès	142
Figure AA.9 – Exemple d'une mesure de la zone D qui échoue	142
Figure AA.10 – Exemple d'une mesure de la zone D qui échoue (à limite)	143
Figure AA.11 – Exemple de piégeage potentiel du PATIENT dans la zone A dans la BARRIÈRE LATÉRALE	143
Figure AA.12 – Exemple de piégeage potentiel du PATIENT dans la zone A sous la BARRIÈRE LATÉRALE	143
Figure AA.13 – Exemple de piégeage potentiel du PATIENT dans la zone B	143
Figure AA.14 – Exemple de piégeage potentiel du PATIENT dans la zone C entre les BARRIERES LATÉRALES scindées	143

Figure AA.15 – Exemple de piégeage potentiel du PATIENT dans la zone C entre la BARRIÈRE LATÉRALE et le PANNEAU TÊTE DE LIT.....	144
Figure AA.16 – Exemple de piégeage potentiel du PATIENT dans la zone D	144
Figure AA.17 – Exemple de piégeage potentiel du PATIENT dans la zone A sous une BARRIÈRE LATÉRALE d'un seul tenant	144
Figure BB.1 – Autres zones d'essais d'impacts éventuels	152
Figure BB.2 – Appareil d'essai aux chocs	153
Figure BB.3 – Présentation schématique de la distance d'isolement sous le LIT MÉDICAL	156
Figure BB.4 – Recommandations et exigences concernant les angles des différentes sections du SOMMIER	158
Figure CC.1 – Outil d'essai de coincement.....	160
Figure CC.2 – Ouverture en forme de V par rapport à B.....	161
Figure CC.3 – Réussite/échec par rapport à la zone B	162
Figure CC.4 – Positionnement de l'outil de coincement.....	163
Figure CC.5 – Réussite/échec par rapport à la zone C entre le panneau tête de lit et le panneau pied de lit	164
Figure CC.6 – Réussite/échec par rapport à la zone C entre les barrières latérales scindées	165
Tableau 201.101 – Protection contre le piégeage du PATIENT	107
Tableau 201.102 – Protection contre les chutes involontaires de PATIENTS.....	127
Tableau 24 – Températures maximales admissibles pour le contact de la peau avec des PARTIES APPLIQUÉES de LITS MÉDICAUX.....	128
Tableau BB.1 – Exigences normatives et informatives pour différents ENVIRONNEMENTS D'APPLICATION 1 to 5	151

COMMISSION ELECTROTECHNIQUE INTERNATIONALE

APPAREILS ÉLECTROMÉDICAUX –

**Partie 2-52: Exigences particulières de sécurité de base
et de performances essentielles des lits médicaux**

AVANT-PROPOS

- 1) La Commission Electrotechnique Internationale (CEI) est une organisation mondiale de normalisation composée de l'ensemble des comités électrotechniques nationaux (Comités nationaux de la CEI). La CEI a pour objet de favoriser la coopération internationale pour toutes les questions de normalisation dans les domaines de l'électricité et de l'électronique. A cet effet, la CEI – entre autres activités – publie des Normes internationales, des Spécifications techniques, des Rapports techniques, des Spécifications accessibles au public (PAS) et des Guides (ci-après dénommés "Publication(s) de la CEI"). Leur élaboration est confiée à des comités d'études, aux travaux desquels tout Comité national intéressé par le sujet traité peut participer. Les organisations internationales, gouvernementales et non gouvernementales, en liaison avec la CEI, participent également aux travaux. La CEI collabore étroitement avec l'Organisation Internationale de Normalisation (ISO), selon des conditions fixées par accord entre les deux organisations.
- 2) Les décisions ou accords officiels de la CEI concernant les questions techniques représentent, dans la mesure du possible, un accord international sur les sujets étudiés, étant donné que les Comités nationaux de la CEI intéressés sont représentés dans chaque comité d'études.
- 3) Les Publications de la CEI se présentent sous la forme de recommandations internationales et sont agréées comme telles par les Comités nationaux de la CEI. Tous les efforts raisonnables sont entrepris afin que la CEI s'assure de l'exactitude du contenu technique de ses publications; la CEI ne peut pas être tenue responsable de l'éventuelle mauvaise utilisation ou interprétation qui en est faite par un quelconque utilisateur final.
- 4) Dans le but d'encourager l'uniformité internationale, les Comités nationaux de la CEI s'engagent, dans toute la mesure possible, à appliquer de façon transparente les Publications de la CEI dans leurs publications nationales et régionales. Toutes divergences entre toutes Publications de la CEI et toutes publications nationales ou régionales correspondantes doivent être indiquées en termes clairs dans ces dernières.
- 5) La CEI elle-même ne fournit aucune attestation de conformité. Des organismes de certification indépendants fournissent des services d'évaluation de conformité et, dans certains secteurs, accèdent aux marques de conformité de la CEI. La CEI n'est responsable d'aucun des services effectués par les organismes de certification indépendants.
- 6) Tous les utilisateurs doivent s'assurer qu'ils sont en possession de la dernière édition de cette publication.
- 7) Aucune responsabilité ne doit être imputée à la CEI, à ses administrateurs, employés, auxiliaires ou mandataires, y compris ses experts particuliers et les membres de ses comités d'études et des Comités nationaux de la CEI, pour tout préjudice causé en cas de dommages corporels et matériels, ou de tout autre dommage de quelque nature que ce soit, directe ou indirecte, ou pour supporter les coûts (y compris les frais de justice) et les dépenses découlant de la publication ou de l'utilisation de cette Publication de la CEI ou de toute autre Publication de la CEI, ou au crédit qui lui est accordé.
- 8) L'attention est attirée sur les références normatives citées dans cette publication. L'utilisation de publications référencées est obligatoire pour une application correcte de la présente publication.
- 9) L'attention est attirée sur le fait que certains des éléments de la présente Publication de la CEI peuvent faire l'objet de droits de propriété intellectuelle ou de droits analogues. La CEI ne saurait être tenue pour responsable de ne pas avoir identifié de tels droits de propriété et de ne pas avoir signalé leur existence.

La Norme internationale CEI 60601-2-52 a été établie par le sous-comité 62D: Appareils électromédicaux, du comité d'études 62 de la CEI: Equipements électriques dans la pratique médicale, et par le comité technique 173 de l'ISO: Produits d'assistance pour personnes en situation de handicap.

Elle est publiée sous double logo.

Cette première édition annule et remplace la première édition de la CEI 60601-2-38, parue en 1996, et son Amendement (1999). Elle constitue une révision technique.

Le texte de cette norme particulière est issu des documents suivants:

FDIS	Rapport de vote
62D/795/FDIS	62D/815/RVD

Le rapport de vote indiqué dans le tableau ci-dessus donne toute information sur le vote ayant abouti à l'approbation de cette norme particulière. A l'ISO, la norme a été approuvée par 20 membres P sur 20 ayant voté.

Cette publication a été rédigée selon les Directives ISO/CEI, Partie 2.

Dans la présente norme, les caractères d'imprimerie suivants sont utilisés:

- Exigences et définitions: caractères romains.
- *Modalités d'essais: caractères italiques.*
- Les indications de nature informative apparaissant hors des tableaux, comme les notes, les exemples et les références: petits caractères. Le texte normatif à l'intérieur des tableaux est également en petits caractères.
- TERMES DÉFINIS A L'ARTICLE 3 DE LA NORME GÉNÉRALE, DANS LA PRÉSENTE NORME PARTICULIÈRE OU COMME NOTÉS: PETITES MAJUSCULES.

Concernant la structure de la présente norme, le terme:

- "article" désigne l'une des dix-sept sections numérotées dans la table des matières, avec toutes ses subdivisions (par exemple, l'Article 7 inclut les paragraphes 7.1, 7.2, etc.);
- "paragraphe" désigne une subdivision numérotée d'un article (par exemple, 7.1, 7.2 et 7.2.1 sont tous des paragraphes appartenant à l'Article 7).

Dans la présente norme, les références à des articles sont précédées du mot "Article" suivi du numéro de l'article concerné. Dans la présente norme particulière, les références aux paragraphes utilisent uniquement le numéro du paragraphe concerné.

Dans la présente norme, la conjonction "ou" est utilisée avec la valeur d'un "ou inclusif", ainsi un énoncé est vrai si une combinaison des conditions quelle qu'elle soit est vraie.

Les formes verbales utilisées dans la présente norme sont conformes à l'usage donné à l'Annexe H des Directives ISO/CEI, Partie 2. Pour les besoins de la présente norme:

- "devoir" mis au présent de l'indicatif signifie que la satisfaction à une exigence ou à un essai est obligatoire pour la conformité à la présente norme;
- "il convient/il est recommandé" signifie que la satisfaction à une exigence ou à un essai est recommandée mais n'est pas obligatoire pour la conformité à la présente norme;
- "pouvoir" mis au présent de l'indicatif est utilisé pour décrire un moyen admissible pour satisfaire à une exigence ou à un essai.

Lorsqu'un astérisque (*) est utilisé comme premier caractère devant un titre, ou au début d'un titre d'alinéa ou de tableau, il indique l'existence d'une ligne directrice ou d'une justification à consulter à l'Annexe AA.

Une liste de toutes les parties de la série CEI 60601, présentées sous le titre général *Appareils électromédicaux*, est disponible sur le site web de la CEI.

Le comité a décidé que le contenu de cette publication ne sera pas modifié avant la date de maintenance indiquée sur le site web de la CEI sous "<http://webstore.iec.ch>" dans les données relatives à la publication recherchée. A cette date, la publication sera

- reconduite,
- supprimée,
- remplacée par une édition révisée, ou
- amendée.

IECNORM.COM : Click to view the full PDF of IEC 60601-2-52:2009

INTRODUCTION

En 1996, la CEI a publié la première édition de la norme particulière relative aux lits d'hôpitaux électriques, la CEI 60601-2-38. La publication a été élaborée en réponse à la demande sur le terrain d'une norme universelle traitant des dangers spécifiques à la sécurité des lits d'hôpitaux. Utilisée conjointement avec une ÉVALUATION DES RISQUES DU FABRICANT, la norme a été estimée comme étant la tendance actuelle pour l'établissement d'une référence de sécurité de base pour l'industrie.

Un amendement de la CEI 60601-2-38 publié en 1999 a reconnu la nécessité d'atténuer un RISQUE de piégeage du PATIENT dans les BARRIERES LATÉRALES, à nouveau en liaison avec l'utilisation de l'ÉVALUATION DES RISQUES DU FABRICANT. Bien que ceci ait amélioré la norme particulière, cet amendement était toujours centré sur les lits d'hôpitaux électriques, et n'a pas permis de prendre en compte les lits d'hôpitaux actionnés manuellement et les produits dans d'autres environnements médicaux.

En 2000, la norme EN 1970 (*Lits réglables pour les PERSONNES HANDICAPÉES – Exigences et méthodes d'essai*) a été publiée, traitant des lits utilisés pour atténuer ou compenser l'incapacité ou le handicap des PERSONNES HANDICAPÉES. Le domaine d'application de la présente norme a été élargi, conjointement avec la CEI 60601-2-38, mais après l'édition de l'Amendement 1 à la CEI 60601-2-38, l'opportunité s'est présentée de combiner les deux normes en une norme internationale commune.

Les travaux ayant commencé sur l'intégration, la CEI a ajusté sa position sur la SECURITÉ DE BASE et les PERFORMANCES ESSENTIELLES, en les intégrant dans la troisième édition de la CEI 60601-1. Il est par conséquent devenu nécessaire d'aligner la nouvelle norme sur la troisième édition. Un nouveau numéro a été donné à la norme particulière, à savoir CEI 60601-2-52, et l'alignement sur la troisième édition a débuté.

La présente norme particulière est par conséquent la réalisation d'un travail important d'alignement et d'ajustement de domaines d'application entre la CEI 60601-2-38, la EN 1970, et la troisième édition de la CEI 60601-1. Elle représente la tendance actuelle en matière de SECURITÉ DE BASE et de PERFORMANCES ESSENTIELLES des LITS MÉDICAUX utilisés pour atténuer les maladies des PATIENTS et le handicap des PERSONNES HANDICAPÉES. Il s'agit d'un effort entrepris par un groupe de travail mixte de la CEI et de l'ISO.

APPAREILS ÉLECTROMÉDICAUX –

Partie 2-52: Exigences particulières de sécurité de base et de performances essentielles des lits médicaux

201.1 Domaine d'application, objet et normes connexes

L'article 1 de la norme générale¹ s'applique avec les exceptions suivantes:

201.1.1 * Domaine d'application

Remplacement:

La présente Norme internationale s'applique à la SÉCURITÉ DE BASE et aux PERFORMANCES ESSENTIELLES des LITS MÉDICAUX destinés aux adultes, désignés ci-après sous le terme LITS MÉDICAUX, tels que définis en 201.3.212.

Si un article ou un paragraphe est spécifiquement destiné à être applicable uniquement à un LIT MÉDICAL ou uniquement aux SYSTÈMES EM, le titre et le contenu de cet article ou de ce paragraphe l'indiquent. Si cela n'est pas le cas, l'article ou le paragraphe s'applique à la fois au LIT MÉDICAL et aux SYSTÈMES EM, selon le cas.

Les DANGERS inhérents à la fonction physiologique prévue des LITS MÉDICAUX ou des SYSTÈMES EM dans le cadre du domaine d'application de la présente norme ne sont pas couverts par des exigences spécifiques contenues dans la présente norme, à l'exception de 7.2.13 et de 8.4.1 de la norme générale.

NOTE Voir aussi 4.2 de la Norme générale.

201.1.2 Objet

Remplacement:

L'objet de la présente norme particulière est d'établir des exigences particulières pour la SÉCURITÉ DE BASE et les PERFORMANCES ESSENTIELLES des LITS MÉDICAUX tels qu'ils sont définis en 201.3.212.

201.1.3 Normes collatérales

Addition:

La présente norme particulière se rapporte aux normes collatérales applicables listées dans l'Article 2 de la CEI 60601-1 et l'Article 2 de la présente norme particulière.

La CEI 60601-1-3, la CEI 60601-1-8 et la CEI 60601-1-10²⁾ ne s'appliquent pas. Toutes les autres normes collatérales publiées dans la série CEI 60601-1 s'appliquent telles qu'elles sont publiées.

¹⁾ La norme générale est la CEI 60601-1:2005, *Appareils électromédicaux – Partie 1: Exigences générales pour la sécurité de base et les performances essentielles*.

²⁾ CEI 60601-1-10 :2007, *Appareils électromédicaux – Partie 1-10: Exigences générales pour la sécurité de base et les performances essentielles – Norme collatérale: Exigences pour le développement des régulateurs physiologiques en boucle fermée*

201.1.4 Normes particulières

Remplacement:

Dans la série CEI 60601, des normes particulières peuvent modifier, remplacer ou supprimer des exigences contenues dans la norme générale et dans les normes collatérales en fonction de ce qui est approprié à l'APPAREIL EM considéré, et elles peuvent ajouter d'autres exigences de sécurité de base et de PERFORMANCES ESSENTIELLES.

Une exigence d'une norme particulière prévaut sur l'exigence correspondante de la norme générale.

Par souci de concision, la CEI 60601-1 est désignée dans la présente norme particulière par le terme "norme générale". Les normes collatérales sont désignées par leur numéro de document.

La numérotation des articles et paragraphes de la présente norme particulière correspond à celle de la norme générale avec le préfixe "201" (par exemple 201.1 dans la présente norme aborde le contenu de l'Article 1 de la norme générale) ou à celle de la norme collatérale applicable avec le préfixe "20x", où x est (sont) le (les) dernier(s) chiffre(s) du numéro de document de la norme collatérale (par exemple 202.4 dans la présente norme particulière aborde le contenu de l'Article 4 de la norme collatérale CEI 60601-1-2, 203.4 dans la présente norme particulière aborde le contenu de l'Article 4 de la norme collatérale CEI 60601-1-3, etc.). Les modifications apportées au texte de la norme générale sont spécifiées par l'utilisation des termes suivants:

"Remplacement" signifie que l'article ou le paragraphe de la norme générale ou de la norme collatérale applicable est remplacé complètement par le texte de la présente norme particulière.

"Addition" signifie que le texte de la présente norme particulière est complémentaire aux exigences de la norme générale ou de la norme collatérale applicable.

"Amendement" signifie que l'article ou le paragraphe de la norme générale ou de la norme collatérale applicable est modifié comme indiqué par le texte de la présente norme particulière.

Les paragraphes, figures ou tableaux qui sont ajoutés à ceux de la norme générale sont numérotés à partir de 201.101. Toutefois, en raison du fait que les définitions dans la norme générale sont numérotées 3.1 à 3.139, les définitions complémentaires dans la présente norme sont numérotées à partir de 201.3.201. Les annexes supplémentaires sont notées AA, BB, etc., et les alinéas supplémentaires aa), bb), etc.

Les paragraphes, figures ou tableaux qui sont ajoutés à ceux d'une norme collatérale sont numérotés à partir de 20x, où "x" est le chiffre de la norme collatérale, par exemple 202 pour la CEI 60601-1-2, 203 pour la CEI 60601-1-3, etc.

L'expression "la présente norme" est utilisée pour faire référence à la norme générale, à toutes les normes collatérales applicables et à la présente norme particulière, considérées ensemble.

Lorsque la présente norme particulière ne comprend pas d'article ou de paragraphe correspondant(e), l'article ou le paragraphe de la norme générale ou de la norme collatérale applicable, bien qu'il puisse être sans objet, s'applique sans modification; lorsqu'il est demandé qu'une partie quelconque de la norme générale ou de la norme collatérale applicable, bien que potentiellement pertinente, ne s'applique pas, cela est expressément mentionné dans la présente norme particulière.

201.2 Références normatives

NOTE Une liste de références informatives est donnée dans la bibliographie à la page 166.

L'article 2 de la norme générale s'applique avec l'exception suivante:

Addition:

CEI 60068-2-31:2008, *Essais d'environnement – Partie 2-31: Essais – Essai Ec: Choc lié à des manutentions brutales, essai destiné en premier lieu aux matériels*

201.3 Termes et définitions

Pour les besoins du présent document, les termes et définitions donnés dans la CEI 60601-1:2005 s'appliquent avec les exceptions suivantes:


NOTE Un index des termes définis est donné à partir de la page 167.

201.3.8

PARTIE APPLIQUÉE

Addition:

La PARTIE APPLIQUÉE comprend toutes les parties du LIT MÉDICAL qui sont à la portée du PATIENT, même si elles sont sous le SOMMIER (voir Figure 201.101)



IEC 2115/09

Légende:

1 Zone de la PARTIE APPLIQUÉE comprenant des ACCESSOIRES

Figure 201.101 – PARTIE APPLIQUÉE

201.3.76

PATIENT

Remplacement:

personne soumise à une procédure de nature médicale, ou PERSONNE HANDICAPÉE

3.131

*** ZONE DE PIÉGEAGE**

Addition:

emplacements au niveau desquels le corps d'un occupant de LIT MÉDICAL peut être emprisonné, enchevêtré, bloqué, ou coincé à l'intérieur ou entre des parties du LIT MÉDICAL, telles que les BARRIERES LATÉRALES, le PANNEAU TÊTE ET PIED DE LIT, le SOMMIER ou le matelas

Addition:

201.3.201

*** ENVIRONNEMENT D'APPLICATION 1**

soins intensifs dispensés dans un hôpital où une surveillance médicale continue et un contrôle constant sont exigés, et où la mise à disposition de systèmes/d'équipements d'assistance vitale utilisés dans les procédures de nature médicale est essentielle pour maintenir ou améliorer les fonctions vitales du PATIENT

201.3.202

*** ENVIRONNEMENT D'APPLICATION 2**

soins de courte durée dispensés dans un hôpital ou autre infrastructure médicale, où une surveillance et un contrôle médicaux sont exigés, et où les APPAREILS EM utilisés dans les procédures de nature médicale sont souvent fournis pour aider à maintenir ou à améliorer l'état du PATIENT

201.3.203

*** ENVIRONNEMENT D'APPLICATION 3**

soins de longue durée dispensés dans une infrastructure médicale, où une surveillance médicale est exigée et où un contrôle est fourni si nécessaire, et où les APPAREILS EM utilisés dans les procédures de nature médicale peuvent être fournis pour aider à maintenir ou à améliorer l'état du PATIENT

NOTE Ceci inclut une utilisation dans les maisons de soins, les centres de réadaptation et les centres de gériatrie.

201.3.204

*** ENVIRONNEMENT D'APPLICATION 4**

soins dispensés à domicile et pour lesquels les APPAREILS EM sont utilisés pour atténuer ou compenser les effets d'une blessure, d'une incapacité ou d'une maladie

NOTE Ceci exclut une utilisation dans tous les autres ENVIRONNEMENTS D'APPLICATION (par exemple maisons de soins, centres de réadaptation et centres de gériatrie), lorsqu'un LIT MÉDICAL est purement conçu pour l'ENVIRONNEMENT D'APPLICATION 4.

201.3.205

*** ENVIRONNEMENT D'APPLICATION 5**

soins externes (ambulatoires), dispensés dans un hôpital ou autre infrastructure médicale, sous surveillance médicale, et où les APPAREILS EM sont fournis pour les besoins des personnes malades, blessées ou handicapées, pour le traitement, le diagnostic ou la surveillance

201.3.206

SOULÈVE-LIT

mécanisme réglable en hauteur sur lequel un SOMMIER peut être monté

NOTE La combinaison d'un SOULÈVE-LIT et d'un LIT non-MÉDICAL compatible spécifié par le FABRICANT est considérée comme un LIT MÉDICAL.

201.3.207

PERSONNE HANDICAPÉE

personne présentant une ou plusieurs déficiences, une ou plusieurs incapacités, une ou plusieurs restrictions de participation, ou une combinaison de ces éléments

[ISO 9999:2007]

201.3.208**PANNEAU TÊTE ET PIED DE LIT**

ensemble(s) monté(s) sur le LIT MÉDICAL, qui identifie(nt) pour le PATIENT le bord de l'extrémité de la tête ou du pied du LIT MÉDICAL et/ou du SOMMIER

NOTE Ils peuvent être utilisés comme des poignées pour pousser un LIT MÉDICAL destiné à transporter des PATIENTS.

201.3.209**POTENCE DE LEVAGE**

ACCESSOIRE fixé à un LIT MÉDICAL et destiné à aider un PATIENT à se tenir lorsqu'il change de position, en fournissant un support de préhension au-dessus du PATIENT


201.3.210**SURMATELAS**

matelas (surface de support) supplémentaire, qui est destiné à être placé sur un matelas existant, et généralement utilisé pour un effet prophylactique ou thérapeutique

201.3.211**SOMMIER**

structure qui soutient un PATIENT (par exemple un matelas)

NOTE Il peut être articulé ou prendre diverses positions pour faciliter les différentes positions de thérapie, de diagnostic et de confort (Voir Figures 201.102 et 201.119 a) à 201.119 d))



IEC 2116/09

Légende

- 1 PANNEAU TÊTE DE LIT
- 2 Section relève-buste
- 3 Section siège
- 4 Section plicature genoux supérieure
- 5 Section plicature genoux inférieure
- 6 PANNEAU PIED DE LIT

**Figure 201.102 – LIT MÉDICAL, configuration générale
(exemple, présentation schématique uniquement)**

201.3.212

*** LIT MÉDICAL**

dispositif pour lequel l'UTILISATION PRÉVUE est le sommeil/repos, qui contient un SOMMIER, et qui est destiné à aider dans le diagnostic, la surveillance, la prévention, le traitement, l'atténuation des effets d'une maladie ou la compensation d'une blessure ou d'un handicap

NOTE 1 Un SOULÈVE-LIT et/ou un SOMMIER amovible en combinaison avec un LIT non-MÉDICAL compatible spécifié par le FABRICANT est (sont) également considéré(s) comme un LIT MÉDICAL.

NOTE 2 Sont exclus les dispositifs pour lesquels l'UTILISATION PRÉVUE est principalement l'examen ou le transport sous surveillance médicale (par exemple brancard, table d'examen).

201.3.213

COMMANDE DE VERROUILLAGE DES MOUVEMENTS

sous-système auxiliaire qui désactive les commandes des mouvements

201.3.214

COMMANDE SUSPENDUE

dispositif portatif, qui a une CONNEXION FONCTIONNELLE avec le LIT MÉDICAL, contrôlant au moins les articulations et/ou les mouvements du LIT MÉDICAL

NOTE Les COMMANDES SUSPENDUES peuvent être à fils ou sans fil, et peuvent intégrer d'autres fonctions (par exemple communications, radio/tv, etc.).

201.3.215

BARRIÈRE LATÉRALE

barrière physique pouvant être un ACCESSOIRE amovible, ou faisant partie intégrante de la construction globale d'un LIT MÉDICAL, et qui est montée sur le(s) côté(s) du LIT MÉDICAL

NOTE Lorsqu'une BARRIÈRE LATÉRALE est en position "haute", elle fournit une barrière physique qui est destinée à réduire le RISQUE pour le PATIENT de glisser accidentellement du matelas ou d'en tomber en se tournant.

201.3.216

MATELAS THÉRAPEUTIQUE

matelas destiné à un effet prophylactique ou thérapeutique

201.3.217

PLANCHE D'ESSAI

planche de chargement plate et rigide, de dimensions spécifiées par le FABRICANT, représentant les dimensions du LIT MÉDICAL

201.3.218

CHASSIS

ensemble des composants du LIT MÉDICAL ou du SOULÈVE-LIT en dessous du SOMMIER

201.4 Exigences générales

L'article 4 de la norme générale s'applique.

201.5 Exigences générales relatives aux essais des APPAREILS EM

L'article 5 de la norme générale s'applique avec les exceptions suivantes:

Paragraphes complémentaires:

201.5.101 * Outils d'essai de piégeage

Les Figures 201.103a et 201.103b montrent des outils d'essai de piégeage (respectivement un outil conique et un outil cylindrique).

Dimensions en millimètres




Figure 201.103a – Outil conique

Dimensions en millimètres




Figure 201.103b – Outil cylindrique

Figure 201.103 – Outils d'essai de piégeage

201.5.102 Patin de charge

Le patin de charge (voir Figure 201.104) est un objet circulaire rigide, de (355 ± 5) mm de diamètre, dont la face présente une courbure sphérique convexe d'un rayon de 800 mm, avec un rayon du bord frontal de 20 mm.




Figure 201.104 – Patin de charge

201.6 Classification des APPAREILS EM et des SYSTÈMES EM

L'article 6 de la norme générale s'applique avec l'exception suivante:

201.6.2 * Protection contre les chocs électriques

Addition:

Pour l'ENVIRONNEMENT D'APPLICATION 4, le LIT MÉDICAL doit être de CLASSE II.

201.7 Identification, marquage et documentation des APPAREILS EM

L'article 7 de la norme générale s'applique avec les exceptions suivantes:

201.7.2.2 Identification

Remplacement du premier alinéa:


Le LIT MÉDICAL doit porter un marquage indiquant le nom ou la marque commerciale et l'adresse du FABRICANT, la RÉFÉRENCE DU MODELE OU DU TYPE et les moyens permettant la traçabilité.

Les composants amovibles doivent porter un marquage indiquant le nom ou la marque commerciale et l'adresse du FABRICANT, la RÉFÉRENCE DU MODELE OU DU TYPE et les moyens permettant la traçabilité, à moins qu'une mauvaise identification ne présente pas de RISQUE inacceptable.

Paragraphes complémentaires:

201.7.2.2.101 * Marquage du poids maximal du PATIENT et de la CHARGE DE FONCTIONNEMENT EN SÉCURITÉ

Le LIT MÉDICAL doit porter le marquage du poids maximal correspondant du PATIENT (voir 201.9.8.3.1) et de la CHARGE DE FONCTIONNEMENT EN SÉCURITÉ (pour le symbole, voir la Figure 201.105).



IEC 2120/09

Figure 201.105 – Symbole graphique du poids maximal du PATIENT et de la CHARGE DE FONCTIONNEMENT EN SÉCURITÉ

Les parties amovibles d'un LIT MÉDICAL d'une masse supérieure à 20 kg doivent être marquées du symbole ISO 7000-1321 (2004-01) : .

201.7.2.2.102 Marquage pour les LITS MÉDICAUX lavables à l'aide de machines par un système de lavage automatique.

Les LITS MÉDICAUX destinés à être utilisés avec un système de lavage automatique doivent porter le texte suivant, pour les distinguer des LITS MÉDICAUX qui ne peuvent pas supporter de telles méthodes de nettoyage: « Avertissement: Pour les besoins de nettoyage, ce lit peut être utilisé avec un système de lavage automatique »

201.7.2.2.103 Marquage pour les LITS MÉDICAUX destinés à un lavage au jet

Les LITS MÉDICAUX destinés à être utilisés avec un lavage au jet doivent porter le texte suivant: «Avertissement: Pour les besoins de nettoyage, ce lit peut être utilisé avec un lavage au jet. »

201.7.2.2.104 Largeur de la structure du SOULÈVE-LIT

Si un SOULÈVE-LIT a une structure à largeur réglable, la plage de réglage doit être marquée, par exemple par un indicateur de mesure linéaire fixé sur les parties réglables.

201.7.2.2.105 * Marquage des matelas de rechange

Les LITS MÉDICAUX conçus pour avoir des matelas de rechange doivent porter l'indication suivante à titre d'avertissement: "Des matelas incompatibles peuvent créer des DANGERS. Lire les instructions d'utilisation", ou un symbole selon le cas, en un emplacement bien en vue sur le SOMMIER, indiquant les matelas compatibles (voir exemple en AA.1).

201.7.2.2.106 * Marquage des BARRIÈRES LATÉRALES amovibles

Les LITS MÉDICAUX conçus pour avoir des BARRIÈRES LATÉRALES amovibles doivent porter l'indication suivante à titre d'avertissement: "Des BARRIÈRES LATÉRALES incompatibles peuvent créer des DANGERS. Lire les instructions d'utilisation" ou utiliser un symbole selon le cas, en un emplacement bien en vue près du point de fixation de la BARRIÈRE LATÉRALE, indiquant les BARRIÈRES LATÉRALES compatibles (voir exemple en AA.2).

201.7.2.4 ACCESSOIRES

Addition:

Lorsqu'une surcharge sur un ACCESSOIRE destiné à supporter des charges peut créer un RISQUE inacceptable, la CHARGE DE FONCTIONNEMENT EN SÉCURITÉ correspondante doit être marquée sur l'ACCESSOIRE.

La conformité est vérifiée par examen.

201.7.4.2 Dispositifs de commande

Addition:

Si la sécurité destinée à empêcher un mouvement involontaire du SOMMIER est assurée par une COMMANDE DE VERROUILLAGE DES MOUVEMENTS qui exige une activation par l'OPÉRATEUR, ceci doit être indiqué par des marquages ou des symboles selon le cas sur l'extérieur du LIT MÉDICAL et qui sont visibles depuis une position d'UTILISATION NORMALE (voir la Figure 201.106).

NOTE Exemple: "Engager la COMMANDE DE VERROUILLAGE DES MOUVEMENTS si un PATIENT peut être blessé en raison de déplacements involontaires du SOMMIER".


La conformité est vérifiée par un examen du LIT MÉDICAL.

201.7.6.3 Symboles pour les organes de commande et les performances

Addition:

Les commandes et/ou les voyants lumineux doivent être marqués, quand cela est possible, en utilisant les symboles qui traduisent la fonction voulue de ces commandes ou voyants lumineux sans nécessiter de texte complémentaire.

NOTE La Figure 201.106 est destinée à servir de ligne directrice lors de la conception de ces symboles. Dans tous les cas où des symboles internationaux normalisés existent, il convient de les utiliser.



IEC 2121/09

Légende

- 1 Fonction abaissement du relève-buste
- 2 Fonction élévation du relève-buste
- 3 Fonction abaissement jambes
- 4 Fonction relève-jambes
- 5 Fonction abaissement SOMMIER
- 6 Fonction élévation SOMMIER
- 7 Fonction Trendelenburg
- 8 Fonction anti-Trendelenburg/Trendelenburg inverse
- 9 Symbole de cadenas (déverrouillé) pour la COMMANDE DE VERROUILLAGE DES MOUVEMENTS
- 10 Symbole de cadenas (verrouillé) pour la COMMANDE DE VERROUILLAGE DES MOUVEMENTS

**Figure 201.106 – Commandes de fonctions et/ou actionneurs du LIT MÉDICAL:
lignes directrices pour la création des symboles graphiques**

201.7.9.2 Instructions d'utilisation

201.7.9.2.1 Généralités

Addition:

Les instructions d'utilisation doivent inclure:

- a) une description conforme à 201.3. de l' (des) ENVIRONNEMENT(S) D'APPLICATION prévu(s);
- b) le poids maximal du PATIENT et la CHARGE DE FONCTIONNEMENT EN SECURITÉ. La CHARGE DE FONCTIONNEMENT EN SECURITÉ est la somme des éléments suivants:
 - le PATIENT;
 - le matelas;
 - les ACCESSOIRES du LIT MÉDICAL (uniquement s'ils sont soutenus par le système de support du LIT MÉDICAL); et
 - la charge supportée par ces ACCESSOIRES (à l'exclusion du poids du PATIENT);
- c) une explication de la manière de désactiver toute fonction du LIT MÉDICAL, si un déplacement causé par cette fonction pouvait blesser le PATIENT;
- d) pour les LITS MEDICAUX destinés à l'ENVIRONNEMENT D'APPLICATION 4, le résultat de la mesure de l'énergie acoustique audible conformément à ISO 3746.

201.7.9.2.2 Avertissement et consignes de sécurité

Addition:

- a) Les instructions d'utilisation doivent contenir une mise en garde indiquant qu'il convient que le LIT MÉDICAL soit laissé dans sa position la plus basse en l'absence de surveillance du PATIENT, de manière à réduire le RISQUE de blessure due aux chutes.
- b) Les instructions d'utilisation doivent contenir une mise en garde sur les DANGERS causés par une manipulation inappropriée du CABLE D'ALIMENTATION, par exemple par des vrillages, des cisaillements ou autres dommages mécaniques.
- c) Les instructions d'utilisation doivent contenir une mise en garde, indiquant que, lorsqu'on achemine des câbles provenant d'autres équipements dans le LIT MÉDICAL, des précautions doivent être prises pour éviter de les écraser entre des parties du LIT MÉDICAL.
- d) Les instructions d'utilisation doivent contenir une mise en garde si un LIT MÉDICAL ne doit être utilisé qu'avec certains dispositifs de hissage, en raison de l'espace limité sous le LIT MÉDICAL.

201.7.9.2.5 Description de l'APPAREIL EM

Addition:

201.7.9.2.5.101 Choix du matelas

Les instructions d'utilisation doivent contenir des informations sur le choix du ou des matelas, y compris les dimensions du matelas et les caractéristiques du matelas (par exemple pour réduire le RISQUE de piégeage et de chutes (voir aussi 201.7.2.2.105)).

201.7.9.2.5.102 Choix des BARRIERES LATERALES

Les instructions d'utilisation doivent contenir des informations sur le choix des BARRIERES LATERALES, y compris les dimensions des BARRIERES LATERALES et les caractéristiques des BARRIERES LATERALES (par exemple pour réduire le RISQUE de piégeage et de chutes (voir aussi 201.7.2.2.106)).

201.7.9.2.5.103 * Angles et hauteur du LIT MÉDICAL

Les instructions d'utilisation doivent indiquer les angles maximaux, pouvant être réalisés par rapport à l'horizontale et en UTILISATION NORMALE par chaque partie du SOMMIER. Elles doivent également indiquer les hauteurs maximale et minimale par rapport au sol, pouvant être atteintes par le SOMMIER en UTILISATION NORMALE. Elles doivent également indiquer toutes les positions d'urgence et les commandes par lesquelles de telles positions sont obtenues.

201.7.9.2.5.104 * Masse maximale du LIT MÉDICAL

Les instructions d'utilisation doivent identifier la masse maximale (en kg) du LIT MÉDICAL. Si le LIT MÉDICAL est destiné à être démonté en plusieurs parties, la masse maximale (en kg) de toutes les parties doit être indiquée dans les instructions d'utilisation.

201.7.9.2.13 Maintenance*Remplacement:*

Les instructions d'utilisation doivent informer l'OPÉRATEUR ou l'ORGANISME RESPONSABLE de manière suffisamment détaillée sur les questions concernant l'examen préventif, la maintenance et la vérification qu'il doit réaliser, y compris la fréquence de cette maintenance.

Les instructions d'utilisation doivent fournir les informations pour la réalisation en toute sécurité d'une telle maintenance de routine nécessaire pour garantir une utilisation continue en sécurité du LIT MÉDICAL.

De plus, les instructions d'utilisation doivent identifier les parties sur lesquelles l'examen préventif et la maintenance doivent être réalisés par le PERSONNEL D'ENTRETIEN, y compris la périodicité d'application et les détails relatifs à la réalisation concrète d'une telle maintenance.

Pour les LITS MÉDICAUX contenant des batteries rechargeables qui sont destinées à une maintenance par d'autres personnes que le PERSONNEL D'ENTRETIEN, les instructions d'utilisation doivent contenir des informations permettant d'assurer une maintenance adéquate.

201.8 Protection contre les DANGERS d'origine électrique provenant des APPAREILS EM

L'article 8 de la norme générale s'applique avec les exceptions suivantes:

201.8.11.3.2 Types*Remplacement:*

Les CABLES D'ALIMENTATION doivent avoir une longueur minimale de 2,5 m, de la prise au périmètre extérieur du LIT MÉDICAL.

Les CABLES D'ALIMENTATION et autres câbles et cordons secteur flexibles externes des LITS MÉDICAUX doivent être de type HD22.10 H05-BQ-F [1]³⁾ ou d'une qualité équivalente pour la robustesse mécanique.

Les CABLES D'ALIMENTATION doivent être équipés d'une compensation de traction et d'une protection relative aux coudes appropriées.

³⁾ Les chiffres entre crochets se réfèrent à la bibliographie.

Les ensembles de CABLES D'ALIMENTATION doivent comporter une prise surmoulée ou tout autre dispositif pour résister à la pénétration d'eau au cours du PROCESSUS de nettoyage auquel le LIT MÉDICAL est destiné.

Le LIT MÉDICAL doit être équipé d'un dispositif permettant de dégager le CABLE D'ALIMENTATION de toute partie ou de tout mécanisme du LIT MÉDICAL en mouvement, afin d'en éviter la détérioration quand le LIT MÉDICAL est utilisé, déplacé ou n'est pas utilisé.

Les CABLES D'ALIMENTATION doivent être protégés de façon appropriée contre les dommages résultant d'un contact avec une ou plusieurs parties mobiles ou de frictions contre des coins tranchants et des arêtes vives à l'intérieur du LIT MÉDICAL.

La conformité est vérifiée par examen.

201.9 Protection contre les DANGERS MÉCANIQUES des APPAREILS EM ET SYSTÈMES EM

L'article 9 de la norme générale s'applique avec les exceptions suivantes:

201.9.1 * DANGERS MÉCANIQUES des APPAREILS EM

Addition:

201.9.1.101 * Protection contre un piégeage du PATIENT dans les parties non mobiles

Toutes les ouvertures ou les zones (A_1 , A_2 , A_3 , A_4 , A_5 , A_6 , B , C , et D) à l'intérieur du système du LIT MÉDICAL et qui sont au-dessus du SOMMIER doivent satisfaire aux exigences de dimensions et de construction des Figures 201.107, 201.108 et du Tableau 201.101. Lorsqu'un RISQUE de piégeage du PATIENT existe et est traité d'une autre façon, ceci doit être justifié par le FABRICANT dans le DOSSIER DE GESTION DES RISQUES.

La conformité est vérifiée avant et après l'application des essais de résistance de la BARRIÈRE LATÉRALE et de fiabilité du verrou (voir 201.9.8.3.3.3).


La conformité est vérifiée avec le SOMMIER à plat, sauf indication contraire figurant dans le Tableau 201.101. L'essai doit être réalisé avec la BARRIÈRE LATÉRALE dans toutes les positions de levage et de blocage.

Tous les essais sont effectués sans le matelas, à l'exception de l'essai pour la Dimension D. Les exigences impliquant le matelas sont vérifiées avec le(s) matelas spécifié(s) par le FABRICANT. Les exigences des Figures 201.107, 201.108 et du Tableau 201.101 impliquant le matelas sont exclues pour les MATELAS THÉRAPEUTIQUES.

Une ÉVALUATION DES RISQUES doit être réalisée pour évaluer:

- les MATELAS THÉRAPEUTIQUES;
- les SURMATELAS;
- les ACCESSOIRES;
- les positions articulées du SOMMIER.

La conformité est vérifiée par les essais suivants et par examen du DOSSIER DE GESTION DES RISQUES.



IEC 2122/09

Ne s'applique que lorsque la zone C ci-dessus est < 60 mm.

A_x représente les différentes zones A_1 , A_2 , A_3 , A_4 , A_5 et A_6

Légende

- 1 Zone de l'OUTIL représentant le diamètre du cou (60 mm).
- 2 Zone de l'OUTIL représentant la largeur du thorax (318 mm).
- 3 Zone de l'OUTIL représentant la largeur de la tête (120 mm).
- 4 SOMMIER
- 5 PANNEAU TÊTE DE LIT
- 6 PANNEAU PIED DE LIT
- 7 Matelas
- 8 BARRIÈRE LATÉRALE

Figure 201.107 – Exemple de LIT MÉDICAL avec une BARRIÈRE LATÉRALE segmentée ou scindée




Tableau 201.101 – Protection contre le piégeage du PATIENT

Zone	Description	Exigence/Méthode de conformité
A ₁	Ouvertures complètement fermées à l'intérieur d'une BARRIÈRE LATÉRALE, d'un PANNEAU TÊTE DE LIT ou d'un PANNEAU PIED DE LIT	L'espace est spécifié comme étant inférieur à 120 mm, tel que défini par l'essai suivant. <i>La conformité est vérifiée par l'essai suivant:</i> <i>A l'exception de A₃, articuler le LIT MÉDICAL et trouver l'ouverture la plus large. Insérer la partie de 60 mm de diamètre de l'outil conique (voir les Figures 201.103a, 201.107 et 201.108) à travers l'ouverture, depuis l'intérieur du système du LIT MÉDICAL. Placer l'outil conique sur l'ouverture concernée. Exercer une force de 250 N appliquée sur l'extrémité cylindrique de 60 mm de l'outil conique dans la direction la plus défavorable.</i> <i>Critère de réussite/d'échec: L'ouverture ne doit pas permettre à la partie de 120 mm de diamètre de l'outil conique d'entrer et de traverser.</i>
A ₂	Ouverture complètement fermée définie par la BARRIÈRE LATÉRALE, ses supports et le SOMMIER	
A ₃	Ouverture partiellement fermée définie par le PANNEAU TÊTE DE LIT, le SOMMIER et la BARRIÈRE LATÉRALE	
A ₄	Ouverture partiellement fermée définie par le PANNEAU PIED DE LIT, le SOMMIER et la BARRIÈRE LATÉRALE (sauf lorsque l'espace entre la BARRIÈRE LATÉRALE et le PANNEAU PIED DE LIT est >318 mm)	
A ₅	Ouverture partiellement fermée entre les BARRIÈRES LATÉRALES segmentées ou scindées et le SOMMIER (sauf lorsque l'espace entre les BARRIÈRES LATÉRALES est >318 mm)	
A ₆	Ouverture partiellement fermée définie par le point le plus bas d'une BARRIÈRE LATÉRALE, le support adjacent de la BARRIÈRE LATÉRALE et le SOMMIER sur l'extérieur des supports de la BARRIÈRE LATÉRALE.	
A	Autre(s) ouverture(s) définie(s) par les ACCESSOIRES (par exemple, tiges porte-sérum, cadres de traction) et les BARRIÈRES LATÉRALES, les PANNEAUX TÊTE ET PIED DE LIT, et/ou le SOMMIER. Ceci n'apparaît pas dans les Figures 201.107 et 201.108 car cela dépend de la position des ACCESSOIRES et de la construction du LIT MÉDICAL.	
B	Distance entre le SOMMIER et le point le plus bas de la BARRIÈRE LATÉRALE à l'extérieur des supports de la BARRIÈRE LATÉRALE. ET Angle entre la BARRIÈRE LATÉRALE et le SOMMIER dans la plage des hauteurs de matelas définie par le FABRICANT ± 2 cm. NOTE $\pm$ la dimension de 2 cm prend en compte la compression du matelas et la hauteur du cou au-dessus du matelas.	Espace <60 mm. ET Angle entre le SOMMIER et l'interface de la BARRIÈRE LATÉRALE >60° sur toute la plage des hauteurs de matelas, de la hauteur de matelas minimale recommandée, moins 2 cm, à la hauteur de matelas maximale recommandée, plus 2 cm. NOTE La GESTION DES RISQUES doit traiter de la possibilité d'utiliser un matelas non spécifié par le FABRICANT. Il convient que la GESTION DES RISQUES pour traiter la condition de piégeage de la zone B (illustrée sur la Figure AA.13) soit réalisée en prenant en considération ce qui suit: i) La forme et la géométrie de la BARRIÈRE LATÉRALE. ii) La distance entre le point le plus bas de la BARRIÈRE LATÉRALE et le SOMMIER. iii) Les propriétés du matériau du matelas. iv) Les dimensions du matelas. v) Le degré d'ajustement entre la BARRIÈRE LATÉRALE, le matelas et le SOMMIER. <i>La conformité est vérifiée par examen du DOSSIER DE GESTION DES RISQUES. Voir aussi l'Annexe CC pour plus d'informations.</i>
C ₁	Espace entre le PANNEAU TÊTE DE LIT et la BARRIÈRE LATÉRALE adjacente.	L'espace entre le PANNEAU TÊTE DE LIT et la BARRIÈRE LATÉRALE adjacente doit être <60 mm.

Zone	Description	Exigence/Méthode de conformité
		<i>La conformité pour un espace < 60 mm est vérifiée par l'essai suivant:</i> <i>L'outil cylindrique (voir Figure 201.103b) doit être orienté parallèlement au sol, avec l'angle le plus défavorable dans le plan horizontal au-dessus de l'espace. L'outil cylindrique de 60 mm doit rester avec le poids total sur l'espace où l'outil cylindrique traverse. Une force verticale supplémentaire ne doit pas être utilisée. L'outil cylindrique ne doit pas être utilisé pour démonter des parties du LIT MÉDICAL.</i> <i>Critère de réussite/d'échec:</i> <i>L'outil cylindrique de 60 mm ne doit pas se glisser dans l'ouverture. Voir l'Annexe CC pour plus de précisions.</i> <i>VOIR AUSSI L'ANNEXE CC POUR PLUS D'INFORMATIONS.</i>
C ₂	Espace entre les BARRIERES LATÉRALES segmentées ou scindées avec les deux BARRIERES LATÉRALES levées.	L'espace entre les BARRIERES LATÉRALES segmentées ou scindées avec les deux BARRIERES LATÉRALES levées doit être <60 mm ou >318 mm. <i>La conformité pour un espace < 60 mm est vérifiée par l'essai suivant:</i> <i>L'outil cylindrique (voir Figure 201.103b) doit être orienté parallèlement au sol, avec l'angle le plus défavorable dans le plan horizontal au-dessus de l'espace. L'outil cylindrique de 60 mm doit rester avec le poids total sur l'espace où le cylindre traverse. Une force verticale supplémentaire ne doit pas être utilisée. L'outil cylindrique ne doit pas être utilisé pour démonter des parties du LIT MÉDICAL.</i> <i>Pour un LIT MÉDICAL avec des BARRIERES LATÉRALES scindées, articuler le SOMMIER pour identifier l'ouverture la plus défavorable entre les BARRIERES LATÉRALES, et effectuer l'essai.</i> <i>Critère de réussite/d'échec:</i> <i>L'outil cylindrique de 60 mm ne doit pas se glisser dans l'ouverture. Voir l'Annexe CC pour plus de précision.</i> OU Pour un espace >318 mm: L'espace doit être > 318 mm pour l'ensemble de la distance verticale. <i>Voir l'Annexe CC pour plus d'informations</i>

Zone	Description	Exigence/Méthode de conformité
C ₃	Espace entre la BARRIÈRE LATÉRALE et le PANNEAU PIED DE LIT. Autre(s) ouverture(s) définie(s) par les ACCESSOIRES (par exemple, tiges porte-sérum, cadres de traction,.....) et les BARRIERES LATÉRALES, le PANNEAU TÊTE DE LIT, le PANNEAU PIED DE LIT, et/ou le SOMMIER.	L'espace entre la BARRIÈRE LATÉRALE et le PANNEAU PIED DE LIT doit être < 60 mm OU >318 mm. <i>La conformité pour un espace < 60 mm est vérifiée par l'essai suivant:</i> <i>L'outil cylindrique (voir Figure 201.103b) doit être orienté parallèlement au sol, avec l'angle le plus défavorable dans le plan horizontal au-dessus de l'espace. L'outil cylindrique de 60 mm doit rester avec le poids total sur l'espace où le cylindre traverse. Une force verticale supplémentaire ne doit pas être utilisée. L'outil cylindrique ne doit pas être utilisé pour démonter des parties du LIT MÉDICAL.</i> <i>Critère de réussite/d'échec:</i> <i>L'outil cylindrique de 60 mm ne doit pas se glisser dans l'ouverture. Voir l'Annexe CC pour plus de précisions.</i> OU Pour un espace >318 mm: L'espace doit être >318 mm pour l'ensemble de la distance verticale. VOIR L'ANNEXE CC POUR PLUS D'INFORMATIONS.
D	Région définie entre la BARRIÈRE LATÉRALE et le matelas.	<i>La conformité est vérifiée par l'essai suivant:</i> <i>Retirer le matelas de la BARRIÈRE LATÉRALE mesurée jusqu'à ce que le système de rétention du matelas ou la BARRIÈRE LATÉRALE opposée arrête le matelas. Tirer vers l'extérieur sur la BARRIÈRE LATÉRALE pour supprimer tout jeu latéral et, pendant l'application de la force, l'outil conique (voir Figure 201.103a) est placé avec son axe longitudinal parallèle à la BARRIÈRE LATÉRALE, sur le matelas dans l'espace horizontal entre la BARRIÈRE LATÉRALE et le matelas. Tourner l'outil conique jusqu'à ce que la ligne sur la surface de l'extrémité de diamètre 120 mm soit horizontale. Laisser l'outil conique s'enfoncer dans l'espace de son propre poids.</i> <i>Si un système de rétention du matelas, un support de BARRIÈRE LATÉRALE ou toute autre structure empêche l'outil conique de s'enfoncer dans l'espace, l'outil conique doit être placé à un endroit différent le long de la BARRIÈRE LATÉRALE, où il n'y a aucune perturbation.</i> <i>Critère de réussite/d'échec:</i> <i>L'extrémité large de l'outil conique ne doit pas s'enfoncer sous la surface du matelas de 50 % ou plus de son diamètre de 120 mm.</i> NOTE L'espace entre le PANNEAU TÊTE DE LIT et le SOMMIER est couvert par la mesure de A (voir la désignation A: autre(s) ouverture(s) définies par les ACCESSOIRES (par exemple, tiges porte-sérum, cadres de traction) et les BARRIERES LATÉRALES, les PANNEAUX TÊTE ET PIED DE LIT, et/ou le SOMMIER. Ceci n'est pas représenté sur la Figure 201.107 ni sur la Figure 201.108, car cela dépend de la construction du LIT MÉDICAL) Espace entre le PANNEAU TÊTE DE LIT et la tête du matelas: les matelas spécifiés par le FABRICANT n'ont normalement pas d'espaces appréciables entre le PANNEAU TÊTE DE LIT et l'extrémité de tête des matelas pour un piégeage possible de la tête.

201.9.2.2 ZONE DE PIÉGEAGE

201.9.2.2.1 Généralités

Addition:

La région entière dans le CHASSIS doit être considérée dans l'ANALYSE DES RISQUES en ce qui concerne les DANGERS de piégeage dus à des mouvements élevés/faibles.

La conformité est vérifiée par examen du DOSSIER DE GESTION DES RISQUES.


201.9.2.2.2 Espaces

Amendement:

Les emplacements identifiés dans les Figures 201.109 et 201.110 doivent être considérés comme des ZONES DE PIÉGEAGE pour les doigts.

Les distances entre les parties mobiles doivent toujours être soit inférieures à 8 mm, soit supérieures à 25 mm (telles que dimensionnées dans la Figure 201.109). La zone hachurée de 200 mm représente les zones de portée normale autour du périmètre du SOMMIER. La distance de 200 mm peut être mesurée en prenant en compte toute barrière empêchant l'accès pour les doigts (voir la Figure 201.110). La région située à l'intérieur de la PARTIE APPLIQUÉE et au-dessus du SOMMIER doit être considérée dans l'ANALYSE DES RISQUES en ce qui concerne l'espacement pour les doigts entre les parties mobiles.

IECNORM.COM : Click to view the full PDF of IEC 60601-2-52:2009




IEC 2124/09

Légende

- A L'espacement entre les parties mobiles doit être inférieur à 8 mm (sans ouverture) ou supérieur à 25 mm.
- B Zone hachurée représentant la zone de portée de piégeage pour les doigts de 200 mm à l'intérieur du bord extérieur.
- C PANNEAU TÊTE DE LIT
- D PANNEAU PIED DE LIT

Figure 201.109 – Espacement admissible pour les doigts dans les zones de portée normale autour du périmètre du SOMMIER




Légende

- 1 Point de pincement (ZONE DE PIÉGEAGE)
- 2 SOMMIER

Figure 201.110 – Exemple utilisant des barrières pour la mesure des distances d'isolement autour du périmètre du SOMMIER, pour réduire les risques de piégeage des doigts du PATIENT

Les emplacements identifiés dans les Figures 201.111 a) et 201.111 b) doivent être considérés comme des ZONES DE PIÉGEAGE pour les pieds.




Légende

Pour la zone où $b \leq 130$ mm

a doit toujours être ≥ 120 mm

NOTE La dimension " a " est uniquement mesurée à partir du sol. La dimension " b " est mesurée à partir du bord extérieur du LIT MÉDICAL. Ceci comprend tous les ACCESSOIRES fixés de façon permanente (par exemple BARRIÈRE LATÉRALE), si applicable.

Figure 201.111 a) – Zone de passage du pied et de l'orteil entre les parties mobiles et le sol



Légende

Pour la zone où b est comprise entre 130 mm et 180 mm

c doit toujours être ≥ 50 mm.

NOTE La dimension " c " est uniquement mesurée à partir du sol. La dimension " b " est mesurée à partir du bord extérieur du LIT MÉDICAL. Ceci comprend tous les ACCESSOIRES fixés de façon permanente (par exemple BARRIÈRE LATÉRALE), si applicable.

Figure 201.111b) – Zone de passage de l'orteil entre les parties mobiles et le sol

Figure 201.111 – Zones de passage

Les mesures doivent être réalisées dans la condition la plus défavorable pour les Figures 201.109, 201.110, 201.111 a) et la Figure 201.111 b).

201.9.2.2.3 Distances de sécurité

Addition:

Pour atteindre une main, la distance de sécurité est de 200 mm (voir la Figure 201.109 et la Figure 201.110).

201.9.2.2.5 Activation continue

Remplacement du point b):

- b) tout mouvement du LIT MÉDICAL ou de ses parties n'est possible que par l'activation du ou des dispositifs de commande qui initient et maintiennent le fonctionnement des éléments du LIT MÉDICAL uniquement si la commande manuelle est actionnée et lorsque la commande manuelle revient automatiquement à la position 'Stop' ou 'Arrêt' lorsqu'elle est relâchée.

NOTE Les mouvements actionnés par la main et le pied sont considérés comme satisfaisant au présent article, tant que la masse et la vitesse permettent un contrôle adéquat du positionnement sans causer un RISQUE inacceptable.

La conformité est vérifiée par examen et par des essais fonctionnels.

201.9.2.3.1 * Mouvement non désiré

Addition:

A l'exception des mouvements du LIT MÉDICAL en situation d'urgence, spécifiés par le FABRICANT, les dispositifs pour désactiver les mouvements des LITS MÉDICAUX (tels que ceux présents dans les commandes des fonctions des COMMANDES SUSPENDUES et/ou des BARRIÈRES LATÉRALES et destinés à être activés par le PATIENT, l'OPÉRATEUR ou le visiteur)

doivent être conçus de telle sorte qu'ils ne puissent pas être réactivés accidentellement par le PATIENT lorsque ce dernier est dans le LIT MÉDICAL.

La conformité est vérifiée par examen.

Les pédales de commande doivent être conçues pour empêcher une activation accidentelle.

Si la conception du LIT MÉDICAL ne peut pas exclure le piégeage ou l'écrasement, il doit y avoir un moyen pour désactiver toutes les pédales de commande utilisées pour tout mouvement du LIT MÉDICAL qui n'est pas actionné manuellement. La commande doit se désactiver sans action de l'OPÉRATEUR après utilisation. On doit prendre en considération une activation involontaire par le PATIENT ou d'autres personnes rampant sous le LIT MÉDICAL, ou par des objets utilisés à proximité.

Si des dispositifs de mouvement du LIT MÉDICAL actionné par le pied sont prévus, les dispositifs de désactivation d'un mouvement du LIT MÉDICAL actionné par le pied doivent être situés ou conçus de telle sorte que le PATIENT ne puisse pas réactiver accidentellement les fonctions, en prenant en compte la mobilité et la surveillance médicale du PATIENT.

La conformité est vérifiée par examen.

201.9.4.2 Basculement dû à l'instabilité

201.9.4.2.2 * Instabilité à l'exclusion du transport

Addition:

Lorsque les essais suivants sont effectués, le LIT MÉDICAL ne doit pas basculer (chavirer) avec la hauteur et la longueur du SOMMIER, les roulettes, les BARRIERES LATÉRALES et autres ACCESSOIRES avec leur CHARGE DE FONCTIONNEMENT EN SECURITÉ dans leur position la plus défavorable d'UTILISATION NORMALE, et le LIT MÉDICAL est équipé du matelas le plus léger spécifié par le FABRICANT ou d'une charge représentant le poids du matelas spécifié, qui est centrée sur le SOMMIER.

La conformité est vérifiée par les essais suivants réalisés avec le SOMMIER à plat et en position horizontale.

Essai de stabilité latérale:

Une charge de 2 200 N est placée sur le bord latéral du SOMMIER et est répartie uniformément sur une surface de 250 mm × 950 mm (voir la Figure 201.112).

Si la charge maximale du PATIENT selon le FABRICANT dépasse 2 200 N, la charge maximale du PATIENT doit être utilisée et est répartie uniformément sur une surface de 950 mm de long et de 250 mm de large (voir la Figure 201.112).

Effectuer l'essai à chaque coin du LIT MÉDICAL.

Essai de stabilité longitudinale:

aa) Si les PANNEAUX PIED DE LIT sont amovibles sans l'aide d'OUTILS:


Retirer le PANNEAU PIED DE LIT; une charge de 2 200 N pour les ENVIRONNEMENTS D'APPLICATION 1, 2, 3, 5 et de 1 850 N pour l'ENVIRONNEMENT D'APPLICATION 4 est répartie uniformément sur une surface de 250 mm sur toute la largeur du LIT MÉDICAL (voir la Figure 201.113).

Si la charge maximale du PATIENT selon le FABRICANT dépasse 2 200 N (ou 1 850 N pour l'ENVIRONNEMENT D'APPLICATION 4), la charge maximale du PATIENT doit être utilisée.

bb) Si les PANNEAUX TÊTE ET PIED DE LIT sont installés de façon permanente ou nécessitent l'utilisation d'OUTILS pour leur retrait:

Deux charges, chacune de 1 100 N, pour les ENVIRONNEMENTS D'APPLICATION 1, 2, 3, 5 et deux charges, chacune de 925 N, pour l'ENVIRONNEMENT D'APPLICATION 4, sont en même temps réparties uniformément sur une surface de 250 mm × 475 mm (Figure 201.114). Effectuer l'essai aux deux extrémités du LIT MÉDICAL.

Dimensions en millimètres




Légende

- 1 Charge de 2 200 N répartie uniformément ou supérieure si la charge maximale du PATIENT excède cette valeur.

Figure 201.112 – Essai de stabilité latérale sur les côtés du LIT MÉDICAL


Dimensions en millimètres



Légende

- 1 Charge de 2 200 N, (1 850 N pour l'ENVIRONNEMENT D'APPLICATION 4), répartie uniformément ou supérieure si la charge maximale du PATIENT excède cette valeur.

Figure 201.113 – Essai de stabilité longitudinale avec PANNEAU PIED DE LIT amovible



Légende

- 1 Deux charges de 1 100 N, (deux charges de 925 N pour l'ENVIRONNEMENT D'APPLICATION 4), réparties uniformément.

Figure 201.114 – Essai de stabilité longitudinale avec PANNEAUX TÊTE ET PIED DE LIT fixes

201.9.4.2.3 Instabilité due à des forces horizontales et verticales

Remplacement du point b):

- b) Le LIT MÉDICAL ne doit pas basculer (chavirer) si une personne s'assoit ou monte dessus.

NOTE Les exigences concernant les surfaces de support PATIENT sont données en 201.9.8.3.

La conformité est vérifiée par examen et par l'essai suivant:

Le LIT MÉDICAL est placé sur un plan horizontal et une force constante dirigée vers le bas de 1 100 N est appliquée au point du moment maximal à toute surface de travail, à l'exclusion du SOMMIER, offrant une prise évidente pour le pied ou une surface pour s'asseoir de 20 cm par 20 cm minimum, et à une hauteur ne dépassant pas 1 m du sol. Avant de conduire l'essai, le LIT MÉDICAL est préparé comme décrit en 201.9.4.2.2.

201.9.4.2.4.2 Force de propulsion

Addition:

Pour les LITS MÉDICAUX destinés au transport des PATIENTS:

- effectuer l'essai avec le LIT MÉDICAL, la CHARGE DE FONCTIONNEMENT EN SECURITÉ étant en place.

Pour les LITS MÉDICAUX non destinés au transport des PATIENTS:

- effectuer l'essai avec le LIT MÉDICAL, sans CHARGE DE FONCTIONNEMENT EN SECURITÉ en place.

201.9.4.2.4.3 * Franchissement d'un seuil

Remplacement:

Les LITS MÉDICAUX MOBILES destinés à transporter des PATIENTS doivent résister aux contraintes résultant d'une manipulation brutale. L'essai de seuil suivant doit être utilisé.

Cette exigence ne s'applique pas à un LIT MÉDICAL spécifié par le FABRICANT pour être déplacé uniquement dans la chambre du PATIENT pour le nettoyage ou l'accès des PATIENTS.

La conformité est vérifiée par l'essai suivant:

Les BARRIERES LATERALES sont levées et verrouillées, avec tous les autres ACCESSOIRES destinés à une UTILISATION NORMALE au cours du transport fixés sur le LIT MÉDICAL et avec la CHARGE DE FONCTIONNEMENT EN SECURITÉ en place et la hauteur dans la position la plus défavorable.

Le LIT MÉDICAL doit être déplacé à une vitesse de 0,8 m/s $\pm$ 0,1 m/s ou, pour les LITS MEDICAUX motorisés pour le transport, la vitesse maximale doit être utilisée, tandis que toutes les roulettes doivent venir heurter et franchir un obstacle fixé à plat au sol, et dont la section est rectangulaire, haute de 20 mm et profonde de 80 mm. Le LIT MÉDICAL, avec toutes les roulettes, doit ensuite franchir à nouveau l'obstacle et revenir au point de départ de l'essai.

Ceci est répété 10 fois.

Critères d'acceptation pour les LITS MEDICAUX MOBILES et motorisés:

A la fin de l'essai, le LIT MÉDICAL, les parties du LIT MÉDICAL et les ACCESSOIRES ne doivent présenter aucune perte de fonction, et aucun déverrouillage/débloccage des BARRIERES LATERALES, ni aucune détérioration physique (par exemple, détérioration des fixations, déverrouillage des ACCESSOIRES s'ils sont fixés sur le LIT MÉDICAL (comme par exemple les BARRIERES LATERALES), qui peut réduire l'UTILISATION NORMALE ou créer un RISQUE comme l'écrasement, ou une déformation permanente, modifiant l'espace pour le piégeage ou les pincements, etc.).

Le LIT MÉDICAL doit passer par-dessus l'obstacle. Le LIT MÉDICAL ne doit pas basculer (chavirer). Le LIT MÉDICAL ou les parties du LIT MÉDICAL ne doivent pas présenter de RISQUE inacceptable. Un RISQUE inacceptable est déterminé par examen du LIT MÉDICAL, de ses parties et des informations correspondantes dans le DOSSIER DE GESTION DES RISQUES.

201.9.4.3.1 * Instabilité pendant le transport

Remplacement du point c):

Un LIT MÉDICAL MOBILE ne doit pas entraîner de RISQUE inacceptable dû à un mouvement horizontal involontaire.

c) La conformité est vérifiée par l'essai suivant:

Avant l'essai, le LIT MÉDICAL est préparé avec les éléments suivants dans la position d'UTILISATION NORMALE la plus défavorable:

- hauteur, articulation et longueur du SOMMIER;*
- roulettes;*
- BARRIERES LATERALES;*
- ACCESSOIRES avec leur CHARGE DE FONCTIONNEMENT EN SECURITÉ en place, y compris en combinaison avec d'autres accessoires (par exemple, une plate-forme avec sa CHARGE DE FONCTIONNEMENT EN SECURITÉ, une potence de levage sans sa CHARGE DE FONCTIONNEMENT EN SECURITÉ);*
- matelas (par exemple hauteur et poids) spécifié par le FABRICANT, ou charge représentant le poids du matelas spécifié, qui est répartie uniformément et centrée sur le SOMMIER.*

Le LIT MÉDICAL MOBILE est placé avec la CHARGE DE FONCTIONNEMENT EN SECURITÉ en place, et le dispositif de verrouillage activé (par exemple des freins), sur un plan recouvert d'un matériau de revêtement en vinyle de 2 mm à 4 mm d'épaisseur, et incliné à 6° par rapport au plan horizontal sur un sol en béton. A la suite d'un mouvement élastique initial, d'une ligne de fuite initiale et d'un pivotement initial des roulettes, il ne

doit pas y avoir de déplacement du LIT MÉDICAL MOBILE supérieur à 50 mm (par rapport au plan incliné). Aucun mouvement initial ne doit donner lieu à un RISQUE inacceptable, en tenant compte de l'UTILISATION NORMALE du LIT MÉDICAL.

201.9.4.3.2 Instabilité à l'exclusion du déplacement

Remplacement du point a):

Voir 201.9.4.3.1.

201.9.4.4 Poignées et autres dispositifs de manutention

Ce paragraphe ne s'applique pas aux ENVIRONNEMENTS D'APPLICATION 1, 2, 3, 5.

201.9.6.2.1 Energie acoustique audible

Addition:

Les résultats de la mesure doivent être indiqués dans les instructions d'utilisation.

Amendement des critères de conformité:

La conformité est vérifiée en mesurant le niveau maximal de pression sonore de valeur A aux distances minimales séparant le PATIENT, l'OPÉRATEUR et d'autres personnes de la source d'énergie acoustique (bruit) en UTILISATION NORMALE, et si nécessaire, en calculant le niveau de pression sonore de valeur A produit par l'APPAREIL EM en conformité avec ISO 3746. Les conditions suivantes s'appliquent.

201.9.8 DANGERS associés aux systèmes de support

201.9.8.1 Généralités

Amendement:

Supprimer le premier tiret.

201.9.8.2 * FACTEUR DE SECURITÉ EN TRACTION

Le présent paragraphe ne s'applique pas (voir 201.9.8.3.2).

201.9.8.3 Résistance des supports PATIENT ou OPÉRATEUR ou des systèmes de suspension

201.9.8.3.1 Généralités

Remplacement:

Les parties des LITS MÉDICAUX servant à supporter ou à immobiliser des PATIENTS doivent être conçues de façon à réduire au minimum le RISQUE de blessures physiques et de relâchement accidentel des fixations.

Pour les ENVIRONNEMENTS D'APPLICATION 1,2, la CHARGE DE FONCTIONNEMENT EN SECURITÉ du LIT MÉDICAL doit être d'au moins 2 000 N. Elle est considérée comme étant la somme des charges minimales suivantes:

- 1 350 N, correspondant approximativement à une masse de 135 kg pour le PATIENT;
- 200 N, correspondant approximativement à une masse de 20 kg pour le matelas;

- 450 N, correspondant approximativement à une masse de 45 kg pour les ACCESSOIRES et pour la masse de la CHARGE DE FONCTIONNEMENT EN SECURITÉ supportée par ces ACCESSOIRES, mais à l'exclusion du poids du PATIENT.

Pour les ENVIRONNEMENTS D'APPLICATION 3, 4, 5, la CHARGE DE FONCTIONNEMENT EN SECURITÉ du LIT MÉDICAL doit être d'au moins 1 700 N. Elle est considérée comme étant la somme des charges minimales suivantes:


- 1 350 N, correspondant approximativement à une masse de 135 kg pour le PATIENT;
- 200 N, correspondant approximativement à une masse de 20 kg pour le matelas;
- 150 N, correspondant approximativement à une masse de 15 kg pour les ACCESSOIRES et pour la masse de la CHARGE DE FONCTIONNEMENT EN SECURITÉ supportée par ces ACCESSOIRES, mais à l'exclusion du poids du PATIENT.

La CHARGE DE FONCTIONNEMENT EN SECURITÉ d'un SOULÈVE-LIT doit être d'au moins 2 200 N. Elle est considérée comme étant la somme des charges minimales suivantes:

- 1 350 N, correspondant approximativement à une masse de 135 kg pour le PATIENT;
- 200 N, correspondant approximativement à une masse de 20 kg pour le matelas;
- 150 N, correspondant approximativement à une masse de 15 kg pour les ACCESSOIRES et pour la masse de la CHARGE DE FONCTIONNEMENT EN SECURITÉ supportée par ces ACCESSOIRES, mais à l'exclusion du poids du PATIENT;
- 500 N, correspondant approximativement à une masse de 50 kg pour les parties du LIT MÉDICAL destinées à être levées par le SOULÈVE-LIT.

Si la CHARGE DE FONCTIONNEMENT EN SECURITÉ spécifiée par le FABRICANT est supérieure à 2 000 N pour les LITS MÉDICAUX pour les ENVIRONNEMENTS D'APPLICATION 1 et 2, 1 700 N pour les LITS MÉDICAUX pour les ENVIRONNEMENTS D'APPLICATION 3, 4 et 5, ou 2 200 N pour les SOULÈVE-LITS, elle doit être utilisée comme base pour les essais.

La CHARGE DE FONCTIONNEMENT EN SECURITÉ doit être placée à la position la plus défavorable autorisée par la configuration ou la fixation des ACCESSOIRES sur les parties support/suspension. La CHARGE DE FONCTIONNEMENT EN SECURITÉ est répartie comme représenté sur la Figure 201.115.



IEC 2131/09

Légende

- 1 Section relève-buste
- 2 Section siège
- 3 Section plicature genoux supérieure
- 4 Section plicature genoux inférieure
- 5 SWL = CHARGE DE FONCTIONNEMENT EN SÉCURITÉ (SAFE WORKING LOAD)

Figure 201.115 – Répartition de la CHARGE DE FONCTIONNEMENT EN SÉCURITÉ pour les essais

Pour un repose-pieds, la partie de la CHARGE DE FONCTIONNEMENT EN SÉCURITÉ représentant la masse des PATIENTS est répartie sur une surface de 0,1 m² ou selon l'espace disponible.

NOTE La position du corps humain varie en fonction de la configuration du système de support/suspension et ainsi, la charge qui agit sur différentes sections variera, et il convient de la prendre en compte. Lorsqu'on analyse les forces et les couples de charge des ensembles supports, la partie de la CHARGE DE FONCTIONNEMENT EN SÉCURITÉ représente la masse définie.

La CHARGE DE FONCTIONNEMENT EN SECURITÉ de la POTENCE DE LEVAGE (y compris la poignée PATIENT) doit être d'au moins 750 N.

201.9.8.3.2 *Forces statiques dues à la charge des personnes

Remplacement:

Le LIT MÉDICAL et le SOULÈVE-LIT doivent être capables de résister à une charge statique uniformément répartie égale à deux fois la CHARGE DE FONCTIONNEMENT EN SECURITÉ ou 4 000 N, en prenant celle des deux valeurs qui est la plus élevée, dans la position la plus défavorable sur le SOMMIER en position horizontale.

Lorsqu'on peut s'attendre à une altération par usure, corrosion, fatigue du matériau ou vieillissement, le facteur de sécurité des éléments supports concernés ne doit pas être inférieur à quatre fois la CHARGE DE FONCTIONNEMENT EN SECURITÉ.

a) Pour les LITS MEDICAUX:

La conformité est vérifiée par l'essai suivant:

Une déformation permanente n'est acceptable que si le LIT MÉDICAL est conforme à sa fonction prévue. La charge statique doit être appliquée pendant au moins 1 min, à moins que le fluage du matériau puisse poser un problème, auquel cas la durée doit être augmentée à au moins 1 h.

b) Pour les SOULÈVE-LITS:

La conformité est vérifiée par l'essai suivant:

Monter une PLANCHE D'ESSAI sur un SOULÈVE-LIT non équipé d'une planche.

Placer le matelas, tel que spécifié par le FABRICANT, sur la planche/PLANCHE D'ESSAI, à plat.

Appliquer une charge verticale égale à deux fois la CHARGE DE FONCTIONNEMENT EN SECURITÉ ou 4 000 N, en prenant celle des deux valeurs qui est la plus élevée (à l'exclusion de la masse du matelas placé sur le LIT MÉDICAL ou de la masse de la PLANCHE D'ESSAI), répartie de façon égale sur le matelas.

Une déformation permanente n'est acceptable que si le SOULÈVE-LIT est conforme à sa fonction prévue. La charge statique doit être appliquée pendant au moins 1 min, à moins que le fluage du matériau puisse poser un problème, auquel cas la durée doit être augmentée à au moins 1 h.

Tous les ACCESSOIRES (y compris ceux ne supportant pas le poids du PATIENT) doivent être conçus pour supporter une charge égale à au moins deux fois la CHARGE DE FONCTIONNEMENT EN SECURITÉ spécifiée pour l'ACCESSOIRE. Cette charge est appliquée dans la direction et la position les plus défavorables sur l'ACCESSOIRE.

La conformité est vérifiée par l'essai suivant:

L'ACCESSOIRE (autre que les POTENCES DE LEVAGE) étant en position la plus défavorable rencontrée en UTILISATION NORMALE, y fixer pendant au moins 1 minute une charge statique égale à deux fois sa CHARGE DE FONCTIONNEMENT EN SECURITÉ, à moins que le fluage du matériau puisse poser un problème, auquel cas la durée doit être augmentée à au moins 1 heure.

Il ne doit pas apparaître de DANGER ou de perte de fonction.

Les fixations des POTENCES DE LEVAGE doivent toujours fonctionner normalement et ne présenter aucun DANGER après les essais suivants. Pour la POTENCE DE LEVAGE, une déformation permanente est acceptable dans le premier essai, mais un mouvement soudain de la POTENCE DE LEVAGE doit être considéré comme un DANGER:

1) *La conformité est vérifiée par l'essai suivant:*

Positionner la POTENCE DE LEVAGE par rapport au LIT MÉDICAL dans sa position la plus défavorable prévue pour l'utilisation.

Appliquer une charge vers le bas égale à deux fois la CHARGE DE FONCTIONNEMENT EN SECURITÉ des POTENCES DE LEVAGE (au moins 1 500 N), sur le point de suspension le plus externe pour la poignée pendant au moins 1 minute, à moins que le fluage du matériau puisse poser un problème, auquel cas la durée doit être augmentée à au moins 1 h.

Examiner la POTENCE DE LEVAGE et ses fixations pendant et après l'application de la charge.

2) *La conformité est vérifiée par l'essai suivant:*

Positionner la POTENCE DE LEVAGE par rapport au LIT MÉDICAL dans sa position la plus défavorable prévue pour l'utilisation. Fixer le LIT MÉDICAL pendant l'essai si nécessaire, en raison de l'instabilité.

Appliquer perpendiculairement au côté du LIT MÉDICAL une force horizontale de 350 N sur le point de suspension le plus externe pour la poignée. Examiner la POTENCE DE LEVAGE et ses fixations pendant et après l'application de la charge.

201.9.8.3.3 * Forces dynamiques dues à la charge des personnes

Remplacement:

201.9.8.3.3.1 Généralités


Lorsque des forces dynamiques (dus à une personne qui s'assoit ou qui se met debout, au PROCESSUS de traitement d'un PATIENT ou situations similaires) peuvent être exercées sur les parties d'un appareil destinées au support ou à la suspension d'un PATIENT en UTILISATION NORMALE, elles ne doivent pas engendrer de RISQUE inacceptable.

La durabilité doit être considérée en ce qui concerne la position la plus défavorable des parties de LIT MÉDICAL destinées au support ou à la suspension d'un PATIENT en UTILISATION NORMALE.

La conformité est vérifiée par les essais suivants:

Condition d'essai:

Mettre le matelas le plus défavorable spécifié par le FABRICANT sur le LIT MÉDICAL, celui-ci étant à plat. Si la hauteur est réglable, ajuster la hauteur sur la position la plus défavorable. Utiliser un patin de charge (Figure 201.104) à l'emplacement A de la Figure 201.116, sur le côté considéré comme étant le plus fragile.



Légende

- 1 Extrémité de la tête du LIT MÉDICAL
- 2 Extrémité du pied du LIT MÉDICAL

Figure 201.116 – Emplacement du patin de charge
(voir la Figure 201.104)

Méthode d'essai:

Appliquer le patin de charge (Figure 201.104) 10 000 fois à l'emplacement A indiqué sur la Figure 201.116, avec une charge de 1 350 N ou la charge maximale du PATIENT, en prenant celle des deux valeurs qui est la plus élevée.

Critères d'acceptation/Méthode de conformité:

Lorsqu'il est essayé conformément aux éléments ci-dessus, le LIT MÉDICAL doit toujours fonctionner normalement et ne présenter aucun DANGER après le retrait du patin de charge.

Retirer le matelas et réaliser ce qui suit:

- évaluer les DANGERS de piégeage conformément à 201.9.1.101;
- évaluer les ZONES DE PIÉGEAGE conformément à 201.9.2.2;
- réaliser l'essai conformément à 201.9.8.3.2.

201.9.8.3.3.2 * Essais dynamiques du mécanisme de réglage en hauteur

Le réglage en hauteur du LIT MÉDICAL ou du SOULÈVE-LIT doit fonctionner normalement et ne présenter aucun RISQUE inacceptable après 3 000 cycles en UTILISATION NORMALE.

La conformité est vérifiée par l'essai suivant:

Mettre le LIT MÉDICAL à plat. Si le SOULÈVE-LIT est séparé d'un SOMMIER, monter une PLANCHE D'ESSAI sur le SOULÈVE-LIT. Appliquer la CHARGE DE FONCTIONNEMENT EN SECURITÉ répartie comme indiqué en 201.9.8.3.1 (Figure 201.115) sur le SOMMIER. Pour les SOULÈVE-LITS, appliquer la CHARGE DE FONCTIONNEMENT EN SECURITÉ moins le poids de la PLANCHE D'ESSAI, répartie comme indiqué en 201.9.8.3.1 (Figure 201.115) sur le SOMMIER.

Le LIT MÉDICAL ou le SOULÈVE-LIT est levé et abaissé complètement 3 000 fois, conformément à la procédure indiquée dans les instructions d'utilisation. Après l'essai, retirer la CHARGE DE FONCTIONNEMENT EN SECURITÉ.

201.9.8.3.3.3 * Résistance de la BARRIÈRE LATÉRALE et fiabilité du verrou

Les blocages/verrouillages des BARRIERES LATÉRALES doivent rester sûrs lorsqu'ils sont soumis aux forces rencontrées en UTILISATION NORMALE.

La conformité est vérifiée par l'essai suivant: Les forces sont appliquées au niveau de l'emplacement accessible le plus défavorable pour le blocage/verrouillage:


Soumettre à des cycles le mécanisme de la BARRIÈRE LATÉRALE (de la position supérieure verrouillée à la position abaissée non verrouillée, puis de nouveau à la position supérieure verrouillée) pendant 30 000 cycles. Une force A ou B ou C ou D ou E ou F (telle que spécifiée à la Figure 201.117) doit ensuite être appliquée dans la position la plus défavorable pour le blocage/verrouillage de la BARRIÈRE LATÉRALE dans la direction de déblocage/déverrouillage. La BARRIÈRE LATÉRALE ne doit pas se débloquer/déverrouiller ni créer tout autre RISQUE inacceptable.

Les BARRIERES LATÉRALES doivent être conçues pour résister aux forces appliquées au cours d'un mauvais usage raisonnablement prévisible pendant le cycle de vie du produit, sans créer de RISQUE inacceptable.

La conformité est vérifiée par l'essai suivant avec la BARRIÈRE LATÉRALE en position supérieure:

- a) *Essai de cycles de force latérale: Exercer une force de 100 N, perpendiculairement à la BARRIÈRE LATÉRALE, sur la partie médiane supérieure de la BARRIÈRE LATÉRALE, dans la direction indiquée par E ou F sur la Figure 201.117. Inverser le sens. Répéter l'opération pendant 3 000 cycles.*
- b) *Essai de cycles de force longitudinale: Exercer une force de 100 N sur la BARRIÈRE LATÉRALE, dans le sens de la longueur de la BARRIÈRE LATÉRALE, comme indiqué par C ou D sur la Figure 201.117. Inverser le sens. Répéter l'opération pendant 3 000 cycles.*
- c) *Essai de cycles de force verticale: Exercer une force de 100 N sur la partie supérieure de la BARRIÈRE LATÉRALE, dans le sens vertical de la BARRIÈRE LATÉRALE, comme indiqué par B sur la Figure 201.117. Répéter l'opération pendant 3 000 cycles. Aucun essai n'est exigé dans la direction A.*
- d) *Une fois les étapes a), b) et c) ci-dessus effectuées, appliquer une charge statique dans les directions de la Figure 201.117, dans les positions les plus défavorables. La BARRIÈRE LATÉRALE ne doit pas se débloquer/déverrouiller ni créer tout autre RISQUE inacceptable.*

NOTE Le FABRICANT doit tenir compte du moment où ces forces peuvent être appliquées dans la position abaissée.



IEC 2133/09

Direction de la force	Force
A	500 N
B	750 N
C	500 N
D	500 N
E	500 N
F	500 N

Figure 201.117 – Application des forces pour l'essai de la BARRIÈRE LATÉRALE

201.9.8.3.3.4 Essais dynamiques de la POTENCE DE LEVAGE

La POTENCE DE LEVAGE et ses fixations doivent encore fonctionner normalement et ne présenter aucun DANGER après l'essai suivant. Un mouvement soudain de la POTENCE DE LEVAGE ou de sa poignée doit être considéré comme un DANGER.

La conformité est vérifiée par l'essai suivant:

Positionner la POTENCE DE LEVAGE par rapport au LIT MÉDICAL dans sa position la plus défavorable prévue pour l'utilisation.

Appliquer à la potence de levage une charge verticale, dirigée vers le bas, de la CHARGE DE FONCTIONNEMENT EN SECURITÉ (au moins 750 N) 1 000 fois sur la poignée de la POTENCE DE LEVAGE. Examiner la POTENCE DE LEVAGE et ses fixations pendant et après l'application de la force et enregistrer la déviation et la déformation.

La déviation de la POTENCE DE LEVAGE ne doit pas être supérieure à 100 mm pendant l'application de la CHARGE DE FONCTIONNEMENT EN SECURITÉ et la déformation permanente ne doit pas dépasser 20 mm après l'essai de durabilité, la mesure étant effectuée par rapport au SOMMIER.

201.9.8.4 Systèmes avec DISPOSITIFS DE PROTECTION MÉCANIQUE

201.9.8.4.1 Généralités

Ce paragraphe ne s'applique pas.

201.9.8.5 Systèmes sans DISPOSITIFS DE PROTECTION MÉCANIQUE

Ce paragraphe ne s'applique pas.


Paragraphe complémentaire:

201.9.101 *Protection contre les chutes involontaires de PATIENTS

Les BARRIÈRES LATÉRALES doivent être conçues avec des exigences de hauteur minimale, tel qu'indiqué par G de la Figure 201.109 et par le Tableau 201.102.

Lorsqu'un MATELAS THÉRAPEUTIQUE ou un SURMATELAS est utilisé et que la BARRIÈRE LATÉRALE ne satisfait pas à G, comme indiqué dans le Tableau 201.102, une ÉVALUATION DES RISQUES doit être effectuée pour garantir une sécurité équivalente.

La conformité est vérifiée par la mesure de G et par examen du DOSSIER DE GESTION DES RISQUES.



IEC 2134/09

Légende

- 1 SOMMIER
- 2 PANNEAU PIED DE LIT
- 3 PANNEAU TÊTE DE LIT
- 4 Matelas
- G Différence de hauteur entre le bord supérieur de la BARRIÈRE LATÉRALE et le dessus du matelas

Figure 201.118 – Hauteur de la BARRIÈRE LATÉRALE

Tableau 201.102 – Protection contre les chutes involontaires de PATIENTS

Désignation	Description	Exigence
G	La hauteur du bord supérieur de la BARRIÈRE LATÉRALE au-dessus du matelas non compressé représente au moins 50 % de la longueur du SOMMIER. D'autres positions de levage et de blocage ne satisfaisant pas à la hauteur ≥ 220 mm ou à la longueur d'au moins 50 % pour les BARRIÈRES LATÉRALES et spécifiées pour des objectifs différents de l'application des BARRIÈRES LATÉRALES (par exemple barre de mobilisation ou poignées) doivent être traitées dans le DOSSIER DE GESTION DES RISQUES du FABRICANT. Lorsque les BARRIÈRES LATÉRALES couvrent moins que la longueur totale du SOMMIER, les parties qui sont au-dessus d'une hauteur de 220 mm doivent être placées stratégiquement pour contribuer à éviter les chutes accidentelles du LIT MÉDICAL. Toutes les dimensions C du Tableau 201.101 doivent être respectées. NOTE Les parties de 220 mm de la barrière peuvent devoir être placées près des parties du corps qui sont les plus élevées (par exemple l'épaule pour un PATIENT couché sur le côté) et des parties qui pourraient amener à glisser du LIT MÉDICAL (jambes) afin d'éviter une chute accidentelle en se tournant ou en glissant. Voir la Figure 201.116, pour des lignes directrices sur la répartition de la masse corporelle.	≥ 220 mm <i>La conformité est vérifiée par des mesures</i>

201.10 Protection contre les DANGERS dus aux rayonnements involontaires ou excessifs

L'article 10 de la norme générale s'applique.

201.11 Protection contre les températures excessives et les autres DANGERS

L'article 11 de la norme générale s'applique avec les exceptions suivantes:

201.11.1.1 Température maximale en UTILISATION NORMALE

Amendement:

Remplacement du Tableau 24:

Tableau 24 – Températures maximales admissibles pour le contact de la peau avec des PARTIES APPLIQUÉES de LITS MEDICAUX

PARTIES APPLIQUÉES des LITS MEDICAUX		Température maximale ^{a b} °C		
		Métal et liquides	Verre, porcelaine, matière vitrifiée	Matière moulée, plastique, caoutchouc, bois
PARTIE APPLIQUÉE en contact avec le PATIENT pendant un temps "t"	$t < 1 \text{ min}$	51	56	60
	$1 \text{ min} \leq t < 10 \text{ min}$	48	48	48
	$10 \text{ min} \leq t$	41	41	41
^a Ces valeurs limites de température sont applicables pour la peau saine d'adultes, mais dans le cas de la limitation de température pour une durée supérieure à 10 min, les PERSONNES HANDICAPÉES ayant une peau sensible ou aucune sensation ont été également prises en compte. Elles ne sont pas applicables lorsque des surfaces importantes de la peau (10 % de la surface totale du corps ou plus) peuvent être en contact avec une surface chaude. Elles ne s'appliquent pas dans le cas de contact de la peau avec plus de 10 % de la surface de la tête. Le cas échéant, des limites appropriées doivent être déterminées et documentées dans le DOSSIER DE GESTION DES RISQUES. ^b Lorsqu'il est nécessaire que les PARTIES APPLIQUÉES dépassent les limites de température du Tableau 201.103 pour offrir un bénéfice clinique, le DOSSIER DE GESTION DES RISQUES doit contenir la documentation montrant que le bénéfice apporté dépasse toute augmentation associée du RISQUE.				

201.11.6.5 Pénétration d'eau ou de corps solides dans les APPAREILS EM et les SYSTÈMES EM

Addition:

201.11.6.5.101 Pénétration d'eau

Les ENVELOPPES du LIT MÉDICAL doivent assurer un degré minimal de protection de IPX4, conformément à la classification de la CEI 60529.

La conformité est vérifiée par les essais de la CEI 60529, le LIT MÉDICAL étant placé dans la position la moins favorable en UTILISATION NORMALE (comme défini dans les instructions d'utilisation) et par examen.

Toutes les parties qui peuvent être enlevées sans l'aide d'un OUTIL doivent être retirées (détachées/ouvertes) avant l'essai. L'essai doit être effectué sans matelas. Après ces procédures, le LIT MÉDICAL ne doit montrer aucun signe de contournement de l'isolation (ou des composants électriques) qui pourrait engendrer un DANGER en CONDITION NORMALE ou en combinaison avec une CONDITION DE PREMIER DÉFAUT (sur la base d'un examen visuel), puis les essais appropriés de tension de tenue et de COURANT DE FUITE sont réalisés.

Le LIT MÉDICAL doit être conforme à son UTILISATION PRÉVUE.

201.11.6.6 Nettoyage et désinfection des APPAREILS EM et des SYSTÈMES EM

Addition:

Si le LIT MÉDICAL est destiné à être utilisé avec des appareils pour le nettoyage et la désinfection, le FABRICANT doit spécifier le PROCESSUS de nettoyage et/ou de désinfection.

201.11.6.6.101 LIT MÉDICAL lavable à l'aide de machines

Les LITS MEDICAUX ou les SOULÈVE-LITS spécifiés par le FABRICANT pour être lavables à l'aide de machines par un système de lavage automatique doivent fonctionner normalement après l'essai.

Les variations par rapport à la procédure d'essai concernant le cycle de l'essai, la température, la durée et les liquides de nettoyage doivent être traitées dans le DOSSIER DE GESTION DES RISQUES du FABRICANT.

La présente méthode d'essai représente une procédure de base pour la désinfection d'un LIT MÉDICAL dans une machine de lavage.

La conformité est vérifiée par l'essai suivant:

- a) *les pièces et les capots d'accès qui peuvent être détachés/ouverts sans l'aide d'un OUTIL doivent être détachés/ouverts;*
- *un traitement de préconditionnement en température est appliqué pendant 10 jours à 65 °C $\pm$ 2 °C ou, si elle est supérieure, à la température maximale de stockage assignée;*
 - *le LIT MÉDICAL doit ensuite être conservé à la température ambiante pendant au moins 16 h;*
- b) *50 cycles d'essais, conformément à la procédure décrite par le FABRICANT dans les instructions d'utilisation, ou se composant de:*
- *2 min de lavage à une eau à 70 °C, à une valeur de pH comprise entre 5 et 8, avec 0,5 % de solution nettoyante et désinfectante, spécifiée par le FABRICANT;*
 - *20 s de rinçage à une eau à 85 °C, à une valeur de pH comprise entre 5 et 8, et avec 0,2 % de solution de rinçage à l'eau claire, conformément aux données du FABRICANT.*
 - *10 min de refroidissement à une température ambiante de 20 °C.*
 - *Critères d'acceptation:*
 - 1) *Immédiatement après les cycles d'essais, le LIT MÉDICAL est branché au réseau. Aucun mouvement involontaire ne doit se produire.*
 - 2) *Le LIT MÉDICAL doit fonctionner tel que spécifié par l'UTILISATION PRÉVUE, aux intervalles suivants:*
 - *immédiatement après les cycles d'essais;*
 - *5 min ($\pm$ 1 min) après les cycles d'essais;*
 - *60 min ($\pm$ 5 min) après les cycles d'essais;*
 - *24 h ($\pm$ 30 min) après les cycles d'essais;*
 - 3) *Effectuer les essais de tension de tenue et de COURANT DE FUITE conformément à 8.8.3 et 8.7, aux intervalles suivants:*
 - *immédiatement après les cycles d'essais;*
 - *24 h ($\pm$ 30 min) après les cycles d'essais.*
 - 4) *Effectuer un examen visuel pour la pénétration d'eau, pouvant entraîner un RISQUE inacceptable (c'est-à-dire court-circuit des barrières d'isolation et violation des lignes de fuite).*

NOTE Pour certains LITS MÉDICAUX, tous les critères d'acceptation ne s'appliquent pas (par exemple pour les LITS MÉDICAUX à commande manuelle, les critères d'acceptation électrique ne s'appliquent pas).

201.11.8 * Coupure de l'alimentation/du RÉSEAU D'ALIMENTATION vers l'APPAREIL EM

Addition:

En situation d'urgence, lorsque la TENSION RÉSEAU a été coupée, les fonctions de PERFORMANCES ESSENTIELLES définies par le FABRICANT doivent pouvoir être actionnées par d'autres moyens, et l'abaissement du relèvement-buste ou toute autre position d'urgence (spécifiée par le FABRICANT) doit pouvoir se faire en l'espace de 30 s.

La conformité est vérifiée par un essai fonctionnel dans les conditions les plus défavorables.

201.12 Précision des commandes et des instruments et protection contre les caractéristiques de sortie présentant des DANGERS

L'article 12 de la norme générale s'applique avec l'exception suivante:

201.12.2 APTITUDE A L'UTILISATION

Addition:

L'analyse de l'APTITUDE À L'UTILISATION doit prendre en compte les aspects de conception du LIT MÉDICAL, tels que la hauteur du SOMMIER, par rapport à l'OPÉRATEUR et au PATIENT, selon l'ENVIRONNEMENT D'APPLICATION.

NOTE Il convient que l'accès au PATIENT pour les besoins du traitement médical ne soit pas retardé de façon significative par les BARRIERES LATÉRALES et leurs systèmes de commande respectifs.

201.13 SITUATIONS DANGEREUSES et conditions de défaut

L'article 13 de la norme générale s'applique avec les exceptions suivantes:

201.13.1.4 * DANGERS MÉCANIQUES particuliers

Addition:

Une protection doit être prévue dans les commandes des mouvements du LIT MÉDICAL électrique, de telle sorte qu'une CONDITION DE PREMIER DÉFAUT ne provoque pas de mouvement sans une intervention humaine.

NOTE Une COMMANDE DE VERROUILLAGE DES MOUVEMENTS peut être utilisée comme un moyen de conformité. Une COMMANDE DE VERROUILLAGE DES MOUVEMENTS peut être activée automatiquement ou par l'OPÉRATEUR. (Voir aussi 201.7.4.2.).

Si un sous-système auxiliaire qui désactive les commandes des mouvements (COMMANDE DE VERROUILLAGE DES MOUVEMENTS) est prévu, il doit rester fonctionnel en CONDITION DE PREMIER DÉFAUT.

Pour les interrupteurs et autres composants qui commandent le mouvement des LITS MÉDICAUX électriques, des COMPOSANTS AUX CARACTÉRISTIQUES A HAUTE FIABILITE sont exigés si leur défaillance peut provoquer un DANGER.

Les sous-systèmes hydrauliques, pneumatiques et mécaniques des LITS MÉDICAUX ne doivent pas provoquer de mouvements inattendus du SOMMIER en CONDITION DE PREMIER DÉFAUT, si des mouvements de ce type peuvent entraîner un DANGER.

La protection de la CONDITION DE PREMIER DÉFAUT pour les LITS MÉDICAUX pneumatiques, hydrauliques ou mécaniques, ou pour les sous-systèmes pneumatiques, hydraulique ou mécaniques des LITS MÉDICAUX électriques, doit être satisfaite par conformité à 9.7.2 de la norme générale, et à d'autres articles applicables de la norme générale et de la présente norme relatifs aux facteurs de sécurité.

201.13.2.2 CONDITION DE PREMIER DÉFAUT électrique

Addition:

201.13.2.2.101 LIT MÉDICAL à déplacement motorisé

Les LITS MÉDICAUX à déplacement motorisé doivent avoir un dispositif pour permettre le transport par l'OPÉRATEUR en CONDITION DE PREMIER DÉFAUT du système de transport.

NOTE Un LIT MÉDICAL à déplacement motorisé comporte un système à propulsion.

201.14 SYSTÈMES ÉLECTROMÉDICAUX PROGRAMMABLES (SEMP)

L'article 14 de la norme générale s'applique.

201.15 Construction de l'APPAREIL EM

L'article 15 de la norme générale s'applique avec les exceptions suivantes:

201.15.3 Résistance mécanique

201.15.3.4.1 APPAREILS EM PORTATIFS

Remplacement:

La COMMANDE SUSPENDUE ne doit pas entraîner de RISQUE inacceptable à la suite d'une chute libre.

La conformité est vérifiée par l'essai suivant:

La COMMANDE SUSPENDUE doit tomber 1 000 fois, tel que spécifié dans la CEI 60068-2-31:2008.

Après l'essai, la COMMANDE SUSPENDUE est vérifiée par examen, par un essai fonctionnel et par les essais de sécurité électrique correspondants.

201.15.3.5 * Essai de manipulations brutales

Remplacement du point c):

c) Choc de chambranle (passage de porte)

Le LIT MÉDICAL est déplacé trois fois dans sa direction normale de déplacement, à une vitesse de $0,4 \text{ m/s} \pm 0,1 \text{ m/s}$, avec le SOMMIER dans la position la plus défavorable, à moins que la position de transport ne soit spécifiée par le FABRICANT, contre un obstacle vertical en bois dur d'une largeur et d'une épaisseur de 40 mm, fixé sur un support rigide vertical (par exemple du béton). La hauteur de l'obstacle vertical doit être au même niveau que le(s) point(s) de contact du LIT MÉDICAL. La direction du mouvement est perpendiculaire à la face de l'obstacle.

Ajouter à la fin de l'alinéa:

Cette exigence ne s'applique pas à un LIT MÉDICAL spécifié par le FABRICANT pour être déplacé uniquement dans la chambre du PATIENT pour le nettoyage ou l'accès des PATIENTS.

201.15.4 Composants et assemblage général des APPAREILS EM

201.15.4.4 * Voyants lumineux

Amendement:

Supprimer les deux premiers alinéas.

201.15.4.6.2 * Limitation du mouvement

Addition:

L'angle (désigné par " γ " sur la Figure 201.119) entre la section relève-buste et la section plicature genoux pour diverses configurations du SOMMIER doit toujours être supérieur à 90° dans des CONDITIONS NORMALES.

Dans l'ENVIRONNEMENT D'APPLICATION 4, l'angle α de la section relève-buste par rapport à l'horizontale doit toujours être supérieur ou égal à 0° .

Si un lit médical dans l'environnement d'application 1, 2, 3 ou 5 est prévu avec position de Trendelenburg, tout le SOMMIER doit pouvoir parvenir à un angle minimal de 12° de telle façon que la tête du PATIENT soit plus basse que le point de circulation central du corps.

La conformité est vérifiée en actionnant les commandes afin d'obtenir l'angle maximal γ , puis en mesurant cet angle.




Figure 201.119a – Angle γ entre la section relève-buste et la section relève-jambes du SOMMIER




Figure 201.119b – Angle γ entre la section relève-buste et la section plicature genoux supérieure du SOMMIER





Figure 201.119c – Angle γ entre la section relève-buste et la section plicature genoux supérieure du SOMMIER

**Légende**

- 1 Section relève-buste
- 2 Section jambes
- 3 Section plicature genoux supérieure
- 4 Section siège
- 5 Section plicature genoux inférieure

Figure 201.119d – Angle γ entre la section relève-buste et la section relève-jambes du SOMMIER

Figure 201.119 – Configurations du SOMMIER

201.15.4.7.1 Résistance mécanique

Amendement au point b):

Cette exigence ne s'applique pas si la pédale de commande est située d'une façon telle que le poids total de l'opérateur ne peut pas être placé sur elle.

La conformité est vérifiée par une application à la pédale de commande, dans sa position d'UTILISATION NORMALE, d'une force de manœuvre de 1 350 N pendant 1 min. La force est appliquée sur une zone de 30 mm de diamètre. Il ne doit y avoir aucun dommage au dispositif entraînant un RISQUE inacceptable.

Addition au point b):

Lorsque le poids total ne peut pas être placé sur les pédales de commande, le DOSSIER DE GESTION DES RISQUES doit alors spécifier les critères de conformité.

La conformité est vérifiée par un examen du DOSSIER DE GESTION DES RISQUES.

201.15.4.7.3 Pénétration de liquides

Amendement:

Les pédales de commande doivent être au moins IPX4.

Paragraphes complémentaires:

201.15.4.101 Ensemble PANNEAU TÊTE ET PIED DE LIT

La conception du LIT MÉDICAL doit permettre un accès immédiat et libre au PATIENT à partir de l'extrémité de la tête du LIT MÉDICAL dans des situations d'urgence. Ceci s'applique pour les ENVIRONNEMENTS D'APPLICATION 1 et 2 uniquement.

NOTE Ceci peut être obtenu par un PANNEAU TÊTE DE LIT, qui peut être retiré par une personne sans l'aide d'un OUTIL.

La conformité est vérifiée par l'essai suivant:

Un OPÉRATEUR ayant une formation appropriée doit pouvoir retirer le PANNEAU TÊTE DE LIT en l'espace de 15 s.

201.15.4.102 Rétention du matelas

Le LIT MÉDICAL doit être conçu pour maintenir le matelas en place en UTILISATION NORMALE, afin d'éviter qu'il ne glisse du SOMMIER. Une ÉVALUATION DES RISQUES doit être réalisée pour évaluer le mouvement maximal du matelas maintenu conformément à la zone "D" du Tableau 201.101.

La conformité est vérifiée par l'examen du DOSSIER DE GESTION DES RISQUES.

201.16 SYSTÈMES EM

L'article 16 de la norme générale s'applique.

201.17 Compatibilité électromagnétique des APPAREILS EM et des SYSTÈMES EM

L'article 17 de la norme générale s'applique.

Annexes

Les annexes de la norme générale s'appliquent.

Annexe AA (informative)

Guide particulier et justifications

AA.1 Guide général

En 1996, la CEI a publié la première édition de la norme particulière relative aux lits d'hôpitaux électriques, la CEI 60601-2-38. La publication a été élaborée en réponse à la demande sur le terrain d'une norme universelle traitant des DANGERS spécifiques à la sécurité des lits d'hôpitaux. Utilisée conjointement avec une ÉVALUATION DES RISQUES DU FABRICANT, la norme a été estimée comme étant la tendance actuelle pour l'établissement d'une référence de SECURITÉ DE BASE pour l'industrie.

Un amendement de la CEI 60601-2-38 publié en 1999 a reconnu la nécessité d'atténuer un RISQUE de piégeage du PATIENT dans les BARRIERES LATÉRALES, à nouveau en liaison avec l'utilisation de l'ÉVALUATION DES RISQUES DU FABRICANT. Bien que ceci ait amélioré la norme particulière, cet amendement était toujours centré sur les lits d'hôpitaux électriques, et n'a pas permis de prendre en compte les lits d'hôpitaux actionnés manuellement et les produits dans d'autres environnements médicaux.

En 2000, la norme EN 1970 (Lits réglables pour les PERSONNES HANDICAPÉES – Exigences et méthodes d'essai) a été publiée, traitant des lits utilisés par les PERSONNES HANDICAPÉES pour atténuer ou compenser une incapacité ou un handicap. Le domaine d'application de la présente norme a été élargi, conjointement avec la CEI 60601-2-38, mais après l'édition de l'Amendement 1 à la CEI 60601-2-38, l'opportunité s'est présentée de combiner les deux normes en une norme internationale commune.

Les travaux ayant commencé sur l'intégration, la CEI a ajusté sa position sur la SECURITÉ DE BASE et les PERFORMANCES ESSENTIELLES, les intégrant dans la troisième édition de la CEI 60601-1. Il est par conséquent devenu nécessaire d'aligner la nouvelle norme avec la troisième édition. Un nouveau numéro a été donné à la norme particulière, à savoir CEI 60601-2-52, et l'alignement sur la troisième édition a débuté.

La présente norme particulière est par conséquent la réalisation d'un travail important d'alignement et d'ajustement de domaines d'application entre la CEI 60601-2-38, la EN 1970, et la troisième édition de la CEI 60601-1. Elle représente la tendance actuelle en matière de SECURITÉ DE BASE et de PERFORMANCES ESSENTIELLES des LITS MÉDICAUX utilisés pour atténuer les maladies des PATIENTS et le handicap des PERSONNES HANDICAPÉES. Il s'agit d'un effort entrepris par un groupe de travail mixte de la CEI et de l'ISO.

AA.2 Justifications pour les articles et paragraphes particuliers

On trouvera ci-dessous les justifications des articles et paragraphes spécifiques de la présente norme particulière, les numéros des articles et des paragraphes correspondant à ceux utilisés dans le corps de la norme.

Paragraphe 201.1.1 – Domaine d'application

Le groupe de travail reconnaît que les définitions des termes "adulte" et "enfant" sont fondées sur des caractéristiques physiques qui varient d'un pays à l'autre. Si le niveau le plus élevé de sécurité doit être atteint pour les PATIENTS et les OPÉRATEURS, il est nécessaire de se fier au jugement professionnel du personnel soignant pour différencier les besoins des enfants de ceux des adultes en liaison avec l'APPAREIL EM, en prenant en compte non seulement les besoins physiques, psychologiques et médicaux de l'individu, mais aussi les préférences du PATIENT. Les exigences dimensionnelles de la présente norme particulière sont fondées sur