

Expression of performance of electrochemical analyzers –

Part 3: Electrolytic conductivity

CORRIGENDUM 1

Page 5

1 Scope

Delete, in the third dash, the three bibliographic references, including the footnote.

Page 6

3.2

In the note, add “= 1 MΩ · cm”, to the formula, in order to read:

$$1 \mu\text{S} \cdot \text{cm}^{-1} = 10^{-4} \text{ S} \cdot \text{m}^{-1} = 1 \text{ M}\Omega \cdot \text{cm}.$$

Page 7

3.7

In the formula, instead of “ Ω ” read “ Ψ ”.

Page 9

6.3 Test solutions

Replace, in the fourth paragraph, “...100 $\mu\text{S} \cdot \text{m}^{-1}$ ” by “...100 $\mu\text{S} \cdot \text{cm}^{-1}$ ”

Page 10

6.4 Test procedures

In item h), instead of “Errors caused by...” read “Uncertainties caused by...”.