

CORRIGENDUM 1

101.3.3 Protection by uncertified current limitation devices or by impedances

Replace the first paragraph of item a) as follows:

- a) An appropriate single component which is constructed, selected, and tested so that safety and reliability for protection against relevant HAZARDS is assured. In particular, the component shall:

Replace item b) 3) as follows:

- 3) meet the applicable CLEARANCE requirements of Annex K for REINFORCED INSULATION between the terminations of the combination of components.

Annex K (normative)

Insulation requirements not covered by 6.7

K.101.4.1 General

Delete note 3.

Table K.102

Replace the existing Table K.102 by the following:

101.3.3 Protection par des limiteurs de courant non certifiés ou des impédances

Remplacer le premier paragraphe du point a) par le suivant:

- a) Un composant individuel approprié qui doit être fabriqué, choisi et essayé pour assurer la sécurité et la fiabilité pour la protection contre les DANGERS appropriés. Ce composant doit notamment avoir:

Remplacer le point b) 3) par le suivant:

- 3) avoir la DISTANCE D'ISOLEMENT entre les extrémités de la combinaison de composants qui respecte les exigences de l'Annexe K pour l'ISOLATION RENFORCÉE.

Annexe K (normative)

Exigences d'isolation non couvertes par 6.7

K.101.4.1 Généralités

Supprimer la note 3.

Tableau K.102

Remplacer le Tableau K.102 existant par le suivant: