

# INTERNATIONAL STANDARD

AMENDMENT 1

**Radionuclide imaging devices – Characteristics and test conditions –  
Part 1: Positron emission tomographs**

IECNORM.COM: Click to view the full PDF of IEC 61675-1:1998/AMD1:2008





## THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2008 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester.

If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

IEC Central Office  
3, rue de Varembe  
CH-1211 Geneva 20  
Switzerland  
Email: [inmail@iec.ch](mailto:inmail@iec.ch)  
Web: [www.iec.ch](http://www.iec.ch)

### About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

### About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

- Catalogue of IEC publications: [www.iec.ch/searchpub](http://www.iec.ch/searchpub)

The IEC on-line Catalogue enables you to search by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, withdrawn and replaced publications.

- IEC Just Published: [www.iec.ch/online\\_news/justpub](http://www.iec.ch/online_news/justpub)

Stay up to date on all new IEC publications. Just Published details twice a month all new publications released. Available on-line and also by email.

- Electropedia: [www.electropedia.org](http://www.electropedia.org)

The world's leading online dictionary of electronic and electrical terms containing more than 20 000 terms and definitions in English and French, with equivalent terms in additional languages. Also known as the International Electrotechnical Vocabulary online.

- Customer Service Centre: [www.iec.ch/webstore/custserv](http://www.iec.ch/webstore/custserv)

If you wish to give us your feedback on this publication or need further assistance, please visit the Customer Service Centre FAQ or contact us:

Email: [csc@iec.ch](mailto:csc@iec.ch)  
Tel.: +41 22 919 02 11  
Fax: +41 22 919 03 00

IECNORM.COM: Click to view full PDF of IEC 61675-1:1999/AMD1:2008

# INTERNATIONAL STANDARD

AMENDMENT 1

**Radionuclide imaging devices – Characteristics and test conditions –  
Part 1: Positron emission tomographs**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

PRICE CODE

**D**

## FOREWORD

This amendment has been prepared by subcommittee 62C: Equipment for radiotherapy, nuclear medicine and radiation dosimetry, of IEC technical committee 62: Electrical equipment in medical practice.

The text of this amendment is based on the following documents:

Enquiry draft	Report on voting
62C/419/CDV	62C/432/RVC

Full information on the voting for the approval of this amendment can be found in the report on voting indicated in the above table.

The committee has decided that the contents of this amendment and the base publication will remain unchanged until the maintenance result date indicated on the IEC web site under "http://webstore.iec.ch" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

A bilingual version of this publication may be issued at a later date.

---

### Introduction to this amendment

Further developments of POSITRON EMISSION TOMOGRAPHS allow most of the tomographs to be operated in fully 3D acquisition mode. To comply with this trend, this amendment describes test conditions in accordance with the acquisition characteristic. It is the intention to simulate 3D imaging without introducing new phantoms or new acquisition or processing protocols. The test does simulate more realistically count rate characteristics for whole body imaging. Measurement of SCATTER FRACTION is not intended with this test. Certain parts of the standard are amended as stated below.

Page 17

#### 3.5.3 Method

*Add, after the first paragraph, the following new text:*

For scatter condition as simulated in 3.5.3.1.3, the total amount of ACTIVITY is the ACTIVITY in the body phantom.


### 3.5.3.1.3 Abdominal imaging

Replace the existing paragraph by the following:

The body phantom without any insert is homogeneously filled with ACTIVITY  $A_{ref}$  and centred in the FOV. The head phantom is filled with the same amount of ACTIVITY  $A_{ref}$  and is positioned in contact with the body phantom. The two arm phantoms are each filled with half the amount of ACTIVITY  $A_{ref}$  and are positioned in contact with the body phantom on the other side. All phantoms will rest on the patient couch. For details on the position of the phantoms and the location of the filling screws, see Figure 13. The arm phantoms are a first order approximation of the abdomen/ legs of the PATIENT.

Page 34

Add the following new Figure 13:



IEC 487/08

**Figure 13 – Phantom position and location of screws for abdominal imaging  
(see 3.5.3.1.3)**

### Index of defined terms

*Add, to the existing list, the following term:*

PATIENT.....rm-62-03

IECNORM.COM: Click to view the full PDF of IEC 61675-1:1998/AMD1:2008

**Withdrawn**