



**International
Standard**

ISO 10109

**Optics and photonics — Guidance
for the selection of environmental
tests**

*Optiques et photonique — Directives relatives au choix des essais
environnementaux*

**Third edition
2024-02**

STANDARDSISO.COM : Click to view the full PDF of ISO 10109:2024

STANDARDSISO.COM : Click to view the full PDF of ISO 10109:2024



COPYRIGHT PROTECTED DOCUMENT

© ISO 2024

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Environmental influences	3
4.1 General.....	3
4.2 Standard environment 1.....	3
4.3 Standard environment 2.....	4
4.4 Standard environment 3.....	4
4.5 Standard environment 4.....	5
4.6 Standard environment 5.....	5
4.7 Standard environment 6.....	5
4.8 Standard environment 7.....	6
4.9 Standard environment 8.....	6
4.10 Standard environment 9.....	7
4.11 Standard environment 10.....	7
4.12 Standard environment 11.....	7
5 Table for environmental tests and test parameters	7
Annex A (informative) Example of selection of environmental tests	13
Annex B (informative) Example of selection of standard environmental tests	20
Bibliography	22

STANDARDSISO.COM : Click to view the full PDF of ISO 10109:2024

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 172, *Optics and photonics*, Subcommittee SC 1, *Fundamental standards*.

This third edition cancels and replaces the second edition ISO 10109:2015 which has been technically revised.

The main changes are as follows:

- standard environments were revised and new standard environments were added.
- example of selection of standard environmental tests were added as [Annex B](#).

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Optical and photonic instruments including additional assemblies from other fields (e.g. mechanical, chemical, and electronic devices) are affected during their use by a number of different environmental and handling parameters, which they are required to resist without significant reduction in performance and to remain within defined specifications. This is what the manufacturer attempts to ensure and the user expects to receive.

This expectation can be assessed by exposure of the instrument to a range of simulated environmental parameters under controlled laboratory conditions. The cumulative combination, degree of severity, and sequence of these conditions can be selected to obtain meaningful results in a relatively short period of time.

Technical requirements as given in the tables of this document are abbreviated and the reader has to consult the referenced standards (i.e. the relevant part of ISO 9022) for the full specification of the technical requirement.

For the purposes of this document, nominal values for properties or performance characteristics are understood to be the manufacturer's internal technical data and do not directly reflect manufacturer's product specifications.

STANDARDSISO.COM : Click to view the full PDF of ISO 10109:2024

[STANDARDSISO.COM](https://standardsiso.com) : Click to view the full PDF of ISO 10109:2024

Optics and photonics — Guidance for the selection of environmental tests

1 Scope

This document contains tables for environmental tests and test parameters which can be used as a guideline for the selection of environmental tests. These include the selection of standardized tests according to ISO 9022 as well as additional parameters not described in ISO 9022 and necessary for the optical or photonic instruments. Ultimately, these tables specify the requirements to be met with regard to the reliability of the optical, mechanical, chemical, and electrical properties or performance characteristics of the instruments when exposed to environmental influences.

Environmental test methods, as specified in ISO 9022 (all parts), can be assigned to the various areas of application for the purpose of ascertaining the suitability of the instruments in the respective area of application.

2 Normative references

There are no normative references in this document.

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1

environmental requirement

specific natural and technical environmental influences between the limiting values of which optical instruments and instruments with optical, mechanical, chemical, and electrical components are to be operable

3.2

technical requirement

defined limiting value for the natural and technical environmental influences occurring in the envisaged area of application

Note 1 to entry: In order to verify whether an instrument fulfils a technical requirement, conditioning methods can be stipulated with degrees of severity whose limiting values are either higher or lower than those specified.

3.3

extent of testing

sum of all required tests to ascertain operability as well as product performance within the intended use and time of life

Note 1 to entry: The extent of design verification/testing is subdivided into

- design verification and production process controlling (not addressed by this document), and
- testing of functional models, prototypes, and production series products.

**3.4
area of application**

classification of the operability of an instrument within a defined environment and intended application

Note 1 to entry: The manufacturer defines the environment parameters. This can be done, e.g. by using pre-defined 'standardized areas of application' such as in [Table 3](#) to [Table 12](#) in full or in part.

EXAMPLE General application in an office environment with air conditioning; or an outdoor application, non-weather protected in arctic climate.

**3.5
conditioning method**

individual or combined environmental influence(s) to which the specimen is submitted during the test, e.g. shock or damp heat

**3.6
degree of severity**

parameter containing all the individual quantities required for the test

Note 1 to entry: See also ISO 9022-2 to ISO 9022-23.

EXAMPLE Temperature, humidity, conditioning time, and others have to be defined for the intended area of application.

**3.7
state of operation**

code that designates the state of operation of a specimen

Note 1 to entry: [Table 1](#) gives states of operation in accordance with ISO 9022 (all parts).

Table 1 — States of operation

State of operation	Comment
0	Specimen in its normal transport and/or storage container as provided by the manufacturer (transport container, shipping package).
1	Specimen is unprotected, ready for operation, power supply not connected.
2	Specimen is in operation during conditioning for a period to be specified in the relevant specification. The mode of the operating status is specified in the relevant specification. During operation, a check is performed to establish if the specimen is functioning as required.

**3.8
status after test**

code that designates the outcome of the test

Note 1 to entry: [Table 2](#) gives the status after test.

Table 2 — Status after test

Status after test	Comment
A	All performance criteria are satisfied.
B	All performance criteria are satisfied. Damage to parts not needed for function or reduced life are possible.
C	Not all performance criteria are satisfied. Damage to parts not needed for function or reduced life are possible.
D	Device may not operate anymore; damage is expected.

4 Environmental influences

4.1 General

Table 3 to Table 13 classify and describe standardized environment areas, which constitute a summary of the various environmental requirements with major influences to products during shipping, transport, storage, and operation.

The values for standard environments 1, 2, 7, 8, 9, 10, and 11 have been compiled from excerpts of IEC 60721-2-1, IEC 60721-3-4, MIL-STD-810H, and MIL-HDBK-310.

Standard environment 5 and 13 are defined for use in weather-protected locations.

Standard environment 2 and 6 are both defined as an open-air climate. The temperature range for standard environment 2 applies for states of operation 0 and 1. The restricted temperature range for standard environment 6 applies for state of operation 2. The restricted limiting values of standard environment 6 are for optical precision measurements and for the use of optical consumer products in the open air.

4.2 Standard environment 1

**Table 3 — Non-weather-protected locations with extremely cold climates
(Arctic or Antarctic climate)**

Environmental influence	Value	Comment
Temperature	-55 °C to +30 °C	This range applies for commonly experienced conditions. In extreme geographical conditions, temperatures as low as -65 °C and as high as 45 °C can occur in the location of use. With temporary or permanent storage in enclosed vehicles, sheds, hangars, or attics, temperatures of over 45 °C can be experienced in strong sunshine, and over 70 °C in extreme cases. The specification does not address these conditions.
Relative humidity	up to 100 %	
Air pressure	70 kPa to 106 kPa	50 kPa to 110 kPa in unfavourable conditions.
Total solar irradiance	up to 1,1 kW/m ²	Intensity of global radiation on earth's surface, depending on the latitude.
Amount of precipitation (rain, snow, or hail)	≤15 mm/min	
Dew or ice build-up	yes	

4.3 Standard environment 2

Table 4 — Global locations, non-weather-protected locations in basic regional type^a

Environmental influence	Value	Comment
Temperature ^b	-33 °C to +50 °C	This range applies for commonly experienced conditions. In extreme geographical conditions, temperatures as low as -45 °C and over 50 °C can be experienced in the location of use. With temporary or permanent storage in enclosed vehicles, sheds, hangars, or attics, temperatures of over 65 °C can occur in strong sunshine, and over 70 °C in extreme cases. The specification does not address these conditions.
Relative humidity	up to 100 %	
Air pressure	70 kPa to 106 kPa	50 kPa to 110 kPa in unfavourable conditions.
Total solar irradiance	up to 1,1 kW/m ²	Intensity of global radiation on earth's surface, depending on the latitude.
Amount of precipitation (rain, snow or hail)	≤5 mm/min	
Dew or ice build-up	yes	

^a Standard environment 8 and 9 have a temperature range up to +55 °C.

^b Testing for temperature range +50 °C is not currently a severity condition in ISO 9022-2.

4.4 Standard environment 3

Table 5 — Global locations, non-weather-protected, with maritime and/or coastal climate

Environmental influence	Value	Comment
Temperature	-20 °C to +35 °C	This range applies for commonly experienced conditions. Along coasts with icing, temperatures below -20 °C can be experienced, and above 35 °C along tropical coasts. With temporary or permanent storage in enclosed vehicles, sheds, hangars, or attics, temperatures over 50 °C can be experienced in strong sunshine, and over 70 °C in extreme cases. The specification does not address these conditions.
Relative humidity	up to 100 %	
Air pressure	90 kPa to 106 kPa	50 kPa to 110 kPa in unfavourable conditions.
Total solar irradiance	up to 1,1 kW/m ²	Intensity of global radiation on earth's surface, depending on the latitude.
Amount of precipitation (rain, snow or hail)	≤15 mm/min	
Dew or ice build-up	yes	

4.5 Standard environment 4

Table 6 — High altitudes of up to 30 000 m

Environmental influence	Value	Comment
Temperature	-65 °C to +55 °C	This range applies for commonly experienced conditions. Above polar zones, temperatures below -65 °C have to be expected. With temporary or permanent storage in enclosed vehicles, sheds, hangars, or attics, temperatures over 55 °C can be experienced in strong sunshine, and over 85 °C in extreme cases. The specification does not address these conditions.
Relative humidity	up to 100 %	
Air pressure	1 kPa to 106 kPa	
Total solar irradiance	up to 1,4 kW/m ²	Intensity of radiation at altitude of 30 000 m.
Amount of precipitation (rain, snow or hail)	≤15 mm/min	
Dew or ice build-up	yes	

4.6 Standard environment 5

Table 7 — Technical climate in weather-protected locations

Environmental influence	Value	Comment
Temperature	+10 °C to +35 °C	This range applies for commonly experienced conditions. In unfavourable conditions, temperatures below 10 °C and over 35 °C may occur. The specification does not address these conditions.
Relative humidity	up to 85 %	
Air pressure	70 kPa to 106 kPa	50 kPa to 110 kPa in unfavourable conditions.
Total solar irradiance	up to 0,9 kW/m ²	Without protection from sunshine, depending on the latitude.

4.7 Standard environment 6

Table 8 — Non-weather-protected locations with restricted limiting values

Environmental influence	Value	Comment
Temperature	-20 °C to +50 °C	The restricted temperature range applies for state of operation 2. The values of standard environment 2 apply for the states of operation 0 and 1.
Relative humidity	up to 100 %	
Air pressure	70 kPa to 106 kPa	50 kPa to 110 kPa in unfavourable conditions.
Total solar irradiance	up to 1,1 kW/m ²	The total solar irradiance depends on the latitude. Beware of critical values in and on the instruments when combining heat and solar irradiance.
Amount of precipitation (rain, snow or hail)	≤5 mm/min	
Dew or ice build-up	yes	

4.8 Standard environment 7

Table 9 — Non-weather-protected locations with cold climates

Environmental influence	Value	Comment
Temperature	-45 °C to +45 °C	This range applies for commonly experienced conditions. In extreme geographical conditions, temperatures as low as -65 °C and as high as 45 °C can occur in the location of use. With temporary or permanent storage in enclosed vehicles, sheds, hangars, or attics, temperatures of over 45 °C can be experienced in strong sunshine, and over 70 °C in extreme cases. The specification does not address these conditions.
Relative humidity	up to 100 %	
Air pressure	70 kPa to 106 kPa	50 kPa to 110 kPa in unfavourable conditions.
Total solar irradiance	up to 1,1 kW/m ²	Intensity of global radiation on earth's surface, depending on the latitude.
Amount of precipitation (rain, snow, or hail)	≤5 mm/min	
Dew or ice build-up	yes	

4.9 Standard environment 8

Table 10 — Non-weather-protected locations with dry hot climates

Environmental influence	Value	Comment
Temperature	-10 °C to +55 °C	This range applies for commonly experienced conditions. In extreme geographical conditions, temperatures as low as -20 °C and over 55 °C can be experienced in the location of use. With temporary or permanent storage in enclosed vehicles, sheds, hangars, or attics, temperatures of over 70 °C can occur in strong sunshine, and over 85 °C in extreme cases. The specification does not address these conditions.
Relative humidity	less than 40 %	
Air pressure	70 kPa to 106 kPa	50 kPa to 110 kPa in unfavourable conditions.
Total solar irradiance	up to 1,1 kW/m ²	Intensity of global radiation on earth's surface, depending on the latitude.
Dew or ice build-up	no	

4.10 Standard environment 9

Table 11 — Non-weather-protected locations with humid hot climates

Environmental influence	Value	Comment
Temperature	+10 °C to +55 °C	This range applies for commonly experienced conditions. In extreme geographical conditions, temperatures over 55 °C can be experienced in the location of use. With temporary or permanent storage in enclosed vehicles, sheds, hangars, or attics, temperatures of over 70 °C can occur in strong sunshine, and over 85 °C in extreme cases. The specification does not address these conditions.
Relative humidity	up to 100 %	
Air pressure	70 kPa to 106 kPa	50 kPa to 110 kPa in unfavourable conditions.
Total solar irradiance	up to 1,1 kW/m ²	Intensity of global radiation on earth's surface, depending on the latitude.
Amount of precipitation (rain, snow or hail)	≤5 mm/min	
Dew or ice build-up	yes	

4.11 Standard environment 10

Table 12 — High altitudes of up to 80 000 m

Environmental influence	Value	Comment
Temperature	-145 °C to +55 °C	This range applies for commonly experienced conditions. Temperatures below -145 °C can be experienced. With temporary or permanent storage in enclosed vehicles, sheds, hangars, or attics, temperatures over 55 °C can be experienced in strong sunshine, and over 85 °C in extreme cases. The specification does not address these conditions.
Relative humidity	up to 100 %	
Air pressure	0,2 Pa to 106 kPa	
Total solar irradiance	up to 1,4 kW/m ²	Intensity of radiation at altitude of 80 000 m.
Amount of precipitation (rain, snow or hail)	≤15 mm/min	
Dew or ice build-up	yes	

4.12 Standard environment 11

Table 13 — Temperature and humidity controlled laboratory

Environmental influence	Value	Comment
Temperature	+17 °C to +23 °C	
Relative humidity	30 % to 60 %	
Air pressure	70 Pa to 106 kPa	

5 Table for environmental tests and test parameters

The wide range of technologies applied in complex instruments often requires specific and individually selected test parameters, test sequences, and combined tests. [Table 14](#) is a collection of mainly climatic and mechanical tests.

The user of this document, e.g. a manufacturer, selects the tests and adapts the parameters to their specific equipment. A specific area of application must be defined for every item of equipment and both the tests and

ISO 10109:2024(en)

the degree of severity also must be adapted to the intended climatic and operating environment and to the customer expectations. It is up to the manufacturer to carefully select the appropriate tests, the combination of the tests, the required technical parameters, the degree of severity, the status of operation, the extent of testing and other parameters. A non-standard environment example is given in [Annex A](#) and a standard environment is given in [Annex B](#).

[Table 14](#) is a comprehensive list of nearly all the environmental tests that could potentially be used for optical applications. This includes the ISO 9022 series and IEC standards. It is not necessary to specify all the tests listed in [Table 14](#). It is most common to only note a subsection of the tests shown.

As a general test result, the status after the test can be shown in the table for each specific test run.

Note that in [Table 14](#), the combination of the three columns “Conditioning method”, “Degree of severity”, and “State of operation” provide the full specification using the coded format as defined in ISO 9022-1. The column labelled “Technical requirement” is only a short description and does not reproduce the full requirement. For a full description, it is necessary to consult the referenced standards.

Table 14 — List of international environmental and handling tests for optics and photonics equipment (the tests may be selected and combined as required)

Test number	Test description	In accordance with standard (leave blank if no accordance)	Conditioning method	Area of application: (e.g. outdoor)			
				Description of application:			
				Degree of severity	State of operation	Technical requirement	Status after test
1	Climate					Select the test and fill in the required parameters	
1.1	Cold	ISO 9022-2	10				
1.2	Heat (dry heat)	ISO 9022-2	11				
1.3	Damp heat	ISO 9022-2	12				
1.4	Condensation	ISO 9022-2	13				
1.5	Slow temperature change	ISO 9022-2	14				
1.6	Temperature shock, rapid temperature change	ISO 9022-2	15				
1.7	Damp heat, cyclic	ISO 9022-2	16				
1.8	Cold, dry heat or temperature change with bump or random vibration	ISO 9022-22	22				
....						
2	Atmosphere					Select the test and fill in the required parameters	
2.1	Combined low pressure and ambient temperature	ISO 9022-23	45				

Table 14 (continued)

Test number	Test description	In accordance with standard (leave blank if no accordance)	Conditioning method	Area of application: (e.g. outdoor)			
				Description of application:			
				Degree of severity	State of operation	Technical requirement	Status after test
2.2	Combined low pressure and dry heat	ISO 9022-23	46				
2.3	Damp heat and low internal pressure, pressure difference low	ISO 9022-23	47				
2.4	Damp heat and low internal pressure, pressure difference medium	ISO 9022-23	48				
2.5	Damp heat and low internal pressure, pressure difference high	ISO 9022-23	49				
2.6	Combined cold, low pressure including hoarfrost and dew	ISO 9022-23	50				
2.7	Combined cold, low pressure without hoarfrost and dew	ISO 9022-23	51				
2.8	Excess pressure						
2.9	High internal pressure	ISO 9022-8	80				
2.10	Low internal pressure	ISO 9022-8	81				
2.11	Immersion	ISO 9022-8	82				
2.12	Dust	ISO 9022-6	52				
2.13	Drip	ISO 9022-7	72				
2.14	Steady rain	ISO 9022-7	73				
2.15	Driving rain	ISO 9022-7	74				
2.16	Dew	ISO 9022-14	75				
2.17	Hoarfrost, followed by process of thawing	ISO 9022-14	76				
2.18	Ice covering, followed by process of thawing	ISO 9022-14	77				
2.19	Precipitation:						
	— Rain	IEC 60529					
	— Snow	IEC 60529					
	— Hail	IEC 60529					
	— Fog	IEC 60529					

Table 14 (continued)

Test number	Test description	In accordance with standard (leave blank if no accordance)	Conditioning method	Area of application: (e.g. outdoor)			
				Description of application:			
				Degree of severity	State of operation	Technical requirement	Status after test
	— Freezing fog	IEC 60529					
	— Hoarfrost	IEC 60529					
2.20	Icing						
2.21	Solar radiation	ISO 9022-9	20				
2.22	Warming due to one-sided sun exposure						
2.23	Wind						
2.24	Dust	IEC 60529					
2.25	Sand						
2.26	Polluted atmosphere — salt	IEC 60068-2-11					
2.27	Polluted atmosphere — marine environment	IEC 60068-2-52					
2.28	Polluted atmosphere — H ₂ S	IEC 60068-2-43					
2.29	Polluted atmosphere — SO ₂	IEC 60068-2-42					
2.30	Polluted atmosphere — NH ₃						
2.31	Polluted atmosphere — Ozone						
....						
3	Mechanical					Select the test and fill in the required parameters	
3.1	Shock	ISO 9022-3	30				
3.2	Handling shock	ISO 9022-3	30/33				
3.3	Bump	ISO 9022-3	31				
3.4	Drop and topple	ISO 9022-3	32				
3.5	Drop in transportation case	ISO 9022-3	33				
3.6	Drop in package for shipping	ISO 9022-3	33				
3.7	Free fall	ISO 9022-3	33				
3.8	Bounce	ISO 9022-3	34				

ISO 10109:2024(en)

Table 14 (continued)

Test number	Test description	In accordance with standard (leave blank if no accordance)	Conditioning method	Area of application: (e.g. outdoor)			
				Description of application:			
				Degree of severity	State of operation	Technical requirement	Status after test
3.9	Steady-state acceleration, centrifugal	ISO 9022-3	35				
3.10	Sinusoidal vibration	ISO 9022-3	36				
3.11	Vibration in transportation case	ISO 9022-3	36/37				
3.12	Random vibration (wide-band) digitally controlled	ISO 9022-3	37				
....						
4	Chemical					Select the test and fill in the required parameters	
4.1	Salt mist	ISO 9022-4	40				
4.2	Humid atmosphere containing sulfur dioxide (SO ₂)	ISO 9022-20	41				
4.3	Humid atmosphere containing hydrogen sulfide (H ₂ S)	ISO 9022-20	42				
4.4	Mould growth	ISO 9022-11	85				
4.5	Contamination: Basic cosmetic substances and artificial hand sweat	ISO 9022-12	86				
4.6	Contamination: Laboratory agents	ISO 9022-12	87				
4.7	Contamination: Production plant resources	ISO 9022-12	88				
4.8	Contamination: Fuels and resources for aircraft, naval vessels and land vehicles	ISO 9022-12	89				
....						
5	Optional tests					individual technical requirements	
....						

Table 14 (continued)

Test number	Test description	In accordance with standard (leave blank if no accordance)	Conditioning method	Area of application: (e.g. outdoor)			
				Description of application:			
				Degree of severity	State of operation	Technical requirement	Status after test
....						

STANDARDSISO.COM : Click to view the full PDF of ISO 10109:2024

Annex A
(informative)

Example of selection of environmental tests

As an example, an area of application has been defined in [Table A.1](#) for a typical “Outdoor use”. If necessary, more “Area of application” are required, additional tables can be added.

The combination of the three columns “Conditioning method”, “Degree of severity”, and “State of operation” provide the full specification using the coded format as defined in ISO 9022-1. The column labelled “Technical requirement” is only a short description and does not reproduce the full requirement. For a full description, it is necessary to consult the referenced standards.

Table A.1 — Example of selection of environmental tests for an outdoor application in non-weather-protected locations in temperate climate

Test number	Test description	In accordance with standard	Conditioning method	Area of application: outdoor			
				Non-weather-protected locations in temperate climate			
				Degree of severity	State of operation	Technical requirement	Status after test
1	Climate					Select the tests and fill in the required parameters	
1.1	Cold	ISO 9022-2	10	04	2	-20 °C, 16 h	A
1.2	Heat (dry heat)	ISO 9022-2	11	03	0 and 1	+55 °C	A
1.3	Damp heat	ISO 9022-2	12	06	2	+55 °C and r.h. 90 % to 95 %	C
1.4	Condensation	ISO 9022-2	13	03	1	2 d, +40 °C, 100 % rel. humidity	C
1.5	Slow temperature change	ISO 9022-2	14	02	0 and 1	from +55 °C to -25 °C	A
1.6	Temperature shock, rapid temperature change	ISO 9022-2	15	02	0 and 1	from +40 °C to -25 °C	A
1.7	Damp heat, cyclic	ISO 9022-2	16	04	2	+23 °C/83 % to 55 °C/95 %	A

^a These are environmental test codes for the methods selected for this combination test. See ISO 9022-1 for the numbering system and ISO 9022-22 for further details.

^b ISO 9022 parts 8 and 23 use hPa. Elsewhere in this document, kPa are used to describe air pressure. Note that 1 hPa = 0,1 kPa.

^c Note that it is permissible in a practical case to have more than one table row for one single test criterion.

^d "x" is a placeholder for the protection against ingress of objects; see IEC 60529 for more detail.

^e "x" is a placeholder for the protection against water ingress; see IEC 60529 for more detail.

^f IEC 60068-2-43 uses units of 1e-6 vol/vol. In this document parts per million, abbreviated ppm

ISO 10109:2024(en)

Table A.1 (continued)

Test number	Test description	In accordance with standard	Conditioning method	Area of application: outdoor			
				Non-weather-protected locations in temperate climate			
				Degree of severity	State of operation	Technical requirement	Status after test
1.8	Cold, dry heat or temperature change with bump or random vibration	ISO 9022-22	22	---	2	Dry heat, 55 °C Vibration 20 Hz to 150 Hz, 0,02 g _n ² /Hz (ISO 9022-11-03/37-01-2) ^a	A
2	Atmosphere					Select the tests and fill in the required parameters	
2.1	Combined low pressure and ambient temperature	ISO 9022-23	45	01	2	23 °C, 800 hPa ^b	A
2.2	Combined low pressure and dry heat	ISO 9022-23	46	03	2	55 °C, 100 hPa ^b , exposure 24 h	A
2.3	Damp heat and low internal pressure, pressure difference low	ISO 9022-23	47	02	2	55 °C, <40 % r.h., 12 cycles, etc.	A
2.4	Damp heat and low internal pressure, pressure difference medium	ISO 9022-23	48	02	2	40 °C, 800 hPa ^b , 90 % r.h., Cond.2: -10 °C, Cond.3: 40 °C	A
2.5	Damp heat and low internal pressure, pressure difference high	ISO 9022-23	49	02	2	40 °C, 90 % r.h., etc.	A
2.6	Combined cold, low pressure including hoarfrost and dew	ISO 9022-23	50	02	1	-40 °C, 600 hPa ^b , 4 h	A
2.7 a) ^c	Combined cold, low pressure without hoarfrost and dew	ISO 9022-23	51	02	1	-40 °C, 600 hPa ^b , 4 h	A

^a These are environmental test codes for the methods selected for this combination test. See ISO 9022-1 for the numbering system and ISO 9022-22 for further details.

^b ISO 9022 parts 8 and 23 use hPa. Elsewhere in this document, kPa are used to describe air pressure. Note that 1 hPa = 0,1 kPa.

^c Note that it is permissible in a practical case to have more than one table row for one single test criterion.

^d "x" is a placeholder for the protection against ingress of objects; see IEC 60529 for more detail.

^e "x" is a placeholder for the protection against water ingress; see IEC 60529 for more detail.

^f IEC 60068-2-43 uses units of 1e-6 vol/vol. In this document parts per million, abbreviated ppm

ISO 10109:2024(en)

Table A.1 (continued)

Test number	Test description	In accordance with standard	Conditioning method	Area of application: outdoor			
				Non-weather-protected locations in temperate climate			
				Degree of severity	State of operation	Technical requirement	Status after test
2.7 b) ^c	Combined cold, low pressure without hoarfrost and dew	ISO 9022-23	51	01	0 and 2	altitude up to 8 500 m	A
2.8	Excess pressure				0 and 2	-1 000 m NOTE Only if used in mines (with or without firedamp).	A
2.9	High internal pressure	ISO 9022-8	80	03	2	40 °C, press diff: 100 hPa ^b , drop: 20 hPa ^b , exposure 10 min	A
2.10	Low internal pressure	ISO 9022-8	81	03	2	40 °C, press diff: 100 hPa ^b , rise: 20 hPa ^b , exposure: 10 min	
2.11	Immersion	ISO 9022-8	82	02	2	immersion depth, 4 m, exposure 2 h	A
2.12	Blowing dust	ISO 9022-6	52	02	1	18 °C to 28 °C, <25 % r.h., 8 m/s to 10 m/s	A
2.13	Drip	ISO 9022-7	72	04	1	Drip rate: 3,5 mm/min, exposure: 5 min	A
2.14	Steady rain	ISO 9022-7	73	02	1	Rain rate: 20 mm/min, exposure: 30 min	A
2.15	Driving rain	ISO 9022-7	74	04	1	Rain rate: 10 mm/min, wind 18 m/s, exposure: 30 min	A
2.16	Dew	ISO 9022-14	75	01	2	10 °C, heat up to 30 °C, 85 % r.h.	A

^a These are environmental test codes for the methods selected for this combination test. See ISO 9022-1 for the numbering system and ISO 9022-22 for further details.

^b ISO 9022 parts 8 and 23 use hPa. Elsewhere in this document, kPa are used to describe air pressure. Note that 1 hPa = 0,1 kPa.

^c Note that it is permissible in a practical case to have more than one table row for one single test criterion.

^d "x" is a placeholder for the protection against ingress of objects; see IEC 60529 for more detail.

^e "x" is a placeholder for the protection against water ingress; see IEC 60529 for more detail.

^f IEC 60068-2-43 uses units of 1e-6 vol/vol. In this document parts per million, abbreviated ppm

ISO 10109:2024(en)

Table A.1 (continued)

Test number	Test description	In accordance with standard	Conditioning method	Area of application: outdoor			
				Non-weather-protected locations in temperate climate			
				Degree of severity	State of operation	Technical requirement	Status after test
2.17	Hoarfrost, followed by process of thawing	ISO 9022-14	76	02	2	-25 °C heat up to 30 °C, 90 % r.h.	A
2.18	Ice covering, followed by process of thawing	ISO 9022-14	77	02	2	-15 °C, 5 mm to 7 mm ice, heat up to 30 °C, 90 % r.h.	A
2.19	Precipitation:						
	— Rain	IEC 60529			0 and 2	IP x4 ^d	A
	— Snow	IEC 60529					
	— Hail	IEC 60529					
	— Fog	IEC 60529					
	— Freezing fog	IEC 60529					
	— Hoarfrost	IEC 60529					
2.20	Icing						
2.21	Solar radiation	ISO 9022-9	20	03	0 and 2	1 000 W/m ² , +40 °C	A
2.22	Warming due to one-sided sun exposure				2	drift, hysteresis	A
2.23	Wind						
2.24	Dust	IEC 60529			0 and 2	IP 5x ^e	B
2.25	Sand					see 2.24, Dust	
2.26	Polluted atmosphere — salt	IEC 60068-2-11				5 %, +35 °C, 24 h	
2.27	Polluted atmosphere — marine environment	IEC 60068-2-52				cyclic: 5 %, severity 2	
2.28	Polluted atmosphere — H₂S	IEC 60068-2-43				21 d; 1 ppm (special) ^f	
2.29	Polluted atmosphere — SO₂	IEC 60068-2-42				21 d; 10 ppm (special) ^f	

^a These are environmental test codes for the methods selected for this combination test. See ISO 9022-1 for the numbering system and ISO 9022-22 for further details.

^b ISO 9022 parts 8 and 23 use hPa. Elsewhere in this document, kPa are used to describe air pressure. Note that 1 hPa = 0,1 kPa.

^c Note that it is permissible in a practical case to have more than one table row for one single test criterion.

^d "x" is a placeholder for the protection against ingress of objects; see IEC 60529 for more detail.

^e "x" is a placeholder for the protection against water ingress; see IEC 60529 for more detail.

^f IEC 60068-2-43 uses units of 1e-6 vol/vol. In this document parts per million, abbreviated ppm

Table A.1 (continued)

Test number	Test description	In accordance with standard	Conditioning method	Area of application: outdoor			
				Non-weather-protected locations in temperate climate			
				Degree of severity	State of operation	Technical requirement	Status after test
2.30	Polluted atmosphere — NH ₃						
2.31	Polluted atmosphere — Ozone						
3	Mechanical					Select the tests and fill in the required parameters	
3.1	Shock	ISO 9022-3	30	02	1	Acceleration 150 m/s ²	A
3.2	Handling shock				1	9 bumps from handling position (1 m) onto hard wood (50 mm) over concrete floor.	B
3.3	Bump	ISO 9022-3	31	02	1	Acceleration 100 m/s ²	A
3.4	Drop and top-ple	ISO 9022-3	32	02	0	Height of overturn: 50 mm	A
3.5	Drop in transportation case	ISO 9022-3	33	05	0	0,5 m drop height, once to all surfaces onto hardwood (50 mm) over concrete floor.	B

^a These are environmental test codes for the methods selected for this combination test. See ISO 9022-1 for the numbering system and ISO 9022-22 for further details.

^b ISO 9022 parts 8 and 23 use hPa. Elsewhere in this document, kPa are used to describe air pressure. Note that 1 hPa = 0,1 kPa.

^c Note that it is permissible in a practical case to have more than one table row for one single test criterion.

^d "x" is a placeholder for the protection against ingress of objects; see IEC 60529 for more detail.

^e "x" is a placeholder for the protection against water ingress; see IEC 60529 for more detail.

^f IEC 60068-2-43 uses units of 1e-6 vol/vol. In this document parts per million, abbreviated ppm

ISO 10109:2024(en)

Table A.1 (continued)

Test number	Test description	In accordance with standard	Conditioning method	Area of application: outdoor			
				Non-weather-protected locations in temperate climate			
				Degree of severity	State of operation	Technical requirement	Status after test
3.6	Drop in package for shipping	ISO 9022-3	33	--	0	0,8 m (0 kg to 10 kg); 0,6 m (10 kg to 20 kg); 0,5 m (20 kg to 30 kg); 0,4 m (30 kg to 40 kg); 0,3 m (40 kg to 50 kg); 0,2 m (50 kg to 100 kg) Once to all surfaces onto hardwood (50 mm) over concrete floor.	B
3.7	Free fall	ISO 9022-3	33	05	0	Height of fall: 500 mm	A
3.8	Bounce	ISO 9022-3	34	02	0	Exposure time: 60 min	A
3.9	Steady-state acceleration, centrifugal	ISO 9022-3	35	02	1	Acceleration 100 m/s ²	A
3.10	Sinusoidal vibration	ISO 9022-3	36	04	2	10 Hz to 500 Hz, 20 m/s ² Exposure time: 30 min	B
3.11	Vibration in transportation case				0	10 Hz to 150 Hz; 2 g _n ; ±0,15 mm	B
3.12	Random vibration (wide band) digitally controlled	ISO 9022-3	37	02	1	20 Hz to 150 Hz, 0,05 g _n ² /Hz	A
4	Chemical					Select the tests and fill in the required parameters	

^a These are environmental test codes for the methods selected for this combination test. See ISO 9022-1 for the numbering system and ISO 9022-22 for further details.

^b ISO 9022 parts 8 and 23 use hPa. Elsewhere in this document, kPa are used to describe air pressure. Note that 1 hPa = 0,1 kPa.

^c Note that it is permissible in a practical case to have more than one table row for one single test criterion.

^d "x" is a placeholder for the protection against ingress of objects; see IEC 60529 for more detail.

^e "x" is a placeholder for the protection against water ingress; see IEC 60529 for more detail.

^f IEC 60068-2-43 uses units of 1e-6 vol/vol. In this document parts per million, abbreviated ppm