
**Optics and optical instruments —
Environmental requirements —**

Part 7:

**Test requirements for optical measuring
instruments**

Optique et instruments d'optique — Conditions d'environnement —

Partie 7: Spécifications d'essai pour instruments de mesure optiques



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

STANDARDSISO.COM : Click to view the full PDF of ISO 10109-7:2001

© ISO 2001

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.ch
Web www.iso.ch

Printed in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this part of ISO 10109 may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

International Standard ISO 10109-7 was prepared by Technical Committee ISO/TC 172, *Optics and optical instruments*, Subcommittee SC 1, *Fundamental standards*.

ISO 10109 consists of the following parts, under the general title *Optics and optical instruments — Environmental requirements*:

- *Part 1: General information, definitions, climatic zones and their parameters*
- *Part 4: Test requirements for telescopic systems*
- *Part 6: Test requirements for medical optical devices*
- *Part 7: Test requirements for optical measuring instruments*
- *Part 8: Test requirements for extreme conditions of use*
- *Part 11: Optical instruments for outdoor conditions of use*

Introduction

Optical measuring instruments as understood by this part of ISO 10109 are instruments, the functioning of which is based on optical principles, such as photometry, interferometry, geometrical optics, holography, refractometry, etc. but which can be supplemented or extended by peripheral methods.

Optical measuring instruments measure physical, geometric or material properties. Examples of optical measuring instruments include:

- chemo-physical analysis instruments (spectrometers, colour-measuring instruments);
- optical measuring instruments such as one-coordinate or multi-coordinate measuring machines, surface-measuring instruments, numerical measuring instruments for machine tool control, autocollimation telescopes, contour-measuring instruments, etc.

Nominal values of properties and performance characteristics as understood in this part of ISO 10109 are predetermined by specifications provided by the manufacturer, technical terms of delivery and instrument standards.

This part of ISO 10109 is the basis for the specification of environmental requirements and environmental tests in instrument standards. If necessary, these requirements and tests may be amended in the instrument standards.

STANDARDSISO.COM : Click to view the full PDF of ISO 10109-7:2001

Optics and optical instruments — Environmental requirements —

Part 7:

Test requirements for optical measuring instruments

1 Scope

This part of ISO 10109 specifies requirements to be met with regard to resistance of the optical, mechanical, chemical and electrical properties or performance data of instruments to environmental influences and hence determines geographical and technical areas of application. It applies to optical instruments and instruments with optical components including accessories in the field of optical metrology for laboratories, production plants and testing.

Environmental test methods as specified in ISO 9022 are assigned to the various areas of application for the purpose of ascertaining the suitability of the instruments in their respective area of application.

This part of ISO 10109 does not apply to instruments which are covered by other parts of ISO 10109.

This part of ISO 10109 does not deal with the requirements to be met by the packaging of the instrument during transport from the manufacturer to the user.

2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this part of ISO 10109. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this part of ISO 10109 are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

ISO 9022-1:1994, *Optics and optical instruments — Environmental test methods — Part 1: Definitions, extent of testing*

ISO 9022-2:1994, *Optics and optical instruments — Environmental test methods — Part 2: Cold, heat, humidity*

ISO 9022-3:1998, *Optics and optical instruments — Environmental test methods — Part 3: Mechanical stress*

ISO 9022-7:1994, *Optics and optical instruments — Environmental test methods — Part 7: Drip, rain*

ISO 9022-12:1994, *Optics and optical instruments — Environmental test methods — Part 12: Contamination*

ISO 9022-20:1997, *Optics and optical instruments — Environmental test methods — Part 20: Humid atmosphere containing sulfur dioxide or hydrogen sulfide*

ISO 10109-1:1994, *Optics and optical instruments — Environmental requirements — Part 1: General information, definitions, climatic zones and their parameters*

3 Terms and definitions

For the purposes of this part of ISO 10109, the terms and definitions given in ISO 9022-1 and ISO 10109-1 apply.

4 Subdivision of the instrument group

The group number of optical measuring instruments to which this part of ISO 10109 refers is 06.

Group number 06 is subdivided into instrument types and the type numbers are given in Table 1.

Table 1 — Subdivision of group 06

Type number	Instrument type ^a
01	Optical measuring instruments for mobile or stationary outdoor use; not weather-protected; e.g. aligning telescopes for industrial use, analytical measuring instruments for mobile use in measuring vehicles; contour-measuring instruments.
02	Optical measuring instruments for mobile or stationary outdoor use; weather-protected; e.g. aligning telescopes for industrial use, analytical measuring instruments for mobile use in measuring vehicles; contour-measuring instruments.
03	Optical measuring systems for use in difficult conditions, weather-protected; e.g. numerical measuring systems, such as linear and rotary encoders on machine tools.
04	Optical measuring instruments for use in rooms without air-conditioning; e.g. length and angle-measuring instruments, measuring and aligning telescopes for industrial use, analytical measuring instruments, contour-measuring instruments.
05	Optical measuring instruments for use at production plants, laboratories and measuring rooms; e.g. high-precision measuring instruments for length, angle, shape and surface measurements, analytical measuring instruments. If required for accuracy or technological reasons, instruments for use in climatic chambers or cleanrooms are also included.
^a Existing product standards or specifications shall primarily be taken into consideration.	

5 Designation of environmental requirements and of environmental test

The relevant specification and other technical documents shall indicate the environmental requirements required by this part of ISO 10109, using the designation in accordance with ISO 10109-1.

An example of the designation of environmental requirements for optical measuring instruments belonging to group 06, of instrument type 02 and requiring the extent of testing T, is:

Environmental requirements ISO 10109-06-02-T

In relevant specifications and other technical documentation, tests carried out in accordance with the environmental requirements given in this part of ISO 10109 shall be designated by the environmental test code as specified in ISO 9022-1.

6 Specification of technical requirements, environmental tests and suitability indices

For the purposes of this part of ISO 10109, the acceleration of free fall shall be taken as $g = 9,81 \text{ m/s}^2$.

Standard climates are specified in ISO 10109-1.

6.1 Type or sample test (extent of testing T)

Table 2 specifies technical requirements, environmental tests and suitability indices for extent of testing T.

Table 3 shows a summary of the tests given in Table 2, as specified in ISO 9022.

6.2 Series test (extent of testing S)

Series tests shall be stipulated in the relevant specification.

6.3 Special requirements

Further technical requirements to be met for instruments for outdoor conditions of use which are not contained in Table 2, e.g. solar radiation, may be selected from the ISO 9022 series, if required, and should be agreed upon separately between the customer and manufacturer.

7 Procedure

Tests shall be performed as specified in ISO 9022. The tests may be performed in any order, if not specified otherwise.

STANDARDSISO.COM : Click to view the full PDF of ISO 10109-7:2001

Table 2 — Technical requirements, environmental tests and suitability indices for extent of testing T

Serial No.	ISO 9022		Instrument type		Instruments not weather-protected			Instruments weather-protected			Instruments use in difficult conditions, weather-protected			Instruments in rooms without air-conditioning			Instruments for measuring rooms					
	Part	Conditioning method	Type number	State of operation	0	1	2	0	1	2	0	1	2	0	1	2	0	1	2			
1	2	10 Cold	State of operation	-40	-	-20	-40	-	-15	-	-40	-	0 ^a	-	-40	-	-	-40	-	-		
			Technical requirement	°C																		
			Degree of severity		08	-	04	08	-	03	08	-	01	08	-	08	-	-	08	-	-	-
			1	E	-	E	-	E	-	E	-	E	-	E	-	E	-	E	-	E	-	-
			2	A	-	B	-	A	-	C	-	A	-	C	-	A	-	A	-	A	-	-
			3	A	-	A	-	A	-	B	-	A	-	B	-	A	-	A	-	A	-	-
2	2	11 Dry heat	Suitability index for standard climates	E	-	E	-	E	-	E	-	E	-	E	-	E	-	E	-	-		
			4	E	-	E	-	E	-	E	-	E	-	E	-	E	-	E	-	-		
			5	A	-	A	-	A	-	A	-	A	-	A	-	A	-	A	-	A	-	-
			6	A	-	A	-	A	-	B	-	A	-	B	-	A	-	A	-	A	-	-
			Technical requirement	Temperature °C	70	-	55	70	-	55 ^a	70	-	70	-	40 ^a	70	-	10 to 40 ^a	70	-	b	
			Degree of severity		05	-	03	05	-	03	05	-	02	05	-	01 to 02	05	-	05	-	-	
3	2	12 Damp heat	1	A	-	A	-	A	-	A	-	A	-	A	-	A	-	A	-	-		
			2	A	-	A	-	A	-	A	-	A	-	C	-	A	-	A	-	-		
			3	A	-	A	-	A	-	A	-	A	-	A	-	A	-	A	-	A	-	
			4	A	-	A	-	A	-	A	-	A	-	C	-	A	-	D	-	A	-	
			5	A	-	A	-	A	-	A	-	A	-	A	-	A	-	A	-	A	-	-
			6	A	-	A	-	A	-	A	-	A	-	B	-	A	-	C	-	A	-	-
3	2	Damp heat	Technical Requirements	Temperature °C	-	40	-	40	-	40	-	40	-	40	-	40	-	40	-	-		
			Degree of severity		-	92	-	92	-	92	-	92	-	92	-	92	-	92	-	-		
			Suitability		-	01	-	01	-	01	-	01	-	01	-	01	-	01	-	-		

The instrument is suitable for the technical requirement if it is operative without restriction after conditioning.

Table 2 (continued)

Serial No.	ISO 9022		Instrument type		Instruments not weather-protected		Instruments weather-protected		Instruments for use in difficult conditions, weather-protected		Instruments in rooms without air-conditioning		Instruments for measuring rooms				
	Part	Conditioning method	Type number		01	02	03	04	05	06	07	08	09	10			
4	2	14 Slow temperature change	State of operation	t_2	0	1	2	0	1	2	0	1	2	0	1	2	
			Technical requirements	t_1	-	40	-	-	40	-	-	40	-	-	40	-	-
			Degree of severity		-	-10	-	-	-10	-	-	-10	-	-	-10	-	-
5	2	15 Temperatures hock	Suitability		-	01	-	-	01	-	-	01	-	-	-		
			Technical requirements	t_2	-	40	-	-	40	-	-	40	-	-	40	-	
			Degree of severity	t_1	-	-25	-	-	-25	-	-	-25	-	-	-25	-	
6	2	16 Damp heat, cyclic	Suitability		-	01	-	-	01	-	-	01	-	-	-		
			Technical requirements	Climate 1 °C/%	-	40/92	-	-	40/92	-	-	40/92	-	-	40/92	-	
			Degree of severity	Climate 2 °C/%	-	23/83	-	-	23/83	-	-	23/83	-	-	23/83	-	
7	3	30 Shock	Suitability		-	01	-	-	01	-	-	01	-	-	-		
			Technical requirements	Acceleration g_n	-	30	-	-	30	-	-	30	-	-	15	-	
			Degree of severity	Duration ms	-	18	-	-	18	-	-	18	-	-	11	-	
8	3	31 Bump	Suitability		-	04 ^c	-	-	04 ^c	-	-	04	-	-	02 ^c	-	
			Technical requirements	Acceleration g_n	25	-	-	25	-	-	25	-	-	10	-	10	-
			Degree of severity	Duration ms	6	-	-	6	-	-	6	-	-	6	-	6	-
			Suitability		05 ^e	-	-	05 ^e	-	-	05	-	-	01 ^c	-		

Table 2 (continued)

Serial No.	ISO 9022		Instrument type	Instruments not weather-protected			Instruments weather-protected			Instruments for use in difficult conditions, weather-protected			Instruments in rooms without air-conditioning			Instruments for measuring rooms				
	Part	Conditioning method		01			02			03			04			05				
9	3	36 Sinusoidal vibration	Type number	0	1	2	0	1	2	0	1	2	0	1	2	0	1	2		
			State of operation	0,5	-	-	0,5	-	-	2	-	-	-	0,5	-	-	-	-	-	
			Technical requirements	Acceleration ϵ_n	10 to 500	-	-	10 to 500	-	-	10 to 500	-	-	10 to 500	-	-	10 to 500	-	-	
10	3	37 Random vibration (wide band), digitally controlled	Degree of severity	01	-	-	01	-	-	04	-	-	01	-	-	01	-	-		
			Suitability	The instrument is suitable for the technical requirement if it is operative without restriction after conditioning.																
			Technical requirements	Rms (g multiples)	-	1,6	-	-	1,6	-	-	-	1,4	-	-	-	-	-	-	
11	20	41 Sulphur dioxide (SO ₂) in humid atmosphere	Degree of severity	-	20 to 500	-	-	20 to 500	-	-	20 to 500	-	-	20 to 500	-	-	-	-		
			Suitability	The instrument is suitable for the technical requirement if it is operative without restriction after conditioning.																
			Technical requirements	Frequency range	-	11	-	-	11 f	-	-	21 f	-	-	-	-	-	-	-	
12	7	74 Driving rain	Degree of severity	-	01	-	-	04	-	-	04	-	-	04	-	-	01	-		
			Suitability	The instrument is suitable for the technical requirement if it is operative without restriction after conditioning.																
			Technical requirements	Wind velocity	-	18	-	-	9	-	-	9	-	-	9	-	-	-	-	
12	7	74 Driving rain	Degree of severity	-	02	-	-	-	-	-	-	-	-	-	-	-	-	-		
			Suitability	The instrument is suitable for the technical requirement if it is operative without restriction after conditioning.																
			Technical requirements	Rain rate	-	2	-	-	-	-	-	-	-	-	-	-	-	-	-	

Table 2 (continued)

Serial No.	ISO 9022		Instrument type		Instruments not weather-protected		Instruments weather-protected		Instruments for use in difficult conditions, weather-protected		Instruments in rooms without air-conditioning		Instruments for measuring rooms			
	Part	Conditioning method	Type number	State of operation	0	1	2	0	1	2	0	1	2	0	1	2
13	12	86	Technical requirements	Ability to be operated for ≥ 5 years if stipulated instructions regarding care and maintenance are complied with.												
			Degree of severity	-	01	-	-	01	-	-	01	-	-	01	-	-
	Suitability	The instrument is suitable for the technical requirement if the representative sample or component has met the requirements of the test.														
14	12	87	Technical requirements	Ability to be operated for ≥ 5 years if stipulated instructions regarding care and maintenance are complied with.												
			Degree of severity	-	02	-	-	02	-	-	02	-	-	02	-	-
	Suitability	The instrument is suitable for the technical requirement if the representative sample or component has met the requirements of the test.														
15	12	88	Technical requirements	Ability to be operated for ≥ 5 years if stipulated instructions regarding care and maintenance are complied with.												
			Degree of severity	-	02	-	-	02	-	-	02	-	-	02	-	-
	Suitability	The instrument is suitable for the technical requirement if the representative sample or component ^d has met the requirements of the test.														
<p>a Existing product standards or specifications shall primarily be taken into consideration.</p> <p>b Technical requirement according to the relevant specification.</p> <p>c Mobile, hand-held units only.</p> <p>d Number of cycles: 2.</p> <p>e In mobile use, built in a test car only state of operation 1.</p> <p>f Only apply degree of severity if no other requirements exist which are found by spectrum analysis at the place of use.</p> <p>g Spray water test in accordance with relevant specification.</p> <p>h See ISO 9022-1.</p>																