



INTERNATIONAL STANDARD ISO 10993-7:2008
TECHNICAL CORRIGENDUM 1

Published 2009-11-15

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Biological evaluation of medical devices —
Part 7:
Ethylene oxide sterilization residuals

TECHNICAL CORRIGENDUM 1

Évaluation biologique des dispositifs médicaux —
Partie 7: Résidus de stérilisation à l'oxyde d'éthylène

RECTIFICATIF TECHNIQUE 1

Technical Corrigendum 1 to ISO 10993-7:2008 was prepared by Technical Committee ISO/TC 194, *Biological evaluation of medical devices*.

Page iv, Foreword

Correct the title of Part 1 to read as follows:

— *Part 1: Evaluation and testing within a risk management process*

Page 10, 5.3 Procedure for product release using residue dissipation curves

Second paragraph, second sentence should read:

Dissipation of EO from most materials and devices follows first-order kinetics, i.e. $(\ln[\text{EO}]) \propto$ (time after sterilization).

ICS 11.100.20

Ref. No. ISO 10993-7:2008/Cor.1:2009(E)

© ISO 2009 – All rights reserved

Published in Switzerland