

Third edition
2015-04-01

AMENDMENT 1
2020-02

Prefilled syringes —

Part 4:

**Glass barrels for injectables and
sterilized subassembled syringes
ready for filling**

AMENDMENT 1

Seringues préremplies —

*Partie 4: Cylindres en verre pour produits injectables et seringues pré-
assemblées stérilisées préremplissables*

AMENDEMENT 1



Reference number
ISO 11040-4:2015/Amd.1:2020(E)

© ISO 2020

STANDARDSISO.COM : Click to view the full PDF of ISO 11040-4:2015/Amd 1:2020



COPYRIGHT PROTECTED DOCUMENT

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 76, *Transfusion, infusion and injection, and blood processing equipment for medical and pharmaceutical use*.

A list of all parts in the ISO 11040 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

STANDARDSISO.COM : Click to view the full PDF of ISO 11040-4:2015/Amd 1:2020

Prefilled syringes —

Part 4:

Glass barrels for injectables and sterilized subassembled syringes ready for filling

AMENDMENT 1

Normative references

Replace the references to ISO 594-1 and ISO 594-2 with the following:

ISO 80369-7, *Small-bore connectors for liquids and gases in healthcare applications — Part 7: Connectors for intravascular or hypodermic applications*

Delete footnote 1).

Replace the reference to ISO 7886-1:1993 with the following:

ISO 7886-1:2017, *Sterile hypodermic syringes for single use — Part 1: Syringes for manual use*

Add the following reference:

ISO 21748:2017, *Guidance for the use of repeatability, reproducibility and trueness estimates in measurement uncertainty estimation*

4.2.1

Replace the text with the following:

Repeatability shall be evaluated for each test method in each laboratory.

Delete EXAMPLE.

5.1.1, second paragraph, second sentence

Replace the references "ISO 594-1" and "ISO 594-2" with "ISO 80369-7".

5.1.1, NOTE, third sentence

Replace the reference "ISO 594 series" with "ISO 80369-7".

5.1.1, NOTE, fourth sentence

Replace the reference "The ISO 594 series" with "ISO 80369-7".

ISO 11040-4:2015/Amd.1:2020(E)

5.1.1, third paragraph, third sentence

Replace the reference "ISO 594" with "ISO 80369-7".

5.1.1, third paragraph, last sentence

Replace the reference "ISO 594:1986" with "ISO 80369-7".

5.1.1, fourth paragraph

Replace the reference "ISO 594" with "ISO 80369-7".

5.1.2, Table 1

Replace header, "Finger flange", third line, under " d_3 ", second column "*nom.*" with "*tol.*".

5.2, second paragraph

Replace the reference "ISO 594-1" with "ISO 80369-7".

6.5.1.3

Replace the reference ISO 7886-1:1993 with "ISO 7886-1:2017".

6.5.2.2

Replace the third list item with the following:

- actual needle length should be in accordance with ISO 7864 (see l in ISO 7864:2016, Figure 2).

6.5.2.2, second paragraph

Replace the paragraph with the following:

When there are particular requirements on needle tip height from the flange or the shoulder, the dimension should be agreed upon between the manufacturer and the customer.

6.5.3.4

Replace the references "ISO 594-1, ISO 594-2" with "ISO 80369-7".

A.2.2, Title

Replace the reference "ISO 594-1" with "ISO 80369-7".