

Third edition
2013-06-01

AMENDMENT 1
2022-06

**Sterilization of health care products —
Radiation —**

Part 2:
Establishing the sterilization dose

AMENDMENT 1

Stérilisation des produits de santé — Irradiation —

Partie 2: Établissement de la dose stérilisante

AMENDEMENT 1

STANDARDSISO.COM : Click to view the full PDF of ISO 11137-2:2013/Amd 1:2022



Reference number
ISO 11137-2:2013/Amd.1:2022(E)

© ISO 2022

STANDARDSISO.COM : Click to view the full PDF of ISO 11137-2:2013/Amd 1:2022



COPYRIGHT PROTECTED DOCUMENT

© ISO 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 198, *Sterilization of health care products*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 204, *Sterilization of medical devices*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

A list of all parts in the ISO 11137 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

STANDARDSISO.COM : Click to view the full PDF of ISO 11137-2:2013/Amd 1:2022

Sterilization of health care products — Radiation —

Part 2: Establishing the sterilization dose

AMENDMENT 1

Clause 2 Normative references

Add the following new normative references:

ISO 11137-1:2006/Amd1:2013, *Sterilization of health care products — Radiation — Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices — Amendment 1*

ISO 11137-1:2006/Amd2:2018, *Sterilization of health care products — Radiation — Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices — Amendment 2: Revision to 4.3.4 and 11.2*

ISO 13004: 20—¹), *Sterilization of health care products — Radiation — Substantiation of selected sterilization dose: Method VD_{max}^{SD}*

6.4

Add the following new subclause:

6.4 If a sterilization dose of 17,5 kGy, 20 kGy, 22,5 kGy, 27,5 kGy, 30 kGy, 32,5 kGy or 35 kGy is established, it shall be substantiated by one of the following methods:

- a) Method VD_{max}^{SD} of ISO 13004:20—;
- b) Method 1 (see Clause 7), subject to the derived sterilization dose taking a value less than or equal to the selected sterilization dose and achieving maximally an SAL of 10^{-6} ;
- c) Method 2 (see Clause 8), subject to the derived sterilization dose taking a value less than or equal to the selected sterilization dose and achieving an SAL of 10^{-6} ; or
- d) a method providing equivalent assurance to that of a), b) or c) above in achieving maximally an SAL of 10^{-6} .

Clause 9

Replace clause title with the following:

9 Method VD_{max} — Substantiation of a selected sterilization dose

9.1

Replace subclause title, add new first paragraph and replace original first paragraph as follows:

1) Under preparation. Stage at the time of publication: ISO/DIS 13004.

9.1 Selected doses and rationale

Rationale and methods for substantiation of sterilization doses of 15 kGy and 25 kGy using Method VD_{max} are provided in Clause 9. Rationale and methods for substantiation of sterilization doses of 17,5 kGy, 20 kGy, 22,5 kGy, 27,5 kGy, 30 kGy, 32,5 kGy or 35 kGy using Method VD_{max} are provided in ISO 13004:20—.

Operationally, Method VD_{max} for substantiation for a selected sterilization dose is similar to dose setting by Method 1 (see Clause 7); it also requires a determination of bioburden and the performance of a verification dose experiment.

STANDARDSISO.COM : Click to view the full PDF of ISO 11137-2:2013/Amd 1:2022