
**Determination of the detection limit and
decision threshold for ionizing radiation
measurements —**

Part 2:

Fundamentals and application to counting
measurements with the influence of sample
treatment

*Détermination de la limite de détection et du seuil de décision des
mesurages des rayonnements ionisants —*

*Partie 2: Principes fondamentaux et application aux mesures par comptage,
avec l'influence du traitement d'échantillon*



Contents

Page

Foreword.....	iii
Introduction.....	iv
1 Scope	1
2 Normative reference	1
3 Terms and definitions.....	1
4 Symbols	3
5 Statistical values and confidence interval.....	4
5.1 Principles	4
5.2 Decision threshold	5
5.3 Detection limit	5
5.4 Confidence interval	6
6 Application of this part of ISO 11929 (see annex A).....	6
6.1 Specified values.....	6
6.2 Assessment of a measuring procedure	6
6.3 Assessment of measured results.....	7
6.4 Documentation.....	7
Annex A (informative) Example of application of this part of ISO 11929: Determination of strontium-90 in soil including chemical separation.....	11
Bibliography	16

© ISO 2000

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization
Case postale 56 • CH-1211 Genève 20 • Switzerland
Internet iso@iso.ch

Printed in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 11929-2 was prepared by Technical Committee ISO/TC 85, *Nuclear energy*, Subcommittee SC 2, *Radiation protection*, Working Group WG 17 (formerly WG 2), *Radioactivity measurements*.

ISO 11929 consists of the following parts, under the general title *Determination of the detection limit and decision threshold for ionizing radiation measurements*:

- *Part 1: Fundamentals and application to counting measurements without the influence of sample treatment*
- *Part 2: Fundamentals and application to counting measurements with the influence of sample treatment*
- *Part 3: Fundamentals and application to counting measurements by high resolution gamma spectrometry, without the influence of sample treatment*
- *Part 4: Fundamentals and application to measurements by use of linear scale analogue ratemeters, without the influence of sample treatment*

Annex A of this part of ISO 11929 is for information only.

Introduction

This part of ISO 11929 addresses the field of ionizing radiation measurements in which events (in particular pulses) on samples are counted after treating them (e.g. aliquotation, solution, enrichment, separation). It considers, besides the random character of radioactive decay and of pulse counting, all other influences arising from sample treatment, (e.g. weighing, enrichment, calibration or the instability of the test setup).

In general, it can be assumed that the results of such measurements follow a mixture of Poisson distributions. Special kinds of mixtures of Poisson distributions are negative binominal distributions [1, 3, 8].

It is assumed part of ISO 11929 that the results from counting of samples and blanks follow a negative binominal distribution. If the influence of sample treatment and instability of the device is small in comparison to statistical counting errors, the results may be Poisson distributed. In this case, ISO 11929-1 can be applied for specification of the statistical values as long as the dispersion test [2] indicates that the Poisson distribution cannot be rejected.

It is also assumed that the duration of measurement is small compared to the half-life of the radionuclides concerned and that the instrument dead time losses are negligible.

This part of ISO 11929 was prepared in parallel with other International Standards prepared by WG 2 (now WG 17): ISO 11932:1996, *Activity measurements of solid materials considered for recycling, re-use or disposal as non-radioactive waste*, and ISO 11929-1, ISO 11929-3 and ISO 11929-4 and is, consequently, complementary to these documents.

The other parts of ISO 11929 deal with counting measurements which do not take sample treatment into consideration, analogue pulse rate-measurement, and specific problems which occur when this part of ISO 11929 is applied (e.g. in the case of spectrometric measurements or continuous monitoring of radioactivity effluents).

Determination of the detection limit and decision threshold for ionizing radiation measurements —

Part 2:

Fundamentals and application to counting measurements with the influence of sample treatment

1 Scope

This part of ISO 11929 specifies suitable statistical values which allow an assessment of the detection capabilities in ionizing radiation measurements with the influence of sample treatment. For this purpose, statistical methods are used to specify the following two statistical values characterizing given probabilities of error.

- The decision threshold, which allows a decision to be made for each measurement with a given probability of error as to whether the registered pulses include a contribution by the sample.
- The detection limit, which specifies the minimum sample contribution which can be detected with a given probability of error using the measuring procedure in question. This consequently allows a decision to be made as to whether a measuring method defined in this part of ISO 11929 satisfies certain requirements and is consequently suitable for the given purpose of measurement.

2 Normative reference

The following normative document contains provisions which, through reference in this text, constitute provisions of this part of ISO 11929. For dated references, subsequent amendments to, or revisions of, such publications do not apply. However, parties to agreements based on this part of ISO 11929 are encouraged to investigate the possibility of applying the most recent edition of the normative document indicated below. For undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

BIPM/IEC/IFCC/ISO/IUPAC/IUPAP/OIML Guide to the Expression of Uncertainty in Measurement. Geneva 1993.

3 Terms and definitions

For the purposes of this part of ISO 11929, the following terms and definitions apply.

3.1

measuring method

combination of a specific sample treatment procedure and the use of a measuring instrument for counting measurements under given conditions

3.2

decision threshold

critical value of a statistical test for the decision between the null hypothesis $\rho_S = \rho_0$ and the alternative hypothesis $\rho_S > \rho_0$.

NOTE It should be the value R_n^* which, when exceeded by the determined value R_n , is taken to indicate that the null hypothesis should be rejected. The statistical test should be designed such that the probability of wrongly rejecting the hypothesis (error of the first kind) is equal to a value α which is fixed prior to commencement of the measurement

3.3

detection limit

smallest expectation value of the net counting rate that can be detected on given probabilities and, therefore, the smallest difference $\rho_n = \rho_s - \rho_0$ associated with the statistical test concerned for the decision between the hypothesis $\rho_s = \rho_0$ and the alternative hypothesis $\rho_s > \rho_0$ and having the following characteristic: if in reality $\rho_n \geq \rho$, the probability of wrongly not rejecting the hypothesis $\rho_s = \rho_0$ (error of second kind) shall be at most equal to a value β which is fixed prior to commencement of the measurement

NOTE The difference between using the decision threshold and using the detection limit is that measured values are to be compared with the decision threshold while the detection limit is to be compared with the guideline value.

3.4

confidence interval

interval including the true value of ρ_n in at least $(1 - \gamma) \times 100$ % of all cases

3.5

sample

whole amount or an aliquot of an inactive material, the content of radioactive nuclides of which has to be determined by ionizing radiation measurement after treatment

3.6

blank

aliquot of an inactive material with properties similar to the sample, which will be treated and measured in the same way as the sample

3.7

background effect

measured counting rate from the blank and radiation from external sources and radionuclides in the detector and shielding

3.8

gross effect

measured counting rate from the sample (sample contribution) and the background radiation

3.9

net effect

(sample contribution) gross effect minus the background effect

3.10

guideline value

value constituted by requirements on measuring procedures arising for scientific, legal or other reasons which are specified, for example, as activity, activity concentration, dose rate, etc.

NOTE If necessary, a calibration factor can be determined using a radioactive reference standard.

4 Symbols

n_0	Number of measured blanks
n_m	Number of measured spiked blanks
n_s	Number of measured samples
n_u	Number of external background measurements using a LSC cocktail with the same amount of demineralized water added instead of strontium solution in the device
$N_{0,i}$	Number of pulses counted with the i -th blank
$N_{m,i}$	Number of pulses counted with the i -th spiked blank
$N_{s,i}$	Number of pulses counted with the i -th sample
$N_{u,i}$	Number of pulses counted during the i -th run of external background measurement using a LSC cocktail with the same amount of demineralized water added instead of strontium solution
t_0	Duration of the measurement with a blank
t_s	Duration of the measurement with a sample
t_m	Duration of the measurement with a spiked blank
t_u	Duration of the measurement with neither a blank nor a sample in the device
$R_{0,i}$	Background effect: counting rate with a blank, quotient of the pulses N_0 counted during the preselected duration of measurement t_0 and the duration of measurement t_0 : $R_{0,i} = N_{0,i}/t_0$
$R_{s,i}$	Gross effect counting rate, quotient of the number of pulses counted during the preselected duration of measurement t_s and the duration of measurement t_s : $R_{s,i} = N_{s,i}/t_s$
$R_{m,i}$	Counting rate with a spiked blank, quotient of the number of pulses counted during the preselected duration of measurement t_m and the duration of measurement t_m : $R_{m,i} = N_{m,i}/t_m$
$R_{u,i}$	External background radiation counting rate using a LSC cocktail with the same amount of demineralized water added instead of strontium solution, quotient of the number of pulses counted during the preselected duration of measurement t_u and the duration of measurement t_u : $R_{u,i} = N_{u,i}/t_u$
$\overline{R_0}$	Mean of all $R_{0,i}$
$\overline{R_s}$	Mean of all $R_{s,i}$
$\overline{R_m}$	Mean of all $R_{m,i}$
$\overline{R_u}$	Mean of all $R_{u,i}$
$\overline{R_n}$	$\overline{R_n} = \overline{R_s} - \overline{R_0}$
ρ_0	Expectation value of all $R_{0,i}$
ρ_s	Expectation value of all $R_{s,i}$
ρ_m	Expectation value of all $R_{m,i}$
ρ_u	Expectation value of all $R_{u,i}$
ρ_n	Expectation value of all $R_{n,i}$
R_n^*	Decision threshold for the net counting rate R_n (see Tables 1 and 2)

ρ_n^*	Detection limit for the expectation value of the net counting rate R_n (see Tables 1 and 2)
σ_0^2	Variance of all $R_{0,i}$
σ_s^2	Variance of all $R_{s,i}$
α	Error of the first kind; the probability of rejecting the null hypothesis $\rho_s = \rho_0$ for the alternative hypothesis $\rho_s > \rho_0$ when the null hypothesis is true
β	Error of the second kind; the probability of accepting the null hypothesis $\rho_s = \rho_0$ against the alternative hypothesis $\rho_s > \rho_0$ when the null hypothesis is false
$1 - \gamma$	Confidence level of the confidence interval for ρ_n (see Tables 1 and 2)
θ	Relative standard deviation of the error due to the influence of the sample treatment and of the device instability
$\left. \begin{array}{l} k_{1-\alpha}, \\ k_{1-\beta}, \\ k_{(1-\gamma/2)} \end{array} \right\}$	Quantiles of the standard normal distribution (see Table 3)
$t_{1-\alpha, f}$	Quantiles of the t -distribution for degree of freedom f (see Table 4)

5 Statistical values and confidence interval

5.1 Principles

5.1.1 General aspects

The definition of the statistical values for decision threshold, detection limit and confidence interval are based on the variances of the measured results. They are a combination of the variations caused by counting statistics, measurement equipment instability and sample treatment. Measurement equipment instability normally can be neglected because usually it is small in comparison to the other influences. The contribution of counting statistics can be calculated by the Poisson formula. The parameter describing the contribution of the sample treatment can be determined by special experiments with spiked blanks, by experimental experience or with reduced quality by the variances of the given measured results.

5.1.2 Model

If device instabilities are neglected, the following model can be applied. The number of pulses, N_0 , counted with a blank is the sum of background radiation (external sources and detector noise) and matrix radiation.

The number of pulses, N_s , counted with a sample is the sum of background, matrix and sample radiation (net counting):

$$N_s = N_0 + N_n \quad (1)$$

It is assumed that, for constant radioactive emission, the numbers of pulses counted are Poisson distributed. In this case, the expectations of the counting rates R_0 and R_s are ρ_0 and $\rho_s = \rho_0 + \rho_n$, respectively.

The variances of R_0 and R_s are ρ_0/t_0 and ρ_s/t_s , respectively.

In a more general case than the previous one, it can be assumed that the sampling error (and errors caused by sample treatment) related to the sum ($\rho_0 + \rho_n$) is a random variable and its relative standard deviation is θ . The expectations of R_0 and R_s are not modified but the variances of them become

$$\text{var}(R_0) = \rho_0/t_0 + (\rho_0 - \rho_u)^2\theta^2 \text{ and } \text{var}(\overline{R_0}) = 1/n_0[\rho_0/t_0 + (\rho_0 - \rho_u)^2\theta^2] \quad (2)$$

$$\text{var}(R_s) = \rho_s/t_s + (\rho_s - \rho_u)^2\theta^2 \text{ and } \text{var}(\overline{R_s}) = 1/n_s[\rho_s/t_s + (\rho_s - \rho_u)^2\theta^2] \quad (3)$$

The net counting rate $R_n = (R_s - R_0)$ has the expectation ρ_n and the variance

$$\text{var}(R_n) = \text{var}(R_0) + \text{var}(R_s) \quad (4)$$

In this case, ρ_0 and ρ_s are unknown parameters; the measured values R_0 , R_s can be taken as estimates and θ can be obtained separate measurements.

5.1.3 Determination of parameter θ

The parameter $\theta^2 \geq 0$ is assumed to be well known, if it is determined using a line of at least $n_m > 20$ spiked blanks. The activity added to the blank for spiking should be high enough to make the variance of counting statistics small in comparison to the expected variance of sample treatment. θ can be calculated using this formula:

$$\theta^2 = \frac{1}{(\overline{R_m} - \rho_u)^2} \left[\frac{\sum_{i=1}^{n_m} (R_{m,i} - \overline{R_m})^2}{n_m - 5/8} - \frac{\overline{R_m}}{t_m} \right] \quad (5)$$

If the result is negative, take $\theta^2 = 0$.

ρ_u should be determined with very high precision.

It is assumed, that θ is the same for a spiked blank and a sample.

5.2 Decision threshold

The decision threshold shall refer to the value R_n^* which, when exceeded by a measured mean net counting rate R_n , is taken to indicate that a sample contribution exists. Otherwise, it shall be assumed in each case that there is no sample contribution.

If this decision rule is observed, a wrong decision occurs with the probability α that there is a sample contribution when in fact only a background effect exists (error of the first kind).

If θ is known the decision threshold is given by

$$R_n^* = k_{1-\alpha} \sqrt{\text{var}(R_n = 0)} \quad (6)$$

where $k_{1-\alpha}$ is a factor given in Table 3.

Formulae for calculation of the decision threshold are given in Table 1 (if θ is known) and in Table 2 (if θ is unknown), respectively.

5.3 Detection limit

The detection limit shall refer to the smallest expectation of the net counting rate ρ for which a wrong decision occurs with the probability β (if the decision rule as specified in 5.2 is applied) that there is no sample contribution but only a background effect (error of the second kind).

To check whether a measuring procedure is suitable for the purpose of measurement, the detection limit shall be compared with a specified guideline value (e.g. specified requirements on the sensitivity of the measuring procedure for scientific, legal or other reasons).

So, if θ is known and with α and β , the detection limit is

$$\rho_n^* = R_n^* + k_{1-\beta} \sqrt{\text{var}(R_n = \rho_n^*)} \quad (7)$$

$$= k_{1-\alpha} \sqrt{\text{var}(R_n = 0)} + k_{1-\beta} \sqrt{\text{var}(R_n = \rho_n^*)} \quad (8)$$

and if $\text{var}(R_n = 0) \approx \text{var}(R_n > 0)$

$$\rho_n^* = (k_{1-\alpha} + k_{1-\beta}) \sqrt{\text{var}(R_n = 0)} \quad (9)$$

If $\alpha = \beta$ as recommended in ISO 11929-1, the detection limit is

$$\rho_n^* = 2 R_n^* \quad (10)$$

where $k_{1-\alpha}$, $k_{1-\beta}$ are factors given in Table 3.

Formulae for calculation of the detection limit under different conditions are given in Table 1 (if θ is known) and in Table 2 (if θ is unknown), respectively.

5.4 Confidence interval

Formulae for calculation of the confidence interval are given in Table 1 (if θ is known) and in Table 2 (if θ is unknown).

6 Application of this part of ISO 11929 (see annex A)

6.1 Specified values

The error probabilities α , β and the confidence level $1 - \gamma$ shall be specified in advance. Frequently cited values are $\alpha = \beta = \gamma = 0,025$. The number of blanks and samples shall be chosen as great as is reasonably achievable. The measuring times shall be chosen such that the detection limit is below the guideline value.

NOTE If $\alpha = \beta = \gamma/2$ are chosen and if $\text{var}(R_n)$ varies just a little with R_n , one obtains for $R_n = R_n^*$ the confidence interval $R_n^* \pm k_{1-\alpha} \sqrt{\text{var}(R_n)}$, that is the interval $(0, \rho_n^*)$. This choice avoids a discontinuity in the expression of results.

6.2 Assessment of a measuring procedure

The decision as to whether a measuring method (3.1) satisfies certain requirements with respect to the detection limit shall be made by comparing the detection limit which has been determined with the specified guideline value (see 5.3).

This may be carried out either in advance, for the assessment of an intended measuring method on the basis of an empirically determined value for the background effect or a separate measurement, or else in retrospect, for the assessment of a measurement already carried out on the basis of the blank results then available.

The detection limit may be calculated by means of the formulae in 5.3 or in Table 1 (if θ is known) or in Table 2 (if θ is unknown), respectively.

If the detection limit thus determined is greater than the guideline value, the measuring procedure is not suitable for the purpose of the measurement.

NOTE Under certain circumstances, a measuring procedure may be suitable for the purpose of measurement, for example, by preselecting a greater duration of measurement or a higher number of pulses, by reducing the background effect, or by increasing sample quantity or chemical enrichment.

6.3 Assessment of measured results

The decision threshold may be calculated by means of the formulae in Table 1 (if θ is known) and in Table 2 (if θ is unknown), respectively.

A measured result shall be compared with the decision threshold thus obtained (see 5.2).

6.4 Documentation

A report on measurements in accordance with this part of ISO 11929 shall be accompanied by details on the probabilities of error, the decision threshold and the detection limit.

For established sample contributions, in addition to the measured value, confidence intervals determined in accordance with the equations in Table 1 (if θ is known) and in Table 2 (if θ is unknown), and the confidence level shall also be reported.

Table 1 — Formulae for the calculation of statistical values if θ is known

$q = \frac{n_s t_s}{n_0 t_0}$	(11)
$c_1 = \frac{1}{n_0 t_0} + \frac{1}{n_s t_s}$	(12)
$c_2 = \frac{1}{n_0} + \frac{1}{n_s}$	(13)
$c_3 = \rho_0 \left(\frac{1}{n_0 t_0} + \frac{1}{n_s t_s} \right) + \theta^2 (\rho_0 - \rho_u)^2 \left(\frac{1}{n_0} + \frac{1}{n_s} \right)$	(14)
$c_4 = \frac{1}{2} (k_{1-\alpha} + k_{1-\beta})$	(15)
$c_5 = \frac{q}{q+1}$	(16)
Decision threshold	
$R_n^* = \frac{k_{1-\alpha}^2 c_5 [c_1 + 2c_2 \theta^2 (\overline{R_0} - \rho_u)]}{2[1 - k_{1-\alpha}^2 \theta^2 c_2 c_5^2]} \times \left[1 + \sqrt{1 + \frac{4[1 - k_{1-\alpha}^2 \theta^2 c_2 c_5^2] [c_1 \overline{R_0} + c_2 \theta^2 (\overline{R_0} - \rho_u)^2]}{k_{1-\alpha}^2 c_5 [c_1 + 2c_2 \theta^2 (\overline{R_0} - \rho_u)]^2}} \right] \quad (17)$	
If $n_s t_s \rho_0$ is big enough and θ is small, the following simplified formula may be applied	
$R_n^* = k_{1-\alpha} \sqrt{\overline{R_0} \left(\frac{1}{n_0 t_0} + \frac{1}{n_s t_s} \right) + \theta^2 (\overline{R_0} - \rho_u)^2 \left(\frac{1}{n_0} + \frac{1}{n_s} \right)} \quad (18)$	

Table 1 (continued)

Detection limit

$$\rho_n^* = \frac{c_4^2 \left[c_3 + \frac{\theta^2}{n_s} (2\rho_0^2 - c_3 c_4^2) \right]}{\rho_0 \left[c_4^2 \frac{\theta^2}{n_s} - 1 \right]^2} \times \left\{ 1 + \sqrt{1 + \frac{\left[c_4^2 \frac{\theta^2}{n_s} - 1 \right]^2 \left[4c_3 \rho_0^2 - c_3^2 c_4^2 \right]}{c_4^2 \left[c_3 + \frac{\theta^2}{n_s} (2\rho_0^2 - c_3 c_4^2) \right]^2}} \right\} \quad (19)$$

If $n_s \cdot t_s \cdot \rho_0$ is big enough and θ is small, the following simplified formula may be applied

$$\rho_n^* = (k_{1-\alpha} + k_{1-\beta}) \sqrt{\rho_0 \left(\frac{1}{n_0 t_0} + \frac{1}{n_s t_s} \right) + \theta^2 (\rho_0 - \rho_u)^2 \left(\frac{1}{n_0} + \frac{1}{n_s} \right)} \quad (20)$$

Estimate for ρ_0 is R_0

Confidence interval

$$\bar{R}_n - a \leq \rho_n \leq \bar{R}_n + a \quad (21)$$

$$a = k_{1-\gamma/2} \sqrt{\frac{\bar{R}_0}{n_0 t_0} + \frac{\bar{R}_s}{n_s t_s} + \theta^2 \left[\frac{(\bar{R}_0 - \rho_u)^2}{n_0} + \frac{(\bar{R}_s - \rho_u)^2}{n_s} \right]} \quad (22)$$

NOTE These formulae are approximations, which are the more exact, the greater the number $t \cdot \rho_0$ or $t \cdot R_0$, respectively. For the values $\alpha = \beta = \gamma/2 = 0,025$, the following applies: The actual probability of an error of the first kind lies between 0,04 and 0,06, in the case where the true value of $\rho_0 \cdot t_s \cdot n_0$ or $\rho_0 \cdot t_0 \cdot n_0$ is ≤ 10 and θ is $< 0,2$.

Under these conditions, the actual probability of an error of the second kind lies between 0,04 and 0,055 and the confidence probability is greater than 0,94.

The range of application of the given formulae is at its broadest for $t_0 \approx t_s$ and $n_0 \approx n_s$, in particular when small numbers of pulses are counted.

With small values of α, β, γ (e.g. $< 0,01$), the actual probabilities of error may deviate significantly more from the specified values. Anyone using this part of ISO 11929 should refer to [1] or his own statistical studies.

In the special case where $\theta = 0$, the formulae are identical to those in ISO 11929-1.

Table 2 — Formulae for the calculation of statistical values if θ , ρ_0 are unknown, based on empirical variations

$$s_0^2 = \frac{1}{n_0 - 1} \sum_{i=1}^{n_0} (R_{0,i} - \overline{R_0})^2 \quad (23)$$

$$s_s^2 = \frac{1}{n_s - 1} \sum_{i=1}^{n_s} (R_{s,i} - \overline{R_s})^2 \quad (24)$$

$$f = n_0 + n_s - 2 \quad (25)$$

Decision threshold

$$R_n^* = t_{1-\alpha, f} \sqrt{\frac{1}{f} \left(\frac{1}{n_0} + \frac{1}{n_s} \right) \cdot [(n_0 - 1)s_0^2 + (n_s - 1)s_s^2]} \quad (26)$$

Detection limit

$$\rho_n^* = \left[t_{1-\alpha, f} + k_{1-\beta} \sqrt{1 + \frac{t_{1-\alpha, f}^2}{2f}} \right] \sqrt{\frac{\sigma_0^2}{n_s} + \frac{\sigma_s^2}{n_0}} \quad (27)$$

Estimates for

$$\sigma_0^2 : s_0^2$$

$$\sigma_s^2 : s_s^2$$

Confidence interval

$$\overline{R_n} - a \leq \rho_n \leq \overline{R_n} + a \quad (28)$$

$$a = t_{(1-\gamma/2), f} \sqrt{\frac{1}{f} \left(\frac{1}{n_0} + \frac{1}{n_s} \right) [(n_0 - 1)s_0^2 + (n_s - 1)s_s^2]} \quad (29)$$

NOTE These formulae are approximations, which are less exact than those in Table 1. These approximations are the more exact, the greater the number $t \cdot \rho_0$ or $t \cdot R_0$ respectively. For the values $\alpha = \beta = \gamma/2 = 0,025$, the following applies: The actual probability of an error of the first kind lies between 0,04 and 0,06 in the case where the true value of $\rho_0 \cdot t_s \cdot n_0$ or $\rho_0 \cdot t_0 \cdot n_s$ is ≤ 10 and θ is $< 0,2$.

Under these conditions, the actual probability of an error of the second kind lies between 0,04 and 0,055 and the confidence probability is greater than 0,94.

The range of application of the given formulae is at its broadest for $t_0 \approx t_s$ and $n_0 \approx n_s$, in particular when small numbers of pulses are counted.

With small values of α , β , γ (e.g. $< 0,01$), the actual probabilities of error may deviate significantly more from the specified values. Anyone using this part of ISO 11929 should refer to [1] or his own statistical studies.

Table 3 — Values $k_{1-\gamma}$, $k_{1-\beta}$, $k_{1-(\gamma/2)}$ as a function of the error probabilities α , β and of the confidence level $1-\gamma$ (quantiles of normal distribution)

Error of probability α or β	Confidence level $1-\gamma$	$k_{1-\alpha}$ $k_{1-\beta}$ $k_{1-(\gamma/2)}$
0,158 6	0,682	1,000
0,100 0	0,800	1,282
0,050 0	0,900	1,645
0,025 0	0,950	1,960
0,022 8	0,955	2,000
0,010 0	0,980	2,326
0,005 0	0,990	2,576
0,001 4	0,997	3,000
0,001 0	0,998	3,090

Table 4 — $t_{1-\alpha, f}$ and $t_{1-\gamma/2, f}$ as a function of error probability α or a confidence level $1-\gamma$ and degree of freedom f (quantiles of t -distribution); for further information see [3]

α $1-\gamma$	0,158 6	0,100 0	0,050 0	0,025 0	0,022 8	0,010 0	0,005 0	0,001 4	0,001 0
f	0,682 8	0,800 0	0,900 0	0,950 0	0,955 4	0,980 0	0,990 0	0,997 2	0,998 0
2	1,32	1,89	2,92	4,30	4,52	6,97	9,93	18,86	22,33
3	1,20	1,64	2,35	3,18	3,30	4,54	5,84	9,22	10,21
4	1,14	1,53	2,13	2,78	2,87	3,75	4,60	6,62	7,17
5	1,11	1,48	2,02	2,57	2,65	3,37	4,03	5,51	5,89
6	1,09	1,44	1,94	2,45	2,51	3,14	3,71	4,90	5,21
7	1,08	1,42	1,90	2,37	2,43	3,00	3,50	4,53	4,79
8	1,07	1,40	1,86	2,31	2,36	2,90	3,36	4,28	4,50
9	1,06	1,38	1,83	2,26	2,32	2,82	3,25	4,10	4,30
10	1,05	1,37	1,81	2,23	2,28	2,76	3,17	3,96	4,14
11	1,05	1,36	1,80	2,20	2,25	2,72	3,11	3,85	4,02
12	1,04	1,36	1,78	2,18	2,23	2,68	3,06	3,76	3,85
13	1,04	1,35	1,77	2,16	2,21	2,65	3,01	3,69	3,79
14	1,04	1,34	1,76	2,15	2,20	2,62	2,98	3,64	3,73
15	1,03	1,34	1,75	2,13	2,18	2,60	2,95	3,59	3,69
16	1,03	1,34	1,75	2,12	2,17	2,58	2,92	3,54	3,65
17	1,03	1,33	1,74	2,11	2,16	2,57	2,90	3,51	3,61
18	1,03	1,33	1,73	2,10	2,15	2,55	2,88	3,48	3,58
19	1,03	1,33	1,73	2,09	2,14	2,54	2,86	3,45	3,55
20	1,03	1,33	1,73	2,09	2,13	2,53	2,85	3,42	3,53
30	1,02	1,31	1,70	2,04	2,09	2,46	2,75	3,27	3,39
40	1,01	1,30	1,68	2,02	2,06	2,42	2,70	3,20	3,31
60	1,01	1,30	1,67	2,00	2,04	2,39	2,66	3,13	3,23
80	1,01	1,29	1,66	1,99	2,03	2,37	2,64	3,10	3,20
100	1,01	1,29	1,66	1,98	2,03	2,36	2,63	3,08	3,17

Annex A (informative)

Example of application of this part of ISO 11929: Determination of strontium-90 in soil including chemical separation

A.1 General

Strontium-90 should be separated from the soil by chemical treatment. The yield of the chemical separation, η , should be determined in advance using a Sr-85 standard added to the soil. After separation, the strontium solution should be mixed into a LSC cocktail. This mixture should be stored to let yttrium-90 build up to isotopic equilibrium. Then the strontium-90 content is measured via the yttrium-90 content. The counting efficiency ε should be determined in advance using a radioactive standard. The external background counting rate R_U can be determined in advance using a LSC cocktail with the same amount of demineralized water as the strontium solution added. The variation coefficient of chemical treatment θ should be determined in advance using five blank material samples spiked with Sr-90, treated in the same manner and measured for $t_m = 30\,000$ s each, or otherwise determined using the results of five samples taken from the soil in question treated and measured in parallel.

NOTE In this example the following two cases are demonstrated:

Case A: Application of this part of ISO 11929, in the case where the external background counting rate R_U and the variation coefficient of sample treatment θ are known (i.e. they are determined in advance by specific measurements, see above)

Case B: Application of this part of ISO 11929, in the case where the variation coefficient of sample treatment θ is unknown (i.e. it is not determined in advance by specific measurements). The variance coefficient should be determined using the results of five samples taken from the soil in question, treated and measured in parallel. In this case, the external background counting rate R_U will be taken from a former measurement.

A.2 Case A: The Variation coefficient of sample treatment is known

A.2.1 Data

Given:

$$\alpha = \beta = 5\%; k_{1-\alpha} = k_{1-\beta} = 1,645; 1 - \gamma = 95\%; k_{1-(\gamma/2)} = 1,96$$

Mass of soil sample $m = 100$ g (0,1 kg)

Number of samples $n_S = 1$; Number of blanks $n_0 = 1$

Durations of measurements: $t_U = 600\,000$ s; $t_S = t_0 = t_m = 30\,000$ s

Determined in advance:

Yield of chemical separation of strontium-90 $\eta = 0,57$ (57 %)

Counting efficiency $\varepsilon = 0,51$ (51 %)

Variation coefficient of chemical separation $\theta^2 = 0,018\,97$ i.e. $\theta = 0,137\,7$

External background counting rate $\rho_U \approx R_U = 0,024\,5$ s⁻¹

Measured:

$$N_0 = 866; R_0 = 866/30\,000 = 0,028\,87$$
 s⁻¹

$$N_S = 1943; R_S = 1943/30\,000 = 0,064\,77$$
 s⁻¹

A.2.2 Calculation of decision threshold

Figures, calculated using the formulae in Table 1:

$$c_1 = 1/15\,000 = 6,667 \times 10^{-5}$$

$$c_2 = 2;$$

$$c_5 = 0,5$$

The decision threshold R_n^* can be calculated using equation (17), Table 1.

With R_0 taken as an estimate for ρ_0 , this will be

$$R_n^* = 0,003\,002\text{ s}^{-1}$$

This equals an activity

$$G_n^* = \frac{0,003\,002\text{ s}^{-1}}{0,51 \times 0,57 \times 0,100\text{ kg}} = 0,103\text{ Bq / kg}$$

NOTE As the variance of chemical separation is not taken into account, the detection limit should be calculated as $G = 0,079\text{ Bq/kg}$

A.2.3 Calculation of detection limit

Figures, calculated using the formulae in Table 1:

$$c_3 = 2,648 \times 10^{-6};$$

$$c_4 = 1,645$$

The detection limit ρ_n^* can be calculated using equation (19), Table 1.

With R_0 as an estimate for ρ_0 , this will be

$$\rho_n^* = 0,010\,22\text{ s}^{-1}$$

This equals an activity of

$$g^* = \frac{0,010\,22\text{ s}^{-1}}{0,51 \times 0,57 \times 0,100\text{ kg}} = 0,35\text{ Bq / kg}$$

NOTE As the variance of chemical separation is not taken into account, the detection limit should be calculated as $g = 0,16\text{ Bq/kg}$.

A.2.4 Result and its confidence interval

The net counting rate is

$$R_n = R_b - R_0 = (0,064\,77 - 0,028\,87)\text{ s}^{-1} = 0,035\text{ s}^{-1}$$

This equals an activity of

$$A = \frac{0,035\text{ s}^{-1}}{0,51 \times 0,57 \times 0,100\text{ kg}} = 1,23\text{ Bq / kg}$$

With $(1 - \gamma) = 0,95$, $k_{(1 - \gamma/2)} = k_{(0,95)} = 1,96$ the term a for the confidence interval can be calculated using formula (22), Table 1, as

$$a = 0,011\,47\text{ s}^{-1}$$

and the confidence interval is

$$0,024\,4\text{ s}^{-1} \leq \rho_n \leq 0,047\,4\text{ s}^{-1}$$

This equals activities of

$$0,84\text{ Bq/kg} \leq A \leq 1,62\text{ Bq/kg}$$

The result reads

$$A = (1,23 \pm 0,39)\text{ Bq/kg}$$

A.3 Case B: The variation coefficient of sample treatment is unknown

A.3.1 Data

Given:

$$\alpha = \beta = 5\%; k_{1-\alpha} = k_{1-\beta} = 1,645; 1 - \gamma = 95\%; k_{1-(\gamma/2)} = 1,96$$

Number of samples taken from the soil in question $n_s = 5$

Mass of soil sample $m = 100\text{ g}$ (0,1 kg) each

Number of measurements $n_s = n_0 = 5$

Duration of measurements: $t_s = t_0 = 30\,000\text{ s}$

Determined in advance:

Yield of chemical separation of strontium-90 $\eta = 0,57$ (57 %)

Counting efficiency $\varepsilon = 0,51$ (51 %)

External background counting rate $\rho_u \approx R_u = 0,024\,5\text{ s}^{-1}$ taken from a former measurement

Measured:

Table A.1 — Measured numbers of pulses

Blanks		Samples	
No.	Number of pulses	No.	Number of pulses
1	966	1	1 832
2	676	2	2 259
3	911	3	2 138
4	856	4	2 320
5	676	5	1 649