

# INTERNATIONAL STANDARD

**ISO**  
**11934**

First edition  
1997-06-01

---

---

## **X and gamma radiation — Indirect- or direct-reading capacitor-type pocket dosimeters**

*Rayonnements X et gamma — Dosimètres individuels à condensateur  
pour lecture directe ou indirecte*



Reference number  
ISO 11934:1997(E)

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 11934 was prepared by Technical Committee ISO/TC 85, *Nuclear energy*, Subcommittee SC 2, *Radiation protection*.

This first edition cancels and replaces ISO 1758:1976, ISO 1759:1976 and ISO 4071:1978 which have been technically revised.

Annex A of this International Standard is for information only.

© ISO 1997

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization  
Case postale 56 • CH-1211 Genève 20 • Switzerland  
Internet central@iso.ch  
X.400 c=ch; a=400net; p=iso; o=isocs; s=central

Printed in Switzerland

# X and gamma radiation — Indirect- or direct-reading capacitor-type pocket dosimeters

## 1 Scope

This International Standard specifies the requirements for direct- and indirect-reading capacitor-type pocket dosimeters and the accessory electrometers used for personal dosimetry of X and gamma radiation.

The tests described in this International Standard are designed to be carried out on the dosimeter equipment associated with the operating accessories specified by the manufacturer.

NOTE — Electrical and mechanical characteristics of accessories are considered to belong to the scope of the International Electrotechnical Commission (Technical Committee 45).

This International Standard is not applicable without qualification to pocket dosimeters used to determine doses due to sources of pulsed radiation or to mixed fields of photons and neutrons. Furthermore, the dosimeters should not be used in radiation fields where the dose rate is likely to exceed their maximum dose rate capability as specified by the manufacturer.

## 2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this International Standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 6980:1996, *Reference beta radiations for calibrating dosimeters and dose-rate meters and determining their response as a function of beta-radiation energy.*

ISO 8529-3:<sup>1)</sup> *Neutron reference radiation — Part 3: Calibration of area and personal dosimeters and determination of response as a function of energy and angle of incidence.*

ISO 9227:1990, *Corrosion tests in artificial atmospheres — Salt spray tests.*

VIM:1993, *International Vocabulary of Basic and General Terms in Metrology*, ISO, OIM

<sup>1)</sup> To be published. (Revision, in part, of ISO 8529:1989)

### 3 Definitions

For the purposes of this International Standard, the definitions given in the VIM (see clause 2) and the following definitions apply.

#### 3.1 direct-reading capacitor-type pocket dosimeter

Device, used for individual monitoring, that permits direct reading of the radiation dose quantity.

NOTE — Such a device consists essentially of an ionization chamber connected to a capacitor. This capacitor is charged by a charging device which may or may not be built into the dosimeter, thus giving the charge indicator a deflection which can be read against a calibrated scale by means of an optical system. If the dosimeter is exposed to ionizing radiation, ionization in the chamber results in a decrease of the capacitor charge.

#### 3.2 indirect-reading capacitor-type pocket dosimeter

Capacitor-type pocket ionization chamber from which the radiation dose quantity can be read indirectly by means of a separate electrometer.

#### 3.3 type test

Test of one or more devices made to a certain design to show that the design meets certain specifications.

#### 3.4 influence quantity

Quantity which may have a bearing on the results of a measurement without being the objective of the measurement. [ISO 4037-3]

#### 3.5 reference conditions

Set of influence quantities for which the calibration factor is valid without any correction. [ISO 4037-3]

#### 3.6 standard test conditions

Range of values of a set of influence quantities under which a calibration or a determination of response is carried out. [ISO 4037-3]

#### 3.7 conventionally true value of a quantity

Best estimate of a quantity determined by a primary or secondary standard or by a reference instrument that has been calibrated against a primary or secondary standard.

#### 3.8 response

<of a detector> Ratio of the quantity evaluated from the detector reading to the conventionally true value of this quantity. [ISO 4037-2]

### 3.9 error of indication

Difference between the indicated value of a quantity and the conventionally true value of that quantity, at the point of interest.

### 3.10 relative intrinsic error

Quotient of the error of indication of a quantity by the conventionally true value of that quantity measured for a reference radiation under specified reference conditions.

It is expressed as a percentage.

### 3.11 calibration quantity

Physical quantity used to establish the characteristics of a dosimeter.

### 3.12 calibration factor

Conventionally true value of a quantity the dosimeter is intended to measure, divided by its indication.

### 3.13 kerma, $K$

Quotient of  $dE_{tr}$  by  $dm$ , where  $dE_{tr}$  is the sum of the initial kinetic energies of all the charged ionizing particles liberated by uncharged ionizing particles in a material of mass  $dm$ .

The SI unit of kerma is the joule per kilogram. The special name for the unit of kerma is gray (Gy).

NOTE 1 Air kerma free in air,  $K_a$ , is generally used in place of the quantity exposure,  $X$ . The SI unit of exposure is coulomb per kilogram, while the former unit is the röntgen,  $1 R = 2,58 \times 10^{-4} C/kg$ .

NOTE 2 Up to photon energies of 3 MeV, it can be assumed that the quantities "air kerma free in air" and "exposure" are approximately equivalent and that a value of  $K_a = 1 Gy$  corresponds to an exposure  $X = 29,45 mC/kg$ . Above 3 MeV and up to 9 MeV, the quantity "air kerma" can still be obtained using small ionization chambers with build-up caps. However, for this higher energy range, absorbed dose in tissue should be used as the calibration quantity. [ICRP 51; ISO 4037-2]

### 3.14 absorbed dose, $D$

Quotient of  $d\varepsilon$  by  $dm$ , where  $d\varepsilon$  is the mean energy imparted by ionizing radiation to matter of mass  $dm$ .

The SI unit of absorbed dose is the joule per kilogram. The special name for the unit of absorbed dose is gray (Gy).

NOTE — In quoting values of absorbed dose it is necessary to specify the material, e.g. tissue.

### 3.15 dose equivalent, $H$

Product of  $Q$  and  $D$  at a point in tissue, where  $D$  is the absorbed dose and  $Q$  is the quality factor at that point.

The SI unit of dose equivalent is joule per kilogram. The special name for the unit of dose equivalent is sievert (Sv).

NOTE — The quality factor for X, gamma and beta radiation is one.

### 3.16 personal dose equivalent, $H_p(d)$

Dose equivalent in soft tissue, at an appropriate depth,  $d$ , below a specified point on the body.

The SI unit of personal dose equivalent is joule per kilogram. The special name for the unit of dose equivalent is sievert (Sv).

### 3.17 ambient dose equivalent, $H^*(d)$

Dose equivalent, at a point in a radiation field, that would be produced by the corresponding expanded and aligned field, in the ICRU sphere at a depth,  $d$ , on the radius opposing the direction of the aligned field.

The SI unit of ambient dose equivalent is joule per kilogram. The special name for the unit of dose equivalent is sievert (Sv).

NOTE — For strongly penetrating radiation, a depth of 10 mm is currently recommended. The ambient dose equivalent for this depth is then denoted by  $H^*(10)$ . [ICRU 51]

## 4 Standard conditions for dosimeter tests

### 4.1 Reference conditions

The reference conditions for indirect or direct reading capacitor-type pocket dosimeters, in accordance with ISO 4037-3, are:

- temperature  $T = 20\text{ °C}$  ;
- relative humidity R.H. = 65 % ;
- atmospheric pressure  $p = 101,3\text{ kPa}$  ;
- radiation background: ambient dose equivalent rate  $\dot{H}^*(10) \leq 0,1\text{ }\mu\text{Sv/h}$ .

### 4.2 Standard test conditions

Except for temperature and humidity tests, the standard test conditions should be in accordance with ISO 4037-2:

- temperature  $T$  between  $18\text{ °C}$  and  $22\text{ °C}$  ;
- relative humidity R.H. between 50 % and 75 % ;

- atmospheric pressure  $p$  between 86 kPa and 106 kPa ;
- radiation background : ambient dose equivalent rate  $\dot{H}^*(10) < 0,25 \mu\text{Sv/h}$ .

The actual conditions should be indicated in the test report. They should not undergo large or rapid changes during a series of measurements.

When a dosimeter ionization chamber is not airtight, the reading of this dosimeter shall be corrected to reference temperature and atmospheric pressure conditions by multiplying it by  $(p_{\text{ref}}/p) \cdot (T_{\text{K}}/T_{\text{K,ref}})$  where  $p_{\text{ref}}$  and  $T_{\text{K,ref}}$  are the reference pressure and temperature,  $p$  and  $T_{\text{K}}$  are the actual measured ones. Here the temperatures are expressed as absolute temperatures (K).

### 4.3 Irradiation conditions

The reference radiations to be used shall be selected from table 1 (see 6.2.9). The calibration quantity shall be measured with a reference instrument which has itself been calibrated in a reference beam traceable to national standards.

Except in the case of special tests, the irradiation should impinge perpendicularly to the principal axis of the dosimeter.

The distance "source to dosimeter" is defined as the distance from the equivalent point source to the geometric centre of the sensitive volume of the dosimeter. When the dosimeter is calibrated on a phantom, it should be positioned with its back face in contact with the phantom.

If a radioactive source is used, the duration of the irradiation must be at least 100 times longer than the time to advance and retract the source. If this condition cannot be met, the amount of interfering irradiation resulting from the movement of the source must be determined.

The irradiation room and calibration devices should meet the following specifications :

- a) table and supports should be made of a low atomic number material and should have minimum mass;
- b) if several dosimeters are irradiated together, the distance between them should be such that the mutual influence on their reading is small. The difference in the readings of a dosimeter irradiated together with other dosimeters and a dosimeter irradiated alone in the same position should be less than 3 %.
- c) in order to subject several dosimeters to the same value of the calibration quantity, their supports should be placed on the same dose rate contour. If sufficient homogeneity cannot be achieved, the support may be made to rotate around the source.

## 5 Quantities

For most of the tests, the calibration quantities shall be air kerma or absorbed dose in tissue, whichever is appropriate for the particular type of radiation.

However, for the tests on the dependence of the dosimeter response on radiation energy and angle of incidence, results shall be reported in terms of personal dose equivalent.

## 6 Requirements concerning performance characteristics and test procedures

### 6.1 General

The following requirements apply to all tests.

- a) The tests shall always be performed on a number of dosimeters randomly selected from a batch.
- b) In order to avoid uncertainties due to geotropism, the readout procedure, including the position of the dosimeter, shall be specified by the manufacturer.
- c) For direct-reading dosimeters, the length of the scale and the number of divisions shall permit a reading that corresponds to 2% of the full-scale value.
- d) For indirect-reading dosimeters, the analog or digital readout shall permit a reading that corresponds to 2% of full range.
- e) The dosimeters shall be capable of being set to zero within a value that corresponds to 2 % of the full-scale value within three attempts by an experienced operator.
- f) For all the tests, a reference radiation from table 1 (see 6.2.9) shall be selected.

### 6.2 Tests with X or gamma radiation

#### 6.2.1 Zero-point stability

##### 6.2.1.1 Requirements

The leakage of charge shall not give rise to a change in the zero point exceeding 2% of the full-scale value for a dosimeter stored in standard test conditions for 8 h.

Direct-reading dosimeters: The change in reading shall not exceed 2 % of the full-scale value when the dosimeter is disconnected from the charging source.

Indirect-reading dosimeters: The change in reading shall not exceed 2 % of the full-scale value when the fully charged dosimeter is reconnected to the reader.

##### 6.2.1.2 Test procedures

Set a series of 10 dosimeters to zero, store them under standard test conditions and read them ( $r_i$ ) after 8 h. If the full-scale value of the dosimeter reading,  $r_{\max}$ , is less or equal to 1 mSv, the readings,  $r_i$ , shall be corrected for natural background dose.

Direct-reading dosimeters: Set a series of 10 dosimeters to zero, disconnect them and read them immediately ( $r_i$ ).

Indirect-reading dosimeters: Charge a series of 10 dosimeters, disconnect them from the charger, immediately reconnect and read them,  $r_i$ .

Calculate the deviations,  $d_i$ , due to leakage or disconnection, in percent:

$$d_i = 100 (r_i / r_{\max}), \text{ where } r_i \text{ are the readings and } r_{\max} \text{ is the full-scale value.}$$

## 6.2.2 Stability of reading

### 6.2.2.1 Requirement

The reading of the dosimeter shall not vary by more than 2 % of the full-scale value after a time of up to 8 h between the irradiation and a reading.

### 6.2.2.2 Test procedures

Irradiate 10 dosimeters to a reading between 50 % and 85 % of the full-scale value. Read them immediately ( $r_0$ ) and read them again ( $r_i$ ) every hour following irradiation, up to 8 h.

For each dosimeter, calculate the relative deviation,  $d_i$ , in percent:

$$d_i = 100 (r_i - r_0) / r_{\max} \text{ for each of the 8 readings, } r_i, \text{ where } r_0 \text{ is the initial reading and } r_{\max} \text{ is the full-scale value.}$$

## 6.2.3 Repeatability

### 6.2.3.1 Requirements

The repeatability of the measurements shall be determined using the same dosimeter subjected to the identical irradiation conditions, including laboratory and operator.

The results of the repeatability test for each dosimeter shall be such that  $2s/\bar{r} < 0,05$  where  $\bar{r}$  is the mean reading of the sampled dosimeter and  $s$  is the standard deviation.

### 6.2.3.2 Test procedure

Set 3 dosimeters to zero. Irradiate them to a reading between 50 % and 85 % of the full-scale value. Read them,  $r_i$ , and reset to zero. Repeat the test 10 times.

Calculate the mean,  $\bar{r}$ , of the 10 readings of each dosimeter, and the standard deviation,  $s$ :

$$\bar{r} = \frac{1}{10} \sum_{i=1}^{10} r_i$$

$$s = \sqrt{\frac{\sum_{i=1}^{10} (r_i - \bar{r})^2}{9}}.$$

## 6.2.4 Batch homogeneity

### 6.2.4.1 Requirements

Batch homogeneity shall be determined by observing the variability among the readings of dosimeters subjected to the same value of the radiation quantity under identical conditions, including laboratory and operator.

The results of the batch homogeneity test shall be such that  $2s/\bar{r} < 0.1$  where  $\bar{r}$  is the mean reading of the dosimeters and  $s$  is the standard deviation.

### 6.2.4.2 Test procedure

Irradiate 10 dosimeters to the same value of the radiation quantity between 50 % and 85 % of the full-scale value and read them ( $r_i$ ).

Calculate the mean of the readings,  $\bar{r}$ , and the standard deviation,  $s$ :

$$\bar{r} = \frac{1}{10} \sum_{i=1}^{10} r_i$$

$$s = \sqrt{\frac{\sum_{i=1}^{10} (r_i - \bar{r})^2}{9}}.$$

## 6.2.5 Lower limit of detection

### 6.2.5.1 Requirements

For capacitor-type pocket dosimeters, the lower limit of detection is given by the lowest scale reading that can be distinguished from the fluctuations of background for a given time interval including leakage. The lower limit of detection,  $r_{\min}$ , is determined by taking twice the standard deviation,  $2s$ , of the mean value of the difference in the readings for a series of dosimeters at the beginning and at the end of an 8 h storage in standard test conditions, rounding up  $2s$  to the next scale division.

After a series of irradiations with values of the radiation quantity corresponding to a significant fraction of the full-scale value of the dosimeter the lower limit of detection shall not vary from the originally determined limit.

### 6.2.5.2 Test procedure

Set 10 dosimeters as close as possible to zero and read them,  $r_{0,i}$ . Store for 8 h in standard test conditions and read again,  $r_i$ .

Calculate the absolute difference between the two readings,  $\Delta r_i = |r_i - r_{0,i}|$ , the mean value,  $\bar{\Delta r}$ , and the standard deviation,  $s$ :

$$\bar{\Delta r} = \frac{1}{10} \sum_{i=1}^{10} \Delta r_i$$

$$s = \sqrt{\frac{\sum_{i=1}^{10} (\Delta r_i - \bar{\Delta r})^2}{9}}$$

The lower limit of detection,  $r_{\min}$ , is then  $2s$  rounded up to the next scale division.

To test the stability of the lower limit of detection take the 10 dosimeters that have been tested for their lower limit. Irradiate them successively 10 times to at least 80 % of the full-scale value and reset to zero. At the end of the irradiation, determine the lower limit of detection  $r'_{\min}$  as described. Compare with the original value  $r_{\min}$ .

### 6.2.6 Relative intrinsic error

#### 6.2.6.1 Requirement

The relative intrinsic errors,  $I$ , shall not exceed 10%.

#### 6.2.6.2 Test procedure

Set 10 dosimeters to zero. Irradiate each to at least 3 different values of the radiation quantity, equally spaced between 20 % and 100 % of the full-scale value. Read,  $r_i$ , and reset the dosimeters to zero after each irradiation.

Calculate the relative intrinsic errors,  $I_i$ , in percent:

$$I_i = 100 (r_i - r_0) / r_0$$

where  $r_0$  is the conventionally true value of the radiation quantity used in the irradiation, and  $r_i$  are the readings of the dosimeters.

### 6.2.7 Linearity

#### 6.2.7.1 Requirements

For the test of linearity the coefficient of variation,  $s_I / \bar{I}$ , for the series of relative intrinsic error measurements made in 6.2.6 shall be less than 0,1, where  $\bar{I}$  is the mean value of the intrinsic error and  $s_I$  is the standard deviation.

### 6.2.7.2 Test procedure

Calculate the mean,  $\bar{I}$ , of the intrinsic errors,  $I_i$ , in 6.2.6 and their standard deviation,  $s_I$ :

$$\bar{I} = \frac{1}{30} \sum_{i=1}^{30} I_i$$

$$s_I = \sqrt{\sum_{i=1}^{30} (I_i - \bar{I})^2 / 29}$$

The coefficient of variation is then  $s_I/\bar{I}$ .

### 6.2.8 Memory effect

#### 6.2.8.1 Requirement

Following an accidental irradiation of the dosimeter to a high dose the drift due to leakage shall not exceed 20 % of the full-scale value following 24 h storage in standard test conditions.

#### 6.2.8.2 Test procedure

Select 3 dosimeters that have been tested for zero point stability (test 6.2.1) and irradiate them once to a value of the radiation quantity corresponding to  $k$  times the full-scale value of the dosimeter where  $k$  is typically 50.

This irradiation shall be performed in more than 1 h and less than 5 h and shall not exceed the dose rate limit specified by the manufacturer.

One hour after the end of the irradiation, reset the dosimeters to zero and store in standard test conditions. Read them after 24 h,  $r_i$ .

Calculate for each dosimeter the drift due to leakage,  $S$ , in percent:

$$S = 100 r_i / r_{\max}, \text{ where } r_i \text{ is the reading after 24 h and } r_{\max} \text{ is the full-scale value of the dosimeters.}$$

### 6.2.9 Energy dependence of response

#### 6.2.9.1 General

This is a type test to determine the response of the dosimeters subjected to X or gamma radiation as a function of radiation energy.

Since pocket dosimeters are used as personal dosimeters, the energy response test should be performed in receptor present conditions. The dosimeters should be irradiated in front of the ISO water phantom, for normal radiation incidence (see ISO 4037-3), and the

radiation quantity used should be the personal dose equivalent,  $H_p(d)$  (see ICRU 47, ICRU 51).

The dosimeter to be tested is placed in a manner that is representative for its use in practice, with the centre of its sensitive volume, as specified by the manufacturer, in contact with and at the centre of the front face of the ISO water phantom in position P. For this calibration point, P, where the conventionally true value of air kerma free in air,  $K_a$ , has been determined, the personal dose equivalent  $H_p(10)$  is calculated using the conversion coefficients in table 1.

**Table 1 — Conversion coefficients (ISO 4037-3)**

Radiation quality	Mean energy or energy keV	Conversion coefficients $H_p(10)/K_a$ Sv/Gy
$^{241}\text{Am}$	59,6	1,89
$^{137}\text{Cs}$	662	1,21
$^{60}\text{Co}$	1250	1,15
Mono-energetic radiation	3000	1,12
	6000	1,11
	10000	1,11
<b>Narrow filtered X-radiation</b>		
N-20	16	0,27
N-40	33	1,17
N-60	48	1,65
N-80	65	1,88
N-100	83	1,88
N-120	100	1,81
N-200	164	1,57
N-250	208	1,48

### 6.2.9.2 Requirements

The response of the dosimeter to any radiation energy, chosen from table 1 within the useful energy range specified by the manufacturer, shall be within 30 % of the response to a calibration reference radiation,  $E_r$ , chosen from the same table.

For dosimeters intended to cover the energy range between 3 MeV and 10 MeV, the response should be within - 50% and + 100% of the response to the calibration reference radiation,  $E_r$ .

### 6.2.9.3 Test procedure

Perform successively, with 3 different dosimeters, the following operation.

Place one dosimeter with the centre of its sensitive volume at the calibration point P in the centre of the front face of the ISO water phantom P. Irradiate the dosimeter to a conventionally true value of personal dose equivalent  $H_p(10)$  between 50 % and 80 % of the full-scale value. All energies  $E_k$  from table 1 which are within the useful energy range specified by the manufacturer shall be used. Read  $r(E_k)$  and reset the dosimeter to zero.

Calculate the response,  $R(E_k)$ , for each energy  $E_k$  :

$$R(E_k) = r(E_k) / H_p(10)$$

Determine the deviation for the response,  $d(E_k)$ , for the energy  $E_i$  in percent:

$$d(E_k) = 100 [R(E_k)/R(E_r) - 1]$$

where  $E_r$  is the calibration reference radiation.

## 6.2.10 Angular dependence of response

### 6.2.10.1 Requirements

This is a type test to determine the angular response of the dosimeters with the photon energies in table 2 which are within the useful energy range specified by the manufacturer.

Within the interval  $0^\circ$  to  $60^\circ$  and for the same personal dose equivalent, the ratio of the dosimeter reading at angle  $\alpha$  to that at normal incidence shall be within 20 % of the values of the ratio  $R_\alpha$  in table 2, for  $^{137}\text{Cs}$  or  $^{60}\text{Co}$  gamma rays and 50% for the N-40, N-60, N-80 narrow spectrum filtered X-rays or  $^{241}\text{Am}$  gamma-rays. The angles  $\alpha$  are set by rotating the phantom around the vertical and horizontal axes at the calibration point, P, this point being at the centre of the sensitive volume of the dosimeter.

### 6.2.10.2 Test procedure

Place the dosimeter on the front face of the ISO water phantom in the same way as for the energy response test 6.2.9. Irradiate under normal incidence to a conventionally true value of personal dose equivalent  $H_p(10)$  between 50 % and 80 % of the full-scale value, read  $r_{0^\circ}$  and reset the dosimeter to zero. Rotate the phantom successively clockwise at angles of  $\alpha = 30^\circ, 45^\circ, 60^\circ$  and anticlockwise at angles of  $\alpha = -30^\circ, -45^\circ, -60^\circ$  around the vertical and horizontal axes passing through P, and irradiate the dosimeter to the same conventionally true value of personal dose equivalent as for  $0^\circ$  and read  $r_\alpha$ . All energies from table 2 which are within the useful energy range specified by the manufacturer shall be used.

Calculate the ratios  $r_\alpha/r_{0^\circ}$  and determine the deviation,  $d_\alpha$ , in percent from the angular response  $R_\alpha = H_p(10)_\alpha/H_p(10)_{0^\circ}$  for the radiation sources in table 2.

$$d_\alpha = 100 \left( \frac{r_\alpha/r_{0^\circ}}{R_\alpha} - 1 \right)$$

Table 2 — Angular response values (ISO 4037-3)

Radiation source	Angular response $R_\alpha$			
	$\alpha = 0^\circ$	$\alpha = 30^\circ$	$\alpha = 45^\circ$	$\alpha = 60^\circ$
N-40	1	0,96	0,87	0,73
N-60	1	0,96	0,89	0,77
N-80 or $^{241}\text{Am}$	1	0,97	0,91	0,79
$^{137}\text{Cs}$	1	1,01	1,01	0,98
$^{60}\text{Co}$	1	1	1,01	0,99

### 6.3 Temperature and humidity tests

#### 6.3.1 Requirements

**Temperature** : within the temperature range - 10 °C to + 40 °C, the dosimeter reading shall not differ by more than + 20 % from its reading in standard test conditions. For dosimeters designed to be used between - 25 °C and +50 °C, the readings shall not differ by more than 50 % from the reading under standard test conditions.

**Humidity**: at a temperature of 35 °C and a relative humidity of 90 % , the dosimeter reading shall not differ by more than 10 % from its reading under standard test conditions.

#### 6.3.2 Test procedure

Use a test apparatus that fits into an environmental chamber supporting 10 dosimeters on the isodose contour of an isotopic gamma source ( $^{137}\text{Cs}$  or  $^{60}\text{Co}$ ).

Condition the dosimeters for 1 h in the chamber at the test conditions before irradiation. Put in the source and irradiate for a fixed time such that the reading of the dosimeters is between 50 % and 80 % of the full-scale value. Remove the source but leave the dosimeters in the chamber until the total time in the test conditions is at least 4 h.

Record the dosimeter reading,  $r_{T,hr}$ , as a function of temperature  $T$  and relative humidity R.H. and apply a temperature and pressure correction to reference conditions for unsealed dosimeters.

These tests should be performed with a series of 10 dosimeters each successively :

- in 65% relative humidity at - 10 °C, 20 °C and 40 °C ;
- in 90 % relative humidity, at 35 °C.

The readings,  $r_r$ , at 20 °C and 65 % relative humidity are taken as reference data. Calculate the ratios :  $r_{-10^\circ,65\%}/r_r$ ,  $r_{40^\circ,65\%}/r_r$  and  $r_{35^\circ,90\%}/r_r$ .

Determine the deviations,  $d_{T,RH}$ , in percent:  $d_{T,RH} = 100 (r_{T,RH}/r_r - 1)$ .

For dosimeters designed to be used between - 25 °C and +50 °C additional tests shall be performed with a series of 10 dosimeters each in 65% relative humidity at - 25 °C and at +50 °C.

Calculate the ratios :  $r_{-25^\circ,65\%}/r_r$  and  $r_{50^\circ,65\%}/r_r$  and the deviations,  $d_{T,RH}$ , in percent.

## 6.4 Mechanical tests

### 6.4.1 Drop test

#### 6.4.1.1 Requirement

The dosimeter shall be able to withstand a drop without affecting its reading in excess of 10 % of the full-scale value. Following this test the dosimeter should function correctly.

#### 6.4.1.2 Test procedure

The dosimeter shall be supported by a device which can release it suddenly, reproducibly and without initial velocity above a hardwood surface, e.g. oak, made from a single piece 5 cm thick, resting on a concrete base. The height of drop should be 1,5 m measured between the upper surface of the board and the lowest point of the dosimeter.

- a) Irradiate 10 dosimeters to 80 % of their full-scale value and read them,  $r_{0,i}$ .
- b) Submit each of them to two drops and read them again,  $r_{1,i}$ .
- c) Calculate the deviation,  $d_{1,i}$ , in percent:  $d_{1,i} = 100(r_{1,i} - r_{0,i})/r_{\max}$  where  $r_{\max}$  is the full-scale value,
- d) Reset the dosimeters to zero, irradiate to the same value as before and read them,  $r_{2,i}$ .
- e) Calculate the deviation,  $d_{2,i}$ , in percent:  $d_{2,i} = 100(r_{2,i} - r_{0,i})/r_{\max}$ .

## 6.4.2 Vibration

### 6.4.2.1 Requirement

The mean instrument reading following the vibration test shall not vary from the initial reading by more than 10% of the full-scale value.

### 6.4.2.2 Test procedure

Irradiate a series of 10 dosimeters to 80 % of the full-scale value and read them,  $r_{0,i}$ .

Submit the dosimeters to a harmonic load of 2g for 15 min in each of three orthogonal directions (1, 2, 3) at one or more frequencies in each of the following ranges : 10 Hz to 21 Hz, 22 Hz to 33 Hz. After each vibration interval of 15 min, read them ( $r_{1,i}$ ,  $r_{2,i}$ ,  $r_{3,i}$ ).

Calculate the mean values  $r_0$ ,  $r_1$ ,  $r_2$ ,  $r_3$  and the deviations,  $d_1$ ,  $d_2$  and  $d_3$ , in percent:

$$d_1 = (r_1 - r_0)/r_{\max}$$

$$d_2 = (r_2 - r_0)/r_{\max}$$

$$d_3 = (r_3 - r_0)/r_{\max}$$

where  $r_{\max}$  is the full-scale value.

Specify any damage revealed by visual inspection.

### 6.4.3 Immersion (for dosimeters specified watertight)

#### 6.4.3.1 Requirement

Immersion at a depth of 30 cm in water for 2 h shall not change the reading of a dosimeter by more than 10 % of the full-scale value.

#### 6.4.3.2 Test procedure

Irradiate a series of 10 dosimeters to 80 % of the full-scale value and read them,  $r_{0,i}$ .

Immerse the dosimeters in water at a depth of 30 cm for 2 h. Remove the dosimeters from the water, dry and read them,  $r_{1,i}$ .

Calculate the deviations,  $d_{1,i}$ , in percent:

$$d_{1,i} = (r_{1,i} - r_{0,i}) / r_{\max}$$

where  $r_{\max}$  is the full-scale value.

### 6.4.4 Atmospheric pressure

#### 6.4.4.1 Requirement

A sealed dosimeter shall achieve the accuracy requirements specified in 6.2.6. An unsealed dosimeter shall withstand the same variations when an air density correction is applied to its reading.

#### 6.4.4.2 Test procedure

Select a series of 10 dosimeters which have successfully passed the relative intrinsic error test (6.2.6). Using the apparatus of test 6.3, irradiate these dosimeters to 80 % of the full-scale value in an environmental chamber, under the following atmospheric pressures  $p_i$ :

$$p_0 = 101,3 + 5 \text{ kPa}$$

$$p_1 = 60 + 5 \text{ kPa}$$

$$p_2 = 120 + 5 \text{ kPa}$$

The temperature should correspond to those under standard test conditions (see 4.2). Read the dosimeters,  $r_i$ , and correct for air density if the dosimeter is unsealed.

Calculate the deviations,  $d_i$ , in percent:

$$d_i = 100 (r_i - r_0) / r_{\max}$$

where

$r_{\max}$  is the full-scale value;

$r_0$  is the dosimeter reading at normal pressure.

## 6.5 Utilization tests

### 6.5.1 Salt spray

#### 6.5.1.1 Requirement

Following the salt spray test, the dosimeter shall meet the requirements of the batch homogeneity test specified in 6.2.4.

#### 6.5.1.2 Test procedure

Place the dosimeters in a spray cabinet complying with ISO 9227.

For 100 h during a one-week period, spray with a sodium chloride solution of mass concentration 50 g/l at 35 °C. For the rest of the week (68 h), stop spraying but maintain the temperature at 35 °C.

Rinse the dosimeters with water, dry them in air at 20 °C and perform the batch homogeneity test in 6.2.4.

### 6.5.2 Decontamination

#### 6.5.2.1 Requirement

The susceptibility of dosimeters to retain contamination should be as low as possible. All exterior surfaces of the instrument should be hard and smooth, with as few joints as practicable.

#### 6.5.2.2 Test procedure

Set 3 dosimeters to zero and read their background after 1 h,  $r_{0,i}$ .

Contaminate each dosimeter with  $^{140}\text{La}$  or another appropriate radionuclide in one physical form: dry dust, moist dust or solution.

Decontaminate the dosimeters in accordance with the manufacturer's directions.

Reset to zero and read the background reading after 1 h,  $r_{1,i}$ .

Compare the readings  $r_{1,i}$  and  $r_{0,i}$ .

## 6.6 Response to beta radiation

### 6.6.1 Requirement

The dosimeter response to beta radiation should be stated.

### 6.6.2 Test procedure

Set 5 dosimeters to zero and irradiate them with a  $^{90}\text{Sr}/^{90}\text{Y}$  source, in standard test conditions and in accordance with the recommendations of ISO 6980, to at least 20 % of the full-scale value.

Following irradiation, read the dosimeters,  $r_i$ .

Calculate the response  $R = \bar{r}/l_0$  where  $l_0$  is the conventionally true value of the beta dose and  $\bar{r}$  is the mean value of the readings,  $r_i$ .

## 6.7 Response to neutrons

### 6.7.1 Requirement

The dosimeter response to neutrons should be stated.

### 6.7.2 Test procedure

Set 5 dosimeters to zero and irradiate them with an  $^{241}\text{Am-Be}$  neutron, in standard test conditions and in accordance with the recommendations of ISO 8529, to at least 20 % of the full-scale value.

Following irradiation, read the dosimeters,  $r_i$ .

Calculate the response  $R = \bar{r}/l_0$  where  $l_0$  is the conventionally true value of the neutron personal dose equivalent and  $\bar{r}$  is the mean value of the readings,  $r_i$ .

## 7 Markings and information

Indirect- or direct-reading capacitor-type pocket dosimeters should bear the following markings and information :

- a) Dosimeter type and identification number.
- b) An indication of the radiation dose quantity it is designed to measure. The scale of the dosimeter should be marked correspondingly in micro- or milligray ( $\mu\text{Gy}$  or  $\text{mGy}$ ) or micro- or millisievert ( $\mu\text{Sv}$  or  $\text{mSv}$ ).
- c) Sealed or unsealed.