
**Ophthalmic optics — Contact lens care
products — Guidelines for determination
of shelf-life**

*Optique ophtalmique — Produits d'entretien pour lentilles de contact —
Lignes directrices pour la détermination de la durée de conservation*

STANDARDSISO.COM : Click to view the full PDF of ISO 13212:2011



STANDARDSISO.COM : Click to view the full PDF of ISO 13212:2011



COPYRIGHT PROTECTED DOCUMENT

© ISO 2011

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 13212 was prepared by Technical Committee ISO/TC 172, *Optics and photonics*, Subcommittee SC 7, *Ophthalmic optics and instruments*.

This second edition cancels and replaces the first edition (ISO 13212:1999), of which it constitutes a minor revision. In particular, the normative references have been updated.

Introduction

The purpose of stability tests of contact lens care products is to obtain sufficient information to enable the manufacturer to establish an appropriate shelf-life and identify any unique storage conditions required to appear on the labelling for the product.

The quality of a contact lens care product is determined by its content of active ingredient(s), its purity and its physicochemical and microbiological properties. It is important to take into account the possible interaction of the container/closure with the contents.

The stability studies are intended to ascertain how the quality of a product varies as a function of time and under the influence of a variety of environmental factors.

On the basis of the information thus obtained, storage conditions are recommended which will guarantee the maintenance of the quality of the product, in relation to its safety, performance and acceptability, throughout the proposed shelf-life.

The design of the finished-product stability studies for a care product is based on the knowledge obtained from studies on the active ingredient(s) and from the development studies.

STANDARDSISO.COM : Click to view the full PDF of ISO 13212:2011

Ophthalmic optics — Contact lens care products — Guidelines for determination of shelf-life

1 Scope

This International Standard provides guidance on the design of stability studies for use in gathering information to enable determination of the shelf-life of contact lens care products.

This International Standard does not address studies designed to obtain information to establish the in-use stability (i.e. notice of discard date) of contact lens care products.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 14729, *Ophthalmic optics — Contact lens care products — Microbiological requirements and test methods for products and regimens for hygienic management of contact lenses*

ISO 14730, *Ophthalmic optics — Contact lens care products — Antimicrobial preservative efficacy testing and guidance on determining discard date*

ISO 18369-1, *Ophthalmic optics — Contact lenses — Part 1: Vocabulary, classification system and recommendations for labelling specifications*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 18369-1 apply.

4 General requirements

4.1 The specified shelf-life of the contact lens care product shall be based on the evaluation of the results of stability studies.

4.2 Analytical methods that have been validated and are stability-indicating shall be used to assay for active ingredients. Validation includes, but is not limited to, being able to differentiate between the active ingredient and its degradation products. The test methods used shall be described in full.

5 Determination of finished-product stability

5.1 Objective

The objective of stability testing on contact lens care products is to provide data for determining the time period during which the product performance characteristics are maintained and to define appropriate storage conditions.

The design of the stability tests is based on the known properties of the active ingredient(s), the properties of the chosen formulation and the recommendations for use of the product.

The relevant assay methods shall be determined prior to the start of the stability testing.

The specifications proposed from the time of manufacture to the end of the proposed shelf-life shall reflect, as far as possible, the results of the stability studies, particularly in relation to any parameters which could have a bearing on performance and safety and on product acceptability.

5.2 Study methods

5.2.1 General

Before starting stability studies, a suitable testing plan should be set up, taking into consideration the properties of the active ingredient(s) as well as the proposed mode of action of the care product.

5.2.2 Real-time studies

These studies should be carried out under a range of controlled test conditions, when applicable, which will enable the shelf-life and the storage requirements which are to appear on the product container label/package insert to be defined. This will normally include studies which are intended to allow the properties of the product at temperatures between 20 °C and 30 °C to be evaluated. However, 25 °C ± 2 °C should be used as the mean kinetic testing temperature.

For each study, the mean temperature, the ranges of temperature and mean humidity, if applicable, shall be stated in the stability report.

These studies are intended to support the initial shelf-life request and, for shelf-life extensions, any changes that could significantly impact the safety and performance of the product (e.g. certain changes in formulation, packaging materials or manufacturing methods).

NOTE Real-time studies are performed in conjunction with accelerated ageing studies to establish an initial shelf-life.

5.2.3 Studies under varying storage conditions

These studies shall be carried out to provide important additional information. They can fulfil a number of objectives, such as

- supporting the initial shelf-life request, by complementing the limited results of the early real-time studies, because decomposition, if it occurs, is likely to be accelerated,
- producing useful data at an early stage of development, demonstrating the effects of adverse storage in the packaging and product, and enabling storage conditions and suitable labelling to be provided, and
- supporting a request to extend the shelf-life.

The various test conditions should be stated. Depending on the nature and objectives of the stability study, the following points may need to be considered:

- a) Various test temperatures: three or more, particularly if long-term real-time data are unavailable. In addition, the effect of low temperatures may need to be considered, such as below $-15\text{ }^{\circ}\text{C}$ (freezer), $2\text{ }^{\circ}\text{C}$ to $8\text{ }^{\circ}\text{C}$ (refrigerator) and freeze-thaw cycling.
- b) High humidity: relative humidity $> 75\%$. Storage under high humidity conditions applies particularly to solid dosage forms. For products such as solutions, suspensions, etc. contained in packs designed to provide a permanent barrier to water loss, storage under high humidity is not necessary. However, low humidity can have an adverse effect on products packaged in semi-permeable containers.
- c) Elevated temperature and humidity in combination: e.g. temperature of $40\text{ }^{\circ}\text{C}$ associated with a relative humidity of 75% , possibly with the effects of cycling between different temperatures and humidities.
- d) Since most contact lens care products are water based, relative humidity less than 40% should be considered: for example, $25\text{ }^{\circ}\text{C}$ and 35% relative humidity.
- e) Light: either natural daylight or defined artificial illumination.

An example of a stability testing plan for the finished product is shown in Annex A.

5.3 Description of the product under study

5.3.1 Number and nature of the batches tested

The number of batches tested shall be stated with the batch number, details of the composition, date of manufacture, size of the batch, and the batch number and the name of the manufacturer of the active ingredient(s) used.

Normally, three batches of the finished product are studied. If the number of batches tested is less than three, it shall be justified.

Satisfactory performance of the product in the smallest size of container with the highest surface/volume ratio shall allow the extension of shelf-life to containers which are up to eight times larger in volume.

5.3.2 Primary container

The product batches shall be packed in the primary containers proposed for marketing. The smallest primary container size should be tested. Satisfactory performance in the smallest size shall allow marketing of the product in containers with eight times the volume present in the smallest size.

Details of the packaging should be stated, including

- a) type(s) of container and closure, and nature of the constituent material(s),
- b) nature of any desiccant if used, and
- c) the complete range of sizes of the product proposed for marketing.

5.4 Characteristics

5.4.1 General

The following characteristics should be studied:

- a) those in the finished-product specification that are likely to be affected by storage, and
- b) those not monitored routinely at the time of manufacture but which may be indicative of the stability/instability of the particular product (e.g. dissolution of tablets).

5.4.2 Physical characteristics of the finished product

The following physical characteristics should be tested:

- a) physical properties specific to the product, such as tablet hardness and hygroscopicity, or pH, colour, clarity and viscosity for solutions;
- b) important quality parameters such as the *in vitro* dissolution, moisture content (e.g. in relation to any desiccant used in the packaging) and particle size;
- c) any other physical characteristics of the product that must be known in order to assess product stability.

5.4.3 Microbial characteristics

The following microbial characteristics shall be tested:

- a) antimicrobial activity of finished products to be marketed for chemical disinfection of contact lenses shall be tested in accordance with ISO 14729, unless otherwise justified;
- b) preservative efficacy of preserved products shall be tested at the end of shelf-life, in accordance with ISO 14730, unless otherwise justified;
- c) sterility of sterile products (or provide valid data to demonstrate maintenance of package integrity);
- d) microbial limits of non-sterile products shall be given.

5.4.4 Chemical characteristics of the finished product

The following chemical characteristics should be determined:

- a) assay of the active ingredient(s), where possible;
- b) consideration of other agents (such as antimicrobial preservatives and antioxidants);
- c) any other chemical characteristics that must be known in order to assess the quality of the product.

5.4.5 Characteristics of primary container interactions

If necessary, carry out a study of the interaction of the container and closure with the contents in any case where this is a risk.

5.4.6 Performance characteristics

If stability cannot be established by chemical method, it should be followed by relevant performance characteristics. Performance tests should mimic as closely as possible the in-use condition; otherwise, the rationale for the design of test(s) should be described.

5.5 Evaluation methods

The test procedures applied to the stability tests on the finished product shall be described in full and validated.

5.6 Presentation of results

The results shall be summarized (e.g. as tables and graphs). For each product batch tested, the initial results (at the time of manufacture) and the results obtained during storage should be given. Results of real-time data should be recorded as they become available, up to the proposed shelf-life.

5.7 Discussion, interpretation and conclusions

The discussion and conclusion shall provide a critical evaluation of the suitability of the test methods used, the results obtained and the proposed shelf-life specification. This should take into account the safety and performance requirements of the product at the end of shelf-life.

If it was necessary to carry out any further studies due to significant changes in relevant properties, an explanation should be given, together with the results of these studies.

A minimum of three months' real-time data at 25 °C should be available, supported by data from accelerated stability studies. Such data would not normally be expected to be suitable for prediction of a shelf-life in excess of two years. Any extension of the shelf-life should be based on additional real-time study results.

Studies under accelerated test conditions will increase the decomposition and may permit some extrapolation of the room-temperature shelf-life from that which would otherwise be acceptable. However, such studies would always need to be supplemented by long-term real-time studies, and normally at least three months' real-time data should be available.

If product batches in test demonstrate a decreasing stability profile, the shelf-life proposed and any overage should be based on the stability of the least stable test result, unless an explanation can be given.

The shelf-life (expiration date) shall be proposed on the primary package to be used for sale.

If there is evidence that batches of the stored product as packed for sale are stable at temperatures up to 30 °C, the product need bear no special temperature storage instructions. However, if there is evidence that the product must be stored under defined conditions of storage, this shall be stated on the container label and the package insert (if included) and the outer carton. The maximum (or minimum) storage temperature should be stated in degrees Celsius (e.g. store below 25 °C; store in a refrigerator at 2 °C to 8 °C; do not refrigerate — store above 8 °C). These storage recommendations on the label/package insert shall reflect conditions found in the country or countries in which the product is to be placed on the market.

Temperatures acceptable for accelerated extrapolation of the expiration date should maintain the same mechanism of decomposition. Generally, temperatures at or below 45 °C will be acceptable.

NOTE For every 10 °C increase in temperature, the rate of decomposition generally increases by a factor of two. Unless otherwise justified on the basis of kinetic evidence, this acceleration factor should be used.

5.8 Ongoing stability

Where data on routine production batches are not provided, ongoing stability studies should be carried out on at least two of the first production batches and the results recorded.