
**Air quality — Determination of ozone
in ambient air — Ultraviolet photometric
method**

*Qualité de l'air — Dosage de l'ozone dans l'air ambiant — Méthode
photométrique dans l'ultraviolet*

STANDARDSISO.COM : Click to view the full PDF of ISO 13964:1998



Contents	Page
1 Scope	1
2 Interferences	1
3 Principle	2
4 Reagents and materials	3
5 Apparatus	3
6 Procedure	6
7 Expression of results	8
8 Test report	9
Annex A (informative) Some reported chemical interferences for UV photometric ozone analysers	10
Annex B (informative) Correction for ambient nitric oxide	11
Annex C (informative) Typical performance specifications for a UV photometric ozone analyser	12
Bibliography	13

STANDARDSISO.COM : Click to view the full PDF of ISO 13964:1998

© ISO 1998

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization
Case postale 56 • CH-1211 Genève 20 • Switzerland
Internet iso@iso.ch

Printed in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 13964 was prepared by Technical Committee ISO/TC 146, *Air quality*, Subcommittee SC 3, *Ambient atmospheres*.

Annexes A, B and C of this International Standard are for information only.

STANDARDSISO.COM : Click to view the full PDF of ISO 13964:1998

[STANDARDSISO.COM](https://standardsiso.com) : Click to view the full PDF of ISO 13964:1998

Air quality — Determination of ozone in ambient air — ultraviolet photometric method

WARNING — This document calls for use of ozone gas. Ozone is a toxic gas and good laboratory practice should limit indoor ozone concentrations to less than $200 \mu\text{g}/\text{m}^3$ (volume fraction 1×10^{-7}). Consult a reference text for more details on hazards and safety of ozone. Any excess should be vented into an activated charcoal scrubber (with negligible back-pressure) or outdoors well away from any sampling intake.

1 Scope

This International Standard specifies an ultraviolet (UV) photometric method for the determination of ozone in ambient air. It is applicable to the determination of ozone concentrations in the range $2 \mu\text{g}/\text{m}^3$ (volume fraction of 1×10^{-9}) to $2 \text{ mg}/\text{m}^3$ (volume fraction of 1×10^{-6}). Furthermore, this International Standard uses the reference conditions of $25 \text{ }^\circ\text{C}$ and $101,25 \text{ kPa}$; however, reference temperatures of $0 \text{ }^\circ\text{C}$ and $20 \text{ }^\circ\text{C}$ are also acceptable.

For calibration, this International Standard specifies ultraviolet photometry as the primary reference procedure because of its proven accuracy and specificity to ozone. The use of secondary reference procedures (often called transfer standards), including non-UV methods, is allowed if they have been previously calibrated by the primary UV reference procedure (see ISO 6879 for definitions).

2 Interferences

The UV photometric method is not subject to interference from any of the common gaseous air pollutants at ambient concentrations less than $0,2 \text{ mg}/\text{m}^3$ (volume fraction of 1×10^{-7}). However, there are reported interferences of about $2 \mu\text{g}/\text{m}^3$ (volume fraction of 1×10^{-9}) and about $8 \mu\text{g}/\text{m}^3$ (volume fraction of 4×10^{-9}) in equivalent ozone for nitrogen dioxide and sulfur dioxide concentrations at a volume fraction of $0,5 \times 10^{-6}$, respectively. Additionally, there are some reported interferences for instruments using a manganese dioxide ozone scrubber; these are tabulated in annex A of this International Standard.

Particulate matter, if not removed, will accumulate in the sampling line and may cause a measurable destruction of ozone.

Ambient nitric oxide in the sample air will to some extent react with the ambient ozone during the residence time of the ambient air in the sampling line; corrections for this effect are given in annex B of this International Standard.

For ozone analysers with single absorption cells, any ambient pollutant that absorbs the photometer's UV radiation and that varies in concentration within the cycle time of the instrument may cause an interference. A similar limitation applies at a site where the ambient ozone concentration itself varies on this time scale, as for example, at traffic intersections.

NOTE In general, if interferences are suspected, it is preferable to use another method (for example, the chemiluminescence method using ethylene, see [2]) rather than attempting to scrub out the interfering agent, because the instability of ozone makes the testing and proving of scrubbers particularly difficult.

3 Principle

Sample air is drawn continuously through an optical absorption cell where it is irradiated lengthwise by monochromatic radiation, centred on 253,7 nm, from a stabilized low-pressure mercury (Hg) discharge lamp. The UV radiation which passes through the absorption cell is measured by a sensitive photodiode or photomultiplier detector and converted to a measurable electrical signal. Absorption of this radiation by the sample air within the absorption cell is a measure of the ambient ozone concentration. An ozone catalytic converter is used to selectively remove the ozone from the sample stream. This can be done either for the sample absorption cell (single-cell configuration) or a separate but identical reference absorption cell (dual-cell configuration); either configuration is acceptable.

The Beer-Lambert law on light absorption is used to relate the measured UV transmittance to the path length of the absorption cell, the ozone absorption coefficient at 253,7 nm and the ozone concentration. The following equation gives this relationship:

$$\text{Transmittance} = I/I_0 = \exp(-aCd) \quad (1)$$

where

I/I_0 is the transmittance of the ozone sample, i.e. the ratio of the irradiance (commonly called light intensity) falling on the detector when the absorption cell contains sample air to the irradiance when the cell contains ozone-scrubbed air;

a is the ozone absorption coefficient at 253,7 nm, in square metres per microgram; ($a = 1,44 \times 10^{-5} \text{ m}^2/\mu\text{g}$; see [3]);

C is the mass concentration of ozone, at the sample temperature and pressure in the absorption cell, in micrograms per cubic metre;

d is the optical path length, in metres.

Most modern commercial ozone analysers measure the temperature and pressure of the sample air in the absorption cell. Using this data, an internal microprocessor automatically calculates the measured ozone concentration relative to some chosen reference condition. For analysers without this automated pressure and temperature compensation, the concentrations need to be corrected manually to the chosen reference condition (T_{ref} °C and 101,25 kPa) according to the following equation:

$$C_{\text{ref}} = \frac{101,25}{P} \cdot \frac{(T + 273,15)}{(T_{\text{ref}} + 273,15)} \cdot C \quad (2)$$

where

C_{ref} is the mass concentration of ozone, at the reference temperature (T_{ref}) and pressure (101,25 kPa), in micrograms per cubic metre;

T is the sample temperature of the absorption cell, in degrees Celsius;

P is the sample pressure in the absorption cell, in kilopascals;

T_{ref} is the reference temperature, in degrees Celsius;

C is the mass concentration of ozone from equation (1), at the sample temperature (T) and pressure (P) in the absorption cell, in micrograms per cubic metre.

4 Reagents and materials

4.1 Sample line, made of material that is inert to ozone, such as glass or fluorocarbon polymer, and shall be as short as possible to keep the residence time to a minimum. Any ambient nitric oxide present in the sample air will react with some of the ozone during the residence time in the sampling line. This decay of ozone is a function of the ambient ozone and nitric oxide concentrations (see note below). Calculations have shown that if the residence time is less than 0,5 s, then the decay in initial ozone will be less than 1 % for most ambient ozone and nitric oxide concentrations encountered. Hence, it is recommended that the sampling line length and flowrate be chosen such that the residence time within the sampling line be as short as possible; a residence time of up to 5 s is allowed.

It is recommended that a particle filter (see 4.2) be installed at the sampling inlet to keep the whole inlet system clean. In case there is a sample line or manifold preceding the inlet particle filter (see 4.2), it shall be clean. Proper precautions shall be taken to prevent condensation inside the sample line, for example, by appropriate heating of the line.

NOTE 1 Annex B describes the correction for the reaction of ambient ozone with nitric oxide in the sampling line.

NOTE 2 Whenever a sampling line is cleaned or replaced, it may take several hours to equilibrate with ambient conditions.

4.2 Particle filter and its support, made of material inert to ozone, such as polytetrafluoroethylene (PTFE), and capable of removing all particles likely to alter the performance of the analyser. It shall be changed on a regular basis, depending on the ambient particle concentrations at the sampling site; however, the period between filter changes should not exceed 14 days. This is necessary because excessive accumulation of particles on the filter can cause loss of ozone from the sample air.

NOTE Frequently, a filter pore size of 5 μm is used.

NOTE Generally, new filters need some time to be conditioned by the ambient atmosphere. As a result, measured ozone concentrations have been observed to decrease temporarily by 5 % to 10 % for periods of 5 min to 15 min immediately following filter changes.

4.3 Zero air, required in the analyser calibration procedure. If synthetic air is used, the oxygen content shall be at the normal atmospheric concentration of $(20,9 \pm 2)\%$.

The zero air shall be free of ozone, nitrogen oxides and any other interfering substance that can cause an undesired measurable positive or negative response in the UV photometer. Different zero air supplies may have different remnant impurities and, in turn, these can cause different UV transmittances of the air samples. Hence, the zero air supplied to the photometer during the I_0 measurement [see equation (1)] shall be the same as that used for generation of the calibration ozone concentrations.

NOTE Details on a system for making zero air from ambient air may be found in [4].

5 Apparatus

5.1 UV photometric ambient ozone analyser

The components of a typical UV photometric ozone measuring system with a single-cell configuration are shown schematically in Figure 1. When assembled, they become part of a UV photometric ambient ozone analyser with specifications conforming to those listed in annex C. Commercial instruments that meet these specifications are readily available in either single or dual cell configurations. The important components of such a system are as follows.

- a) **Ultraviolet absorption cell**, constructed of material inert to ozone, such as fluorocarbon polymer, borosilicate glass, fused silica or fluorocarbon-coated metal. It shall be mechanically stable so that any optical alignments are not affected by vibration or change in ambient temperature. Provision shall be made for measuring the temperature and pressure of the gas in the absorption cell [see 5.1 h) and 5.1 i)].

- b) **Ultraviolet source lamp**, i.e. a low-pressure mercury vapour discharge lamp that emits monochromatic UV radiation centred at 253,7 nm. The UV radiance of this lamp shall be electronically stabilized so that the analyser meets the required performance specifications (see annex C for typical values); any variation in lamp radiance during the measurement cycles will result in a measurement bias or noise. The lamp's mercury line at 185 nm (which photolyses oxygen to produce ozone) shall be eliminated by means of a high-silica glass enclosure or shield.

NOTE Spectral data show that ambient water vapour should not interfere in the UV wavelength region of interest here (about 200 nm to 300 nm). However, it has been observed in some analysers of one manufacturer that the quartz window of the UV source lamp was covered with a significant number of microscopic scratches. This resulted in a measurable variation in light scattering at the internal window surface due to changing humidity in the sample air. Replacement of such faulty windows eliminated this interference (see [1]).

- c) **UV detector**.

Quantitatively, 99,5% of the radiation sensed by the detector shall be at 253,7 nm. The response of this sensor and its associated electronics shall be sufficiently stable so that the analyser meets the required performance specifications.

NOTE Vacuum photodiodes with cesium telluride sensitization meet the requirement of sensitivity at 253,7 nm and negligible sensitivity at other mercury emission wavelengths. Other detectors, such as photomultiplier tubes, also satisfy this purpose.

- d) **Ozone-specific scrubber (converter) with by-pass valve**, whose active component shall be a material that selectively catalyses the destruction of ozone in the ambient air sample stream.

NOTE 1 Manganese dioxide on a metal substrate has been found to do this satisfactorily. Even though this particular compound does not remove any of the other more common pollutants, it has been found to partly remove some organic compounds which may be present in some ambient air. Annex A lists such interfering chemical compounds.

A three-way by-pass valve is used to direct ambient air alternately either through or around the ozone-specific scrubber. It shall be made of material that is inert to ozone, such as a fluorocarbon polymer.

NOTE 2 A significant decrease in response to ambient ozone of a UV ozone analyser may be an indication of scrubber failure. Normally, manufacturers give an average lifetime of such scrubbers; however, the actual lifetime will depend on the sampling location. For example, a city location, which is subject to higher pollutant concentrations than a remote location, may deactivate the scrubber prematurely.

- e) **Air pump**, installed at the end of the system (see Figure 1) to pull sample air through the instrument; capable of maintaining a flowrate of several litres per minute.
- f) **Air flow control device**, installed immediately ahead of the air pump so that the air flow through the analyser can be properly set.
- g) **Air flowmeter**, installed after the UV absorption cell (see Figure 1) to measure the sample flowrate in the instrument, capable of measuring air flow up to several litres per minute.
- h) **Temperature indicator**, capable of measuring the temperature of the sample air in the ultraviolet absorption cell with an accuracy of $\pm 0,5$ °C.
- i) **Pressure indicator**, capable of measuring the pressure of the sample air in the absorption cell with an accuracy of $\pm 0,2$ kPa.




Figure 1— Schematic diagram of a typical UV photometer system

5.2 Apparatus for calibration by primary reference procedure

A simplified schematic of an ozone calibration system is shown in Figure 2, and comprises the following.

NOTE An equivalent system would be one in which an extra zero air line, from the same zero air source, bypasses the ozone generator so as to facilitate switching zero or ozonized air into the manifold or directly into the UV calibration photometer. This would require additional inert gas lines, connections and valves.

- a) **Primary UV calibration photometer**, used for only that purpose and meticulously maintained under stable laboratory conditions.

Similar to the UV photometer of the ambient analyser, it shall consist of a low pressure mercury discharge lamp, a single or dual absorption cell, and detection and signal processing electronics. To prevent ozone generation within the absorption cell, a high-silica glass window (or equivalent) shall be inserted between the mercury lamp and the absorption cell to remove the 185 nm Hg wavelength (which photolyzes oxygen to produce ozone) and yet allow transmission of the 253,7 nm Hg measurement wavelength. The detector shall be capable of measuring the transmittance through the absorption cell at the 253,7 nm wavelength, with less than 0,5 % of the radiation detected being from other wavelengths (vacuum diodes with cesium telluride sensitization meet this requirement). The length of the light path through the absorption cell shall be known to

an accuracy better than $\pm 0,5\%$, and the cell and associated plumbing shall be designed to minimize surface loss of ozone. Provision shall be made for measuring the temperature and pressure of the gas in the absorption cell [see 5.2 g) and 5.2 h)]. Figure 1 provides a simplified schematic of a UV photometer system, except that the ozone-specific scrubber with by-pass valve is not used.

- b) **Ozone generator**, to generate the stable ozone concentrations that are measured by the primary UV calibration photometer described in 5.2 a).

The ozone generator shall be capable of producing steady ozone concentrations in the range of interest at the required flowrate and throughout the period of calibration. If a variable ozone generator is not available, the calibration system shall include a means of suitably diluting the ozone with additional zero air (see 4.3), and a mixing chamber shall be installed before the output manifold.

- c) **Air flow controllers**, capable of maintaining steady air flows throughout the period of calibration.
- d) **Air flowmeters**, capable of measuring the required air flows.

NOTE The actual air flows are not used in the calculations but are used only for achieving the nominal concentrations required in the calibration procedure.

- e) **Air pump (suction)** to draw the required sample flow through the absorption cell. A nominal flowrate of 2 l/min is recommended.
- f) **Output manifold**, made of materials inert to ozone, such as borosilicate glass or fluorocarbon polymer. It shall be of sufficient diameter and be vented to ensure equal pressure inside and outside the manifold. The vent shall be located so as to prevent intrusion of ambient air.
- g) **Temperature indicator**, for the absorption cell in the primary UV calibration photometer, accurate within $\pm 0,5\text{ }^\circ\text{C}$.
- h) **Pressure indicator**, for the absorption cell in the primary UV calibration photometer, accurate within $\pm 0,2\text{ kPa}$.




Figure 2 — Schematic diagram of a typical ozone calibration flow system for use with UV photometry

6 Procedure

6.1 Operation of the ambient ozone analyser

Install the instrument in a suitable location and, if necessary, provide temperature control of the measurement room to minimize any temperature dependence of the instrument. Follow the manufacturer's operating instructions for the analyser to set the various operating parameters correctly, including the UV source lamp intensity, the sample flowrate and, if applicable, activation of the electronic temperature and pressure compensation. Also, follow the diagnostic procedure as outlined in the manufacturer's operations manual for verifying that the important instrument functions are within their performance specifications. Introduce sample air into the instrument and record the ozone

concentration by means of a suitable recording device (for example, chart recorder, electronic data acquisition system, etc.).

During continuous operation of the instrument, checks on the instrument zero, span and operational parameters shall be made at least once per week. This is necessary to reveal any change in instrument stability or ozone scrubber efficiency. It is further recommended that a multipoint calibration be carried out every 3 to 4 months.

6.2 Calibration of the ambient ozone analyser

6.2.1 Principle

A schematic diagram of a calibration system is shown in Figure 2. Various ozone concentrations (in air) are generated and accurately measured with the primary UV calibration photometer. These ozone calibration atmospheres are simultaneously sampled by the ambient ozone analyser via a common manifold. The analyser responses are plotted against the ozone concentrations as measured by the primary UV calibration photometer. Alternatively, a secondary reference procedure (often called transfer standard) can be calibrated against the primary UV calibration photometer and then used to calibrate the ambient analyser at the sampling location.

NOTE See [4] for details on the principles of operation of a UV photometer.

6.2.2 Calibration procedure using the primary UV calibration photometer

Follow the steps below for the detailed calibration procedure. During this procedure, operate the ambient ozone analyser at its normal flowrate and temperature. The calibration includes measurements using zero air (see 4.3) and at least five ozone concentrations [using the ozone generator as described in 5.2 b)] which shall be reasonably spaced to cover the ambient range. For all calibrations, the input flow to the manifold shall exceed, by at least 20 %, the total flow required by the instruments attached to the manifold, with the excess appropriately vented at atmospheric pressure.

Carry out the following steps in the calibration procedure:

- a) assemble the apparatus as shown in Figure 2; for ozone analysers with the automated electronic temperature and pressure compensation, ensure that this circuit has been activated so as to yield corrected output values;
- b) introduce zero air into the manifold; if necessary, change the zero control setting of the ambient ozone analyser to indicate an output reading close or equal to zero; this will become the zero offset of the instrument;
- c) record the output values of the ozone analyser and the primary UV photometer; for ozone photometers and analysers without the automated electronic temperature and pressure compensation, measure and record the temperature of the absorption cell and its internal pressure (usually measured at the UV absorption cell outlet);
- d) adjust the ozone generator to produce the highest ozone concentration required; record all values as in step c) above; calculate the corresponding ozone concentration according to equation (3) of 7.1; if so desired, change the span control setting of the ambient ozone analyser to indicate an output reading close to or equal to the ozone concentration found from equation (3) above, and if the span and zero settings are not independent, then also repeat the foregoing sequence of steps b) to d);
- e) generate and measure, according to step d) above, at least four other ozone concentrations over the range of interest; or else, the ozone generator may be kept constant at its highest intensity and the zero air flow may be varied to accomplish the same purpose;
- f) plot the ozone analyser output values versus the corresponding calculated ozone concentrations at the chosen reference conditions of temperature and pressure;
- g) determine the linear calibration function of the ozone analyser by means of a linear regression analysis and report the appropriate response factor or slope (for example, in units of micrograms per cubic metre per volt), the intercept or zero offset, and their associated errors.

NOTE 1 Repeating this calibration to estimate the precision of the calibration is optional.

NOTE 2 A small fraction of the ozone may be lost upon contact with the internal surfaces of the absorption cell of the primary UV photometer. Therefore, for maximum accuracy, such losses of ozone should be quantitatively determined and used to correct the calculated concentration. (See [4] and [7]).

NOTE 3 If the calibration response factor [determined in step g) above] is unexpectedly different from the performance specification of the instrument, it is recommended that the source of this difference be investigated and rectified.

6.2.3 Calibration using a secondary reference procedure

Give consideration to the use of a secondary reference procedure (transfer standard) whenever the primary reference procedure is not readily available, such as at the sampling site. Therefore, whenever such a secondary reference procedure is used, modify the calibration procedure of 6.2.2 to suit the particular calibration method in use (see [4], [5] and [6] which describe in detail such alternative secondary reference procedures).

The secondary reference procedure shall meet the following requirements:

- calibration against the primary UV calibration photometer shall be done at least once per year;
- its accuracy shall be established against this same primary UV calibration photometer and it shall be maintained within $\pm 5\%$ between successive primary calibrations.

NOTE 1 The recommended secondary reference procedure for calibration is a second (portable) UV photometer system with its own zero air and ozone supply (see [7] and [8]). Another acceptable secondary reference procedure is the gas-phase titration of excess nitric oxide by ozone, or vice versa (for details, see [5], [6] and [9]).

NOTE 2 There are other laboratory-based secondary reference procedures for calibration that are possibly less accurate and precise (for example, see [6], [10] and [11]). These include:

- the neutral buffered potassium iodide — potassium bromide - thiosulfate method (KIBRT);
- the boric acid-buffered potassium iodide (BAKI) method;
- the neutral-buffered potassium iodide — potassium bromide (KIBR) method;
- a stable ozone generator with a variable output.

7 Expression of results

7.1 Calibration ozone concentration

Using the measured parameters of the primary UV calibration photometer, calculate the ozone concentration in the manifold at the reference conditions (T_{ref} °C and 101,25 kPa) from the following equation:

$$C_{\text{ref}} = \frac{-101,25}{P} \cdot \frac{(T + 273,15)}{(T_{\text{ref}} + 273,15)} \cdot \frac{\ln(I/I_0)}{1,44 \times 10^{-5}} \cdot \frac{-1}{d} \quad (3)$$

where

C_{ref} is the calibration ozone concentration at the reference conditions, in micrograms per cubic metre;

d is the optical path length of the primary UV calibration photometer cell, in metres;

I/I_0 is the transmittance of the ozone sample, i.e. the ratio of the irradiance (commonly called light intensity) falling on the detector when the absorption cell contains sample air to the irradiance when the cell contains ozone-scrubbed air;

P is the pressure in the photometer cell, in kilopascals;

T is the temperature of the photometer cell, in degrees Celsius;

T_{ref} is the reference temperature, in degrees Celsius.

NOTE To calculate the ozone volume fraction or mixing ratio, see 7.2.

7.2 Ambient ozone concentration

Convert the ambient analyser readings to ozone concentration by using the appropriate linear calibration function or response factor [determined in step g) of 6.2.2]. Report the results as micrograms per cubic metre at the chosen reference conditions, or as a volume fraction. For ozone, at 101,3 kPa, a volume fraction of $1,0 \times 10^{-6}$ is equivalent to $2\,141 \mu\text{g}/\text{m}^3$ at $0\text{ }^\circ\text{C}$, $1\,995 \mu\text{g}/\text{m}^3$ at $20\text{ }^\circ\text{C}$, and $1\,962 \mu\text{g}/\text{m}^3$ at $25\text{ }^\circ\text{C}$.

7.3 Repeatability

The repeatability shall be less than 5 % (with a probability of 95 %) for the generation and UV photometric measurement of ozone calibration atmospheres; the same shall apply for the secondary reference procedure.

NOTE Studies have shown that for UV photometric ozone calibrations the repeatability can be as low as $\pm 1\%$ or $\pm 1 \times 10^{-9}$ when expressed as a volume fraction, whichever is greater, if care is taken in operating the system. Even though there are some variabilities associated with the ambient analyser and calibration system, it is expected that the repeatability of the ozone analysis of ambient air can be within $\pm 3,5\%$ (see [11]).

7.4 Accuracy

The accuracy of the method is estimated to be better than $\pm 4\%$ of the measured concentration.

NOTE The accuracy of the ambient ozone UV photometric method will be dependent on the accuracy of the procedure to which it is referenced. The accuracy of the UV photometer primary reference procedure will be limited by the accuracy of the coefficient of absorption ($\pm 1,5\%$) and the measurements of the UV transmittance, pressure, temperature, ozone losses, etc. (see [3]). When a secondary reference procedure is used, additional inaccuracies inherent to the procedure will be introduced.

8 Test report

The test report shall contain the following information:

- a) a reference to this International Standard;
- b) documentation on the calibration photometer or method, and its operator;
- c) the date and results of the calibration (see clause 7);
- d) any peculiarities encountered during the calibration.

Annex A (informative)

Some reported chemical interferents for UV photometric ozone analysers

Interfering compound (at a volume fraction of 1×10^{-6})	Response (in % of concentration)	Bibliographic reference
Styrene	20	12
<i>trans</i> - β -Methylstyrene	>100	12
Benzaldehyde	5	12
<i>o</i> -Cresol	12	12
Nitrocresol	100	12

NOTE In [12], no response was reported at volume fractions of up to 1×10^{-6} for the following compounds: toluene, peroxyacetyl nitrate, biacetyl, peroxybenzoyl nitrate, methyl nitrate, n-propyl nitrate, n-butyl nitrate, methanethiol, methyl sulfide and ethyl sulfide. More recently, however, another study (see [13]) found that UV analysers responded positively by about $0,1 \times 10^{-6}$ volume fraction of equivalent ozone for about 1×10^{-6} volume fraction of toluene.