

INTERNATIONAL
STANDARD

ISO
14402

First edition
1999-09-01

**Water quality — Determination of phenol
index by flow analysis (FIA and CFA)**

*Qualité de l'eau — Détermination de l'indice phénol par analyse en flux (FIA
et CFA)*

STANDARDSISO.COM : Click to view the full PDF of ISO 14402:1999



Reference number
ISO 14402:1999(E)

Contents	Page
1 Scope	1
2 Normative references	1
3 Determination of phenol index (without distillation) after extraction	1
3.1 Principle	1
3.2 Interferences	1
3.3 Reagents	2
3.4 Apparatus	4
3.5 Sampling	6
3.6 Procedure	7
3.7 Calculation of results	8
4 Determination of phenol index (without extraction) after distillation	9
4.1 Principle	9
4.2 Interferences	9
4.3 Reagents	9
4.4 Apparatus	10
4.5 Sampling	13
4.6 Procedure	13
4.7 Calculation of results	14
5 Expression of results	14
6 Precision and accuracy	14
7 Test report	14
Annex A (informative) Statistical data	15
Bibliography	17

© ISO 1999

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization
 Case postale 56 • CH-1211 Genève 20 • Switzerland
 Internet iso@iso.ch

Printed in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 14402 was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 2, *Physical, chemical biochemical methods*.

Annex A of this International Standard is for information only.

STANDARDSISO.COM : Click to view the full PDF of ISO 14402:1999

Introduction

Methods for determination of water quality using flow analysis and automatic wet chemical procedures are particularly suitable for the processing of large sample series at a high analysis frequency.

Differentiation is needed between flow injection analysis (FIA) [1, 2] and continuous flow analysis (CFA) [3]. Both methods include automatic dosage of the sample into a flow system (manifold) where the analytes in the sample react with the reagent solutions on their way through the manifold. The sample preparation may be integrated in the manifold. The reaction product is measured in a flow detector.

Phenol index is an analytical convention. It represents a group of aromatic compounds which under the specific reaction conditions form coloured condensation products. The analytical result is expressed in terms of phenol concentration.

This International Standard describes two methods: the determination of phenol index (without distillation) after extraction, and the determination of phenol index (without extraction) after distillation.

It should be investigated whether and to what extent particular problems will require the specification of additional marginal conditions.

STANDARDSISO.COM : Click to view the full PDF of ISO 14402:1999

Water quality — Determination of phenol index by flow analysis (FIA and CFA)

1 Scope

This International Standard specifies two methods for the determination of the phenol index in waters of different origin (such as ground waters, surface waters, seep waters, and waste waters) in mass concentrations of 0,01 mg/l to 1 mg/l (in the undiluted sample). In particular cases, the range of application may be adapted by varying the operating conditions. Clause 3 describes the determination of phenol index (without distillation) after extraction, and in clause 4 the determination of phenol index (without extraction) after distillation is given.

2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this International Standard. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards:

ISO 3696:1987, *Water for analytical laboratory use — Specification and test methods*.

ISO 5667-3:1994, *Water quality — Sampling — Part 3: Guidance on sample handling and preservation*.

ISO 6439:1990, *Water quality — Determination of phenol index — 4-Aminoantipyrine spectrometric methods after distillation*.

3 Determination of phenol index (without distillation) after extraction

3.1 Principle

The sample is fed into a continuously flowing carrier stream and mixed with also continuously flowing solutions of 4-aminoantipyrine and potassium peroxodisulfate. Phenolic compounds in the sample are oxidized by potassium peroxodisulfate, and the resulting quinones react with 4-aminoantipyrine, forming coloured condensation products. These are extracted in a flow extraction unit from the aqueous phase into chloroform. The chloroform phase is separated by a suitable phase separator (e.g. a hydrophobic semipermeable membrane), and the absorbance of the organic phase is measured spectrometrically in a flow spectrometer at 470 nm to 475 nm. More information on this analytical technique is given in the references [6 to 9].

It is absolutely essential that the test described in this International Standard be carried out by suitably qualified staff.

3.2 Interferences

3.2.1 Chemical interferences

Under the prevailing reaction conditions, aromatic amines will also form condensation products with 4-aminoantipyrine, leading to positive bias.

Interferences can occur when the sample, after the addition of the reagent solutions, does not reach a pH of 10,0 to 10,5. In particular this may occur in the cases of strongly acidic, strongly alkaline and buffered samples. In these cases, the sample is adjusted to a pH between 5 and 7 prior to addition of the reagent solutions.

Further information on interferences is given in [5].

3.2.2 Physical interferences arising from applying CFA and FIA

If the samples contain particulate matter, refer to 3.5 (last paragraph). Turbid samples do not cause interferences with the determination. In the event of coloured samples, check whether the colour can be extracted with chloroform, and determine the sample blank without the addition of reagents R1 and R2. The difference in response between the two measurements shall be taken into account with the evaluation (according to 3.7).

The interlaboratory trial (see clause 6 and annex A) has shown that detergents in waste water can strongly influence the determination, because the foam produced in the flow system can disturb on the one hand the steam distillation of volatile phenols (phenol index after distillation, see clause 4, and on the other hand the phase segmentation and phase separation procedures (phenol index after extraction, see clause 3). In general such interferences can easily be discovered.

In the case of significant detergent content, this International Standard is only applicable for phenol mass concentrations above 0,1 mg/l.

3.3 Reagents

Use only reagents of recognized analytical grade quality. The reagent blank value shall regularly be checked (see 3.6.3). The solutions used for the flow system shall be degassed. If not stated otherwise, it is recommended to degas the solutions under reduced pressure, because by this procedure the solutions are simultaneously purified.

WARNING — Phenol is toxic and can easily be absorbed through the skin. Chloroform is toxic and cancerogenic. Waste containing these substances should be disposed of appropriately.

3.3.1 Water, of grade 1 in accordance with ISO 3696

3.3.2 Potassium hydroxide, KOH

3.3.3 Sodium hydrogencarbonate, NaHCO₃

3.3.4 4-aminoantipyrine (4-amino-2,3-dimethyl-1-phenyl-3-pyrazolin-5-one), C₁₁H₁₃N₃O

3.3.5 Potassium peroxydisulfate, K₂S₂O₈

3.3.6 Phenol, C₆H₅OH

3.3.7 Boric acid, H₃BO₃

3.3.8 Ethanol, C₂H₅OH, 96 % mass fraction

3.3.9 2-Propanol, C₃H₇OH, 100 % mass fraction

3.3.10 Sulfuric acid, $\rho(\text{H}_2\text{SO}_4) = 1,84 \text{ g/ml}$

3.3.11 Hydrochloric acid, HCl, 50 % mass fraction

3.3.12 Potassium hydroxide solution, $c(\text{KOH}) = 1,0 \text{ mol/l}$

3.3.13 Buffer solution

Dissolve in a 1 000 ml graduated flask in approximately 500 ml of water (3.3.1): 23 g of sodium hydrogencarbonate (3.3.3), 27 g of boric acid (3.3.7), and 35 g of potassium hydroxide (3.3.2) and make up to volume with water.

The pH of the buffer solution is approximately 10,3. The solution is stable for 1 month.

3.3.14 Carrier solution (symbol C in Figure 1)

Use water (3.3.1) degassed under reduced pressure.

3.3.15 4-Aminoantipyrine solution I (symbol R1 in Figures 1 and 2)

Dissolve in a 100 ml graduated flask 0,5 g of 4-aminoantipyrine (3.3.4) in approximately 50 ml of buffer solution (3.3.13), and make up to volume with buffer solution (3.3.13).

Degas the solution, e.g. by membrane filtration.

Prepare fresh solution every day.

3.3.16 Potassium peroxodisulfate solution (symbol R2 in Figures 1 and 2)

Dissolve in a 100 ml graduated flask 5 g of potassium peroxodisulfate (3.3.5) in approximately 90 ml of water (3.3.1), adjust to pH 11 with potassium hydroxide solution (3.3.12) and make up to volume with water.

Degas the solution, e.g. by membrane filtration.

Prepare fresh solution daily.

3.3.17 Chloroform, CHCl_3 (symbol Org in Figures 1 and 2)

Degas the chloroform solution either by membrane filtration or for 3 min in an ultrasonic bath.

3.3.18 Phenol stock solution, $\rho = 1\ 000\ \text{mg/l}$

Dissolve in a 1 000 ml graduated flask 1,000 g of phenol (3.3.6) in water (3.3.1) and make up to volume with water. Use only colourless phenol crystals.

The cooled solution (2 °C to 5 °C) is stable for one month.

3.3.19 Phenol standard solution I, $\rho = 10\ \text{mg/l}$

Pipette 1 ml of the stock solution (3.3.18) into a 100 ml graduated flask, and make up to volume with water (3.3.1).

The cooled solution (2 °C to 5 °C) is stable for one week.

3.3.20 Phenol standard solution II, $\rho = 1\ \text{mg/l}$

Pipette 10 ml of the standard solution I (3.3.19) into a 100 ml graduated flask, and make up to volume with water (3.3.1).

The cooled solution (2 °C to 5 °C) is stable for one week.

3.3.21 Calibration solutions

Prepare the calibration solutions according to the origin of the sample and the expected concentrations by diluting the phenol standard solution I or II respectively (3.3.19 or 3.3.20).

Prepare a minimum of at least five calibration solutions per working range.

Proceed as follows for working ranges I and II, if using e.g. six calibration solutions:

a) Working range I, (0,1 mg/l to 1 mg/l):

Into each of a series of 100 ml graduated flasks pipette 1 ml, 3 ml, 5 ml, 6 ml, 8 ml and 10 ml respectively of the standard solution I (3.3.19), and make up to volume with water (3.3.1).

The concentration of phenol in these calibration solutions is 0,1 mg/l, 0,3 mg/l, 0,5 mg/l, 0,6 mg/l, 0,8 mg/l and 1,0 mg/l, respectively.

b) Working range II (0,01 mg/l to 0,1 mg/l):

Into each of a series of 100 ml graduated flasks pipette 1 ml, 3 ml, 5 ml, 6 ml, 8 ml, and 10 ml respectively of the standard solution II (3.3.20), and make up to volume with water (3.3.1).

The concentration of phenol in these calibration solutions is 0,01 mg/l, 0,03 mg/l, 0,05 mg/l, 0,06 mg/l, 0,08 mg/l and 0,1 mg/l, respectively.

Prepare fresh calibration solutions each day.

3.4 Apparatus

3.4.1 Flow injection analysis system (FIA)

The flow injection system (FIA) shall comprise the following components (see Figure 1):


- a) reagent reservoirs;
- b) low pulsation pump with specific pump tubing, for flowrates as shown in Figure 1, as an example;
- c) displacement bottle for the feeding of the chloroform;
- d) sample injector with suitable injection volumes;
- e) extraction cell with phase segmentor and phase separator (e.g. hydrophobic semipermeable membrane of PTFE).

EXAMPLES Membrane thickness: 150 μm to 200 μm ; pore size: 0,5 μm to 2 μm ; porosity: 75 %.

- f) transport tubes and reaction coils, internal diameter 0,5 mm to 0,8 mm, tube connections and T-connections of chemically inert plastic, with a minimal dead volume;
- g) spectrometric detector with flow cell, of optical path length 10 mm, wavelength 470 nm to 475 nm.
- h) recording unit (e.g. strip chart recorder, integrator or printer/plotter).

NOTE In general, peak heights are measured.

- i) autosampler, if required.

**Key**

- C: Carrier solution (3.3.14)
 R1: 4-aminoantipyrine solution I (3.3.15)
 R2: Potassium peroxodisulfate solution (3.3.16)
 Org: Chloroform (3.3.17)
 1 Pump (flowrates in ml/min)
 2 Injector
 600 µl [working range 0,01 to 0,1 mg/l phenol]
 200 µl [working range 0,1 to 1,0 mg/l phenol]
 3 Reaction coil: 60 cm/∅ int. 0,5 mm
 4 Reaction coil: 80 cm/∅ int. 0,5 mm
 5 Extraction unit: 160 cm/∅ int. 0,7 mm
 5.1 Phase segmentor,
 5.2 Phase separator
 6 Detector: optical pathlength: 1 cm, wavelength: 470 nm to 475 nm
 7 Waste

Figure 1 — Example of a flow injection system for the determination of 0,01 mg/l to 1,0 mg/l phenol index without distillation and with extraction (according to 3.4.1)


3.4.2 Continuous flow analysis (CFA)

The continuous flow analysis system shall comprise the following components (see Figure 2):

- autosampler allowing a reproducible introduction of the sample or of the carrier liquid;
- reagent reservoirs;
- low pulsation pump with specific, chemically inert pump tubes, with flowrates as shown in Figure 2, as an example;
- displacement bottle for the feeding of the chloroform, if required;
- manifold with highly reproducible gas bubble, sample, and reagent introduction, with appropriate transport systems and extraction systems, and connection assemblies, e.g. of glass, chemically inert plastics or metal, and with appropriate separator for the separation of the organic phase from the aqueous phase;
- spectrometric detector with flow cell, optical pathlength 0,5 cm to 5 cm, wavelength 470 nm to 475 nm
- recording unit (e. g. strip chart recorder, integrator or printer/plotter).

NOTE 1 In general, peak heights are measured.

NOTE 2 A CFA system with an internal diameter of 2 mm is described in Figure 2. Other internal diameters (e.g. approximately 1 mm) may also be used.



Key

- | | | | |
|-----|---|-----|--|
| R1 | 4-aminoantipyrine solution I (3.3.15) | 4.2 | Phase separator |
| R2 | Potassium peroxodisulfate solution (3.3.16) | 4.3 | Reaction coil: 150cm/Ø int. 2 mm |
| Org | Chloroform (3.3.17) | 5 | Detector: optical pathlength: 0,5 cm to 5 cm, wavelength: 470 nm to 475 nm |
| 1 | Pump (flowrates in ml/min) | 6 | Segmentation gas (air) |
| 2 | Reaction coil: 40 cm/Ø int. 2 mm | 7 | Sample |
| 3 | Reaction coil: 40 cm/Ø int. 2 mm | 8 | Resample |
| 4 | Extraction unit | 9 | Waste |
| 4.1 | Phase segmentor | | |

Figure 2 — Example of a continuous flow system for the determination of 0,01 mg/l to 1,0 mg/l of phenol index without distillation and with extraction (according to 3.4.2)

3.4.3 Additional apparatus

- a) **Graduated flasks**, 100 ml and 1 000 ml;
- b) **Graduated pipettes**, 1 ml to 10 ml;
- c) **Membrane filter assembly** with membrane filters, pore size 0,45 µm;
- d) **pH measuring device** (e. g. pH electrode).

3.5 Sampling

Use glass or polytetrafluoroethylene (PTFE) containers for sampling.

Prior to use, rinse all containers and devices with which the sample may come into contact, with sulfuric acid of pH approximately 2.

Analyse the samples immediately after their collection. Alternatively, adjust to a pH of approximately 2 with sulfuric acid (3.3.10 or diluted solution) or hydrochloric acid (3.3.11 or diluted solution), store in the dark at a temperature of 2 °C to 5 °C, and analyse within 24 h.

In exceptional cases, after acidification and membrane (pressure) filtration of the sample, a storage of up to two weeks is possible. The applicability of this preservation method shall be checked for the individual case of examination. For more information on sample preservation, see ISO 5667-3 and [10].

Filtration of the sample prior to measurement is necessary, if there is a risk of clogging the transport tubes.

3.6 Procedure

3.6.1 Preparation for measurement

Prior to measurement, continuously run the reagent solutions C (3.3.14), R1 (3.3.15), R2 (3.3.16) and Org (3.3.17) through the flow analysis system, wait for the baseline to stabilize, and zero the baseline.

Consider the system is ready to operate, when the baseline remains stable (no drift). A satisfying signal-to-noise ratio should be obtained.

Verify that a signal-to-noise ratio is obtained that has no significant effect on the results.

The most frequent reasons for a poor signal-to-noise ratio are defective separator membranes or traces of water at the walls of the cell. Traces of water adhering to the walls of the cell can be removed by rinsing the cell with ethanol (3.3.8) or 2-propanol (3.3.9).

Monitor the blank of the reagent and control the membrane function as described in 3.6.3. Carry out the calibration according to 3.6.4.

3.6.2 Checking of the flow system

With the measuring system adjusted to working range II, using a calibration solution (3.3.21) with a concentration of 0,05 mg/l, an absorbance per 1 cm cell length of at least 0,01 cm⁻¹ shall be obtained. Otherwise the flow system is not suitable, and it shall be replaced by a system which fulfils this requirement.

NOTE If the photometric detector (3.4.1, 3.4.2) is not designed for measurement of absorbance values, the absorbance can be determined by an external photometer designed to measure absorbance values.

3.6.3 Checking of the reagent blank

Wait for the baseline to stabilize.

In place of the reagent solutions R1 (3.3.15) and R2 (3.3.16), run water through the system until a stable signal is obtained. Record the change in the absorbance.

If the absorbance (per centimetre cell length) decreases by more than 0,05 cm⁻¹, it can be assumed that self-condensation products have been formed. In this case the preparation of the solutions, the checking of the flow system (see 3.6.2.) and the monitoring of the reagent blank (see 3.6.3) shall be repeated.

Subsequently, transport reagent solutions R1 (3.3.15) and R2 (3.3.16) again.

3.6.4 Calibration

Select working range I or II as appropriate, and prepare the calibration solutions for the working range selected. Carry out a separate calibration for each working range.

Use for working range I and FIA (3.4.1), for example, an injection volume of 200 µl, and for working range II, for example, an injection volume of 600 µl.

For the working ranges I and II and CFA (3.4.2), choose the cell length and the flowrate to obtain the highest possible response for the calibration solution of the highest concentration.

Before starting the calibration, zero the instrument, if need be, in accordance with the manufacturer's instructions.

Calibrate by sequentially applying the calibration solutions (at least five, see 3.3.21) and reagent blanks.

Obtain the measured values corresponding to the calibration solutions applied.

The test conditions for the calibration and the measurement of samples (3.6.5) shall be the same. The magnitude of the signal measured is proportional to the mass concentration of phenol.

Establish the regression line for the series of measured values according to equation (1):

$$y = b \cdot \rho + a \quad (1)$$

where

y is the measured value, in terms of instrument-related units;

b is the slope of the calibration function, in instrument-related units \times l/mg or instrument-related units \times l/ μ g

ρ is the mass concentration of phenol, in milligrams per litre or in micrograms per litre, in the calibration solutions;

a is the ordinate intercept of the reference function, in instrument-related units.

For further approaches see 3.6.5.

3.6.5 Sample measurement

Analyse the samples in the same way as the calibration solutions with the flow analysis system FIA or CFA (3.4.1 or 3.4.2) respectively.

If the mass concentrations to be determined exceed the validity of the selected working range, dilute the sample or analyse using the other working range.

Verify the validity of the calibration function of the respective working range after each sample series, but at least after the measurement of 10 to 20 samples, using one calibration solution each for the lower and upper parts of the respective working range. Establish a new calibration, if necessary.

3.7 Calculation of results

Determine the mass concentration of the determinand in the measuring solution using the measured value obtained as described in 3.6.5, from the calibration function [equation (1), 3.6.4].

For the evaluation, use the appropriate calibration function. Do not extrapolate beyond the working range selected. Calculate ρ using equation (2):

$$\rho = (y - a) / b \quad (2)$$

where ρ is the mass concentration of phenol index, in milligrams per litre or in micrograms per litre, expressed as phenol;

For an explanation of the other symbols, see equation (1).

Take all dilution steps into account.

4 Determination of phenol index (without extraction) after distillation

4.1 Principle

The sample is fed into a continuously flowing carrier stream, mixed with phosphoric acid, and in-line-distilled at pH 1,4. The distillate, containing steam-volatile phenolic compounds, is then mixed with continuously flowing solutions of 4-aminoantipyrine and potassium hexacyanoferrate(III). Phenolic compounds in the distillate are oxidized by hexacyanoferrate(III), and the resulting quinones react with 4-aminoantipyrine forming yellow condensation products, which are measured spectrometrically in a flow spectrometer at 505 nm to 515 nm.

It is absolutely essential that the test described in this International Standard be carried out by suitably qualified staff.

More information on this analytical technique is given in reference [11].

Automatic off-line distillation devices are also applicable.

4.2 Interferences

4.2.1 Chemical interferences

Distillation can be performed at several pH values (pH = 0,5, 1,4, 4). At pH 4 aromatic amines will also distil and form, under the conditions of the reaction condensation products with 4-aminoantipyrine, leading to positive bias. Because steam-volatile phenols are exclusively determined, the distillation is performed at pH 1,4.

More information on interferences is found in references [5, 11, 12].

4.2.2 Physical interferences arising from applying CFA and FIA.

Interferences caused by clogging of the distillation capillary may occur when the salt content of the sample exceeds 10 g/l. In these cases, dilute the sample with water.

With samples containing particulate matter, refer to 3.5 (last paragraph). Turbid or coloured samples and samples preserved by acidification (see 3.6) will not interfere with the determination.

The interlaboratory trial (see annex A) has shown that detergents in waste water can strongly influence the determination, because the foam produced in the flow system can disturb both the steam distillation of volatile phenols (for the determination of phenol index after distillation, see clause 4) and the phase segmentation and phase separation procedures (determination of phenol index after extraction, see clause 3). In general such interferences can easily be discovered in CFA flow systems.

In the case of significant detergent content, this International Standard is only applicable for phenol mass concentrations above 0,1 mg/l.

4.3 Reagents

See also 3.3. The reagent blank value shall regularly be checked (see 4.6.3). In addition to those reagents listed in 3.3, the following reagents are required:

4.3.1 Phosphoric acid, H_3PO_4 , 85 % mass fraction

4.3.2 Potassium hexacyanoferrate(III), $\text{K}_3\text{Fe}(\text{CN})_6$

4.3.3 Potassium chloride, KCl

4.3.4 Surfactant: Polyethylene glycol dodecyl ether, $\text{C}_{16}\text{H}_{30}\text{O}_3$, F 33 °C to 41 °C, solution, 30 % mass fraction

The solution is stable for approximately 4 weeks.

4.3.5 Distillation reagent (symbol Acid in Figures 3 and 4)

Dissolve in a 100 ml graduated flask, while cooling, 10 ml of phosphoric acid (4.3.1) in approximately 80 ml of water (3.3.1) and make up to volume with water.

Degas, e.g. by membrane filtration.

Prepare a fresh solution every day.

4.3.6 Carrier solution (symbol C in Figure 3)

Use water degassed under reduced pressure.

4.3.7 4-Aminoantipyrine solution II (symbol R3 in Figures 3 and 4)

Dissolve in a 100 ml graduated flask 65 mg of 4-aminoantipyrine (3.3.4) in approximately 80 ml of water (3.3.1), add 0,5 ml of surfactant (4.3.4), and make up to volume with water.

Degas the solution e.g. by membrane filtration.

Prepare a fresh solution every day.

4.3.8 Potassium hexacyanoferrate(III) solution (symbol R4 in Figures 3 and 4)

Dissolve in a 100 ml graduated flask 0,2 g of potassium hexacyanoferrate(III) (4.3.2), 0,3 g of boric acid (3.3.7), and 0,5 g of potassium chloride (4.3.3) in approximately 80 ml of water (3.3.1).

Adjust the pH to 10,3 with potassium hydroxide solution (3.3.12).

Make up to volume with water.

Degas the solution, e.g. by membrane filtration.

Prepare fresh solution every day.

4.4 Apparatus

4.4.1 Flow injection analysis (FIA)

The flow injection system shall comprise the following components (see Figure 3):

- a) reagent reservoirs;
- b) low pulsation pump;
- c) specific pump tubing, with flowrates as shown in Figure 3, as an example;
- d) sample injectors with suitable injection volumes;
- e) flow distillation assembly, equipped with a distillation glass capillary, heatable up to 155 °C;

EXAMPLE Length approximately 80 cm, internal diameter approximately 1,5 mm.

- f) transport tubes and reaction coils, internal diameter 0,5 mm to 0,8 mm, tube connections and T-connections of chemically inert plastics, having a minimal dead volume;
- g) spectrometric detector with flow cell, optical pathlength 0,5 cm to 5 cm, wavelength 505 nm to 515 nm;
- h) recording unit (e.g. strip chart recorder, integrator or printer/plotter);

NOTE 1 In general, peak heights are evaluated.

NOTE 2 Figure 3 describes a combined system which contains components of a flow injection and a continuous flow system.

- i) autosampler, if required.

e) flow distillation assembly, equipped with a distillation glass capillary, heatable up to 155 °C;


EXAMPLE Length approximately 80 cm, internal diameter approximately 1,5 mm.

f) spectrometric detector with flow cell, optical pathlength 0,5 cm to 5 cm, wavelength 505 nm to 515 nm;

g) recording unit (e. g. strip chart recorder, integrator or printer/plotter);

NOTE 1 : In general peak heights are evaluated.

NOTE 2 : A CFA system with an internal diameter of 1 mm is described in Figure 4. Other internal diameters (e.g. approximately 2 mm) may also be used (see note 2 in 3.4.2).



Key

R3	4-aminoantipyrine solution II (4.3.7)	6	Reaction coil / 50 cm/Ø int. 1 mm
R4	K ₃ Fe(CN) ₆ solution (4.3.8)	7	Detector: optical pathlength: 0,5 cm to 5 cm, wavelength: 505 nm to 515 nm
1	Pump (flowrates in ml/min)	8	Sample
2	Reaction coil: 10 cm/Ø int. 1 mm	9	Segmentation gas (air)
3	Reaction coil: 60 cm/Ø int. 1 mm	10	Distillation reagent (acid) (4.3.5)
4.1	Heating bath 155 °C	11	Resample
4.2	Distillation unit	12	Waste
5	Reaction coil / 50 cm/Ø int. 1 mm		

Figure 4 — Example of a continuous flow system for determination of 0,01 mg/l to 1,0 mg/l of phenol index with distillation and without liquid-liquid extraction (according to 4.4.2)

4.4.3 Additional apparatus

(See 3.4.3.)

4.5 Sampling

(See 3.5.)

Alternatively, stabilize samples as described in ISO 5667-3.

4.6 Procedure

4.6.1 Preparation for measurement

Heat the distillation bath of the flow analysis apparatus to a temperature of 155 °C, and start feeding water (3.3.1) and distillation reagent (symbol Acid, 4.3.5). As soon as an uniform condensate has formed on the walls of the condenser of the distillation unit, start continuously feeding, instead of water, the reagent solutions 4-aminoantipyrine II (symbol R3, 4.3.7) and potassium hexacyanoferrate(III) (symbol R4, 4.3.8) through the flow analysis system. Wait for the baseline to stabilize, and zero the baseline.

Consider the system ready to operate when the baseline remains stable (no drift). A satisfying signal-to-noise ratio should be obtained.

Verify that a signal-to-noise ratio is obtained that has no significant effect on the results.

The most frequent cause for a poor signal-to-noise ratio is an insufficient formation of condensate, leading to a poor liberation of air bubbles. In this event, check the temperature of the cooling water. The reagent blank is monitored as described in 4.6.3

Perform calibration as described in 4.6.4.

4.6.2 Checking of the flow system

With the measuring system adjusted to working range II, using a calibration solution (3.3.21) with a concentration of 0,05 mg/l, an absorbance per 1 cm cell length of at least 0,01 shall be obtained. Otherwise consider the flow system is not suitable. Replace it by a system which fulfils this requirement.

Consider the note in 3.6.2.

4.6.3 Checking of the reagent blank

Wait for the baseline to stabilize.

In place of reagent solutions 4-aminoantipyrine II (symbol R3, 4.3.7) and potassium hexacyanoferrate(III) (R4, 4.3.8), run water (3.3.1) through the system until a stable response is obtained. Record the change in the absorbance.

If the absorbance (per centimetre cell length) decreases by more than 0,05 cm⁻¹, assume that self-condensation products have been formed.

In this case, the preparation of the solutions, the checking of the flow system (4.6.2), and the monitoring of the reagent blank (4.6.3) shall be repeated.

Subsequently, feed reagent solutions (R3, R4) again.

4.6.4 Calibration

Select the working range I or II, and prepare the calibration solutions (3.3.21) for the respective working range. Carry out a separate calibration for each working range.

For the working ranges I and II with FIA (4.4.1), use an injection volume of 800 µl.