

---

---

**Milk and milk products — Determination  
of nitrate and nitrite contents —**

Part 2:  
**Method using segmented flow analysis  
(Routine method)**

*Lait et produits laitiers — Détermination des teneurs en nitrates et en  
nitrites —*

*Partie 2: Méthode par analyse d'écoulement segmentée (Méthode de  
routine)*



**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. Neither the ISO Central Secretariat nor the IDF accepts any liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies and IDF national committees. In the unlikely event that a problem relating to it is found, please inform the ISO Central Secretariat at the address given below.

STANDARDSISO.COM : Click to view the full PDF of ISO 14673-2:2001

© ISO and IDF 2001

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO or IDF at the respective address below.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.ch](mailto:copyright@iso.ch)  
Web [www.iso.ch](http://www.iso.ch)

International Dairy Federation  
41 Square Vergote • B-1030 Brussels  
Tel. + 32 2 733 98 88  
Fax + 32 2 733 04 13  
E-mail [info@fil-idf.org](mailto:info@fil-idf.org)  
Web [www.fil-idf.org](http://www.fil-idf.org)

Printed in Switzerland

## Contents

Page

Foreword.....	iv
1 Scope .....	1
2 Normative references .....	1
3 Terms and definitions .....	1
4 Principle.....	2
4.1 Nitrate determination.....	2
4.2 Nitrite determination.....	2
5 Reagents.....	2
6 Apparatus .....	5
7 Sampling.....	5
8 Preparation of test sample.....	6
8.1 Cheese .....	6
8.2 Dried milk products and infant food.....	6
8.3 Milk and liquid milk products .....	6
9 Procedure .....	6
9.1 Preparation of the reduction column.....	6
9.2 Checking the reducing capacity of the column.....	7
9.3 Determination of nitrate content .....	7
9.4 Determination of nitrite content .....	8
9.5 Calibration curve.....	9
9.6 Checking the drift of standard solutions .....	9
10 Calculation and expression of results.....	9
10.1 Calculation.....	9
10.2 Expression of results .....	10
11 Precision.....	10
11.1 General.....	10
11.2 Nitrite.....	10
11.3 Nitrate.....	10
12 Test report .....	11
Bibliography .....	14

## Foreword

**ISO (the International Organization for Standardization)** is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this part of ISO 14673|IDF 189 may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

International Standard ISO 14673-2|IDF 189-2 was prepared by Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 5, *Milk and milk products*, and the International Dairy Federation (IDF), in collaboration with AOAC International. It is being published jointly by ISO and IDF and separately by AOAC International.

This first edition of ISO 14673-2|IDF 189-2, together with ISO 14673-1|IDF 189-1 and ISO 14673-3|IDF 189-3, cancels and replaces ISO 4099:1984, ISO 6736:1982, ISO 6739:1988, ISO 6740:1985 and ISO 8195:1987, which have been technically revised.

ISO 14673|IDF 189 consists of the following parts, under the general title *Milk and milk products — Determination of nitrate and nitrite contents*:

- Part 1: Method using cadmium reduction and spectrometry
- Part 2: Method using segmented flow analysis (Routine method)
- Part 3: Method using cadmium reduction and flow injection analysis with in-line dialysis (Routine method)

## Foreword

**IDF (the International Dairy Federation)** is a worldwide federation of the dairy sector with a National Committee in every member country. Every National Committee has the right to be represented on the IDF Standing Committees carrying out the technical work. IDF collaborates with ISO and AOAC International in the development of standard methods of analysis and sampling for milk and milk products.

Draft International Standards adopted by the Action Teams and Standing Committees are circulated to the National Committees for voting. Publication as an International Standard requires approval by at least 50 % of National Committees casting a vote.

International Standard ISO 14673-2|IDF 189-2 was prepared by Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 5, *Milk and milk products*, and the International Dairy Federation (IDF), in collaboration with AOAC International. It is being published jointly by ISO and IDF and separately by AOAC International.

All work was carried out by the Joint ISO/IDF/AOAC Action Team, *Minerals and minor compounds*, of the Standing Committee on *Minor components characterization of physical properties*, under the aegis of its project leader Mr G. Bråthen (NO).

STANDARDSISO.COM : Click to view the full PDF of ISO 14673-2:2001



# Milk and milk products — Determination of nitrate and nitrite contents —

## Part 2: Method using segmented flow analysis (Routine method)

**WARNING** — The use of this International Standard may involve hazardous materials, operations and equipment. This standard does not purport to address all the safety problems associated with its use. It is the responsibility of the user of this standard to establish safety and health practices and determine the applicability of regulatory limitations prior to use.

### 1 Scope

This part of ISO 14673|IDF 189 specifies a routine method for the determination of the nitrate and nitrite contents of milk and milk products by segmented flow analysis. The method is applicable to milk, cheese, and liquid and dried milk products and infant foods.

### 2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this part of ISO 14673|IDF 189. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this part of ISO 14673|IDF 189 are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

ISO 648, *Laboratory glassware — One-mark pipettes*

ISO 835-1, *Laboratory glassware — Graduated pipettes — Part 1: General requirements*

ISO 1042, *Laboratory glassware — One-mark volumetric flasks*

### 3 Terms and definitions

For the purposes of this part of ISO 14673|IDF 189, the following terms and definitions apply.

#### 3.1

##### **nitrate content**

mass fraction of nitrate determined by the procedure specified in this part of ISO 14673|IDF 189

**NOTE** The nitrate content is expressed as the mass in milligrams of nitrate ions ( $\text{NO}_3^-$ ) per kilogram of product.

#### 3.2

##### **nitrite content**

mass fraction of nitrite determined by the procedure specified in this part of ISO 14673|IDF 189

**NOTE** The nitrite content is expressed as the mass in milligrams of nitrite ions ( $\text{NO}_2^-$ ) per kilogram of product.

## 4 Principle

### 4.1 Nitrate determination

A test portion is suspended in water. Part of the suspension is transferred to the analyser for dialysis. The nitrate ions are reduced to nitrite. The nitrite content is determined by a spectrometric method.

Standard nitrate solutions are determined by the same procedure. The nitrate content is calculated by comparing the reading obtained from the test portion with the readings from the standard solutions.

NOTE Any nitrite present is determined as nitrate. The amount of nitrite is generally small compared to the amount of nitrate. Infant food based on soy-proteins can be an exception to the rule. A correction for the nitrite present can be applied after determination of the nitrite content.

### 4.2 Nitrite determination

A test portion is suspended in a solution of ammonium and sodium chloride. Part of the suspension is transferred to the analyser for dialysis. The nitrite content is determined by a spectrometric method.

Standard nitrite solutions are determined by the same procedure. The nitrite content is calculated by comparing the reading obtained from the test portion with that from the standard solution.

## 5 Reagents

Use only reagents of recognized analytical grade unless otherwise specified.

**5.1 Water**, distilled or deionized, or water of equivalent purity, free from nitrate and nitrite ions.

To avoid the possible inclusion of small gas bubbles in the copperized cadmium column (6.9), freshly boil the distilled or deionized water and cool to room temperature. Use the thus-prepared water for the preparation of the column (9.1), to check the reducing capacity of the column (9.2), and for regeneration of the column.

**5.2 Cadmium granules**, of diameter 0,3 mm to 0,8 mm.

Prepare cadmium granules, if not available commercially, as follows.

Place a suitable number of zinc rods in a beaker. Cover the rods with cadmium sulfate solution (5.3). Scrape the cadmium sponge from the rods from time to time over a period of 24 h. Remove the zinc rods and decant the liquid until only sufficient remains to cover the cadmium sponge. Wash the sponge two or three times with water. Transfer the cadmium sponge to a laboratory blender together with 400 ml of the dilute hydrochloric acid (5.6) and blend for a few seconds to obtain granules of the required size. Return the contents of the blender to the beaker and leave to stand for several hours, while stirring occasionally to remove bubbles. Decant most of the liquid and immediately copperize the granules as described in 9.1.2.

**WARNING** — Because of its toxicity, the used cadmium should be delivered as chemical waste to the relevant authorities.

**5.3 Cadmium sulfate solution**,  $c(\text{CdSO}_4 \cdot 8\text{H}_2\text{O}) = 40 \text{ g/l}$ .

Dissolve 40 g of cadmium sulfate in water in a 1 000 ml volumetric flask (6.4). Dilute to the mark with water and mix.

**5.4 Copper(II) sulfate solution**,  $c(\text{CuSO}_4 \cdot 5\text{H}_2\text{O}) = 20 \text{ g/l}$ .

Dissolve 2 g of copper(II) sulfate in water in a 100 ml volumetric flask (6.4). Dilute to the mark with water and mix.

**5.5 Hydrochloric acid (HCl)**, ( $\rho_{20} = 1,19 \text{ g/ml}$ ).

**5.6 Dilute hydrochloric acid**,  $c(\text{HCl}) \approx 1 \text{ mol/l}$ .

Carefully add 80 ml of hydrochloric acid (5.5) to about 700 ml water in a 1 000 ml volumetric flask (6.4) while regularly swirling the contents. Cool the contents to room temperature. Dilute to the mark with water and mix carefully.

**5.7 Disodium ethylenediaminetetraacetate dihydrate (EDTA) solution** ( $\text{Na}_2\text{C}_{10}\text{H}_{14}\text{N}_2\text{O}_8 \cdot 2\text{H}_2\text{O}$ ).

Dissolve 33,5 g of EDTA in water in about 900 ml water in a 1 000 ml volumetric flask (6.4). Dilute to the mark with water and mix.

**5.8 Poly(oxyethylene lauryl ether) solution**, 30 % mass fraction.

**5.9 Ammonium and sodium chloride solution I.**

Dissolve 40 g of ammonium chloride ( $\text{NH}_4\text{Cl}$ ) and 200 g of sodium chloride ( $\text{NaCl}$ ) in about 950 ml water in a 1 000 ml volumetric flask (6.4). Use concentrated ammonia to adjust the pH to 8,5. Dilute to the mark with water. Add 2 ml of poly(oxyethylene lauryl ether) solution (5.8) and mix.

**5.10 Ammonium and sodium chloride solution II.**

Dissolve 10 g of ammonium chloride ( $\text{NH}_4\text{Cl}$ ) and 50 g of sodium chloride ( $\text{NaCl}$ ) in 950 ml water in a 1 000 ml volumetric flask (6.4). Add 20 ml EDTA solution (5.7) and mix. Use concentrated ammonia to adjust the pH to 8,5. Dilute to the mark with water. Add 2 ml of poly(oxyethylene lauryl ether) solution (5.8) and mix.

**5.11 Ammonium and sodium chloride solution III.**

Dissolve 10 g of ammonium chloride ( $\text{NH}_4\text{Cl}$ ) and 50 g of sodium chloride ( $\text{NaCl}$ ) in 950 ml water in a 1 000 ml volumetric flask (6.4). Use concentrated ammonia to adjust the pH to 8,5. Dilute the contents to the 1 000 ml mark with water. Add 2 ml of poly(oxyethylene lauryl ether) solution (5.8) and mix.

**5.12 Phosphoric acid solution** ( $\text{H}_3\text{PO}_4$ ), 85 %, ( $\rho_{20} = 1,71 \text{ g/ml}$ ).

**5.13 Colour reagent.**

Add 100 ml of phosphoric acid solution (5.12) to 800 ml water in a 1 000 ml volumetric flask (6.4). Add in the following order: 10 g sulfanilamide ( $\text{NH}_2\text{C}_6\text{H}_4\text{SO}_2\text{NH}_2$ ) and 0,5 g *N*-1-naphthylethylene diamine dihydrochloride ( $\text{C}_{10}\text{H}_7\text{NHCH}_2\text{CH}_2\text{NH}_2 \cdot 2\text{HCl}$ ) and mix. Dilute to the mark with water. Add 0,5 ml of poly(oxyethylene lauryl ether) solution (5.8) and mix again.

If stored in a refrigerator, the colour reagent solution may be kept for up to 1 month.

**5.14 Sodium nitrate stock solution** ( $\text{NaNO}_3$ ),  $c(\text{NO}_3^-) = 0,400 \text{ g/l}$ .

Dry an amount of sodium nitrate in the oven (6.18) set at 110 °C to 120 °C for 2 h. Dissolve 0,548 4 g of the dry sodium nitrate in water in a 1 000 ml volumetric flask (6.4). Dilute to the mark with water and mix.

**5.15 Sodium nitrate working standard solution**,  $c(\text{NO}_3^-) = 40 \mu\text{g/ml}$ .

Pipette (6.5) 10 ml of the sodium nitrate stock solution (5.14) into a 100 ml volumetric flask (6.4). Dilute to the mark with water and mix.

**5.16 Sodium nitrate standard solutions.**

Prepare a series of standard sodium nitrate solutions with an increasing concentration of nitrate by pipetting into six 200 ml volumetric flasks respectively, 1 ml, 2 ml, 3 ml, 6 ml, 9 ml and 12 ml of the sodium nitrate working standard solution (5.15). Dilute each standard solution to the mark with water and mix. The nitrate ( $\text{NO}_3^-$ ) contents of the obtained sodium nitrate standard solutions are 0,2  $\mu\text{g/ml}$ , 0,4  $\mu\text{g/ml}$ , 0,6  $\mu\text{g/ml}$ , 1,2  $\mu\text{g/ml}$ , 1,8  $\mu\text{g/ml}$  and 2,4  $\mu\text{g/ml}$  respectively.

### 5.17 Sodium nitrite stock solutions (NaNO<sub>2</sub>).

Dry an amount of sodium nitrite in the oven (6.18) set at 110 °C to 120 °C for 2 h.

#### 5.17.1 Sodium nitrite stock solutions I, $c(\text{NO}_2^-) = 0,297 \text{ mg/ml}$ .

Dissolve 0,445 g of the dry sodium nitrite (5.17) in water in a 1 000 ml volumetric flask (6.4). Dilute to the mark with water and mix. Stock solution I may be stored for up to 1 day.

#### 5.17.2 Sodium nitrite, stock solutions II, $c(\text{NO}_2^-) = 1,001 \text{ mg/ml}$ .

Dissolve 1,502 g of sodium nitrite (5.17) in water in a 1 000 ml volumetric flask (6.4). Dilute to the mark with water and mix. If stored in a refrigerator, stock solution II may be kept for up to 1 month.

### 5.18 Sodium nitrite working standard solutions.

#### 5.18.1 Sodium nitrite working standard solution I, $c(\text{NO}_2^-) = 0,89 \mu\text{g/ml}$ .

Transfer with a pipette (6.5) 3 ml of the sodium nitrite stock solution I (5.17.1) to a 1 000 ml volumetric flask (6.4). Dilute to the mark with water and mix.

#### 5.18.2 Sodium nitrite working standard solution II, $c(\text{NO}_2^-) = 0,100 \mu\text{g/ml}$ .

Pipette (6.5) 10 ml of the sodium nitrite stock solution II (5.17.2) into a 1 000 ml volumetric flask (6.4). Dilute to the mark with water and mix. Pipette 10 ml of this solution into another 1 000 ml volumetric flask (6.4). Dilute to the mark with ammonium and sodium chloride solution III (5.11) and mix.

### 5.19 Sodium nitrite standard solutions.

Prepare a series of standard sodium nitrite standard solutions with an increasing concentration of nitrite by pipetting into five 200 ml volumetric flasks (6.4) respectively 10 ml, 20 ml, 40 ml, 100 ml and 200 ml of sodium nitrite working standard solution II (5.18.2). Except for the 200 ml solution, dilute each standard solution to the 200 ml mark with ammonium and sodium chloride solution III (5.11) and mix. The sodium nitrite standard solutions have a nitrite (NO<sub>2</sub><sup>-</sup>) content of 0,005 μg/ml, 0,010 μg/ml, 0,020 μg/ml, 0,050 μg/ml and 0,100 μg/ml, respectively.

### 5.20 Alkaline detergent.

Use one of the solutions mentioned below.

#### 5.20.1 Extran solution, 5 % volume fraction.

Pipette (6.5) 50 ml of extran solution into a 1 000 ml volumetric flask (6.4). Dilute to the mark with water and mix.

#### 5.20.2 Sodium hydroxide solution, $c(\text{NaOH}) = 0,1 \text{ mol/l}$ .

Dissolve 4 g of sodium hydroxide (NaOH) in water in a 1 000 ml volumetric flask (6.4). Dilute to the mark with water and mix.

### 5.21 Regenerating fluid.

Add 50 ml of EDTA solution (5.7) and 2,0 ml of dilute hydrochloric acid (5.6) to 1 litre of water (5.1) and mix.

### 5.22 Reference test sample.

Use as a reference test sample, samples from a milk powder with a known nitrate content.

### 5.23 pH indicator paper.

## 6 Apparatus

Clean all glassware thoroughly and rinse with distilled water to ensure that it is free from nitrate and nitrite ions.

Usual laboratory equipment and, in particular, the following.

- 6.1 **Analytical balance**, capable of weighing to the nearest 1 mg, with a readability of 0,1 mg
- 6.2 **Sample container**, provided with an airtight lid.
- 6.3 **Conical flasks**, of capacity 250 ml and 500 ml.
- 6.4 **Volumetric flasks**, of nominal capacity 100 ml, 500 ml and 1 000 ml respectively, complying with the requirements of ISO 1042, class B.
- 6.5 **Pipettes**, capable of delivering 1 ml, 2 ml, 3 ml, 6 ml, 9 ml, 10 ml, 12 ml, 20 ml, 40 ml, 50 ml and 100 ml respectively, complying with the requirements of ISO 648, class A, or ISO 835-1. Where appropriate, burettes may be used instead of pipettes.
- 6.6 **Graduated cylinders**, of capacities 100 ml and 150 ml.
- 6.7 **Glass beakers**, of capacities 100 ml and 150 ml.
- 6.8 **Glass funnels**, with a stem of about internal diameter 2 mm.
- 6.9 **Suitable tube**, U-shaped, of length about 30 cm and internal diameter 2 mm.
- 6.10 **Analyser**, capable of determining the nitrate and nitrite contents according to the described procedure.
- 6.11 **Suspension apparatus**, appropriate to suspend the test samples.
- 6.12 **Grinding device**, appropriate for grinding the laboratory sample, if necessary.  
To avoid loss of moisture the device should not produce undue heat. A hammer mill shall not be used.
- 6.13 **Laboratory mixer or homogenizer**, with glass containers of capacity 250 ml or 400 ml, suitable for suspending test portions of cheese and whey cheese.
- 6.14 **Glass wool**.
- 6.15 **Spatula**.
- 6.16 **Magnetic stirrer**.
- 6.17 **Water baths**, capable of being maintained at 35 °C to 40 °C and of boiling water.
- 6.18 **Oven**, capable of being maintained at 110 °C to 120 °C.

## 7 Sampling

Sampling is not part of the method specified in this part of ISO 14673 | IDF 189. A recommended sampling method is given in ISO 707.

It is important that the laboratory receive a sample which is truly representative and has not been damaged or changed during transport or storage.

Store the test sample in such a way that deterioration and change in composition are prevented.

## 8 Preparation of test sample

### 8.1 Cheese

8.1.1 Prior to analysis, remove the rind or mouldy surface layer of the test sample in such a way as to obtain a sample representative of the cheese as it is usually consumed.

8.1.2 Grind the test sample by means of an appropriate device (6.12). Mix the ground mass quickly and, if possible, grind a second time and again mix thoroughly. Clean the device after grinding of each sample. If the test sample cannot be ground, mix it thoroughly by intensive stirring and kneading.

8.1.3 As soon as possible after grinding, transfer the test sample to an airtight sample container (6.2) to await the determination which, preferably, should be carried out without delay. If a delay is unavoidable, take all precautions to ensure proper conservation of the test sample and to prevent preservation of moisture on the inside surface of the container.

8.1.4 Do not examine ground cheese showing unwanted mould growth or beginning to deteriorate.

### 8.2 Dried milk products and infant food

Transfer the test sample into a sample container (6.2) of capacity about twice the volume of the sample. Close the container immediately. Mix the test sample thoroughly by repeatedly shaking and inverting the container.

### 8.3 Milk and liquid milk products

In the case of a test sample from non-skimmed milk products, adjust the temperature of the test sample to 35 °C to 40 °C in the water bath (6.19). Mix the test sample thoroughly, but gently, by repeatedly inverting the sample bottle without causing frothing or churning, and cool quickly to approximately 20 °C.

## 9 Procedure

### 9.1 Preparation of the reduction column

9.1.1 Weigh about 10 g cadmium granules (5.2) in a glass beaker (6.7). Add about 25 ml of dilute hydrochloric acid (5.6). Wash the cadmium granules by swirling the glass beaker for several minutes. Decant the hydrochloric acid solution and subsequently wash the cadmium granules with water until the washing is pH neutral. Check the pH with indicator paper (5.23).

9.1.2 Transfer about 50 ml copper(II) sulfate solution (5.4) to the washed cadmium granules (9.1.1) in the glass beaker. Allow the obtained mixture to stand for 5 min while regularly swirling the beaker. Decant the copper sulfate solution and wash the cadmium granules 10 times with water. Keep the cadmium granules immersed in water.

9.1.3 Connect the glass funnels (6.8) with the tube (6.9), folded in a U-shape, and fill almost completely with water. By means of a spatula (6.15) transfer small portions of the cadmium granules through the funnel into the tube. Tap the tube to pack the column.

9.1.4 After the column has been filled almost completely, put glass wool plugs (6.14) of length about 1 cm into both ends of the tube.

9.1.5 Before use, run the sodium nitrate solution with the highest nitrate concentration (5.16) through the column for 1 h.

## 9.2 Checking the reducing capacity of the column

Using the analyser (6.10), measure sequentially the absorbance of two cups with the sodium nitrate standard solution of  $1,2 \mu\text{g NO}_3^-$  per millilitre (5.16), two cups with water, and two cups with a sodium nitrite standard solution of  $0,89 \mu\text{g NO}_2^-$  per millilitre (5.18.1).

The ratio,  $r$ , of the average absorption of the nitrate standard solution to the average absorption of the nitrite standard solution shall be at least 0,95.

If the ratio,  $r$ , is lower than 0,95, prepare a new column.

## 9.3 Determination of nitrate content

### 9.3.1 Preparation of test portion

#### 9.3.1.1 Cheese

Weigh, to the nearest 5 mg, 2,5 g of test sample (8.1) in a glass beaker (6.7). Add 50 ml of water at about 50 °C and suspend the test portion with the aid of the suspension apparatus (6.11). Immediately cool the glass beaker with the test solution. Fill a cup of the sample exchanger with the test solution, in such a way that the separated fat remains in the glass beaker.

#### 9.3.1.2 Dried milk products and infant food

Weigh, to the nearest 5 mg, 2,5 g of test sample (8.2) in a glass beaker (6.7). Add 50 ml water and suspend the test portion with the aid of the suspension apparatus (6.11). Fill a cup of the sample exchanger with the test solution.

#### 9.3.1.3 Liquid milk products

Weigh, to the nearest 10 mg, 10 g of test sample (8.3) in a 50 ml volumetric flask (6.4). Dilute to the mark with water and mix. Fill a cup of the sample exchanger with the test solution.

### 9.3.2 Determination

**9.3.2.1** Start the pump, the spectrometer, the recorder and the data-processing equipment of the analyser (6.10). Run the reagents through the analyser according to Figure 1 for 15 min.

**9.3.2.2** Fill the sample exchanger of the analyser in the following order: start with the six cups filled with the sodium nitrate standard solutions (5.16), placed in order of increasing nitrate content, followed by the cups with the test portion (9.3.1.1, 9.3.1.2 or 9.3.1.3).

**9.3.2.3** Start the sample exchanger of the analyser. Activate the data-processing equipment according to the directions for use. When the determination is finished, remove the reduction column (9.1) from the system and flush the system without column with an alkaline detergent (5.20) for at least 15 min and subsequently with water for at least 15 min.

**9.3.2.4** Then flush the reduction column with the regenerating fluid (5.21) with a flow rate of about 2 ml/min for 15 min and subsequently again with water at the same flow rate for 10 min. Then close the column carefully.

### 9.3.3 Dilutions

If the nitrate content of the test solution is higher than the nitrate content of the highest sodium nitrate standard solution, repeat the analysis in duplicate by again preparing two new test solutions according to 9.3.1.

Dilute the new test solutions in such a way that the expected results given by the recorder will be between the result of the lowest-but-one sodium nitrate standard solution and that of the highest standard solution. Then transfer the thus diluted test solution to a cup.

#### **9.3.4 Reference sample**

Examine the reference test sample (5.22) at least in duplicate, according to 9.3.2.1 to 9.3.2.4.

### **9.4 Determination of nitrite content**

#### **9.4.1 Preparation of test portion**

##### **9.4.1.1 Cheese**

Weigh, to the nearest 5 mg, 2,5 g of the test sample (8.1) in a glass beaker (6.7). Add 50 ml of ammonium and sodium chloride solution III (5.11), at about 50 °C, and suspend the test portion with the aid of the suspension apparatus (6.11). Immediately cool the glass beaker with the test solution. Fill a cup of the sample exchanger with the test solution, in such a way that the separated fat remains in the glass beaker.

##### **9.4.1.2 Dried milk products and infant foods**

Weigh, to the nearest 5 mg, 2,5 g of the test sample (8.2) in a glass beaker (6.7). Add 50 ml of ammonium and sodium chloride solution III (5.11) and suspend the test portion with the aid of the suspension apparatus (6.11). Fill a cup of the sample exchanger with the test solution.

##### **9.4.1.3 Milk and liquid milk products**

Weigh, to the nearest 10 mg, 10 g of the test sample (8.3) in a 50 ml volumetric flask (6.4). Make up to the mark with ammonium and sodium chloride solution III (5.11) and mix. Fill a cup of the sample exchanger with the test solution.

#### **9.4.2 Determination**

**9.4.2.1** Start the pump, the spectrometer, the recorder and the data-processing equipment of the analyser (6.10). Run the reagents through the analyser according to Figure 2 for 15 min.

**9.4.2.2** Fill the sample exchanger of the analyser in the following order: start with the five tubes filled with sodium nitrite standard solutions (5.19), placed in order of increasing nitrite content, followed by the tubes with the test solution (9.4.1.1, 9.4.1.2 or 9.4.1.3).

**9.4.2.3** Start the sample exchanger of the analyser. Activate the data-processing equipment according to the directions for use. When the determination is finished, flush the system with an alkaline detergent (5.20) for at least 15 min and subsequently with water for at least 15 min.

#### **9.4.3 Dilutions**

If the nitrite content of the test solution is higher than the nitrite content of the highest sodium nitrite standard solution, repeat the analysis in duplicate by again preparing two new test solutions according to 9.3.1.

Dilute the new test solutions in such a way that the expected results given by the recorder will be between the result of the lowest-but-one sodium nitrite standard solution and that of the highest standard solution. Then transfer the thus diluted test solutions to a cup.

#### **9.4.4 Reference test sample**

Examine the reference test sample (5.22) at least in duplicate, according to 9.4.2.1 to 9.4.2.3.

## 9.5 Calibration curve

Calculate the most suitable calibration curve and the correlation coefficient for both the sodium nitrate and the sodium nitrite standard solutions. The obtained correlation coefficient shall be greater than 0,998 5.

## 9.6 Checking the drift of standard solutions

Regularly check the drift using the appropriate standard solution mentioned in 9.2. The relative difference between two successive standards shall not exceed 10 %. The drift path between the highest and lowest results obtained on one day shall not exceed 20 %. If these values are exceeded, check the cause and repeat the examinations.

## 10 Calculation and expression of results

### 10.1 Calculation

Convert the results given by the recorder for the test solution to micrograms of nitrate or nitrite ions per millilitre, read from the calibration curve of the standard solutions.

Calculate the nitrate or nitrite contents of the test sample by using one of the following equations.

a) For dried milk products:

$$\omega_n = \omega_t \times d \times f_d$$

b) For liquid milk products:

$$\omega_n = \omega_t \times 5 \times f_l$$

c) For cheese:

$$\omega_n = \omega_t \times 20,5 \times f_c$$

where

$\omega_n$  is the numerical value of the nitrate or nitrite content of the test sample, in milligrams per kilogram;

$\omega_t$  is the numerical value of the nitrate or nitrite content of the test solution read from the calibration curve, in micrograms per millilitre;

$d$  is the numerical value of the dilution factor;

$f_d$  is the multiplying factor for dried milk and infant food obtained by the following formula ( $f_d = 20,6$ ):

$$f_d = \frac{50 \times 1,03 \times 400}{1000}$$

$f_l$  is the multiplying factor for liquid milk products obtained by the following formula ( $f_l = 5$ ):

$$f_l = \frac{50 \times 100}{1000}$$

$f_c$  is the multiplying factor for cheese obtained by the following formula ( $f_c = 20,5$ ):

$$f_c = \frac{50 \times 1,025 \times 400}{1000}$$

where

- 50 is the conversion value to change the expression from milligrams per millilitre to milligrams per 50 ml;
- 1,03 is the ratio of the volume of the test solution (9.3.1.2 or 9.4.1.2) (2,5 g + 50 ml) and the volume of the standard solutions (50 ml);
- 1,025 is the correction value for the moisture content of cheese (average moisture content: 50 %);
- 400 is the conversion value to change the expression from 2,5 g to 1 kg;
- 1 000 is the conversion value to change the expression from micrograms to milligrams.

## 10.2 Expression of results

Express the results for nitrite content to 1 decimal place and for nitrate content to whole figures.

## 11 Precision

### 11.1 General

The values for repeatability and reproducibility limits are expressed for the 95 % probability level and may not be applicable to concentration ranges and matrices other than those given.

### 11.2 Nitrite content

#### 11.2.1 Repeatability

The absolute difference between two independent single test results, obtained using the same method on identical test material in the same laboratory by the same operator using the same equipment within a short interval of time, will in not more than 5 % of cases be greater than 0,5 mg/kg of product.

#### 11.2.2 Reproducibility

The absolute difference between two single test results, obtained using the same method on identical test material in different laboratories with different operators using different equipment, will in not more than 5 % of cases be greater than 1,0 mg/kg of product.

### 11.3 Nitrate content

#### 11.3.1 Repeatability

The absolute difference between two independent single test results, obtained using the same method on identical test material in the same laboratory by the same operator using the same equipment within a short interval of time, will in not more than 5 % of cases be greater than:

- for samples with a nitrate content < 100 mg/kg: 10 mg/kg;
- for samples with a nitrate content  $\geq$  100 mg/kg: 15 % of the arithmetic mean of the results.

### 11.3.2 Reproducibility

The absolute difference between two single test results, obtained using the same method on identical test material in different laboratories with different operators using different equipment, will in not more than 5 % of cases be greater than:

- for samples with a nitrate content < 100 mg/kg: 15 mg/kg;
- for samples with a nitrate content  $\geq$  100 mg/kg: 20 % of the arithmetic mean of the results.

## 12 Test report

The test report shall specify:

- all information required for the complete identification of the sample;
- the sampling method used, if known;
- the test method used, with reference to this part of ISO 14673 | IDF 189;
- all operating details not specified in this part of ISO 14673 | IDF 189, or regarded as optional, together with details of any incident which may have influenced the result(s);
- the test result(s) obtained or, if the repeatability has been checked, the final quoted results obtained.

STANDARDSISO.COM : Click to view the full PDF of ISO 14673-2:2001