



**International
Standard**

ISO 15883-3

Washer-disinfectors —

**Part 3:
Requirements and tests for washer-
disinfectors employing thermal
disinfection for human waste
containers**

Laveurs désinfecteurs —

*Partie 3: Exigences et essais pour laveurs désinfecteurs destinés à
la désinfection thermique de récipients pour déjections humaines*

**Second edition
2024-11**

STANDARDSISO.COM : Click to view the full PDF of ISO 15883-3:2024



COPYRIGHT PROTECTED DOCUMENT

© ISO 2024

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Performance requirements	2
4.1 General.....	2
4.2 Chemical dosing systems.....	2
4.3 Emptying.....	3
4.4 Cleaning.....	3
4.5 Disinfecting.....	4
4.6 Drying.....	4
4.7 Cooling.....	4
5 Mechanical and control requirements	4
5.1 Instrumentation and control.....	4
5.2 Process.....	4
5.3 Venting, draining, and purging.....	5
5.4 Water quality.....	5
6 Testing for conformity	5
6.1 General.....	5
6.2 Test for drain seal integrity.....	6
6.3 Test for flushing of non-absorbent material.....	6
6.4 Test for flushing of absorbent material (toilet tissue).....	6
6.5 Test for loading/emptying of human waste containers.....	6
6.5.1 Automatic emptying.....	6
6.5.2 Manual emptying.....	6
6.6 Tests for soil removal from load, load carrier(s) and chamber walls.....	6
7 Information to be provided	7
8 Information to be requested from the purchaser by the supplier of the WD	7
Annex A (informative) Summary of test programmes	8
Annex B (normative) Specification for toilet tissue for use in tests on washer-disinfectors (WD) for human waste containers	9
Annex C (informative) Measurement of water absorbency of toilet tissue by the basket method	10
Bibliography	11

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 198, *Sterilization of health care products*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 102, *Sterilizers and associated equipment for processing of medical devices*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 15883-3:2006), which has been technically revised.

The main changes are as follows:

- revision of cross-references to clauses in ISO 15883-1:2024 and ISO 15883-5:2021;
- alignment of terms and definitions with ISO 11139:2018+Amd 1:2024;
- addition of [4.1.2](#) listing exemptions from the requirements of ISO 15883-5:2024;
- alignment of text in clauses on cleaning and disinfecting with revised clauses of ISO 15883-2;
- addition of rinsing from former 4.6 to [4.4](#) and [4.5](#);
- addition of [4.7](#);
- revision of references in the Bibliography.

A list of all parts in the ISO 15883 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

This document is the third part in the ISO 15883 series of standards specifying the performance of washer-disinfectors (WD) and the general requirements for performance applicable to WD for human waste containers. The requirements given in this document apply to WD used for emptying, flushing, washing, and thermally disinfecting human waste containers intended for reuse such as:

- portable sanitary pans;
- supports for single-use bedpans;
- hospital bowls for containment of human waste;
- urine collection containers;
- suction bottles;
- products similar to the above and used for similar purposes.

Fields of application within the scope of the ISO 15883 series of standards can include laboratory, veterinary, dental and pharmaceutical applications, and other specific applications, such as WD for bedsteads and transport carts and the disinfection of crockery and cutlery intended for use with immunologically compromised patients.

Requirements for WD for other applications are specified in other parts of the ISO 15883 series of standards.

In order to reduce the risk of spillage and the generation of aerosols, most machines incorporate means to empty human waste containers automatically, e.g. by the action of closing the WD door.

Where equipment does not provide automatic emptying facilities, extra care is needed by the user to avoid exposure to human waste and contamination of the work environment including the generation of aerosols.

The reliability of a human waste container WD can be adversely affected if the machine is connected to a poorly designed or constructed drainage system. The purchaser is therefore recommended to ensure that the drainage system complies with the manufacturer's recommendations in all respects.

Safety requirements for WD are given in IEC 61010-2-040.

NOTE Local or national regulations can apply in respect of the potential adverse effects on the quality of water intended for human consumption or environmental impacts caused by the WD and its intended use.

[STANDARDSISO.COM](https://standardsiso.com) : Click to view the full PDF of ISO 15883-3:2024

Washer-disinfectors —

Part 3:

Requirements and tests for washer-disinfectors employing thermal disinfection for human waste containers

1 Scope

This document specifies requirements for washer-disinfectors (WD) that are intended to be used for emptying, flushing, washing and thermal disinfection of non-critical devices in the form of human waste containers by one operating cycle.

This document is intended to be used in conjunction with the general requirements specified in ISO 15883-1:2024, except for those specified in [4.1.1](#), and with the requirements of ISO 15883-5:2021, except for those specified in [4.1.2](#).

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 15883-1:2024, *Washer-disinfectors — Part 1: General requirements, terms and definitions and tests*

ISO 15883-5:2021, *Washer-disinfectors — Part 5: Performance requirements and test method criteria for demonstrating cleaning efficacy*

ISO 17664-2:2021, *Processing of health care products — Information to be provided by the medical device manufacturer for the processing of medical devices — Part 2: Non-critical medical devices*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 15883-1 and the following apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1

non-critical device

<washer-disinfector> item processed in a washer-disinfector, whose surface(s) are intended to contact intact skin of a body but do not penetrate it, or device not intended for direct patient contact

EXAMPLE Blood pressure cuffs, wheelchairs, trays, bowls, dishes, glassware, receivers, containers for transit.

Note 1 to entry: National regulations can use alternative wording for the definition of this term when applied to medical devices.

[SOURCE: ISO 11139:2018/Amd 1:2024, 3.357]

4 Performance requirements

4.1 General

4.1.1 The requirements of ISO 15883-1:2024 apply with the exception of:

- ISO 15883-1:2024, 4.1.13 (which refers to minimum pressure inside the pipework during the washing stage);
- ISO 15883-1:2024, 4.3.2 (which refers to chemical disinfection);
- ISO 15883-1:2024, 5.7.4 (which refers to verification of the dose of process chemical admitted);
- ISO 15883-1:2024, 5.7.5 (which specifies the accuracy and reproducibility of chemical dosing systems);
- ISO 15883-1:2024, 5.8 (which refers to load temperature protection);
- ISO 15883-1:2024, 5.9 (which refers to control of temperatures on the load and chamber walls, see [4.5.3](#) and [4.5.4](#) of this document);
- ISO 15883-1:2024, 5.21 b) (which refers to microprocessor control systems);
- ISO 15883-1:2024, 6.8.5 (which refers to tests for load temperature protection);
- ISO 15883-1:2024, 6.10.3 (which refers to protein residue tests).

4.1.2 The requirements of ISO 15883-5:2021 apply with the exception of:

- ISO 15883-5:2021, 4.1.4 (requirement that verification and documentation of the full cleaning stage does not interfere with analyte detection is not applicable);
- ISO 15883-5:2021, 4.4.3 (which refers to assay criteria);
- ISO 15883-5:2021, 5.1.1 (which refers to validation of cleaning test methods);
- ISO 15883-5:2021, 5.1.3 (which refers to validation of detection methods);
- ISO 15883-5:2021, 5.1.4 (which refers to analyte assay methods);
- ISO 15883-5:2021, 5.3.4 and 5.4.4 (requirements for action levels specified in 4.4.3 for the load, and conformance of cleaning efficacy for the test load with the alert levels in 4.4.3, are not applicable).

4.1.3 The WD shall be designed to process one or more human waste containers in an operating cycle.

NOTE This can require the use of a variety of types of load carrier(s).

4.1.4 The temperature attained on the surfaces of the load during the disinfection stage shall not be less than the lower limit of the disinfection temperature profile.

4.2 Chemical dosing systems

4.2.1 Provision shall be made for the installation of specific chemical dosing system(s), when specified by the purchaser, to allow for the dosing of process chemical(s).

4.2.2 The volume of process chemical(s) admitted shall be adjustable by a means that shall deliver the set volume to an accuracy of $\pm 10\%$ or better.

4.2.3 The WD shall either be fitted with means to ensure that a fault is indicated when insufficient process chemical(s) has/have been admitted, or it shall be possible for the operator to visually verify that the required amount of process chemical(s) has/have been used.

4.3 Emptying

4.3.1 The manufacturer shall require the purchaser to specify whether the human waste containers are required to be emptied manually or automatically.

Manual emptying of human waste containers should be avoided whenever possible.

4.3.2 When the human waste container(s) are to be emptied automatically, the emptying system shall ensure that there is no spillage of the human waste container contents or discharge of aerosols of the contents of human waste containers during automatic emptying.

Check for conformance in accordance with [6.5.1](#).

4.3.3 When the human waste container(s) are to be emptied manually into the WD, the door aperture and load support system shall be designed to enable the human waste container to be emptied and then located in the load carrier without spillage or splashing.

Check for conformance in accordance with [6.5.2](#).

4.4 Cleaning

4.4.1 Cleaning shall be tested in accordance with the requirements of ISO 15883-1:2024, 4.2 and 6.10 and the performance requirements and test method criteria specified in ISO 15883-5:2021, Clauses 4 and 5, and Annex B.

NOTE 1 Some suitable test soils for human waste containers are described in Reference [7] and Reference [8].

The cleaning process shall also meet the requirements of the test specified in [6.6](#).

Where applicable, any treatment required prior to cleaning in the WD shall be performed in accordance with the instructions for use for the load specified in ISO 17664-2:2021, 6.4, or for the WD specified in ISO 15883-1:2024, 8.1 a).

NOTE 2 ISO 17664-1 can be applicable if the human waste container is intended to contact broken skin.

4.4.2 Pre-wash (flushing)

The human waste containers shall be flushed with sufficient water to remove the gross soiling.

4.4.3 Washing

The human waste containers shall be washed on both their inner and outer surfaces.

NOTE 1 The water used to wash the human waste containers can be discharged without recirculation or be recirculated during a single washing stage within one operating cycle.

During the washing stage:

- the washing time shall start when the temperature at the control sensor of the WD reaches the lower limit of the temperature profile;
- the temperatures recorded on the surface of the load and load carrier(s) for the washing stage follow the temperature profile defined for this stage and are within ± 5 °C of the relevant set temperature for each holding time of the stage (see ISO 15883-1:2024, 4.2.3).

NOTE 2 A washing stage can include two or more washing temperatures and washing temperature bands.

4.5 Disinfecting

4.5.1 When tested in accordance with the method specified in ISO 15883-1:2024, 6.8.2, 6.8.3 and 6.8.4, each operating cycle shall include a thermal disinfection stage during which the specified minimum temperature for the specified minimum (holding) time, or the equivalent lethality of an A_0 of at least 60, is achieved on all surfaces which are required to be disinfected.

NOTE Thermal disinfection can be achieved by rinsing the load with hot water, exposure to steam or combination of the two.

4.5.2 Higher A_0 values than 60 may be specified for the inner surface of the human waste container, the outer surfaces of the human waste container or the walls of the WD chamber.

4.5.3 The temperature on the surface of the load shall be within 0 °C to 15 °C of the disinfection temperature throughout the time specified for disinfection when this has been specified as a time-temperature relationship.

4.5.4 The temperature recorded on the surface of the chamber wall shall be within 0 °C to 15 °C of the set temperature throughout the time specified for disinfection when this has been specified as a time-temperature relationship.

4.5.5 If rinsing between cleaning and disinfection is omitted, then the requirements of ISO 15883-1:2024, 4.2.4 apply.

4.5.6 If there is a final rinsing after the disinfection stage, then ISO 15883-1:2024, 4.4 applies.

4.6 Drying

The provision of a separate drying stage within the operating cycle is optional.

Hot air or compressed air used for drying shall be of a quality which shall not impair the cleanliness of, nor introduce microbial contamination to, the load.

NOTE ISO 8573-1, ISO 29463-1 and EN 1822-1 provide guidance and specifications on air quality.

4.7 Cooling

The provision of a separate cooling stage within the operating cycle is optional. Means used to cool the load shall not impair the cleanliness of, nor introduce microbial contamination to, the load.

5 Mechanical and control requirements

5.1 Instrumentation and control

5.1.1 The WD shall either be fitted with a display showing chamber temperature [see ISO 15883-1:2024, 5.11.4 a)] or an indicating light to show attainment of a pre-set disinfection temperature. The sensor shall be located as specified in ISO 15883-1:2024, 5.12.6.

5.1.2 Provision shall be made for the installation of a temperature recorder when specified by the purchaser. When a recorder is fitted, this shall be deemed to meet the requirements of [5.1.1](#).

5.2 Process

5.2.1 The inner surfaces of the chamber shall be cleaned and disinfected during the process.

5.2.2 Means shall be provided to pre-set the washing temperature and time over a specified range. Adjustment shall be by means of an access device.

5.2.3 Means shall be provided to pre-set the disinfection temperature and time over a specified range. Adjustment shall be by means of an access device.

NOTE The facility to adjust the pre-set time and temperature over the ranges specified is to allow for disinfection for short time periods at high temperature or, to obtain equivalent effect, prolonged times at lower temperatures; this can be required for processing human waste containers which will not withstand higher temperatures.

The combination of disinfection time and disinfection temperature shall be set to achieve the specified A_0 value.

5.2.4 If process fluid is recirculated, means shall be provided to prevent the spread of contaminants and microorganisms to a subsequent process step.

NOTE The water used to wash the human waste containers can be discharged without recirculation or be recirculated during a single flushing stage within one operating cycle.

5.3 Venting, draining, and purging

5.3.1 For free standing and single door WD the drainage outlet shall be placed towards the rear of the WD.

This should be done in such a manner that it is possible to make connection with a drainage outlet in any position within a horizontal angle of 180°.

5.3.2 When tested in accordance with [6.2](#), the water seal in the trap between the chamber and the drain shall not be broken during five (5) consecutive operating cycles.

NOTE When it is not practicable to vent to the outside atmosphere, a condenser can be used for indirect venting. This can produce a backpressure which, if excessive, will break the water seal between the chamber and the drain. The test is intended to verify that this does not occur.

5.3.3 After the completion of the washing stage of the operating cycle, the trap shall be clear of soil and waste and filled with fresh water.

Conformance shall be tested in accordance with [6.3](#) and [6.4](#).

After each of five (5) tests carried out in accordance with [6.3](#), there shall be no test spheres retained within the chamber or trap.

After each of five (5) tests carried out in accordance with [6.4](#), there shall be no toilet tissues retained within the chamber and the trap shall be clear of soil and waste and filled with fresh water.

5.4 Water quality

The tests required by ISO 15883-1:2024, 5.23 and 6.4.2.2 shall be limited to determination of hardness [expressed as millimole per litre (mmol/l) of CaCO_3] when the water is from a potable supply.

6 Testing for conformity

6.1 General

Testing for conformity shall be carried out in accordance with ISO 15883-1:2024, Clause 6.

NOTE [Annex A](#) includes a summary of test programmes for WD for human waste containers in addition to those recommended in ISO 15883-1:2024, Annex A.

6.2 Test for drain seal integrity

Fit a transparent test trap of the same type and dimensions as that normally fitted to the WD in place of the normal trap and connect it to a suitable outlet.

Pour sufficient water into the chamber of the WD to charge the trap and ensure that there are no leaks.

Carry out five (5) consecutive operating cycles with a full load of human waste containers without opening the door between consecutive operating cycles. On completion of the five (5) cycles, observe the trap and check that the water seal in the trap has not been broken.

6.3 Test for flushing of non-absorbent material

One non-absorbent test sphere shall be used for each human waste container that the WD is intended to process in a single operating cycle. Drop the non-absorbent test spheres [relative density of 1 075 to 1 080, diameter (43 ± 0,5) mm] into the otherwise empty chamber, close the door and start the operating cycle. At the end of the cleaning stage and before the disinfection stage, stop the machine and examine the chamber and the trap for retained spheres.

Perform the test for flushing of non-absorbent materials a total of five (5) times.

6.4 Test for flushing of absorbent material (toilet tissue)

Twelve (12) sheets of toilet tissue in accordance with [Annex B](#) shall be used for each human waste container that the WD is intended to process in a single operating cycle. Crumple the required number of sheets of toilet tissue and drop into the chamber of the WD containing the maximum intended load of human waste containers. Close the door and start the cycle. At the end of the cleaning (washing) stage, interrupt the operating cycle and record the number of tissue sheets remaining in the chamber and load; inspect the trap and record the number of tissues retained in the trap.

Perform the test for flushing of absorbent material a total of five (5) times.

NOTE See [Annex C](#) for details of measurement of water absorbency of toilet tissue by the basket method.

6.5 Test for loading/emptying of human waste containers

6.5.1 Automatic emptying

For WD conforming with [4.3.2](#), each type of human waste container which the WD is designed to process shall be tested. Fill each human waste container with water to no less than 75 % of its brim-full capacity and locate it in the load carrier in accordance with the instructions for use. Load the chamber to the maximum recommended capacity. Close the door. Observe whether any liquid has been spilled or splashed outside the WD.

6.5.2 Manual emptying

For WD conforming with [4.3.3](#), each type of human waste container which the WD is designed to process shall be tested. Fill each human waste container with water to no less than 75 % of its brim-full capacity and empty it into the WD and locate it in the load carrier in accordance with the instructions for use. Load the chamber to the maximum recommended capacity. Close the door. Observe whether any liquid has been spilled or splashed outside the WD.

6.6 Tests for soil removal from load, load carrier(s) and chamber walls

The tests shall be carried out in accordance with ISO 15883-1:2024, 6.10 and methods specified in ISO 15883-5:2021. Refer to ISO 15883-5:2021, Table A.1 and Bibliography for references.

NOTE 1 Local regulations can require the use of particular test soils and methods.

NOTE 2 The user's choice of test soil(s) and method(s) for performance testing can indicate a need to carry out similar testing before the WD is supplied.

ISO 15883-3:2024(en)

The test soils used for the load, load carrier(s) and chamber wall can be different. Where different test soils are used, the rationale for the choice of test soil shall be documented.

7 Information to be provided

In addition to the information listed in ISO 15883-1:2024, Clause 8, the purchaser shall be provided with the following information:

- a) the maximum load or capacity and acceptable loading combinations;
- b) the range of load supports available and required;
- c) the quantity of water required per operating cycle;
- d) any details of ventilation discharge requirements e.g. air volume, system pressure drop and maximum temperature;
- e) the energy required per operating cycle;
- f) the following information obtained by testing in accordance with ISO 15883-1:2024, 6.13:
 - 1) the time for an operating cycle from a cold start;
 - 2) the time for each of five (5) consecutive operating cycles, all having a hot start;
 - 3) the location and temperatures of the coolest and hottest parts of the load and the chamber wall during thermal disinfection;
- g) the disinfection temperature band.

8 Information to be requested from the purchaser by the supplier of the WD

In addition to the information listed in ISO 15883-1:2024, Clause 10, the supplier of the WD shall request the following information from the purchaser:

- a) the number, type, and capacity of human waste containers to be processed;
- b) whether human waste containers are required to be emptied automatically;
- c) whether additional process chemical dosing system(s) is/are required;
- d) the duct work connections available;
- e) the A_0 value, or the combination of time and temperature, to be attained for thermal disinfection (if the A_0 value has not been defined by the purchaser, see [4.5.1](#)).

Annex A
(informative)

Summary of test programmes

[Table A.1](#) summarizes the recommended test programmes applicable to WD for human waste containers in addition to those recommended in ISO 15883-1:2024. Other tests or schedules of tests providing equivalent assurance are equally acceptable.

Table A.1 — Summary of test programmes for WD for human waste containers

Brief description of test	Requirements subclause	Test subclause	Type test	Works test	Operational qualification	Performance qualification	Routine test
Drain seal integrity	5.3.2	6.2	X	O	O	B	B
Flushing of non-absorbent materials	5.3.3	6.3	X	B	B	B	B
Flushing of absorbent materials	5.3.3	6.4	X	B	X	B	B
Safety of loading and emptying human waste containers	4.3.2 or 4.3.3	6.5.1 or 6.5.2	X	B	B	X	B
Key X = recommended B = not recommended O = optional test which may be requested by the purchaser or user. Optional tests may be carried out at the discretion of the user or can be required by local regulation.							