

---

---

**Workplace air quality — Sampling and  
analysis of volatile organic compounds  
by solvent desorption/gas  
chromatography —**

**Part 2:  
Diffusive sampling method**

*Qualité de l'air des lieux de travail — Échantillonnage et analyse des  
composés organiques volatils par désorption au solvant/chromatographie  
en phase gazeuse —*

*Partie 2: Méthode d'échantillonnage par diffusion*



**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

STANDARDSISO.COM : Click to view the full PDF of ISO 16200-2:2000

© ISO 2000

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 734 10 79  
E-mail [copyright@iso.ch](mailto:copyright@iso.ch)  
Web [www.iso.ch](http://www.iso.ch)

Printed in Switzerland

## Contents

	Page
1	<b>Scope</b> ..... 1
2	<b>Normative references</b> ..... 1
3	<b>Principle</b> ..... 1
4	<b>Reagents and materials</b> ..... 2
5	<b>Apparatus</b> ..... 4
6	<b>Sampling</b> ..... 5
7	<b>Procedure</b> ..... 6
7.1	<b>Desorption</b> ..... 6
7.2	<b>Analysis</b> ..... 6
7.3	<b>Determination of desorption efficiency</b> ..... 6
7.4	<b>Calibration of uptake rate</b> ..... 7
8	<b>Calculations</b> ..... 8
8.1	<b>General</b> ..... 8
8.2	<b>Mass concentration of analyte</b> ..... 8
8.3	<b>Volume concentration of analyte</b> ..... 9
8.4	<b>Uptake rates</b> ..... 9
9	<b>Interferences</b> ..... 9
10	<b>Precision and bias</b> ..... 9
11	<b>Storage and transport</b> ..... 10
12	<b>Test report</b> ..... 10
13	<b>Quality control</b> ..... 10
	<b>Annex A (informative) Description of sorbent types</b> ..... 11
	<b>Annex B (informative) Diffusive sampling rates (cm<sup>3</sup>/min)</b> ..... 12
	<b>Annex C (informative) Equivalence of gas chromatographic stationary phases</b> ..... 20
	<b>Annex D (informative) Suppliers of charcoal-based organic vapour diffusive samplers</b> ..... 21
	<b>Annex E (informative) Specific information on sampler type A</b> ..... 22
	<b>Annex F (informative) Specific information on sampler type B</b> ..... 23
	<b>Annex G (informative) Specific information on sampler type C</b> ..... 24
	<b>Annex H (informative) Specific information on sampler type D</b> ..... 25
	<b>Annex I (informative) Specific information on sampler type E</b> ..... 26
	<b>Annex J (informative) Retention indices of selected VOCS on BP-1 and BP-10 phases</b> ..... 27
	<b>Bibliography</b> ..... 31

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this part of ISO 16200 may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

International Standard ISO 16200-2 was prepared by Technical Committee ISO/TC 146, *Air quality*, Subcommittee SC 2, *Workplace atmospheres*.

ISO 16200 consists of the following parts, under the general title *Workplace air quality — Sampling and analysis of volatile organic compounds by solvent desorption/gas chromatography*:

- *Part 1: Pumped sampling method*
- *Part 2: Diffusive sampling method*

Annexes A to J of this part of ISO 16200 are for information only.

# Workplace air quality — Sampling and analysis of volatile organic compounds by solvent desorption/gas chromatography —

## Part 2: Diffusive sampling method

### 1 Scope

This part of ISO 16200 gives general guidance for the sampling and analysis of volatile organic compounds (VOCs) in air.

This part of ISO 16200 is applicable to a wide range of VOCs, including hydrocarbons, halogenated hydrocarbons, esters, glycol ethers, ketones and alcohols. A number of devices and sorbents are recommended for the sampling of these VOCs, each sorbent having a different range of applicability.

**NOTE** Activated coconut shell charcoal is frequently used. Very polar compounds may require derivatization; very low boiling compounds will only be partially retained by the sorbents and can only be estimated qualitatively. Semi-volatile compounds will be fully retained by the sorbents, but may only be partially recovered.

This part of ISO 16200 is valid for the measurement of airborne vapours of VOCs in a concentration range of approximately  $1 \text{ mg/m}^3$  to  $1000 \text{ mg/m}^3$  individual organic for an exposure time of 8 h.

The upper limit of the useful range is set by the sorptive capacity of the sorbent used and, subject to dilution of the analysed solution, by the linear dynamic range of the gas chromatograph column and detector or by the sample splitting capability of the analytical instrumentation used. The lower limit of the useful range depends on the noise level of the detector and on blank levels of analyte and/or interfering artefacts on the sampling devices or in the desorption solvent. Artefacts are typically sub-nanogram for activated charcoal, but higher levels of aromatic hydrocarbons have been noted in some batches.

### 2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this part of ISO 16200. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this part of ISO 16200 are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

EN 838:1995, *Workplace atmospheres — Diffusive samplers for the determination of gases and vapours — Requirements and test methods*.

EN 1540, *Workplace atmospheres — Terminology*.

### 3 Principle

Diffusive samplers consist of a sorbent separated from ambient air by some form of diffusion resistance, commonly a controlled air gap and draught shield. The diffusive sampler (or samplers) is exposed to air for a measured time

period. The rate of sampling is determined by prior calibration in a standard atmosphere (see 7.4). Volatile organic compounds migrate into the sampler by diffusion and are collected on the sorbent, normally activated carbon. The collected vapour is desorbed by a solvent, typically carbon disulfide, and the solution is analysed with a gas chromatograph equipped with a flame ionization detector, mass spectrometer or other selective detector.

## 4 Reagents and materials

During the analysis, use only reagents of recognized analytical reagent grade.

### 4.1 Volatile organic compounds

A wide range of VOCs are required as reagents for calibration purposes.

### 4.2 Desorption solvent

The desorption solvent, commonly carbon disulfide, should be of chromatographic quality. It shall be free from compounds co-eluting with the substances of interest. Check the purity of each new batch of solvent.

NOTE Carbon disulfide is normally recommended for the desorption of non-polar compounds from activated carbon. For polar compounds and mixtures of polar and non-polar compounds, there is no ideal universal desorption solvent. Dichloromethane, methanol, higher alcohols, dimethylformamide and acetonitrile have been used as eluants, either singly or mixed with each other or carbon disulfide. Dichloromethane may cause corrosion in the flame ionization detector.

The use of carbon disulfide desorption solvent can result in problems when polar analytes are collected from humid atmospheres. Polar analytes may be soluble in a water phase which forms following desorption with carbon disulfide when sufficient water is collected with the sample. A desorption solvent modifier shall be present at a sufficient concentration to result in a homogeneous solution in desorbed samples; dimethylformamide may be suitable for this purpose.

### 4.3 Sorbents

#### 4.3.1 Activated charcoal

A particle size of 0,35 mm to 0,85 mm is recommended. Before packing the samplers, the charcoal shall be heated in an inert atmosphere, e.g. high-purity nitrogen, at approximately 600 °C for 1 h. To prevent recontamination of the charcoal, it shall be kept in a clean atmosphere during cooling to room temperature, storage and loading into the samplers. Samplers prepacked by the manufacturer with pre-conditioned charcoal are also available and require no further conditioning.

NOTE 1 Activated charcoal is usually processed from coconut shells. Some manufacturers recommend synthetic carbons as alternatives to charcoal of biological origin (see annexes A and B).

NOTE 2 The sorptive capacity and desorption efficiency of different batches of activated charcoal may vary. Commercial samplers, if used, should be purchased from the same batch and in sufficient number to provide consistent performance for a definite period of time.

#### 4.3.2 Other sorbents

Sorbents other than charcoal may be used for certain applications (see annex B).

NOTE A description of sorbent types is given in annex A. Equivalent sorbents may be used.

## 4.4 Calibration standards

### 4.4.1 General

Calibration blend solutions are required in order to compare the concentrations of desorbed solutions (7.2) with those calibration standards in the gas chromatographic analysis. Such solutions should be prepared in a way that is traceable to national standards.

An internal standard, for example trifluorotoluene or 3-bromofluorobenzene, is optional. If used, it should not interfere with the compounds of interest and it should not be removed from the elution solvent by the sorbent. In the context of this method, the purpose of the internal standard is to correct for small variations in the injection volume. The use of an internal standard as a surrogate to correct for desorption efficiency (e.g. *n*-propyl acetate in the analysis of *n*-butyl acetate) is not recommended. Desorption efficiency should be determined directly with the compounds of interest (7.3).

Storage times for calibration solutions vary according to application. Typically, carbon disulfide dilutions should be prepared fresh weekly, or more frequently if evidence is noted of decomposition or evaporation.

**NOTE** In the analysis of complex mixtures, calibration blends of the pure compounds may be prepared before dilution with the elution solvent. Examples of three calibration blends are listed here. These have been used in the analysis of mixed solvents in paints, thinners, adhesives, cleaning fluids and miscellaneous commercial products. The components are arranged to give resolved peaks on both BP-1 and BP-10 phases<sup>1)</sup>. Other blends may be more appropriate on different columns or in other applications.

- a) Blend 1 consists of: *n*-hexane, *n*-heptane, *n*-octane, *n*-decane, *n*-undecane, *n*-dodecane, benzene, toluene, *o*-xylene, *p*-xylene, *n*-propylbenzene, isopropylbenzene, *o*-ethyltoluene, *m*-ethyltoluene, *p*-ethyltoluene, 1,2,4-trimethylbenzene, 1,3,5-trimethylbenzene, *n*-propyl acetate, *n*-butyl acetate, isobutyl acetate, butoxyethyl acetate.
- b) Blend 2 consists of: isopropanol, isobutanol, *n*-butanol, 1-methoxy-2-propanol, butoxyethanol, toluene, ethylbenzene, 1,2,3-trimethylbenzene, ethyl acetate, ethoxyethyl acetate.
- c) Blend 3 consists of: acetone, 2-butanone, 4-methylpentan-2-one, cyclohexanone, 2-methylcyclohexanone, 3-methylcyclohexanone, 4-methylcyclohexanone, isopropyl acetate, *n*-nonane, toluene.

In the above examples, calibration blends 1-3 are stable for at least one year when stored in dark glass bottles with polytetrafluoroethylene (PTFE)-lined screw-caps at less than 4 °C.

#### 4.4.2 Solution containing approximately 10 mg/ml of each liquid component

Weigh 1 g of the substance or substances of interest into a 100 ml volumetric flask, starting with the least volatile substance. Make up to 100 ml with desorption solvent (4.2), stopper and shake to mix.

#### 4.4.3 Solution containing approximately 1 mg/ml of liquid components

Introduce 50 ml of desorption solvent into a 100 ml volumetric flask. Add 10 ml of solution 4.4.2. Make up to 100 ml with desorption solvent, stopper and shake to mix.

#### 4.4.4 Solution containing approximately 100 µg/ml of each liquid component.

Weigh 10 mg of the substance or substances of interest into a 100 ml volumetric flask, starting with the least volatile substance. Make up to 100 ml with desorption solvent (4.2), stopper and shake to mix.

#### 4.4.5 Solution containing approximately 10 µg/ml of liquid components

Introduce 50 ml of desorption solvent into a 100 ml volumetric flask. Add 10 ml of solution 4.4.4. Make up to 100 ml with desorption solvent, stopper and shake to mix.

#### 4.4.6 Solution containing approximately 1 mg/ml of gas components

For gases, e.g. ethylene oxide, a high level calibration solution may be prepared as follows. Obtain pure gas at atmospheric pressure by filling a small plastic gas-bag from a gas cylinder. Fill a precision 1 ml gas-tight syringe

<sup>1)</sup> BP-1 and BP-10 are examples of suitable products available commercially. This information is given for the convenience of users of this part of ISO 16200 and does not constitute endorsement by ISO of these product. Equivalent products may be used if they can be shown to lead to the same results. Annex C gives a non-exclusive list of products that are believed to be equivalent.

(5.5) with 1 ml of the pure gas and close the valve of the syringe. Using a septum vial of suitable capacity, add 2 ml desorption solvent and close with the septum cap. Insert the tip of the syringe needle through the septum cap into the desorption solvent. Open the valve and withdraw the plunger slightly to allow the desorption solvent to enter the syringe. The action of the gas dissolving in the desorption solvent creates a vacuum, and the syringe fills with solvent. Return the solution to the flask. Flush the syringe twice with the solution and return the washings to the flask. Calculate the mass of gas added using the gas laws, i.e. 1 mole of gas at STP occupies 22,4 l.

#### 4.4.7 Solution containing approximately 10 µg/ml of gas components

For gases, e.g. ethylene oxide, a low level calibration solution may be prepared as follows. Obtain pure gas at atmospheric pressure by filling a small plastic gas-bag from a gas cylinder. Fill a precision 10 µl gas-tight syringe (5.5) with 10 µl of the pure gas and close the valve of the syringe. Using a septum vial of suitable capacity, add 2 ml desorption solvent and close with the septum cap. Insert the tip of the syringe needle through the septum cap into the desorption solvent. Open the valve and withdraw the plunger slightly to allow the desorption solvent to enter the syringe. The action of the gas dissolving in the desorption solvent creates a vacuum, and the syringe fills with solvent. Return the solution to the flask. Flush the syringe twice with the solution and return the washings to the flask. Calculate the mass of gas added using the gas laws, i.e. 1 mole of gas at STP occupies 22,4 l.

#### 4.5 Calibration blend atmospheres (for 4.6 and 7.4)

Prepare standard atmospheres of known concentrations of the compound(s) of interest by a recognized method. See for example ISO 6141, ISO 6145 and ISO 6349 [1-3]. If the procedure is not applied under conditions that will allow the establishment of full traceability of the generated concentrations to primary standards, confirm the delivered concentrations using an independent procedure.

#### 4.6 Standards for desorption efficiency (for 7.3)

Prepare standards by exposure of the samplers to the standard atmosphere (4.5) for an accurately known time. After exposure, remove and seal the samplers.

If the generation of standard atmospheres is not practicable, the standards may be prepared by a liquid spiking procedure, provided that the accuracy of the spiking technique is established by using procedures giving spiking levels traceable to primary standards of mass and/or volume, or is confirmed by an independent procedure. This is the procedure usually recommended by manufacturers; follow the manufacturer's guidance for specific instructions. These will vary significantly with the sampler type, and some examples are given in annexes E to J. In principle, load the devices by injecting aliquots of standard solutions (4.4) of accurately known mass or volume at three or more levels onto clean samplers, seal the samplers and leave to equilibrate.

## 5 Apparatus

Ordinary laboratory apparatus and the following.

### 5.1 Diffusive samplers.

A number of solvent-desorption diffusive samplers are available commercially. Information on available devices is given in annex D and information of available sorbent types is given in annex A. Manufacturer-supplied data on the characteristics of some typical sampler types are given in annexes E to J.

NOTE Self-packed samplers should not be used unless they can be shown to have reproducible and constant sampling rates.

The desorption efficiency (*D*) for each batch of samplers shall be checked by one of the methods described in 4.6 and 7.3.

Some diffusive samplers have a back-up section, which can be used as a check on overload of the sampling section.

**5.2 Gas chromatograph**, fitted with a flame ionization (FID), photoionization detector, mass spectrometric or other suitable detector, capable of detecting an injection of 0,5 ng toluene, with a signal-to-noise ratio of at least 5:1.

NOTE Dichloromethane may cause corrosion in the FID of some instruments.

The gas chromatograph column shall be capable of separating the analytes of interest from other components. Examples of suitable choices are 50 m x 0,22 mm fused silica columns with BP-1 or BP-10 stationary phases. A typical film thickness is in the range 0,5  $\mu\text{m}$  to 2,0  $\mu\text{m}$ . Typical operating conditions for these columns might be temperature programming from 50 °C to 200 °C at 5 °C/min with a carrier gas flowrate of 0,7 ml/min to 0,8 ml/min helium. Annex C gives a list of equivalent phases.

### 5.3 Autosampler.

Autosamplers are commercially available with liquid-chilled sample trays, suitable for the analysis of volatile solvents.

### 5.4 Volumetric glassware.

Precision volumetric flasks of an accurately known volume, to be used for the preparation of calibration blend solutions (4.4). These should be obtained from suppliers issuing certificates of calibration traceable to primary standards or be traceably calibrated in the laboratory by weighing of the solvent(s) applied.

### 5.5 Syringes (for 4.4.6 and 4.4.7).

Precision gas-tight syringes of accurately known volumes of 1,0 ml and 10  $\mu\text{l}$ , readable to 0,01 ml and 0,1  $\mu\text{l}$  respectively.

## 6 Sampling

Select a diffusive sampler appropriate for the compound or mixture to be sampled. Guidance on the availability of suitable samplers is given in 5.1. Annex B gives information on the availability of manufacturer-supplied diffusive uptake rates for a range of VOCs and, where appropriate, any special conditions, e.g. a variation from the basic device design. If a device is not listed for a particular VOC, it will be necessary to calibrate the device according to 7.4.

Follow the manufacturer's guidance for specific sampling instructions. These will vary significantly with the sampler type, and some examples are given in annexes E to J. In principle, any protective cover is removed before the diffusive sampler is exposed to the target atmosphere, and the sampler is re-sealed again at the end.

When intended for personal sampling, mount the sampler in the breathing zone as defined in accordance with EN 1540. When used for fixed-location sampling, choose a suitable sampling site. In either case, the sampler should have unrestricted access to the sampled atmosphere, i.e. it should not be obscured by the wearer's clothing or other objects.

NOTE Some designs of diffusive sampler are affected by air velocity. See 7.4 for details.

The exposure time recommended for the VOCs covered by this part of ISO 16200 is normally 8 h for workplace monitoring. If the maximum exposure time recommended is less than 8 h, this is indicated in annex B. Sampling over shorter periods is possible, down to 30 min for workplace monitoring, but the limits of the measurable concentration range increase accordingly. For example, for a 4-h sampling period, the concentration range is approximately 2  $\text{mg}/\text{m}^3$  to 2 000  $\text{mg}/\text{m}^3$ .

Samplers should be uniquely labelled. Solvent-containing paints and markers or adhesive labels should not be used to label the samplers.

Record air temperature and barometric pressure periodically during sampling if it is desired either to express concentrations reduced to specific conditions (8.2) or to express concentrations as volume fractions (8.3).

Field blanks should be prepared by using samplers identical to those used for sampling and subjecting them to the same handling procedure as the samples except for the actual period of sampling. Label these as blanks.

## 7 Procedure

**CAUTION** — This part of ISO 16200 does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this part of ISO 16200 to establish appropriate health and safety practices and determine the applicability of regulatory limitations prior to use.

### 7.1 Desorption

If samples are not to be analysed within 8 h, they shall be placed in a clean, uncoated, sealed metal or glass container.

In each case, carry out the desorption in a clean atmosphere in a fume hood. Desorb the sample blanks in the same way as the samples.

Follow the manufacturer's guidance for specific desorption instructions. These will vary significantly with the sampler type, and some examples are given in annexes D to H. In principle, the collected VOC is extracted from the sorbent (4.3) with a suitable solvent (4.2). In some cases, the desorption is done without disassembling the device; in others, the sorbent is removed and desorbed in a separate vessel.

If there is a back-up section of sorbent, this should be desorbed separately.

NOTE In some circumstances, a higher  $D$  may be obtained with ultrasonic extraction as opposed to mechanical shaking.

### 7.2 Analysis

Set up the gas chromatograph for the analysis of volatile organic compounds. A variety of chromatographic columns may be used for the analysis of these compounds (see 5.2). The choice will depend largely on which compounds, if any, are present that might interfere in the chromatographic analysis.

Inject a known fixed volume (1  $\mu\text{l}$  to 5  $\mu\text{l}$ ) of each standard solution (4.4) into the gas chromatograph. A standardized injection technique should be used so that repeatable peak heights or areas are obtained. Typically, for a series of replicate injections, the relative standard deviation should be better than  $\pm 2\%$ .

NOTE Autosamplers normally achieve better than  $\pm 1\%$ .

Inject the same fixed volume of solution from the desorbed sample into the gas chromatograph. Read from the calibration graph the concentration of the analyte in the desorbed sample. Analyse the sample blank and the samples used to determine desorption efficiency in the same way.

Correspondence of retention time on a single column should not be regarded as proof of identity. The retention indices of about 160 VOCs on BP-1 and BP-10 phases are given in annex K. They are a useful guide to elution order on these phases or their near equivalent, but are not definitive, since exact values depend on temperature programme, carrier flowrate and other factors.

If a back-up section contains more than 10 % of the sample, discard the sample as unreliable.

### 7.3 Determination of desorption efficiency

The desorption efficiencies ( $D$ ) of VOCs can vary with the type and batch of sorbent used. Thus it is necessary for each type of sorbent and for each analyte to determine  $D$  over the sample concentration range. Samples are prepared as described in 4.6 and analysed as described in 7.2. Prepare at least three samples at each load level.  $D$  is then the amount recovered divided by the amount applied.

Alternatively to the liquid spiking procedure (4.6), the phase-equilibrium method may be applied, in which accurately known volumes of standard solutions are added to unused blank samplers with the difference in concentration before and after addition being determined.

If the desorption efficiency data can be shown to be a homogeneous set, e.g. by applying the Bartlett test, then  $D$  is given by the pooled mean. Otherwise the data should be examined to determine whether it can be modelled using a smooth non-linear equation, with  $D$  increasing with the ratio of analyte mass to sorbent mass. In such cases,  $D$  can be estimated using this curve.

If the  $D$  at the load level is less than 0,75 (75 %), a sample result corresponding to that level should be discarded (but see note 2).

NOTE 1 Indicative values of  $D$  for single compounds may be obtained from the manufacturer. Actual values should always be determined at the time of analysis. The desorption efficiency will vary with the mass loading of compound on the sorbent tube; the variation is usually significant where the average value is below 90 %.

NOTE 2 Where mixtures of non-polar analytes are desorbed with pure carbon disulfide, the mutual concentration effect on  $D$  is generally negligible. If the composition of a mixture of polar and non-polar analytes is known approximately,  $D$  values should be established with a similar mixture. It may not be possible to achieve greater than 75 %  $D$  for all components of such a mixture with a single desorption solvent. Provided that it can be established that the  $D$  is consistent and that no better solvent has been found, then a compromise is acceptable, although where possible, the taking of a second sample and optimizing desorption conditions for both polar and non-polar analytes is preferred.

NOTE 3 The liquid spiking and phase-equilibrium methods may not take account of high humidity at the time of sampling. Adsorbed water vapour is a factor which can be simulated by addition of water to the sorbent. This should be investigated when sampling water-soluble compounds from atmospheres of high humidity.

NOTE 4 The phase-equilibrium method may give rise to incorrect values for  $D$  [4-7].

## 7.4 Calibration of uptake rate

Diffusive uptake rates for some typical samplers are given in annex B. Data on approximately 200 compounds have been compiled from the manufacturers' latest sources available [8-13]. Unless otherwise specified, rates are appropriate for standard conditions (25 °C and 101 kPa). Some rates already include an allowance for desorption efficiency. Uptake rates listed as type C evaluation were calculated by the manufacturers using geometric constants and diffusion coefficients either known experimentally [14] or estimated from empirical equations [15-18].

If an uptake rate for a particular compound or device is not available, it shall be determined experimentally. Expose samplers for a measured time to a calibration blend atmosphere (4.4.8) containing the compound or compounds of interest. The concentrations and times of exposure should be typical of the intended use of the sampler. Analyse the samplers according to 7.2 and calculate the diffusive sampling rate as mass collected per unit concentration per unit time. Equation (7) can then be used to convert the value to cubic centimetres per minute ( $\text{cm}^3/\text{min}$ ). Full details of an appropriate procedure to be followed are given in EN 838.

The uptake rate of samplers is not significantly affected by air movement, provided the air velocity exceeds a threshold value which depends on design. Generally, air velocities greater than 0,1 m/s are sufficient for the samplers described in annexes E to J. Other samplers may have different characteristics [19]. Manufacturer documentation should be consulted for any specific recommendations.

For an ideal diffusive sampler, the dependence of  $U$  on absolute temperature and pressure is governed by that of the diffusion coefficient,  $D'$ , of the analyte. The latter dependence is given by:

$$D' = f(T^{n+1}, P^{-1}) \quad (1)$$

with  $0,5 < n < 1,0$ .

Hence, the dependence of  $U$ , expressed in units of cubic centimetres per minute or equivalent is:

$$U = f(T^{n+1}, P^{-1}) \quad (2)$$

When  $U'$  is expressed in units of pg/ppb-min or equivalent (see 8.3), then the dependence is given by:

$$U' = f(T^n) \quad (3)$$

In the latter case, the dependence will be of the order of 0,2 %/K to 0,4 %/K. In the case of a non-ideal sampler, the temperature dependence of  $U'$  may be compensated by the temperature dependence of the sorption coefficient of the analyte. In any case, accurate knowledge of the average temperature and pressure during the sampling period is important for a correct application of equations (4) and (5) (clause 8).

## 8 Calculations

### 8.1 General

Prepare a log-transformed calibration graph by plotting the base-ten logarithm of the areas of the analyte peaks corrected for blank levels on the vertical scale against the base-ten logarithm of the concentration of the analyte, in micrograms per millilitre ( $\mu\text{g/ml}$ ), in the injected aliquot of the calibration blend solutions.

NOTE Other methods of weighting calibration points, such as linear, exponential or polynomial plots, may be more or less suitable, depending on the linearity of the detector response and the software available.

### 8.2 Mass concentration of analyte

Calculate the concentration of the analyte in the sampled air, in milligrams per cubic metre, by means of the following equation:

$$c_m = \frac{m_1 + m_2 - m_3}{D \cdot U \cdot t} \times 10^6 \quad (4)$$

where

$c_m$  is the concentration of analyte in the air sampled, in milligrams per cubic metre;

$D$  is the desorption efficiency as found in 7.3, for a mass  $m_1$ ;

$m_1$  is the mass of analyte present in the sample as found in 7.2, in milligrams;

$m_2$  is the mass of analyte present in the sample (back-up section, if used) as found in 7.2, in milligrams;

$m_3$  is the mass of analyte present in the blank tube, in milligrams;

$U$  is the diffusive uptake rate, in cubic centimetres per minute (annex B or 7.4)

$t$  is the exposure time, in minutes.

The value of  $U$  applied should be that for the temperature and pressure of sampling (see 7.4).

If it is desired to express concentrations reduced to specified conditions, e.g. 25 °C and 101 kPa, then:

$$c_c = c_m \cdot \frac{101}{p} \cdot \frac{T + 273}{298} \quad (5)$$

where

$c_c$  is the concentration of analyte in the air sampled, reduced to specified conditions, in milligrams per cubic metre;

$p$  is the actual pressure of the air sampled, in kilopascals;

$T$  is the actual temperature of the air sampled, in degrees Celsius.

### 8.3 Volume concentration of analyte

Calculate the volume fraction of the analyte in air, in millilitres per cubic metre, by means of the following equation:

$$c_V = \frac{m_1 + m_2 - m_3}{D \cdot U' \cdot t} \times 10^6 \quad (6)$$

where

$c_V$  is the volume fraction of the analyte in air, in millilitres per cubic metre;

$U'$  is the diffusive uptake rate, in nanograms per (millilitres per cubic metre) per minute.

### 8.4 Uptake rates

Uptake rates in cubic centimetres per minute ( $U$ ) and in nanograms per (millilitres per cubic metre) per minute ( $U'$ ) are related by:

$$U = U' \cdot \frac{24,5}{M} \cdot \frac{101}{p} \cdot \frac{T + 273}{298} \quad (7)$$

where

$M$  is the molecular mass of the analyte of interest, in grams per mole;

24,5 is the molar volume at 25° C and 101 kPa.

## 9 Interferences

Organic components which have the same or nearly the same retention time as the analyte of interest during the gas chromatographic analysis will interfere. Interferences can be minimized by proper selection of gas chromatographic columns and conditions.

High humidity may affect the recovery of some compounds from samplers, particularly for those using activated charcoal. The method description should be consulted for specific advice.

## 10 Precision and bias

The procedure described in this part of ISO 16200, when used for the determination of concentrations of VOCs in workplace air, shall meet the requirements of EN 838 or an equivalent test procedure. For the purposes of this part of ISO 16200, the HSE protocol [20] and the NIOSH protocol [21, 22] are taken to be equivalent to EN 838. EN 838 assigns levels of evaluation defined as:

- **1A:** Full evaluation of uptake rate, including effects of time, concentration, temperature, humidity, back-diffusion, storage, desorption efficiency and air velocity; overall uncertainty including bias and random errors  $\leq 30\%$ .
- **1B:** Partial evaluation of an analogue within a homologous series in which upper and lower members have been shown to comply with level 1A.

Evaluations to EN 838 level 1A or 1B are time-consuming. This part of ISO 16200 acknowledges that in the absence of experimental data, empirical data based on the ideal uptake rate could be used, subject to certain limitations. Annex B summarizes the levels of evaluation as follows:

- **A:** Full (EN 838 level 1A or NIOSH protocol or close equivalent).

- **B:** Partial (EN 838 level 1B, or other tests in which experimental uptake rates were measured over a more limited range than that specified by level 1A or 1B, as permitted in EN 482 [23]).
- **C:** Theoretical or ideal uptake rates are calculated from known or estimated diffusion coefficients and a geometric constant characteristic of the sampling device (the ratio of the effective area to the diffusion path length,  $A/l$  cm). The geometric constant may itself be estimated from selected experimental diffusion coefficients and uptake rates if the device moderates diffusion through a porous barrier or operates by radial diffusion.

There is no general consensus on the tests to be included in a partial level B evaluation. The significance of certain tests, such as back-diffusion or desorption efficiency, depends on the type of sampler and the application. One manufacturer (SKC Inc) distinguishes between its evaluation of an analogue within a homologous series, designated *bi-level*, in which the lower member complies fully with EN 838 level 1A, and more limited tests designated *partial* [23]. For the purposes of this part of ISO 16200, field evaluations can also comply with level B, subject to comparison of the sampler with an independent method which has also been validated according to an established protocol, e.g. a pumped sorbent tube or a different diffusive sampler method.

NOTE 1 For definitions of precision and related terms, see for example ISO 5725 [24] or IUPAC [25].

NOTE 2 Level C rates should be used with caution, and confirmed experimentally as soon as practicable by the procedure in 7.4.

## 11 Storage and transport

The long-term stability of hydrocarbons and some chloroalkanes on charcoal is good. The long-term stability of many polar compounds on charcoal is unknown. Use of a refrigerator or freezer for storage and transport will generally improve stability.

## 12 Test report

The test report shall contain at least the following information:

- a) complete identification of the sample;
- b) reference to this part of ISO 16200 or any supplementary International Standard;
- c) the sampling location, sampling time period and volume of air sampled;
- d) the barometric pressure and temperature, if required in clause 6;
- e) the test result;
- f) any unusual features noted during the determination;
- g) any operation not included in this part of ISO 16200 or in the International Standard to which reference is made or which is regarded as optional.

## 13 Quality control

An appropriate level of quality control should be employed (see [26, 27] or equivalent).

The field blank is acceptable if artefact peaks are no greater than 10 % of the typical areas of the analytes of interest.

## Annex A (informative)

### Description of sorbent types

The sorbents are examples of suitable products available commercially. This information is given for the convenience of users of this part of ISO 16200 and does not constitute endorsement by ISO of these product. Equivalent products may be used if they can be shown to lead to the same results.

Sorbent	Type
Carbon	Coconut shell
Carbon	Petroleum-based
Anasorb <sup>2)</sup> 727	Beaded microporous polymer with hydrophobic surface
Chromosorb <sup>3)</sup> 106	Beaded microporous polymer with hydrophobic surface
Anasorb <sup>2)</sup> 747	Beaded active carbon derived from petroleum precursors
Silica gel	
Tenax <sup>4)</sup> TA	Poly(diphenyl oxide)
Porapak <sup>5)</sup> R	

STANDARDSISO.COM : Click to view the full PDF of ISO 16200-2:2000

---

2) Anasorb<sup>TM</sup> is a trademark of SKC Inc., USA. Anasorb 727 and Chromosorb 106 are believed to be equivalent.

3) Chromosorb<sup>TM</sup> is a trademark of Manville Corp., USA. Anasorb 727 and Chromosorb 106 are believed to be equivalent.

4) Tenax<sup>TM</sup> is a trademark of Enka Research Institute NV, NL.

5) Porapak<sup>TM</sup> is a trademark of Waters Associates Inc., USA.

## Annex B (informative)

### Diffusive sampling rates (cm<sup>3</sup>/min)

Compound	Type A sampler		Type B sampler		Type C sampler		Type D sampler		Type E sampler	
	Level <sup>a</sup>	Uptake rate <sup>b</sup>	Level <sup>a</sup>	Uptake rate	Level <sup>a</sup>	Uptake rate <sup>c</sup>	Level <sup>a</sup>	Uptake rate <sup>d</sup>	Level <sup>a</sup>	Uptake rate
<b>Hydrocarbons</b>										
1,3-butadiene	C	7,61	A	42,8 <sup>e</sup>					A	8,03
n-pentane	C	6,32	B	35,3 <sup>e</sup>	A	14,9	A	74	C	6,15
1-pentene			C	35,9	C	16,3				6,39
2-methylpentane			C	31,4	C	14,1				5,51
2-methyl-1,3-butadiene	C	6,51	C	36,6 <sup>e</sup>						6,67
n-hexane	B	5,49	A	32,0	B	14,3	A	66	A	5,51
n-hexane			A	31,7 <sup>f</sup>	B	14,3 <sup>g</sup>	A	58		
n-heptane	B	4,83	B	28,9	B	13,9 <sup>g</sup>			B	5,01
1-heptene			C	29,3	C	13,1 <sup>g</sup>	A	53		5,15
n-octane	B	4,62	B	26,6	B	12,7 <sup>g</sup>			B	4,61
n-nonane	C	4,32	B	24,6	B	10,6 <sup>g</sup>			B	4,28
n-decane	C	4,04	B	23,1	C	10,2	A	43	C	4
n-dodecane			B	21,5						3,55
cyclopentane			C	36,2 <sup>e</sup>						6,7
cyclopentadiene			C	39,5					C	7,3
dicyclopentadiene			C	23,6	C	11,8				
cyclohexane	B	5,58	B	32,4	B	15,6	A	47	B	5,92
cyclohexene	B	5,72	B	32,3	C	15,4			C	6,15
methylcyclohexane	C	5,09	B	28,9	B	14,2			B	5,33
<i>trans</i> -1,2-dimethyl-cyclohexane			C	25,4	C	12,4				4,86
4-vinyl-1-cyclohexene			B	27,9					C	5,15
benzene	A	6,44	B	35,5	A	16,0	A	80	A	6,76
toluene	A	5,72	A	31,4	B	14,5	A	74	A	6,01
ethylbenzene	B	5,20	C	27,3	B	12,9				5,34
<i>m</i> -xylene	B	5,03	B	27,3	B	12,5 <sup>g</sup>	A	61	A	5,4
<i>o</i> -xylene	B	5,45	B	27,3	B	11,9 <sup>g</sup>	A	61	A	5,4
<i>p</i> -xylene	B	5,04	B	27,3	B	12,8 <sup>g</sup>	A	61	A	5,4
styrene	B	5,26	A	28,9	A	13,7 <sup>h</sup>	A	61	C	5,52
styrene					A	13,7 <sup>j</sup>				
divinyl benzene			C	23,3						4,72
vinyltoluene			C	25,1	C	12,3 <sup>h</sup>			B	5,01
$\alpha$ -methylstyrene	C	4,88	B	25,0	A	12,6 <sup>h</sup>			B	5,02
$\alpha$ -methylstyrene					A	12,6 <sup>j</sup>				
isopropylbenzene (cumene)	C	5,08	B	24,5	B	12,8	A	58	C	4,87
isopropenylbenzene	C	4,88								5,02
2-ethyltoluene	C	4,78	C	24,5						4,87
3-ethyltoluene	C	4,80	C	24,6						4,87
4-ethyltoluene	C	4,79	C	24,5						4,87

Compound	Type A sampler		Type B sampler		Type C sampler		Type D sampler		Type E sampler	
	Level <sup>a</sup>	Uptake rate <sup>b</sup>	Level <sup>a</sup>	Uptake rate	Level <sup>a</sup>	Uptake rate <sup>c</sup>	Level <sup>a</sup>	Uptake rate <sup>d</sup>	Level <sup>a</sup>	Uptake rate
1,2,3-trimethylbenzene	C	4,95	C	24,3	C	12,0				4,86
1,2,4-trimethylbenzene	C	4,95	C	24,4	C	12,1			C	4,86
1,3,5-trimethylbenzene	C	4,95	B	26,3	C	12,1			B	4,86
1,2,3,4-tetramethylbenzene			C	22,2	C	11,1			C	4,48
1,2,3,5-tetramethylbenzene					C	11,2			C	4,48
1,2,4,5-tetraethylbenzene					C	11,2			C	4,48
p-t-butyltoluene	C	4,28	B	20,7	B	10,4 <sup>g</sup>			C	4,19
naphthalene	C	4,87	C	24,6	C	12,2			A	5,05
divinylbenzene			C	23,3					C	4,72
α-pinene	C	4,26	C	22,8	A	11,4 <sup>j</sup>			C	4,48
β-pinene	C	4,26	C	22,7	B	11,4 <sup>j</sup>			C	4,45
Δ <sup>3</sup> -carene			C	22,0	B	11,4 <sup>j</sup>				4,37
limonene	C	4,24	C	21,9	C	11,4 <sup>j</sup>			C	4,34
dicyclopentadiene			C	23,6					A	4,69
dodecene			C	21,9					C	3,61
2,2,4-trimethylpentane			C	27,1			A	55	C	4,6
phenyl-cyclohexane			C	20,0					C	4,09
phenyl-cyclohexene			C	20,3					C	4,18
propane									C	8,26
4-vinyl-cyclohexane			C	26,0					C	5,15
<b>Halocarbons</b>										
methyl chloride	C	9,57				k			C	10,68
methyl bromide	C	8,22	C	40,9 <sup>e</sup>					C	9,39
methyl iodide	C	7,24	C	36,7	C	18,7			C	8,46
dichloromethane	B	7,78	A	37,9 <sup>e</sup>	A	14,7	B	90 <sup>l</sup>	A	8,04
chlorobromomethane	C	7,15	B	34,4	C	15,4	B	70	C	8,18
chlorotrifluoromethane										8,55
bromoform	C	5,75	C	29,3	C	21,2			C	6,62
chloroform	C	6,66	C	33,5	B	13,0 <sup>g</sup>			B	7,3
carbon tetrachloride	C	6,21	B	30,2	B	14,1 <sup>g</sup>			B	6,43
carbon tetrabromide			C	26,6						5,76
vinyl chloride	B	8,29	B	40,8					B	9,1
vinyl bromide			C	37,0	C	18,2			B	8,21
bromoethane		6,95	B	36,4	C	18,1			C	7,75
1,2-dibromoethane		6,20	B	29,6	C	14,7			B	6,36
1,1-dichloroethane		6,89	C	33,2					B	7,2
1,2-dichloroethane	C	6,80	B	33,2	B	14,2 <sup>g</sup>			C	7,36
1,1-dichloroethene (vinylidene chloride)	C	6,89	C	35,1	B	12,3 <sup>g</sup>			B	7,61
1,2-dichloroethene	C	6,83	B	35,2	A	14,8			B	7,66
trichloroethene	B	6,56	B	31,1	A	14,9	A	65	C	6,4
1,1,1-trichloroethane	B	5,96	A	30,9	B	14,1 <sup>g</sup>	A	47	B	6,36
1,1,2-trichloroethane	C	5,94	B	29,7	B	12,5 <sup>g</sup>			B	6,41
tetrachloroethene	B	5,98	A	28,3	A	12,9	A	65	C	5,96

Compound	Type A sampler		Type B sampler		Type C sampler		Type D sampler		Type E sampler	
	Level <sup>a</sup>	Uptake rate <sup>b</sup>	Level <sup>a</sup>	Uptake rate	Level <sup>a</sup>	Uptake rate <sup>c</sup>	Level <sup>a</sup>	Uptake rate <sup>d</sup>	Level <sup>a</sup>	Uptake rate
1,1,2,2-tetrachloroethane	C	5,42	C	28,4	B	11,8 <sup>g</sup>			B	5,72
hexachloroethane	C	4,56	C	26,4	C	11,5				4,81
1-bromobutane	C	5,92	C	29,0						5,9
bromopropane			A	31,7	A	145			B	6,18
halothane	B	5,70	C	30,2						
halothane			B	24,0 <sup>m</sup>						
halothane			B	23,1 <sup>n</sup>						
enflurane	B	5,31	C	28,3	C	13,8 <sup>h</sup>			B	5,52
isoflurane	B	5,30	C	28,3	B	13,7 <sup>h</sup>			B	5,56
sevoflurane	C	5,03	C	27,3	C	13,1 <sup>h</sup>			B	5,16
desflurane			C	30,1 <sup>e</sup>	C	14,8 <sup>h</sup>			B	5,88
1,1-dichloro-2,2,2-trifluoroethane (HCFC 123)			B	30,9					C	6,36
1,1,1,2-tetrafluoroethane (HFC134a)			B	37,1					C	7,93
1,1,2-trichloro-1,2,2-trifluoroethane	C	5,47	C	29,1 <sup>e</sup>	C	14,1			B	5,72
2-chloro-1,1,1,2-tetrafluoroethane (HCFC 124)			B	35,8					C	6,9
1,1,1,2-tetrachloro-2,2-difluoroethane			C	27,5 <sup>e</sup>					B	5,37
1,1,2,2-tetrachloro-1,2-difluoroethane	C	5,11	C	28,2 <sup>e</sup>					C	5,37
1,2-dichloropropane (propylene dichloride)	C	5,73	B	30,6	B	14,3 <sup>g</sup>	A	66	C	6,41
3-chloropropene (allyl chloride)			C	35,1	C	17,8			C	7,54
1,2,3-trichloropropane	C	5,16	C	27,4	B	11,9 <sup>g</sup>			C	5,79
cis-1,3-dichloropropene			C	30,7	C	15,2				6,57
2-chloro-1,3-butadiene (chloroprene)	C	6,23	C	32,2					C	6,83
1-chloro-2,3-epoxypropane (epichlorohydrin)	C	6,18	C	29,6	C	16,0 <sup>h</sup>			C	8,19
chlorobenzene	B	5,60	B	29,3	C	14,2			B	6,01
benzyl chloride			C	27,2	C	12,3			B	5,43
o-dichlorobenzene	C	5,01	B	27,8	C	12,6			B	5,44
m-dichlorobenzene			C	26,7	C	12,7				5,44
p-dichlorobenzene	C	5,03	B	27,8	C	12,7			B	5,44
α-chlorotoluene	C	5,35							C	5,43
o-chlorotoluene			C	27,3	C	13 <sup>g</sup>			B	5,39
o-chlorostyrene			C	26,0	A	9,8 <sup>g,h,j</sup>			B	5,05
trifluoromethyl benzene			C	27,8	B	13,3 <sup>g</sup>				
1-chloro-4-(trifluoromethyl) benzene					B	11,8 <sup>g</sup>				

Compound	Type A sampler		Type B sampler		Type C sampler		Type D sampler		Type E sampler	
	Level <sup>a</sup>	Uptake rate <sup>b</sup>	Level <sup>a</sup>	Uptake rate	Level <sup>a</sup>	Uptake rate <sup>c</sup>	Level <sup>a</sup>	Uptake rate <sup>d</sup>	Level <sup>a</sup>	Uptake rate
1,1-dichloro-1-fluoroethane (HCFC 141b)			C	33,0 <sup>e</sup>					B	6,89
dichloro(1,3)pentafluoropropane									C	5,33
dichloro(3,3)pentafluoropropane									C	5,33
dichlorodifluoromethane (CFC12)			C	36,5					B	7,67
dichlorofluoromethane (CFC21)			C	36,9					B	8,08
dichlorotetrafluoroethane (CFC114)			C	31,1					B	6,13
ethyl chloride (chloroethane)			C	38,8					C	8,52
fluorotrichloromethane (CFC11)			C	33,2					B	6,98
hexachlorobutadiene			C	22,9					C	4,39
hexachlorocyclopentadiene			C	22,1					C	4,26
methoxyflurane (metofane)			C	27,8					C	5,52
trifluoroethanol			C	34,3					C	7,64
<b>Esters</b>										
methyl formate	C	8,17	C	45,0 <sup>e</sup>					B	8,64
ethyl formate	C	7,32	C	38,8					C	7,27
methyl acetate	C	7,34	B	37,0 <sup>e</sup>					C	7,28
ethyl acetate	B	6,46	B	34,5	C	15,6	A	64	B	6,34
n-propyl acetate	C	5,76	B	30,1	C	14,6			B	5,65
isopropyl acetate	C	5,78	C	31,7	C	14,1			B	5,65
n-butyl acetate	B	5,04	C	31,6	C	12,7	A	60	B	5,12
isobutyl acetate	C	4,97	B	31,0	C	12,8	A	63	B	5,12
s-butyl acetate	C	4,98	B	28,6	C	12,9			C	5,12
t-butyl acetate	C	5,01	C	29,4	C	12,9			C	5,12
n-amyl acetate	C	4,58	B	26,0	C	11,8			C	4,71
isoamyl acetate	C	4,60	C	27,2	C	11,8			B	4,71
s-amyl acetate			C	27,2	C	11,9			C	4,71
1,3-dimethylbutyl acetate (s-hexyl acetate)			C	25,5	C	11,1				4,35
ethylhexyl acetate			C	22,9	C	9,8			C	3,81
ethyl propionate	C	5,42	C	31,2	C	14,0				5,65
methyl acrylate	C	6,17	C	35,8	A	15,7 <sup>h</sup>			B	6,61
ethyl acrylate	C	5,52	C	32,2	B	13,7 <sup>g,h</sup>			C	5,85
n-butyl acrylate	C	4,69	C	27,3	B	11,7 <sup>g,h</sup>			C	4,83
isobutyl acrylate					C	12,1 <sup>h</sup>				4,82
methyl methacrylate	C	5,56	C	31,8	B	13,1 <sup>g,h</sup>	A	68	B	5,86
ethyl methacrylate			C	29,4	C	13,1 <sup>h</sup>			C	5,28
methoxyethyl acetate (methyl cellosolve acetate)	C	5,14	B	29,0	C	13,1	A	64	B	5,34

Compound	Type A sampler		Type B sampler		Type C sampler		Type D sampler		Type E sampler	
	Level <sup>a</sup>	Uptake rate <sup>b</sup>	Level <sup>a</sup>	Uptake rate	Level <sup>a</sup>	Uptake rate <sup>c</sup>	Level <sup>a</sup>	Uptake rate <sup>d</sup>	Level <sup>a</sup>	Uptake rate
2-ethoxyethyl acetate (cellosolve acetate)	C	4,57	B	26,6	C	12,0	A	54	B	4,88
1-methoxy-2-propyl acetate (propylene glycol monomethyl ether) acetate)	C	5,26	B	25,2	C	12,2	A	60	B	4,87
2-methoxy-1-propyl acetate					C	12,0			C	5,35
2-butoxyethyl acetate (butyl cellosolve acetate)	C	4,39	B	24,3	C	10,5	A	41	C	4,18
vinyl acetate	C	6,20	C	35,8 <sup>p</sup>	A	161 <sup>h</sup>			B	6,59
benzyl acetate			C	22,6	C	11,3			C	4,59
ethyl lactate			C	29,1					B	5,35
<b>Alcohols and glycol ethers</b>										
ethanol	C	8,91	B	43,7 <sup>e</sup>	C	20,9 <sup>h</sup>			B	9,05
2-chloroethanol (ethylene chlorohydrin)	C	6,68	C	33,9					B	7,73
n-propanol	C	7,44	B	39,7	C	18,5 <sup>h</sup>			B	7,53
isopropanol	B		A	39,4 <sup>e</sup>	C	17,8 <sup>h</sup>	A	52 <sup>q</sup>	B	7,53
2-propen-1-ol (allyl alcohol)	C	7,66	C	40,4	C	18,4 <sup>h</sup>			C	7,93
n-butanol	B	6,46	B	34,3	C	15,5 <sup>h</sup>	A	74	C	6,52
isobutanol	B	6,08	B	35,9	C	15,6 <sup>h</sup>	A	77	B	6,51
s-butanol	B	6,73	C	34,8	C	15,6 <sup>h</sup>			C	6,51
t-butanol	C	6,55	C	35,2	C	15,8 <sup>h</sup>			C	6,5
n-amyl alcohol			B	31,2	C	13,9 <sup>h</sup>				5,78
isoamyl alcohol	C	5,46	B	32,3	C	13,9 <sup>h</sup>			B	5,78
s-amyl alcohol			C	31,2						5,77
hexyl alcohol			C	28,5	C	12,6				5,23
methyl amyl alcohol (methyl isobutyl carbinol)			C	29,2	C	12,8 <sup>h</sup>			C	5,22
2-ethylhexanol	C	4,38	C	25,2	C	10,9			C	4,42
isooctyl alcohol	C	4,32	C	25,1	C	11,1			C	4,41
nonyl alcohol			C	23,8	C	10,2				4,12
decyl alcohol			C	22,7	C	9,6				3,86
dodecyl alcohol			C	20,8	C	8,7				3,45
2-methoxyethanol	C	6,34	B	36,3	C	16,1 <sup>h</sup>			B	6,92
2-ethoxyethanol	C	5,91	B	32,4	C	14,4	A	55	B	6,08
isopropoxyethanol			C	29,5					C	5,77
2-methoxy-1-propanol					C	14,4 <sup>h</sup>				6,08
1-methoxy-2-propanol (propylene glycol monomethyl ether)	B	5,72	B	32,4	C	14,5 <sup>h</sup>	A	55 <sup>q</sup>	B	6,08
1-ethoxy-2-propanol (propylene glycol monoethyl ether)			C	29,6			A	68 <sup>q</sup>	C	4,96
2-butoxyethanol	B	4,76	B	28,2	C	12,0 <sup>h</sup>	A	56 <sup>q</sup>	C	7,29

Compound	Type A sampler		Type B sampler		Type C sampler		Type D sampler		Type E sampler	
	Level <sup>a</sup>	Uptake rate <sup>b</sup>	Level <sup>a</sup>	Uptake rate	Level <sup>a</sup>	Uptake rate <sup>c</sup>	Level <sup>a</sup>	Uptake rate <sup>d</sup>	Level <sup>a</sup>	Uptake rate
2,3-epoxy-1-propanol (glycidol)			C	37,1	C	16,7 <sup>h</sup>				8,15
ethylene glycol			C	37,9	C	17,4 <sup>h</sup>				4,25
ethylene glycol monohexyl ether			C	24,3	C	10,5			B	4,38
dipropylene glycol methyl ether	C	4,25	C	25,3	C	10,8 <sup>h</sup>			C	5,58
cyclohexanol	C	5,11	B	29,5	C	13,5			C	5,07
methyl cyclohexanol			C	25,3	C	12,5 <sup>h</sup>			C	5,54
benzene-1,3-diol (resorcinol)			C	25,8						4,12
terpineol			C	20,0	C	10,5 <sup>j</sup>			C	6,34
furfuryl alcohol			C	30,6					B	5,13
diacetone alcohol	C	5,05	B	28,2	C	12,4 <sup>h</sup>			B	4,18
2-butoxyethanol (butyl cellosolve acetate)			C	24,4					C	3,58
butyl carbitol acetate (diglycol butyl ether acetate)			C	21,6					C	3,58
carbitol acetate			C	23,6					C	4,03
dibutylcarbitol (diethyleneglycol dibutyl ether)			C	20,2					C	3,26
diethyleneglycol ethyl ether [carbitol]			C	26,1					C	4,4
ethyleneglycol diethyl ether (ethyl Cellosolve) (1,2-diethoxyethane)			C	28,0					C	4,88
ethylene glycol monomethyl ether (methoxy ethanol)			C	36,3						6,92
methyl alcohol (methanol)				k					C	11,64
n-octyl alcohol			C	25,1					C	4,42
<b>Ketones</b>										
acetone	B	7,87	B	40,1 <sup>e,p</sup>	A	15,2 <sup>h</sup>	A	77 <sup>q,r</sup>	A	8
2-butanone	C	6,77	A	36,3 <sup>p</sup>	B	17,1 <sup>g,h</sup>	A	57	A	6,76
2-pentanone	C	5,95	B	33,0	B	15,7 <sup>h</sup>			B	6,04
3-pentanone (diethyl ketone)			C	32,7	C	14,8 <sup>h</sup>			C	6,04
methyl isopropyl ketone			C	32,8	C	14,8 <sup>h</sup>			C	6,03
2-hexanone	C	7,10	B	29,7 <sup>p</sup>	C	13,4 <sup>h</sup>			B	5,41
3-heptanone (ethyl butyl ketone)			C	28,0	C	12,3 <sup>h</sup>			C	4,93
4-heptanone (dipropyl ketone)			C	27,8	C	12,3 <sup>h</sup>				4,93
5-methyl-2-hexanone	C	4,92	C	28,0						4,93
2,6-dimethyl heptan-4-one (diisobutyl ketone)	C	4,24	B	24,6	B	10,3 <sup>g,h</sup>			C	4,23

Compound	Type A sampler		Type B sampler		Type C sampler		Type D sampler		Type E sampler	
	Level <sup>a</sup>	Uptake rate <sup>b</sup>	Level <sup>a</sup>	Uptake rate	Level <sup>a</sup>	Uptake rate <sup>c</sup>	Level <sup>a</sup>	Uptake rate <sup>d</sup>	Level <sup>a</sup>	Uptake rate
4-methylpentan-2-one (MIBK)	B	5,27	B	30,0	B	13,5 <sup>h</sup>	A	64	B	5,4
4-methylpentan-3-ene-2-one (mesityl oxide)	C	5,70	B	31,2	C	13,7			C	5,6
methyl n-amyl ketone (2-heptanone)	C	4,82	C	27,9	C	12,2 <sup>h</sup>			B	4,94
methyl isoamyl ketone			C	28,0	C	12,2 <sup>h</sup>			C	4,93
ethyl amyl ketone (5-methyl-3-heptanone)			C	26,4	C	11,4 <sup>h</sup>			C	4,54
2,4-pentanedione			C	31,7					C	5,9
butyrolactone			C	33,7	C	15,8 <sup>h</sup>				6,98
cyclohexanone	C	6,02	B	28,9	B	15,1 <sup>j</sup>	A	60 <sup>g</sup>	B	5,58
isophorone	C	4,51	B	21,7	C	11,3 <sup>h</sup>			B	4,5
2,4-dimethyl-3-pentanone (diisopropyl ketone)			C	28,1					C	4,93
methyl cyclohexanone			C	26,5					C	5,24
<b>Ethers</b>										
diethyl ether	C	6,89	C	36,8 <sup>e</sup>	C	16,3			A	6,5
diisopropyl ether	C	5,12	C	31,2 <sup>e</sup>	C	13,2			B	5,21
dichloroethyl ether	C	5,21	C	26,1	C	12,7			C	5,38
1-dichloro-2-difluoroethyl ether	C	5,23								5,07
1,4-dioxane	C	6,90	C	34,5	C	16,0 <sup>h</sup>			B	6,62
dimethoxy methane	C	6,65	C	37,9 <sup>e</sup>	C	17,1			B	6,9
tetrahydrofuran	C	7,00	C	37,2	C	17,4 <sup>h</sup>			B	7,14
isopropyl glycidyl ether			C	29,1	C	12,8			C	5,11
butyl glycidyl ether			C	27	C	11,6			C	4,06
phenyl glycidyl ether			C	22,2	C	11,1			C	4,58
methyl t-butyl ether			A	30,8	A	13,6	B	65	C	5,77
ethyl t-butyl ether			C	29,9	B	13,1 <sup>G</sup>	B	61		5,21
methyl t-amyl ether			C	29,6	B	13,1 <sup>G</sup>				5,21
diphenyl ether	C	3,93	C	20,3	C	10,4			B	4,23
1,1-dimethoxyethane			C	33,5					C	6,07
1,2-dimethoxyethane			C	33,0					C	6,07
dimethyl ether									C	9,01
<b>Miscellaneous</b>										
acetonitrile	C	8,86	C	48,2 <sup>e</sup>	C	22,4 <sup>h</sup>			B	9,64
acrylonitrile	C	7,94	C	43,8	A	20,4 <sup>h</sup>	B	75	C	8,36
camphor	C	4,10	C	21,4	C	10,8 <sup>h</sup>			B	4,26
carbon disulfide	B	7,60	C	42,8 <sup>e</sup>					C	9,04
ethyl mercaptan			C	41,1					B	8,07
ethylene oxide	C	8,96	B	49,3 <sup>s</sup>					C	9,75
propylene oxide	C	7,42	B	37,7 <sup>e,s</sup>	C	19,9			B	7,96
furfural			C	34,3					C	6,64
morpholine			C	33,1					C	6,13
N,N-dimethylaniline					C	12,0			C	4,99
dimethylformamide			C	35,5	C	16,4 <sup>h</sup>			C	7,1

Compound	Type A sampler		Type B sampler		Type C sampler		Type D sampler		Type E sampler	
	Level <sup>a</sup>	Uptake rate <sup>b</sup>	Level <sup>a</sup>	Uptake rate	Level <sup>a</sup>	Uptake rate <sup>c</sup>	Level <sup>a</sup>	Uptake rate <sup>d</sup>	Level <sup>a</sup>	Uptake rate
dimethyl acetamide			C	32,0					C	6,22
pyridine	C	6,44	C	34,9	C	16,3			C	6,99
N-methyl-2-pyrrolidone			C	28,8					C	4,22
1,1-dichloronitroethane			C	28,5					C	5,96
1-chloro-1-nitropropane										5,96
acetic acid										8,7
dimethyl sulfoxide										7,45
hydrogen sulfide										8,03
nitrobenzene										5,68
n-propyl nitrate			C	32,7						6,6

<sup>a</sup> For an explanation of evaluation levels A to C, see clause 10.

<sup>b</sup> For type A samplers, the effective  $A/l$  ratio has been revised from 1,41 cm to 1,25 cm ( $0,80 \text{ cm}^{-1}$ ). Manufacturer data are the latest available and have been adjusted relative to published data in references [8-10].

<sup>c</sup> For type C samplers, in most cases Anasorb 747 can be used in place of charcoal with no change in the uptake rates.

<sup>d</sup> For type D samplers, the uptake rates include a correction for desorption efficiency.

<sup>e</sup> Type B sampler: with back-up section.

<sup>f</sup> Type B sampler: MDHS 74 [28].

<sup>g</sup> Type C sampler: bi-level validation [22].

<sup>h</sup> Type C sampler: with Anasorb 747.

<sup>j</sup> Type C sampler: with Anasorb 727.

<sup>k</sup> Type B sampler: not recommended for methanol or methyl chloride.

<sup>l</sup> Type D sampler: 8 h exposure was at 1/3 of ACGIH limit value.

<sup>m</sup> Type B sampler: re-calculated from data of Mazur *et al.* [29].

<sup>n</sup> Type B sampler: HSE Internal Report [30].

<sup>p</sup> Type B sampler: refrigerate and analyse as quickly as possible if sampled under high humidity.

<sup>q</sup> Type D sampler: If exposure was prolonged at >60 % R.H. add water to standards to match water content of charcoal desorbate.

<sup>r</sup> Type D sampler: If not exposed for more than 6 h at limit value.

<sup>s</sup> Type B sampler: special ethylene oxide monitor [31].

## Annex C (informative)

### Equivalence of gas chromatographic stationary phases

Company	Equivalent phase	Equivalent phase
SGE	BP-1	BP-10
Chrompack	CP-Sil 5 CB	CP-Sil 19 CB
J & W	DB-1	DB-1701
Supelco	SPB-1	SPB-1701
Hewlett-Packard	HP-1	HP-1701
Restek	Rtx-1	Rtx-1701
Quadrex	007-1	007-1701

In each table column, the phases are believed to be equivalent. For similar retention times on the two GC columns, the film thickness of the BP-10 column (or equivalent) should be about half that of the BP-1 column (or equivalent).

## Annex D (informative)

### Suppliers of charcoal-based organic vapour diffusive samplers

Arelco GABIE sampler	Arelco A.R.C. 2 avenue Ernest Renan, 94120 Fontenay-sous-Bois, France + 33 1 43 94 06 09
Assay Technology ChemDisk 541 sampler	(not available)
Dräger ORSA-5	Dräger Ltd Kitty Brewster industrial estate, Blyth, Northumberland NE24 4RG, UK + 44 1670 352891
Perkin-Elmer diffusive tube sampler (with charcoal sorbent)	Perkin-Elmer Ltd Post Office Lane, Beaconsfield, Bucks HP9 1QA, UK + 44 1494 676161
Radiello	Fondazione Salvatore Maugheri Via Svizzera, 16, 35127 Padova, Italy + 39 049 806 4511
SKC Organic vapour sampler 575	SKC Inc Valley View Road, Box 334, Eighty Four, PA 15330-9614, USA + 1-800-752-8472
3M Organic vapour monitor 3500/3520	3M Company, 3M Center, Bldg. 275-6W-01, St Paul, MN 55144-1000, USA (tel not available)

The devices in this list are examples of suitable products available commercially. This information is given for the convenience of users of this part of ISO 16200 and does not constitute endorsement by ISO of these product. Equivalent products may be used if they can be shown to lead to the same results.

## Annex E (informative)

### Specific information on sampler type A

#### E.1 Specification

The diffusive sampler type A (ORSA-5) consists of an open-ended glass tube containing granular coconut shell charcoal. Two porous plugs of cellulose acetate at both ends of the tube fix the charcoal in the tube and act as diffusion barriers during sampling. To protect the sampler from contamination during storage and transport, the tube is placed in a glass vial which is sealed by a PTFE-coated (or equivalent) screw cap. The sampler is provided with a clip holder made of polyethylene.

Glass tube	outside diameter	10 mm
	inside diameter	8 mm
	length	28 mm
Plug	diameter	8,5 mm
	length	5 mm
	pressure drop	300 Pa at 1 l/min
Charcoal	mass	400 mg
	particle size	0,4 mm to 0,8 mm

#### E.2 Sampling instructions

Immediately before sampling, remove the diffusive sampler from its transportation jar and place in the holder. At the end of a measured time of exposure, remove the sampling tube from its holder and return it to the transportation jar. Seal the jar carefully with the screw cap.

#### E.3 Desorption instructions

Remove the sampling tube from the transportation jar. Remove the porous plug from the sampler and tip the charcoal sorbent into a septum vial (5 ml to 15 ml capacity). Close the vial and add elution solvent (2 ml to 10 ml, depending on application) through the septum. Agitate the vial occasionally over a period of 30 min to ensure maximal desorption.

#### E.4 Determination of desorption efficiency

Remove the porous plug and inject known amounts of analyte with a microlitre syringe into the sorbent bed at three or more levels. Replace the porous plug and leave for at least 16 h [32].

#### E.5 Diffusive sampling rates

See annex B.

## Annex F (informative)

### Specific information on sampler type B

#### F.1 Specification

The diffusive sampler type B (3M) consists of a nylon circular body, a white polypropylene membrane and a metal clip for attachment.

Body	nominal outside diameter	30 mm
Charcoal	proprietary product	

#### F.2 Sampling instructions

Immediately before sampling, remove the diffusive sampler from its protective metal can. At the end of a measured time of exposure, remove the white membrane and retaining ring from the sampler and replace with the closure cap with the ports firmly closed. Then replace the sampler in its protective metal can for transport.

#### F.3 Desorption instructions

Open the centre port on the closure cap and pipette 1,5 ml of elution solvent into the sampler. Close the port and agitate the sampler occasionally over a period of 30 min to ensure maximal desorption.

#### F.4 Determination of desorption efficiency

With a microlitre syringe, inject known amounts of analyte into the samplers at three or more levels through one of the filling ports, seal and leave for at least 16 h [33].

#### F.5 Diffusive sampling rates

See annex B.

## Annex G (informative)

### Specific information on sampler type C

#### G.1 Specification

(Information not available)

#### G.2 Sampling instructions

Immediately before sampling, remove the diffusive sampler type C (SKC) from its protective bag. At the end of a measured time of exposure, remove the sampler and seal as follows: place the O-ring provided on the sampler face and press on the cap, ensuring the O-ring is sealed all around. Send the sampler together with the remaining accessories (including the operating instructions) to the analysing laboratory.

#### G.3 Desorption instructions

Do not remove the cap from the face of the sampler. Open the two ports at the rear using a sharp knife or other means. Slowly pipette 2,0 ml of elution solvent into the sampler. Press the plugs into place to close the ports and desorb for 1 h using an appropriate shaker. Open the ports and either remove aliquots for direct syringe injection or transfer the elution solvent to an autosampler vial through the length of PTFE tube provided (see operating instructions). Approximately 1,5 ml can be transferred by the latter method.

#### G.4 Determination of desorption efficiency

With a microlitre syringe, inject known amounts of analyte into the samplers at three or more levels through one of the filling ports, seal and leave for at least 16 h [33].

#### G.5 Diffusive sampling rates

See annex B.

## Annex H (informative)

### Specific information on sampler type D

#### H.1 Specification

The diffusive sampler type D (Radiello) consists of a sintered microporous polyethylene cylinder as a diffusive surface containing an inner coaxial stainless steel cylindrical adsorbing cartridge containing activated charcoal. The sampler is provided with a triangular polycarbonate supporting plate with hanging clip and transparent label pocket.

Outer tube:	outside diameter	16 mm
	inside diameter	12,6 mm
	length	50 mm
	pore diameter	(25 ± 5) μm
Inner tube:	outside diameter	5,9 mm
	length	60 mm
Charcoal:	mass	530 mg
	particle size	0,5 mm to 0,7 mm

#### H.2 Sampling instructions

Immediately before sampling, remove the sorbent cartridge from its glass storage tube and insert in the diffusive body, making sure it is aligned centrally. Attach the diffusive body to the support plate. At the end of a measured time period of exposure, remove the sorbent cartridge, replace in its glass storage tube and cap.

#### H.3 Desorption instructions

Remove the cap from a clean glass storage tube and pipette in 2,0 ml elution solvent. Insert the sorbent cartridge in the solvent and leave for 30 min to 60 min, occasionally agitating to ensure maximal desorption.

#### H.4 Determination of desorption efficiency

Take an unused blank sorbent cartridge in its glass storage tube and replace the polyethylene cap with a silicone septum. With a microlitre syringe, inject known amounts of analyte into the glass storage tube at three or more levels and leave for at least 16 h. Alternatively, if suitable septa are not available, the phase equilibrium method may be acceptable, subject to limitations described in references [4] to [7].

#### H.5 Diffusive sampling rates

See annex B.

## Annex I (informative)

### Specific information on sampler type E

#### I.1 Specification

The diffusive sampler type E (Assay Technology) consists of a circular plastics housing incorporating a charcoal wafer encased within a wafer dish which snaps into a sampling grid containing a multiplicity of cone-shaped sampling ports. The number of sampling ports determines the sampler's uptake rate. When not in use, sampling ports are covered by an impermeable sampler cap which snaps over the face of the sampling grid.

Housing:            diameter        31 mm  
                         thickness        7 mm

Sampling ports:    1 mm x 3 mm nominal

Charcoal:            proprietary product

#### I.2 Sampling instructions

No special requirements.

#### I.3 Desorption instructions

Remove monitor from the plastics return container. Pry with spatula to separate the sampling grid from the clear plastics wafer tray, exposing the carbon sampling wafer. Quickly transfer the carbon sampling wafer to weighing paper which has been creased in the middle, grasp the wafer through the weighing paper, break the wafer in two along the score mark, and pour the wafer pieces into a glass desorption vial. Pipette 2,0 ml of desorbing solution and cap the vial with an inert, gas-tight closure. Agitate the vial continuously for 1 h using an orbital shaker or equivalent. Reserve for gas chromatographic analysis.

#### I.4 Determination of desorption efficiency

Break an unused carbon sampling wafer along the score mark and place in a glass desorption vial. Inject a known amount of analyte onto the wafer surface using a microsyringe, cap the vial, and hold overnight. Repeat with several different quantities of analyte representing expected air concentrations. Add desorbing solution to the vial, and proceed with analysis.

#### I.5 Diffusive sampling rates

See annex B.