
**Respiratory protective devices —
Methods of test and test equipment —**

**Part 14:
Measurement of sound level**

*Appareils de protection respiratoire — Méthodes d'essai et
équipement d'essai —*

Partie 14: Mesurage du niveau sonore

STANDARDSISO.COM : Click to view the full PDF of ISO 16900-14:2015



STANDARDSISO.COM : Click to view the full PDF of ISO 16900-14:2015



COPYRIGHT PROTECTED DOCUMENT

© ISO 2015, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

	Page
Foreword.....	iv
Introduction.....	vi
1 Scope.....	1
2 Normative references.....	1
3 Terms and definitions.....	1
4 Prerequisites.....	1
5 General test requirements.....	2
6 Method A — Measurement of the noise level.....	2
6.1 Principle.....	2
6.2 Equipment.....	2
6.3 Procedure.....	2
6.4 Test report for method A.....	3
7 Method B — Measurement of warning sound level.....	3
7.1 Principle.....	3
7.2 Equipment.....	4
7.3 Procedure.....	4
7.4 Test report for method B.....	4
Annex A (normative) Application of uncertainty of measurement.....	5
Bibliography.....	7

STANDARDSISO.COM : Click to view the full PDF of ISO 16900-14:2015

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#).

The committee responsible for this document is ISO/TC 94, *Personal safety — Protective clothing and equipment*, Subcommittee SC 15, *Respiratory protective devices*.

ISO 16900 consists of the following parts, under the general title *Respiratory protective devices — Methods of test and test equipment*:

- *Part 1: Determination of inward leakage*
- *Part 2: Determination of breathing resistance*
- *Part 3: Determination of particle filter penetration*
- *Part 4: Determination of gas filter capacity and migration, desorption and carbon monoxide dynamic testing*
- *Part 5: Breathing machine, metabolic simulator, RPD headforms and torso, tools and verification tools*
- *Part 6: Mechanical resistance/strength of components and connections*
- *Part 7: Practical performance tests methods*
- *Part 8: Measurement of RPD air flow rates of assisted filtering RPD*
- *Part 9: Determination of carbon dioxide content of the inhaled air*
- *Part 10: Resistance to ignition, flame, radiant heat and heat*
- *Part 11: Determination of field of vision*
- *Part 12: Determination of volume-averaged work of breathing and peak respiratory pressures*
- *Part 13: RPD using regenerated breathable gas and special application mining escape RPD: Consolidated test for gas concentration, temperature, humidity, work of breathing, breathing resistance, elastance and duration*

— *Part 14: Measurement of sound level*

STANDARDSISO.COM : Click to view the full PDF of ISO 16900-14:2015

Introduction

This part of ISO 16900 is intended as a supplement to the respiratory protective devices (RPD) performance standards. Test methods are specified for complete devices or parts of devices. If deviations from the test method given in this part of ISO 16900 are necessary, these deviations will be specified in the performance standards.

The following definitions apply in understanding how to implement an ISO International Standard and other normative ISO deliverables (TS, PAS, IWA):

- “shall” indicates a requirement;
- “should” indicates a recommendation;
- “may” is used to indicate that something is permitted;
- “can” is used to indicate that something is possible, for example, that an organization or individual is able to do something.

3.3.1 of the ISO/IEC Directives, Part 2 (sixth edition, 2011) defines a requirement as an “expression in the content of a document conveying criteria to be fulfilled if compliance with the document is to be claimed and from which no deviation is permitted.”

3.3.2 of the ISO/IEC Directives, Part 2 (sixth edition, 2011) defines a recommendation as an “expression in the content of a document conveying that among several possibilities one is recommended as particularly suitable, without mentioning or excluding others, or that a certain course of action is preferred but not necessarily required, or that (in the negative form) a certain possibility or course of action is deprecated but not prohibited.”

STANDARDSISO.COM : Click to view the full PDF of ISO 16900-14:2015

Respiratory protective devices — Methods of test and test equipment —

Part 14: Measurement of sound level

1 Scope

This part of ISO 16900 specifies a laboratory test method for determining the sound level generated by the complete respiratory protective device (RPD) and RPD warning sounds measured on a headform to which the RPD is fitted.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 16972, *Respiratory protective devices — Terms, definitions, graphical symbols and units of measurement*

IEC 61260-1, *Octave-band and fractional-octave-band filters — Part 1: Specifications*

IEC 61672-1, *Electroacoustics — Sound level meters — Part 1: Specifications*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 16972 apply.

3.1

A-weighted equivalent continuous sound pressure level

$L_{p,A,eqT}$

10 times the logarithm to the base 10 of the ratio of the time average of the square of the A-weighted sound pressure, p_A , during a stated time interval of duration T (starting at t_1 and ending at t_2) to the square of the reference value, p_0 , expressed in decibels

$$\text{Note 1 to entry: } L_{p,A,eqT} = 10 \lg \left[\frac{\frac{1}{T} \int_{t_2}^{t_1} p_A^2(t) dt}{p_0^2} \right] \text{ dB, where the reference value, } p_0, \text{ is } 20 \mu\text{Pa.}$$

[SOURCE: ISO 9612:2009 [1], 3.1 modified]

4 Prerequisites

In order to implement this part of ISO 16900, at least the following parameters need to be specified in the relevant device standard:

- a) the number of RPD specimens;
- b) any prior conditioning or testing;
- c) the RPD operating conditions;

- d) test duration;
- e) any deviations from the method(s).

5 General test requirements

Unless otherwise specified, the values stated in this part of ISO 16900 are expressed as nominal values. Except for temperature limits, values which are not stated as maxima or minima shall be subject to a tolerance of $\pm 5\%$. Unless otherwise specified, the ambient conditions for testing shall be between $16\text{ }^{\circ}\text{C}$ and $32\text{ }^{\circ}\text{C}$ and $(50 \pm 30)\%$ RH. Any temperature limits specified shall be subject to an accuracy of $\pm 1\text{ }^{\circ}\text{C}$.

Where the assessment of the pass/fail criterion depends on a measurement, an uncertainty of measurement as specified in [Annex A](#) shall be reported.

6 Method A — Measurement of the noise level

6.1 Principle

The RPD is operated on a RPD headform and torso connected to a breathing machine. Microphones are placed to measure the A-weighted continuous equivalent sound pressure level generated by the device, exclusive of any warning device, at the right and left ears.

6.2 Equipment

- a) The required RPD headform with soft surface finish shall be used. Additionally, RPD torso can be used where necessary. An omnidirectional microphone that has a profile of no more than 5 mm above the surface of the RPD headform when fitted in the test ear position inside the RPD shall be used, see [Figure 1](#).
- b) An instrument that meets the IEC 61672-1 Class 2 standard for sound level meters, capable of measuring the integrated A-weighted sound pressure level in decibels, indicated as dBA.
- c) A suitable associated sound calibrator for the microphone system and sound level meter.
- d) A breathing machine or metabolic simulator as appropriate.

The test environment should not include surfaces that may reflect sound from the specimen.


6.3 Procedure

- a) The sound level meter and microphone system shall be calibrated using the associated sound calibrator.
- b) The microphone shall be placed so that it faces outwards and lies flat on the external ear (pinnae) of the RPD headform at the position shown in [Figure 1](#). This position corresponds to the centre of the external ear and level with the tragus. The microphones shall be held in place so that they do not move during the test. To avoid sound reflections caused by the measurement equipment, any microphone supporting elements, electrical leads, etc. within 30 mm of the measurement point shall occupy an area not exceeding 5 mm^2 in any plane above the surface of the RPD headform. It shall be ensured that any electrical leads are held in place flat to the surface of the RPD headform.

NOTE Putty or sticky-tack has been successfully used to stick the microphone to the RPD headform.

If part of the microphone is inset into the RPD headform, the sensing element of the microphone shall not be below the surface of the RPD headform and the opening shall be sealed against the side of the microphone.

- c) The RPD shall be fitted on the RPD headform and torso, making all the necessary adjustments to ensure a good fit with reference to the manufacturer's instructions for use. The RPD shall not make contact with the microphone assembly.
- d) The RPD shall be operated as specified by the manufacturer at the manufacturer's maximum flow conditions.
- e) Operate the breathing machine or metabolic simulator at the appropriate workrate for the class of the RPD.
- f) The $L_{p,A,eqT}$ (A-weighted equivalent continuous sound pressure level) at both ear positions shall be measured. Measurements shall be made over the period of time as required by the performance standard during continuous operation of the RPD.
- g) The background noise level in the test room, measured without the RPD fitted to the RPD headform, but with the breathing machine operating (when used), shall be a maximum of 70 dBA.
- h) The measurements shall be repeated three times, refitting the RPD on the RPD headform and torso in between each measurement.



Key

- 1 required RPD headform
- 2 microphone position in flattened pinna
- 3 sound measurement device

Figure 1 — Typical arrangement of the sound level measurement

6.4 Test report for method A

The test report shall include information regarding those parameters specified in [Clause 4](#), together with the background sound pressure level and the maximum measured $L_{p,A,eqT}$ value. The uncertainty of measurement shall be reported.

7 Method B — Measurement of warning sound level

7.1 Principle

The RPD is operated on a RPD headform and torso connected to a breathing machine. Microphones are placed to measure the sound spectrum generated by the RPD warning device at the right and left ears.

7.2 Equipment

- a) The required RPD headform and torso shall be used. An omnidirectional microphone that has a profile of no more than 5 mm above the surface of the RPD headform when fitted in the test ear position inside the RPD shall be used.
- b) A sound spectrum analyser specified as conforming to IEC 61260-1.
- c) A suitable associated sound calibrator for the microphone system and sound spectrum analyser.
The test environment should not include surfaces that can reflect sound from the specimen.
- d) A breathing machine or metabolic simulator as appropriate.

7.3 Procedure

- a) The sound level meter and microphone system shall be calibrated using the associated sound calibrator.

The microphone shall be placed so that it faces outwards and lies flat on the external ear (pinnae) at the position shown in [Figure 1](#). This position corresponds to the centre of the external ear and level with the tragus. The microphones shall be held in place so that they do not move during the test. To avoid sound reflections caused by the measurement equipment, any microphone supporting elements, electrical leads, etc. within 30 mm of the measurement point shall occupy an area not exceeding 5 mm² in any plane above the surface of the RPD headform. It shall be ensured that any electrical leads are held in place flat to the surface of the RPD headform.

NOTE Putty or sticky-tack has been successfully used to stick the microphone to the RPD headform.

If part of the microphone is inset into the RPD headform, the sensing element of the microphone shall not be below the surface of the head and the opening shall be sealed against the side of the microphone.

- b) The RPD shall be fitted on the RPD headform and torso, making all the necessary adjustments to ensure a good fit with reference to the manufacturer's instructions for use.
- c) The RPD shall be operated as specified by the manufacturer at the workrate for its class.
- d) The sound level at the octave bands specified by the performance standard shall be measured during the operation of the warning device at both left and right ears. Measurements shall be made over a period as required by the performance standard. While the RPD is functioning, the background sound level in the test room, at the octave bands specified by the performance standard, shall be a maximum of 80 dBC.
- e) The measurements shall be repeated three times, refitting the RPD on the RPD headform and torso in between each measurement.

7.4 Test report for method B

The test report shall include information regarding those parameters specified in [Clause 4](#), together with the background sound level and the minimum measured sound levels at the specified octave band generated by the warning device at the specified conditions. The uncertainty of measurement shall be reported.


Annex A (normative)

Application of uncertainty of measurement

A.1 Determination of compliance

In order to determine compliance or otherwise of the measurement made in accordance with this test method, when compared to the specification limits given in the performance standard, the following protocol shall be applied.

If the test result \pm the uncertainty of measurement, U , falls completely inside or outside of the specification zone for the particular test given in the performance standard, then the result shall be deemed to be a straightforward pass or fail (see [Figures A.1](#) and [A.2](#)).



Key

- 1 lower specification limit
- 2 specification zone
- 3 upper specification limit
- 4 uncertainty of measurement, U
- 5 measured value

Figure A.1 — Result pass