

---

---

**Solid mineral fuels — Determination of  
sulfur by IR spectrometry**

*Combustibles minéraux solides — Détermination du soufre par  
spectrométrie infrarouge*

STANDARDSISO.COM : Click to view the full PDF of ISO 19579:2006



**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

STANDARDSISO.COM : Click to view the full PDF of ISO 19579:2006

© ISO 2006

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland

**Contents**

Page

Foreword.....	iv
Introduction .....	v
1 Scope .....	1
2 Normative references .....	1
3 Principle .....	1
4 Apparatus .....	1
5 Reagents .....	2
6 Preparation of sample .....	2
7 Procedure .....	3
8 Calibration check .....	3
9 Expression of results .....	3
10 Precision .....	3
11 Test report .....	4

STANDARDSISO.COM : Click to view the full PDF of ISO 19579:2006

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 19579 was prepared by Technical Committee ISO/TC 27, *Solid mineral fuels*, Subcommittee SC 5, *Methods of analysis*.

STANDARDSISO.COM : Click to view the full PDF of ISO 19579:2006

## Introduction

Sulfur is normally present in coal in three forms; inorganic sulfides such as pyrite ( $\text{FeS}_2$ ), inorganic sulfates associated with the mineral matter and organic sulfur in the carbonaceous substance.

This International Standard describes a high-temperature combustion/infrared absorption method of analysis, which is used to determine the total sulfur content of coal.

STANDARDSISO.COM : Click to view the full PDF of ISO 19579:2006



# Solid mineral fuels — Determination of sulfur by IR spectrometry

## 1 Scope

This International Standard specifies an alternative method of determining the total sulfur content of hard coal, brown coal, and lignite by high-temperature combustion and infrared (IR) absorption using commercially available instruments.

This method has been shown to be applicable to coal samples having an ash yield of less than 40 %.

## 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 334, *Solid mineral fuels — Determination of total sulfur — Eschka method*

ISO 351, *Solid mineral fuels — Determination of total sulfur — High temperature combustion method*

ISO 687, *Solid mineral fuels — Coke — Determination of moisture in the general analysis test sample*

ISO 5068-2, *Brown coals and lignites — Determination of moisture content — Part 2: Indirect gravimetric method for moisture in the analysis sample*

ISO 11722, *Solid mineral fuels — Hard coal — Determination of moisture in the general analysis test sample by drying in nitrogen*

## 3 Principle

The coal sample is combusted at 1 350 °C in a stream of oxygen. Particulates and water vapour are removed from the gas stream by traps of glass wool and magnesium perchlorate. The gas stream then passes through a cell in which the sulfur dioxide is measured by an infrared absorption detector, connected to a microprocessor. The instrument is calibrated with standard reference materials. The percentage of sulfur in the sample is calculated from this prior calibration by the microprocessor.

## 4 Apparatus

**4.1 Instrument**, commercially available, consisting of the following:

- a) resistance furnace, capable of maintaining a temperature of approximately 1 350 °C in the combustion zone;
- b) combustion tube, of ceramic material, to contain the sample and combustion gases;

- c) combustion boats, of ceramic material, into which the sample is weighed and then combusted;
- d) combustion gas-purification train, through which the combustion gases pass from the combustion tube to the detector (the train consists of a glass wool filter and a trap containing anhydrous magnesium perchlorate);
- e) detector/microprocessor, a non-dispersion infrared measurement system.

4.2 **Balance**, capable of weighing to 1 mg.

## 5 Reagents

5.1 **Magnesium perchlorate**, anhydrous granular.

5.2 **Oxygen**, compressed industrial grade, 99,5 % by volume.

5.3 **Glass wool**.

5.4 **Reference coals or cokes**.

Certified reference materials (CRMs) are coal or coke samples of known sulfur concentration analysed in accordance with ISO 334 or ISO 351. The reference materials shall be traceable to an internationally recognized certifying organization.

NOTE 1 Calibration with coals or cokes that are similar in composition to the analysis samples is necessary. It has been noted that some coals containing high levels of carbonates and some low-rank coals containing high levels of carboxylic salts of calcium can require longer combustion/integration times, as the release of sulfur dioxide can be delayed during combustion.

If coal or coke samples (other than CRMs) are used for calibrations, then it is necessary to verify the calibration. This is done by analysing an appropriate CRM after calibration. If the results of the analysis of the CRM do not agree with the certified value of the sulfur content, the instrument should be recalibrated.

NOTE 2 Reference materials including CRMs can contain moisture and are usually certified on a dry basis. It is necessary to determine the moisture content in accordance with ISO 11722, ISO 687 and ISO 5068-2 as appropriate. This moisture is used to calculate the dry-basis sulfur value to an "as analysed" value,  $w_{S,aa}$ , expressed in percent, as given in Equation (1):

$$w_{S,aa} = w_{S,d,ac} \times (100 - w_{H_2O})/100 \quad (1)$$

where

$w_{S,d,ac}$  is the dry-basis sulfur content, expressed in mass percent;

$w_{H_2O}$  is the moisture content, expressed in mass percent.

## 6 Preparation of sample

The coal or coke used for the determination of total sulfur content is the analysis sample ground to pass a sieve of 212  $\mu\text{m}$  aperture.

Expose the sample in a thin layer for the minimum time required for the moisture content to reach approximate equilibrium with the laboratory atmosphere. Before commencing the determination, mix the air-dried sample.

After weighing the test portion (see Clause 7), determine the moisture content using a further portion of the test sample by the method described in ISO 11722, ISO 687 or ISO 5068-2, as appropriate.

## 7 Procedure

Prepare the instrument according to the manufacturer's instructions. It is advisable to process two samples to condition the instrument prior to calibration. Calibrate the instrument using a reference coal or coke sample or reference coal or coke samples if a multi-point calibration is possible. The calibration should cover the range of sulfur concentrations below and above that present in the samples.

Use the calibration procedure recommended by the instrument manufacturer. Accurately weigh, to 0,001 g, 0,2 g to 0,5 g of the reference coal or coke into the combustion boat and enter the weight into the processor (computer). Push the boat plus sample into the hot zone of the furnace. The combustion products are pumped through the traps to the IR detector, where the absorption signal due to sulfur dioxide is integrated. At completion of the burn (indicated by the instrument), remove the boat containing the combustion residue.

**NOTE** Some instruments are designed for automatic operation and manual sample introduction and removal is not necessary.

Repeat the process with a further two replicates of the reference coal or coke. Check the accuracy of the calibration by analysing the calibrating reference coal or coke and another reference coal or coke of lower sulfur concentration as unknown samples. If the results obtained for the reference coals or cokes are not within the repeatability of the method (see 10.1), repeat the calibration procedure.

Determine the sulfur concentration in the sample by the same procedure as used for the reference coals or cokes.

## 8 Calibration check

A calibration check shall be completed periodically (it is recommended that this be after every five determinations). A standard of known sulfur content, not necessarily a CRM, shall be analysed.

If the result does not agree with the concentration of sulfur in the standard within the repeatability of the method, the preceding results shall be discarded and the calibration repeated (Clause 7) and the samples reanalysed.

## 9 Expression of results

The sulfur content of the sample as analysed is calculated by the microprocessor and expressed as a mass percent. The result, the mean of duplicate determinations, shall be reported to the nearest 0,01 mass %.

## 10 Precision

### 10.1 Repeatability

The results of duplicate determinations carried out at different times in the same laboratory, by the same operator, with the same apparatus, on representative portions taken from the same analysis sample, shall not differ by more than the value of the repeatability,  $r$ , as given by Equation (2):

$$r = 0,02 + 0,03 \bar{x} \quad (2)$$

where  $\bar{x}$  is the mean of the intra-laboratory results.