
**Traditional Chinese medicine — Skin
electrical resistance measurement
devices**

*Médecine traditionnelle chinoise — Détecteurs de résistance
électrique aux points d'acupuncture*

STANDARDSISO.COM : Click to view the full PDF of ISO 20495:2018



STANDARDSISO.COM : Click to view the full PDF of ISO 20495:2018



COPYRIGHT PROTECTED DOCUMENT

© ISO 2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Technical requirements	2
4.1 Measurable range.....	2
4.2 Accuracy of measurement.....	2
4.3 Detection electrical signal.....	3
4.4 Electrode force indication device.....	3
4.5 Force sensing switch of exploratory electrode.....	3
4.6 Detection voltage.....	4
4.7 Detection current.....	4
4.8 Effective area of auxiliary electrode.....	4
4.9 Electrical safety.....	4
4.10 Electromagnetic compatibility.....	4
4.11 Biological evaluation.....	4
Annex A (informative) Particular guidance and rationale	6
Bibliography	9

STANDARDSISO.COM : Click to view the full PDF of ISO 20495:2018

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 249, *Traditional Chinese medicine*.

Traditional Chinese medicine — Skin electrical resistance measurement devices

1 Scope

This document specifies performance and test methods for skin electrical resistance measurement devices.

This document is applicable to skin electrical resistance measurement devices through which an electrical signal is applied on acupuncture points of the human body or specific points of the body surface in a non-invasive way so as to detect the resistance of the body.

This document is not applicable to software or functions for medical analysis of the test results.

Additional guidance and rationale are given in [Annex A](#).

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 10993-1, *Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process*

IEC 60601-1, *Medical electric equipment — Part 1: General requirements for basic safety and essential performance*

IEC 60601-1-2, *Medical electric equipment — Part 1-2: General requirements for basic safety and essential performance — Collateral standard: Electromagnetic compatibility — Requirements and tests*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <http://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1

skin electrical resistance

direct current (DC) resistance value or alternating current (AC) resistance value between acupuncture points on the skin surface of the patient (or specific points) and a reference point of the skin surface (point on the skin in contact with auxiliary electrode)

3.2

skin electrical resistance measurement device

device that can detect electrical resistance on acupuncture points or specific points of the body surface under the condition of electrical signals being applied to the human body

3.3

detection electrical signal

voltage signal or current signal on device or for match use applied to patient for electrical resistance detection

3.4

auxiliary electrode

electrode that is fixed on specific points of the patient (e.g. hand-held or fixed on palm or forehead) for reference use, which is relatively fixed during detection

3.5

exploratory electrode

electrode that directly contacts acupuncture points or specific points of the patient's body surface, constituting a pair of electrodes with an auxiliary electrode for resistance detection

3.6

detection voltage

root mean square (RMS) of maximum open-circuit voltage between the auxiliary electrode and the exploratory electrode in the detection electrical signal

3.7

detection current

RMS of maximum current between the auxiliary electrode and the exploratory electrode in detection of the electrical signal

3.8

electrode force indication device

device on the exploratory electrode or auxiliary electrode whose function is to control or display the applied force during detection, including an exploratory electrode with force switch

3.9

force sensing switch of exploratory electrode

switch mounted in the exploratory electrode to reduce the error of detection result during use

3.10

effective dimension of electrode

accessible part that has electrical energy transmitting to the patient during detection of auxiliary electrode and exploratory electrode.

4 Technical requirements

4.1 Measurable range

The skin electrical resistance measurement device shall be capable of measuring skin electrical resistance between 100 Ω and 10 K Ω .

Conformity inspection of technical documentation is provided by the manufacturer.

For the accuracy of measurable value, see [A.3](#).

4.2 Accuracy of measurement

The error between any measured value and actual value shall be no more than $\pm 10\%$. The measured value may be expressed as electrical resistance (Ω), voltage (V), current (A) or a relative value.

Conformity is determined by carrying out the following test.

Within the measuring range of skin electrical resistance specified by manufacturers, choose five measuring values including boundary values for detection. For a device with multiple measuring

ranges, each measuring range is tested separately. Choose five measuring values including boundary value for detection specified by manufacturers within each measuring range.

Measure the resistance on the equivalent circuit of five measuring values for more than five times, non-reactive 1 % resistor corresponding as close as possible to each of the measuring points selected. Each measured value shall be within ± 10 % of the stated value of the selected resistor.

4.3 Detection electrical signal

For either a detection voltage or detection current adjustable device, the error between the set value and actual value shall be no more than ± 10 %.

Conformity is determined by carrying out the following test.

According to the technical documentation provided by the manufacturer, the output voltage value or the current value of detection electrical signal is set separately for each measuring range. The actual open circuit output voltage value of the device is detected or the actual short-circuit output current value is measured. The error percentage for each measuring range is calculated using [Formulae \(1\)](#) and [\(2\)](#):

$$V_{\text{error}} = \frac{V_{\text{actual}} - V_{\text{set}}}{V_{\text{set}}} \times 100 \% \quad (1)$$

$$I_{\text{error}} = \frac{I_{\text{actual}} - I_{\text{set}}}{I_{\text{set}}} \times 100 \% \quad (2)$$

where

V_{error} is the voltage error between set value and actual value;

V_{actual} is the actual voltage value of applied voltage;

V_{set} is the set value of applied voltage;

I_{error} is the current error between set value and actual value;

I_{actual} is the actual current value of applied current;

I_{set} is the set value of applied current.

The error percentage (V_{error} or I_{error}) shall not exceed ± 10 %.

4.4 Electrode force indication device

For an auxiliary electrode or exploratory electrode with force indication, the error between the indicated value and the actual value shall be no more than ± 10 %.

Conformity is checked by applying a force generating device for detection. The error between the actual value and the set value shall be no more than ± 10 %.

4.5 Force sensing switch of exploratory electrode

The repeatability error when activating the force sensing switch shall be no more than ± 10 %.

Conformity is determined by carrying out the following test.

Gradually apply an increasing force until the force sensing switch is activated. Record the applied force value. Repeat the procedure 10 times.

Identify the maximum applied force value (F_{\max}) and the minimum applied force value (F_{\min}) of the 10 measurements and calculate the mean (F_n) of the 10 measurements. Calculate the repeatable error (R_{error}) using [Formula \(3\)](#):

$$R_{\text{error}} = \frac{F_{\max} - F_{\min}}{F_n} \times 100 \% \quad (3)$$

where

F_{\max} is the maximum force value of the ten measurements;

F_{\min} is the minimum force value of the ten measurements;

F_n is the mean of ten measurements.

4.6 Detection voltage

The root mean square (RMS) value of the detection voltage shall not exceed 24 V.

Conformity is determined by carrying out the following test.

Adjust the detection voltage to its maximum value. Measure the RMS value of the open-circuit output voltage between the auxiliary electrode and the exploratory electrode.

The measured value shall not exceed 24 V.

4.7 Detection current

The RMS value of the detection current shall not exceed 0,5 mA.

Conformity is determined by carrying out the following test.

Adjust the detection current to its maximum value. Measure the RMS value of the short-circuit current between the auxiliary electrode and the exploratory electrode. For a device with short-circuit protection, the detection current is the RMS value of the maximum current that does not activate (trip) the short-circuit protection.

The measured value shall not exceed 0,5 mA RMS.

4.8 Effective area of auxiliary electrode

The effective area of the auxiliary electrode shall be greater than or equal to 300 mm².

4.9 Electrical safety

The skin electrical resistance measurement device shall meet the applicable electrical safety requirements of IEC 60601-1.

4.10 Electromagnetic compatibility

The skin electrical resistance measurement device shall meet the applicable requirements of IEC 60601-1-2.

4.11 Biological evaluation

Conformity shall be confirmed if records provided by the manufacturer establish that the biocompatibility of the materials selected has been demonstrated. The biocompatibility of the contact part of the auxiliary electrode and the exploratory electrode intended to come into contact with the patient shall be assessed and documented according to the guidance and principles given in ISO 10993-1.

Conformity is determined by:

- a) analogy with published data, or
- b) selection of materials intended to come into contact with patient already shown to be biocompatible by proven clinical use in a similar application, or
- c) experience with similar devices already on the market, together with evidence of traceability to the materials used in those devices, or
- d) conformity with published procedures for biological evaluation of medical devices.

STANDARDSISO.COM : Click to view the full PDF of ISO 20495:2018

Annex A (informative)

Particular guidance and rationale

A.1 General guidance


It has been found that resistance on acupuncture points and other specific points of the body can reflect the health status of the human body based on the theory of TCM and clinical practice. A skin electrical resistance measurement device is a device through which an electrical signal is applied on acupuncture points of the human body or specific points of the body surface in a non-invasive way so as to detect resistance of body.

Acupuncture points conform to parts of the human body defined in *WHO Standardization of Acupuncture Nomenclature* published by WHO Western Pacific Region in 1984^[1]. Other specific points of the body surface refer to Ryodo points which are not the same as acupuncture points and specific points of medicinal value according to doctors.

A.2 Skin electrical resistance measurement device

This subclause describes a model of skin electrical resistance measurement devices. The model can be seen in [Figure A.1](#).

STANDARDSISO.COM : Click to view the full PDF of ISO 20495:2018

**Key**

- 1 output device
- 2 electrical signal source
- 3 exploratory electrode
- 4 auxiliary electrode

NOTE 1 The output device can display the detection value. It can be included in the device and also in software.

NOTE 2 An electrical signal source can provide electrical energy for detection, usually, AC source or DC source.

Figure A.1 — Skin electrical resistance measurement device

A.3 Accuracy of measurable value

The measurable range of electrical resistance specified in this document is based on a general measurable range applied widely in similar clinical devices. This document only specifies the minimum measurable range.

During detection of the accuracy of measurable value, choosing three values besides boundary values considers the linearity of the device. This document does not specify the choice of three values besides boundary values with the purpose of reducing the customized cost of non-reactive load resistance. The three values are evenly chosen within the range of boundary as much as possible.

A.4 Electrode force indication device

During detection of electrical resistance, the force applied on the exploratory electrode and auxiliary electrode will affect the result of detection. Electrode force indication displays the force applied on the exploratory electrode and auxiliary electrode.