

---

---

**Prosthetics and orthotics —  
Foot orthotics — Uses, functions  
classification and description**

*Prothèses et orthèses — Orthèses de pied — Utilisations, fonctions,  
classification et description*

STANDARDSISO.COM : Click to view the full PDF of ISO 21064:2017



STANDARDSISO.COM : Click to view the full PDF of ISO 21064:2017



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2017, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Ch. de Blandonnet 8 • CP 401  
CH-1214 Vernier, Geneva, Switzerland  
Tel. +41 22 749 01 11  
Fax +41 22 749 09 47  
copyright@iso.org  
www.iso.org

# Contents

	Page
<b>Foreword</b> .....	<b>iv</b>
<b>Introduction</b> .....	<b>v</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>1</b>
<b>4 Uses</b> .....	<b>2</b>
<b>5 Functions</b> .....	<b>2</b>
<b>6 Classification and description of devices</b> .....	<b>2</b>
6.1 Classification.....	2
6.2 Description of foot orthoses (used in conjunction with footwear).....	3
6.3 Description of orthotic footwear.....	3
6.3.1 Type of fabrication.....	3
6.3.2 Prefabricated orthotic footwear.....	3
6.3.3 Custom fabricated orthotic footwear.....	4
6.4 Description of adaptations (modifications) to footwear.....	4
<b>Bibliography</b> .....	<b>5</b>

STANDARDSISO.COM : Click to view the full PDF of ISO 21064:2017

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: Foreword - Supplementary information

This document was prepared by Technical Committee ISO/TC 168, *Prosthetics and orthotics*.

## Introduction

No internationally accepted method is available to describe devices applied to the foot to address neuromusculoskeletal impairments.

This situation has created problems for practitioners prescribing these devices, for manufacturers when describing their products and for those reporting on the devices used in the treatment of their patients and their effects.

This document proposes the adoption of the term “foot orthotics” for this field of orthotic practice. It also specifies the clinical objectives for the use of these devices and describes their functions and constituent components.

Manufacturers’ trade names and details of materials and manufacturing processes have been avoided.

STANDARDSISO.COM : Click to view the full PDF of ISO 21064:2017

[STANDARDSISO.COM](https://standardsiso.com) : Click to view the full PDF of ISO 21064:2017

# Prosthetics and orthotics — Foot orthotics — Uses, functions classification and description

## 1 Scope

This document establishes a method of classifying and describing the devices that are used in the field of foot orthotics. It does not describe the materials or manufacturing methods used for their fabrication.

## 2 Normative references

There are no normative references in this document.

## 3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <http://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

### 3.1

#### **foot orthoses**

orthoses that encompass the whole or part of the foot intended to address the effect of a neuromusculoskeletal impairment(s)

Note 1 to entry: They can be custom fabricated or prefabricated.

Note 2 to entry: This definition supersedes the definition of “foot orthoses” in ISO 8549-3.

### 3.2

#### **orthotic footwear**

footwear intended to address the effect of a neuromusculoskeletal impairment(s)

Note 1 to entry: These can encompass the ankle joint. These devices are sometimes referred to as “orthopaedic footwear”. They can be custom fabricated or prefabricated.

#### 3.2.1

##### **custom fabricated orthotic footwear**

orthotic footwear designed and manufactured to meet the clinical needs of the individual user based on information such as moulds, lasts and measurements

Note 1 to entry: They can include a custom fabricated foot orthosis.

#### 3.2.2

##### **prefabricated orthotic footwear**

orthotic footwear designed and manufactured to meet the clinical needs of a particular group of users and which may be available in a range of sizes, styles and materials

Note 1 to entry: They can require adjustments and can include a foot orthosis in order to meet the clinical needs of the individual user.

### 3.3

#### **adaptations (modifications) to footwear**

adaptations (modifications) to either standard footwear or prefabricated orthotic footwear intended to meet the clinical needs of the individual user

## 4 Uses

Foot orthotic devices may be used to achieve the following clinical objectives:

- a) to relieve pain;
- b) to promote balanced standing and walking;
- c) to manage deformities which are:
  - 1) reducible (flexible) (e.g. pes planus);
  - 2) irreducible (fixed) (e.g. fixed equinus);
- d) to control joint instability (e.g. ankle ligament insufficiency);
- e) to compensate for abnormal limb length and/or foot length or shape;
- f) to protect tissues (e.g. diabetic neuropathy and inflammatory arthropathy);
- g) to promote healing (e.g. foot ulcer);
- h) to stimulate motor-sensory feedback.

## 5 Functions

The function of the device may be:

- a) for deformities:
  - 1) to reduce or correct a deformity (i.e. to move a reducible (flexible) deformity into an improved alignment and hold the correction obtained);
  - 2) to hold a deformity (i.e. to prevent a fixed deformity from increasing);
- b) to control the alignment of a joint;
- c) to control the motion of a joint;
- d) to compensate for abnormal limb length and/or foot shape and size;
- e) to reduce or redistribute the load on the foot;
- f) to apply increased pressure on a specific area(s) of the foot surface.

## 6 Classification and description of devices

### 6.1 Classification

Three classes of device are recognised:

- a) foot orthoses;
- b) orthotic footwear;
- c) adaptations (modifications) to footwear.

## 6.2 Description of foot orthoses (used in conjunction with footwear)

Types of foot orthoses include:

- a) heel raises;
- b) heel cups/shells (which may be extended);
- c) foot orthoses with:
  - 1) flares;
  - 2) supramalleolar extensions;
  - 3) medial/longitudinal arch supports;
  - 4) metatarsal supports;
  - 5) stiffening;
  - 6) cushioning;
  - 7) toe supports, spacers and protectors;
  - 8) medial and lateral wedges;
  - 9) local pressure relief;
  - 10) increased local pressure shaping;
  - 11) toe extension shaping.

## 6.3 Description of orthotic footwear

### 6.3.1 Type of fabrication

Orthotic footwear can be:

- a) prefabricated orthotic footwear;
- b) custom fabricated orthotic footwear.

### 6.3.2 Prefabricated orthotic footwear

Types include footwear with:

- a) extra width and/or depth;
- b) protective toe caps (excluding industrial footwear);
- c) stiffened soles;
- d) rocker soles;
- e) extended openings;
- f) easy closures;

and footwear to

- g) partially unload the foot;
- h) accommodate vulnerable (at risk) feet;