

---

---

**Cosmetics — Microbiology — General  
instructions for microbiological  
examination**

*Cosmétiques — Microbiologie — Instructions générales pour les  
examens microbiologiques*

STANDARDSISO.COM : Click to view the full PDF of ISO 21148:2005



**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

STANDARDSISO.COM : Click to view the full PDF of ISO 21148:2005

© ISO 2005

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland

**Contents**

Page

<b>Foreword</b> .....	<b>iv</b>
<b>Introduction</b> .....	<b>v</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Terms and definitions</b> .....	<b>1</b>
<b>3 Premises</b> .....	<b>1</b>
<b>4 Equipment</b> .....	<b>3</b>
<b>5 Strains of microorganisms</b> .....	<b>6</b>
<b>6 Personnel</b> .....	<b>6</b>
<b>7 Preparation of the apparatus and glassware</b> .....	<b>6</b>
<b>8 Preparation and sterilization of culture media and reagents</b> .....	<b>8</b>
<b>9 Laboratory samples</b> .....	<b>10</b>
<b>10 Operating practices</b> .....	<b>12</b>
<b>11 Expression of results</b> .....	<b>13</b>
<b>12 Neutralization of the antimicrobial properties of the product</b> .....	<b>13</b>
<b>Annex A (informative) Basic identification techniques</b> .....	<b>14</b>
<b>Annex B (informative) Basic techniques for counting and plating</b> .....	<b>19</b>
<b>Annex C (informative) Preparation and calibration of inoculums</b> .....	<b>20</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 21148 was prepared by Technical Committee ISO/TC 217, *Cosmetics*.

STANDARDSISO.COM : Click to view the full PDF of ISO 21148:2005

## Introduction

The purpose of this International Standard is to help ensure that the general techniques used for conducting cosmetic microbiological examinations are the same in other laboratories that adopt these standards, to help achieve homogeneous results in different laboratories and to contribute towards the protection of the health of the laboratory personnel by preventing risk of infection.

When conducting microbiological examinations for cosmetic products, it is especially important that:

- only those microorganisms which are present in the samples be isolated or enumerated;
- the microorganisms do not contaminate the environment.

In order to achieve this, it is necessary to pay attention to personal hygiene and to use working techniques which ensure, as far as possible, exclusion of extraneous contamination.

Since, in this International Standard it is possible to give only a few examples of the precautions to be taken during microbiological examinations, a thorough knowledge of the microbiological techniques and of the microorganisms involved is essential. It is important that the analyses be conducted as accurately as possible, including calculation of the number of microorganisms.

A large number of manipulations can, for example, unintentionally lead to cross-contamination and the analyst should always verify the accuracy of the results given by his/her technique. It is necessary to take special precautions, not only for reasons of hygiene, but also to ensure good reproducibility of the results. It is not possible to specify all the precautions to be taken in all circumstances, but this International Standard at least provides the main measures to be taken when preparing, sterilizing and storing the media and the equipment.

The given recommendations will allow enumeration and detection of mesophilic microorganisms which may grow under aerobic conditions.

The recommendations are applicable to the determination of the absence of, or limited occurrence of specified microorganisms that are of interest for cosmetic products.

The test methods are described in the individual standards. Alternative microbiological procedures can be used provided that their equivalence has been demonstrated or the method has been otherwise validated. The choice of a specific method, or combination of methods mentioned in these International Standards will depend on the purpose for performing the test and it is for the user to decide which approach is best for his/her application.



# Cosmetics — Microbiology — General instructions for microbiological examination

## 1 Scope

This International Standard gives general instructions for carrying out microbiological examinations of cosmetic products, in order to ensure their quality and safety, in accordance with an appropriate risk analysis (e.g. low water activity, hydro-alcoholic, extreme pH values).

Because of the large variety of products and potential uses within this field of application, these instructions might not be appropriate for some products in every detail (e.g. certain water-immiscible products).

## 2 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

### 2.1

#### **product**

portion of an identified cosmetic product received in the laboratory for testing

### 2.2

#### **sample**

portion of the product (at least 1 g or 1 ml) which is used in the test to prepare the initial suspension

### 2.3

#### **initial suspension**

suspension (or solution) of a sample in a defined volume of an appropriate liquid (diluent, neutralizer, broth or combination of them)

### 2.4

#### **sample dilution**

dilution of the initial suspension

## 3 Premises

### 3.1 Test areas

The areas required for the specific operation of a microbiology laboratory are as follows:

- receipt, storage, preparation and processing of the samples;
- preparation and sterilization of culture media, apparatus and glassware;
- performance of analyses: weighing, dilutions, inoculations, subculturing, incubation, maintenance of the strain, etc.;
- decontamination and cleaning of apparatus, glassware, and processing of the analysis waste.

### 3.2 Additional areas

The areas included in this category are, for example:

- entrances, corridors, stairways, lifts;
- administrative areas (e.g. secretarial, offices, documentation rooms, etc.);
- cloakrooms and toilets;
- archive rooms;
- stores.

### 3.3 Location of the premises

The environment within which the microbiological analyses are carried out shall not affect the reliability of the analyses.

Care shall be taken to locate the premises so as to avoid risk of cross-contamination.

Care shall be taken to ensure protection against extreme conditions such as excess temperature, dust, humidity, steam, noise, vibration, exposure to direct sunlight, etc.

The surface area shall be sufficiently large to keep the work areas clean and orderly.

During the course of the tests, care shall be taken to limit access to the test areas to only those persons required to conduct the tests.

Separate rooms and/or separate areas and/or specific enclosures should be provided for the following:

- receipt, storage and preparation of samples;
- manipulation of microbial cultures;
- preparation of culture media, apparatus and glassware;
- decontamination and washing area;
- sterilization;
- incubators, refrigerators and freezers.

### 3.4 Equipping the premises

**3.4.1** The test premises shall be fitted out in the following ways in order to reduce the risks of contamination by dust and therefore by microorganisms:

- walls, ceilings and floors should be smooth, non-porous, easy to clean, and resistant to detergents and disinfectants used in laboratories;
- overhead pipes conveying fluids should not cross the premises unless they are hermetically enclosed;
- sun-protection systems, when used, shall be installed on the outside of the windows, where practicable;

— windows and doors shall be able to be closed when conducting the test in order to minimize draughts. Furthermore, they shall be designed so as to avoid the formation of dust traps and hence, to facilitate the cleaning.

**3.4.2** The ambient temperature and air quality (microorganism content, humidity, dust-spreading rate, etc.) shall be compatible with carrying out the tests.

According to needs, a filter-ventilation and/or a microbiological cabinet are recommended for this purpose.

**3.4.3** The laboratory bench tops and furniture shall be made of smooth, non-porous impermeable materials, which are easy to clean and disinfect. Cabinet and equipment tops should be accessible for cleaning.

Non-fixed laboratory furniture shall be designed so as to facilitate cleaning the floors.

It is desirable that documents or books that are not frequently used be kept outside the test areas.

### **3.5 Maintenance**

The floors, walls, ceiling, laboratory bench tops and furniture shall be maintained in good order to avoid cracks where dirt might particularly accumulate and thus cause a source of contamination.

Regular cleaning and, when relevant, disinfection shall be carried out in order to keep the premises in a condition suitable for conducting tests.

The ventilation systems and their filters shall be regularly maintained and filters changed when necessary.

## **4 Equipment**

### **4.1 General**

In general, all equipment shall be kept clean and in proper working condition.

Maintenance operations should be monitored. The measurement instruments and apparatus shall be regularly verified according to an appropriate timetable and results recorded.

### **4.2 Microbiological cabinets**

Cabinets are of two types:

- a) clean-air cabinets, which are intended to protect the product from extraneous contamination and to minimize contamination due to the operator;
- b) safety cabinets, which are intended to protect the product from extraneous contamination, and also to protect the operator and the environment.

Either cabinet can be used. Safety cabinets should be used for all work involving risk for the operator.

A cabinet is a dust-free workstation equipped with vertical laminar airflow. In microbiology, a safety cabinet is used to retain the microorganisms on filters.

### **4.3 Balances**

A microbiology laboratory for analyses of cosmetic products should be equipped with balances of the required range and accuracy for the different products to be weighed. Generally, the accuracy required for weighing the samples to be analysed and some components of the culture media and reagents, is  $\pm 0,01$  g.

#### 4.4 Homogenizer

This equipment (e.g. blender, stomacher, etc.) may be used to prepare the initial suspension from the test samples of non-liquid products.

#### 4.5 pH-meter

The pH-meter should be capable of measuring to an accuracy of  $\pm 0,1$  pH units and its minimum measuring threshold shall be 0,01 pH units.

#### 4.6 Autoclave

The autoclave shall be kept in good operating condition and shall regularly be inspected by the competent departments in accordance with the manufacturer's instructions and proper documentation should be recorded.

The autoclave shall not be used to sterilize both clean materials and also to decontaminate used materials at the same time. Wherever possible, separate autoclaves for these two processes should be used.

#### 4.7 Incubator

Incubators shall be equipped with a regulation system which allows the temperature to be kept even and stable over their entire working volume.

If the ambient temperature is close to, or higher than, that of the incubator, use an incubator with a cooling system.

Incubators should be protected from direct sunlight.

If possible, incubators should not be completely filled in one single operation because the culture media will take a long time to equilibrate to temperature, whatever type of incubator is used (forced-air convection or otherwise).

The temperature shall be checked and recorded at least every working day.

#### 4.8 Water baths

Water baths are of two types:

- thermostatically-controlled baths, suitable for incubation of inoculated culture media, for identification tests, etc.;
- temperature-controlled water baths for maintenance of sterile agar media in a molten state for later use in specified procedures.

The required temperature and accuracy are stipulated in each method of application.

#### 4.9 Refrigerator or cold-storage room

The temperature, unless otherwise specified, shall be  $5\text{ }^{\circ}\text{C} \pm 3\text{ }^{\circ}\text{C}$ .

#### 4.10 Freezer

The temperature, unless otherwise specified, shall be below  $-18\text{ }^{\circ}\text{C}$ .

#### 4.11 Sterilizing oven

A sterilizing oven is a chamber which allows the destruction of microorganisms by dry heat.

The temperature shall be evenly distributed within the chamber.

The oven shall be equipped with:

- a thermostat;
- a thermometer or a recording thermocouple;
- a duration indicator or a programmer/timer.

#### 4.12 Colony-counting device

A colony-counting device may be used.

#### 4.13 Other equipment

**WARNING — Volumetric glassware shall not be sterilized in a sterilizing oven.**

Other equipment and apparatus for everyday use, include the following:

- a) filtration apparatus (see below);
- b) glass or plastic containers (test tubes, flasks, bottles);
- c) glass or plastic Petri dishes (most commonly between 85 mm and 100 mm in diameter);
- d) glass or plastic pipettes (10 ml, 2 ml, 1 ml), automatic pipettes;
- e) sampling instruments;
- f) wires and loops (of nickel/ chromium, platinum/ iridium or disposable plastic, etc.);
- g) optical microscope;
- h) gas burner or wire incinerator;
- i) dispenser for culture media and reagents;
- j) mechanical stirrer.

If the membrane filtration method is used, the equipment shall also include:

- a membrane filtration system or filtration apparatus constructed of a suitable material, with a filter holder of at least 50 ml, and suitable for use of filters with diameter 47 mm to 50 mm and not more than 0,45 µm pore size;
- the type of membrane material is chosen in such a way that the bacteria are not affected by the residual components of the sample to be investigated;
- a vacuum source able to give an even filtration flow rate (the device shall be set as to obtain the filtration of 100 ml of liquid in less than 2 min).

## 5 Strains of microorganisms

The strains needed for the validation of the methodology are indicated in each method of application.

## 6 Personnel

### 6.1 Competence

All personnel working in a microbiology laboratory shall have received adequate training to enable them to conduct properly the operations entrusted to them.

The personnel who perform the tests shall have a good knowledge and sufficient practical experience with microbiological techniques and the microorganisms sought. The person in charge shall be able to interpret the accuracy and precision required to yield acceptable results.

### 6.2 Hygiene

In the field of personal hygiene, the following precautions shall be taken not only in order to avoid contaminating the samples and culture media, but also in order to avoid risk of personnel infection:

- wear laboratory clothing that is light-coloured, clean and in good condition, manufactured from a fabric which limits the risks of flammability; this clothing shall not be worn outside the work areas;
- keep nails perfectly clean and well groomed, and preferably short;
- wash hands, before and after microbiological examinations and immediately after visiting the toilets or eating; for drying hands, use disposable paper or disposable cloth towels;
- when inoculating, avoid speaking, coughing, etc;
- do not smoke, drink or eat in the test areas;
- do not put food for personal consumption in the laboratory refrigerators;
- special precautions shall be taken by persons having infections or illnesses that are likely to contaminate samples with microorganisms and may invalidate results.

## 7 Preparation of the apparatus and glassware

### 7.1 Preparation

The apparatus and glassware used in microbiology shall be prepared in such a manner as to guarantee its cleanliness and/or sterility up until the time of use.

Stopper tubes and cap bottles by appropriate material prior to sterilization.

Stopper the pipettes with cotton or any other appropriate material.

If necessary, the apparatus and glassware to be sterilized should be placed in special containers or wrapped in an appropriate material (e.g. special paper, aluminium, etc.).

## 7.2 Sterilization

### 7.2.1 Sterilization by dry heat

Heat in sterilizing oven for at least 1 h at 170 °C to 180 °C or use any combination of time and temperature if they are validated.

Indicators can be used in order to make certain that the sterilization has been achieved.

### 7.2.2 Sterilization by moist heat

Heat for at least 15 min at minimum of 121 °C in an autoclave. Indicators can be used in order to make certain that the sterilization has been achieved.

## 7.3 Disposable apparatus

Disposable apparatus may be used in the same way as the re-usable glassware (Petri dishes, pipettes, tubes, etc.) if the specifications are similar.

In this case, contact the manufacturer to determine that the proposed apparatus and glassware is suitable for use in microbiology (in particular sterility) and that the material contains no substances that inhibit the growth of microorganisms.

Disposable apparatus shall be decontaminated prior to its disposal. Other than the methods described in 7.6, and depending upon national regulations, incineration may be used. If there is an incinerator on the premises, decontamination and disposal may be achieved in a single operation.

## 7.4 Management of clean apparatus and glassware

Clean apparatus and glassware shall be protected against external contamination during storage, under conditions which maintain their cleanliness.

## 7.5 Management of sterile apparatus and glassware

Prior to use, the apparatus and glassware shall be stored under conditions which allow them to remain sterile. Disposable apparatus and glassware shall be stored in accordance to the manufacturer's specifications, without any damage to the packaging; laboratory-prepared apparatus and glassware shall be stored in clean containers.

When sterilizing apparatus and glassware intended for microbiology, an expiry date (or a manufacturing date) shall be put on each packaging. Hermetically sealed apparatus and glassware can be stored for up to 3 months prior to use unless otherwise specified.

## 7.6 Treatment of contaminated material

After use (culture of microorganisms or material in contact with microorganisms), the apparatus and glassware and their contents shall be treated for destruction of microorganisms, prior to cleaning or disposal, whatever the microorganism involved.

According to the nature of the materials, disinfection (10.1), sterilization (7.2) or incineration of disposable material (7.3) may be used.

## 7.7 Washing

Wash apparatus and glassware after they have been treated (7.6).

Empty the containers of their contents.

Prior to washing, separate seals from stoppers or caps, as appropriate.

Rinse detergent residue from equipment with tap water. Rinse clean equipment with water 8.2.

In the absence of any commercial product, a sodium carbonate solution of 0,125 % (mass fraction) can be used, followed by immersion in diluted acid [e.g. hydrochloric acid  $c(\text{HCl}) = 0,1 \text{ mol/l}$ ].<sup>[6]</sup>

Specialized apparatus and glassware may be used in order to facilitate cleaning operations (e.g. pipette washers, dishwashers, ultrasonic baths, etc.).

## 8 Preparation and sterilization of culture media and reagents

### 8.1 General

The accurate preparation of culture media is one of the fundamental steps in microbiological analysis and it shall be given special care.

### 8.2 Water

**WARNING — Water processed through an ion exchanger (de-ionized), may have a high microorganism content; it is therefore advisable not to use such water without verifying that the microorganism content of the water is low. Consult the manufacturer for the best way to minimize microbial contamination. Heavily contaminated deionized water that has been filter sterilized may still contain substances inhibitory to the growth of some microorganisms.**

Use distilled water or water of equivalent quality i.e. purified water <sup>[2] [9] [11]</sup> or deionized water <sup>[8]</sup>. If the distilled water is prepared from chlorinated water, neutralize the chlorine prior to the distillation.

The water shall be stored in containers manufactured of inert materials (e.g. neutral glass, polyethylene, etc.).

### 8.3 Preparation of culture media

#### 8.3.1 General

Two types of culture media preparation exist:

- from basic ingredients, dehydrated or not; or
- from dehydrated complete media.

Follow the manufacturer's specified storage conditions and expiration date.

Do not use the culture media beyond the stated shelf life.

Protect laboratory culture media that is in dehydrate from absorbing additional moisture from the environment during storage and use.

#### 8.3.2 Rehydration

Follow the manufacturer's recommendation for rehydration.

#### 8.3.3 Measurement of pH

Measure the pH using a pH-meter (4.5) and adjust it, if necessary, so that, after sterilization and cooling down to room temperature, the medium is at the required  $\text{pH} \pm 0,2$  units, unless otherwise stated.

NOTE The adjustment is normally carried out using a solution of approximately 40 g/l (about 1 mol/l) of sodium hydroxide (NaOH) or approximately 36,5 g/l (about 1 mol/l) of hydrochloric acid (HCl).<sup>[6]</sup>

#### 8.3.4 Dispensing

Dispense the medium into appropriate containers, either manually or using automatic apparatus.

### 8.4 Sterilization

#### 8.4.1 General

The sterilization of culture media and reagents can be carried out using various techniques, including:

- sterilization by moist heat;
- sterilization by filtration.

According to the technique used, once sterilized, the media should be monitored, in particular with respect to pH, colour, sterility and microbiological performance.

#### 8.4.2 Sterilization by moist heat

Use an autoclave (4.6) for sterilization. Generally, the sterilization step takes 15 min or longer at 121 °C. Adapt the sterilization cycle as necessary for the volume and number of containers, the loading pattern, and the media type.

The performance of the sterilization shall be verified using appropriate means.

#### 8.4.3 Sterilization by filtration

Sterilization by filtration can be performed under vacuum or at pressurized conditions.

Use sterile membranes and filter elements with a pore diameter of 0,22 µm (except in certain cases, where 0,45 µm can be used). Refer to the manufacturer's instructions regarding the use of filter elements or membranes which have been purchased in a sterile condition.

Sterilize the different components of the filtration apparatus, assembled or not, in the autoclave for 15 min at 121 °C. If necessary, aseptic assembly can be performed in a microbiological cabinet after autoclaving. Certain appliances can be purchased in a sterile condition.

### 8.5 Storage

#### 8.5.1 General

Each package of bottles, tubes and Petri dishes shall be labelled and shall bear the following details:

- name of the medium;
- date of preparation and/or expiry date.

#### 8.5.2 Laboratory-prepared culture media and reagents

Culture media dispensed in tubes or bottles and reagents which are not used immediately shall be protected against light and desiccation using appropriate inert stoppers or screw-on lids with inert liners.

They shall be kept under conditions which prevent their composition from being modified.

Never use media which have become dehydrated.

Prior to use, it is desirable that the temperature of the culture media be equilibrated to that of the laboratory, unless otherwise specified.

EXAMPLE TSA medium, prepared in the laboratory, is to be kept in flasks and stored in the dark, generally for no more than two months, unless otherwise specified.

### **8.5.3 Ready-to-use culture media and reagents**

It is necessary to comply with the manufacturer's instructions:

- expiry date,
- storage temperature and conditions,
- conditions for use (pH, etc.), and
- efficiency control.

### **8.6 Melting of agar culture media**

Melt a culture medium by placing it in boiling water bath or by any other process which gives identical result (e.g. a steam flow-through autoclave).

Avoid over-heating and remove the culture medium, as soon as it has melted. Keep the culture medium in a molten state in a water bath kept at no more than 48 °C until such time as it is to be used. Never use culture medium at a temperature higher than 48 °C. It is preferable not to keep a molten medium more than 8 h. In the case of particularly sensitive culture media, the melting duration shall be shortened.

No unused medium shall be resolidified for subsequent use.

### **8.7 Preparation of Petri dishes**

Pour the molten agar culture medium into sterile Petri dishes so as to obtain a thickness of at least 3 mm to 4 mm (e.g. for 90 mm diameter dishes, 15 ml to 20 ml of agar medium are normally required).

Allow the agar medium to cool and solidify by placing the Petri dishes on a cool, horizontal surface.

Use the thus-prepared Petri dishes immediately or store them under (in the dark and in the appropriate temperature and duration) conditions which prevent their composition from being modified. Label the dishes as described in 8.5.

Drying may be necessary before use.

Ready-prepared agar medium plates are available commercially. Store and use them according to the manufacturer's instructions.

## **9 Laboratory samples**

### **9.1 General**

Product and sample are defined in 2.1 and 2.2 respectively.

## 9.2 Sampling the cosmetic product

It is important that the laboratory receive a product which is truly representative of the cosmetic product and has not been damaged during transport or storage.

Sampling shall have been carried out in accordance with the specific appropriate documents. If there is no specific document, it is recommended that the parties concerned come to an agreement on this subject.

## 9.3 Transport

During their transportation to the laboratory, products to be tested shall be kept under the conditions which minimize changes in the microbial content.

## 9.4 Receipt and storage

The laboratory personnel shall check the condition of the products on receipt and confirm that condition and quantity is satisfactory. If their condition is unsatisfactory or if their quantity is insufficient, the laboratory cannot perform the prescribed test.

However, in special circumstances if the test is completed, the personnel shall record the condition of the product and the reason for testing.

The products admitted into the laboratory shall be documented in such a manner that their progress through to the time of drafting the test report can be monitored.

The following information shall be noted:

- date of receipt;
- characteristics of the sampling operation (sampling date, sampling conditions, etc.);
- name, reference, origin and requesting party;
- characteristics of the product.

If necessary, store products to be tested at room temperature. Do not incubate, refrigerate or freeze products (2.1) and samples (2.2) before or after analysis.

## 9.5 Handling products and samples

In order to avoid contamination of the environment and of the products and samples, handle the products in such a manner as to avoid any risk of contamination. To achieve this, follow aseptic techniques including, for example:

- any instrument that is used to open the packaging shall be sterile;
- any instrument that is used to remove the sample from the product shall be sterile;
- if required, sanitize the package and cover to be opened as appropriate.

## 9.6 Conservation and destruction of products

Except for special cases, keep the products until all results have been obtained, or longer if necessary.

Discard products from which samples have been taken, except if the nature and the level of contamination justify that they be treated as a contaminated material (7.6).

## 10 Operating practices

### 10.1 Hygienic precautions during the testing

Precautions shall be taken in order to conduct the work as far as possible under aseptic conditions, for example:

- a) make sure that the work area is clean and that there are no draughts (doors and windows closed);
- b) before and after the work, decontaminate the work surface with an appropriate disinfectant;
- c) prior to starting, make sure that everything required for carrying out the work is available;
- d) in the case of work conducted in a microbiological cabinet, use sterile gloves or decontaminate hands prior to starting work and avoid crossing forearms and hands;
- e) when not working in a microbiological cabinet, open the sample container in the vicinity of a flame, holding them in the most inclined position possible;
- f) carry out the work as quickly as possible without making any unnecessary movements;
- g) if the total contents of a bag of disposable pipettes, Petri dishes, etc., is not used during the course of an examination, make sure that the container is properly closed after taking the appropriate number of units;
- h) sterilize loops and inoculation wires, etc., before and after use with a flame; in order to avoid splatter of substances and microorganisms, preferably use a wire incinerator whenever possible or use disposable sterile loops and wires;
- i) place used pipettes, spatulas, etc. in specific receptacles containing an appropriate disinfectant (e.g. a sodium hypochlorite solution for the pipettes) prior to treatment (7.6);
- j) place reusable equipment which may contain microorganisms into specific containers prior to sterilization, before the washing operation;
- k) place used disposable apparatus in appropriate containers prior to sterilization or incineration (7.6);
- l) immediately mop up spillage by means of cotton pads or any other appropriate material impregnated with appropriate disinfectant, then clean and disinfect the work surface prior to continuing.

The manipulation of products and subsequent cultures likely to contain pathogenic bacteria requires special precautions. The following are recommended:

- a microbiological cabinet for all manipulations required for conducting the analysis;
- automatic pipettes (pipetting by mouth suction is strictly prohibited).

NOTE Droplets are a major cause of environmental contamination and of infection. Droplets can be formed for example:

- when using shakers, syringes, etc.;
- when emptying pipettes by blowing;
- when sterilizing wet inoculation loops or needles.

It is therefore necessary to minimize their formation.

## 10.2 Preparation of the initial suspension and of sample dilutions

### 10.2.1 General

In the case of the preparation of an initial suspension and of sample dilutions, the time that elapses between the end of the preparation and the moment the inoculum comes into contact with the culture medium shall not exceed 45 min, unless specifically mentioned in the relevant documents.

The initial suspension is prepared from a sample of at least 1 g or 1 ml of the well-mixed product under test.

Record the exact mass or volume of the sample,  $S$ .

### 10.2.2 Water-miscible product

Transfer the sample ( $S$ ) of product into any suitable container and make a precise and appropriate dilution according to the standard to be applied.

Record the dilution factor,  $d$ .

### 10.2.3 Water-immiscible products

Transfer the sample ( $S$ ) of product into any suitable container containing an appropriate amount of solubilizing agent (e.g. polysorbate 80) and make a precise and appropriate dilution according to the standard to be applied.

Record the dilution factor,  $d$ .

## 10.3 Counting methods

Follow the specifications of the standard method to be applied. For information, see Annex B.

## 10.4 Detection methods

Follow the specifications of the standard method to be applied.

## 11 Expression of results

Follow the specifications of the standard method to be applied.

## 12 Neutralization of the antimicrobial properties of the product

Prior to detection or enumeration viable microorganisms in a cosmetic product, the possible inhibition of microbial growth by the sample shall be neutralized. In all cases and whatever the methodology, the neutralization of the antimicrobial properties of the product shall be checked and validated.

Follow the specifications of the standard method to be applied.

## Annex A (informative)

### Basic identification techniques

#### A.1 Preparation of a pure culture

##### A.1.1 General

Begin the preparation of a pure culture by the selection of a colony on or in an agar medium which has been inoculated with a dilution of the test sample or with a culture.

Then inoculate the selected colony on a non-selective agar culture medium. After incubation, select a well-isolated colony. Repeat the operation if necessary.

Use the plating-out techniques described in A.1.2. Different methods may prove to be necessary in specific cases.

##### A.1.2 Plating out

###### A.1.2.1 General

Take a small quantity from the surface of a well-isolated colony using the tip of a sterile loop.

Then plate out, either directly with the cells present on the loop (A.1.2.2) or after having prepared a suspension of these cells (A.1.2.3).

###### A.1.2.2 Direct method: Example

Using the tip of the loop, inoculate, in close streaks, a portion of about one-third of the agar medium surface area. Sterilize and cool down the loop. From the edge of the inoculated area, make another series of streaks, less close together than in the first case, over half of the surface area not yet inoculated. Repeat the operation over the remaining surface area making more-spread-out streaks (see Figure A.1).




Figure A.1 — Example of plating out: Direct method

**A.1.2.3 Method using dilution**

Suspend the cells in 1 ml to 2 ml the selected dilution, rubbing the inoculated loop against the wall of the tube at the surface of the liquid, then mix well.

Sterilize and cool down the loop. Using the loop, take a small portion of the microbial suspension and proceed as stated in A.1.2.2.

**A.1.3 Incubation**

Invert the inoculated Petri dishes and place them in the incubator for the chosen time at the chosen temperature.

**A.1.4 Selection**

After incubation, select a well-isolated colony from the dish, either for subsequent plating out, or for the tests to be performed.

If possible, the final tests should be carried out using cells stemming from one single colony. If there is insufficient cell material in one colony, it should first be subcultured in a liquid medium or on the agar medium slant, after which the subculture can be used for the tests to be performed.

**A.2 Gram's stain (modified Hucker technique)****A.2.1 General**

This staining of bacterial cells allows description for the morphology of the bacteria and classification of them into two groups as a function of whether or not they are capable of retaining the violet stain of crystal violet under the test conditions. This division results mainly from differences in the structure of the cell walls of the two groups and it is correlated with other major differences between the two groups. There are a number of ways to conduct a Gram's stain, but all follow the sequences given below.

**A.2.2 Solutions****A.2.2.1 General**

Commercially available solutions may be used. In this case, follow the manufacturer's recommendations.

**A.2.2.2 Crystal violet solution****A.2.2.2.1 Composition**

Crystal violet	2,0 g
Ethanol (95 %)	20 ml
Ammonium oxalate (C <sub>2</sub> H <sub>8</sub> N <sub>2</sub> O <sub>4</sub> )	0,8 g
Water	80 ml

**A.2.2.2.2 Preparation**

Dissolve the crystal violet in the ethanol and the ammonium oxalate in the distilled water. Mix the two solutions and allow the mixture to stand for 24 h prior to use.

**A.2.2.3 Iodine solution**

**A.2.2.3.1 Composition**

Iodine	1,0 g
Potassium iodide (KI)	2,0 g
Water	100 ml

**A.2.2.3.2 Preparation**

Dissolve the potassium iodide in 10 ml of distilled water, add the iodine in fractions.

After dissolution, make up to 100 ml in a volumetric flask.

**A.2.2.4 Safranin solution**

**A.2.2.4.1 Composition**

Safranin O	0,25 g
Ethanol (95 %)	10 ml
Water	100 ml

**A.2.2.4.2 Preparation**

Dissolve the safranin in the ethanol then mix with the distilled water. Make up volumetrically to a final volume of 100 ml.

When crystal violet is used, the stability of the solution should be verified. For the verification, mix one drop of the crystal violet solution with one drop of iodine solution on a glass slide to see a chemical reaction. If the crystallization is seen on the glass slide, do not use the crystal violet solution.

**A.2.2.5 Staining technique**

After fixing (e.g. with a flame) the bacterial film on the microscope slide prepared from a culture 18 h to 24 h, or when the broth is turbid, cover the film with the crystal violet solution (A.2.2.2). Allow it to react for 1 min.

Gently rinse the inclined slide with water for a few seconds.

Cover the slide with the iodine solution (A.2.2.3). Allow it to react for 1 min. Gently rinse the inclined slide with water for a few seconds.

Pour gently and continuously a film of ethanol (95 %) onto the inclined slide over a period of no more than 30 s and until no more of the violet colour is emitted.

Gently rinse the inclined slide with water in order to eliminate the ethanol.

Cover the slide with the solution of safranin (A.2.2.4) for 10 s.

Gently rinse the inclined slide with water.

Dry the slide.