

---

---

**Workplace air — Gases and vapours  
— Requirements for evaluation of  
measuring procedures using diffusive  
samplers**

*Air des lieux de travail — Gazes et vapeurs — Exigences pour  
l'évaluation des procédures pour le mesurage à l'aide de dispositifs de  
prélèvement par diffusion*

STANDARDSISO.COM : Click to view the full PDF of ISO 23320:2022



STANDARDSISO.COM : Click to view the full PDF of ISO 23320:2022



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

	Page
Foreword.....	v
Introduction.....	vi
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>1</b>
<b>4 Symbols and abbreviated terms</b> .....	<b>1</b>
<b>5 Types of samplers</b> .....	<b>3</b>
<b>6 Requirements</b> .....	<b>3</b>
6.1 General.....	3
6.2 Sampler requirements.....	3
6.2.1 Nominal uptake rate.....	3
6.2.2 Air velocity/sampler orientation.....	3
6.2.3 Sampler leak test.....	4
6.2.4 Shelf life.....	4
6.2.5 Sampler identification (for commercially available diffusive samplers).....	4
6.2.6 Marking.....	4
6.2.7 Instructions for use.....	4
6.3 Measuring procedure requirements.....	5
6.3.1 Sampling procedure requirements.....	5
6.3.2 Analytical procedure requirements.....	5
6.3.3 Expanded uncertainty.....	6
6.3.4 Method description.....	6
<b>7 General test conditions</b> .....	<b>7</b>
7.1 Reagents.....	7
7.2 Apparatus.....	7
7.3 Independent method.....	7
7.4 Generation of a calibration gas mixture.....	8
7.4.1 General.....	8
7.4.2 Determination of mass concentration.....	8
<b>8 Test methods</b> .....	<b>9</b>
8.1 General.....	9
8.2 Sampler test methods.....	9
8.2.1 Determination of (nominal) uptake rate.....	9
8.2.2 Air velocity.....	10
8.2.3 Sampler leak test.....	11
8.2.4 Shelf life (for Type A impregnated supports).....	11
8.2.5 Sampler identification.....	12
8.2.6 Marking.....	12
8.2.7 Instructions for use.....	12
8.3 Measuring procedure test methods.....	12
8.3.1 Determination of the sampling conditions.....	12
8.3.2 Analytical procedure test methods.....	13
8.3.3 Method recovery and method precision.....	15
8.4 Uncertainty of measurement.....	17
8.4.1 Identification of random and non-random uncertainty components.....	17
8.4.2 Estimation of individual uncertainty components.....	17
8.4.3 Calculation of expanded uncertainty.....	19
<b>9 Test report</b> .....	<b>19</b>
<b>Annex A (informative) Fundamentals of diffusive sampling</b> .....	<b>20</b>
<b>Annex B (informative) Estimation of uncertainty of measurement</b> .....	<b>23</b>

<b>Annex C (informative) Calculation of uptakes rates from diffusion coefficients</b> .....	<b>33</b>
<b>Annex D (informative) Example of estimation of expanded uncertainty</b> .....	<b>35</b>
<b>Bibliography</b> .....	<b>38</b>

STANDARDSISO.COM : Click to view the full PDF of ISO 23320:2022

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 146, *Air quality*, Subcommittee SC 2, *Workplace atmospheres*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 137, *Assessment of workplace exposure to chemical and biological agents*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

## Introduction

This document provides a framework for assessing the performance of procedures for measuring gases and vapours against the general requirements for the performance of procedures for measuring chemical agents in workplace atmospheres as specified in ISO 20581. These performance criteria include maximum values of expanded uncertainty achievable under prescribed laboratory conditions for the methods to be used.

This document enables manufacturers and users of diffusive samplers and developers and users of procedures for measuring gases and vapours to adopt a consistent approach to method validation.

This document is based on EN 838:2010, published by the European Committee for Standardization (CEN) and is also complementary to ISO 16107.

STANDARDSISO.COM : Click to view the full PDF of ISO 23320:2022

# Workplace air — Gases and vapours — Requirements for evaluation of measuring procedures using diffusive samplers

## 1 Scope

This document specifies performance requirements and test methods under prescribed laboratory conditions for the evaluation of diffusive samplers (see Reference [1]) and of procedures using these samplers for the determination of gases and vapours in workplace atmospheres (see Reference [2]).

This document is applicable to diffusive samplers and measuring procedures using these samplers, such as ISO 16200-2 and ISO 16017-2, in which sampling and analysis are carried out in separate stages.

This document is not applicable to

- diffusive samplers which are used for the direct determination of concentrations, and
- diffusive samplers which rely on sorption into a liquid.

This document addresses requirements for method developers and/or manufacturers.

NOTE For the purposes of this document a manufacturer can be any commercial or non-commercial entity.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 20581, *Workplace air — General requirements for the performance of procedures for the measurement of chemical agents*

ISO 22065, *Workplace air — Gases and vapours — Requirements for evaluation of measuring procedures using pumped samplers*

ISO 18158, *Workplace air — Terminology*

ISO 8655-2, *Piston-operated volumetric apparatus — Part 2: Piston pipettes*

ISO 8655-6, *Piston-operated volumetric apparatus — Part 6: Gravimetric methods for the determination of measurement error*

## 3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 18158 and ISO 20581 apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

## 4 Symbols and abbreviated terms

NOTE See 8.4 and Annex C for symbols used in conjunction with uncertainty of measurement only.

$A$	cross-sectional area of sorption surface, in square centimetres (cm <sup>2</sup> )
CRM	certified reference material
$D_a$	diffusion coefficient of an analyte, in square centimetres per minute (cm <sup>2</sup> · min <sup>-1</sup> )
$D_{a1}$	diffusion coefficient of analyte 1, in square centimetres per minute (cm <sup>2</sup> · min <sup>-1</sup> )
$D_{a2}$	diffusion coefficient of analyte 2, in square centimetres per minute (cm <sup>2</sup> · min <sup>-1</sup> )
$l$	length of static air layer in sampler (or equivalent for permeation types), in centimetres (cm)
$m_b$	mass of analyte desorbed from blank sampler, in nanograms (ng)
$m_d$	mass of analyte desorbed, in nanograms (ng)
$m_s$	mass of the analyte which can diffuse to a suitable sorbent within a certain time, i.e. the mass uptake of a diffusive sampler, in nanograms (ng)
$\dot{m}_1$	mass loss from permeation tube, in micrograms per minute (µg · min <sup>-1</sup> )
$M_a$	molar mass of analyte, in grams per mole (g · mol <sup>-1</sup> )
$n$	number of replicate samples
OELV	occupational exposure limit value
$p_{at}$	actual pressure of the test atmosphere sampled, in kilopascals (kPa)
$R$	recovery
$R_{an}$	analytical recovery
RH	relative humidity of the test atmosphere sampled, in percent (%)
$t_e$	exposure time, in minutes (min)
$T_{at}$	temperature of the test atmosphere sampled, in Kelvin (K)
$\dot{U}_d$	uptake rate, in cubic centimetres per minute (cm <sup>3</sup> · min <sup>-1</sup> )
$(\dot{U}_d)'$	uptake rate, in nanograms per parts per million (volume fraction) per minute (ng · ppm <sup>-1</sup> · min <sup>-1</sup> )
$\dot{U}_{d1}$	uptake rate of analyte 1, in cubic centimetres per minute (cm <sup>3</sup> · min <sup>-1</sup> )
$\dot{U}_{d2}$	uptake rate of analyte 2, in cubic centimetres per minute (cm <sup>3</sup> · min <sup>-1</sup> )
$\dot{v}$	flow rate into the exposure chamber, for example, in litres per minute (l · min <sup>-1</sup> )
$\beta_a$	mass concentration of the analyte in the calibration gas mixture, in milligrams per cubic metre (mg · m <sup>-3</sup> )
$(\beta_a)'$	mass concentration in parts per million (ppm);
$\beta_{a1}$	mass concentration of the given analyte at the beginning of the diffusion layer (i.e. at the distance $l$ from the surface of the sorbent), in milligrams per cubic metre (mg · m <sup>-3</sup> )

$\beta_{a2}$	mass concentration of the given analyte at the end of the diffusion layer (i.e. at the surface of the sorbent), in milligrams per cubic metre ( $\text{mg} \cdot \text{m}^{-3}$ )
$\bar{\beta}_{a,R}$	mean mass concentration of the analyte recovered from the test gas atmosphere, in milligrams per cubic metre ( $\text{mg} \cdot \text{m}^{-3}$ );
$\beta_{cg}$	mass concentration of the calibration gas mixture, in milligrams per cubic metre ( $\text{mg} \cdot \text{m}^{-3}$ )
$\vartheta_{at}$	temperature of the test atmosphere sampled, in degree Celsius ( $^{\circ}\text{C}$ )
$K_v$	coefficient of variation (CV)(The predecessor term "relative standard deviation" is deprecated and has been replaced by the term "coefficient of variation". See also ISO 3534-1:2006, 2.38, Note 2.)
$\vartheta_a$	volume fraction of the analyte, in microlitres per litre ( $\mu\text{l} \cdot \text{l}^{-1}$ )

## 5 Types of samplers

Samplers for gases and vapours can be divided into type A samplers and type B samplers:

Type A samplers rely on sorption onto a solid or onto a support impregnated with a reagent, desorption with solvent, and subsequent analysis of the desorbate. They are usually made of polypropylene or glass and consist of one or more sorbent layers and contain an active sorbent (e.g. activated carbon) or a support impregnated with reagent.

Type B samplers rely on sorption onto a solid or onto a support impregnated with a reagent, thermal desorption, and analysis of the desorbate. They are usually made of glass or metal, are sealed with removable fittings and consist of one or more beds of sorbent (e.g. porous polymer resin).

## 6 Requirements

### 6.1 General

Some requirements (see [6.2](#)) shall be initially verified by the manufacturer once for each type of sampler. Other requirements (see [6.3](#)) shall be verified for each combination sampler/chemical agent.

Measuring procedures shall meet the requirements for measuring procedures specified in [6.3](#). When use of a sampler for measurement of a particular gas or vapour is claimed, the sampler shall meet the requirements specified in [6.2](#).

NOTE 1 No useful performance requirements can be given for the effect of interferents (with the exception of relative humidity). The effect of interferents is difficult to predict for a non-ideal sorbent without adsorption isotherm data on mixed systems which is normally unavailable. However, the user of diffusive samplers is cautioned that the adsorption of water vapour on certain sorbents, e.g. activated carbon and silica gel, can have a large effect on sampler capacity and analytical recovery.

NOTE 2 Because of the known effect of pressure on diffusion coefficients, a pressure test is not necessary.

### 6.2 Sampler requirements

#### 6.2.1 Nominal uptake rate

The nominal uptake rate and the coefficient of variation shall be provided by the manufacturer in accordance with [8.2.1](#) and [Annex A](#).

#### 6.2.2 Air velocity/sampler orientation

The manufacturer shall test the working range of air velocity and the influence of sampler orientation in accordance with [8.2.2](#).

### 6.2.3 Sampler leak test

When tested in accordance with [8.2.3](#), any additional analyte determined above the blank value (see [6.3.2.3](#)) shall be less the one-third of the calculated mass uptake by the sampler for 30 min exposure to a concentration of 0,1 OELV.

### 6.2.4 Shelf life

The manufacturer shall specify the shelf life of the diffusive sampler when stored in its original package. During this period the sampler shall fulfil all requirements.

### 6.2.5 Sampler identification (for commercially available diffusive samplers)

Samplers shall be uniquely identified.

### 6.2.6 Marking

Diffusive samplers shall be marked with at least the following:

- a) manufacturer's name;
- b) product identification;
- c) batch identification;
- d) shelf life (if applicable);
- e) number of this document.

If required due to limited space, the marking may be placed on the packaging of the diffusive sampler. However, the manufacturer's name and product identification shall be indicated on the diffusive sampler.

### 6.2.7 Instructions for use

The instructions for use supplied with the diffusive sampler shall be written in the principal language(s) used in the countries where the diffusive sampler is to be marketed. Where appropriate, the instructions for use shall contain directly and by reference to an online document, at least the following information:

- a) designated use (general purpose for a number of gases and vapours or, specific, for a particular gas or vapour, see [6.1](#));
- b) blank value (only when used for a particular gas or vapour, see [6.1](#));
- c) nominal uptake rate for the substances for which the diffusive sampler is intended to use;
- d) directions for proper handling of the diffusive sampler, including opening and closing;
- e) general information on the principle of use, for example, sorbent type, reaction of the reagent impregnated solid, desorption method;
- f) information on storage and transport;
- g) air velocity range in which the sampler can be used;
- h) orientation;
- i) information on health or environmental hazards and method of disposal.

The general information on the principle of use can be given in additional literature.

## 6.3 Measuring procedure requirements

### 6.3.1 Sampling procedure requirements

#### 6.3.1.1 Sampling time

Sampling time shall be established according to concentration range of the compounds of interest over which measurements are to be made, i.e. up to two times the OELV, see ISO 20581, and taking into account the nominal or theoretical uptake rate.

#### 6.3.1.2 Bias due to the selection of a non-ideal sorbent (back diffusion)

When tested in accordance with [8.3.1.1](#), the bias shall be  $\leq 10\%$ .

#### 6.3.1.3 Uptake rate

If it is possible to calculate the ideal steady-state value in accordance with [8.2.1](#), the nominal uptake rate, determined in accordance with [8.2.1](#), shall be within  $\pm 25\%$  of the steady-state value.

#### 6.3.1.4 Storage conditions after sampling

The storage conditions after sampling shall be specified. When tested in accordance with [8.3.1.3](#), the mean value of the recovery after storage shall not differ by more than 10 % from the value before storage.

### 6.3.2 Analytical procedure requirements

#### 6.3.2.1 Limit of quantification

The limit of quantification as determined in [8.3.2.1](#) for long-term OELVs shall be lower than the mass collected by the sampler at a concentration of 0,1 OELV for 8 h.

The limit of quantification for short-term OELVs shall be lower than the mass collected by the sampler at a concentration of 0,5 OELV for 15 min.

If the sampler is to be used for shorter reference periods, the limit of quantification also shall be able to be measured at those periods.

#### 6.3.2.2 Analytical recovery

When tested in accordance with [8.3.2.2](#) the analytical recovery  $R_{an}$  shall be

for type A samplers:  $R_{an} \geq 75\%$  with  $K_v \leq 10\%$  at each loading, and

for type B samplers:  $R_{an} \geq 95\%$  with  $K_v \leq 10\%$  at each loading.

Where the analytical recovery cannot be achieved the user shall take care that the measuring procedure meets all other requirements of this document and ISO 20581.

#### 6.3.2.3 Blank value

In order to obtain acceptable values for the quantification limit of the method, the blank value of the sampling media should be as low as technically possible.

The blank value as tested in [8.3.2.3](#) shall be less than the limit of quantification as determined in accordance with [8.3.2.1](#) or otherwise it can be subtracted from the result, but the standard deviation of the blank value will contribute to the expanded uncertainty of the measuring procedure.

Where it is known that the blank value is significant and varies between batches of samplers, it shall be checked on each batch of samplers.

To eliminate any contamination which could occur during storage before use Type B samplers should be cleaned before sampling by taking them through the thermal desorption procedure.

This cleaning process should be carried out as close as possible to the time when the samplers will be used.

### 6.3.3 Expanded uncertainty

When tested in accordance with 8.3 the expanded uncertainty calculated in accordance with 8.4 shall meet the requirements given in ISO 20581.

The expanded uncertainty requirement shall be met from 10 °C to 40 °C and at relative humidities from 20 % to 80 %. Above 30 °C the use of correction factors is permitted to meet this requirement.

In addition, the performance criteria should also be met under a wider variety of environmental influences, representative of workplace conditions.

### 6.3.4 Method description

#### 6.3.4.1 Scope of the measuring procedure

The scope of the measuring procedure shall include information on the following:

- a) principle of the method;
- b) chemical agents covered by the measuring procedure;
- c) analytical technique used;
- d) working ranges;
- e) chemical agents for which the measuring procedure is known to be adequate but not completely validated according to this document, especially in case of compounds of the same chemical family or homologous series;
- f) chemical agents for which the measuring procedure is known to be inadequate;
- g) any known interferences.

#### 6.3.4.2 Method performance

The measuring procedure shall comprise information about method performance, including the following:

- a) the chemical agents for which measuring procedure has been shown to be effective;
- b) the range of concentrations of chemical agents in air, sample volume, uptake rates, exposure time and range of environmental conditions over which the measuring procedure has been shown to meet the performance criteria for expanded uncertainty prescribed in ISO 20581;
- c) the limit of quantification of the measuring procedure for chemical agents of interest;
- d) full details of any known interferences, including suitable and sufficient information on how to minimise their effects.

### 6.3.4.3 Apparatus and reagents

With regard to apparatus and reagents the measuring procedure shall

- a) specify that the diffusive sampler used complies with the provisions of this document,
- b) specify the required characteristics of analytical instruments to be used,
- c) specify the quality of the reagents to be used.

### 6.3.4.4 Safety information

The measuring procedure shall provide suitable and sufficient information on the safety hazards associated with the reagents and equipment used.

## 7 General test conditions

### 7.1 Reagents

Use reagents of analytical grade, where possible.

### 7.2 Apparatus

Usual laboratory apparatus and the following:

**7.2.1 Dynamic system** for generating, pre-mixing and delivering a known concentration of a test gas or vapour in air (see ISO 6145-1, ISO 6145-4, ISO 6145-6 and ISO 6145-10), including at least:

- an exposure chamber constructed of inert materials such as glass or polytetrafluorethylene (PTFE), through which the generated test atmosphere is passed, of sufficient capacity to accommodate simultaneously at least six test samplers and six samplers of one independent method (see 7.3) positioned in such a manner that there is no interference between each sampler;
- provisions for measuring, controlling and varying the air flow rate through the chamber and the concentration, temperature and relative humidity of the calibration gas mixture.

NOTE It is also possible to use a smaller exposure chamber and to carry out repeat experiments to obtain at least six pairs of data.

**7.2.2 Micropipettes or syringes**, for applying known volumes of standard solutions, conforming with the requirements of ISO 8655-2 and with a calibration checked in accordance with ISO 8655-6.

**7.2.3 Instruments for analysing the gas, vapour or a characteristic reaction product** collected by either the test sampler or an independent sampling method.

### 7.3 Independent method

The concentration of the generated calibration gas mixture in the exposure chamber shall be verified as follows:

- a) by an independent method, which has been validated using an established protocol, for example a pumped sampler method, bubbler method, or a different diffusive sampler method; or
- b) by using an independently calibrated on-line instrument, e.g. a flame ionization detector, or an infrared spectrometer.

If a pumped sampler procedure is used as the independent method, the method shall conform with all requirements of ISO 22065.

## 7.4 Generation of a calibration gas mixture

### 7.4.1 General

Set up a calibration gas mixture (see ISO 6141, ISO 6143, ISO 6144 and Reference [3]) at the concentration and values of temperature, relative humidity, etc. specified in the appropriate test methods in [Clause 8](#).

Ensure that the flow rate into the exposure chamber exceeds the combined sampling rate of all samplers by at least 25 %.

### 7.4.2 Determination of mass concentration

**7.4.2.1** Calculate the mass concentration of the calibration gas mixture,  $\beta_{cg}$ , given in milligrams per cubic metre ( $\text{mg} \cdot \text{m}^{-3}$ ), from the test atmosphere generation parameters. For example, for a permeation cell system, the delivered mass concentration is:

$$\beta_{cg} = \frac{\dot{m}_1}{\dot{v}} \quad (1)$$

where

$\dot{m}_1$  is the mass loss from permeation tube, in micrograms per minute ( $\mu\text{g} \cdot \text{min}^{-1}$ );

$\dot{v}$  is the flow rate into the exposure chamber, for example, in litres per minute ( $\text{l} \cdot \text{min}^{-1}$ ).

NOTE 1 The example does not give a preference for permeation systems for generating calibration gas mixtures of gases and vapours.

NOTE 2 This value is the calculated inlet value of the exposure chamber concentration.

**7.4.2.2** Measure the mass concentrations at the inlet and outlet of the exposure chamber using the independent method described in [7.3](#) with all samplers within the test chamber, including both the test and independent method functioning.

Determine whether the measured outlet mass concentration differs by more than 5 % from the measured inlet mass concentration. If it does, then the calibration gas mixture generation system shall be changed e.g. by increasing the flow rate or chamber volume, until the difference is less than 5 %.

When the difference is less than 5 %, calculate the mean mass concentration in the test atmosphere within the exposure chamber either from the mean of the calculated inlet and outlet values, or from the mean calculated inlet value adjusted for (half of) the experimentally determined depletion.

**7.4.2.3** Determine the mean mass concentration of the test atmosphere within the exposure chamber experimentally using the results of the independent method described in [7.3](#). A correction may be applied for any known bias in the independent method.

Compare the experimentally determined mass concentration with the calculated value (see [7.4.2.2](#)). If the experimentally determined value is within  $\pm 10$  % of the calculated value of the mass concentration of the delivered test atmosphere, take the calculated value as the true value. If this requirement is not met, then make adjustments or use an alternative generation method or verify the independent method.

If it is not possible to calculate the mass concentration of the calibration gas mixture, for example, for reactive gases, the value determined by the independent method shall be used as the true value.

## 8 Test methods

### 8.1 General

If it is known in advance that a certain type of diffusive sampler is unaffected by an environmental influence, then the relevant tests in 8.3.3.1 to 8.3.3.5 may be modified to examine only the factors likely to have an influence.

If not otherwise specified in the test procedure, the sampler orientation shall be as specified by the manufacturer.

There are different levels of evaluation. These levels are specified as follows:

- a) level 1: A measuring procedure evaluated for the analyte of interest in accordance with the normative part of this document;
- b) level 2: A measuring procedure deemed to be compliant with the normative part of this document on the basis that the analyte of interest is an analogue within a homologous series, both upper and lower members of which have been tested and shown to comply with level 1. Such an evaluation shall include at least the nominal uptake rate determination as specified in 8.2.1 and the determination of the analytical recovery as specified in 8.3.2.2.

NOTE 1 Some special groups of substances (for example toluene, xylenes) usually isomers, can be treated as homologous when it is known that their chemical and physical properties are very similar.

NOTE 2 To reduce the number of experiments a factorial design can be applied, see References [4] to [7].

### 8.2 Sampler test methods

#### 8.2.1 Determination of (nominal) uptake rate

Expose a set of six diffusive samplers to a test atmosphere under the following exposure conditions:

- concentration: 1 OELV;
- time: 4 h;
- relative humidity:  $(50 \pm 5)$  %;
- temperature: 20 °C to 25 °C;
- air velocity:  $0,5 \text{ m} \cdot \text{s}^{-1}$ .

Analyse the diffusive samplers by reference to standard solutions or to samplers spiked with known amounts of analyte.

Calculate the (nominal) uptake rate according to Formula (2):

$$\dot{U}_d = \frac{m_d - m_b}{R_{an} \times \beta_a \times t_e} \quad (2)$$

where

$\dot{U}_d$  is the uptake rate, in cubic centimetres per minute ( $\text{cm}^3 \cdot \text{min}^{-1}$ );

$m_d$  is the mass of analyte desorbed, in nanograms (ng);

$m_b$  is the mass of analyte desorbed from blank sampler, in nanograms (ng);

$R_{an}$  is the analytical recovery.

$\beta_a$  is the mass concentration of the analyte in the calibration gas mixture, in milligrams per cubic metre ( $\text{mg} \cdot \text{m}^{-3}$ )

$t_e$  is the exposure time, in minutes (min)

NOTE 1 If the mass concentration is given as  $10^{-6}$  (parts per million), use  $(\beta_a)'$  and  $(\dot{U}_d)'$  instead of  $\beta_a$  and  $\dot{U}_d$ .

Calculate the mean (nominal) uptake rate and the coefficient of variation.

NOTE 2 For the calculation of uptake rates from diffusion coefficients see [Annex C](#).

### 8.2.2 Air velocity

The manufacturer is responsible for ensuring that this test is carried out for at least one chemical agent.

Expose a set of six diffusive samplers to a test atmosphere under the following exposure conditions:


- concentration: 1 OELV;
- time: 4 h;
- relative humidity:  $(50 \pm 5) \%$ ;
- temperature:  $20 \text{ }^\circ\text{C}$  to  $25 \text{ }^\circ\text{C}$ ;
- air velocity:  $0,01 \text{ m} \cdot \text{s}^{-1}$  to  $4,0 \text{ m} \cdot \text{s}^{-1}$ ;
- sampler orientation: to be tested parallel and perpendicular to the flow direction.

Analyse the set by reference to standard solutions or to samplers spiked with known amounts of analyte.

Calculate the observed mass concentration (see [8.3.3.1](#)) and plot the mean value against air velocity, assuming linear flow. Determine the air velocity corresponding to an observed mass concentration of about 90 % and 110 % of its maximal (plateau) value for each sampler orientation (see [Figure 1](#)). Test the samplers and use under conditions where air velocities are in the range of the plateau area.

As the influence of air movement on diffusive sampler performance is dependent on sampler geometry and not on the analyte selected, it is necessary to perform this test only on a given diffusive sampler using a representative chemical agent.

Samplers which are intended only for personal monitoring need to be tested only over the range from  $0,1 \text{ m} \cdot \text{s}^{-1}$  to  $1,5 \text{ m} \cdot \text{s}^{-1}$  (indoor workplaces only) or over the range from  $0,1 \text{ m} \cdot \text{s}^{-1}$  to  $4,0 \text{ m} \cdot \text{s}^{-1}$  (indoor or outdoor workplaces).

**Key**

X	air velocity around diffusive sampler	1	minimum air velocity
Y	observed mass concentration of the analyte	2	maximum air velocity
<sup>a</sup>	$\beta_a$ at plateau.		

**Figure 1 — Typical relationship between air velocity and observed mass concentration for diffusive samplers**

**8.2.3 Sampler leak test**

Expose a set of six sealed diffusive samplers to a test atmosphere under the following exposure conditions using a representative chemical agent:

- concentration: 2 OELV;
- time: 4 h;
- relative humidity:  $(50 \pm 5) \%$ ;
- temperature: 20 °C to 25 °C;
- air velocity: about  $0,5 \text{ m} \cdot \text{s}^{-1}$ .

Analyse the set to determine any leakage.

**8.2.4 Shelf life (for Type A impregnated supports)**

Store the diffusive sampler at the limits of the environmental conditions specified by the manufacturer and/or in the measuring procedure. At the end of the specified shelf-life, test the diffusive sampler under the following exposure conditions using a representative chemical agent:

- concentration: 2 OELV;
- time: 8 h;
- relative humidity:  $(80 \pm 5) \%$ ;
- temperature:  $(40 \pm 2) \text{ }^\circ\text{C}$ ;
- air velocity: above minimum specified in [8.2.2](#).

Compare with the requirement in [6.2.4](#).

### 8.2.5 Sampler identification

Perform a visual check.

### 8.2.6 Marking

Perform a visual check.

### 8.2.7 Instructions for use

Perform a visual check.

## 8.3 Measuring procedure test methods

### 8.3.1 Determination of the sampling conditions

#### 8.3.1.1 Bias due to the selection of a non-ideal sorbent

Expose diffusive samplers in two sets of at least six replicates to an atmosphere of the most volatile chemical agent for which the sampler can be used at 2 OELV and  $(80 \pm 5) \% \text{ RH}$  for 30 min. Then one set is capped, and the other set exposed to clean air [also at  $(80 \pm 5) \% \text{ RH}$ ] for a further 7,5 h.

NOTE 1 Diffusive samplers will normally be unbiased, since they are calibrated against calibration gas mixture. However, bias can result from the use of non-ideal sorbent (see [Annex A](#)) or from the effects of environmental influences, such as temperature and relative humidity. This test determines the magnitude of any bias due to back diffusion. Both sets of samplers are assumed to have been exposed to a time-weighted average concentration of 0,125 OELV for 8 h, since the test represents the worst-case situation in which a 30 min exposure occurs either at the beginning or end of an 8 h period. The difference between the mass uptake of the two sets of samplers, caused by back-diffusion, represents the maximum bias that can be encountered in a real non-constant atmosphere.

NOTE 2 If the sampler is intended for use at higher temperatures than standard temperature, the higher temperature can be used in this test.

Calculate the mean mass uptakes for the two sets of samplers and the difference, in percent (%), between the means. Compare with the requirement in [6.3.1.2](#).

#### 8.3.1.2 Determination of uptake rates

Determine the uptake rate experimentally according to [8.2.1](#) or use the nominal uptake rate provided by the manufacturer.

#### 8.3.1.3 Storage after sampling

##### 8.3.1.3.1 Direct method

Use two sets of at least six diffusive samplers and sample from a test atmosphere under the following exposure conditions:

- concentration: 0,1 OELV and 2 OELV;
- time: 8 h;
- relative humidity:  $(80 \pm 5) \%$ ;
- temperature: 20 °C to 25 °C;
- air velocity: above minimum specified in [8.2.2](#).

Analyse one set within one day and the other set after two weeks storage at room temperature, or as otherwise directed by the manufacturer.

Calculate the mean for each of the two sets of test results and the difference between the means, in percent (%). Compare with the requirement in [6.3.1.4](#). If this requirement is not met repeat the test with a shorter storage time or by using different storage conditions.

NOTE An alternative approach can be to carry out a more comprehensive set of experiments determining the recovery after a range of different storage times, for example, one day, three days, seven days, 10 days and two weeks.

### 8.3.1.3.2 Sampling media spiking method

Using two sets of at least six diffusive samplers, spike directly the sampling media with an equivalent loading as in [8.3.1.3.1](#) and add an amount of water equivalent to an exposure to air for 8 h at  $(80 \pm 5)$  % RH at a temperature of 20 °C to 25 °C for the appropriate time. The amount of water to be added can be calculated from moisture uptake rate data supplied by the manufacturer. In the absence of such data, expose samplers to clean air at 20 °C to 25 °C and  $(80 \pm 5)$  % RH before spiking with the analyte. Analyse one set within one day and the other set after two weeks storage at room temperature, or as otherwise directed by the manufacturer.

Calculate the mean for each of the two sets of test results and the difference between the means, in percent (%). Compare with the requirement in [6.3.1.4](#). If this requirement is not met repeat the test with a shorter storage time or by using different storing conditions.

## 8.3.2 Analytical procedure test methods

### 8.3.2.1 Analytical quantification limit

For type A samplers, spike ten unused diffusive samplers with appropriate masses of the analyte of interest, such that the test solutions produced from them will have mass concentrations near their respective anticipated detection limit and analyse under repeatability conditions.

For type B samplers, spike ten unused diffusive samplers with appropriate masses of the analyte of interest near its respective anticipated detection limit and analyse under repeatability conditions.

Estimate the quantification limit for each of the analytes of interest as the mean result plus ten times the standard deviation of the mean result. Compare with the requirement in [6.3.2.1](#).

### 8.3.2.2 Determination of the analytical recovery

#### 8.3.2.2.1 Sampling media spiking method from the liquid phase

Conduct the determination at four different loadings, ranging from the lowest loading to the highest loadings as indicated in [Table 1](#). Add a known mass of analyte to at least six sampling media for each loading, using a micropipette or syringe (see [7.2.2](#)) and diluting in a non-interfering solvent, if necessary. The analyte may either be applied directly to the sorbent or be allowed to diffuse from a spiked glass-fibre filter in a closed system. Desorb the analyte or a reaction product, if appropriate. Analyse the samples by reference to liquid standards prepared directly.

**Table 1 — Sample loadings for determination of analytical recovery**

Concentration	Sampling time	Sample loading
0,1 OELV	30 min	lowest (0,1 OELV × uptake rate × 30 min)
2 OELV	8 h	highest (2 OELV × uptake rate × 8 h)

Calculate the analytical recovery, by dividing the mean mass recovered at each loading by the mass applied, and the coefficient of variation of replicates. Compare with the requirement in [6.3.2.2](#).

#### 8.3.2.2.2 Phase equilibrium method (for type A non-impregnated diffusive samplers)

Prepare at least six sets of four pairs of the solutions corresponding to four different sample loadings within the range given in [Table 1](#) using the same volume of solvent used for the desorption of the samplers. Add the sorbent from an unused diffusive sampler to one solution of each pair and allow to equilibrate for at least 30 min. Analyse all solutions.

Calculate the analytical recovery by dividing the concentrations of the solutions to which sorbent has been added by the concentrations of the corresponding solutions without added sorbent and calculate the mean and the coefficient of variation of the replicate samples. Compare with the requirements given in [6.3.2](#).

If the mean analytical recovery measured by the phase equilibrium method is less than 95 % or the analytical recovery measured at any level is less than 90 %, only the test given in [8.3.2.2.1](#) shall be used.

#### 8.3.2.2.3 Sampling media spiking method (for Type B samplers)

Add a known mass of analyte to at least six sampling media at each loading, corresponding to the loadings in [8.3.2.2.1](#) and using the method described in [8.3.2.2.1](#).

Calculate the analytical recovery by dividing the mean mass recovered at each loading by the mass applied and calculate the coefficient of variation of the replicate samples. Compare with the requirement given in [6.3.2.2](#).

Type B samplers are part of the injection system of commercial thermal desorption instruments. A direct method is to compare recovery with the spiked sampler in-line versus the response from the introduction of analyte directly onto the gas chromatograph column. Absolute recovery for Type B samplers cannot normally be determined in this way unless the manufacturer of the thermal desorber has provided a direct injection facility that does not perturb any gas flow set with the sampler in-line. If a direct injection facility is not available, the following method may be used:

Load the analyte on sampling media, together with an internal standard known to have a recovery of 100 % under the applied desorption conditions. n-pentane or n-hexane are suitable. Compare the relative detector response obtained from thermal desorption with the relative response obtained by a direct liquid injection of the analyte with the internal standard.

NOTE Thermal desorption of an analyte from a Type B sampler is a non-equilibrium process. Analytical recovery is close to 100 % unless the desorption time is too short under the applied conditions of temperature and carrier gas velocity or the desorption temperature is too low, or the analyte undergoes partial decomposition due to a chemical reaction with, for example, the sorbent or its catalytic or oxidising impurities, or due to a reaction with any other material in the flow path.

#### 8.3.2.3 Determination of the blank value

Analyse six unused samplers. Calculate the mean and the standard deviation. Compare with the requirements given in [6.3.2.3](#).

When tested in accordance with [8.3.2.3](#) the blank value shall be negligible or sufficiently consistent that a correction can be applied.

NOTE The variation in a non-negligible blank will contribute to the expanded uncertainty of the measuring procedure.

### 8.3.3 Method recovery and method precision

#### 8.3.3.1 General

The method recovery and method precision tests given in 8.3.3.2 to 8.3.3.5 require calculation of the mass concentration of the analyte,  $\beta_a$ , from the mass of analyte recovered from the samplers and the volume of test atmosphere sampled by using the nominal uptake rate according to [Formula \(3\)](#):

$$\beta_a = \frac{m_d - m_b}{\dot{U}_d \times t_e \times R_{an}} \quad (3)$$

where

$m_d$  is the mass of analyte desorbed, in nanograms (ng)

$m_b$  is the mass of analyte desorbed from blank sampler, in nanograms (ng)

$\dot{U}_d$  is the uptake rate, in cubic centimetres per minute ( $\text{cm}^3 \cdot \text{min}^{-1}$ )

$t_e$  is the exposure time, in minutes (min)

$R_{an}$  is the analytical recovery

NOTE 1 If the mass concentration is given as  $10^{-6}$  (parts per million), use  $(\beta_a)'$  and  $(\dot{U}_d)'$  instead of  $\beta_a$  and  $\dot{U}_d$ .

NOTE 2 A manufacturer's value or a calculated value (see 8.2.1) can be used instead of the experimentally determined value (see [Annex C](#)) of the nominal uptake rate.

NOTE 3 The mass concentration adjusted to specified conditions,  $\beta_{a,\text{corr}}$ , for example 20 °C (= 293 K) and 101,3 kPa, can be calculated according to [Formula \(4\)](#):

$$\beta_{a,\text{corr}} = \beta_a \frac{101,3}{p_{at}} \times \frac{T_{at}}{293} \quad (4)$$

where

$p_{at}$  is the actual pressure of the test atmosphere sampled, in kilopascals (kPa)

$T_{at}$  is the temperature of the test atmosphere sampled, in Kelvin (K)

The concentration of the analyte, given as a volume fraction,  $\vartheta_a$ , can be calculated according to [Formula \(5\)](#):

$$\vartheta_a = \beta_{a,\text{corr}} \frac{24,05}{M_a} \quad (5)$$

where

$\vartheta_a$  is the volume fraction of the analyte, in microlitres per litre ( $\mu\text{l} \cdot \text{l}^{-1}$ )

24,05 is the molar volume at 293 K and 101,3 kPa, in litres per mole (l/mol);

$M_a$  is the molar mass of analyte, in grams per mole ( $\text{g} \cdot \text{mol}^{-1}$ )

For type testing the manufacturer's value of the uptake rate shall be used.

### 8.3.3.2 Effect of exposure time

Using at least six diffusive samplers, sample from a test atmosphere under the following exposure conditions:

- concentration: 1 OELV;
- time: 0,5 h, 4 h, 8 h;
- relative humidity:  $(50 \pm 5)$  %;
- temperature: 20 °C to 25 °C;
- air velocity: above minimum specified in [8.2.2](#).

NOTE This test can be part of [8.3.3.3](#).

Analyse the diffusive samplers by reference to standard solutions or to samplers spiked with known amounts of analyte. For each exposure combination, calculate the measured concentration (see [8.3.3.1](#)) for each of the six (or more) replicate diffusive samplers. Divide each by the reference concentration of the test atmosphere (see [7.4](#)). Calculate the mean method recovery for each exposure combination and the coefficient of variation of the replicate samples for each sample loading; and also calculate the overall method recovery and coefficient of variation of the means.

### 8.3.3.3 Effect of exposure concentration

Using at least six diffusive samplers for each concentration, sample from a test atmosphere under the following exposure conditions:

- concentration: 0,1 OELV, 0,5 OELV, 1 OELV and 2 OELV;
- time: 4 h;
- relative humidity:  $(50 \pm 5)$  %;
- temperature: 20 °C to 25 °C;
- air velocity: above minimum specified in [8.2.2](#).

Analyse the diffusive samplers by reference to standard solutions or to samplers spiked with known amounts of analyte. For each exposure combination, calculate the measured concentration (see [8.3.3.1](#)) for each of the six (or more) replicate diffusive samplers. Divide each by the reference concentration of the test atmosphere (see [7.4](#)). Calculate the mean method recovery and the coefficient of variation of the replicate samples for each sample loading; and also calculate the overall method recovery and coefficient of variation of the means.

### 8.3.3.4 Effect of the relative humidity of the test atmosphere

Using at least six diffusive samplers for each combination of concentration and relative humidity, sample from a test atmosphere under the following exposure conditions:

- concentration: 0,1 OELV and 2 OELV;
- time: 4 h;
- relative humidity:  $(20 \pm 5)$  % and  $(80 \pm 5)$  %;
- temperature: 20 °C to 25 °C;
- air velocity: above minimum specified in [8.2.2](#).

NOTE The high and low values of the relative humidity are given for guidance only. If it is known that the samplers are to be used in wider, or more restricted, ranges of relative humidity, the values can be adjusted accordingly.

Analyse the diffusive samplers by reference to standard solutions or to samplers spiked with known amounts of analyte. For each exposure combination, calculate the measured concentration (see 8.3.3.1) for each of the six (or more) replicate diffusive samplers. Divide each by the reference concentration of the test atmosphere (see 7.4). Calculate the mean results for each combination of concentration and relative humidity, and the differences between the means at relative humidities of  $(80 \pm 5) \%$  and  $(20 \pm 5) \%$  for each concentration, to estimate the effect of the relative humidity on method recovery.

### 8.3.3.5 Effect of the temperature of the test atmosphere

Using at least six samplers for each temperature, sample from a test atmosphere under the following exposure conditions:

- concentration: 2 OELV;
- time: 4 h;
- relative humidity:  $(50 \pm 5) \%$ ;
- temperature:  $(10 \pm 2) ^\circ\text{C}$  and  $(40 \pm 2) ^\circ\text{C}$ ;
- air velocity: above minimum specified in 8.2.2.

The minimum range of temperature is  $(10 \pm 2) ^\circ\text{C}$  to  $(30 \pm 2) ^\circ\text{C}$ . If the diffusive sampler is tested outside this range, the requirements in 6.3.3 can be fulfilled by using correction factors.

NOTE The high and low values of the temperature are given for guidance only. If it is known that the samplers are to be used in wider or more restricted ranges, the values can be adjusted accordingly.

Analyse the diffusive samplers by reference to standard solutions or to samplers spiked with known amounts of analyte. For each exposure combination, calculate the measured concentration (see 8.3.3.1) for each of the six (or more) replicate diffusive samplers. Divide each by the reference concentration of the test atmosphere (see 7.4). Calculate the mean results for each temperature and the mean difference to estimate the effect of the temperature on method recovery.

## 8.4 Uncertainty of measurement

### 8.4.1 Identification of random and non-random uncertainty components

Identify all random and non-random uncertainty components of the measuring procedure, for example, by constructing a cause and effect diagram (see ISO/IEC Guide 98-3 and References [8], [9] and [10]).

NOTE See B.1 for a list of random and non-random uncertainty components that typically need to be considered.

### 8.4.2 Estimation of individual uncertainty components

#### 8.4.2.1 General

For each of the significant uncertainty components identified in 8.4.1, estimate individual uncertainties or calculate them from experimental data as prescribed in 8.4.2.2 to 8.4.2.5, referring to the guidance in Annex B.

#### 8.4.2.2 Uncertainty associated with sampled air volume

Estimate the random and non-random uncertainty components of the sampled air volume, referring to the guidance in B.2.

If the uncertainty of measurement is being estimated for the general use of a published method, make a worst-case estimate of the uncertainty components concerned.

If the uncertainty of measurement is being estimated for the use of the method under specific conditions, for example, by a particular organisation using particular sampling equipment and a particular sampling protocol, estimate the uncertainty components for the specific equipment concerned (for example, uptake rate, timer), taking account of any specific additional requirements of the sampling protocol (for example, sampling time).

**8.4.2.3 Uncertainty associated with sample storage and transport**

Estimate the non-random uncertainty components associated with sample storage and transport, using the results of the test in [8.3.1.3](#), referring to the guidance in [B.4](#).

**8.4.2.4 Uncertainty associated with method recovery**

Estimate method bias and the non-random uncertainty components associated with method recovery, using the results of the test in [8.3.3](#), referring to the guidance in [B.5](#).

**8.4.2.5 Uncertainty associated with method variability**

Estimate the random uncertainty components associated with method variability, using the results of the test in [8.3.3.3](#), referring to the guidance in [B.6](#).

**8.4.2.6 Calculation of the combined standard uncertainty**

Calculate the combined standard uncertainty, expressed as a percentage, according to [Formulae \(6\)](#) to [\(8\)](#):

$$u_{c,r} = \sqrt{u_{s,r}^2 + u_{a,r}^2} \tag{6}$$

$$u_{c,nr} = \sqrt{u_{s,nr}^2 + u_{a,nr}^2} \tag{7}$$

$$u_c = \sqrt{u_{c,r}^2 + u_{c,nr}^2} \tag{8}$$

where

- $u_{s,r}$  is the random sampling uncertainty;
- $u_{s,nr}$  is the non-random sampling uncertainty;
- $u_{a,r}$  is the random analytical uncertainty;
- $u_{a,nr}$  is the non-random analytical uncertainty;
- $u_{c,r}$  is the combined random standard uncertainty (associated with sampling and analysis);
- $u_{c,nr}$  is the combined non-random standard uncertainty (associated with sampling and analysis); and
- $u_c$  is the combined standard uncertainty.

### 8.4.3 Calculation of expanded uncertainty

Calculate the expanded uncertainty of the measuring procedure,  $U$ , using a coverage factor  $k = 2$  according to [Formula \(9\)](#).

$$U = 2 \times u_c \quad (9)$$

NOTE An example for estimation of expanded uncertainty is given in [Annex D](#).

## 9 Test report

The test report shall include at least the following information:

- a) reference to this document (i.e. ISO 23320:2022);
- b) complete identification of the test atmosphere and its verification;
- c) the type of diffusive sampler used;
- d) the independent test method used;
- e) all validation data obtained in the tests under [8.2](#) and/or [8.3](#) as applicable and the determined values of the performance characteristics;
- f) the statistical analysis of the test results;
- g) the calculated values of the uncertainty components and the expanded uncertainty;
- h) whether the acceptance criteria are met;
- i) the level of evaluation;
- j) any unusual features noted during the determinations;
- k) any operation not included in this document that could have influenced the results;
- l) the technical justification of omitting any tests;
- m) the date of test report and testing period.

## Annex A (informative)

### Fundamentals of diffusive sampling

#### A.1 Principles

The mass of the analyte which can diffuse to a suitable sorbent within a certain time, i.e. the mass uptake of a diffusive sampler,  $m_s$ , in nanograms (ng), is determined by [Formula \(A.1\)](#) which is derived from Fick's first law of diffusion:

$$m_s = \frac{A \times D_a \times (\beta_{a1} - \beta_{a2}) \times t_e}{l} \quad (\text{A.1})$$

where

- $A$  is the cross-sectional area of sorption surface, in square centimetres (cm<sup>2</sup>)
- $D_a$  is the diffusion coefficient of an analyte, in square centimetres per minute (cm<sup>2</sup> · min<sup>-1</sup>)
- $\beta_{a1}$  is the mass concentration of the given analyte at the beginning of the diffusion layer (i.e. at the distance  $l$  from the sorption surface), in milligrams per cubic metre (mg · m<sup>-3</sup>)
- $\beta_{a2}$  is the mass concentration of the given analyte at the end of the diffusion layer (i.e. at the sorption surface), in milligrams per cubic metre (mg · m<sup>-3</sup>)
- $t_e$  is the exposure time, in minutes (min)
- $l$  is the length of static air layer in sampler (or equivalent for permeation types), in centimetres (cm)

Ideally  $\beta_{a1}$  is equal to the mass concentration of the given analyte in the air outside the diffusive sampler and  $\beta_{a2}$  equals zero ("zero sink"-condition). In that case the magnitude of the compound-specific parameter,  $\frac{A \times D_a}{l}$  (see [Formula \(A.1\)](#)), is dependent only on the diffusion coefficient of the given analyte and on the geometry of the diffusive sampler used.

However, in practice the parameters  $\beta_{a1}$  and  $\beta_{a2}$ , as well as the temperature, relative humidity, external air movement and also the presence of other compounds can influence the uptake rate. In particular, if external air movement is insufficient, diffusive uptake into the sampler can reduce the external concentration and hence decrease the concentration gradient and increase the effective value of  $l$  (see air velocity/sampler orientation test in [8.2.2](#)). Also, if a non-ideal sorbent is used,  $\beta_{a2}$  will be non-zero, except for  $t_e = 0$ , and again the sampling rate will be reduced (see [A.3](#)), hence the back-diffusion test (see [8.3.1.1](#)).

For type A and type B diffusive samplers, analytical recovery is an additional factor which can influence their suitability.

A special form of dependence of the uptake rate on the mass concentration  $\beta_{a1}$  can appear, as a result of the influence of the kinetics of the reaction between a given analyte and applied reagent, and also as a result of sorption of non-reacted analyte or as a result of side reactions.

NOTE More details on principles are given in EN 13528-3.

## A.2 Dimensions of uptake rate

For a given mass concentration  $\beta_a$ , in milligrams per cubic metre ( $\text{mg} \cdot \text{m}^{-3}$ ) of gas or vapour, the diffusive nominal uptake rate  $\dot{U}_d$ , in cubic centimetres per minute ( $\text{cm}^3 \cdot \text{min}^{-1}$ ), is given by [Formula \(A.2\)](#):

$$\dot{U}_d = \frac{m_s}{\beta_a \times t_e} \quad (\text{A.2})$$

where

$m_s$  is the mass of the analyte which can diffuse to a suitable sorbent within a certain time, i.e. the mass uptake of a diffusive sampler, in nanograms (ng)

$\beta_a$  is the mass concentration of the analyte in the calibration gas mixture, in milligrams per cubic metre ( $\text{mg} \cdot \text{m}^{-3}$ )

$t_e$  is the exposure time, in minutes (min)

NOTE 1 The diffusive nominal uptake rate  $\dot{U}_d$  is equivalent to the active sampling flow in cubic centimetres per minute ( $\text{cm}^3 \cdot \text{min}^{-1}$ ), at which active sampling would collect the same amount of molecules of a given gas.  $\dot{U}_d$  depends on the diffusion coefficient of the analyte and the geometry of the considered sampler.

NOTE 2 Although the uptake rate has dimensions of cubic centimetres per minute ( $\text{cm}^3 \cdot \text{min}^{-1}$ ) this is actually a reduction of nanograms per milligrams per cubic metre per minute ( $\text{ng} \cdot \text{mg}^{-1} \cdot \text{m}^{-3} \cdot \text{min}^{-1}$ ) and does not indicate a real volumetric flow of air.

NOTE 3 Some manufacturers quote diffusive uptake in  $\text{ng} \cdot \text{ppm}^{-1} \cdot \text{min}^{-1}$ . These are practical dimensions, since most occupational hygienists use parts per million ( $10^{-6}$ ) for concentrations of gases and vapours. Uptake rates expressed in these dimensions are also relatively unaffected by temperature and pressure. Thus, for a given volume fraction  $\varphi_a$ , given in parts per million, of gas or vapour, the uptake rate is given by [Formula \(A.3\)](#):

$$(\dot{U}_d)' = \frac{m_s}{\varphi_a \times t_e} \quad (\text{A.3})$$

where  $\varphi_a$  is the volume fraction of the analyte.

NOTE 4  $\dot{U}_d$  in cubic centimetres per minute ( $\text{cm}^3 \cdot \text{min}^{-1}$ ) and  $(\dot{U}_d)'$ , the uptake rate at ambient conditions, in  $\text{ng} \cdot \text{ppm}^{-1} \cdot \text{min}^{-1}$ , are related by [Formula \(A.4\)](#).

$$(\dot{U}_d)' = \dot{U}_d \times \frac{M_a}{24,05} \times \frac{293}{T_{at}} \times \frac{p_{at}}{101,3} \quad (\text{A.4})$$

where

$M_a$  is the molar mass of analyte, in grams per mole ( $\text{g} \cdot \text{mol}^{-1}$ )

$p_{at}$  is the actual pressure of the test atmosphere sampled, in kilopascals (kPa)

$T_{at}$  is the temperature of the test atmosphere sampled, in Kelvin (K)

### A.3 Bias due to the selection of a non-ideal sorbent

The performance of a diffusive sampler depends critically on the selection and use of a sorbent or collection medium which has high sorption efficiency. The residual vapour pressure of the sampled compound at the sorption surface will then be very small in comparison to the ambient pressure, and the observed uptake rate will be close to its ideal steady-state value, which can usually be calculated from the geometry of the sampler and the diffusion coefficient of the analyte in air.

STANDARDSISO.COM : Click to view the full PDF of ISO 23320:2022

## Annex B (informative)

### Estimation of uncertainty of measurement

#### B.1 General

Methods for measurement of chemical agents in workplace atmospheres usually involve two major steps: sampling and analysis. The following is a typical, but non-exclusive, list of random and non-random uncertainty components:

- a) sampling
  - 1) uncertainty associated with mass uptake (see [B.2](#));
  - 2) uncertainty associated with sampling efficiency (see [B.3](#));
  - 3) uncertainty associated with sample storage and transport (see [B.4](#));
- b) analysis
  - 1) uncertainty associated with method recovery (see [B.5](#));
  - 2) uncertainty associated with method variability (see [B.6](#));
  - 3) uncertainty associated with the calibration (see [B.6.3](#) and [B.6.4](#));
  - 4) uncertainty associated with instrument response drift (see [B.6.6](#)).

#### B.2 Uncertainty associated with mass uptake

##### B.2.1 Sources of uncertainty

For diffusive sampling, mass uptake has the following sources of uncertainty:

- uptake rate (see [B.2.2](#)); and
- sampling time (see [B.2.3](#)).

##### B.2.2 Uptake rate

The random and non-random uncertainty components associated with the uptake rate should be estimated from the results of the replicate samples collected from a test atmosphere at relative humidity of  $(50 \pm 5) \%$  and a temperature of  $20 \text{ }^\circ\text{C}$  to  $25 \text{ }^\circ\text{C}$  (see [8.3.1.2](#)) corrected by analytical recovery, if applicable.

The random and non-random uncertainty components associated with the uptake rate are given by [Formula \(B.1\)](#):

$$u_{\text{ur}} = \sqrt{\frac{(K_{\text{v,r}})^2}{n} + (u_{\text{rc}})^2} \quad (\text{B.1})$$

where

- $u_{ur}$  is the relative standard uncertainty associated with the uptake rate, in percent (%);
- $K_{v,r}$  is the coefficient of variation of the replicate samples collected from the test atmosphere, in percent (%);
- $n$  is the number of replicate samples;
- $u_{rc}$  is the relative standard uncertainty of the reference concentration of the test atmosphere, in percent (%).

### B.2.3 Sampling time

Sampling time can be measured very exactly with a radio-controlled clock, a quartz clock or stopwatch. The major source of uncertainty in measurement of sampling time is the accuracy with which the reading is taken, i.e. to the nearest minute or second.

If the reading is taken to the nearest second, the non-random uncertainty component is very small for both long-term and short-term measurements and can be negligible. If the reading is taken to the nearest minute, the non-random component is very small for long-term measurements (for example, more than 2 h) and can be disregarded, but for short-term measurements it needs to be taken into account.

For example, if time is recorded to the nearest minute, the coefficient of variation is 1,4 % for a sampling time of 30 min (summing the maximum 0,5 min biases at the start and end of the sampling period and dividing by the sampling time and  $\sqrt{6}$ , assuming a triangular probability distribution).

## B.3 Uncertainty associated with sampling efficiency

### B.3.1 Back diffusion

The adsorption of gases and vapours on diffusive samplers can be influenced by the pressure, relative humidity and temperature of the sampled air and the concentration of chemical agents in the sampled air. These factors can affect the adsorption capacity, the performance of the adsorption process and the uptake rate.

For diffusive samplers, back diffusion can occur if there is significant variation in the air concentration of the analyte during the sampling time, in which case the sampling efficiency could be less than 100 % and the uncertainty of the back diffusion needs to be taken into account.

Assuming a rectangular probability distribution, the uncertainty associated with back diffusion is given by [Formula \(B.2\)](#):

$$u_{bd} = \frac{\Delta_{bd}}{\sqrt{3}} \tag{B.2}$$

where

- $u_{bd}$  is the relative standard uncertainty associated with back diffusion;
- $\Delta_{bd}$  is the difference between the mean results of replicate samples analysed in [8.3.1.1](#), in percent (%).

### B.3.2 Exposure time

The non-random uncertainty component associated with exposure time can be estimated by the analysis of replicate samples collected from a test atmosphere (see 8.3.3.2). Assuming a rectangular probability distribution, the uncertainty associated with exposure time is given by [Formula \(B.3\)](#):

$$u_{te} = \frac{\Delta_{te}}{\sqrt{3}} \quad (\text{B.3})$$

where

$u_{te}$  is the relative standard uncertainty associated with exposure time;

$\Delta_{te}$  is the highest difference between the mean results of replicate samples collected from the test atmosphere at different exposure times, in percent (%).

### B.4 Uncertainty associated with sample storage and transport

The non-random uncertainty component associated with sample storage and transport can be estimated by the analysis of replicate samples collected from a test atmosphere or prepared by spiking sampling collection media with the chemical agent of interest (see 8.3.1.3).

Assuming a rectangular probability distribution, the uncertainty associated with sample storage is given by [Formula \(B.4\)](#):

$$u_{st} = \frac{\Delta_{st}}{\sqrt{3}} \quad (\text{B.4})$$

where

$u_{st}$  is the relative standard uncertainty associated with sample storage;

$\Delta_{st}$  is the difference between the mean results of replicate samples analysed immediately after sampling or preparation and replicate samples analysed after the maximum storage time, in percent (%).

When samples are transported in an appropriate manner as specified in the measuring procedure, it is not necessary to take into consideration any component of uncertainty other than those associated with storage.

### B.5 Uncertainty associated with method recovery

#### B.5.1 General

Method recovery is influenced by several factors (see [B.3](#)). The study of their influence is carried out following the tests described in [8.3.3.2](#), [8.3.3.3](#), [8.3.3.4](#) and [8.3.3.5](#) by the use of test atmospheres.

The experimental data collected when carrying out these tests give representative information about the factors causing variation and bias (relative to a reference value) that occur in routine applications of the specified method of measurement, for example, concentration, temperature and humidity. These data can be used to estimate the method uncertainty components as described in [B.5.2](#) to [B.5.6](#).

Measurement procedures for chemical agents using diffusive samplers usually prescribe the correction of the results for analytical recovery. In this case, method recovery is estimated from the results of the samples taken from the test atmospheres corrected for analytical recovery.

## B.5.2 Analytical recovery

**B.5.2.1** Analytical recovery can be calculated from the results of the analysis of replicate samples with known mass of the compound of interest (known samples), dividing the mass of analyte recovered by the known mass. The known samples can be certified reference materials (CRMs) or sampling media spiked samples at several loadings covering the range of the application of the method.

The random and non-random uncertainty components associated with the analytical recovery can be estimated from the results of the test in [8.3.2.2](#)

**B.5.2.2** When the results are corrected for analytical recovery and it is not level dependent, the random uncertainty component associated with this correction is given by [Formula \(B.5\)](#):

$$u_{ar} = \sqrt{\frac{(K_{v,rks})^2}{n} + (u_{ks})^2} \quad (\text{B.5})$$

where

- $u_{ar}$  is the relative standard uncertainty of the analytical recovery, in percent (%);
- $K_{v,rks}$  is the coefficient of variation of the replicate known samples, in percent (%);
- $n$  is the number of replicate samples;
- $u_{ks}$  is the relative standard uncertainty of the known samples, in percent (%).

When CRMs are available,  $u_{ks}$  should be estimated from the CRM certificate.

**EXAMPLE** The uncertainty of spiked samples with pure compound, assuming that the effect of temperature on the dispensed volume is negligible, can be estimated by [Formula \(B.6\)](#):

$$u_{ks} = \sqrt{(u_{pu})^2 + \frac{(B_{max,sy})^2}{3} + \frac{(K_{v,sy})^2}{n}} \quad (\text{B.6})$$

where

- $u_{pu}$  is the relative standard uncertainty of the purity of analyte, in percent (for example, if the purity is  $\geq 99\%$ , the relative uncertainty is  $((100 - 99)/\sqrt{3})\%$ );
- $B_{max,sy}$  is the maximum bias of the volume dispensed by the syringe used to spike the blank sampling media, in percent (%);
- $K_{v,sy}$  is the coefficient of variation of the volume dispensed by the syringe used to spike the blank sampling media, in percent (%);
- $n$  is the number of replicate samples

If the amount of analyte spiked to the blank sampling media is weighed, the uncertainty of the nominal value is estimated as a combination of the uncertainty of the balance used and the coefficient of variation of the reading of the weight.

**B.5.2.3** If analytical recovery correction is not applied to the results, the analytical bias should be estimated and treated as an uncertainty component. The non-random uncertainty component associated with incomplete recovery can be estimated as the difference of the mean analytical recovery

of the replicate samples of all concentrations from 100 % and converted to a standard uncertainty. The relative standard uncertainty of the analytical bias is given by [Formula \(B.7\)](#), see Reference [11].

$$u_{ab} = \sqrt{\left(\frac{B_a}{k}\right)^2 + \frac{(K_{v,rks})^2}{n} + (u_{ks})^2} \quad (\text{B.7})$$

where

- $u_{ab}$  is the relative standard uncertainty of the analytical bias, in percent (%);
- $B_a$  is the bias of the mean results of replicate analyses for the known samples, in percent (%);
- $k$  is the coverage factor used in the calculation of the expanded uncertainty;
- $K_{v,rks}$  is the coefficient of variation of the replicate known samples, in percent (%);
- $n$  is the number of replicate samples;
- $u_{ks}$  is the relative standard uncertainty of the known samples, in percent (%).

When CRMs are available  $u_{ks}$  should be estimated from the CRM certificate.

### B.5.3 Method bias

Method bias can be estimated from the results of the replicate samples collected from a test atmosphere at relative humidity of  $(50 \pm 5)$  % and temperature of 20 °C to 25 °C (see [8.3.3.3](#)) corrected for analytical recovery, if applicable. When the method bias is significant, bias is estimated and treated as an uncertainty component.

The non-random uncertainty component can be estimated as the difference, in percent (%), of the mean results of the replicate samples from 100 %. The relative standard uncertainty associated with the method bias is given by [Formula \(B.8\)](#), see Reference [11].

$$u_{mb} = \sqrt{\left(\frac{B_m}{k}\right)^2 + \frac{(K_{v,r})^2}{n} + (u_{rc})^2} \quad (\text{B.8})$$

where

- $u_{mb}$  is the relative standard uncertainty associated with the method bias, in percent (%);
- $B_m$  is the bias of the mean results of the reference concentration, in percent (%) (see [7.4.2](#));
- $k$  is the coverage factor used in the calculation of the expanded uncertainty;
- $K_{v,r}$  is the coefficient of variation of the replicate samples collected from the test atmosphere, in percent (%);
- $n$  is the number of replicate samples;
- $u_{rc}$  is the relative standard uncertainty of the reference concentration of the test atmosphere, in percent (%), see [B.5.4](#).

### B.5.4 Reference concentration

In a properly designed and performed experiment, the random and non-random uncertainty components associated with the test atmosphere concentrations are expected to be small. They depend on the system used for generation and can be calculated by a propagation of errors from the uncertainty

of the parameters of the test atmosphere generation. For example, for a dynamic system the random uncertainty associated with the reference concentration of the test atmosphere is usually less than 3 %.

### B.5.5 Effect of relative humidity

The non-random uncertainty component associated with the effect of relative humidity can be estimated from the difference between the mean results of replicate samples collected from the test atmospheres at relative humidities of  $(20 \pm 5)$  % and  $(80 \pm 5)$  % (see [8.3.3.4](#)).

Assuming a rectangular probability distribution, the uncertainty associated with the effect of relative humidity is given by [Formula \(B.9\)](#):

$$u_h = \frac{\Delta_h}{\sqrt{3}} \quad (\text{B.9})$$

where

$u_h$  is the relative standard uncertainty associated with effect of relative humidity on the recovery;

$\Delta_h$  is the higher of the differences between the mean results of replicate samples collected from the test atmospheres at relative humidities of  $(20 \pm 5)$  % and  $(80 \pm 5)$  %, in percent (%).

### B.5.6 Effect of temperature

The non-random uncertainty component associated with the effect of temperature can be estimated from the difference between the mean results of replicate samples collected from the test atmospheres at temperatures of  $(10 \pm 2)$  °C and  $(40 \pm 2)$  °C (see [8.3.3.5](#)).

Assuming a rectangular probability distribution, the uncertainty associated with the effect of temperature is given by [Formula \(B.10\)](#):

$$u_T = \frac{\Delta_T}{\sqrt{3}} \quad (\text{B.10})$$

where

$u_T$  is the relative standard uncertainty associated with effect of temperature on the recovery;

$\Delta_T$  is the differences between the mean results of replicate samples collected from the test atmospheres at temperatures of  $(10 \pm 2)$  °C and  $(40 \pm 2)$  °C, in percent (%).

## B.6 Uncertainty associated with method variability

### B.6.1 General

The uncertainty associated with method variability can be estimated from method precision data obtained from the results of the replicate samples collected from the test atmospheres used in [8.3.3.2](#) as described in [B.6.2](#). Separate uncertainty estimates need to be made for any sources of systematic error, where applicable, for example non-random uncertainty associated with the concentration of calibration solutions (see [B.6.3](#)), calibration function (see [B.6.4](#)), dilution of the sample solutions (see [B.6.5](#)) and instrument response drift (see [B.6.6](#)).

The uncertainty associated with analytical variability is included in the method variability.

Independent uncertainty estimates associated with analytical variability can be made either from analytical precision data either obtained under repeatability conditions (see [B.6.7.1](#)) or from data obtained under reproducibility conditions (see [B.6.7.2](#)). In both cases, separate uncertainty estimates need to be made for any sources of systematic error, where applicable, for example non-random

uncertainty associated with the concentration of calibration solutions (see B.6.3), calibration function (see B.6.4), dilution of the sample solutions (see B.6.5) and instrument response drift (see B.6.6). When the analytical precision is determined from within-laboratory reproducibility data, for example, using quality control data, most random and randomized uncertainty components of the analytical variability are included. See ISO 21748 for further guidance. When within-laboratory reproducibility data are used the values obtained for the analytical precision can be higher than when repeatability data are used because, in this case, between days precision are included.

## B.6.2 Method precision

Method precision can be calculated from the results of the replicate samples collected from a test atmosphere at a relative humidity of  $(50 \pm 5) \%$  and a temperature of  $20 \text{ }^\circ\text{C}$  to  $25 \text{ }^\circ\text{C}$  (see 8.3.3.3).

The random uncertainty component can be estimated by Formula (B.11), see also ISO/IEC Guide 98-3:

$$u_{\text{mp}} = \sqrt{(K_{\text{v,m}})^2 + \left(1 - \frac{1}{n}\right) + (K_{\text{vp,r}})^2} \quad (\text{B.11})$$

where

- $u_{\text{mp}}$  is the relative standard uncertainty associated with the method precision, in percent (%);
- $K_{\text{v,m}}$  is the coefficient of variation of the means, in percent (%);
- $n$  is the number of replicate samples, see also Formula (B.12);
- $K_{\text{vp,r}}$  is the pooled coefficient of variation of the replicate samples, in percent (%), calculated according to Formula (B.13).

For an unequal number of replicate samples  $n$  can be estimated by Formula (B.12), see Reference [12]:

$$n = \frac{N^2 - \sum_{j=1}^J (n_j)^2}{(J-1)N} \text{ with } N = \sum_{j=1}^J n_j \quad (\text{B.12})$$

where

- $N$  is the total number of replicate samples at all concentration levels;
- $J$  is the number of concentration levels;
- $n_j$  is the number of replicate samples at the concentration level  $j$ .

For both equal and unequal numbers of replicate samples  $K_{\text{vp,r}}$  can be estimated by Formula (B.13):

$$K_{\text{vp,r}} = \sqrt{\frac{[(n_1-1) \times (K_{\text{v},1})^2] + [(n_2-1) \times (K_{\text{v},2})^2] + [(n_3-1) \times (K_{\text{v},3})^2] + [(n_4-1) \times (K_{\text{v},4})^2]}{(n_1-1) + (n_2-1) + (n_3-1) + (n_4-1)}} \quad (\text{B.13})$$

where

- $K_{\text{v},1}, K_{\text{v},2}, K_{\text{v},3}, K_{\text{v},4}$  are the coefficients of variation at the four tested concentrations;
- $n_1, n_2, n_3, n_4$  are the numbers of replicate samples at each test concentration.

NOTE When the number of replicate samples at all concentration levels are equal, then  $n = n_j$  and

$$K_{\text{vp,r}} = \sqrt{\frac{[(K_{\text{v},1})^2 + (K_{\text{v},2})^2 + (K_{\text{v},3})^2 + (K_{\text{v},4})^2]}{4}}$$

### B.6.3 Concentration of calibration solutions

The non-random uncertainty components associated with the concentration of the calibration solutions can be estimated from one or more of the following:

- a) the certificate provided by the manufacturer of a pressurised test gas;
- b) the purity of the starting material (for example, purity of more than 99,5 %);
- c) the uncertainties in the weighing of chemical agents and solutions, i.e. the uncertainty of a balance;
- d) the uncertainties in the preparation of a test gas;
- e) the random uncertainty components associated with a dilution procedure.

**EXAMPLE** The relative standard uncertainty associated with the concentrations of the calibration solutions,  $u_{cc}$ , in percent (%), can be estimated from the uncertainty of the mass of the pure compound and the uncertainty of the micropipette or syringe used to prepare the calibration solutions, using [Formula \(B.14\)](#):

$$u_{cc} = \sqrt{(u_m)^2 + (K_{v, sy})^2 + \frac{(B_{max, p})^2}{3}} \quad (\text{B.14})$$

where

$u_{cc}$  is the relative standard uncertainty of the mass of pure compound weighed, estimated from its purity, the calibration certificate of the balance and the coefficient of variation of the balance readings, in percent (%);

$K_{v, sy}$  is coefficient of variation of the micropipette or syringe used to prepare the calibration solutions, in percent (%), for example, estimated from the certificate provided by the manufacturer;

$B_{max, p}$  is the maximum bias of the micropipette or syringe used to prepare the calibration solutions, in percent (%), for example, estimated from the confidence interval given on the certificate provided by the manufacturer.

### B.6.4 Calibration function

The random uncertainty component associated with the calibration function can be calculated from parameters obtained by the least-squares linear regression. See Reference [9].

2 % is a reasonable estimate of the random uncertainty component associated with the calibration function and can be used in most cases. This was the value used in the EU project BC/CEN/ENTR/000/2002-16 Analytical methods for chemical agents, see Reference [13].

### B.6.5 Dilution of the sample solutions (if applicable)

If sample solutions are diluted before analysis it is necessary to take into consideration the random and non-random uncertainty components associated with the dilution process.

The random uncertainty component is the relative uncertainty of the solution volume dispensed by the micropipette used in dilution of the sample solutions.