
Sharps injury protection — Requirements and test methods — Sharps protection features for single-use hypodermic needles, introducers for catheters and needles used for blood sampling

Protection contre les blessures par perforants — Exigences et méthodes d'essai — Dispositifs de protection des aiguilles hypodermiques, des introducteurs pour cathéters et des aiguilles utilisées pour les prélèvements sanguins, non réutilisables

STANDARDSISO.COM : Click to view full PDF of ISO 23908:2011



STANDARDSISO.COM : Click to view the full PDF of ISO 23908:2011



COPYRIGHT PROTECTED DOCUMENT

© ISO 2011

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction.....	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Requirements.....	2
4.1 General	2
4.2 Activation of the sharps injury protection feature	3
4.3 Security of safe mode protection	3
5 Test methods	3
5.1 General	3
5.2 Testing activation of a sharps injury protection feature	3
5.2.1 Principle.....	3
5.2.2 Apparatus	4
5.2.3 Procedure.....	4
5.3 Challenging the device in safe mode	5
5.3.1 General	5
5.3.2 Principle.....	5
5.3.3 Apparatus	5
5.3.4 Procedure	5
5.4 Testing access to the sharp in safe mode	6
5.5 Testing simulated clinical use	6
5.6 Test report.....	6
6 Information supplied by the manufacturer	6
6.1 General	6
6.2 Marking/labelling	7
6.3 Instructions for use.....	7
Annex A (informative) Guidance on simulated user studies	8
Annex B (informative) Method for testing access to the sharp in safe mode	10
Bibliography.....	11

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 23908 was prepared by Technical Committee ISO/TC 84, *Devices for administration of medicinal products and intravascular catheters*.

STANDARDSISO.COM : Click to view the full PDF of ISO 23908:2011

Introduction

This International Standard addresses sharps injury protection systems designed to protect users of medical devices. These sharps injury protection features are intended to prevent, or reduce the potential for, disease transmission which could result from accidental, post-use sharps injuries.

This International Standard is aimed at addressing devices primarily intended for human use, of a wide range of product types, including, among others, hollow-bore needles for injection or infusion of therapeutics into the body, or sampling of fluids from the body, and hollow bore or solid-core needles used for blood sampling (e.g. lancing devices). It addresses sharps injury protection systems which are either active or passive in their activation after the medical device's intended use. It does not cover solid-core needles used for surgery (e.g. suture needles).

Given the broad variation in product design and sharps protection technology, the variety of different types of devices, and in order to avoid unnecessarily restricting innovation, this International Standard has been developed as "horizontal" in nature, which means it provides for general design, testing and labelling requirements, rather than specific physical and prescriptive design requirements. It therefore differs from more "vertical" standards, which list specific maximum forces, detailed test fixture designs, test systems to be used or detailed test measures, as such prescriptive details cannot cover the variety of designs and devices, and may impede continuing innovation in new products, features and/or protection mechanisms that lead to future improvements in healthcare.

This International Standard presumes that the product developer would use a risk-based approach (consistent with ISO 14971) to determine the device design that best meets the needs of a target user population and expected use settings. Through this risk-based approach, the sharps injury protection system would have performance requirements appropriate to the foreseeable risks associated with the intended use of the device, expected user interfaces, and the settings in which these safety features are expected to be used.

This International Standard provides guidelines to enable the manufacturer to verify that the design of the sharps injury protection systems complies with the design intent spelled out in the design specification. As part of this verification, the manufacturer is expected to demonstrate that the performance of the sharps injury protection system is appropriate to the intended users and settings through the use of appropriate simulated or clinical use studies. These simulated or clinical use studies allow the manufacturer to demonstrate that, when used in accordance with the instructions for use, in settings representative of real-life intended use and by intended or foreseeable users, the device functions as intended.

Existing products and those currently under development may not fulfil some of the requirements given by this International Standard. However, manufacturers would be well advised to follow its provisions when improving existing products or developing new products to obtain an even higher level of quality.

[STANDARDSISO.COM](https://standardsiso.com) : Click to view the full PDF of ISO 23908:2017

Sharps injury protection — Requirements and test methods — Sharps protection features for single-use hypodermic needles, introducers for catheters and needles used for blood sampling

1 Scope

This International Standard gives requirements and test methods for evaluating the performance parameters of sharps injury protection features, whether active or passive in design, for medical devices containing (sharp) hypodermic needles for single use, introducers for catheters and lancets, and other needles used in blood sampling. The sharps injury protection devices it covers may be provided integral to the device or combined with the device prior to use to achieve the sharps injury protection.

It does not give requirements for the storage and handling of the sharps protection before its intended use, or for the medical device itself.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 2859 (all parts), *Sampling procedures for inspection by attributes*

ISO 3951 (all parts), *Sampling procedures for inspection by variables*

ISO 14971, *Medical devices — Application of risk management to medical devices*

ISO 16269-6, *Statistical interpretation of data — Part 6: Determination of statistical tolerance intervals*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

activation

deployment of the sharps protection mechanism

3.2

active safety feature

sharps protection feature that requires an additional step by the user to activate, separate from any action needed to perform the primary intended function of the device

3.3

accidental sharps injury

unintentional penetration into human tissue by the sharp after the intended use

3.4
passive safety feature
sharps protection feature that does not require an additional step by the user to activate, separate from any action needed to perform the primary intended function of the device

3.5
safe mode
state of the device after activation of the safety feature

3.6
sharp
part of the device that can penetrate human tissue

3.7
sharps injury protection feature
feature that prevents accidental sharps injury

4 Requirements

4.1 General

4.1.1 Where the requirements do not specify forces for activation of the safety feature, the appropriate force shall be determined by using a risk-based approach in accordance with ISO 14971, supported by simulated user studies that mimic actual clinical use by using patient substitutes (e.g. instructional models) rather than actual patients. The study design should be based on statistical considerations and should have clear acceptance criteria. Guidance on conducting simulated user studies is outlined in Annex A.

4.1.2 Once in safe mode, the safety feature(s) of the device shall provide protection against accidental sharps injury until safe disposal of the sharp under expected conditions of use.

4.1.3 It shall be apparent to the user as to when the device is in safe mode.

Activation/safe mode shall be communicated to the user in a clear and unmistakable manner by either visual, tactile and/or audible means. If the manufacturer determines that the user environment requires a permanent indication of safe mode, then a visual indication shall be included.

4.1.4 Activation of the sharps protection feature shall permit the user's hand(s) to remain behind the exposed contaminated sharp.

Safety features may be operated either actively or passively. If active operation is required, one-handed operation is recommended.

4.1.5 The safety result shall

- not negatively affect the intended performance characteristics or proper disposal of the device,
- not impede or adversely affect the intended clinical performance of the device,
- resist inadvertent activation under expected conditions of use.

4.1.6 The performance of the safety feature as described in 4.1.2 to 4.1.5 shall be demonstrated through appropriate simulated or clinical use studies for the specified conditions indicated under the conditions of use.

NOTE 1 Appropriate simulated or clinical use studies may be helpful in establishing specifications to meet the requirements of Clause 5.

NOTE 2 Annex A contains guidance for simulated or clinical use studies.

NOTE 3 IEC 62366 covers the application of usability engineering to medical devices.

4.2 Activation of the sharps injury protection feature

Active sharps injury protection features shall be able to be activated immediately after intended use.

Passive sharps injury protection features shall enter safe mode immediately after intended use.

The sharps injury protection feature shall be able to be activated by a force appropriate for the intended users of the device (e.g. patients, health care professionals or family members). An appropriate force shall be selected that eases actuation and avoids unintended actuation.

The appropriate activation forces shall be determined using a risk-based approach in accordance with ISO 14971. The manufacturer shall confirm that these force values are the values at which the sharps injury protection feature can be activated. These force values shall be obtained using the methodology outlined in Clause 5.

4.3 Security of safe mode protection

Once in safe mode, the safety feature shall

- a) resist forces so as to prevent unintended exposure to the sharps, when tested in accordance with 5.3, and
- b) minimize the risk of accidental access to the sharp, when tested in accordance with 5.4.

Using a risk-based approach in accordance with ISO 14971, the manufacturer shall determine appropriate minimum overriding forces. These force values shall be obtained using the methodology outlined in Clause 5.

5 Test methods

5.1 General

Unless otherwise specified in the relevant device standard(s), all tests and test evaluations shall be performed at the following standard atmosphere conditions:

- temperature: $(23 \pm 5) ^\circ\text{C}$;
- relative humidity: $(50 \pm 25) \%$.

The device with integrated sharps injury protection or stand-alone sharps injury protection device that is tested shall have been subjected to storage for at least 4 h under these conditions immediately prior to testing/evaluation.

The repeatability and reproducibility of the test apparatus shall be no greater than 20 % of the allowed tolerance band for any given set of measurements.

When a sharps protection means is integral to a device covered by any other standard, or when combined with such a device prior to use, it shall be subjected to the same preconditioning requirements set out for the device by that other standard.

5.2 Testing activation of a sharps injury protection feature

5.2.1 Principle

Test pieces shall be chosen to test the intended means of activation by either tensile, compressive or torsional force applied directly (but smoothly) to the safety mechanism. The resulting activation forces/torques shall be recorded as specified in 5.6.

5.2.2 Apparatus

A test apparatus shall be used that can hold the device firmly, can actuate the safety feature at various speeds and at appropriate angles, and that can display the resulting activation force/torque in a reproducible and repeatable manner. The safety feature shall be confirmed to activate consistently when set up as per the instructions for use.

5.2.3 Procedure

Set up the number, n , of devices (see Annex A) outlined in the instructions for use and test as follows.

- a) Insert the device in a fixture.
- b) Prepare the fixture as required.
- c) Start the test cycle.
- d) Read the force gauge.
- e) Record the test results. Attribute sampling shall be conducted in accordance with ISO 2859 and variable sampling shall be conducted in accordance with ISO 3951.
- f) Repeat for n devices (as specified).

Test values obtained shall be used to calculate a statistical tolerance (see ISO 16269-6) using the methodology outlined below.

If it is assumed that the peak force values of activation of the sharps injury protection feature obtained in accordance with steps a) to f) are independent from each other, and normally distributed, then the statistical tolerance around the mean of peak force values can be evaluated. The 95 % confidence level, two-sided statistical tolerance interval can be calculated using the average (\bar{x}) plus or minus the standard deviation (s) multiplied by a tolerance limit factor (k):

$$\bar{x} \pm k \times s$$

where

- \bar{x} is the average of the sample values;
- k is the tolerance limit factor;
- s is the standard deviation of the sample values.

The factor (k) is determined based upon the confidence level (95 %), probability content (p), and the number of measurements (n) taken as per ISO 16269-6.

Test force values satisfy the requirement when, for a given test set, the following expressions are fulfilled:

$$\bar{x} + (k \times s) \leq \text{USL}$$

and

$$\bar{x} - (k \times s) \geq \text{LSL}$$

where USL and LSL are the upper and lower specification limits, respectively. These specification limits are to be determined from the risk assessment, including human factors considerations.

NOTE ISO 16269-6 also addresses one-sided tolerances and other non-normal distributions.

5.3 Challenging the device in safe mode

5.3.1 General

Tests to challenge the sharps injury protection feature after activation or in locked mode shall be carried out using the following method.

5.3.2 Principle

Using a risk-based approach in accordance with ISO 14971, the manufacturer shall determine minimum overriding or unlocking forces. The manufacturer shall confirm that these force values are the values at which the sharps injury protection feature cannot be overridden once in safe mode. These force values shall be obtained with the methodology outlined in 5.2.3 steps a) to f). These force values shall be used to calculate a statistical tolerance (see ISO 16269-6) with a similar methodology to that given in 5.2.

As specified in Clause 4, the manufacturer shall demonstrate that the sharps injury protection feature can withstand overriding forces once in safe mode, appropriate to the target population for which the device is intended and for any other individuals (e.g. health care professionals or family members who administer the therapeutics) who may incidentally come in contact with the locked device prior to its safe disposal.

Test pieces are chosen and a force/torque applied to the safety mechanism in a manner consistent with the failure modes identified in the risk assessment. The resulting unlocking or overriding forces are then recorded as specified in 5.6.

5.3.3 Apparatus

The test apparatus used shall

- hold the device firmly,
- challenge the safety feature at suitable or appropriate speed and angles, and
- display the resulting unlocking force in a reproducible and repeatable manner.

The safety feature shall be engaged or in safe mode as per the instructions for use prior to testing.

5.3.4 Procedure

Activate the safety feature on n devices as outlined in the instructions for use, and for each failure mode identified in 5.3.2, carry out the following.

- a) Insert the device in a fixture.
- b) Prepare the fixture as required.
- c) Start the test cycle.
- d) Read the force gauge.
- e) Record the test results. Attribute sampling shall be conducted in accordance with ISO 2859 and variable sampling shall be conducted in accordance with ISO 3951.
- f) Repeat for n devices (as specified).

5.4 Testing access to the sharp in safe mode

It shall be demonstrated that, once the device is in safe mode, the risk of accidental access to the sharps is minimized.

No single test method can cover the variety of designs and working principles associated with sharps protection features. Based upon the individual design of the engineered sharps injury protection device, the manufacturer shall demonstrate that such risk of accidental access to the sharps is minimized, whether by a specific test method and/or by engineering study/computer aided design analysis.

Considerations of usability engineering (for medical devices) shall be made that assess and mitigate risks caused by usability problems associated with correct use and use errors.

The test method described in Annex B, when found to be suitable for a particular design of an engineered sharps injury protection device and for the anticipated usage setting, is an appropriate method for demonstrating compliance with the requirement specified in 4.3 b).

NOTE See Reference [2] for hand/finger sizes.

5.5 Testing simulated clinical use

To satisfy most general requirements, the manufacturer shall demonstrate that the safety feature does not impede or adversely affect the intended clinical performance of the device, does not activate prematurely under expected conditions of use, and provides protection against unintentional sharps injury until safe disposal of the sharp.

As evidence, the manufacturer shall conduct simulated user studies (SUS) that mimic actual clinical use by using patient substitutes (e.g. instructional models) rather than actual patients. Where appropriate, the study design should be based on statistical considerations and shall have clear acceptance criteria.

5.6 Test report

The test report shall include the following information:

- identity of the device;
- test equipment description;
- the safety mechanism activation force;
- the force used to challenge the safe mode (overriding force);
- test results for verifying access to the sharps in safe mode.

6 Information supplied by the manufacturer

6.1 General

The device with integrated sharps injury protection or stand-alone sharps injury protection device shall be accompanied by the information needed for its safe and proper use, taking account of the training and knowledge of the potential users, and the information needed to identify the manufacturer.

Information needed for the safe use of the sharps injury protection device shall be set out on the device itself and/or on the packaging for each unit or, where appropriate, on the sales packaging. If individual packaging of each unit is not practicable, the information shall be set out in the leaflet supplied with one or more devices.

NOTE In some countries, national regulations exist whose requirements may supersede or complement the marking, labelling and information specified in this clause.

6.2 Marking/labelling

Any marking of a device with integrated sharps injury protection or stand-alone sharps injury protection device that is essential for the safe use of the device shall be visible, and easily legible after being subjected to the preconditioning specified in 5.1. This shall be checked by visual inspection by normal, or corrected-to-normal, vision at environmental lighting condition of (215 ± 20) lux.

The marking/labelling of the device with integrated sharps injury protection or stand-alone sharps injury protection device shall at least comprise the following particulars:

- a) the name or trade name and address of the manufacturer;
- b) the details strictly necessary to identify the device and contents of the packaging, especially for the user;
- c) where appropriate, the word "STERILE";
- d) the batch code, preceded by the word "LOT", or the harmonized symbol, or the serial number;
- e) where appropriate, an indication of the date by which the device should be used, in safety, expressed as the year and month (e.g. YYYY-MM);
- f) where appropriate, an indication that the device is for single use;
- g) any special storage and/or handling conditions;
- h) where applicable, the method of sterilization;
- i) the intended purpose of the device, if not obvious to the user, clearly stated by the manufacturer.

6.3 Instructions for use

The sharps protection feature shall be accompanied by sufficient information on its safe use.

The instructions for use shall at least contain information on the following:

- a) precautions to be taken and any warnings;
- b) if the sharps protection feature is to be connected to the device with the sharp in order to operate as required for its intended purpose, sufficient details of its characteristics to identify the correct devices in order to obtain a safe combination;
- c) description of special features;
- d) instructions for use that clearly describe how the safety mechanism is activated and that define user ergonomics so that the user's hands remain behind the contaminated sharp;
- e) proper instructions for disposal of the actuated sharps protection device, provided by the manufacturer;
- f) the details given in 6.2 with the exception of d) and e);
- g) date of issue or the latest revision of the instructions for use.

Annex A (informative)

Guidance on simulated user studies

A.1 General

Simulated use testing mimics actual clinical use by using patient substitutes (e.g. instructional models) rather than actual patients. Simulated use testing helps

- isolate problems with the device,
- optimize the device design,
- identify deficiencies in labelling, and
- evaluate the type of training needed for device users.

There are no standardized, validated methods for simulating clinical use of sharps injury protection features. It is recommended that a protocol specific for the device be developed. Whenever possible, protocols should be based on statistical considerations, such as sample size, response variables, pass/fail criteria, comprehensive report forms/questionnaires, proper controls, and appropriate statistical test methods. The protocol should be comprehensive; it should

- state a clear objective,
- include a determination of sample size,
- explain how the number of evaluators was determined,
- explain how evaluators were selected,
- define terms and evaluation parameters used, and
- explain how the data will be analysed.

A.2 Study design

The evaluators should include a variety of individuals representing the user. These may represent the patients, anybody providing assistance, e.g. parents or other family members, or health care professionals who routinely use, or help in the use of the type of device being tested. A simulated clinical use study using volunteers that mimic intended use generally helps avoid learning curve artefacts. It is recommended that the study also include observers who comment on the evaluators' adherence to protocol and their technique.

Bias should be minimized by selecting a sufficient number of evaluators so that each evaluator uses a large enough sample of devices (such as 1/8 of the total number per evaluator) to allow him/her to gain familiarity with the device and thus provide an objective opinion. The evaluators should have no conflicting financial interest in the device, but they may be compensated for their time. Studies conducted at more than one test site will decrease test bias.

The device should be tested under conditions that simulate the critical clinical variables (e.g. models to simulate patients, gloved hands, dry and wet fingers, one-handed technique).