

INTERNATIONAL  
STANDARD

ISO  
27105

IDF  
216

First edition  
2016-04-01

---

---

**Milk and cheese — Determination  
of hen's egg white lysozyme  
content by high performance liquid  
chromatography**

*Lait et fromages — Détermination de la teneur en lysozyme de blanc  
d'oeuf par chromatographie liquide haute performance*

STANDARDSISO.COM : Click to view the full PDF of ISO 27105:2016



Reference numbers  
ISO 27105:2016(E)  
IDF 216:2016(E)

© ISO and IDF 2016

STANDARDSISO.COM : Click to view the full PDF of ISO 27105:2016



**COPYRIGHT PROTECTED DOCUMENT**

© ISO and IDF 2016, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Ch. de Blandonnet 8 • CP 401  
CH-1214 Vernier, Geneva, Switzerland  
Tel. +41 22 749 01 11  
Fax +41 22 749 09 47  
copyright@iso.org  
www.iso.org

International Dairy Federation  
Silver Building • Bd Auguste Reyers 70/B • B-1030 Brussels  
Tel. + 32 2 325 67 40  
Fax + 32 2 325 67 41  
info@fil-idf.org  
www.fil-idf.org

# Contents

Page

<b>Forewords</b> .....	<b>iv</b>
<b>Introduction</b> .....	<b>vi</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Terms and definitions</b> .....	<b>1</b>
<b>3 Principle</b> .....	<b>1</b>
<b>4 Reagents and reference substances</b> .....	<b>1</b>
4.1 Reagents and materials .....	1
4.2 Lysozyme .....	2
<b>5 Apparatus</b> .....	<b>2</b>
<b>6 Sampling</b> .....	<b>3</b>
<b>7 Procedure</b> .....	<b>3</b>
7.1 Preparation of lysozyme standard solution .....	3
7.1.1 Lysozyme standard stock solution .....	3
7.1.2 Lysozyme standard working solution .....	3
7.2 Test portion .....	3
7.2.1 Milk .....	3
7.2.2 Cheese .....	3
7.2.3 Preparation of test solution .....	3
7.3 HPLC determination .....	4
7.3.1 Chromatographic conditions .....	4
<b>8 Calculation and expression of results</b> .....	<b>5</b>
8.1 Single-point calibration .....	5
8.2 Expression of results .....	6
<b>9 Precision</b> .....	<b>6</b>
9.1 Interlaboratory test .....	6
9.2 Repeatability .....	6
9.3 Reproducibility .....	6
<b>10 Test report</b> .....	<b>6</b>
<b>Annex A (informative) Verification by LC/MS</b> .....	<b>8</b>
<b>Annex B (informative) Interlaboratory test</b> .....	<b>10</b>
<b>Bibliography</b> .....	<b>11</b>

## Forewords

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#)

The committee responsible for this document is ISO/TC 34, *Food products*, Subcommittee SC 5, *Milk and milk products* and the International Dairy Federation (IDF) and it is being published jointly by ISO and IDF.

This first edition of ISO 27105|IDF 216 cancels and replaces ISO/TS 27105|IDF/RM 216:2009, which has been technically revised.

IDF (the International Dairy Federation) is a non-profit private sector organization representing the interests of various stakeholders in dairying at the global level. IDF members are organized in National Committees, which are national associations composed of representatives of dairy-related national interest groups including dairy farmers, dairy processing industry, dairy suppliers, academics and governments/food control authorities.

ISO and IDF collaborate closely on all matters of standardization relating to methods of analysis and sampling for milk and milk products. Since 2001, ISO and IDF jointly publish their International Standards using the logos and reference numbers of both organizations.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. IDF shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

ISO 27105|IDF 216 was prepared by the IDF Standing Committee on *Analytical Methods for Additives and contaminants* and ISO Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 5, *Milk and milk products*.

All work was carried out by the ISO/IDF Project Group on *Determination of hen's egg white lysozyme content by high performance liquid chromatography* of the Standing Committee on *Analytical Methods for Additives and Contaminants* under the aegis of the project leaders, T. Berger (CH) and Prof. L. Pellegrino (IT).

This first edition of ISO 27105|IDF 216 cancels and replaces ISO/TS 27105|IDF/RM 216:2009, which has been technically revised.

STANDARDSISO.COM : Click to view the full PDF of ISO 27105:2016

## Introduction

Lysozyme (EC 3.2.1.17, muramidase) is an enzyme widely dispersed in nature; it appears, e.g. in hen's egg white (approximately 3 % to 4 %), saliva and in tear liquid. Lysozyme has a preservative effect because of the lytic activity on the cell wall of some bacteria. Hen's egg white lysozyme is used in cheese making to prevent late blowing of semi-hard and hard cheeses.

STANDARDSISO.COM : Click to view the full PDF of ISO 27105:2016

# Milk and cheese — Determination of hen's egg white lysozyme content by high performance liquid chromatography

## 1 Scope

This International Standard specifies a method for the quantitative determination of hen's egg white lysozyme content in milk and cheese.

The method is suitable for measuring low levels of hen's egg white lysozyme with a quantification limit of 10 mg/kg.

## 2 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

### 2.1

#### hen's egg white lysozyme content

mass fraction of substance determined by the procedure specified

Note 1 to entry: The lysozyme content is expressed as milligram per kilogram.

## 3 Principle

Casein and denatured whey proteins from milk and cheese are precipitated isoelectrically at pH 4,3 (cheese) or at pH 2,2 (milk). Acid-soluble hen's egg white lysozyme is then determined by reversed-phase high-performance liquid chromatography (HPLC) and fluorescence detection. The lysozyme peak can be verified by LC/MS (see [Annex A](#)).

## 4 Reagents and reference substances

Use only reagents of recognized analytical grade, unless otherwise specified, and only distilled water or water of equivalent purity.

### 4.1 Reagents and materials

#### 4.1.1 Sodium chloride solution, $c(\text{NaCl}) = 1 \text{ mol/l}$ .

Dissolve 58,44 g of sodium chloride in 1 l water.

#### 4.1.2 Hydrochloric acid, $c(\text{HCl}) = 1 \text{ mol/l}$ .

Dissolve 4,0 ml of hydrochloric acid with mass fraction 37 % in a 50 ml one-mark volumetric flask. Dilute to the mark with water.

#### 4.1.3 Sodium hydroxide, $c(\text{NaOH}) = 1 \text{ mol/l}$ .

Dissolve 2,6 ml of sodium hydroxide with a mass fraction of 50 % in a 50 ml one-mark volumetric flask. Dilute to the mark with water.

#### 4.1.4 Trifluoroacetic acid ( $\text{CF}_3\text{COOH}$ ), analytical grade.

4.1.5 **Acetonitrile** (CH<sub>3</sub>CN), HPLC grade.

4.1.6 **Water**, HPLC grade.

## 4.2 Lysozyme

Pure hen's egg white lysozyme<sup>1)</sup>. Sufficiently pure and characterized lysozyme is difficult to obtain. A control on lot-to-lot variability is needed.

## 5 Apparatus

Usual laboratory equipment and, in particular, the following.

5.1 **pH-meter**, accurate to 0,1 unit.

5.2 **Fluted filter**, diameter 15 cm<sup>2)</sup>.

5.3 **Membrane filter**, pore size 0,22 µm<sup>3)</sup>.

5.4 **Balance**, capable of weighing to the nearest 100 mg, with a readability of 10 mg.

5.5 **Analytical balance**, capable of weighing to the nearest 0,1 mg, with a readability of 0,01 mg.

5.6 **Magnetic stirrer**.

5.7 **Homogenizer**<sup>4)</sup>, capable of spinning at a rotational frequency of 3 000 rpm to 3 500 rpm.

5.8 **HPLC equipment**.

5.8.1 **Gradient pumping system**, capable of operating with a speed of 1,0 ml/min.

5.8.2 **Manual or automatic injector**, capable of injecting amounts of 50 µl.

5.8.3 **Column heater**, capable of maintaining a column temperature of 45 °C ± 2 °C.

5.8.4 **Column**, reversed phase<sup>5)</sup>, 5 µm, 250 mm × 4,6 mm.

---

1) Lysozyme SIGMA L-6876 is an example of a suitable product available commercially. This information is given for the convenience of users of this document and does not constitute an endorsement by either ISO or IDF of this product.

2) Fluted filter Schleicher&Schuell 595 ½ is an example of a suitable product available commercially. This information is given for the convenience of users of this document and does not constitute an endorsement by either ISO or IDF of this product.

3) Millipore Millex-GV PVDF 0,22 µm is an example of a suitable product available commercially. This information is given for the convenience of users of this document and does not constitute an endorsement by either ISO or IDF of this product. Equivalent products may be used if they can be shown to lead to the same results.

4) Ultra turrax® is an example of a suitable product available commercially. This information is given for the convenience of users of this document and does not constitute an endorsement by either ISO or IDF of this product. Equivalent products may be used if they can be shown to lead to the same results.

5) PLRP-S 300 Å (Polymer Laboratories, UK) is an example of a suitable product available commercially. This information is given for the convenience of users of this document and does not constitute an endorsement by either ISO or IDF of this product.

**5.8.5 Fluorescence detector**, capable of operating at 280 nm excitation and at 340 nm emission.

## 6 Sampling

Sampling is not part of the method specified in this International Standard. A recommended sampling method is given in ISO 707|IDF 50.<sup>[1]</sup>

It is important that the laboratory receives a sample, which is representative and has not been damaged or changed during transport or storage.

## 7 Procedure

### 7.1 Preparation of lysozyme standard solution

#### 7.1.1 Lysozyme standard stock solution

Weigh to the nearest 0,01 mg, 10 mg of lysozyme (4.2) into a 10 ml one-mark volumetric flask and wait for complete dissolution. Dilute to the mark with sodium chloride solution (4.1.1).

Prepare fresh the standard stock solutions daily.

#### 7.1.2 Lysozyme standard working solution

Pipette 80 µl of the lysozyme standard stock solution (see 7.1.1) into a 10 ml one-mark volumetric flask. Dilute to the mark with sodium chloride solution (4.1.1).

The obtained lysozyme standard working solution contains 8,0 mg of lysozyme per litre.

### 7.2 Test portion

#### 7.2.1 Milk

Weight to the nearest 0,01 g, 10,00 g of test sample into a 100 ml beaker.

#### 7.2.2 Cheese

Before weighing, grate the test samples of cheese. Weight to the nearest 0,01 g, 2,00 g of test sample into a 100 ml beaker.

NOTE Soft cheese can be grated after freezing.

#### 7.2.3 Preparation of test solution

Add 20 ml of sodium chloride solution (4.1.1) to the test portion (see 7.2.1 or 7.2.2) and mix. Adjust the pH of the obtained solution drop wise with sodium hydroxide solution (4.1.3) to pH 6,0.

Homogenize the test solution for 30 s by using the homogenizer (5.7) at a speed rate of 2 500 rpm to 3 000 rpm. Rinse the homogenizer in a separate 100 ml beaker using 10 ml of sodium chloride solution (4.1.1). Add the rinsing to the test solution.

Stir the beaker containing the test solution on a magnetic stirrer at room temperature for 1 h. Adjust the pH of the test solution obtained from the test portion of milk (see 7.2.1) to pH 2,2 or that obtained from the test portion of cheese to pH 4,3 by using hydrochloric acid (4.1.2).

Transfer the test solution from the 100 ml beaker to a 50 ml one-mark volumetric flask. Use sodium chloride solution (4.1.1) to rinse the 100 ml beaker and add the rinsing to the test solution. Dilute to the mark with the sodium chloride solution (4.1.1) and mix.

Allow the test solution to stand at room temperature for at least 15 min.

Firstly, filter the test solution through a fluted filter (5.2) and then through the membrane filter (5.3) directly into a HPLC vial.

### 7.3 HPLC determination

#### 7.3.1 Chromatographic conditions

Prepare the following stock solutions:

- stock solution I: 1 ml of trifluoroacetic acid (4.1.4) in 1 l of water (4.1.6);
- stock solution II: 1 ml of trifluoroacetic acid (4.1.4) in 1 l of acetonitrile (4.1.5).

Use the following elution solvents for HPLC:

- elution solvent A contains: stock solution I:stock solution II = 72,25:27,75 (w/w);
- elution solvent B contains: stock solution II.

Table 1 — Suggested gradient elution


Time (min)	Elution solvent A <sup>a</sup> (%)	Elution solvent B <sup>a</sup> (%)
0,0	100	0
20,0	100	0
21,0	50	50
22,0	50	50
23,0	100	0
35,0	100	0

<sup>a</sup> The gradient elution might require slight modification in order to achieve the resolution shown in Figure 1 and Figure 2. Isocratic elution in the range of the lysozyme peak, as used in this International Standard, leads to a cleaner chromatogram but to larger peak width. The use of a gradient may lead to smaller peak widths and more stable retention times but also to more peaks in the lysozyme range with a risk of misinterpretation.

Set the flow rate of the gradient elution pumping system of the HPLC equipment at 1,0 ml/min.

Set the temperature of the column heater at 45 °C. Determine the equilibration time by monitoring the column elution. The detector response at the end of the run (baseline) should be equal to its initial value. An isocratic flushing of 15 min is usually sufficient.


Injection: Use a manual or automatic injector to inject 50 µl of the solutions into the column.



**Key**

1 lysozyme

**Figure 1 — HPLC fluorescence signal from the standard working solution (7.1.2)**



**Key**

1 lysozyme

**Figure 2 — HPLC fluorescence signal from a hard cheese sample**

## 8 Calculation and expression of results

### 8.1 Single-point calibration

Calculate the lysozyme content of the test sample,  $w$ , expressed in milligrams per kilogram, by using Formula (1):

$$w = \frac{H_t \times c_s \times V_t}{H_s \times m_t} \quad (1)$$

where

- $c_s$  is the concentration, in milligrams per litre, of the standard working solution (see [7.1.2](#));
- $H_t$  is the numerical value of the peak height or area of the test solution (see [7.2](#)), in counts;
- $H_s$  is the numerical value of the peak height or area of the standard working solution (see [7.1.2](#)), in counts;
- $m_t$  is the mass, in grams, of the test portion (see [7.2.1](#) or [7.2.2](#));
- $V_t$  is the volume, in millilitres, of the test solution (see [7.2.3](#)).

Check equipment linearity and reagent blank regularly. This may be needed in case of lysozyme contents of more than 100 mg/kg of cheese which can lead to a non-linear response. In such cases, the use of a multiple point calibration is necessary.

## 8.2 Expression of results

Express the results to one decimal place. Express test results of below 10 mg/kg as: "less than 10 mg/kg", or equivalent expression.

## 9 Precision

### 9.1 Interlaboratory test

The values for repeatability and reproducibility derived from an interlaboratory test<sup>[7]</sup> were determined in accordance with ISO 5725-1<sup>[2]</sup> and ISO 5725-2<sup>[3]</sup>

The values for the repeatability and the reproducibility limit are expressed for the 95 % probability level and may not be applicable to concentration ranges and matrices other than those given. Details of the interlaboratory test on the precision of the method are given in [Annex B](#).

### 9.2 Repeatability

The absolute difference between two individual single test results, obtained with the same method on identical test material in the same laboratory by the same operator using the same equipment within a short interval of time, shall in not more than 5 % of the cases be greater than 7,7 % (95 % confidence intervals 6,2 % to 8,5 %).

### 9.3 Reproducibility

The absolute difference between two individual single test results, obtained with the same method on identical test material in different laboratories with different operators using different equipment, shall in not more than 5 % of the cases be greater than 33,8 % (95 % confidence intervals 20,3 % to 57,5 %).

Running the chromatographic system on another temperature, using a different column type, or other differing conditions do not result automatically in false or less sensitive results but need thoroughly to be validated.

## 10 Test report

The test report shall contain at least the following information:

- all information necessary for the complete identification of the sample;
- the sampling method used, if known;
- the test method used, with reference to this International Standard, i.e. ISO 27105|IDF 216;

- d) all operational details not specified in this International Standard, or regarded as optional, together with details of any incident which may have influenced the test result(s);
- e) the test result(s) obtained, and, if the repeatability has been checked, the final quoted results obtained.

STANDARDSISO.COM : Click to view the full PDF of ISO 27105:2016

## Annex A (informative)

### Verification by LC/MS

#### A.1 General

In case of lysozyme peaks in the test solutions, findings can be verified using liquid chromatography/mass spectrometry LC/MS. MS parameters for lysozyme have to be optimized and tuned according to the system manual of the instrument (e.g. needle voltage 3,0 kV, probe 550 °C, cone voltage 130 V).

Use an LC/MS equipment capable of running the conditions described in A.2 and A.3 [different from HPLC (see 7.3) because of non-useful trifluoroacetic acid] and measure the mass signals  $m/z = 1\,431 [M+H_{10}]^{10+}$ ;  $1\,590 [M+H_9]^{9+}$  and  $1\,788 [M+H_8]^{8+}$ , retention time is approx. 12,5 min.

Verify the presence of lysozyme by homogeneous distribution of these masses all over the presumed lysozyme peak.

#### A.2 LC/MS equipment

**A.2.1 Gradient pumping system**, capable of operating at 0,8 ml/min.

**A.2.2 Manual or automatic injector**, capable of injecting amounts of 5 µl.

**A.2.3 Column heater**, capable of maintaining the column temperature at  $40\text{ °C} \pm 2\text{ °C}$ .

**A.2.4 Column**, reversed phase<sup>6)</sup>, 5 µm, 250 mm × 4,6 mm, conditioned for several hours without trifluoroacetic acid.

**A.2.5 Mass spectrometer detector**, capable of operating in ion mode ESI+ at  $m/z$  1 431; 1 590 and 1 788.

#### A.3 Chromatographic conditions

Use the following elution solvents for LC/MS:

- elution solvent A containing 5 ml of formic acid (analytical grade) in 1 l of water (4.1.6);
- elution solvent B containing 5 ml of formic acid in 1 l of acetonitrile (4.1.5).

Set the flow rate of the pumping system of the HPLC equipment at 0,80 ml/min. The flow split should be: 0,5 ml/min MS; 0,3 ml/min waste (when 25 % B), motor valve 0,0 min eluent MS; 0,5 min eluent waste; 7,0 min eluent MS.

Using a manual or automatic injector inject 5 µl.

---

6) PLRP-S 300Å (Polymer Laboratories, UK) is an example of a suitable product available commercially. This information is given for the convenience of users of this document and does not constitute an endorsement by either ISO or IDF of this product.