
**Surface active agents — Detergents —
Determination of cationic-active matter
content —**

Part 1:
**High-molecular-mass cationic-active
matter**

*Agents de surface — Détergents — Détermination de la teneur en
matière active cationique —*

Partie 1: Matière active cationique à haute masse moléculaire



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

STANDARDSISO.COM : Click to view the full PDF of ISO 2871-1:2010



COPYRIGHT PROTECTED DOCUMENT

© ISO 2010

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 2871-1 was prepared by Technical Committee ISO/TC 91, *Surface active agents*.

This second edition cancels and replaces the first edition (ISO 2871-1:1988), of which it constitutes a minor revision.

ISO 2871 consists of the following parts, under the general title *Surface active agents — Detergents — Determination of cationic-active matter content*:

- *Part 1: High-molecular-mass cationic-active matter*
- *Part 2: Cationic-active matter of low molecular mass (between 200 and 500)*

STANDARDSISO.COM : Click to view the full PDF of ISO 2871-1:2010

Surface active agents — Detergents — Determination of cationic-active matter content —

Part 1: High-molecular-mass cationic-active matter

1 Scope

This part of ISO 2871 specifies a method for the determination of high-molecular-mass cationic-active materials such as

- a) quaternary ammonium compounds in which two of the alkyl groups each contain 10 or more carbon atoms, e.g. distearyl-dimethyl-ammonium chlorides, or
- b) salts of imidazoline or 3-methylimidazoline in which long-chain acylaminoethyl and alkyl groups are substituted in the 1- and 2-positions, respectively.

The method is applicable to solids or to aqueous solutions of the active material when the relative molecular mass of the cationic-active matter is known or when it has been previously determined if its content is expressed as a percentage by mass.

The method is not applicable if anionic surface active agents are present.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 385, *Laboratory glassware — Burettes*

ISO 607, *Surface active agents and detergents — Methods of sample division*

ISO 1042, *Laboratory glassware — One-mark volumetric flasks*

ISO 2271:1989, *Surface active agents — Detergents — Determination of anionic-active matter by manual or mechanical direct two-phase titration procedure*

3 Principle

Titration of an aliquot portion of a standard solution of sodium lauryl sulfate with a solution of the test sample, according to the direct two-phase titration procedure described in ISO 2271.

4 Reagents

During the analysis, use only reagents of recognized analytical grade and only distilled water or water of at least equivalent purity.

4.1 Chloroform, $\rho_{20} = 1,48$ g/ml, distilling between 59,5 °C and 61,5 °C.

4.2 Propan-2-ol.

4.3 Sodium lauryl sulfate, standard volumetric solution, $c(\text{C}_{12}\text{H}_{25}\text{NaO}_4\text{S}) = 0,004$ mol/l.

See 4.5 of ISO 2271:1989.

4.4 Mixed indicator solution.

See 4.8 of ISO 2271:1989.

5 Apparatus

Ordinary laboratory apparatus and the following:

5.1 Flask or measuring cylinder, 100 ml capacity, with ground-glass stopper.

5.2 Burette, 25 ml capacity, complying with the specifications for class A of ISO 385.

5.3 One-mark volumetric flask, 1 000 ml capacity, with ground-glass stopper, complying with class A of ISO 1042.

6 Sampling

The laboratory sample of the detergent shall be prepared and stored in accordance with the instructions given in ISO 607.

7 Procedure

7.1 Test portion

Weigh, to the nearest 1 mg, a sufficient amount of the laboratory sample to contain between 0,002 mol and 0,004 mol of cationic-active matter.

NOTE This will give a titration volume of between 10 ml and 20 ml.

Possible interferences:

- Low relative molecular mass sulfonates of toluene and xylene present as hydrotropes do not interfere when present in concentrations up to and including 15 % (mass fraction) with respect to the active material. At higher levels, their influence should be evaluated in each particular case.
- Non-ionic surface active agents, soap, urea and the salts of (ethylenedinitrilo)tetraacetic acid do not interfere.
- Typical inorganic components of detergent formulations, such as sodium chloride, sulfate, borate, tripolyphosphate, perborate, silicate, etc., do not interfere, but bleaching agents other than perborate shall be destroyed before the analysis, and the sample shall be completely soluble in water.

7.2 Determination

Dissolve the test portion (7.1) in 20 ml of the propan-2-ol (4.2), warming if necessary. Add approximately 50 ml of water and stir. Transfer to the 1 000 ml one-mark volumetric flask (5.3) and dilute to the mark with water. This is test solution A.

By means of a pipette, transfer 10 ml of the standard volumetric sodium lauryl sulfate solution (4.3) into the flask or measuring cylinder (5.1). Add 10 ml of water, 15 ml of the chloroform (4.1) and 10 ml of the mixed indicator solution (4.4).

Fill the burette (5.2) with test solution A.

Titrate the sodium lauryl sulfate solution with test solution A; stopper the flask or measuring cylinder and shake well after each addition.

The lower layer of chloroform will be coloured pink. As the end point approaches, the emulsions formed during shaking tend to break easily and the coloration of the chloroform layer becomes fainter.

From this point, continue the titration drop by drop, shaking after each addition of titrant, until the end point is reached. This is at the moment when the pink colour is completely discharged from the chloroform layer, which turns a faint greyish blue. With excess of test solution A, the chloroform layer turns blue. Note the volume of test solution A required for the titration.

8 Expression of results

8.1 Calculation

The content, expressed as a percentage by mass, of cationic-active matter is given by the formula

$$\frac{cM_r \times 10 \times 1000 \times 100}{1000 Vm_0} = \frac{cM_r \times 1000}{Vm_0}$$

where

c is the actual concentration, expressed in moles of $C_{12}H_{25}NaO_4S$ per litre, of the sodium lauryl sulfate solution (4.3);

M_r is the relative molecular mass of the cationic-active matter;

V is the volume, in millilitres, of test solution A used for the titration;

m_0 is the mass, in grams, of the test portion (7.1).

8.2 Precision

8.2.1 Repeatability

The difference between the results of two determinations carried out on the same sample simultaneously or in rapid succession by the same analyst using the same apparatus should not exceed 1,5 % of the mean value.

8.2.2 Reproducibility

The difference between the results obtained on the same sample in two different laboratories should not exceed 3 % of the mean value.

9 Test report

The test report shall include the following information:

- a) all the indications necessary for the complete identification of the sample;
- b) the method used (reference to this part of ISO 2871);
- c) the results obtained and the way in which they have been expressed;
- d) any unusual features noted during the determination;
- e) any operation not specified in this part of ISO 2871 or in the International Standards to which reference is made, as well as any operation regarded as optional.

STANDARDSISO.COM : Click to view the full PDF of ISO 2871-1:2010