
**Wheat, rye and respective flours, durum
wheat and durum wheat semolina —
Determination of the Falling Number
according to Hagberg-Perten**

*Blés tendres, seigles et leurs farines, blés durs et leurs semoules —
Détermination de l'Indice de Chute selon Hagberg-Perten*

STANDARDSISO.COM : Click to view the full PDF of ISO 3093:2004



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

STANDARDSISO.COM : Click to view the full PDF of ISO 3093:2004

© ISO 2004

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	iv
1 Scope.....	1
2 Normative references	1
3 Terms and definitions.....	1
4 Principle	2
5 Reagents	2
6 Apparatus.....	2
7 Sampling	3
8 Preparation of test sample	3
8.1 Whole grain.....	3
8.2 Flour and semolina samples.....	4
9 Procedure.....	4
9.1 Determination of the moisture content.....	4
9.2 Test portion	4
9.3 Determination of Falling Number	5
9.4 Calculation	6
10 Precision	7
10.1 Interlaboratory tests	7
10.2 Repeatability	7
10.3 Reproducibility	7
11 Test report.....	8
Annex A (normative) Formulae for altitude correction of Falling Number values.....	9
Annex B (informative) Results of interlaboratory tests	10
Bibliography	13

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

ISO draws attention to the fact that it is claimed that compliance with this document may involve the use of a patent concerning the Falling Number apparatus given in 6.1.

ISO takes no position concerning the evidence, validity and scope of this patent right.

The holder of this patent right has assured the ISO that he is willing to negotiate licences under reasonable and non-discriminatory terms and conditions with applicants throughout the world. In this respect, the statement of the holder of this patent right is registered with the ISO. Information may be obtained from:

Perten Instruments AB
P.O. Box 5101
S-141 05 HUDDINGE
Sweden

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights other than those identified above. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 3093 was prepared by Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 4, *Cereals and pulses*.

This third edition cancels and replaces the second edition (ISO 3093:1982), which has been technically revised.

Wheat, rye and respective flours, durum wheat and durum wheat semolina — Determination of the Falling Number according to Hagberg-Perten

1 Scope

1.1 This International Standard describes the determination of the α -amylase activity of cereals by the Falling Number method according to Hagberg-Perten.

This method is applicable to cereal grains, in particular to wheat and rye and their respective flours, durum wheat and its semolina. For the purposes of this International Standard the term "flour" includes semolina and ground grain (wholemeal), the particle sizes of which are defined.

This method is not applicable for the determination of low levels of α -amylase activity, which can be carried out in accordance with ISO 7973.

1.2 By converting the Falling Number into a Liquefaction Number (LN), it is possible to use this method to estimate the composition of mixtures of grain, flour or semolina with known Falling Number values necessary to produce a sample of a required Falling Number.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 565, *Test sieves — Woven metal wire cloth, perforated metal plate and electroformed sheet — Nominal sizes of openings*

ISO 712, *Cereals and cereals products — Determination of moisture content — Routine reference method*

ISO 3310-1, *Test sieves — Technical requirements and testing — Part 1: Test sieves of metal wire cloth*

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

Falling Number

FN

total time, in seconds, required to activate a viscometer stirrer and allow it to fall a predetermined distance through an aqueous gel prepared from heating a mixture of flour or semolina and water in a viscometer tube, and which is undergoing liquefaction due to attack by the enzyme α -amylase

NOTE Time is counted from immersion in the water bath.

3.2

Liquefaction Number

LN

result of a simple calculation to convert the Falling Number into a value used to estimate the composition of mixtures of grain, flour or semolina necessary to produce a sample of the required Falling Number

NOTE LN values, unlike Falling Numbers are additive.

4 Principle

The α -amylase activity is estimated using the starch present in the sample as a substrate. The determination is based on the ability of an aqueous suspension of flour, semolina or wholemeal cereal product to gelatinize rapidly in a boiling water bath, and on the measurement of starch liquefaction by the α -amylase present in the sample.

Liquefaction affects the thickness of the starch gel and, hence, the resistance to the viscometer stirrer and the time taken for it to fall a defined distance.

5 Reagents

5.1 Water, produced by distillation or demineralization, complying with grade 3 in accordance with ISO 3696.

6 Apparatus

Usual laboratory apparatus and, in particular, the following.

6.1 Apparatus for the determination of Falling Number¹⁾, comprising the following components.

6.1.1 Water bath, with integral heating unit, cooling system and water level indicator.

6.1.2 Electronic timer.

6.1.3 Viscometer stirrer, metallic, able to move freely within the ebonite stopper.

Its rods shall be straight, without ring and the wheels free from distortion and wear.

6.1.4 Precision viscometer tubes, manufactured from special glass, with the following dimensions:

- inner diameter of 21,00 mm \pm 0,02 mm;
- outer diameter of 23,80 mm \pm 0,25 mm;
- inner height of 220,0 mm \pm 0,3 mm.

6.1.5 Rubber stoppers, to fit the viscometer tubes.

6.2 Automatic dispenser or pipette, allowing a volume of 25,0 ml \pm 0,2 ml to be delivered.

1) The "Falling Number" apparatus having a specifically designed viscometer stirrer produced by Perten Instruments is an example of suitable device available commercially. This device forms the subject of a patent.

This information is given for the convenience of users of this International Standard and does not constitute an endorsement by ISO of this product.

6.3 Analytical balance, capable of weighing to the nearest 0,01 g.

6.4 Laboratory mill²⁾, hammer type, and fitted with a 0,8 mm screen allowing the production of a wholemeal product meeting the particle size specification shown in 8.1.3.

The performance of the mill should be checked periodically using a well-mixed sample of ground grain (as produced in 8.1.2).

The mill should be equipped with an automatic feeding device, particularly for the grinding of high moisture content grain.

6.5 Laboratory sieve, of aperture 800 µm, and conforming to ISO 565 and ISO 3310-1.

7 Sampling

Sampling is not part of the method specified in this International Standard. A recommended sampling method is given in ISO 13690.

It is important the laboratory receive a sample which is truly representative and has not been damaged or changed during transport or storage. Storage time and storage conditions of the sample in the laboratory may have a significant effect on Falling Number values.

8 Preparation of test sample

8.1 Whole grain

8.1.1 Elimination of impurities

If necessary, the sample should be cleaned to eliminate impurities (e.g. stones, dust, husks and other cereal grains). A 300 g of representatively sampled grain shall be taken from the laboratory sample.

A smaller sample of about 200 g, although it provides less reproducible results, may be used for routine inspections. If the sample is less than 200 g, the results risk being marred by errors.

8.1.2 Grinding of grain samples

Feed the laboratory mill (6.4) carefully with grain to avoid heating and overloading. The feed to the mill may be controlled automatically with an automatic feeding device. Grinding should be continued for 30 s to 40 s after the last of the sample has entered the mill. Discard the bran particles remaining inside the mill, provided these do not represent more than 1 % of the quantity of grain sampled for grinding. Thoroughly mix all of the milled product before use.

It is recommended (especially in the case of successive grindings) to allow the meal to cool for 1 h before proceeding with the test.

8.1.3 Ground sample specification

WARNING — The Falling Number values can be affected by the particle size of the ground grain.

2) The LM 3100 and LM 120 mills from Perten Instruments are examples of suitable devices available commercially.

This information is given for the convenience of users of this International Standard and does not constitute an endorsement by ISO of these products.

The milled product shall comply with the specification given in Table 1.

Table 1 — Ground sample specification

Aperture of sieve	Amount of ground sample passing through sieve
µm	%
710	100
500	95 to 100
210 to 200	80 or less

The particle size distribution of the ground material should be checked regularly using a well-mixed sample of ground grain as produced in 8.1.2.

The following method may be used to do this. Select the appropriate sieves, as specified in Table 1, and arrange in order of decreasing aperture size with a suitable sieving aid in each sieve and a receiver pan on the bottom. Weigh out a representative sample of 50,0 g and place on the top sieve. Sieve in a horizontal plane, manually for at least 5 min until nothing passes through the sieve, or mechanically for a period of 10 min. Weigh the material retained on each sieve and that contained in the receiver pan. Calculate the percentage of ground grain which passes through each sieve.

8.2 Flour and semolina samples

Flour samples shall not contain lumps. If necessary, sift the flour using the laboratory sieve (6.5) in order to remove lumps or foreign bodies.

In the case of coarse commercial wholemeal flours or semolina, a sample shall be ground using the laboratory mill (6.4) in order to produce a test sample which conforms to the particle size specification shown in Table 1. The ground sample shall be mixed thoroughly before use.

9 Procedure

9.1 Determination of the moisture content

The Falling Number should be determined on flour or ground material with a moisture content of 15%.

Therefore, determine the moisture content of the prepared test material (8.1 and 8.2) using the method described in ISO 712.

Alternatively, a rapid instrumental procedure may be used (e.g. near infrared reflectance) provided this has been calibrated using ISO 712.

9.2 Test portion

Carry out the determination on two test portions simultaneously or rapidly one after the other.

Refer to Table 2, column (2) which shows the required mass of sample to be taken, at different moisture contents, in order to ensure that a constant ratio of dry matter is used for the Falling Number determination.

If greater differentiation of the Falling Number values is required for samples with very high α -amylase activity, as is normally the case for rye, refer to column (3).

Weigh the test portion to the nearest 0,05 g.

Table 2 — Test portion as a function of moisture content of the sample

Moisture content %	Test portion, g		Moisture content %	Test portion, g	
	for a nominal mass of 7 g at 15 % moisture content	for a nominal mass of 9 g at 15 % moisture content		for a nominal mass of 7 g at 15 % moisture content	for a nominal mass of 9 g at 15 % moisture content
(1)	(2)	(3)	(1)	(2)	(3)
9,0	6,40	8,20	13,6	6,85	8,80
9,2	6,45	8,25	13,8	6,90	8,85
9,4	6,45	8,25	14,0	6,90	8,85
9,6	6,45	8,30	14,2	6,90	8,90
9,8	6,50	8,30	14,4	6,95	8,90
10,0	6,50	8,35	14,6	6,95	8,95
10,2	6,55	8,35	14,8	7,00	8,95
10,4	6,55	8,40	15,0	7,00	9,00
10,6	6,55	8,40	15,2	7,00	9,05
10,8	6,60	8,45	15,4	7,05	9,05
11,0	6,60	8,45	15,6	7,05	9,10
11,2	6,60	8,50	15,8	7,10	9,10
11,4	6,65	8,50	16,0	7,10	9,15
11,6	6,65	8,55	16,2	7,15	9,20
11,8	6,70	8,55	16,4	7,15	9,20
12,0	6,70	8,60	16,6	7,15	9,25
12,2	6,70	8,60	16,8	7,20	9,25
12,4	6,75	8,65	17,0	7,20	9,30
12,6	6,75	8,65	17,2	7,25	9,35
12,8	6,80	8,70	17,4	7,25	9,35
13,0	6,80	8,70	17,6	7,30	9,40
13,2	6,80	8,75	17,8	7,30	9,40
13,4	6,85	8,80	18,0	7,30	9,45

9.3 Determination of Falling Number

9.3.1 Fill the water bath (6.1.1) with water to the level dictated by the overflow. Turn on the cooling system and ensure that cold water is flowing through the cooling lid. Switch on the Falling Number apparatus and bring the water to the boil. The water bath shall be boiling vigorously before any determinations are carried out and during the entire test period.

9.3.2 Transfer the weighed test portion (9.2) to a clean, dry viscometer tube (6.1.4). Add 25 ml \pm 0,2 ml of water (5.1) at 22 °C \pm 2 °C using the automatic dispenser or pipette (6.2).

9.3.3 Immediately stopper (6.1.5) the viscometer tube (6.1.4) and shake it vigorously up and down 20 to 30 times in order to obtain a uniform suspension³⁾. Ensure that dry flour or ground material is not trapped at the top of the tube against the stopper. If necessary, this can be freed by easing the stoppers slightly upwards and reshaking as necessary.

3) The "SHAKE-MATIC" apparatus from Perten Instruments is convenient for the homogenization of the suspension.

This information is given for the convenience of users of this International Standard and does not constitute an endorsement by ISO of this product.

9.3.4 Remove the stopper (6.1.5), scrape any material remaining on the bottom of the stopper back into the tube (6.1.4) and, with the viscometer stirrer (6.1.3), scrape down any material adhering to the sides of the tube. Leave the stirrer in the tube.

For double-tube instruments, operations 9.3.2 to 9.3.4 should be performed within 30 s of the addition of water and may be conducted on two tubes simultaneously.

9.3.5 Immediately place the viscometer tube (6.1.4), together with the stirrer (6.1.3), through the hole in the lid into the boiling water bath (6.1.1). Activate the stirrer head (single or double head) following the manufacturer's instructions. The equipment will then automatically carry out the operations to complete the test. The test is considered to be complete when the viscometer stirrer has reached the bottom of the gelatinized suspension. Record the time displayed on the timer (6.1.2). This constitutes the Falling Number.

9.3.6 Swing the stirrer head across or press the "stop" button to withdraw the stirrer head. Remove the tube and its stirrer carefully as they are hot. Clean the tubes and stirrers thoroughly, ensuring that no material is left in the recess of the ebonite top which could interfere with the descending stirrer during the next test. Rinse the tubes and allow to drain. Ensure that the viscometer stirrer is dry before re-use.

9.4 Calculation

9.4.1 Falling Number (FN)

The Falling Number value is affected by the boiling temperature of the water, which is linked to the atmospheric pressure and the altitude of the laboratory. No adjustment of the boiling temperature of the water bath should be made as this leads to errors in the results.

For laboratories located below an altitude of 600 m, the uncorrected Falling Number values should be used for ground grain samples, and below an altitude of 750 m uncorrected values should be used for flour or semolina.

For those laboratories located above these altitudes, apply the appropriate correction formula given in Annex A.

Take as result the arithmetic mean of two determinations if the conditions of repeatability given in Table 3 and Table 4 are satisfied.

9.4.2 Liquefaction number (LN)

The relationship between Falling Number values and α -amylase activity is not linear and therefore Falling Numbers cannot be used to calculate the composition of mixtures of grain, flour or semolina. The non-linear relationship may be converted into a linear one which makes it possible to calculate the theoretical Falling Number of a blend of wheat, flour or semolina arithmetically or graphically. Transform Falling Number values into Liquefaction Numbers (LN) using the empirical formula given below:

$$NL = \frac{6\,000}{IC - 50}$$

where

FN is the Falling Number value;

6 000 is a constant value;

50 is a constant value, corresponding approximately to the time, expressed in seconds, required for the starch to gelatinize sufficiently to be susceptible to attack by enzymes.

Liquefaction Numbers are proportional to α -amylase activity over the normal range found in commercial flour.

10 Precision

10.1 Interlaboratory tests

The results of the interlaboratory tests on the precision of the method are given in Annex B. The values derived from these interlaboratory tests may not be applicable to concentration ranges and matrices other than those given.

10.2 Repeatability

The absolute difference between two independent single test results, obtained using the same method on identical test material in the same laboratory by the same operator using the same equipment within a short interval of time, will in not more than 5 % of cases be greater than the values given in Table 3 and Table 4.

If the results of two tests are outside these limits, carry out two determinations again.

10.3 Reproducibility

The absolute difference between two single test results, obtained using the same method on identical test material in different laboratories with different operators using different equipment, will in not more than 5 % of cases be greater than the values given in Table 3 and Table 4.

Table 3 — Repeatability and reproducibility limits for Hagberg Falling Numbers for flour (deduced from Table B.1)

Values in seconds (s)

Hagberg Falling Number value	Repeatability limit	Reproducibility limit
	<i>r</i>	<i>R</i>
60 to 199	5	10
200 to 229	9	24
230 to 259	12	27
260 to 289	15	30
290 to 319	19	33
320 to 349	22	36
350 to 379	25	39
380 to 409	28	42
410 to 439	31	45
440 to 469	35	48
470 to 499	38	51
≥ 500	40	60

Table 4 — Repeatability and reproducibility limits for Hagberg Falling Number for wheat (deduced from Table B.2)

Values in second (s)

Hagberg Falling Number value	Repeatability limit <i>r</i>	Reproducibility limit <i>R</i>
60 to 79	10	10
80 to 109	13	21
110 to 139	15	30
140 to 169	17	38
170 to 199	19	46
200 to 229	21	54
230 to 259	23	62
260 to 289	25	70
290 to 319	27	78
320 to 349	30	86
350 to 379	32	94
≥ 380	40	100

11 Test report

The test report shall specify:

- all information necessary for the complete identification of the sample;
- the sampling method used, if known;
- the test method used, in particular the nominal test mass taken;
- all operating details not specified in this International Standard, or regarded as optional, together with details of any incidents which may have influenced the test result(s);
- the test result(s) obtained, and whether any correction has been made for altitude;
- if the repeatability has been checked, the final result obtained.

Annex A (normative)

Formulae for altitude correction of Falling Number values

A.1 Ground grain samples

For laboratories located at an altitude over 600 m (2 000 ft) where the boiling temperature of the water bath is lower than 98 °C, correct the Falling Numbers (FN) using the following formula:

$$F_n = 10^X$$

where

$$X = 1,0 \times \lg F_{\text{alt}} - 1,630\,93 \times 10^{-4} \times A + 2,635\,76 \times 10^{-8} \times A^2 + 5,750\,30 \times 10^{-5} \times \lg F_{\text{alt}} \times A - 1,069\,223 \times 10^{-8} \times \lg F_{\text{alt}} \times A^2$$

F_n is the calculated Falling Number value corresponding to the value at sea level;

F_{alt} is the original value as measured at a given altitude;

A is the altitude of the laboratory, in feet above sea level. If the altitude is measured in metres, a factor of 3,28 shall be included to convert metres to feet.

A.2 Flour and semolina

For laboratories located at an altitude over 750 m (2 500 ft) where the boiling temperature of the water bath is lower than 98 °C, correct the Falling Numbers (FN) using the following formula:

$$F_n = 10^X$$

where

$$X = -849,41 + 0,425\,6 \times 10^{-5} \times A^2 + 454,19 \times \lg F_{\text{alt}} \times A - 0,212\,9 \times 10^{-5} \times \lg F_{\text{alt}} \times A^2$$

F_n is the calculated Falling Number value corresponding to the value at sea level;

F_{alt} is the original value as measured at a given altitude;

A is the altitude of the laboratory, in feet above sea level. If the altitude is measured in metres a factor of 3,28 shall be included to convert metres to feet.

The calculation of Falling Number from the formulae given in A.1 and A.2 may be replaced by reading from a conversion table, drawn up to correct for a specific altitude, in order that corrected Falling Number values may be obtained for each measured value.

Annex B
(informative)

Results of interlaboratory tests

B.1 Wheat flours

A test conducted by the International Association of Cereal Science and Technology (ICC) using 11 laboratories (10 retained after elimination of outliers) on 10 samples of flours gave the statistical results (assessed according to Reference [7]) shown in Table B.1.

Table B.1 — Data for Falling Number for wheat flour

Values in seconds (s)

	Sample									
	1	2	3	4	5	6	7	8	9	10
Mean value	210,6	218,6	244,5	263,9	296,5	333,5	373,9	408,3	457,2	485,9
Repeatability standard deviation, s_r	4,51	4,22	5,03	4,73	5,28	6,74	8,31	10,96	10,88	16,70
Relative standard deviation of repeatability, %	2,14	1,93	2,06	1,79	1,78	2,02	2,22	2,68	2,38	3,44
Repeatability limit, r ($2,8 s_r$)	12,63	11,82	14,08	13,25	14,79	18,88	23,27	30,70	30,48	46,76
Reproducibility standard deviation, s_R	5,66	9,12	10,21	8,20	14,98	16,00	13,23	17,82	13,81	18,89
Relative standard deviation of reproducibility, %	2,69	4,17	4,18	3,11	5,05	4,80	3,54	4,37	3,02	3,89
Reproducibility limit, R ($2,8 s_R$)	15,84	25,53	28,59	22,97	41,94	44,82	37,04	49,91	38,67	52,90

The relative standard deviation of repeatability was less than or equal to 3,44 %.

The relative standard deviation of reproducibility was less than or equal to 5,05 %.