
**Petroleum and liquid petroleum
products — Measurement of level
and temperature in storage tanks by
automatic methods —**

**Part 2:
Measurement of level in marine
vessels**

*Pétrole et produits pétroliers liquides — Mesurage du niveau et de
la température dans les réservoirs de stockage par des méthodes
automatiques —*

Partie 2: Mesurage du niveau dans les citernes de navire



STANDARDSISO.COM : Click to view the full PDF of ISO 4266-2:2023



COPYRIGHT PROTECTED DOCUMENT

© ISO 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword.....	v
Introduction.....	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Precautions	3
4.1 Safety precautions.....	3
4.1.1 General.....	3
4.1.2 Equipment precautions.....	3
4.2 General recommendations.....	3
4.2.1 Accuracy and performance.....	3
4.2.2 Speed of response.....	3
4.2.3 Protection from mechanical damage.....	4
4.2.4 Manual gauging.....	4
4.2.5 Minimum measurable level.....	4
4.2.6 Trim and list.....	4
4.2.7 Product temperatures.....	4
4.2.8 Compatibility.....	4
4.2.9 Entrained air and vapour.....	4
4.2.10 Vessel motion.....	4
4.3 Use of marine ALGs in fiscal/custody transfer.....	5
5 Accuracy	5
5.1 Intrinsic error of ALGs.....	5
5.2 Calibration prior to installation.....	5
5.3 Initial shipyard adjustment.....	5
5.4 Error caused by operating conditions.....	5
5.5 Overall accuracy.....	6
5.5.1 General.....	6
5.5.2 Use of ALGs for fiscal/custody transfer purposes.....	6
6 Installation of marine ALGs	6
6.1 General.....	6
6.2 Location of ALG.....	6
6.3 Location of manual calibration check point.....	6
6.4 Gauging of inerted tanks.....	6
7 Onboard verification of marine ALGs	7
7.1 General precautions.....	7
7.1.1 Check for smooth operation of level-sensing elements— at the shipyard.....	7
7.1.2 ALG technology-specific considerations.....	7
7.2 Verification by innage gauging or ullage gauging.....	7
7.3 Initial verification.....	7
7.4 Subsequent verification.....	7
7.4.1 General.....	7
7.4.2 Agreement between ALG reading and manual gauge reading.....	7
7.4.3 Use of average gauge readings.....	7
7.4.4 Adjustment of the ALG.....	8
7.5 Verification by alternate methods.....	8
7.6 Schedule for regular ALG verification.....	8
7.7 Record keeping.....	8
8 Data communication and receiving	8
8.1 General.....	8
8.2 Use of remote readout in fiscal or custody transfer.....	8

8.3 Telemetry and readout equipment.....	9
Annex A (informative) Accuracy limitations of marine level measurement.....	10
Annex B (informative) Accuracy limitations of marine volume measurement.....	11
Bibliography.....	12

STANDARDSISO.COM : Click to view the full PDF of ISO 4266-2:2023

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 28 *Petroleum and related products, fuels and lubricants from natural or synthetic sources*, Subcommittee SC 2, *Measurement of petroleum and related products*.

This second edition cancels and replaces the first edition (ISO 4266-2:2002), which has been technically revised.

The main changes are as follows:

- in [Clause 2](#), the reference to ISO 8697 has been removed as it is a withdrawn document;
- in [4.2.7](#), it has been clarified that the tank level should be measured and recorded simultaneously with the temperature;
- in [7.4.4.1](#), it has been clarified regarding the procedure prior to making adjustments against manual measurements;
- added addition documents in Bibliography.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Marine automatic level gauges (ALGs) are not normally used in fiscal or custody transfer applications because of the limitations described in [Annexes A](#) and [B](#). However, level measurement by marine ALGs may be used in fiscal or custody transfer when no other alternative, reliable measurement is available. The use of marine-vessel-based ALGs in fiscal or custody transfer normally requires mutual contractual agreement between the buyer and the seller and can be subject to government regulations.

STANDARDSISO.COM : Click to view the full PDF of ISO 4266-2:2023

Petroleum and liquid petroleum products — Measurement of level and temperature in storage tanks by automatic methods —

Part 2: Measurement of level in marine vessels

1 Scope

This document gives guidance on the accuracy, installation, calibration and verification of automatic level gauges (ALGs), both intrusive and non-intrusive, for measuring the level of petroleum and liquid petroleum products having a Reid vapour pressure less than 100 kPa, transported aboard marine vessels (i.e. tankers and barges).

This document gives guidance for buyers and sellers who mutually agree to use marine ALGs for either fiscal and/or custody transfer applications.

This document is not applicable to the measurement of level in refrigerated cargo tanks.

NOTE For information on the measurement of level in refrigerated cargo tanks, please see ISO 18132-1 and ISO 18132-3.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1998-1, *Petroleum industry — Terminology — Part 1: Raw materials and products*

ISO 1998-2, *Petroleum industry — Terminology — Part 2: Properties and tests*

ISO 1998-3, *Petroleum industry — Terminology — Part 3: Exploration and production*

ISO 1998-4, *Petroleum industry — Terminology — Part 4: Refining*

ISO 1998-5, *Petroleum industry — Terminology — Part 5: Transport, storage, distribution*

ISO 1998-6, *Petroleum industry — Terminology — Part 6: Measurement*

ISO 1998-7, *Petroleum industry — Terminology — Part 7: Miscellaneous terms*

ISO 1998-99, *Petroleum industry — Terminology — Part 99: General and index*

ISO 4512, *Petroleum and liquid petroleum products — Equipment for measurement of liquid levels in storage tanks — Manual methods*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 1998-1, ISO 1998-2, ISO 1998-3, ISO 1998-4, ISO 1998-5, ISO 1998-6, ISO 1998-7, ISO 1998-99 and the following apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1

automatic level gauge

ALG

automatic tank gauge

ATG

instrument that continuously measures liquid height [*dip* (3.2) or *ullage* (3.5)] in storage tanks

3.2

dip

innage

vertical distance between the dipping datum point and the liquid level

3.3

innage-based automatic level gauge

automatic level gauge (3.1) designed and installed to measure the liquid *dip* (3.2) directly

3.4

still-pipe

stilling-well

stand pipe

guide pole

vertical, perforated pipe built into a tank to contain the liquid-level-detecting element in order to reduce measurement errors arising from liquid turbulence, surface flow or agitation of the liquid, and to provide a stable mounting point for an automatic level gauge

3.5

ullage

outage

distance between the liquid level and the upper reference point, measured along the vertical measurement axis

3.6

ullage-based automatic level gauge

automatic level gauge (3.1) designed and installed to measure the distance from the ALG reference point to the liquid surface

3.7

intrusive automatic level gauge

automatic level gauge (3.1) where the level-sensing device intrudes within the tank and makes physical contact with the liquid, e.g. a float and servo-operated-type automatic tank gauge

3.8

non-intrusive automatic level gauge

automatic level gauge (3.1) where the level-sensing device can intrude within the tank, but does not make physical contact with the liquid, e.g. a microwave or radar-type automatic tank gauge

4 Precautions

4.1 Safety precautions

4.1.1 General

In addition, the manufacturers' recommendations on the use and installation of the equipment should be followed.

NOTE International Standards and government regulations, classification societies and ISGOTT on safety and material-compatibility precautions can apply when using marine ALG equipment.

4.1.2 Equipment precautions

4.1.2.1 All marine ALGs should be capable of withstanding the pressure, temperature and other environmental conditions likely to be encountered in marine service. When an ALG is installed in a corrosive service, any parts exposed to the liquid or vapours should be of durable, corrosion-resistant construction.

4.1.2.2 All ALGs should be sealed to withstand the vapour pressure of liquid in the tank. ALGs mounted on vessels with an inert gas system (IGS) should be designed to withstand the operating pressure of the IGS.

4.1.2.3 All marine ALGs should be specified and installed in accordance with the appropriate national and/or international (IMO, IEC, CENELEC, ISGOTT, ISO, etc.) marine electrical safety standards. ALGs should be certified for use in the hazardous-area classification appropriate to their installation.

All ALG equipment should be maintained in safe operating condition and the manufacturers' maintenance instructions should be complied with.

NOTE 1 The design and installation of ALGs can be subject to the approval of the national measurement organization and classification societies, who can have issued a general type approval for the design of the ALG for the particular service for which it is to be employed. Type approval is normally issued after an ALG has been subjected to a specific series of tests, and is subject to the ALG being installed in an approved manner.

NOTE 2 Type-approval tests can include the following: visual inspection, performance, vibration, humidity, dry heat, inclination, fluctuations in power supplies, insulation, resistance, electromagnetic compatibility, and high voltage.

4.1.2.4 The ALGs should provide security to prevent unauthorized adjustment or tampering. For ALGs to be used in fiscal or custody transfer application, the ALG should provide facilities to allow sealing of the calibration adjustment.

4.2 General recommendations

4.2.1 Accuracy and performance

The general recommendations given in [4.2.2](#) to [4.2.10](#) affect the accuracy and performance of all types of marine ALGs and should be observed where they are applicable.

4.2.2 Speed of response

Marine ALGs should have sufficient dynamic response to track the liquid level during maximum tank filling or emptying rates.

4.2.3 Protection from mechanical damage

Marine ALGs should be designed to withstand damage caused by waves in the tanks due to ship movement. They should also be able to withstand damage from high velocity jets of water or oil used to wash the tanks.

NOTE 1 This protection can require mounting the ALGs in still-pipes.

NOTE 2 Alternately, this protection can require that the ALGs with a float or displacer-type level-sensing element be raised to a "store" position when it is not being used. Such ALGs cannot be used during tank washing.

4.2.4 Manual gauging

When an ALG is set or verified by manual gauging, the manual gauging shall be performed to obtain the highest accuracy in accordance with ISO 4512.

4.2.5 Minimum measurable level

The ALG should be able to measure levels as near to the bottom of the tank as possible. This may require provision of a sump in the tank bottom in vessels with double bottoms.

NOTE The minimum measurable level of certain types of ALGs can limit their ability to measure small volumes of remaining on board (ROB)/on board quantity (OBQ).

4.2.6 Trim and list

For the best accuracy, the vessel should be on an even keel and upright. In situations where both trim and list exist, every effort should be made to eliminate at least one condition, preferably list.

Trim and list corrections are not required on vessel tanks of cuboid (i.e. rectangular prism) shape, provided that the ALG is located at the geometric centre of the deck area for the tank. Where the ALG is not so located, correction will be required. On vessel tanks that have curvature(s), such as the aft and forward wing tanks, trim and list corrections are recommended. Correction for trim, list and wedge is permissible by table or calculation, using the procedure described in API MPMS Chapter 17.4.

4.2.7 Product temperatures

Product temperatures should be measured and recorded simultaneously, or as close as practical, as the tank level is measured. The temperature should be representative of the tank contents and should be measured as described in ISO 4266-5.

4.2.8 Compatibility

All parts of the ALG in contact with the product should be compatible with the product, to avoid product contamination.

4.2.9 Entrained air and vapour

Sufficient time should be allowed before gauging a tank to permit the liquid to free itself of entrained air or gas vapours.

4.2.10 Vessel motion

During lightering or offshore operation, or when the vessel is at an exposed berth, vessel motion causes waves on the surface of the product. At least three readings should be taken as quickly as possible and the readings should be averaged. If the vessel is in heavy motion due to large swells or waves, at least five readings should be taken.

Some ALGs provide internal filtering algorithms as part of the readout to average the level readings over a time interval. The filtering time can be fixed or made adaptive to the encountered motion.

4.3 Use of marine ALGs in fiscal/custody transfer

Marine ALGs are not normally used in fiscal or custody transfer application because of the limitations described in [Annexes A](#) and [B](#). However, level measurement by marine ALGs can be used in fiscal or custody transfer when no other alternative, reliable measurement is available. The use of marine-vessel-based ALGs in fiscal or custody transfer normally requires mutual contractual agreement between the buyer and the seller.

NOTE Government regulations can apply.

5 Accuracy

5.1 Intrinsic error of ALGs

The measurement accuracy of all marine ALGs is affected by the inherent error of the ALG, i.e. the error of the ALGs when tested under controlled conditions as specified by the manufacturers.

5.2 Calibration prior to installation

Marine ALGs to be used in a fiscal or custody transfer application should, prior to installation (i.e. in the factory and in a testing laboratory), agree with a certified reference within 3 mm at a minimum of three points over the intended operating range of the ALG. The certified reference should be traceable to national standards.

The uncertainty of the certified reference should not exceed 0,5 mm, with correction applied. If the reference is a certified gauge tape, refer to ISO 4512 for tape calibration.

NOTE Metrological requirements for the uncertainty of the calibration reference can be more stringent.

5.3 Initial shipyard adjustment

The shipyard adjustment (i.e. setting) procedure should be performed in accordance with the ALG manufacturer's instructions. The procedure normally consists of setting the ALG so that it reads correctly at the minimum level and at the setting point where the zero adjustment took place. The distances from the zero adjustment point to the zero point for the tank capacity table should be referenced.

The setting of the ALG should also confirm that the remote readout reads the same level (within ± 1 mm) as the level transmitter (if the ALG is provided with a local deck readout).

5.4 Error caused by operating conditions

The total error of the measurement by an ALG, due to operating conditions, varies and it is difficult to quantify on a general basis. For examples, refer to [Annex A](#).

5.5 Overall accuracy

5.5.1 General

The overall accuracy of level measurement by ALGs, as installed, is limited by the intrinsic accuracy of the ALG equipment, the effect of installation methods, and the effect of changes in the operating conditions.

NOTE Depending on the overall accuracy of the ALG as installed (“installed accuracy”), ALGs can be used in either fiscal or custody transfer, or for operational purposes. The use of ALGs in fiscal/custody transfer requires the highest possible accuracy. The use of ALGs for operational purposes often permits a lower degree of accuracy.

5.5.2 Use of ALGs for fiscal/custody transfer purposes

The ALG should meet the calibration tolerance prior to installation (see [5.2](#)).

The ALG should meet the calibration tolerance after installation and initial setting at the shipyard (see [5.3](#)).

The ALG should meet the onboard verification tolerance (see [Clause 7](#)).

6 Installation of marine ALGs

6.1 General

Marine ALGs should be installed in accordance with the ALG manufacturer’s instructions.

6.2 Location of ALG

The ALG should be located to minimize the effect of surface turbulence and waves in the cargo tank. The location should be designed to avoid damage during tank washing. Mechanical, float-operated ALGs which cannot withstand waves or tank washing should be provided with the ability to “store” the float when the ALG is not in use.

For vessel tanks of cuboid (i.e. rectangular prism) shape, the ALG should be located at the geometric centre of the tank, which usually eliminates the need for trim and list corrections. On vessel tanks that have curvature side(s), such as the aft and forward wing tanks, the ALG should be located nearer to the inboard bulkhead to avoid interference with the curvature of the vessel's tank bottom.

Presently, the combination of vessel design and ALG technology limits the use of ALGs to the measurement of partially or fully loaded vessels. Therefore, other measurement locations are required to accommodate the measurement of small volumes (ROB/OBQ) (see API MPMS, Chapter 17.4).

6.3 Location of manual calibration check point

To permit accurate comparison of the product level measured between manual gauging and automatic tank gauging, a manual calibration check point should be provided close (i.e. within 1 m) to the ALG.

6.4 Gauging of inerted tanks

On tanks with an inert gas system (IGS), the ALG should be designed and installed so that it can be maintained and calibrated without depressurising the IGS.

7 Onboard verification of marine ALGs

7.1 General precautions

7.1.1 Check for smooth operation of level-sensing elements— at the shipyard

After mounting ALGs which use intrusive level-sensing elements and prior to calibration, the level-sensing elements should be checked to ensure free operation over the entire operating range. This check should be done slowly to simulate actual operation and to avoid damaging the ALG.

7.1.2 ALG technology-specific considerations

It is possible that there are additional technology-specific considerations which can affect ALG verification. Specific, additional steps can be required to prepare the ALG prior to the verification. Technology-specific considerations include, for example, the effect of physical and electrical properties of the liquid and vapour in the tank, the need for checking free movement of the level sensor, and others (refer to the ALG manufacturers' documentation for further instruction).

7.2 Verification by innage gauging or ullage gauging

Marine ALGs that measure the ullage level from the top down should be verified by manual ullage gauging. Marine ALGs that measure innage level from the bottom up should be verified by manual innage gauging.

7.3 Initial verification

After leaving the shipyard and before the maiden voyage, most vessels fill certain tanks with water to check the operation of the vessel's pumps, valves and lines. During this initial tank filling, the verification of the ALGs should be checked against manual tank gauging.

7.4 Subsequent verification

7.4.1 General

After completion of cargo loading at a load port, and prior to discharge at a discharge port, the normal practice is to check the reading of the ALG against manual tank gauging. The manual gauging levels are normally used for the cargo reports. To minimize the effect of vessel motion and adverse external conditions, verification should be performed with a stable liquid level. The liquid level at which the ALG is verified should be within the intended operating range of the ALG.

Where the reference gauge point for manual gauging is different from the ALG, apply the appropriate correction.

7.4.2 Agreement between ALG reading and manual gauge reading

If the reading by ALG and the reading by manual gauging agree within 6 mm, no further action should be required.

7.4.3 Use of average gauge readings

If the reading by ALG and the reading by manual gauging differ by more than 6 mm, the ALG reading and the manual gauging should be repeated three times (or five times if there are waves in the tank). The average of the manual gauge readings and the ALG readings should be compared. For best accuracy, the vessel should be on an even keel and upright. In situations where both trim and list exist, every effort should be made to eliminate at least one condition, preferably list.

7.4.4 Adjustment of the ALG

7.4.4.1 If the difference between the average ALG reading and the average of the manual gauge readings exceeds 6 mm, the ALG should be investigated and differences recorded over at least 3 levels, if difference exists over all levels, it should be re-adjusted or reset to agree with manual gauging. These adjustments and the reasons for them should be recorded in the vessel's equipment maintenance log.

7.4.4.2 After adjustment, the reading by ALG should be compared with the reading by manual gauging as described in [7.4.3](#). If the difference between the average ALG reading and the average manual gauge reading is less than 6 mm, no further action should be required.

7.4.4.3 If the ALG cannot be adjusted to agree with the average manual gauge reading, a correction may be used. The correction value should be posted near the ALG readout and used to correct the ALG reading. This correction should be recorded in the vessel documents.

7.5 Verification by alternate methods

In practice, due to the operating constraints (e.g. closed or restricted gauging requirement), a lack of properly located manual gauging point(s), or adverse sea conditions (e.g. swell, vessel motion), reliable manual gauging often cannot be obtained to verify ALG measurement. As an alternative method, onboard verification can be performed by comparing the level measured by the ALG against predetermined, stable reference(s) at the tank or ALG support pipe. The method and procedure, which can vary depending on the ALG type, should follow the ALG manufacturer's recommendation.

7.6 Schedule for regular ALG verification

Vessels which do not routinely confirm ALG accuracy (by comparison with manual gauging or by comparison with reference) should perform the verification on a quarterly basis as a minimum.

7.7 Record keeping

ALG verification records should be documented and the record should be ready for inspection by involved parties. The records should be kept for a minimum of one year or twenty voyages whichever is the longer.

8 Data communication and receiving

8.1 General

This clause gives recommendations for the specification of the communication between level transmitter(s) and receiver(s), or vice versa.

8.2 Use of remote readout in fiscal or custody transfer

The remote readout of an acceptable ALG may be used in fiscal or custody transfer providing that the whole system, including the remote readout, complies with the required calibration tolerances given in [Clause 5](#).

NOTE Some readout equipment can be programmed to alarm at high or low levels. Some equipment can also look up the tank-capacity table, by applying the appropriate coefficients of expansion and calculate the standard volumes.

8.3 Telemetry and readout equipment

The ALG should be designed and installed such that data transmission and receipt should:

- not compromise the accuracy of the measurement, i.e. the difference between the levels displayed by the remote receiving unit and displayed (or measured) by the level transmitter at the tank should not exceed 1 mm,
- not compromise the resolution of the measurement output signal,
- provide proper security and protection of the measured data to ensure its integrity,
- provide adequate speed to meet the update time required for the receiving unit, and
- be electromagnetically immune.

STANDARDSISO.COM : Click to view the full PDF of ISO 4266-2:2023

Annex A (informative)

Accuracy limitations of marine level measurement

Level measurement using marine ALGs is affected by the following inherent limitations, regardless of the ALGs used.

- a) Measurement of small volumes (ROB or OBQ).

Measurement of small volumes of ROB and OBQ by an ALG can be difficult.

- b) Accurate determination of trim and list.

Accurate determination of trim and list is difficult, and the trim and list corrections affect the accuracy of the marine level measurement. Due to hog and sag, twists and bends, it may be necessary to take a multi-point draft, then using trim correction which is appropriate to that of the tank. If automatic correction for trim and/or list is provided as part of the ALG readout, the correction should be used in accordance with MPMS API Chapter 17.4.

- c) Effect of the vessel motion causing waves in the tanks.

Waves in the tanks make it difficult to measure an average level. Many ALGs read the instantaneous level at the point of measurement, whereas manual level gauging tends to measure the height of the wave crests, making calibration of ALGs difficult when waves are present in the tanks. Some ALGs provide internal filtering algorithms as part of the readout to average the level readings over a time interval. The filtering time can be fixed or made adaptive to the encountered motion.

- d) Change of the tank dimensions due to oil or water temperature.

The dimensions of a vessel's tanks change with the water and oil temperature and other factors. This affects the conversion from tank level to volume. The change in the vertical tank dimension also changes the reference height which affects the level accuracy of ALGs which are mounted on the top deck structure.

- e) Change of the tank dimensions due to hog or sag.

Hogging or sagging changes the reference height which affects the level accuracy of ALGs which are mounted on the top deck structure.

The limitations listed above can have significant impact on the overall accuracy of level measurement by all types of marine automatic level gauges.