
**Laboratory glassware — Boiling flasks
with conical ground joints**

*Verrerie de laboratoire — Fioles coniques et ballons à col muni d'un
assemblage conique rodé*

STANDARDSISO.COM : Click to view the full PDF of ISO 4797:2004



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

STANDARDSISO.COM : Click to view the full PDF of ISO 4797:2004

© ISO 2004

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 4797 was prepared by Technical Committee ISO/TC 48, *Laboratory glassware and related apparatus*, Subcommittee SC 6, *Laboratory and volumetric ware*.

This second edition cancels and replaces the first edition (ISO 4797:1981), which has been technically revised to incorporate the following changes:

- a) dimensions and joint sizes have been adapted to the current state of manufacturing;
- b) two series have been introduced for conical flasks;
- c) height tolerances have been added to Series 1;
- d) conical flask of 10 ml nominal volume has been added;
- e) round-bottom flasks of 10 ml, 25 ml, 5 l, 6 l and 10 l nominal volume have been added.

[STANDARDSISO.COM](https://standardsiso.com) : Click to view the full PDF of ISO 4797:2004

Laboratory glassware — Boiling flasks with conical ground joints

1 Scope

This International Standard specifies requirements for an internationally acceptable series of boiling flasks with conical ground joints for general laboratory purposes.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 383, *Laboratory glassware — Interchangeable conical ground joints*

ISO 1773, *Laboratory glassware — Narrow-necked boiling flasks*

ISO 3585, *Borosilicate glass 3.3 — Properties*

3 Types

Three types of boiling flasks with conical ground joints are specified:

- a) conical flasks;
- b) flat-bottom flasks;
- c) round-bottom flasks.

4 Series of capacities

Two series are specified for each type of boiling flask with conical ground joints. The series differ in height and in selection of joint sizes. It is recommended that, in national standards, one of these series is chosen.

5 Material

Boiling flasks shall be made from borosilicate glass 3.3 in accordance with ISO 3585, and shall be free from visible defects which might affect performance and free from any internal stress which would impair the performance of the flask.

6 Dimensions

The external diameter of body of round-bottom flasks and flat-bottom flasks, the external diameter of body at the widest point of conical flasks and the minimum wall thickness shall comply with the dimensions specified in ISO 1773.

The nominal overall height of the boiling flasks with conical ground joints shall be as specified in Tables 1 to 3.

7 Ground glass joints

The sizes of the conical joints fitted to the boiling flasks shall be as given in Tables 1 to 3. The joints shall comply with the requirements of ISO 383, k6 series.

8 Marking

The following inscriptions shall be permanently and legibly marked on all laboratory boiling flasks with conical ground joints:

- the nominal volume of the boiling flask, for example "100 ml";
- the size of the conical ground joint, for example "29/32";
- the manufacturer's and/or vendor's name and/or mark;
- an area with a surface suitable for marking with a pencil.

It is recommended that reference be made on each flask to this International Standard, for example by the inscription "ISO 4797".

Table 1 — Overall height and joint sizes for conical flasks

| Nominal volume ml | Series 1 | | Series 2 | |
|----------------------|----------------------|----------------------------------|------------------------------|----------------|
| | Overall height mm | Joint sizes | Nominal overall height mm | Joint sizes |
| 10 | 60 ± 3 | 14/23 | — | — |
| 25 | 70 ± 3 | 14/23 | 70 | 14/23 19/26 |
| 50 | 85 ± 6 | 19/26 | 85 | 14/23 19/26 |
| 100 | 100 ± 6 | 14/23 19/26 24/29 29/32 | 105 | 24/29 29/32 |
| 250 | 140 ± 6 | 19/26 24/29 | 135 | 19/26 24/29 |
| 500 | 175 ± 6 | 29/32 | 170 | 29/32 34/35 |
| 1 000 | 220 ± 7 | 24/29 29/32 | 210 | 24/29 29/32 |
| 2 000 | 270 ± 7 | 34/35 | 275 | 34/35 |
| 3 000 | — | — | 310 | 34/35 |
| 5 000 | — | — | 365 | 45/40 |