



**INTERNATIONAL STANDARD ISO 4823:2000**  
**TECHNICAL CORRIGENDUM 1**

Published 2004-07-15

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

## **Dentistry — Elastomeric impression materials**

### **TECHNICAL CORRIGENDUM 1**

*Art dentaire — Produits pour empreintes, à base d'élastomères*

*RECTIFICATIF TECHNIQUE 1*

Technical Corrigendum 1 to ISO 4823:2000 was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 2, *Prosthetic materials*.

*Page 22*

Replace existing Figure 6 with the new Figure 6 shown below, in which the dimension line indicating the height of the test pedestal ( $21,5 \text{ mm} \pm 0,05 \text{ mm}$ ) has been elongated to correctly include the slotted table portion.