

INTERNATIONAL STANDARD

ISO 5223

Third edition
1995-08-15

AMENDMENT 1
1999-12-15

Test sieves for cereals

AMENDMENT 1 Additional sizes

Tamis de contrôle pour céréales

*AMENDEMENT 1
Dimensions supplémentaires*

STANDARDSISO.COM : Click to view the full PDF of ISO 5223:1995/Amd 1:1999



Reference number
ISO 5223:1995/Amd.1:1999(E)

© ISO 1999

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

STANDARDSISO.COM : Click to view the full PDF of ISO 5223:1995/Amd 1:1999

© ISO 1999

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 734 10 79
E-mail copyright@iso.ch
Web www.iso.ch

Printed in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the element of this Amendment may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

Amendment 1 to International Standard ISO 5223:1995 was prepared by Technical Committee ISO/TC 34, *Agricultural food products*, Subcommittee SC 4, *Cereals and pulses*.

STANDARDSISO.COM : Click to view the full PDF of ISO 5223:1995/Amd 1:1999

Test sieves for cereals

AMENDMENT 1 Additional sizes

Page 1, clause 1

Replace the whole clause with the following text.

This International Standard specifies requirements for test sieves to be used for the laboratory determination of undesirable substances in a sample of cereals and which pass through test sieves of the following nominal sizes:

a) test sieves with long rounded apertures:

1,00 mm × 20,0 mm

1,50 mm × 20,0 mm

1,60 mm × 20,0 mm

1,70 mm × 20,0 mm

1,80 mm × 20,0 mm

1,90 mm × 20,0 mm

2,00 mm × 20,0 mm

2,20 mm × 20,0 mm

2,25 mm × 20,0 mm

2,50 mm × 20,0 mm

2,80 mm × 20,0 mm

3,50 mm × 20,0 mm

3,55 mm × 20,0 mm

b) test sieves with round apertures:

diameter 1,40 mm

diameter 1,80 mm

diameter 4,50 mm

Test sieves with long rounded apertures listed in a) are used in particular for separating "shrivelled" kernels from rye, triticale, durum wheat, common wheat and barley. Exceptions are those with apertures of 1,50 mm and 1,60 mm which are used for grading rice, as well as those with apertures of 2,50 mm and 2,80 mm which are usually used for the calibration of malting barley.

Test sieves with rounded apertures of diameter 1,40 mm are used for separating chips (small fragments of kernels) from rice; those with diameter 1,80 mm are used for sorghum; those with diameter 4,50 mm are used for separating broken grains from maize.

Page 2, subclause 4.2.2

Replace the list with the following:

either

- nominal diameter of aperture (w): 1,40 mm,
- aperture tolerance: $\pm 0,08$ mm,
- pitch (centres) (p): 2,6 mm nominal; 3,0 mm max.; 2,2 mm min.

or

- nominal diameter of aperture (w): 4,50 mm,
- aperture tolerance: $\pm 0,14$ mm,
- pitch (centres) (p): 6,3 mm nominal; 7,2 mm max.; 5,3 mm min.

Page 2, Table 1

Replace table 1 with the following table.

Table 1 — Characteristics of test sieves with long rounded apertures and linear perforations

Dimensions in millimetres

Dimensions of apertures				Pitches ¹⁾					Plate thickness
Width	Tolerance on width	Length	Tolerance on length		Normal tolerance	Reduced tolerance		Tolerance	
w_1	$\pm \Delta w_1$	w_2	$\pm \Delta w_2$	p_1	$\pm \Delta p_1$	$\pm \Delta p_1$	p_2	$\pm \Delta p_2$	
1,00	0,03	20,0	0,2	3,0	0,20	0,10	25,0	0,5	0,5 to 0,6
1,50	0,04	20,0	0,2	4,0	0,24	0,12	25,0	0,5	0,8 to 0,9
1,60	0,04	20,0	0,2	4,0	0,24	0,12	25,0	0,5	0,8 to 0,9
1,70	0,04	20,0	0,2	4,0	0,24	0,12	25,0	0,5	0,8 to 0,9
1,80	0,04	20,0	0,2	4,2	0,24	0,12	25,0	0,5	0,8 to 0,9
1,90	0,04	20,0	0,2	4,3	0,24	0,12	25,0	0,5	0,8 to 0,9
2,00	0,04	20,0	0,2	4,5	0,26	0,13	25,0	0,5	0,8 to 0,9
2,20	0,05	20,0	0,2	4,9	0,26	0,13	25,0	0,5	0,8 to 0,9
2,25	0,05	20,0	0,2	4,9	0,26	0,13	25,0	0,5	0,8 to 0,9
2,50	0,05	20,0	0,2	4,9	0,26	0,13	25,0	0,5	0,8 to 0,9
2,80	0,05	20,0	0,2	4,9	0,26	0,13	25,0	0,5	0,8 to 0,9
3,50	0,06	20,0	0,2	6,8	0,34	0,17	25,0	0,5	0,8 to 0,9
3,55	0,06	20,0	0,2	6,8	0,34	0,17	25,0	0,5	0,8 to 0,9

¹⁾ See figure 1.