



**International
Standard**

ISO 5840-2

**Cardiovascular implants — Cardiac
valve prostheses —**

Part 2:
**Surgically implanted heart valve
substitutes**

AMENDMENT 1

*Implants cardiovasculaires — Prothèses valvulaires —
Partie 2: Prothèses valvulaires implantées chirurgicalement
AMENDEMENT 1*

**Second edition
2021-01**

**AMENDMENT 1
2025-03**



COPYRIGHT PROTECTED DOCUMENT

© ISO 2025

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 150, *Implants for surgery*, Subcommittee SC 2, *Cardiovascular implants and extracorporeal systems*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 285, *Non-active surgical implants*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

A list of all parts in the ISO 5840 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

STANDARDSISO.COM : Click to view the full PDF of ISO 5840-2:2021/Amd 1:2025

Cardiovascular implants — Cardiac valve prostheses —

Part 2: Surgically implanted heart valve substitutes

AMENDMENT 1

Clause 2, Normative references

Add the following reference:

ISO/PAS 7020:2023, *Sizing parameters of surgical valve prostheses: Requirements regarding the application of ISO 5840-2*

Clause 3, Terms and definitions

3.5

Remove Note 2 to entry.

3.6

Remove Note 2 to entry.

3.8

Replace the term "supra-annulus" with "supra-annular region".

3.8

supra-annular region

region wholly above the patient's annulus

Note 1 to entry: See Figure 1.

Note 2 to entry: See also 3.4, 3.5, and 3.7.

6.3.3

Replace the entire subclause with the following text.

Refer to ISO 5840-1:2021, 6.3.4.

Refer to ISO/PAS 7020:2023 for requirements pertaining to labelling of surgical valves.

Annex D contains a list of terms that can be used to describe various valve models.

F.2.1.1

Replace the entire subclause with the following text.

Each pressure measurement (e.g. ventricular pressure, aortic pressure) should have an upper frequency limit (−3 dB cut-off) of at least 30 Hz and a measurement accuracy of at least ±0,26 kPa (±2 mmHg). The flow meter should have an upper frequency limit (−3 dB cut-off) of at least 30 Hz.

Annex F, Table F.2

Replace the entire table with the following:

Beat rate cycles/min	Systolic duration %	Cardiac output l/min	Differential pressure across closed valve ^a
45	30	5	Hypotensive, normotensive, severe hypertensive
70	35	5	Hypotensive, normotensive, severe hypertensive
120	50	5	Hypotensive, normotensive, severe hypertensive

^a Refers to the mean differential pressure across the closed valve.

F.2.3.4

Replace the list item g) with the following:

- g) regurgitant volume, including the closing volume and leakage volume (see ISO 5840-1:2021, Figure 2), expressed in millilitres and as a percentage of forward flow volume, and the corresponding differential pressure across closed valve (i.e. mean back pressure).