
International Standard



6974

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Natural gas — Determination of hydrogen, inert gases and hydrocarbons up to C₈ — Gas chromatographic method

Gaz naturel — Détermination de l'hydrogène, des gaz inertes et des hydrocarbures jusqu'en C₈ — Méthode par chromatographie en phase gazeuse

First edition — 1984-10-15

STANDARDSISO.COM : Click to view the full PDF of ISO 6974:1984

UDC 543.544 : 547.211/ .217.2 + 546.29

Ref. No. ISO 6974-1984 (E)

Descriptors : natural gas, gas analysis, determination of content, hydrogen, helium, hydrocarbons, gas phase chromatography, quantitative analysis, apparatus.

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Every member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work.

Draft International Standards adopted by the technical committees are circulated to the member bodies for approval before their acceptance as International Standards by the ISO Council. They are approved in accordance with ISO procedures requiring at least 75 % approval by the member bodies voting.

International Standard ISO 6974 was prepared by Technical Committee ISO/TC 158, *Analysis of gases*.

STANDARDSISO.COM : Click to view the full PDF of ISO 6974:1984

Natural gas — Determination of hydrogen, inert gases and hydrocarbons up to C₈ — Gas chromatographic method

0 Introduction

This International Standard describes a precise and accurate method for the analysis of natural gas.

The method requires the use of two columns which can be put into one or two gas chromatographs. The constituents of the eluent of the first column are detected by a thermal conductivity detector (TCD). The constituents of the eluent of the second column are detected by a TCD and flame ionization detector (FID) in series.

NOTE — If the two columns are put in one chromatograph the procedure given in clause A.6 should be followed.

The detectors and associated electronics shall have a rapid response time — a time constant of not greater than 0,1 s is desired.

The method allows measurement of helium, hydrogen, oxygen, nitrogen, carbon dioxide and C₁ up to C₈ hydrocarbons.

If the content of hydrocarbons only up to C₄ or C₅ is required, reference can be made either to ISO 6568, which describes a simple method for the analysis up to C₅ of constituents of dry natural gas by gas chromatography, or to ISO 6569, which describes a rapid evaluation up to C₄ of constituents of natural gas by gas chromatography.

For analysis of natural gas for hydrocarbons from butane (C₄) up to hexadecane (C₁₆), see ISO 6975.

1 Scope and field of application

This International Standard specifies a method of quantitative analysis of the following constituents of natural gas

- helium
- hydrogen
- oxygen
- nitrogen

- carbon dioxide
- hydrocarbons from C₁ up to C₈

The gas sample shall not contain any hydrocarbon condensate and/or water.

2 References

ISO 683/13, *Heat-treated steels, alloy steels and free-cutting steels — Part 13: Wrought stainless steels.*¹⁾

ISO 6142, *Gas analysis — Preparation of calibration gas mixtures — Weighing methods.*

ISO 6143, *Gas analysis — Determination of composition of calibration gas mixtures — Comparison methods.*

ISO 6146, *Gas analysis — Preparation of calibration gas mixtures — Manometric method.*

ISO 6568, *Natural gas — Simple analysis by gas chromatography.*

ISO 6569, *Natural gas — Rapid analysis by gas chromatography.*

ISO 6975, *Natural gas — Determination of hydrocarbons from butane (C₄) up to hexadecane (C₁₆) — Gas chromatographic method.*²⁾

3 Principle

Separation of the components of natural gas by means of two chromatographic columns. Use of a molecular sieve 13 X³⁾ column with a TCD for the separation and detection of helium, hydrogen, oxygen and nitrogen, and a Porapak R³⁾ column with a TCD and an FID in series for the separation and detection of nitrogen, carbon dioxide and hydrocarbons from C₁ up to C₈. The two analyses are carried out independently and the results are combined.

If oxygen is seen to be present at a concentration greater than 0,02 % (mol/mol) when measuring on molecular sieve, then

1) At present at the stage of draft. (Revision of ISO 683/13-1974.)

2) At present at the stage of draft.

3) 13 X and Porapak R are trade names for commercially available products. At present, no other products intended for this purpose are known to be available commercially. This information is given for the convenience of the user of this International Standard and does not constitute an endorsement of these products by ISO.

the nitrogen value shall be taken from the molecular sieve analysis. If less than 0,02 % (mol/mol) of oxygen is present, the nitrogen value is taken from the Porapak R analysis. Quantitative results are achieved by calibration of the detectors with calibration gas mixtures. The response factors of helium, hydrogen, oxygen, nitrogen, carbon dioxide and the saturated hydrocarbons from methane up to butane are derived from an analysis of calibration gas mixtures. The relative response factors to butane of isobutane and higher hydrocarbons are given in table 5.

The resulting composition of the natural gas is normalized to 100 % (see 8.1).

4 Materials

4.1 Materials for the determination of helium, hydrogen, oxygen and nitrogen

4.1.1 Carrier gas — argon, $\geq 99,99$ % pure¹⁾, free from oxygen and water.

4.1.2 Reference materials.

4.1.2.1 Helium, oxygen and hydrogen, > 99 % pure.¹⁾

4.1.2.2 Nitrogen and methane, $> 99,99$ % pure¹⁾, free from helium, oxygen and hydrogen.

4.2 Materials for the determination of nitrogen, carbon dioxide, and hydrocarbons from C₁ up to C₈

4.2.1 Carrier gas — helium, $\geq 99,99$ % pure¹⁾, free from oxygen and water.

4.2.2 Auxiliary gases.

4.2.2.1 Hydrogen, 99,99 % pure¹⁾, free from corrosive gases, water and organic compounds.

4.2.2.2 Air, free from hydrocarbon impurities.

4.2.3 Reference materials.

4.2.3.1 Nitrogen and methane, $> 99,99$ % pure.

4.2.3.2 Carbon dioxide, ethane, propane and butane, $> 99,95$ % pure¹⁾.

Other reference materials which may be used for peak identification are pentane, hexane, heptane, octane, benzene, toluene, cyclohexane, isobutane, 2-methylbutane, neopentane and methylcyclohexane.

5 Apparatus

5.1 Apparatus for the determination of helium, hydrogen, oxygen and nitrogen

5.1.1 Gas chromatograph, capable of temperature programmed operation and equipped with a TCD.

5.1.1.1 Column oven and temperature control

5.1.1.1.1 Column oven, temperature range 35 to 350 °C, capable of being maintained to within $\pm 0,5$ °C at any temperature in the range during an analysis. To obtain a temperature of 35 °C an accessory for cooling with liquid carbon dioxide or liquid nitrogen may be necessary.

NOTE — Alternative procedures for analysis on the molecular sieve 13 X column are given in clause A.4.

5.1.1.1.2 Temperature control : The oven shall be provided with a linear programmer suitable for providing a rate of temperature increase of 30 °C/min over the specified range.

5.1.1.2 Flow regulators, to give suitable carrier gas flow rates.

NOTE — A more accurate flow controller may be used.

5.1.2 Injection device, bypass-type injector (gas sampling valve), capable of injecting a volume of 1 ml.

5.1.3 Columns

Two columns with the same type of packing and with the same dimensions (see 7.1.1.1.2).

NOTE — The second column is normally used for drift compensation during a temperature programme. If drift compensation is done by means of the electronic integrator a second column is not necessary.

5.1.3.1 Tube

- Nature : stainless steel, number 20 according to ISO 683/13 (AISI type 316), cleaned and degreased
- Length : 1 m
- Diameter : 2 mm internal diameter
- Form : appropriate to the chromatograph
- Radius : appropriate to the chromatograph

5.1.3.2 Packing

5.1.3.2.1 Molecular sieve 13 X

Particle size : 150 to 180 μm (80 to 100 ASTM mesh)

1) If the purity of the gas is less than that specified, it is essential to check that the type of impurity present does not interfere with the analysis. Also, even if the carrier gases argon and/or helium fall within the specification, some of the impurities present in these gases can nevertheless interfere with the analysis. Under these circumstances, appropriate purification is essential.

After packing, treatment overnight at approximately 350 °C with carefully dried carrier gas flowing is necessary for good separation.

5.1.3.2.2 Method of packing

Any method which results in uniform column packing may be used.

NOTE — The following method is suitable.

The column outlet is closed with a sintered disc or glass-wool plug. A reservoir containing rather more packing than is needed to fill the column is connected to the inlet and a pressure of 0,4 MPa of nitrogen is applied to this reservoir. The flow of packing into the column is assisted by vibration. When the column is full, allow the pressure to decay slowly before disconnecting the reservoir.

5.1.3.2.3 Efficiency of resolution

Inject a sample containing equivalent amounts (in the region of 0,4 %) of hydrogen and helium so that the height of the valley between the peaks above the baseline is not greater than 10 % of the height of the larger peak under the operation conditions stated in 7.1. If this criterion is not met continue the activation for a longer period or prepare a new column.

5.1.4 Thermal conductivity detector, with a time constant of not greater than 0,1 s.

5.1.5 Potentiometric recorder

- Range : appropriate to the detector (usually 1 to 10 mV)
- Impedance : > 2 000 Ω
- Response time : ≤ 0,5 s

NOTE — Alternative devices may be used provided that the requirements for response are satisfied.

5.1.6 Integrator

Wide range : 0 to 1 V

The integrator shall be capable of baseline tracking and of measuring peaks on a sloping baseline.

5.2 Apparatus for the determination of nitrogen, carbon dioxide and hydrocarbons from C₁ up to C₈

5.2.1 Gas chromatograph, for the analysis of gases suitable for dual column application equipped with a TCD and an FID in series. The FID shall have a linear dynamic range of six decades and a lower limit of detection of 10⁻¹² g/s of carbon.

5.2.1.1 Column oven and temperature control

5.2.1.1.1 Column oven, temperature range 35 to 300 °C, capable of being maintained to within ± 0,5 °C at any temperature in the range during an analysis. To obtain a

temperature of 35 °C an accessory for cooling with liquid carbon dioxide or liquid nitrogen may be necessary.

5.2.1.1.2 Temperature control : the oven shall be provided with a linear programmer suitable for providing a rate of temperature change of 15 °C/min over the specified range.

5.2.1.2 Flow regulators, to give suitable carrier gas flow rates.

NOTE — A more accurate flow controller may be used.

5.2.2 Injection device, bypass-type injector (gas sampling valve), with heating device, capable of injecting a volume of 0,5 ml.

5.2.3 Columns

Two columns with the same type of packing and with the same dimensions (see 7.2.1.1.2).

NOTE — The second column is normally used for drift compensation during a temperature programme. If drift compensation is done by means of the electronic integrator a second column is not necessary.

5.2.3.1 Tube

- Nature : stainless steel, number 20 according to ISO 683/13 (AISI type 316), cleaned and degreased
- Length : 3 m
- Diameter : 2 mm internal diameter
- Form : appropriate to the chromatograph
- Radius : appropriate to the chromatograph

5.2.3.2 Packing

5.2.3.2.1 Porapak R

Particle size : 150 to 180 μm (80 to 100 ASTM mesh)

After packing, treatment overnight at approximately 230 °C with carefully dried carrier gas flowing is necessary for good separation.

5.2.3.2.2 Method of packing

Any method which results in uniform column packing may be used. A suitable method is described in the note to 5.1.3.2.2.

5.2.3.2.3 Efficiency of resolution

Inject a natural gas sample so that the height of the valley between the 2-methylbutane and pentane peaks above the baseline is not greater than 10 % of the height of the larger peak under the operation conditions stated in 7.2. If this criterion is not met continue the activation for a longer period or prepare a new column.

5.2.4 Detectors

- For components including hydrocarbons up to C₃ : TCD
- For hydrocarbons from C₄ up to C₈ : FID

Ethane and propane can be detected by an FID if the concentration is less than 1 %.

In either case the time constant shall not be greater than 0,1 s. The TCD and FID detectors shall be connected in series.

5.2.5 Potentiometric recorder(s)

In the case of non-automatic change-over from TCD to FID, a dual pen recorder is necessary to detect both signals in parallel.

In the case of automatic change-over from TCD to FID, a single pen recorder is sufficient.

- Range : appropriate to detector (usually 1 to 10 mV)
- Response time : < 0,5 s

NOTE — Alternative devices may be used provided that the requirements for response are satisfied.

5.2.6 Integrator

Wide range : 0 to 1 V

The integrator shall be capable of baseline tracking and of measuring peaks on a sloping baseline.

6 Samples

The method is intended for natural gases containing constituents in the ranges listed in table 1. They do not represent the limit of detection, but the limits within which the stated precision of the method applies. If one or more components in a sample is not detectably present, the method can still be used.

Table 1 — Concentration ranges of natural gas components

Component	Range [% (mol/mol)]
Helium	0,01 to 0,5
Hydrogen	0,01 to 0,5
Oxygen	0,1 to 0,5
Nitrogen	0,1 to 40
Carbon dioxide	0,1 to 30
Methane	50 to 100
Ethane	0,1 to 15
Propane	0,001 to 5
C ₄ hydrocarbons up to octane	0,001 to 0,5

7 Procedure

If the apparatus has been used for a previous determination ensure that it has returned to the starting conditions in 7.1.1 before injecting a sample or calibration gas mixture.

Inject the sample into two columns according to 7.1 and 7.2.

Modifications of the procedure that are not part of the standard are described in annex A.

7.1 Determination of helium, hydrogen, oxygen and nitrogen

7.1.1 Control of the apparatus

Set up the gas chromatograph according to the manufacturer's instructions.

Set the heater of the bypass injector to 50 °C.

7.1.1.1 Oven and column

7.1.1.1.1 Temperature programme :

- a) initial temperature 35 °C for 7 min;
- b) heat at 30 °C/min to 250 °C;
- c) maintain at 250 °C for 10 min.

NOTE — Alternative procedures for analysis on the molecular sieve column are given in clause A.4.

7.1.1.1.2 Carrier gas flow rate

10 ml of argon per minute

7.1.1.2 Detector

Detector temperature setting : between 140 and 160 °C.

Set the detector parameters to the maximum setting allowed by the manufacturer at the conditions given above.

7.1.1.3 Recorder

Set the chart speed to a minimum of 0,5 cm/min.

7.1.1.4 Integrator

Set up the integrator in accordance with the manufacturer's instructions. Attenuate the recorder signal, if desired, independently of the integrator input signal.

7.1.2 Calibration

7.1.2.1 At the beginning of each period of use, after the initial sample injection described in 7.1.3, inject the calibration gas mixture (7.1.2.2) twice. The repeatability shall fall within the

relative values quoted in 8.2.1, given the composition of the calibration gas mixture. If this is not achieved, inject the calibration gas mixture a third time, and check the repeatability of the second and third injection. When two successive injections fulfil the repeatability requirement, take the mean of the two as the calibration value.

Recalibration shall be performed at intervals of not greater than 4 h.

NOTE — If more than three replicates are necessary this implies that there may be errors in the procedure or faults in the apparatus.

7.1.2.2 Use gravimetric high pressure calibration gas mixtures (table 2) prepared according to ISO 6142.

NOTE — For alternative methods for the preparation of calibration gas mixtures, see clause A.5.

Table 2 — Calibration gas mixtures

Component	Gas mixture 1 % (mol/mol)	Gas mixture 2 % (mol/mol)
Helium	0,4	0,4
Oxygen	0,4	—
Hydrogen	—	0,4
Nitrogen ¹⁾		
Methane ²⁾	balance	balance

1) The nitrogen concentration has to be chosen so that it is sufficiently similar to the nitrogen content of the sample (see table 4).

2) In preparing calibration gas mixtures containing oxygen and methane, it is essential that the proportions of oxygen and methane lie outside the explosive limits.

7.1.3 Test

Purge the sample valve with the gas to be analysed, using at least 20 times the volume of the valve and associated pipework.

Stop the purge so as to allow the gas to reach the temperature of the valve and ambient pressure, then inject.

Inject the appropriate calibration gas mixture in the same way.

If this volume of sample is not enough to purge the valve, contamination by air or by the previous sample will be evident. In this case, use a larger volume of sample for purging.

NOTE — It is also possible to introduce the sample into a previously evacuated sample loop and allow it to equilibrate to ambient pressure before injection.

7.1.4 Examination of the chromatogram

A typical chromatogram is shown in figure 1.

The order of elution of components and information on relative retention times for a properly prepared column are given in table 3. Table 3 has to be used with care because relative retention times are unreliable on a molecular sieve column as they depend upon the conditions of preparation and state of use of the column. The column can be regenerated by overnight heating at 350 °C (see 5.1.3.2.1).

Table 3 — Order of elution of components

Component	Relative retention times
Helium	0,17
Hydrogen	0,21
Oxygen	0,38
Nitrogen	0,69
Methane	1,00
Ethane	3,07

Measure the areas of the peaks due to components in the sample and in the calibration gas mixture. Where a component has been measured at different attenuations in sample and calibration gas mixture, convert the measurements to the same attenuation.

7.1.5 Preliminary expression of results

Calculate the concentration of each component of the sample according to the equation

$$X_i = X_{i,c} \left(\frac{A_i}{A_{i,c}} \right) \quad \dots (1)$$

where

X_i is the mole fraction, expressed as a percentage, of substance i in the sample;

$X_{i,c}$ is the mole fraction, expressed as a percentage, of substance i in the calibration gas mixture;

A_i is the area of the peak for substance i on the chromatogram of the sample;

$A_{i,c}$ is the area of the peak for substance i on the chromatogram of the calibration gas mixture.

7.2 Determination of nitrogen, carbon dioxide and hydrocarbons from C₁ up to C₈

7.2.1 Control of the apparatus

Set up the gas chromatograph according to the manufacturer's instructions.

Set the heater of the bypass injector to 100 °C.

7.2.1.1 Oven and column

7.2.1.1.1 Temperature programme :

- initial temperature 35 °C for 3 min;
- heat at 15 °C/min to 200 °C;
- maintain at 200 °C for 30 min.

7.2.1.1.2 Carrier gas flow rate

35 ml of helium per minute

7.2.1.2 Detectors

- TCD temperature setting : between 240 and 260 °C
- Set the bridge current according to manufacturer's instructions.
- FID temperature setting : between 290 and 310 °C

Adjust the hydrogen and air flow rates so as to give optimum response as recommended by the manufacturer. Light the flame in the FID.

7.2.1.3 Recorder(s)

Set the chart speed to a minimum of 0,5 cm/min.

7.2.1.4 Integrator(s)

Set up the integrator(s) in accordance with the manufacturer's instructions. Attenuate the recorder signal independently of the integrator input signal.

7.2.2 Calibration

7.2.2.1 At the beginning of each period of use, after the initial sample injection described in 7.1.3, inject the calibration gas mixture (7.2.2.2) twice. The repeatability shall fall within the relative values quoted in 8.2.1, given knowledge of the composition of the calibration gas mixture. If this is not achieved, inject the calibration gas mixture a third time, and check the repeatability of the second and third injection. When two successive injections fulfil the repeatability requirement, take the mean of the two as the calibration value.

Recalibration shall be performed at intervals of not greater than 4 h.

NOTE — If more than three replicates are necessary this implies that there may be errors in the procedure or faults in the apparatus.

7.2.2.2 Use gravimetric high pressure calibration gas mixtures prepared according to ISO 6142.

The mixture should contain nitrogen, carbon dioxide, methane, ethane, propane and butane.

In order to obtain maximum accuracy, the tolerances between the concentration of the indicated components of the calibration gas mixture and the components of the sample shall be as given in table 4.

Table 4 — Tolerance between concentrations of components in the calibration gas mixture and sample

Sample actual component concentration	Calibration gas mixture deviation of component concentration
% (mol/mol)	% relative to sample concentration
0,001 to 0,1	± 100
0,1 to 1	± 50
1 to 10	± 10
10 to 50	± 5
50 to 100	± 3

7.2.3 Test

Proceed in a similar way to that described in 7.1.3.

7.2.4 Examination of the chromatogram

A typical chromatogram is shown in figure 2.

NOTES

1 Different batches of Porapak R often show slight variation in performance. For example the retention sequence of benzene and cyclohexane may be reversed.

It is therefore recommended that the retention times of benzene and cyclohexane be determined from time to time, and certainly after new columns have been installed.

2 Baseline stability may be checked as follows :

First, raise the oven to the final temperature (7.2.1.1.1) to clear any accumulated contamination.

Second, cool to the initial temperature.

Third, inject a calibration gas mixture containing a low concentration of butane and start the temperature programme (7.2.1.1.1).

Fourth, at the end of the calibration gas run, cool to the initial temperature. Perform a blank run by injecting carrier gas in place of a sample (see 7.1.3) and start the temperature programme.

Fifth, calculate the concentrations of the constituents of which the peaks are recognized by the integrator in the C₅ up to C₈ region by comparison with the butane in the calibration gas.

No individual peak should originate from a constituent with a concentration exceeding 0,04 % (mol/mol). If larger peaks are seen, repeat blank runs until satisfactory. If necessary, prepare new columns, preferably from a different batch of Porapak R.

Measure the areas of the peaks due to components in the sample and in the calibration gas mixture. Where a component has been measured at different attenuations in sample and calibration gas mixture, convert the measurements to the same attenuation.

7.2.5 Preliminary expression of results

Calculate the concentration of each component of the sample according to equation (1).

To determine the cyclic and acyclic hydrocarbon components with 4 up to 8 carbon atoms (these components are not present in the calibration gas mixture), apply relative response factors of the FID for these components compared to butane, either experimentally determined or those given in table 5.

Calculate the concentration of the cyclic and acyclic hydrocarbon components with 4 up to 8 carbon atoms according to the equation

$$X_i = F_i \left(\frac{A_i}{A_b} \right) X_b \quad \dots (2)$$

where

X_i is the mole fraction, expressed as a percentage, of substance i in the sample;

X_b is the mole fraction, expressed as a percentage, of butane in the sample;

F_i is the response factor of the FID of substance i compared to butane;

A_i is the area of the peak for substance i on the chromatogram of the sample;

A_b is the area of the peak for butane on the chromatogram of the sample.

Table 5 — Relative response factors of the FID for natural gas components

Component	F_i (1)2)3)
Propane	1,333
Isobutane	1,000
Butane	1,000
Pentanes	0,800
Hexanes	0,667
Heptanes	0,573
Octanes	0,501
Benzene	0,669
Cyclohexane	0,668
Methylcyclohexane	0,572
Toluene	0,572

1) Butane = 1,000

2) These figures are derived from : KAISER, R. *Chromatographie in der Gasphase*, Part III (1961), p. 136.

3) The values can only be obtained when using the FID under the conditions of 7.2.1.2.

8 Expression of results

For the final calculations of the concentrations of the components in the samples and for the final presentation of the results, two cases can be distinguished :

- a) if less than 0,02 % (mol/mol) of oxygen is present in the samples, then the nitrogen concentration of the samples shall be taken from the Porapak R analysis;

- b) if the oxygen concentration of the samples is greater than 0,02 % (mol/mol), then the nitrogen concentration of the samples shall be taken from the molecular sieve analysis.

8.1 Normalization

If the sum of the concentrations of all the components, $\sum_{i=1}^n X_i$,

determined in the samples through the molecular sieve and/or Porapak R analysis, adds up to between 99 and 101 %, then the individual concentrations, X_i , shall be normalized to 100 % by means of equation (3)

$$X'_i = \frac{X_i}{\sum_{i=1}^n X_i} \times 100 \quad \dots (3)$$

where

X'_i is the mole fraction, expressed as a percentage, of substance i in the sample after normalization;

X_i is the mole fraction, expressed as a percentage, of substance i found from either equation (1) or (2);

n is the total number of components determined separately on both the molecular sieve 13 X and Porapak R column.

The concentrations of the components shall be expressed as mole percentages to three significant figures.

Normally, natural gas does not contain oxygen. If, however, natural gas samples are found to contain oxygen and if this is due to contamination by improper sampling of the gas then the concentration of nitrogen and all the other components shall be corrected according to the following procedure :

- a) the sample contains more than 0,02 % (mol/mol) oxygen. If $X_{O_2} > 0,02$ % (mol/mol), correct the nitrogen concentration according to the equation

$$X_{N_2,c} = X'_{N_2} - \frac{78}{21} X'_{O_2} \quad \dots (4)$$

where

$X_{N_2,c}$ is the mole fraction, expressed as a percentage, of nitrogen after correcting the concentration for air contamination;

X'_{N_2} is the mole fraction, expressed as a percentage, of nitrogen in the sample after normalization;

X'_{O_2} is the mole fraction, expressed as a percentage, of oxygen in the sample after normalization.

NOTE — In equation (4) it is assumed that the TCD responses for N_2 and O_2 are equal.

b) the sample contains less than 0,02 % (mol/mol) oxygen. If $X_{O_2} < 0,02$ % (mol/mol) correct the nitrogen concentration according to 1) or 2)

- 1) If the nitrogen analysis has been carried out on the molecular sieve 13 X column, equation (4) shall be applied.
- 2) If the nitrogen analysis has been carried out on the Porapak R column, apply equation (5)

$$X_{N_2,c} = X'_{N_2} - \frac{100}{21} X'_{O_2} \quad \dots (5)$$

NOTE — In equation (5) it is assumed that the TCD responses for N_2 , O_2 and Ar are equal.

The concentrations X'_i of all the components except oxygen, shall be normalized to 100 % according to the equation

$$X''_i = \frac{X_i}{\sum_{i=1}^{n-2} X'_i + X_{N_2,c}} \times 100 \quad \dots (6)$$

where

X''_i is the mole fraction, expressed as a percentage, of substance i in the sample;

X'_i is the mole fraction, expressed as a percentage, of substance i in the sample after normalization;

n is the total number of components;

$n - 2$ is the total number of components except oxygen and nitrogen, determined separately on both the molecular sieve 13 X and Porapak R column.

8.2 Precision and accuracy

8.2.1 Repeatability

The repeatability of the method has been evaluated. The data are only valid within the ranges quoted.

Table 6 — Repeatability

Concentration range	Repeatability ¹⁾	
	% (mol/mol)	% (mol/mol)
10 to 100	± 0,1	± 0,1
1 to 10	± 1	± 0,02
0,1 to 1	± 10	± 0,01
0,01 to 0,1	± 30	± 0,005
0,001 to 0,01	± 50	± 0,002

1) Whichever value is the larger.

8.2.2 Reproducibility

At present the determination of reproducibility is still under investigation.

8.2.3 Accuracy

It is very difficult to give figures on the accuracy of the method here because the accuracy is mainly determined by the choice of calibration gas and the linearity of the detectors. Generally, to obtain best results, i.e. with the highest accuracy, there has to be a close resemblance between the calibration gas and the sample to be analysed.

9 Test report

The test report shall include the following information :

- a) a reference to this International Standard;
- b) all information necessary for complete identification of the sample, for example
 - the date of sampling;
 - the place in the pipeline system at which the sample was taken;
- c) the sampling method used (including the size and type of material of the high pressure cylinder);
- d) any deviation from the procedure specified;
- e) a complete list of all components determined or detected;
- f) if possible the precision of the determined concentrations of the components of the sample, including the number of determinations;
- g) any unusual features noted during the determination (i.e. in the chromatogram).




Figure 1 — Analysis of helium, hydrogen, oxygen and nitrogen on the molecular sieve 13 X column, with indication of the absolute retention time, in minutes

STANDARDISO.COM : Click to view the full PDF of ISO 6974:1984




Figure 2 – Analysis of nitrogen (oxygen), carbon dioxide and hydrocarbons from C₁ up to C₈ on the Porapak R column