

INTERNATIONAL STANDARD

**ISO
7494**

First edition
1990-04-01

Dental unit

Unit dentaire

STANDARDSISO.COM : Click to view the full PDF of ISO 7494:1990



Reference number
ISO 7494:1990(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 7494 was prepared by Technical Committee ISO/TC 106, *Dentistry*.

STANDARDSISO.COM : Click to view the full PDF of ISO 7494:1990

© ISO 1990

All rights reserved. No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization
Case Postale 56 • CH-1211 Genève 20 • Switzerland

Printed in Switzerland

Introduction

This International Standard takes priority over IEC 601-1 as specified in the individual clauses of this International Standard.

Only the specifications laid down in this International Standard are applicable.

This International Standard refers to IEC 601-1:1977, the basic standard on safety of medical electrical equipment, wherever relevant, by stating the respective clause numbers of IEC 601-1.

STANDARDSISO.COM : Click to view the full PDF of ISO 7494:1990

PROVISIONAL TITLE

When the standard is published, the following text will appear in the title page of the standard:

Application of the standard is intended for use in the following circumstances:

1. The standard is intended for use in the following circumstances:

2. The standard is intended for use in the following circumstances:

STANDARDSISO.COM : Click to view the full PDF of ISO 7494:1990

Dental unit

1 Scope

This International Standard applies to all dental units whether they are electrically equipped or not.

It specifies requirements, test methods and marking details for items of equipment used for delivering and storing dynamic and static instruments such as handpieces, syringes, saliva ejectors, cuspidors, high volume evacuators, but specifically excludes Class III equipment.

Specifications for instruments will be the subject of future, separate International Standards.

2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this International Standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 1942-4:1989, *Dental vocabulary — Part 4: Dental equipment*.

ISO 4211:1979, *Furniture — Assessment of surface resistance to cold liquids*.

ISO 9687:—¹⁾, *Dental equipment — Graphical symbols*.

IEC 601-1:1977, *Safety of medical electrical equipment — Part 1: General requirements*.

1) To be published.

3 Definitions

For the purposes of this International Standard, the following definition and, where relevant, definitions given in clause 2 of IEC 601-1:1977 apply.

dental unit: Item of dental equipment through which electrical power and/or various fluids or gases are supplied to a number of dental instruments and devices. It is usually fitted with conveniently orientated instrument holders and controls, and consists of interconnected sub-units of dental equipment and instruments providing a functional unit for dental use.

4 Classification

4.1 According to type of protection against electric shock

Dental units may be classified as follows:

a) Class I equipment

Equipment in which protection against electric shock does not rely on basic insulation only, but which includes an additional safety precaution in such a way that means are provided for the connection of accessible conductive parts to the protective (earth) conductor in the fixed wiring of the installation in such a way that accessible conductive parts cannot become live in the event of a failure of the basic insulation.

b) Class II equipment

Equipment in which protection against electric shock does not rely on basic insulation only, but in which additional safety precautions such as double insulation or reinforced insulation are provided, there being no provision for protective earthing or reliance upon installation conditions.

4.2 According to degree of protection against electric shock

Dental units may be grouped by type as follows:

a) Type B equipment

Class I or II equipment or equipment with an internal electrical power source providing an adequate degree of protection against electric shock particularly regarding:

- allowable leakage currents;
- reliability of the protective earth connection (if present).

Type B equipment is, for example, suitable for intentional external and internal application to the patient, excluding direct cardiac application.

b) Type BF equipment

Type B equipment with an F-type isolated (floating) applied part.

4.3 According to mode of operation

Dental units are a type of equipment with intermittent operation.

4.4 According to degree of electrical connection between dental unit and patient

4.4.1 Equipment with an applied part but not specifically designed for applications where a functional conductive connection to the patient is made, for example, ultrasonic equipment, dental patient chairs.

4.4.2 Equipment without an applied part in conductive connection with the patient, but specifically designed for applications where an electrical or mechanical connection is made with equipment having an applied part with a conductive connection to the patient, for example, computers, displays.

4.5 Marking or identification

Marking or identification of classifications shall comply with clause 10.

5 Requirements

This clause specifies all requirements relevant to dental units. Many of these characteristics are quantitatively verifiable as detailed in the relevant subclause of clause 7. Some are not but are still objectively verifiable by visual inspection. Some re-

quire a subjective decision by qualified test personnel as concerns compliance with the requirement. It is envisaged that these tests will be replaced by quantitative tests as soon as results of relevant research work are available.

Electrical requirements are only applicable to dental units with electrical devices. There are, however, general requirements in IEC 601-1 referred to which are applicable to non-electrical dental units as well.

All terms used in connection with electrical requirements shall conform in definition to clause 2 of IEC 601-1:1977.

In the following clauses reference is made to clauses of IEC 601-1:1977. These clauses not only cover the respective requirements, but also the tests required.

5.1 Environmental conditions

5.1.1 Operation

Clause 1.4 b) of IEC 601-1:1977 applies.

5.1.2 Power supply

The dental unit shall operate from a supply mains with the following characteristics:

- a) a rated voltage not exceeding 250 V single-phase;
- b) a sufficiently low internal impedance so as not to interfere with proper performance of the unit;
- c) voltage fluctuations generally not exceeding $\pm 10\%$ of the nominal voltage, not including short-time fluctuations (for example, duration less than 1 s) at irregular intervals such as caused by operation of X-ray generators or similar equipment;
- d) voltages which are practically sinusoidal and forming a practically symmetrical supply system in case of polyphase supply;
- e) a frequency of not more than 1 kHz;
- f) a frequency which does not deviate by more than 1 Hz from the nominal value up to 100 Hz and by more than 1 % from 100 Hz to 1 kHz;
- g) the protective measures to be specified in a future IEC International Standard on electrical installations in hospital and in medically used rooms outside hospitals.

Testing shall be carried out in accordance with 7.2, or by using readily available measuring instruments.

5.2 General design

5.2.1 Recommendations

5.2.1.1 The dental unit should incorporate provisions to hold or store one or more dental instruments and/or accessories such as syringes, lights, high volume evacuators, saliva ejectors, instrument tray holders, in such a manner and orientation as to provide for safety and convenience during normal dental procedures, and for good ergonomic working conditions.

5.2.1.2 Dental units should be designed, constructed and manufactured so that, when properly transported, stored, installed, used and maintained according to the manufacturer's instructions, they cause no danger which could reasonably be foreseen, to the patient, to the operating personnel, or to the surroundings in normal use and in single fault condition.

5.2.1.3 The dental unit should have the strength and rigidity necessary to resist the stresses encountered in normal dental practice without introducing fire, electrical shock, or total or partial collapse of the dental unit, which can cause injury to the patient and operating personnel. The unit shall be constructed so as to resist spillage and ingress of moisture which could result in deterioration of electrical components when tested in accordance with IEC 601-1:1977, clause 44.3.

5.2.1.4 Edges and corners of components and parts accessible to the patient or personnel should be finished in such a manner as to avoid injury to the patient or operator.

5.2.2 Testing

Testing shall be carried out in accordance with 7.2.

If the product also passes all the following tests described in this International Standard, it shall be considered that these requirements are fulfilled.

5.3 Operating controls

Operating controls should be designed and located to minimize accidental activation.

For marking, see 10.3.

Testing shall be carried out in accordance with 7.2.

5.4 Cuspidors

The cuspidor bowl, if provided, shall be of a material that is capable of undergoing frequent antiseptic cleaning in accordance with instructions provided by the manufacturer without signs of deterioration.

Testing shall be carried out in accordance with 7.3.

5.5 Solids collector

The dental unit shall contain a solids collector in the unit's waste system. The solids collector should be capable of filtering anything 2 mm or larger on its lesser diameter.

Testing shall be carried out in accordance with 7.2.

5.6 Amalgam separator device

The dental unit shall be capable of being connected to an amalgam separator device in the waste system.

Testing shall be carried out in accordance with 7.2.

5.7 Fallsafe device

The dental unit shall incorporate failsafe devices such as mechanical limits which prevent injury to patients and/or operating personnel, and to avoid damage to the dental unit due to the failure of a power actuator and/or failsafe switch.

The failsafe device shall also offer protection against hazards which might arise from the close connection with a dental chair to which the dental unit is attached.

Testing shall be carried out in accordance with 7.4.

5.8 Moving parts

Moving parts that may constitute a hazard shall be protected or guarded. The distance between power-activated moving parts which are accessible to hands or fingers shall be less than 10 mm when fully opened, or a minimum of 20 mm when fully closed. Adequate safety features shall be provided for feet or legs which are accessible to power-activated moving parts. Such precautions could include limit switch plates with mechanical limits.

Testing shall be carried out in accordance with 7.2.

5.9 Bursting pressure

Pressure systems used in a dental unit the rupture of which would constitute a hazard to the patient and operating personnel shall be strong enough to withstand without bursting or leaking the working pressure specified by the manufacturer; see 8.3.

Testing shall be carried out in accordance with 7.5.

5.10 Pressure relief

Means for safe relief of pressure shall be provided for all parts in which pressure might be generated, e.g. in the event of fire. Such means may include pressure-relief devices, fused plugs, soldered joints, or non-metallic tubing.

Testing shall be carried out in accordance with 7.2.

5.11 Cleaning and disinfection

All exterior parts of the dental unit shall be capable of undergoing cleaning and disinfection without causing deterioration to the unit's surface or markings by using agents recommended by the manufacturer.

Testing shall be carried out in accordance with 7.6.

NOTE 1 The requirements in 5.12 to 5.34 are primarily applicable to electrically equipped dental units.

5.12 Power input

Clause 7 of IEC 601-1:1977 applies.

5.13 Single fault conditions

Clause 12 of IEC 601-1:1977 applies.

5.14 Protection against electric shock hazards

Clause 13 of IEC 601-1:1977 applies.

5.15 Requirements related to classification

5.15.1 Class I equipment

Clause 14.1 of IEC 601-1:1977 applies.

5.15.2 Class II equipment

Clause 14.2 of IEC 601-1:1977 applies.

5.15.3 Classes I and II equipment

Clause 14.4 of IEC 601-1:1977 applies, limited to classes I and II.

5.15.4 Types B and BF

Clause 14.6 of IEC 601-1:1977 applies, limited to types B and BF.

5.16 Limitation of voltage and/or current

Clause 15 of IEC 601-1:1977 applies.

5.17 Enclosures and protective covers

Clause 16 of IEC 601-1:1977 applies.

5.18 Insulation and protective impedances

Clause 17 of IEC 601-1:1977 applies.

5.19 Earthing and potential equalization

Clauses 18 a) to g) of IEC 601-1:1977 apply.

5.20 Continuous leakage currents and patient auxiliary currents

The maximum allowable values shall be as specified in table 1.

Testing shall be carried out in accordance with 7.7.

5.21 Dielectrical strength

The dielectrical strength of the components and groups of components of the dental unit shall be sufficient to withstand the voltages specified.

Testing shall be carried out in accordance with 7.8.

5.22 Stability and transportability

Clause 24 of IEC 601-1:1977 applies.

5.23 Excessive temperatures

Clause 42 of IEC 601-1:1977 applies.

5.24 Spillage and splash-proofness

Clauses 44.3 and 44.6 of IEC 601-1:1977 apply.

5.25 Leakage

Clause 44.4 of IEC 601-1:1977 applies.

5.26 Human errors

Clauses 46.1 to 46.4 of IEC 601-1:1977 apply.

5.27 Interruption of power supply

Clauses 49.1 to 49.3 of IEC 601-1:1977 apply.

5.28 Accuracy of operating data

Clause 50.1 of IEC 601-1:1977 applies.

Table 1 — Maximum allowable values of continuous leakage currents and patient auxiliary currents

Values in milliamperes

Current path	Type B		Type BF	
	N.C. 1)	S.F.C. 2)	N.C. 1)	S.F.C. 2)
Earth leakage current	0,5	1 ³⁾ , 4)	0,5	1 ³⁾ , 4)
Enclosure leakage current	0,1	0,5	0,1	0,5
Patient leakage current	0,1	0,5	0,1	0,5
Patient leakage current (mains voltage on the signal input part or signal output part)	—	5	—	—
Patient leakage current (mains voltage on the applied part)	—	—	—	5
Patient auxiliary current	0,01 0,1	0,5	0,01 0,1	0,5

1) N.C.: Normal condition.
2) S.F.C.: Single fault condition.
3) The only single fault condition for the earth leakage current is the interruption of one supply conductor at a time [see clause 19.2 a) and figure 16 of IEC 601-1:1977].
4) See clause 19.3 e) of IEC 601-1:1977.

5.29 Fault conditions causing overheating and/or mechanical damage

Clause 52 of IEC 601-1:1977 applies.

5.30 Enclosures and covers

Clause 55.2 of IEC 601-1:1977 applies.

5.31 Components and general assembly

Clauses 56.1 b) and d) of IEC 601-1:1977 apply.

Clauses 56.2 a), b), d), e), f) to 56.10 of IEC 601-1:1977 apply.

Clause 56.11 b) of IEC 601-1:1977 applies.

5.32 Mains parts, components and layout

Clause 57 of IEC 601-1:1977 applies.

5.33 Protective earth terminals

Clause 58 of IEC 601-1:1977 applies.

5.34 Construction and layout

Clause 59 of IEC 601-1:1977 applies.

6 Sampling

At least one dental unit for each model series shall be evaluated for compliance with this International Standard.

7 Testing

7.1 General information for tests

7.1.1 General provisions

The sequence of tests shall be according to IEC 601-1:1977, appendix C.

All tests described in this International Standard are type tests. Unless otherwise specified, tests shall not be repeated. This applies specifically to the dielectric strength tests, which shall be made only on the manufacturer's premises or in independent test laboratories.

Since some of the tests described are destructive tests, the dental unit tested is not to be used afterwards.

The rating of components shall be inspected to check that they are appropriate for the application intended.

Where a component or equipment part has specified ratings exceeding those appropriate to its use in the equipment, it does not have to be tested for such a wider range.

Compliance is considered to be fulfilled if all relevant tests of this International Standard are passed successfully.

Dental units, or parts thereof, using materials or having forms of construction different from those detailed in this International Standard shall be acceptable if it can be demonstrated that an equivalent degree of safety is obtained.

7.1.2 Ambient temperature, humidity, atmospheric pressure

7.1.2.1 After the dental unit being tested has been set up for normal use, tests shall be carried out under operating conditions at

- a) an ambient temperature within the range from 15 °C to 35 °C;
- b) a relative humidity within the range from 45 % to 75 %;
- c) an atmospheric pressure within the range from 860 mbar to 1060 mbar (645 mmHg to 795 mmHg).

7.1.2.2 The equipment shall be protected from other conditions, which might affect the validity of the tests (for example, draughts).

7.1.2.3 In cases where ambient temperatures cannot be maintained, the test conditions shall be consequently modified and the results adjusted accordingly.

7.1.3 Other conditions

Clauses 4.6 a), b), d) of IEC 601-1:1977 apply.

7.1.4 Supply and test voltages, type of current, nature of supply, frequency

Clause 4.7 of IEC 601-1:1977 applies.

7.1.5 Preconditioning

Clause 4.8 of IEC 601-1:1977 applies.

7.1.6 Repairs and modifications

Clause 4.9 of IEC 601-1:1977 applies.

7.1.7 Moisture preconditioning treatment

Clause 4.10 of IEC 601-1:1977 applies.

7.2 Visual inspection

Visually inspect the equipment with the naked eye at normal visual acuity to determine compliance with the requirements.

7.3 Cuspidor

The cuspidor shall be treated 20 times with disinfecting agents recommended by the manufacturer before being visually inspected for compliance with the requirement.

7.4 Fallsafe device

On dental units which are power-actuated and controlled by limit switches, such limit switches shall be purposely bypassed, one at a time, while the dental unit operates through its full range of motion, without resulting in collapse of the dental unit that would be harmful to the patient and/or operating personnel.

7.5 Bursting pressure

Two samples each of the pressure systems shall be subjected to a hydrostatic pressure.

Connect the system to a suitable hydraulic pump. Raise the pressure gradually to 40 % of the final test pressure (three times the working pressure specified by the manufacturer) and hold at that pressure for 60 s. Then increase the pressure to three times the specified working pressure and hold for 3 min. The results are unacceptable if the sample either bursts or leaks.

7.6 Cleaning and disinfection

Cleaning and disinfection test shall be in accordance with ISO 4211. Cleaning and disinfection agents shall be applied for 24 h.

7.7 Continuous leakage currents and patient auxiliary currents

The earth leakage current, the enclosure leakage current, the patient leakage current and the patient auxiliary current shall be tested

- a) after the dental unit has been brought up to normal operating temperature in accordance with the requirements of clause 7 of IEC 601-1:1977;
- b) after the moisture preconditioning treatment as described in clause 4.10 of IEC 601-1:1977. The measurements shall be carried out with equipment located outside the humidity cabinet and shall commence 1 h after equipment has been

taken out of this cabinet, and has been placed in an environment with a temperature less than or equal to the temperature of the humidity cabinet. During testing, those measurements which do not energize equipment shall be made first.

7.8 Dielectrical strength

The test voltage for single-phase equipment and for three-phase equipment (to be tested as single-phase equipment) shall be applied to the insulation parts as described in clauses 20.1 and 20.2 of IEC 601-1:1977 during 1 min and according to tables V, VI and VII of IEC 601-1:1977:

- a) immediately after warming up to operating temperature and switching off the equipment, and for heating elements by application of the circuit of figure 28 of IEC 601-1:1977, and
- b) immediately after the moisture preconditioning treatment (as described in clause 4.10 of IEC 601-1:1977) with the equipment de-energized during the test and kept in the humidity cabinet, and
- c) after any required sterilization procedure with the equipment de-energized (see clause 44.7 of IEC 601-1:1977).

Initially, not more than half the prescribed voltage shall be applied, then it shall be raised over a period of 10 s to the full value, which shall be maintained for 1 min.

8 Manufacturer's instructions

Dental units shall be accompanied by documents containing relevant information as specified in the following sub-clauses.

Testing of compliance of the requirements of 8.1 to 8.3 shall be carried out in accordance with 7.2.

8.1 General

Clause 6.8.1 of IEC 601-1:1977 applies.

Information should be available in the language of the country where the product is to be used.

8.2 Instructions for use

Clauses 6.8.2 a), c), d), e), f), g) of IEC 601-1:1977 apply.

In addition, the full range of motion shall be quoted.

If mounting accommodation for other manufacturer's attachments are provided, the manufacturer shall stipulate the maximum loading capabilities for those

attachments in the instructions for use included with each dental unit.

8.3 Technical description

Clauses 6.8.3 a), b) and c) of IEC 601-1:1977 apply.

In addition, the following information shall be provided by the manufacturer:

- a) overall dimensions of the dental unit;
- b) overall dimensions of the base plate and service location interfaces, if applicable;
- c) details of interface surfaces and methods of retention (bolts, etc.) and electrical supplies and other services;
- d) information on the assembly and mounting of the dental unit;
- e) electrical characteristics (voltage, frequency, fuse values);
- f) step-by-step procedures and operation and routine maintenance of the dental unit, including illustrations showing the location of and explanation of each control and other features relating to safety consideration in the intended use;
- g) instructions for use shall also include warning statements on the use of the dental unit in conjunction with other equipment that may move;
- h) directions for cleaning and disinfecting the dental unit;
- i) dental units shall contain data on the mass and maximum movement of the unit and its accessories generated at the mounting position, with the unit and accessories in the most unfavourable position;
- j) the mass of non-chair mounted units shall also be indicated;
- k) the maximum safe load capabilities of the working surface in its most unfavourable position;
- l) standard attachments that the dental unit is designed to accept and loading capabilities for these attachments;
- m) the colour code of the tubing used in the dental unit;
- n) full fluid characteristics for the input and output connections to the dental unit;

- o) a list of spare parts that would be required in general use;
- p) schematic wiring diagrams;
- q) minimum spare requirements and recommendations for unit installation within the dental operator;
- r) working pressures of pressure systems used in the dental unit.

9 Packaging

Testing of compliance of the requirements of 9.1 and 9.2 shall be carried out in accordance with 7.2.

9.1 Dental units shall be packaged for transportation in such a way that no damage may occur during anticipated transport conditions.

9.2 Component parts shall be specifically identified on the outside to facilitate correct assembly.

10 Marking

Testing of compliance of the requirements of 10.1 to 10.6 shall be carried out in accordance with 7.2.

10.1 Marking on outside of mains-operated units

Marking on the outside of dental unit or unit parts and on the outside of the dental unit chairs or dental unit parts without direct supply mains connection shall be as specified in table 2.

Mains-operated dental units, including separable components thereof, which have a mains part, shall be provided at least with permanently affixed and clearly legible markings on the outside of the major part of the dental units as described in table 2.

10.1.1 Indication of origin

Clause 6.1 e) of IEC 601-1:1977 applies.

10.1.2 Model of type reference

Clause 6.1 f) of IEC 601-1:1977 applies.

10.1.3 Connection of supply

Clause 6.1 g) of IEC 601-1:1977 applies.

10.1.4 Supply frequency (in hertz)

Clause 6.1 h) of IEC 601-1:1977 applies.

Table 2 — Marking on outside of dental unit

Requirements as specified in clauses	Subject	Mains-operated equipment
10.1.1	Indication of origin	x
10.1.2	Model or type reference	x
10.1.3	Connection to the supply	x ¹⁾
10.1.4	Supply frequency (Hz)	x ¹⁾
10.1.5	Power input	x ¹⁾
10.1.6	Classification	x ²⁾
10.1.7	Mode of operation	x ²⁾
10.1.8	Fuses	x ²⁾
10.1.9	Mechanical stability	x ²⁾

1) For non-permanently installed equipment only. For other equipment, see also 6.2 a) of IEC 601-1:1977.
2) If applicable.

10.1.5 Power input

Clause 6.1 j) of IEC 601-1:1977 applies.

10.1.6 Classification

Clause 6.1 l) of IEC 601-1:1977 applies.

10.1.7 Mode of operation

Clause 6.1 m) of IEC 601-1:1977 applies.

10.1.8 Fuses

Clause 6.1 n) of IEC 601-1:1977 applies.

10.1.9 Mechanical stability

Clause 6.1 u) of IEC 601-1:1977 applies.

10.2 Marking on inside of dental units or their parts

Clauses 6.2 a), e), f), h) j), k), l), m) of IEC 601-1:1977 apply.

10.3 Marking of operating controls

Clauses 6.3 b), c), e) of IEC 601-1:1977 apply.

The mains switch shall be clearly identified.