

---

---

**Microscopes — Values, tolerances and  
symbols for magnification**

*Microscopes — Valeurs, tolérances et symboles de grossissement*

STANDARDSISO.COM : Click to view the full PDF of ISO 8039:2014



STANDARDSISO.COM : Click to view the full PDF of ISO 8039:2014



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2014

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

	Page
Foreword .....	iv
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>1</b>
<b>4 Symbols for magnification of imaging components</b> .....	<b>1</b>
<b>5 Values and tolerances for magnification</b> .....	<b>1</b>
5.1 Values for magnification .....	1
5.2 Tolerance of values of magnification for imaging components .....	1
<b>Bibliography</b> .....	<b>3</b>

STANDARDSISO.COM : Click to view the full PDF of ISO 8039:2014

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT), see the following URL: [Foreword — Supplementary information](#).

The committee responsible for this document is ISO/TC 172, *Optics and photonics*, Subcommittee SC 5, *Microscopes and endoscopes*.

This third edition cancels and replaces the second edition (ISO 8039:2012), of which it constitutes a minor revision.

# Microscopes — Values, tolerances and symbols for magnification

## 1 Scope

This International Standard specifies a series of values, tolerances, and symbols for the magnification of imaging components of light microscopes and defines a number of imaging components and magnifying systems to which they apply.

## 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 10934-1, *Optics and optical instruments — Vocabulary for microscopy — Part 1: Light microscopy*

## 3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 10934-1 apply.

## 4 Symbols for magnification of imaging components

[Table 1](#) gives the symbols that shall be used when referring to the magnification of imaging components and combinations thereof and gives examples of methods of expression.

## 5 Values and tolerances for magnification

### 5.1 Values for magnification

Values for the magnification of imaging components or magnifying systems should be one of the values given in [Table 2](#). The products or quotients of any two values in the table are also to be considered as values within the table. The table can be extended by a factor of 10 per row.

### 5.2 Tolerance of values of magnification for imaging components

Tolerances of values of magnification shall be as given in [Table 3](#).

**Table 1 — Symbols for magnification and methods of expression**

Component	Symbol	Methods of expression	
		Preferred	Alternative
Objective: a) corrected for finite primary image distance b) corrected for infinite primary image distance	$M_O$ $M_{O\infty}$	$M_O = 25 : 1$ $M_{O\infty} = 25 \times$	25 : 1 or 25 25 ×
Eyepiece	$M_E$	$M_E = 10 \times$	10 ×
Tube lens	$q$	$q = 1,25 \times$	1,25 ×
Projection lens	$M_{PHOT}$	$M_{PHOT} = 2,5 : 1$	2,5 : 1
Total (lateral) magnification of microscope: a) for visual observation b) for real image	$M_{TOT VIS}$ $M_{TOT PROJ}$	$M_{TOT VIS} = 500 \times$ $M_{TOT PROJ} = 500 : 1$	500 × 500 : 1

**Table 2 — Values for magnification**

					...0,32	0,4	0,5	0,63	0,8
1	1,25	1,6	2	2,5	3,2	4	5	6,3	8
10	12,5	16	20	25	32	40	50	63	80
100	125	160	200	250	320	400	500	630	800
1 000	1 250	1 600	2 000...						

NOTE 1 The values have been taken from the R10 series in ISO 3:1973.  
 NOTE 2 The value 0,32 has been rounded from its R10 series value.  
 NOTE 3 Besides the values in this table, the following values are also in use: 1,5 – 15 – 30 – 60 – 150.

**Table 3 — Tolerances on magnification**

System/component	Tolerance %
Objective	±5
Tube lens	±2
Projection lens	±2
Eyepiece	±5