

# INTERNATIONAL STANDARD

# ISO 8041

First edition  
1990-07-15

**AMENDMENT 1**  
1999-11-01

---

---

## Human response to vibration — Measuring instrumentation

### AMENDMENT 1

*Réponse des individus aux vibrations — Appareillage de mesure*  
AMENDEMENT 1

STANDARDSISO.COM : Click to view the full PDF of ISO 8041:1990/Amd 1:1999



Reference number  
ISO 8041:1990/Amd.1:1999(E)

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Amendment 1 to ISO 8041:1990 was prepared by Technical Committee ISO/TC 108, *Mechanical vibration and shock*, Subcommittee SC 3, *Use and calibration of vibration and shock measuring instruments*.

This amendment is a result of the resolution taken by ISO/TC 108/SC 3, September 1995, to revise ISO 8041:1990. In this first step, ISO 8041 is made compatible with the revised ISO 2631-1.

This amendment specifies instrumentation characteristics for measurement of whole-body vibration in accordance with redefined frequency weightings and alternative evaluation procedures introduced in ISO 2631-1:1997. Specification, calibration and verification tests for instrumentation have not been agreed for all evaluation procedures and these will be addressed in a future revision of ISO 8041.

The instrumentation specifications for measurement according to ISO 5349 (hand-arm) and ISO 2631-2 (whole-body combined) vibration are unchanged. However, data for these frequency weightings are now presented in the style used throughout this amendment.

© ISO 1999

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization  
Case postale 56 • CH-1211 Genève 20 • Switzerland  
Internet iso@iso.ch

Printed in Switzerland

# Human response to vibration — Measuring instrumentation

## AMENDMENT 1

Page 1

### 2 Normative references

Add year of publication, 1990, to ISO 2041.

Change year of publication of ISO 2631-1 to 1997 and change the title to *Mechanical vibration and shock — Evaluation of human exposure to whole-body vibration — Part 1: General requirements*.

Change year of publication of ISO 5805 to 1997 and change title to *Mechanical vibration and shock — Human exposure — Vocabulary*.

Delete ISO 2631-3.

Replace the two lines concerning IEC 225 by:

IEC 61260:1995, *Electroacoustics — Octave-band and fractional-octave-band filters*

IEC 60651:1979, *Electroacoustics — Sound level meters*

IEC 60651:1979/Amd.1:1993, Amendment 1.

Page 2

### 3.1 weighted vibration

Replace first paragraph on page 2 by:

The acceleration is weighted in accordance with one of the eight frequency weightings listed in table 1 and specified in tables 4 to 11. Exact definitions are presented in annex B.

Replace table 1 by the following table:

**Table 1 — Frequency weightings**

Characteristics of vibration	Nominal frequency range Hz	International Standard
Whole-body vibration: Principal weightings: Whole body, vertical direction $z$ , designated $W_k$	0,5 to 80	ISO 2631-1:1997
Whole body, horizontal directions $x, y$ , designated $W_d$	0,5 to 80	ISO 2631-1:1997
Whole body, motion sickness, vertical direction $z$ , designated $W_f$	0,1 to 0,5	ISO 2631-1:1997
Whole body, combined directions, designated W.B.combined	1 to 80	ISO 2631-2:1989
Additional weightings: Whole body, seat back horizontal direction $x$ , designated $W_c$	0,5 to 80	ISO 2631-1:1997
Whole body, rotational vibration $r_x, r_y, r_z$ , designated $W_e$	0,5 to 80	ISO 2631-1:1997
Head of recumbent person, vertical direction $x$ , designated $W_j$	0,5 to 80	ISO 2631-1:1997
Hand-arm vibration: All directions $x, y, z$ , designated $W_h$	8 to 1000	ISO 5349:1986

NOTE Additional applications of the frequency weightings are described in ISO 2631-1.

Change the paragraph under table 1 to:

The frequency-weighted acceleration shall be integrated to a root-mean-square value over the duration of the measurement using linear averaging (see 3.3.1). The weighted root-mean-square acceleration may be calculated from acceleration spectra (see 4.3.2). The weighted acceleration may also be integrated to a running root-mean-square acceleration (see 3.3.3), vibration dose value (see 3.3.4), or motion sickness dose value (see 3.3.5). When quoting the weighted acceleration, the frequency weighting and the averaging method, including the appropriate integration time or time constant, shall be indicated.

Change NOTE to NOTE 1; add note 2:

NOTE 2 For the evaluation of comfort in some environments (e.g. rail vehicles), a frequency weighting designated  $W_b$  may be applied in vertical direction ( $z$ ). This weighting curve deviates from the  $W_k$  weighting (see annex B) and is defined in BS 6841:1987.

**3.3 Equivalent continuous vibration value and level**

Replace the whole clause by the following.

**3.3 Root-mean-square and fourth-power values**

**3.3.1 weighted root-mean-square acceleration value**

The weighted r.m.s. acceleration  $a_w$ , in metres per second squared or radians per second squared, is defined by the expression:

$$a_w = \sqrt{\frac{1}{T} \int_0^T a_w^2(t) dt} \tag{1}$$

where

$a_w(t)$  is the weighted acceleration (translational or rotational) as a function of time (time history), in metres per second squared ( $m/s^2$ ) or radians per second squared ( $rad/s^2$ ), respectively;

$T$  is the duration of the measurement.

### 3.3.2 weighted root-mean-square acceleration level

The weighted r.m.s. acceleration level expressed in decibels, is defined by

$$L_{a,w} = 10 \cdot \lg(a_w/a_0)^2 \text{ dB} \quad (2)$$

where

$a_w$  is defined in 3.3.1;

$a_0$  is the reference acceleration ( $10^{-6}$  m/s<sup>2</sup>).

### 3.3.3 weighted running root-mean-square acceleration value

The weighted running r.m.s. acceleration  $a_w(t_0)$ , in metres per second squared, at the instantaneous observation time  $t_0$  is defined by the expression:

$$a_w(t_0) = \sqrt{\frac{1}{\tau} \int_{t_0-\tau}^{t_0} a_w^2(t) dt} \quad (3)$$

where

$a_w(t)$  is the weighted instantaneous acceleration magnitude, in m/s<sup>2</sup>;

$\tau$  is the integration time of the measurement, in seconds, preferably 1 s;

$t$  is the time (integration variable).

Exponential averaging may be used for the running r.m.s. method, as an approximation of the linear averaging. The exponential averaging is defined as follows (see also annex D):

$$a_w(t_0) = \sqrt{\frac{1}{\tau} \int_{-\infty}^{t_0} a_w^2(t) \exp\left(-\frac{t-t_0}{\tau}\right) dt} \quad (4)$$

where

$a_w(t)$  is the weighted instantaneous acceleration magnitude, in m/s<sup>2</sup>;

$\tau$  is the time constant of the measurement, in seconds, preferably 1 s;

$t$  is the time (integration variable).

The difference in the result is very small for application to shocks of a short duration compared to  $\tau$ , it is somewhat larger (up to 30 %) when applied to shocks and transients of longer duration.

The maximum of the running r.m.s. value is denoted MTVV (maximum transient vibration value).

NOTE In this International Standard there is no test procedure or tolerances defined for instrumentation to be used for the measurement of MTVV.

### 3.3.4 fourth-power vibration dose value (VDV)

The integral of the fourth power of the weighted instantaneous acceleration  $a_w(t)$ , in  $\text{m/s}^{1.75}$ , as defined by the expression:

$$\text{VDV} = \sqrt[4]{\int_0^T a_w^4(t) dt} \quad (5)$$

where  $T$  is the duration of the measurement, in seconds.

The vibration dose value is more sensitive to peaks than is the r.m.s. value.

NOTE In this International Standard there is no test procedure or tolerances defined for instrumentation to be used for the measurement of VDV.

### 3.3.5 motion sickness dose value (MSDV)

The integral of the squared weighted instantaneous acceleration  $a_w(t)$ , in  $\text{m/s}^{1.5}$ , as defined by the expression:

$$\text{MSDV} = \sqrt{\int_0^T a_w^2(t) dt} \quad (6)$$

where  $T$  is the duration of the measurement, in seconds.

The motion sickness dose value may be obtained from the weighted r.m.s. acceleration through multiplication by  $\sqrt{T}$ .

NOTE In this International Standard there is no test procedure or tolerances defined for instrumentation to be used for the measurement of MSDV.

### 3.4 crest factor

Replace definition by:

The ratio of the peak signal value evaluated over the duration of measurement to the r.m.s. value over the same time interval. The peak value is the largest deviation from the arithmetic mean of the frequency-weighted acceleration signal.

Delete NOTE.

### 3.8 reference calibration frequency

Replace table 2 by the following table:

**Table 2 — Preferred reference calibration frequencies**

Characteristics of vibration	Reference calibration frequency		Weighting	
	$\omega$ s <sup>-1</sup>	$f$ Hz	factor	dB
Principal whole-body:				
Whole-body, vertical, $W_k$	50	7,958	1,037 1	0,32
Whole-body, horizontal $x$ and $y$ , $W_d$	50	7,958	0,254 5	-11,89
Whole-body, motion sickness, $W_f$	2,5	0,398	0,388 8	-8,20
Whole-body, combined, W.B.combined	50	7,958	0,581 2	-4,71
Additional whole-body:				
Whole-body, seat back, horizontal $x$ , $W_c$	50	7,958	0,893 3	-0,98
Whole-body, rotational, $W_e$	50	7,958	0,126 1	-17,98
Head of recumbent person, $W_j$	50	7,958	1,016	0,14
Hand-arm:				
All directions, $W_h$	500	79,58	0,202	-13,89

Page 3:

## 4 Characteristics

Last line: Replace IEC 225 by IEC 61260.

### 4.3 Weighting characteristics

Replace this subclause by the following.

#### 4.3 Weighting characteristics

##### 4.3.1 Frequency weighting of acceleration time history

For integration of the frequency weighted acceleration time history in order to assess human response to vibration, one or more of the following frequency weightings shall be applied. Frequency weighting characteristics are defined in annex B and shown in the tables indicated.

a) Principal weightings for whole-body vibration related to health, comfort and perception:

$W_k$  for vertical  $z$  direction and for vertical recumbent direction (table 4);

$W_d$  for horizontal  $x$  and  $y$  directions and for horizontal recumbent direction (table 5);

W.B.combined for vibrations in buildings (table 7).

b) Principal weighting for whole-body vibration related to motion sickness:

$W_f$  for vertical  $z$  direction (table 6).

c) Additional weightings for whole-body vibration:

$W_c$  for horizontal  $x$  direction of seat back vibration (table 8);

$W_e$  for all axes of rotational seat vibration (table 9);

$W_j$  for vertical  $x$  vibration under the head of a recumbent person (table 10).

d) Weighting related to hand-arm vibration:

$W_h$  for all directions of hand-arm vibration (table 11).

Figures C.1 to C.8 in annex C illustrate the frequency weighting curves. Here, as well as in tables 4 to 11, the band limiting is included.

The frequency weightings may be realized by either analog filters or by digital methods applied in frequency or time domain. They are defined in a mathematical form familiar to filter designers in annex B. Equations for the frequency band limitation are expressed separately to enable a two-step filtering procedure.

Lower and upper frequency band limitation shall be achieved by two-pole high-pass and low-pass filters, respectively, with Butterworth characteristics having an asymptotic slope of 12 dB per octave. The corner frequencies of the band-limiting filters are one-third of an octave outside the nominal frequency range of the relevant band.

Band limiting filters are high-pass at 0,4 Hz and low-pass at 100 Hz for weightings  $W_c$ ,  $W_d$ ,  $W_e$ ,  $W_j$  and  $W_k$ , whereas the frequency weighting  $W_i$  has a high-pass filter at 0,08 Hz and a low-pass filter at 0,63 Hz. The W.B.combined weighting has a high-pass filter at 0,8 Hz and a low-pass filter at 100 Hz. For hand-arm filters, the high-pass and low-pass limits are 6,3 Hz and 1 250 Hz respectively.

Other optional weighting characteristics may be included.

If such an optional weighting characteristic is designated "flat", its frequency response with respect to the input signal, for example acceleration or velocity, shall be constant but imposed by the appropriate band limiting characteristic. A flat characteristic enables the instrumentation to function as a preamplifier for an auxiliary device or to measure the unweighted signal.

Weighting and amplifier circuits shall satisfy the requirements of 5.1. When the flat response is provided, the manufacturer shall specify its frequency range and tolerances. The tolerances shall not be greater than those for the frequency-weighting characteristics (tables 4 to 11).

#### 4.3.2 Frequency weighting of acceleration spectra

The acceleration may be analysed as either constant bandwidth or proportional bandwidth spectra (e.g. as one-third-octave bands or narrower) of the unweighted acceleration. Any form of frequency analysis, analog or digital, direct one-third-octave band or summation of narrow band data may be used.

In the case of one-third-octave bands, the centre frequencies shall be as stated in tables 4 to 11. In the case of constant bandwidth analysis, the frequency resolution (i.e. the interval between adjacent spectral lines) should be no larger than the bandwidth of the narrowest one-third-octave band within the nominal frequency range of the appropriate frequency weighting.

The frequency weighted r.m.s. acceleration shall be determined by frequency-weighting and appropriate summation of squared narrow band or one-third-octave band data. The overall weighted r.m.s. acceleration in m/s<sup>2</sup> or rad/s<sup>2</sup> shall be determined in accordance with the following equation:

$$a_w = \sqrt{\sum_i (W_i a_i)^2} \quad (7)$$

where

- $a_w$  is the weighted r.m.s. acceleration, in metres per second squared or radians per second squared;
- $W_i$  is the weighting factor for the  $i^{\text{th}}$  frequency band or spectral line;
- $a_i$  is the r.m.s. acceleration in the  $i^{\text{th}}$  frequency band or spectral line, over the duration of the measurement.

For narrow band r.m.s. data, weighting factors  $W_i$  may be obtained for each spectral line using the equations in annex B.

In the case of one-third-octave-band analysis, the filter characteristics shall be consistent with the one-third-octave-band filter specifications in IEC 61260.

If the acceleration spectrum is estimated from line spectra (i.e. FFTs), an appropriate method should be used to smooth the spectral estimate (e.g. subdivision of the acceleration record into overlapping segments and use of a tapering window to suppress side-lobe leakage).

In the case of narrow-band analysis of random vibration, the  $i^{\text{th}}$  squared acceleration  $a_i^2$  in  $\text{m}^2/\text{s}^4$  shall be obtained from power spectral density (PSD) data  $\text{PSD}_i$  through multiplication by the frequency resolution  $\Delta f$ :

$$a_i^2 = \text{PSD}_i \cdot \Delta f \quad (8)$$

For narrow band (FFT) analysis consideration shall be given to ensure that the frequency resolution, windowing function, averaging, overlapping and other parameters are adequate.

For the application of equation (7) to one-third-octave data for whole-body vibration, weighting factors found in tables 3 and 4 of ISO 2631-1:1997 shall be applied. These are based on nominal mid-frequencies.

#### 4.3.3 Integration in the time domain

Vibration-measuring instrumentation for human response shall, in the time domain, at least provide a linear integrated root-mean-square value of the frequency-weighted acceleration signal over a selectable time period (including 60 s and longer periods). The integration time shall be indicated by the instrumentation.

An instrument providing exponential averaging shall include a time constant of 1 s. If additional time constants are provided, these should preferably be 1/8 s or 8 s. The time constant used shall be indicated by the instrument.

When provided, the peak characteristic allows the vibration measuring instrumentation to indicate the maximum peak of the vibratory signal whether it is positive or negative.

The averaging method, linear or exponential, shall be indicated.

For the purpose of measuring the maximum transient vibration value (MTVV) (running r.m.s. acceleration), the instrumentation shall in addition provide a continuous or continuously sampled measurement of the short-time linear or exponentially integrated r.m.s. value. The selection of integration time (time constant) shall include 1 s. The SLOW characteristics defined for sound level meters in IEC 60651 may be used. The increment of a sampled observation time  $t_0$  shall be smaller than the integration time. The maximum transient vibration value (MTVV) of the acceleration shall be indicated. The integration time used shall be indicated by the instrumentation.

NOTE 1 Results obtained using different time constants,  $\tau$ , should not be compared.

For the purpose of measuring the fourth-power vibration dose (VDV), the instrumentation shall in addition to the integrated r.m.s. acceleration provide a fourth-power dose value according to equation (5). The measurement may be interrupted and continued. The total integration time used shall be indicated by the instrumentation.

For the purpose of measuring the motion sickness dose value (MSDV), the instrumentation must be capable of providing an accurately integrated value of the squared low frequency acceleration using a long integration time (perhaps several hours). The total integration time used shall be indicated by the instrumentation.

The crest factor should be used to indicate if the basic integrated r.m.s. acceleration is suitable for describing the severity of the vibration or if the additional integration methods above should be applied as well. For this

purpose the instrumentation may be equipped with a facility to read out the peak value of the frequency-weighted instantaneous acceleration.

NOTE 2 The integration times specified should not be taken to be necessarily representative of an integration time of the human body.

Page 4

## 5.1 General

Replace the complete subclause by the following.

The complete instrumentation chain (comprising the transducer, amplifier, weighting network and detector-indicator) for measurements according to ISO 5349 (hand-arm) or ISO 2631-2 (whole-body combined) shall have either or both of the characteristics and tolerances given in tables 7 and 11.

The characteristics and tolerances of instrumentation used for measurement according to ISO 2631-1:1997 and given in tables 4, 5, 6, 8, 9 and 10 apply only to the combined frequency weighting and band-limiting.

Graphs and analytical expressions for each of the tables are given in annex B and annex C respectively.

Provisions for external filter connection may be included.

STANDARDSISO.COM : Click to view the full PDF of ISO 8041:1990/Amd.1:1999

Pages 5 to 12:

Replace tables 4 to 8 by the following tables 4 to 11 and accordingly renumber tables 9 to 12 as tables 12 to 15.

**Table 4 — Frequency weighting  $W_k$  for vertical whole body vibration (z-axis)  
Seated, standing or recumbent person — Based on ISO 2631-1**

x	Frequency, Hz		Band-limit factor	Band limit dB	Weighting factor	Weighting dB	$W_k$ factor	$W_k$ dB	Tolerance %	Tolerance dB
	Nominal	True								
-10	0,1	0,100 0	0,062 4	-24,10	0,500	-6,01	0,031 2	-30,11	+26	+2
-9	0,125	0,125 9	0,098 6	-20,12	0,500	-6,02	0,049 3	-26,14	+26	+2
-8	0,16	0,158 5	0,155	-16,19	0,500	-6,02	0,077 6	-22,21	+26	+2
-7	0,2	0,199 5	0,241	-12,34	0,500	-6,02	0,121	-18,37	+26	+2
-6	0,25	0,251 2	0,367	-8,71	0,499	-6,03	0,183	-14,74	+26 -21	±2
-5	0,315	0,316 2	0,530	-5,51	0,499	-6,04	0,264	-11,55	+26 -21	±2
-4	0,4	0,398 1	0,704	-3,05	0,498	-6,06	0,350	-9,11	+12 -11	±1
-3	0,5	0,501 2	0,843	-1,48	0,497	-6,08	0,419	-7,56	+12 -11	±1
-2	0,63	0,631 0	0,928	-0,65	0,495	-6,12	0,459	-6,77	+12 -11	±1
-1	0,8	0,794 3	0,969	-0,27	0,492	-6,17	0,477	-6,44	+12 -11	±1
0	1	1,000	0,987	-0,11	0,489	-6,22	0,482	-6,33	+12 -11	±1
1	1,25	1,259	0,995	-0,04	0,487	-6,25	0,485	-6,29	+12 -11	±1
2	1,6	1,585	0,998	-0,02	0,494	-6,12	0,493	-6,13	+12 -11	±1
3	2	1,995	0,999	-0,01	0,531	-5,50	0,531	-5,50	+12 -11	±1
4	2,5	2,512	1,000	0,00	0,634	-3,96	0,633	-3,97	+12 -11	±1
5	3,15	3,162	1,000	0,00	0,807	-1,86	0,807	-1,86	+12 -11	±1
6	4	3,981	1,000	0,00	0,965	-0,31	0,965	-0,31	+12 -11	±1
7	5	5,012	1,000	0,00	1,039	0,33	1,039	0,33	+12 -11	±1
8	6,3	6,310	1,000	0,00	1,054	0,46	1,054	0,46	+12 -11	±1
9	8	7,943	1,000	0,00	1,037	0,32	1,037	0,32	0	0
10	10	10,00	1,000	0,00	0,988	-0,10	0,988	-0,10	+12 -11	±1
11	12,5	12,59	1,000	0,00	0,899	-0,92	0,899	-0,93	+12 -11	±1
12	16	15,85	1,000	0,00	0,775	-2,22	0,774	-2,22	+12 -11	±1
13	20	19,95	0,999	-0,01	0,638	-3,91	0,637	-3,91	+12 -11	±1
14	25	25,12	0,998	-0,02	0,511	-5,83	0,510	-5,84	+12 -11	±1
15	31,5	31,62	0,995	-0,04	0,405	-7,85	0,403	-7,89	+12 -11	±1
16	40	39,81	0,988	-0,11	0,320	-9,90	0,316	-10,01	+12 -11	±1
17	50	50,12	0,970	-0,27	0,253	-11,95	0,245	-12,21	+12 -11	±1
18	63	63,10	0,929	-0,64	0,200	-13,98	0,186	-14,62	+12 -11	±1
19	80	79,43	0,846	-1,46	0,158	-16,01	0,134	-17,47	+12 -11	±1
20	100	100,0	0,707	-3,01	0,125	-18,03	0,088 7	-21,04	+12 -11	±1
21	125	125,9	0,534	-5,46	0,100	-20,04	0,053 1	-25,50	+26 -21	±2
22	160	158,5	0,370	-8,64	0,079 0	-22,05	0,029 2	-30,69	+26 -21	±2
23	200	199,5	0,244	-12,27	0,062 7	-24,05	0,015 3	-36,32	+26	+2
24	250	251,2	0,157	-16,11	0,049 8	-26,06	0,007 79	-42,16	+26	+2
25	315	316,2	0,100	-20,04	0,039 5	-28,06	0,003 93	-48,10	+26	+2
26	400	398,1	0,063 0	-24,02	0,031 4	-30,06	0,001 98	-54,08	+26	+2

NOTE x is the frequency band number according to IEC 61260:1995.  
The table is based on true frequencies (exact values as defined in IEC 61260:1995).  
Tolerance applies to the combination  $W_k$  of band-limiting and frequency weighting.

**Table 5 — Frequency weighting  $W_d$  for horizontal whole body vibration ( $x$ - or  $y$ -axis)  
Seated, standing or recumbent person — Based on ISO 2631-1**

$x$	Frequency, Hz		Band-limit factor	Band limit dB	Weighting factor	Weighting dB	$W_d$ factor	$W_d$ dB	Tolerance %	Tolerance dB
	Nominal	True								
-10	0,1	0,100 0	0,062 4	-24,10	1,001	0,01	0,062 4	-24,09	+26	+2
-9	0,125	0,125 9	0,098 6	-20,12	1,001	0,01	0,098 7	-20,12	+26	+2
-8	0,16	0,158 5	0,155	-16,19	1,001	0,01	0,155	-16,18	+26	+2
-7	0,2	0,199 5	0,241	-12,34	1,002	0,02	0,242	-12,32	+26	+2
-6	0,25	0,251 2	0,367	-8,71	1,004	0,03	0,368	-8,68	+26 -21	±2
-5	0,315	0,316 2	0,530	-5,51	1,006	0,05	0,533	-5,47	+26 -21	±2
-4	0,4	0,398 1	0,704	-3,05	1,009	0,07	0,710	-2,98	+12 -11	±1
-3	0,5	0,501 2	0,843	-1,48	1,013	0,11	0,854	-1,37	+12 -11	±1
-2	0,63	0,631 0	0,928	-0,65	1,018	0,15	0,944	-0,50	+12 -11	±1
-1	0,8	0,794 3	0,969	-0,27	1,023	0,20	0,991	-0,08	+12 -11	±1
0	1	1,000	0,987	-0,11	1,024	0,20	1,011	0,10	+12 -11	±1
1	1,25	1,259	0,995	-0,04	1,012	0,11	1,007	0,06	+12 -11	±1
2	1,6	1,585	0,998	-0,02	0,973	-0,24	0,971	-0,26	+12 -11	±1
3	2	1,995	0,999	-0,01	0,892	-0,99	0,891	-1,00	+12 -11	±1
4	2,5	2,512	1,000	0,00	0,774	-2,23	0,773	-2,23	+12 -11	±1
5	3,15	3,162	1,000	0,00	0,640	-3,88	0,640	-3,88	+12 -11	±1
6	4	3,981	1,000	0,00	0,514	-5,78	0,514	-5,78	+12 -11	±1
7	5	5,012	1,000	0,00	0,408	-7,78	0,408	-7,78	+12 -11	±1
8	6,3	6,310	1,000	0,00	0,323	-9,83	0,323	-9,83	+12 -11	±1
9	8	7,943	1,000	0,00	0,255	-11,87	0,255	-11,87	0	0
10	10	10,00	1,000	0,00	0,202	-13,91	0,202	-13,91	+12 -11	±1
11	12,5	12,59	1,000	0,00	0,160	-15,93	0,160	-15,93	+12 -11	±1
12	16	15,85	1,000	0,00	0,127	-17,95	0,127	-17,95	+12 -11	±1
13	20	19,95	0,999	-0,01	0,100	-19,96	0,100	-19,97	+12 -11	±1
14	25	25,12	0,998	-0,02	0,079 7	-21,97	0,079 6	-21,98	+12 -11	±1
15	31,5	31,62	0,995	-0,04	0,063 3	-23,97	0,063 0	-24,01	+12 -11	±1
16	40	39,81	0,988	-0,11	0,050 3	-25,97	0,049 6	-26,08	+12 -11	±1
17	50	50,12	0,970	-0,27	0,039 9	-27,98	0,038 7	-28,24	+12 -11	±1
18	63	63,10	0,929	-0,64	0,031 7	-29,98	0,029 5	-30,62	+12 -11	±1
19	80	79,43	0,846	-1,46	0,025 2	-31,98	0,021 3	-33,43	+12 -11	±1
20	100	100,0	0,707	-3,01	0,020 0	-33,98	0,014 1	-36,99	+12 -11	±1
21	125	125,9	0,534	-5,46	0,015 9	-35,98	0,008 48	-41,43	+26 -21	±2
22	160	158,5	0,370	-8,64	0,012 6	-37,98	0,004 67	-46,62	+26 -21	±2
23	200	199,5	0,244	-12,27	0,010 0	-39,98	0,002 44	-52,24	+26	+2
24	250	251,2	0,157	-16,11	0,007 96	-41,98	0,001 25	-58,09	+26	+2
25	315	316,2	0,100	-20,04	0,006 32	-43,98	0,000 629	-64,02	+26	+2
26	400	398,1	0,063 0	-24,02	0,005 02	-45,98	0,000 316	-70,00	+26	+2

NOTE  $x$  is the frequency band number according to IEC 61260:1995.  
The table is based on true frequencies (exact values as defined in IEC 61260:1995).  
Tolerance applies to the combination  $W_d$  of band-limiting and frequency weighting.

**Table 6 — Frequency weighting  $W_f$  for vertical whole body vibration ( $z$ -axis)  
Motion sickness, seated or standing person — Based on ISO 2631-1**

$x$	Frequency, Hz		Band-limit factor	Band limit dB	Weighting factor	Weighting dB	$W_f$ factor	$W_f$ dB	Tolerance %	Tolerance dB
	Nominal	True								
-17	0,02	0,019 95	0,062 1	-24,14	0,388	-8,23	0,024 1	-32,37	+26	+2
-16	0,025	0,025 12	0,098 1	-20,17	0,388	-8,23	0,038 0	-28,40	+26	+2
-15	0,031 5	0,031 62	0,154	-16,23	0,390	-8,18	0,060 2	-24,41	+26	+2
-14	0,04	0,039 81	0,240	-12,38	0,400	-7,96	0,096 2	-20,34	+26	+2
-13	0,05	0,050 12	0,365	-8,75	0,431	-7,31	0,157	-16,06	+26 -21	±2
-12	0,063	0,063 10	0,528	-5,54	0,506	-5,91	0,267	-11,45	+26 -21	±2
-11	0,08	0,079 43	0,702	-3,07	0,646	-3,79	0,454	-6,86	+12 -11	±1
-10	0,1	0,100 0	0,842	-1,49	0,826	-1,67	0,695	-3,16	+12 -11	±1
-9	0,125	0,125 9	0,927	-0,66	0,971	-0,25	0,900	-0,92	+12 -11	±1
-8	0,16	0,158 5	0,967	-0,29	1,038	0,33	1,004	0,04	+12 -11	±1
-7	0,2	0,199 5	0,982	-0,15	1,011	0,09	0,993	-0,06	+12 -11	±1
-6	0,25	0,251 2	0,983	-0,15	0,865	-1,26	0,850	-1,41	+12 -11	±1
-5	0,315	0,316 2	0,968	-0,29	0,635	-3,94	0,615	-4,22	+12 -11	±1
-4	0,4	0,398 1	0,928	-0,65	0,419	-7,57	0,388	-8,22	0	0
-3	0,5	0,501 2	0,845	-1,47	0,263	-11,59	0,223	-13,05	+12 -11	±1
-2	0,63	0,631 0	0,706	-3,02	0,164	-15,71	0,116	-18,73	+12 -11	±1
-1	0,8	0,794 3	0,532	-5,47	0,102	-19,82	0,054 3	-25,30	+26 -21	±2
0	1	1,000	0,369	-8,66	0,063 8	-23,91	0,023 5	-32,57	+26 -21	±2
1	1,25	1,259	0,243	-12,29	0,039 9	-27,97	0,009 70	-40,26	+26	+2
2	1,6	1,585	0,156	-16,13	0,025 1	-32,01	0,003 92	-48,14	+26	+2
3	2	1,995	0,099 2	-20,07	0,015 8	-36,04	0,001 57	-56,11	+26	+2
4	2,5	2,512	0,062 8	-24,04	0,009 9	-40,05	0,000 624	-64,10	+26	+2
5	3,15	3,162	0,039 7	-28,03	0,006 26	-44,06	0,000 248	-72,10	+26	+2
6	4	3,981	0,025 0	-32,03	0,003 95	-48,07	0,000 099	-80,10	+26	+2

NOTE  $x$  is the frequency band number according to IEC 61260:1995.  
The table is based on true frequencies (exact values as defined in IEC 61260:1995).  
Tolerance applies to the combination  $W_f$  of band-limiting and frequency weighting.

**Table 7 — Frequency weighting W.B.combined for whole-body vibrations in buildings  
All directions — Based on ISO 2631-2**

x	Frequency, Hz		Band-limit factor	Band limit dB	Weighting factor	Weighting dB	W.B. combined factor	W.B. combined dB	Tolerance %	Tolerance dB
	Nominal	True								
-10	0,1	0,100 0	0,015 8	-36,00	1,000	0,00	0,015 8	-36,00	+26	+2
-9	0,125	0,125 9	0,025 1	-32,00	1,000	0,00	0,025 1	-32,00	+26	+2
-8	0,16	0,158 5	0,040	-28,01	1,000	0,00	0,040	-28,01	+26	+2
-7	0,2	0,199 5	0,063	-24,02	0,999	-0,01	0,063	-24,02	+26	+2
-6	0,25	0,251 2	0,100	-20,04	0,999	-0,01	0,099	-20,05	+26	+2
-5	0,315	0,316 2	0,157	-16,11	0,998	-0,01	0,156	-16,12	+26	+2
-4	0,4	0,398 1	0,244	-12,27	0,998	-0,02	0,243	-12,29	+26	+2
-3	0,5	0,501 2	0,370	-8,64	0,996	-0,03	0,368	-8,67	+26	+2
-2	0,63	0,631 0	0,534	-5,45	0,994	-0,05	0,530	-5,51	+26	+2
-1	0,8	0,794 3	0,707	-3,01	0,990	-0,08	0,700	-3,09	+26	+2
0	1	1,000	0,846	-1,46	0,985	-0,13	0,833	-1,59	+26	+2
1	1,25	1,259	0,929	-0,64	0,976	-0,21	0,907	-0,85	+12	+1
2	1,6	1,585	0,970	-0,27	0,963	-0,33	0,934	-0,59	+12	+1
3	2	1,995	0,988	-0,11	0,944	-0,50	0,932	-0,61	+12	+1
4	2,5	2,512	0,995	-0,04	0,915	-0,77	0,910	-0,82	+12	+1
5	3,15	3,162	0,998	-0,02	0,874	-1,17	0,872	-1,19	+12	+1
6	4	3,981	0,999	-0,01	0,819	-1,73	0,818	-1,74	+12	+1
7	5	5,012	1,000	0,00	0,750	-2,50	0,750	-2,50	+12	+1
8	6,3	6,310	1,000	0,00	0,669	-3,49	0,669	-3,49	+12	+1
9	8	7,943	1,000	0,00	0,582	-4,70	0,582	-4,70	0	0
10	10	10,00	1,000	0,00	0,494	-6,12	0,494	-6,12	+12	+1
11	12,5	12,59	1,000	0,00	0,412	-7,71	0,411	-7,71	+12	+1
12	16	15,85	1,000	0,00	0,338	-9,43	0,337	-9,44	+12	+1
13	20	19,95	0,999	-0,01	0,274	-11,25	0,274	-11,25	+12	+1
14	25	25,12	0,998	-0,02	0,221	-13,12	0,220	-13,14	+12	+1
15	31,5	31,62	0,995	-0,04	0,177	-15,04	0,176	-15,09	+12	+1
16	40	39,81	0,988	-0,11	0,141	-16,99	0,140	-17,10	+12	+1
17	50	50,12	0,970	-0,27	0,113	-18,96	0,109	-19,23	+12	+1
18	63	63,10	0,929	-0,64	0,089 7	-20,94	0,083 4	-21,58	+12	+1
19	80	79,43	0,846	-1,46	0,071 4	-22,93	0,060 4	-24,38	+26	+2
20	100	100,0	0,707	-3,01	0,056 7	-24,92	0,040 1	-27,93	+26	+2
21	125	125,9	0,534	-5,46	0,045 1	-26,92	0,024 1	-32,37	+26	+2
22	160	158,5	0,370	-8,64	0,035 8	-28,91	0,013 3	-37,55	+26	+2
23	200	199,5	0,244	-12,27	0,028 5	-30,91	0,006 94	-43,18	+26	+2
24	250	251,2	0,157	-16,11	0,022 6	-32,91	0,003 54	-49,02	+26	+2
25	315	316,2	0,099 5	-20,04	0,018 0	-34,91	0,001 79	-54,95	+26	+2
26	400	398,1	0,063 0	-24,02	0,014 3	-36,91	0,000 899	-60,92	+26	+2
27	500	501,2	0,039 8	-28,01	0,011 3	-38,91	0,000 451	-66,91	+26	+2
28	630	631,0	0,025 1	-32,00	0,009 01	-40,91	0,000 226	-72,91	+26	+2
29	800	794,3	0,015 8	-36,00	0,007 16	-42,91	0,000 113	-78,91	+26	+2

NOTE x is the frequency band number according to IEC 61260:1995.  
The table is based on true frequencies (exact values as defined in IEC 61260:1995).  
Tolerance applies to the total frequency weighting W.B.combined for the whole instrument chain (see 5.1).

**Table 8 — Frequency weighting  $W_c$  for horizontal whole body vibration ( $x$  axis)  
Seat back, seated person — Based on ISO 2631-1**

$x$	Frequency, Hz		Band-limit factor	Band-limit dB	Weighting factor	Weighting dB	$W_c$ factor	$W_c$ dB	Tolerance %	Tolerance dB
	Nominal	True								
-10	0,1	0,100 0	0,062 4	-24,10	1,000	0,00	0,062 4	-24,10	+26	+2
-9	0,125	0,125 9	0,098 6	-20,12	1,000	0,00	0,098 6	-20,12	+26	+2
-8	0,16	0,158 5	0,155	-16,19	1,000	0,00	0,155	-16,19	+26	+2
-7	0,2	0,199 5	0,241	-12,34	1,000	0,00	0,241	-12,34	+26	+2
-6	0,25	0,251 2	0,367	-8,71	1,000	0,00	0,367	-8,71	+26 -21	±2
-5	0,315	0,316 2	0,530	-5,51	1,000	0,00	0,530	-5,51	+26 -21	±2
-4	0,4	0,398 1	0,704	-3,05	1,001	0,01	0,704	-3,05	+12 -11	±1
-3	0,5	0,501 2	0,843	-1,48	1,001	0,01	0,844	-1,47	+12 -11	±1
-2	0,63	0,631 0	0,928	-0,65	1,001	0,01	0,929	-0,64	+12 -11	±1
-1	0,8	0,794 3	0,969	-0,27	1,002	0,02	0,972	-0,25	+12 -11	±1
0	1	1,000	0,987	-0,11	1,004	0,03	0,991	-0,08	+12 -11	±1
1	1,25	1,259	0,995	-0,04	1,006	0,05	1,000	0,00	+12 -11	±1
2	1,6	1,585	0,998	-0,02	1,008	0,07	1,006	0,06	+12 -11	±1
3	2	1,995	0,999	-0,01	1,012	0,11	1,012	0,10	+12 -11	±1
4	2,5	2,512	1,000	0,00	1,018	0,15	1,017	0,15	+12 -11	±1
5	3,15	3,162	1,000	0,00	1,023	0,19	1,023	0,19	+12 -11	±1
6	4	3,981	1,000	0,00	1,024	0,21	1,024	0,21	+12 -11	±1
7	5	5,012	1,000	0,00	1,013	0,11	1,013	0,11	+12 -11	±1
8	6,3	6,310	1,000	0,00	0,974	-0,23	0,974	-0,23	+12 -11	±1
9	8	7,943	1,000	0,00	0,894	-0,97	0,894	-0,97	0	0
10	10	10,00	1,000	0,00	0,776	-2,20	0,776	-2,20	+12 -11	±1
11	12,5	12,59	1,000	0,00	0,643	-3,84	0,643	-3,84	+12 -11	±1
12	16	15,85	1,000	0,00	0,517	-5,73	0,517	-5,74	+12 -11	±1
13	20	19,95	0,999	-0,01	0,410	-7,74	0,410	-7,75	+12 -11	±1
14	25	25,12	0,998	-0,02	0,324	-9,78	0,324	-9,80	+12 -11	±1
15	31,5	31,62	0,995	-0,04	0,256	-11,83	0,255	-11,87	+12 -11	±1
16	40	39,81	0,988	-0,11	0,203	-13,86	0,200	-13,97	+12 -11	±1
17	50	50,12	0,970	-0,27	0,161	-15,89	0,156	-16,15	+12 -11	±1
18	63	63,10	0,929	-0,64	0,127	-17,91	0,118	-18,55	+12 -11	±1
19	80	79,43	0,846	-1,46	0,101	-19,92	0,085 4	-21,37	+12 -11	±1
20	100	100,0	0,707	-3,01	0,080 1	-21,93	0,056 7	-24,94	+12 -11	±1
21	125	125,9	0,534	-5,46	0,063 6	-23,93	0,033 9	-29,39	+26 -21	±2
22	160	158,5	0,370	-8,64	0,050 5	-25,93	0,018 7	-34,57	+26 -21	±2
23	200	199,5	0,244	-12,27	0,040 1	-27,93	0,009 8	-40,20	+26	+2
24	250	251,2	0,157	-16,11	0,031 9	-29,94	0,004 99	-46,04	+26	+2
25	315	316,2	0,100	-20,04	0,025 3	-31,94	0,002 52	-51,98	+26	+2
26	400	398,1	0,063 0	-24,02	0,020 1	-33,94	0,001 27	-57,95	+26	+2

NOTE  $x$  is the frequency band number according to IEC 61260:1995.  
The table is based on true frequencies (exact values as defined in IEC 61260:1995).  
Tolerance applies to the combination  $W_c$  of band-limiting and frequency weighting.

**Table 9 — Frequency weighting  $W_e$  for rotational whole body vibration (all directions)  
Seated person — Based on ISO 2631-1**

x	Frequency, Hz		Band-limit factor	Band limit dB	Weighting factor	Weighting dB	$W_e$ factor	$W_e$ dB	Tolerance %	Tolerance dB
	Nominal	True								
-10	0,1	0,100 0	0,062 4	-24,10	1,002	0,02	0,062 5	-24,08	+26	+2
-9	0,125	0,125 9	0,098 6	-20,12	1,004	0,03	0,098 9	-20,09	+26	+2
-8	0,16	0,158 5	0,155	-16,19	1,006	0,05	0,156	-16,14	+26	+2
-7	0,2	0,199 5	0,241	-12,34	1,009	0,07	0,244	-12,27	+26	+2
-6	0,25	0,251 2	0,367	-8,71	1,013	0,11	0,371	-8,60	+26 -21	±2
-5	0,315	0,316 2	0,530	-5,51	1,018	0,15	0,539	-5,36	+26 -21	±1
-4	0,4	0,398 1	0,704	-3,05	1,023	0,20	0,720	-2,86	+12 -11	±1
-3	0,5	0,501 2	0,843	-1,48	1,024	0,20	0,864	-1,27	+12 -11	±1
-2	0,63	0,631 0	0,928	-0,65	1,012	0,10	0,939	-0,55	+12 -11	±1
-1	0,8	0,794 3	0,969	-0,27	0,972	-0,25	0,942	-0,52	+12 -11	±1
0	1	1,000	0,987	-0,11	0,891	-1,00	0,880	-1,11	+12 -11	±1
1	1,25	1,259	0,995	-0,04	0,772	-2,25	0,768	-2,29	+12 -11	±1
2	1,6	1,585	0,998	-0,02	0,638	-3,90	0,637	-3,91	+12 -11	±1
3	2	1,995	0,999	-0,01	0,513	-5,80	0,513	-5,80	+12 -11	±1
4	2,5	2,512	1,000	0,00	0,407	-7,80	0,407	-7,81	+12 -11	±1
5	3,15	3,162	1,000	0,00	0,322	-9,85	0,322	-9,85	+12 -11	±1
6	4	3,981	1,000	0,00	0,254	-11,89	0,254	-11,89	+12 -11	±1
7	5	5,012	1,000	0,00	0,201	-13,93	0,201	-13,93	+12 -11	±1
8	6,3	6,310	1,000	0,00	0,159	-15,95	0,159	-15,95	+12 -11	±1
9	8	7,943	1,000	0,00	0,126	-17,97	0,126	-17,97	0	0
10	10	10,00	1,000	0,00	0,100	-19,98	0,100	-19,98	+12 -11	±1
11	12,5	12,59	1,000	0,00	0,079 6	-21,99	0,079 5	-21,99	+12 -11	±1
12	16	15,85	1,000	0,00	0,063 2	-23,99	0,063 1	-23,99	+12 -11	±1
13	20	19,95	0,999	-0,01	0,050 1	-25,99	0,050 1	-26,00	+12 -11	±1
14	25	25,12	0,998	-0,02	0,039 8	-28,00	0,039 7	-28,01	+12 -11	±1
15	31,5	31,62	0,995	-0,04	0,031 6	-30,00	0,031 5	-30,04	+12 -11	±1
16	40	39,81	0,988	-0,11	0,025 1	-32,00	0,024 8	-32,11	+12 -11	±1
17	50	50,12	0,970	-0,27	0,020 0	-34,00	0,019 4	-34,26	+12 -11	±1
18	63	63,10	0,929	-0,64	0,015 8	-36,00	0,014 7	-36,64	+12 -11	±1
19	80	79,43	0,846	-1,46	0,012 6	-38,00	0,010 6	-39,46	+12 -11	±1
20	100	100,0	0,707	-3,01	0,010 0	-40,00	0,007 07	-43,01	+12 -11	±1
21	125	125,9	0,534	-5,46	0,007 94	-42,00	0,004 24	-47,46	+26 -21	±2
22	160	158,5	0,370	-8,64	0,006 31	-44,00	0,002 33	-52,64	+26 -21	±2
23	200	199,5	0,244	-12,27	0,005 01	-46,00	0,001 22	-58,27	+26	+2
24	250	251,2	0,157	-16,11	0,003 98	-48,00	0,000 62	-64,11	+26	+2
25	315	316,2	0,100	-20,04	0,003 16	-50,00	0,000 315	-70,04	+26	+2
26	400	398,1	0,063 0	-24,02	0,002 51	-52,00	0,000 158	-76,02	+26	+2

NOTE x is the frequency band number according to IEC 61260:1995.  
The table is based on true frequencies (exact values as defined in IEC 61260:1995).  
Tolerance applies to the combination  $W_e$  of band-limiting and frequency weighting.

**Table 10 — Frequency weighting  $W_i$  for vertical head vibration ( $x$ -axis)  
Recumbent person — Based on ISO 2631-1**

$x$	Frequency, Hz		Band-limit factor	Band limit dB	Weighting factor	Weighting dB	$W_i$ factor	$W_i$ dB	Tolerance %	Tolerance dB
	Nominal	True								
-10	0,1	0,100 0	0,062 4	-24,10	0,497	-6,08	0,031 0	-30,18	+26	+2
-9	0,125	0,125 9	0,098 6	-20,12	0,497	-6,08	0,049 0	-26,20	+26	+2
-8	0,16	0,158 5	0,155	-16,19	0,497	-6,08	0,077 0	-22,27	+26	+2
-7	0,2	0,199 5	0,241	-12,34	0,497	-6,08	0,120	-18,42	+26	+2
-6	0,25	0,251 2	0,367	-8,71	0,496	-6,08	0,182	-14,79	+26 -21	±2
-5	0,315	0,316 2	0,530	-5,51	0,496	-6,09	0,263	-11,60	+26 -21	±2
-4	0,4	0,398 1	0,704	-3,05	0,496	-6,09	0,349	-9,15	+12 -11	±1
-3	0,5	0,501 2	0,843	-1,48	0,495	-6,11	0,418	-7,58	+12 -11	±1
-2	0,63	0,631 0	0,928	-0,65	0,494	-6,12	0,459	-6,77	+12 -11	±1
-1	0,8	0,794 3	0,969	-0,27	0,493	-6,15	0,478	-6,42	+12 -11	±1
0	1	1,000	0,987	-0,11	0,491	-6,19	0,484	-6,30	+12 -11	±1
1	1,25	1,259	0,995	-0,04	0,488	-6,24	0,485	-6,28	+12 -11	±1
2	1,6	1,585	0,998	-0,02	0,484	-6,30	0,483	-6,32	+12 -11	±1
3	2	1,995	0,999	-0,01	0,482	-6,33	0,482	-6,34	+12 -11	±1
4	2,5	2,512	1,000	0,00	0,489	-6,21	0,489	-6,22	+12 -11	±1
5	3,15	3,162	1,000	0,00	0,525	-5,60	0,525	-5,60	+12 -11	±1
6	4	3,981	1,000	0,00	0,625	-4,08	0,625	-4,08	+12 -11	±1
7	5	5,012	1,000	0,00	0,795	-1,99	0,795	-1,99	+12 -11	±1
8	6,3	6,310	1,000	0,00	0,947	-0,47	0,947	-0,47	+12 -11	±1
9	8	7,943	1,000	0,00	1,016	0,14	1,016	0,14	0	0
10	10	10,00	1,000	0,00	1,030	0,26	1,030	0,26	+12 -11	±1
11	12,5	12,59	1,000	0,00	1,026	0,22	1,026	0,22	+12 -11	±1
12	16	15,85	1,000	0,00	1,019	0,16	1,019	0,16	+12 -11	±1
13	20	19,95	0,999	-0,01	1,013	0,11	1,012	0,10	+12 -11	±1
14	25	25,12	0,998	-0,02	1,008	0,07	1,006	0,06	+12 -11	±1
15	31,5	31,62	0,995	-0,04	1,005	0,05	1,000	0,00	+12 -11	±1
16	40	39,81	0,988	-0,11	1,003	0,03	0,991	-0,08	+12 -11	±1
17	50	50,12	0,970	-0,27	1,002	0,02	0,972	-0,25	+12 -11	±1
18	63	63,10	0,929	-0,64	1,001	0,01	0,930	-0,63	+12 -11	±1
19	80	79,43	0,846	-1,46	1,001	0,01	0,846	-1,45	+12 -11	±1
20	100	100,0	0,707	-3,01	1,001	0,00	0,708	-3,01	+12 -11	±1
21	125	125,9	0,534	-5,46	1,000	0,00	0,534	-5,45	+26 -21	±2
22	160	158,5	0,370	-8,64	1,000	0,00	0,370	-8,64	+26 -21	±2
23	200	199,5	0,244	-12,27	1,000	0,00	0,244	-12,26	+26	+2
24	250	251,2	0,157	-16,11	1,000	0,00	0,157	-16,11	+26	+2
25	315	316,2	0,100	-20,04	1,000	0,00	0,100	-20,04	+26	+2
26	400	398,1	0,063 0	-24,02	1,000	0,00	0,063 0	-24,02	+26	+2

NOTE  $x$  is the frequency band number according to IEC 61260:1995.

The table is based on true frequencies (exact values as defined in IEC 61260:1995).

Tolerance applies to the combination  $W_i$  of band-limiting and frequency weighting.

**Table 11 — Frequency weighting  $W_h$  for hand-arm vibration (all directions) — Based on ISO 5349**

x	Frequency, Hz		Band-limit factor	Band-limit dB	Weighting factor	Weighting dB	$W_h$ factor	$W_h$ dB	Tolerance %	Tolerance dB
	Nominal	True								
-1	0,8	0,794 3	0,015 8	-36,00	1,001	0,01	0,015 9	-36,00	+26	+2
0	1	1,000	0,025 1	-32,00	1,001	0,01	0,025 1	-31,99	+26	+2
1	1,25	1,259	0,039 8	-28,01	1,002	0,01	0,039 8	-27,99	+26	+2
2	1,6	1,585	0,063 0	-24,02	1,003	0,02	0,063 1	-23,99	+26	+2
3	2	1,995	0,099	-20,04	1,004	0,04	0,100	-20,01	+26	+2
4	2,5	2,512	0,157	-16,11	1,007	0,06	0,158	-16,05	+26	+2
5	3,15	3,162	0,244	-12,27	1,010	0,09	0,246	-12,18	+26	+2
6	4	3,981	0,370	-8,64	1,015	0,13	0,375	-8,51	+26 -21	±2
7	5	5,012	0,534	-5,46	1,021	0,18	0,545	-5,27	+26 -21	±2
8	6,3	6,310	0,707	-3,01	1,028	0,24	0,727	-2,77	+26 -21	±2
9	8	7,943	0,846	-1,46	1,032	0,28	0,873	-1,18	+26 -21	±2
10	10	10,00	0,929	-0,64	1,024	0,21	0,951	-0,43	+12 -11	±1
11	12,5	12,59	0,970	-0,27	0,987	-0,11	0,958	-0,38	+12 -11	±1
12	16	15,85	0,988	-0,11	0,907	-0,85	0,896	-0,96	+12 -11	±1
13	20	19,95	0,995	-0,04	0,786	-2,09	0,782	-2,14	+12 -11	±1
14	25	25,12	0,998	-0,02	0,648	-3,76	0,647	-3,78	+12 -11	±1
15	31,5	31,62	0,999	-0,01	0,520	-5,69	0,519	-5,69	+12 -11	±1
16	40	39,81	1,000	0,00	0,411	-7,72	0,411	-7,72	+12 -11	±1
17	50	50,12	1,000	0,00	0,324	-9,78	0,324	-9,78	+12 -11	±1
18	63	63,10	1,000	0,00	0,256	-11,83	0,256	-11,83	+12 -11	±1
19	80	79,43	1,000	0,00	0,202	-13,88	0,202	-13,88	0	0
20	100	100,0	1,000	0,00	0,160	-15,91	0,160	-15,91	+12 -11	±1
21	125	125,9	1,000	0,00	0,127	-17,93	0,127	-17,93	+12 -11	±1
22	160	158,5	1,000	0,00	0,101	-19,94	0,101	-19,94	+12 -11	±1
23	200	199,5	1,000	0,00	0,079 9	-21,95	0,079 9	-21,95	+12 -11	±1
24	250	251,2	0,999	-0,01	0,063 4	-23,95	0,063 4	-23,96	+12 -11	±1
25	315	316,2	0,998	-0,02	0,050 4	-25,96	0,050 3	-25,98	+12 -11	±1
26	400	398,1	0,995	-0,04	0,040 0	-27,96	0,039 8	-28,00	+12 -11	±1
27	500	501,2	0,988	-0,11	0,031 8	-29,96	0,031 4	-30,07	+12 -11	±1
28	630	631,0	0,970	-0,27	0,025 2	-31,96	0,024 5	-32,23	+12 -11	±1
29	800	794,3	0,929	-0,64	0,020 0	-33,96	0,018 6	-34,60	+12 -11	±1
30	1 000	1 000	0,846	-1,46	0,015 9	-35,96	0,013 5	-37,42	+26 -21	±2
31	1 250	1 259	0,707	-3,01	0,012 6	-37,96	0,008 94	-40,97	+26 -21	±2
32	1 600	1 585	0,534	-5,46	0,010 0	-39,96	0,005 36	-45,42	+26 -21	±2
33	2 000	1 995	0,370	-8,64	0,007 98	-41,96	0,002 95	-50,60	+26 -21	±2
34	2 500	2 512	0,244	-12,27	0,006 34	-43,96	0,001 54	-56,23	+26	+2
35	3 150	3 162	0,157	-16,11	0,005 03	-45,96	0,000 788	-62,07	+26	+2
36	4 000	3 981	0,099 5	-20,04	0,004 00	-47,96	0,000 398	-68,01	+26	+2
37	5 000	5 012	0,063 0	-24,02	0,003 18	-49,96	0,000 200	-73,98	+26	+2
38	6 300	6 310	0,039 8	-28,01	0,002 52	-51,96	0,000 100	-79,97	+26	+2
39	8 000	7 943	0,025 1	-32,00	0,002 00	-53,96	0,000 050	-85,97	+26	+2
40	10 000	10 000	0,015 8	-36,00	0,001 59	-55,96	0,000 025	-91,97	+26	+2

NOTE x is the frequency band number according to IEC 61260:1995.  
 The table is based on true frequencies (exact values as defined in IEC 61260:1995).  
 Tolerance applies to the total frequency weighting  $W_h$  for the whole instrument chain (see 5.1).

Page 17

**Annex A****A.3 Time constants**

Replace paragraph by the following.

The instrumentation shall allow the duration of measurement to be long enough to ensure reasonable statistical precision. The appropriate International Standard should be consulted for further guidance.

Page 18

**Annex B**

Replace by annex B by the following.

**Annex B**  
(normative)

**Mathematical definition of the frequency weightings**

**Table B.1 — Parameters of the transfer functions of frequency weightings**

Weighting	a) Band-limiting		b) Acceleration-velocity transition			c) Upward step			
	$f_1$ Hz	$f_2$ Hz	$f_3$ Hz	$f_4$ Hz	$Q_4$	$f_5$ Hz	$Q_5$	$f_6$ Hz	$Q_6$
$W_k$	0,4	100	12,5	12,5	0,63	2,37	0,91	3,35	0,91
$W_d$	0,4	100	2,0	2,0	0,63	$\infty$	—	$\infty$	—
$W_f$	0,08	0,63	$\infty$	0,25	0,86	0,062 5	0,80	0,1	0,80
W.B.combined	0,794 3	100	—	5,684	—	$\infty$	—	$\infty$	—
$W_c$	0,4	100	8,0	8,0	0,63	$\infty$	—	$\infty$	—
$W_e$	0,4	100	1,0	1,0	0,63	$\infty$	—	$\infty$	—
$W_j$	0,4	100	$\infty$	$\infty$	—	3,75	0,91	5,32	0,91
$W_h$	6,310	1 258,9	15,915	15,915	0,64	$\infty$	—	$\infty$	—

The frequencies  $f_1$  to  $f_6$  and the resonant quality factors  $Q_4$  to  $Q_6$  are parameters of the transfer functions determining the overall frequency weighting, referred to acceleration as input quantity. The transfer function is expressed as a product of several factors.

- a) Band-limiting (two-pole filter with Butterworth characteristic,  $Q_1 = Q_2 = 1/\sqrt{2}$ )

high-pass:

$$|H_h(p)| = \left| \frac{1}{1 + \sqrt{2}\omega_1 / p + (\omega_1 / p)^2} \right| = \sqrt{\frac{f^4}{f^4 + f_1^4}}$$

low-pass:

$$|H_l(p)| = \left| \frac{1}{1 + \sqrt{2}p / \omega_2 + (p / \omega_2)^2} \right| = \sqrt{\frac{f_2^4}{f^4 + f_2^4}}$$

- b) a-v transition (proportionality to acceleration at lower frequencies, proportionality to velocity at higher frequencies)

$$|H_t(p)| = \left| \frac{1 + p / \omega_3}{1 + p / (Q_4\omega_4) + (p / \omega_4)^2} \right| = \sqrt{\frac{f^2 + f_3^2}{f_3^2}} \cdot \sqrt{\frac{f_4^4 \cdot Q_4^2}{f^4 \cdot Q_4^2 + f^2 \cdot f_4^2(1 - 2Q_4^2) + f_4^4 \cdot Q_4^2}}$$

For W.B.combined this expression is replaced by

$$|H_t(p)| = \left| \frac{1}{1 + p / \omega_4} \right| = \sqrt{\frac{1}{1 + f^2 / f_4^2}}$$

- c) Upward step (steepness approximately 6 dB per octave, proportionality to jerk)

$$|H_s(p)| = \left| \frac{1 + p / (Q_5\omega_5) + (p / \omega_5)^2}{1 + p / (Q_6\omega_6) + (p / \omega_6)^2} \cdot \left( \frac{\omega_5}{\omega_6} \right)^2 \right| = \frac{Q_6}{Q_5} \cdot \sqrt{\frac{f^4 \cdot Q_5^2 + f^2 \cdot f_5^2(1 - 2Q_5^2) + f_5^4 \cdot Q_5^2}{f^4 \cdot Q_6^2 + f^2 \cdot f_6^2(1 - 2Q_6^2) + f_6^4 \cdot Q_6^2}}$$

where

$$\omega_n = 2\pi f_n \quad (n = 1 \text{ to } 6)$$

$$p = j\omega \text{ (imaginary angular frequency).}$$

The product  $H_h(p) \cdot H_l(p)$  represents the band-limiting transfer function. Its corner frequencies (intersection of asymptotes) are given by  $f_1$  and  $f_2$ .

The product  $H_t(p) \cdot H_s(p)$  represents the actual frequency weighting transfer function for a certain application.  $H_t(p) = 1$  for frequency weighting  $W_j$ ;  $H_s(p) = 1$  for frequency weightings  $W_c$ ,  $W_d$ ,  $W_e$ ,  $W_h$  and W.B.combined. This is indicated by infinity frequencies and absence of quality factors in table B.1.

The total weighting function is

$$H(p) = H_h(p) \cdot H_l(p) \cdot H_t(p) \cdot H_s(p)$$

The weighting curves in annex C show the modulus (magnitude)  $|H|$  of  $H$  versus the frequency  $f$  in a double-logarithmic scale.

NOTE 1 If the equations are interpreted in the time domain, the imaginary angular frequency  $p$  leads directly to the differential operator  $dr$ , which may be approximated by  $\Delta t$  in a digital realization of the weighting if the sampling interval  $\Delta t$  is small enough.