



INTERNATIONAL STANDARD ISO 8041:2005
TECHNICAL CORRIGENDUM 1

Published 2007-11-01

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Human response to vibration — Measuring instrumentation

TECHNICAL CORRIGENDUM 1

Réponse des individus aux vibrations — Appareillage de mesure

RECTIFICATIF TECHNIQUE 1

Technical Corrigendum 1 to ISO 8041:2005 was prepared by Technical Committee ISO/TC 108, *Mechanical vibration, shock and condition monitoring*, Subcommittee SC 3, *Use and calibration of vibration and shock measuring instruments*.

Page 26, 12.7, after Equation (15)

Add:

“NOTE The error, ε , is a relative measure, i.e. a relative error expressed as percentage.”

Page 30, 12.1.1, after Table 15

Add:

“NOTE The errors, ε , mentioned in 12.11 are relative measures, i.e. relative errors expressed as percentages.”