



INTERNATIONAL STANDARD ISO 8185:1997
TECHNICAL CORRIGENDUM 1

Published 2001-05-15

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Humidifiers for medical use — General requirements for humidification systems

TECHNICAL CORRIGENDUM 1

Humidificateurs médicaux — Exigences générales relatives aux systèmes d'humidification

RECTIFICATIF TECHNIQUE 1

Technical Corrigendum 1 to International Standard ISO 8185:1997 was prepared by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 3, *Lung ventilators and related equipment*.

Page 17

Section 10

56.3 Connections – General

Section 56.3 is not intended to exclude the use of 22 mm connections for humidifiers for paediatric use. Thus, the second sentence should be modified to read as follows:

“If intended for adult use, a 22 mm connector shall be provided.”

Delete the word “should” from the middle of the next paragraph, to read

“...tubes which comply with ISO 5367, ...”