
**Thermal insulation — Determination of
steady-state thermal resistance and
related properties — Heat flow meter
apparatus**

AMENDMENT 1

*Isolation thermique — Détermination de la résistance thermique et des
propriétés connexes en régime stationnaire — Méthode fluxmétrique*

AMENDEMENT 1

STANDARDSISO.COM : Click to view the full PDF of ISO 8301:1991/Amd 1:2010



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

STANDARDSISO.COM : Click to view the full PDF of ISO 8301:1991/Amd 1:2010



COPYRIGHT PROTECTED DOCUMENT

© ISO 2010

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

Amendment 1 to ISO 8301:1991 was prepared by Technical Committee ISO/TC 163, *Thermal performance and energy use in the built environment*, Subcommittee SC 1, *Test and measurement methods*.

STANDARDSISO.COM : Click to view the full PDF of ISO 8301:1991/Amd.1:2010

STANDARDSISO.COM : Click to view the full PDF of ISO 8301:1991/Amd 1:2010

Thermal insulation — Determination of steady-state thermal resistance and related properties — Heat flow meter apparatus

AMENDMENT 1

Page 1, 1.2

Delete footnote 1.

Page 5, 1.5.3.3

Add the following note at the end of the subclause:

NOTE Comparative tests under the European Key Mark scheme have shown that it is possible to achieve reproducibility of 1,5 % using careful calibration and careful operational procedures.

Pages 5 and 6, 1.5.4.2

At the end of the last sentence, replace “in accordance with 1.5.4.1.” with “in accordance with 1.5.4.1 and 2.5.4.”

Page 6, 1.7.2.1

At the end of the second sentence, add “(e.g. the use of three-dimensional finite elements calculations for evaluation of edge effects).”

Pages 9 and 10, 2.2.1.2

Retain the first three lines and delete the rest of the text in the subclause.

Page 13, 2.2.3.1.1

Replace the following text in the last three sentences of the second paragraph:

The number of such thermocouples on each side should be not less than $10\sqrt{A}$, or 2, whichever is greater, where A is the metering area section in square metres. One thermocouple per surface has been found adequate on existing plates having surface areas of less than 0,04 m², provided that either the thermocouples are changed frequently or the thermocouple calibration is checked regularly. A minimum of two thermocouples is required for all new apparatus.

with the following:

The number of such thermocouples on each side shall not be not less than $10\sqrt{A}$, or 2, whichever is greater, where A is the metering area section in square metres. One thermocouple per surface has been found adequate when testing high resistance specimens using plate thermocouples only without contact sheets.

Page 14, 2.2.3.1.3

Delete the last sentence of the last paragraph.

Page 14, 2.2.3.2.2

Replace the first part of the first paragraph "When the apparatus is used for high-level performance in measurements or reporting testing, the measuring system shall have the following capabilities, regardless of whether a wide-range or a multirange measuring system is used: " with: "The measuring system shall have the following capabilities, regardless of whether a wide-range or a multirange measuring system is used."

Page 14, 2.2.3.2.2

Delete list item f).

Page 15, 2.2.5.1

Add at the end of the first paragraph: "Finite element three-dimensional simulation shall be carried out to ensure one-dimensional heat flow in the metering area".

Page 15, 2.2.5.1

In the third paragraph, change "5 K" to "2 K" and add the following note:

NOTE For specimens with fine capillaries or pores, up to 5 K can be needed.

Pages 15 and 16, 2.2.5.3

Delete the subclause.

Page 16, 2.3.4

Add at the end of the subclause "Finite element three-dimensional simulation shall be carried out to ensure one-dimensional heat flow in the metering area."

Page 19, 2.5.3.1.4

Delete the subclause.