

---

---

**Piston-operated volumetric apparatus —**

**Part 2:  
Piston pipettes**

*Appareils volumétriques à piston —*

*Partie 2: Pipettes à piston*

STANDARDSISO.COM : Click to view the full PDF of ISO 8655-2:2002



**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

STANDARDSISO.COM : Click to view the full PDF of ISO 8655-2:2002

© ISO 2002

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.ch](mailto:copyright@iso.ch)  
Web [www.iso.ch](http://www.iso.ch)

Printed in Switzerland

**Contents**

	Page
1 Scope .....	1
2 Normative references .....	1
3 Terms and definitions .....	2
4 Principle of operation .....	2
5 Design .....	2
6 Type, designation .....	5
7 Metrological performance requirements .....	5
8 Marking .....	8

**Annexes**

<b>A</b> Information to be provided by the supplier for the user of piston pipettes and tips .....	<b>9</b>
<b>B</b> Possible sources of error for piston pipettes with air interface .....	<b>10</b>
Bibliography.....	11

STANDARDSISO.COM : Click to view the full PDF of ISO 8655-2:2002

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this part of ISO 8655 may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

International Standard ISO 8655-2 was prepared by Technical Committee ISO/TC 48, *Laboratory glassware and related apparatus*, Subcommittee SC 1, *Volumetric instruments*.

ISO 8655 consists of the following parts, under the general title *Piston-operated volumetric apparatus*:

- *Part 1: Terminology, general requirements and user recommendations*
- *Part 2: Piston pipettes*
- *Part 3: Piston burettes*
- *Part 4: Dilutors*
- *Part 5: Dispensers*
- *Part 6: Gravimetric methods for the determination of measurement error*

The following part is under preparation:

- *Part 7: Non-gravimetric methods for the determination of measurement error*

Annex A forms a normative part of this part of ISO 8655. Annex B is for information only.

## Introduction

ISO 8655 addresses the needs of:

- suppliers, as a basis for quality control including, where appropriate, the issuance of supplier's declarations;
- test houses and other bodies, as a basis for independent certification;
- users of the equipment, to enable routine checking of accuracy.

The tests specified should be carried out by trained personnel.

STANDARDSISO.COM : Click to view the full PDF of ISO 8655-2:2002

STANDARDSISO.COM : Click to view the full PDF of ISO 8655-2:2002

# Piston-operated volumetric apparatus —

## Part 2: Piston pipettes

### 1 Scope

This part of ISO 8655 specifies

- metrological requirements,
- maximum permissible errors,
- requirements for marking and
- information to be provided for users,

for air-displacement (type A) and positive-displacement (type D) single-channel and multi-channel piston pipettes, complete with their selected tip(s) and any other essential, consumable parts, designed to deliver their specified nominal volume (Ex).

NOTE General requirements and definitions of terms of piston-operated volumetric apparatus are given in ISO 8655-1. Conformity testing (type evaluation) of piston-operated volumetric apparatus is given in ISO 8655-6. Alternative test methods such as photometric and titrimetric methods will be the subject of a future Part 7 to ISO 8655. For all other tests (e.g. quality assurance by the supplier, analytical and measuring equipment quality assurance by the user) see ISO 8655-6 or alternative test methods. For safety requirements of electrically powered piston pipettes, see IEC 61010-1.

### 2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this part of ISO 8655. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this part of ISO 8655 are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

ISO 8655-1:2002, *Piston-operated volumetric apparatus — Part 1: Terminology, general requirements and user recommendations*

ISO 8655-6:2002, *Piston-operated volumetric apparatus — Part 6: Gravimetric methods for the determination of measurement error*

### 3 Terms and definitions

For the purposes of this part of ISO 8655, the terms and definitions given in ISO 8655-1 and the following apply.

#### 3.1

##### **nominal volume**

⟨variable-volume piston pipette⟩ greatest volume selectable by the user and specified by the manufacturer

NOTE This means that a variable-volume piston pipette with a useful volume range (see ISO 8655-1) of 10  $\mu\text{l}$  to 100  $\mu\text{l}$  has the nominal volume of 100  $\mu\text{l}$ .

#### 3.2

##### **nominal volume**

⟨multi-channel piston pipette⟩ greatest volume selectable by the user and specified by the manufacturer per channel

### 4 Principle of operation

The tip made of plastic or glass is attached to the piston pipette. With the piston at the lower aspiration limit, the tip is dipped into the liquid to be dispensed as a measured volume. When moved to the upper aspiration limit, the piston aspirates the liquid. The liquid volume to be dispensed is then expelled by depressing or sliding the piston between the volume-defining limits. Some air-displacement piston pipettes (see 5.1, type A) have an extra air volume which can be used to expel the last drop of liquid.

See also Figure 1.

### 5 Design

#### 5.1 Types of piston pipette

A piston pipette may be designed as follows:

- fixed volume, designed by the manufacturer to dispense only its nominal volume, for example 100  $\mu\text{l}$ ;
- variable volume, designed by the manufacturer to dispense volumes selectable by the user within its specified useful volume range, for example between 10  $\mu\text{l}$  and 100  $\mu\text{l}$ .

The piston may


- either have a body of air contained between the piston and the surface of the liquid (air displacement – type A);
- or be in direct contact with the surface of the liquid (positive or direct displacement – type D).

In the case of the type D piston pipette, either the plunger or the capillary, or both may be reusable (type D1) or disposable (type D2).

See Figure 1 for details.

#### 5.2 Adjustment

**5.2.1** A piston pipette shall be adjusted by its manufacturer for the delivery (Ex) of its nominal volume (or selected volume, in the case of a variable-volume model), for the standard reference temperature of 20 °C, a relative air humidity of 50 % and a barometric pressure of 101 kPa (1,01 bar), using grade 3 water as specified in ISO 3696.

**Type D**

Positive (direct) displacement

**Type A**

Air displacement

**a) Single-channel piston pipettes****b) Multi-channel piston pipette****Figure 1 — Piston pipettes**

**5.2.2** Some piston pipettes have provision for adjustment by the user when, for example, it is found in routine calibration that the volume delivered is not within specification. Such user adjustment shall be made according to the manufacturer's instructions and by reference to the gravimetric test method specified in ISO 8655-6.

Any piston pipette so adjusted shall have clear, visible evidence that the initial adjustment has been modified. This information shall also be recorded on any certificate of conformity and the procedure should be included in the user's quality manual.

**5.2.3** Some piston pipettes are designed to have their factory-preset adjustment altered by the user so that they will dispense their specified volume when used with liquids with physical properties differing from those of water (see annex B for details). In such cases, the design shall prevent unintentional readjustment. Such user adjustment shall be made according to the manufacturer's instructions and by reference to the gravimetric test method specified in ISO 8655-6, but using the selected liquid instead of water.

If the user readjusts the piston pipette, it shall be clearly and unequivocally indicated on the outside of the piston pipette that readjustment has been effected. The user shall mark the outside of the piston pipette with the name of the liquid for which the nominal volume now applies. This information shall also be recorded on any certificate of conformity. The procedure should also be included in the user's quality manual.

**5.3 Transfer of hand warmth**

The constructional design of piston pipettes and the materials used for their manufacture shall be chosen in such a way that any heat transmitted from the user's hand to the apparatus during periods of use or of test in accordance with ISO 8655-6 may be ignored.

NOTE Transfer of hand warmth will appear as a systematic drift of results during the series of tests to ISO 8655-6.

## 5.4 Pipette tips

### 5.4.1 General

**5.4.1.1** The dispensing orifice of the tip shall be shaped in such a way that consistent dispensing of the liquid to be measured is ascertained. When the tip is touched against the wall of a vessel in successive operations, any amount of liquid remaining in or around the dispensing orifice of the tip shall be consistent.

**5.4.1.2** In the case of sterilizable pipette tips the sterilization procedures indicated as appropriate by the manufacturer in user information or on packaging (see annex A) shall not negatively affect the metrological characteristics of the tips such as shape, seal and wettability.

NOTE This requirement can be assessed by comparing errors of measurement using tips which have and have not been sterilized.

### 5.4.2 Air-displacement pipette tips

**5.4.2.1** Air-displacement pipette tips shall be disposable parts, usually made of plastic, which fit on the tip holder of the pipette and prevent the instrument from contact with the aspirated liquid. To assure their metrological performance they shall conform to 5.4.1.1, 5.4.1.2, 5.4.2.2 and 5.4.2.3.

**5.4.2.2** Tips for piston pipettes with air interface shall be fitted in accordance with the pipette supplier's instructions to form a good seal between the tip and the tip cone of the piston pipette.

NOTE Variability of amount of externally retained liquid or an incomplete seal will contribute to poor precision when testing to ISO 8655-6.

Pipette tips made of plastic for piston pipettes with air interface are designed for single use. They shall not be cleaned for reuse as their metrological characteristics will no longer be reliable.

**5.4.2.3** The form of the pipette tips to be used with a multi-channel piston pipette shall be so straight that all tips fitted are positioned with parallel axes in the same plane in order to allow for even liquid dispensing in the target vessels, e.g. the adjacent wells of a microtiter plate. The bottoms of properly fitted tips shall not vary in spacing from their nominal axes, nor from their common plane by more than  $\pm 1,0$  mm for nominal volumes up to 350  $\mu\text{l}$  and  $\pm 1,5$  mm for nominal volumes exceeding 350  $\mu\text{l}$ .

### 5.4.3 Positive-displacement pipette tips

**5.4.3.1** Positive-displacement pipette tips shall consist of a plunger and a capillary which fit on the tip holder of the pipette. Various material may be used for the plunger, such as metal or plastic, and the capillary, such as plastic or glass. These pipette tips may be reusable or disposable (both plunger and capillary is changed at each sampling).

**5.4.3.2** The shape and material of the plunger and capillary shall confer a good seal of the tip, as well as a smooth action between the plunger and the capillary, to ensure consistent dispensing of the liquid.

**5.4.3.3** Sterilizability shall be in accordance with 5.4.1.2.

## 5.5 Type evaluation

Type testing of the design requirements specified in 5.2 to 5.4 shall satisfy the metrological performance requirements specified in clause 7 when the pipette is tested in accordance with ISO 8655-6.

## 6 Type, designation

Designation of a type D1 single-channel piston pipette (positive displacement with reusable plunger) with a nominal volume of 100  $\mu\text{l}$ :

**Pipette ISO 8655 - D1 - 100**

Designation of a type D2 variable-volume single-channel piston pipette (positive displacement with disposable plunger/capillary), volume range variable from 20  $\mu\text{l}$  to 200  $\mu\text{l}$ :

**Pipette ISO 8655 - D2 - 20-200**

Designation of a type A variable-volume single-channel piston pipette with air interface, volume range variable from 10  $\mu\text{l}$  to 100  $\mu\text{l}$ :

**Pipette ISO 8655 - A - 10-100**

Designation of a 8-channel piston pipette with air interface (A) and with a nominal volume of 200  $\mu\text{l}$ :

**Pipette ISO 8655 - A - 200 $\times$ 8**

Designation of a 12-channel piston pipette with air interface (A), volume range variable from 20  $\mu\text{l}$  to 200  $\mu\text{l}$ :

**Pipette ISO 8655 - A - 20-200 $\times$ 12**

## 7 Metrological performance requirements

### 7.1 General

#### 7.1.1 Conformity tests

The conformity test specified in ISO 8655-6 evaluates the total system of piston pipette with its selected tip or tip and piston combination and its operator. The conformity test shall be carried out in accordance with ISO 8655-6 by a test house or other authorized body prior to the issuance of a certificate of conformity.

The type test shall be carried out by the supplier prior to the issuance of a supplier's declaration of conformity.

For conformity tests, the maximum permissible errors of Tables 1 and 2 shall apply.

#### 7.1.2 User tests

Users should establish a regular testing routine (see ISO 8655-1:2002, 7.3) for their piston pipettes using either ISO 8655-6 or alternative test methods (see note to clause 1) and having regard to:

- accuracy requirements of the liquid delivery undertaken;
- frequency of use;
- number of operators using the piston pipette;
- number of dispensings performed on each occasion of use;
- nature of the liquid dispensed (corrosiveness, solvent power etc.);
- any recommendations made by the supplier.

Annex B lists parameters which influence the metrological performance of piston pipettes and recommendations for their handling.

**7.2 Fixed-volume piston pipettes of types A and D1**

For air-displacement piston pipettes (type A) with fixed volume and for positive-displacement piston pipettes with reusable plunger and capillary (type D1) and with fixed volume, the maximum permissible errors of Table 1 apply.

For the maximum permissible errors of piston pipettes with intermediate nominal volumes between those given in Table 1, the absolute values for the next greater nominal volume shall apply.

NOTE This means that the maximum permissible systematic error of a piston pipette with a nominal volume of 25 µl is ± 0,5 µl and its maximum permissible random error is ± 0,2 µl.

**Table 1 — Maximum permissible errors for types A and D1**

Nominal volume µl	Maximum permissible systematic error		Maximum permissible random error	
	± %	± µl <sup>a</sup>	± % <sup>b</sup>	± µl <sup>c</sup>
1	5,0	0,05	5,0	0,05
2	4,0	0,08	2,0	0,04
5	2,5	0,125	1,5	0,075
10	1,2	0,12	0,8	0,08
20	1,0	0,2	0,5	0,1
50	1,0	0,5	0,4	0,2
100	0,8	0,8	0,3 <sup>d</sup>	0,3 <sup>d</sup>
200	0,8	1,6	0,3 <sup>d</sup>	0,6 <sup>d</sup>
500	0,8	4,0	0,3	1,5
1 000	0,8	8,0	0,3	3,0
2 000	0,8	16	0,3	6,0
5 000	0,8	40	0,3	15,0
10 000	0,6	60	0,3	30,0

<sup>a</sup> Expressed as the deviation of the mean of a tenfold measurement from the nominal or selected volume (see ISO 8655-6:2002, 8.4).

<sup>b</sup> Expressed as the coefficient of variation of a tenfold measurement (see ISO 8655-6:2002, 8.5).

<sup>c</sup> Expressed as the repeatability standard deviation of a tenfold measurement (see ISO 8655-6:2002, 8.5).

<sup>d</sup> For piston pipettes of type D1 the maximum permissible errors may be ± 0,4 %.

**7.3 Fixed-volume piston pipettes of type D2**

For positive-displacement piston pipettes with disposable plunger and capillary (type D2) and with fixed volume, the maximum permissible errors of Table 2 apply.

For the maximum permissible errors of piston pipettes with intermediate nominal volumes between those given in Table 2, the absolute values for the next greater nominal volume shall apply.

**7.4 Variable-volume piston pipettes**

In the case of variable-volume piston pipettes, the nominal volume is the greatest possible user-selectable volume which is specified by the manufacturer. For example, a piston pipette with a useful volume range from 10 µl to 100 µl has a nominal volume of 100 µl.

In the conformity test, the maximum permissible errors for the nominal volumes in Tables 1 and 2 apply to every selectable volume throughout the useful volume range of the piston pipette; i.e. the maximum permissible systematic errors of variable-volume piston pipette types A and D1 with a useful volume range of 10 µl to 100 µl are ± 0,8 µl and the maximum permissible random errors are ± 0,3 µl for every measured volume.

NOTE It is recognized that the errors of measurement of intermediate volumes in the useful volume range of a variable-volume piston pipette may be considerably better than those specified for the nominal volume of the piston pipette.

Table 2 — Maximum permissible errors for type D2

Nominal volume $\mu\text{l}$	Maximum permissible systematic error		Maximum permissible random error	
	$\pm \%$	$\pm \mu\text{l}^{\text{a}}$	$\pm \%^{\text{b}}$	$\pm \mu\text{l}^{\text{c}}$
5	2,5	0,13	1,5	0,08
10	2,0	0,2	1,0	0,1
20	2,0	0,4	0,8	0,16
50	1,4	0,7	0,6	0,3
100	1,5	1,5	0,6	0,6
200	1,5	3,0	0,4	0,8
500	1,2	6,0	0,4	2,0
1 000	1,2	12,0	0,4	4,0

<sup>a</sup> Expressed as the deviation of the mean of a tenfold measurement from the nominal or selected volume (see ISO 8655-6:2002, 8.4).  
<sup>b</sup> Expressed as the coefficient of variation of a tenfold measurement (see ISO 8655-6:2002, 8.5).  
<sup>c</sup> Expressed as the repeatability standard deviation of a tenfold measurement (see ISO 8655-6:2002, 8.5).

Repetitive piston pipettes with a continuously selectable volume shall be considered as variable-volume piston pipettes of type D2 and the maximum permissible errors of Table 2 shall apply.

## 7.5 Multi-channel piston pipettes

The maximum permissible systematic and random errors of multi-channel piston pipettes shall be equal to twice the values specified in Table 1 for single-channel piston pipettes. Each channel of the multi-channel piston pipette, considered independently, shall meet these specifications.

In the case of variable-volume multi-channel piston pipettes, the maximum permissible errors for the nominal volume apply to every selectable volume throughout the useful volume range (see 7.4 for further details).

## 7.6 Pipette tips

The maximum permissible errors always apply to the total system of piston pipette and tip. When using pipette tips not delivered by the pipette supplier, the supplier's declaration or the certificate of conformity does not apply. In such cases, the evaluation of the maximum permissible errors shall be performed using the alternative tips with the method of test specified in ISO 8655-6.

Before the conformity evaluation of the piston pipette with the alternative tips, it shall be ascertained, that the piston pipette in combination with the original tips conforms to the metrological requirements specified in this part of ISO 8655.

Users may test the metrological performance for routine quality assurance procedures of the total system of

- piston pipette and tip (type A) and operator, or
- tip and piston combination (type D) and operator

in accordance with ISO 8655-6 or with alternative test methods (see note to clause 1).

## 8 Marking

### 8.1 Piston pipettes

The following data shall be permanently marked on each piston pipette:

- a) nominal volume or, for variable-volume piston pipettes, useful volume range;
- b) unit of measurement, e.g.  $\mu\text{l}$  or ml;
- c) supplier's name and/or trademark;
- d) supplier's apparatus name or type;
- e) apparatus identification serial number or an equivalent unique identifier.

In addition, the following information should be marked:

- f) abbreviation "Ex" and the reference temperature "20 °C";
- g) reference to this part of ISO 8655, i.e. "ISO 8655-2".

Where appropriate, a conformity mark may be affixed to the piston pipette.

### 8.2 Pipette tips

The following information shall be printed on the packaging of each saleable unit of pipette tips:

- a) supplier's name and/or trademark;
- b) maximum volume to be dispensed with the tip;
- c) statement of whether or not the tips are suitable for sterilization.

The supplier should indicate, where appropriate, procedures for sterilization of the pipette tips.