
**Piston-operated volumetric
apparatus —**

**Part 5:
Dispensers**

*Appareils volumétriques à piston —
Partie 5: Distributeurs*

STANDARDSISO.COM : Click to view the full PDF of ISO 8655-5:2022



STANDARDSISO.COM : Click to view the full PDF of ISO 8655-5:2022



COPYRIGHT PROTECTED DOCUMENT

© ISO 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword.....	iv
Introduction.....	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principle of operation	2
5 Adjustment	2
5.1 Basis of adjustment.....	2
5.2 Initial adjustment.....	3
5.3 Subsequent adjustment.....	3
6 Metrological performance requirements	3
6.1 General.....	3
6.2 Calculation of maximum permissible errors not given in Tables 1 and 2	3
6.3 Dispenser tips.....	6
7 User information	6
8 Marking	6
8.1 Dispenser.....	6
8.2 Dispenser tips.....	7
Bibliography	8

STANDARDSISO.COM : Click to view the full PDF of ISO 8655-5:2022

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 48, *Laboratory equipment*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 332, *Laboratory equipment*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 8655-5:2002), which has been technically revised. It also incorporates the Technical Corrigendum ISO 8655-5:2002/Cor.1:2008.

The main changes are as follows:

- ISO 8655-7 and ISO 8655-8 have been added as normative references;
- metrological performance requirements for dispenser tips have been further specified;
- [Tables 1](#) and [2](#) have been revised.

A list of all parts in the ISO 8655 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

The ISO 8655 series addresses the needs of:

- manufacturers, as a basis for quality control including, where appropriate, the issuance of manufacturer's declarations;
- calibration laboratories, test houses, users of the equipment and other bodies as a basis for independent calibration, testing, verification and routine tests.

The tests specified in the ISO 8655 series are intended to be carried out by trained personnel.

STANDARDSISO.COM : Click to view the full PDF of ISO 8655-5:2022

[STANDARDSISO.COM](https://standardsiso.com) : Click to view the full PDF of ISO 8655-5:2022

Piston-operated volumetric apparatus —

Part 5: Dispensers

1 Scope

This document specifies

- metrological requirements,
- maximum permissible errors,
- requirements for marking and
- information to be provided for users,

for dispensers. It is applicable to dispensers with nominal volumes from 1 µl up to 200 ml, designed to deliver their volume (Ex).

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3696:1987, *Water for analytical laboratory use — Specification and test methods*

ISO 8655-1, *Piston-operated volumetric apparatus — Part 1: Terminology, general requirements and user recommendations*

ISO 8655-6, *Piston-operated volumetric apparatus — Part 6: Gravimetric reference measurement procedure for the determination of volume*

ISO 8655-7, *Piston operated volumetric apparatus — Part 7: Alternative measurement procedures for the determination of volume*

ISO 8655-8, *Piston-operated volumetric apparatus — Part 8: Photometric reference measurement procedure for the determination of volume*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 8655-1 apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

4 Principle of operation

Dispensers are used for the accurate repetitive delivery (dispensing) of preset liquid volumes. They are of two types:

- single-stroke dispensers providing a single delivery from each filling stroke;
- multiple delivery dispensers or ratchet-based systems providing multiple deliveries from each filling stroke.

The piston can be operated manually, electrically, pneumatically or hydraulically. Drive mechanism, piston and cylinder can be a single unit or can be separable by simple hand actions, so that different pistons and cylinders (change-over units) can be used with the same drive mechanism.

During operation, the aspiration tube dips into the reservoir containing the fluid to be dispensed. After the system has been primed with fluid, assuring that it is free of any air, the piston aspirates fluid by moving in one direction and delivers the fluid to be measured by moving in the opposite direction.

Dispensers can be constructed with or without valves, see [Figure 1](#). The metrological characteristics of dispensers are dependent on, among other things, the material and workmanship of the aspiration and delivery tubing used. The dispenser shall therefore be provided together with the necessary tubing or suggestion or statement about the characteristics the tubing shall have in order to meet the metrological demands placed on the system.

Manufacturers' instruction manuals should contain detailed and specific information about the proper operation of dispensers.




Figure 1 — Schematic drawing of dispensers with and without valves

5 Adjustment

5.1 Basis of adjustment

A dispenser shall be adjusted for the delivery (Ex) of its nominal volume (or selected volume, in the case of a variable-volume model).

For countries that have adopted the standard reference temperature of 20 °C, the adjustment shall be for the temperature of 20 °C, a relative air humidity of 50 % and a barometric pressure of 101,3 kPa, when handling grade 3 water as specified in ISO 3696:1987.

For those countries that have adopted a standard reference temperature of 27 °C, the adjustment shall be for the temperature 27 °C, a relative air humidity of 50 % and a barometric pressure of 101,3 kPa, when handling grade 3 water as specified in ISO 3696:1987.

5.2 Initial adjustment

A dispenser shall be provided with an initial adjustment.

5.3 Subsequent adjustment

Some dispensers have provision for adjustment when, for example, it is found in routine checks that the volume delivered is not within the specification. Such adjustment shall be made according to the manufacturer's instructions and by reference to a measurement procedure in accordance with ISO 8655-6, ISO 8655-7 or ISO 8655-8.

Any dispenser so adjusted shall have clear, visible evidence that the initial adjustment has been modified. This information shall also be recorded.

6 Metrological performance requirements

6.1 General

In order to state the metrological trueness and precision of the total system of the dispenser and thus determine its systematic and random errors, a reference measurement procedure as specified in ISO 8655-6 and ISO 8655-8 or a measurement method in accordance with ISO 8655-7 shall be used. The total system consists of the drive and, if applicable, the change-over unit (see [Clause 4](#)) including aspiration and delivery tubing and shall be included in the measurements. The maximum permissible errors given in [Tables 1](#) and [2](#) shall apply.

If metrological performance data are issued for change-over units by the manufacturer, the manufacturer shall specify the combination of drive and change-over unit to which the metrological performance data applies.

6.2 Calculation of maximum permissible errors not given in [Tables 1](#) and [2](#)

The calculation of maximum permissible systematic and random error in the usable volume range, not included in [Tables 1](#) and [2](#), shall be made by dividing the nominal volume by the selected volume and multiplying the result by the maximum permissible errors at nominal volume. This calculation does not apply to volumes below 10 % of the nominal volume.

[Formula \(1\)](#) shall be applied for the calculation:

$$e_{V_s} = \frac{V_{\text{nom}}}{V_s} \times e_{V_{\text{nom}}} \quad (1)$$

where

V_{nom} is the nominal volume;

V_s is the selected volume;

$e_{V_{\text{nom}}}$ is the maximum permissible error (either systematic or random) at nominal volume;

e_{V_s} is the maximum permissible error (either systematic or random) at the selected volume.

If the calculated value exceeds 25 %, then the value of 25 % shall be applied as the maximum permissible error.

EXAMPLE Single-stroke dispenser with a nominal volume of 10 ml and a usable volume range of 1–10 ml. Calculation of maximum permissible systematic error at a selected volume of 2 ml:

$$e_{V_{\text{nom}}} = 0,60 \%$$

$$V_{\text{nom}} = 10 \text{ ml}$$

$$V_s = 2 \text{ ml}$$

$$e_{V_s} = \frac{V_{\text{nom}}}{V_s} \times e_{V_{\text{nom}}}$$

$$e_{V_s(2 \text{ ml})} = \frac{10 \text{ ml}}{2 \text{ ml}} \times 0,6 \%$$

$$e_{V_s(2 \text{ ml})} = 5 \times 0,6 \%$$

$$e_{V_s(2 \text{ ml})} = 3 \%$$

Table 1 — Maximum permissible errors for single-stroke dispensers

Nominal volumes ml	Dispensing volume Setting as a proportion of the nominal volume %	Maximum permissible sys- tematic error ^a	Maximum permissible ran- dom error ^a
		±%	% ^b
0,01	100	2,0	1,0
	50	4,0	2,0
	10	20	10
> 0,01 to 0,02	100	2,0	0,50
	50	4,0	1,0
	10	20	5,0
> 0,02 to 0,05	100	1,5	0,40
	50	3,0	0,80
	10	15	4,0
> 0,05 to 0,1	100	1,5	0,30
	50	3,0	0,60
	10	15	3,0
> 0,1 to 0,2	100	1,0	0,30
	50	2,0	0,60
	10	10	3,0
> 0,2 to 0,5	100	1,0	0,20
	50	2,0	0,40
	10	10	2,0
> 0,5 to 200	100	0,60	0,20
	50	1,2	0,40
	10	6,0	2,0

^a To calculate errors in units of millilitres, multiply the maximum permissible errors by the selected volume.

^b Expressed as the coefficient of variation according to ISO 8655-6, ISO 8655-7, or ISO 8655-8.

Table 2 — Maximum permissible errors for multiple-delivery dispensers

Dispensing volume		Maximum permissible systematic error ^a ±%	Maximum permissible random error ^a % ^b
Nominal volumes ml	Setting as a proportion of the nominal volume %		
0,001 to 0,002	100	5,0	5,0
	50	10	10
	10	25	25
> 0,002 to 0,003	100	2,5	3,5
	50	5,0	7,0
	10	25	25
> 0,003 to 0,01	100	2,0	2,5
	50	4,0	5,0
	10	20	25
> 0,01 to 0,02	100	1,5	2,0
	50	3,0	4,0
	10	15	20
> 0,02 to 0,05	100	1,0	1,5
	50	2,0	3,0
	10	10	15
> 0,05 to 0,2	100	1,0	1,0
	50	2,0	2,0
	10	10	10
> 0,2 - 0,5	100	1,0	0,60
	50	2,0	1,2
	10	10	6,0
> 0,5 to 1	100	1,0	0,40
	50	2,0	0,80
	10	10	4,0
> 1 to 2	100	0,80	0,40
	50	1,6	0,80
	10	8,0	4,0
> 2 to 5	100	0,60	0,30
	50	1,2	0,60
	10	6,0	3,0
> 5 to 25	100	0,50	0,30
	50	1,0	0,60
	10	5,0	3,0
> 25 to 200	100	0,50	0,25
	50	1,0	0,50
	10	5,0	2,5

^a To calculate errors in units of millilitres, multiply the maximum permissible errors by the selected volume.

^b Expressed as the coefficient of variation according to ISO 8655-6, ISO 8655-7, or ISO 8655-8.

6.3 Dispenser tips

The maximum permissible errors always apply to the total system of dispenser and tip. Before the dispenser tip is placed on the market, it shall be proven that the system, including the tips, fulfils the requirements and maximum permissible errors of this document. The different batches of tips as well as the different moulds/cavities used shall be taken into account.

A set of measurements shall be performed in accordance with ISO 8655-6, ISO 8655-7 or ISO 8655-8.

In the case of sterilisable dispenser tips, the sterilisation procedures indicated as appropriate by the manufacturer in user information or on packaging (see [Clause 7](#)) shall not negatively affect the metrological characteristics of the tips such as shape, seal and wettability.

NOTE This requirement can be checked by comparing errors of measurement using tips which have been and have not been sterilised.

7 User information

Information essential to the proper use of the apparatus and its accessories (see ISO 8655-1) shall be provided when making a dispenser available on the market and shall be as follows:

- a) the basis of adjustment (Ex) at reference conditions according to ISO 8655-1;
- b) nominal volume; where this is not practicable [see [Clause 8](#), item a)], information shall be provided to enable the nominal volume to be correctly identified from markings on the appropriate unit or module;
- c) smallest volume which can be delivered observing the maximum permissible errors in accordance with [Clause 6](#);
- d) the correct method of use;
- e) the error limits of the systematic and random error of measurement at the nominal volume, at 50 % of the nominal volume and either at 10 % of the nominal volume or the smallest selectable volume whichever is greater;
- f) information regarding the care, cleaning and maintenance of the dispenser;
- g) advice about air-bubble-free filling;
- h) suggestions as to the basis on which a minimum routine testing protocol can be established including a reference to this document, i.e. ISO 8655-5;
- i) upon request, information regarding the interaction of the materials of the dispenser with organic and inorganic solutions, solvents and caustic chemicals;
- j) recommendations for the proper storage of the dispenser.

8 Marking

8.1 Dispenser

The following data shall be permanently marked on each dispenser:

- a) nominal volume; where this is not practicable due to a modular design, the volumetric modules shall be marked either with their nominal volumes or information to enable the user to obtain the nominal volume from the user information [see [Clause 7](#), item b)];
- b) unit of measurement;
- c) brand name and/or trademark;