

INTERNATIONAL
STANDARD

ISO
9022-10

First edition
1994-07-15

**Optics and optical instruments —
Environmental test methods —**

Part 10:

Combined sinusoidal vibration, dry heat or cold

*Optique et instruments d'optique — Méthodes d'essais
d'environnement —*

Partie 10: Essai combiné de vibrations sinusoïdales-chaleur sèche ou froid



Reference number
ISO 9022-10:1994(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 9022-10 was prepared by Technical Committee ISO/TC 172, *Optics and optical instruments*, Subcommittee SC 1, *Fundamental standards*.

ISO 9022 consists of the following parts, under the general title *Optics and optical instruments — Environmental test methods*:

- Part 1: *Definitions, extent of testing*
- Part 2: *Cold, heat, humidity*
- Part 3: *Mechanical stress*
- Part 4: *Salt mist*
- Part 5: *Combined cold, low air pressure*
- Part 6: *Dust*
- Part 7: *Drip, rain*
- Part 8: *High pressure, low pressure, immersion*
- Part 9: *Solar radiation*
- Part 10: *Combined sinusoidal vibration, dry heat or cold*

© ISO 1994

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization
Case Postale 56 • CH-1211 Genève 20 • Switzerland

Printed in Switzerland

- Part 11: Mould growth
- Part 12: Contamination
- Part 13: Combined shock, bump or free fall, dry heat or cold
- Part 14: Dew, hoarfrost, ice
- Part 15: Combined random vibration wide band: reproducibility medium, in dry heat or cold
- Part 16: Combined bounce or steady-state acceleration, in dry heat or cold
- Part 17: Combined contamination, solar radiation
- Part 18: Combined damp heat and low internal pressure
- Part 19: Temperature cycles combined with sinusoidal or random vibration
- Part 20: Humid atmosphere containing sulfur dioxide or hydrogen sulfide

STANDARDSISO.COM : Click to view the full PDF of ISO 9022-10:1994

Introduction

Optical instruments are affected during their use by a number of different environmental parameters which they are required to resist without significant reduction in performance.

The type and severity of these parameters depend on the conditions of use of the instrument (for example, in the laboratory or workshop) and on its geographical location. The environmental effects on optical instrument performance in the tropics and subtropics are totally different from those found when they are used in the arctic regions. Individual parameters cause a variety of different and overlapping effects on instrument performance.

The manufacturer attempts to ensure, and the user naturally expects, that instruments will resist the likely rigours of their environment throughout their life. This expectation can be assessed by exposure of the instrument to a range of simulated environmental parameters under controlled laboratory conditions. The severity of these conditions is often increased to obtain meaningful results in a relatively short period of time.

In order to allow assessment and comparison of the response of optical instruments to appropriate environmental conditions, ISO 9022 contains details of a number of laboratory tests which reliably simulate a variety of different environments. The tests are based largely on IEC standards, modified where necessary to take into account features special to optical instruments.

It should be noted that, as a result of continuous progress in all fields, optical instruments are no longer only precision-engineered optical products, but, depending on their range of application, also contain additional assemblies from other fields. For this reason, the principal function of the instrument must be assessed to determine which International Standard should be used for testing. If the optical function is of primary importance, then ISO 9022 is applicable, but if other functions take precedence then the appropriate International Standard in the field concerned should be applied. Cases may arise where application of both ISO 9022 and other appropriate International Standards will be necessary.

Optics and optical instruments — Environmental test methods —

Part 10:

Combined sinusoidal vibration, dry heat or cold

1 Scope

This part of ISO 9022 specifies methods for the testing of optical instruments and instruments containing optical components under equivalent conditions, for their ability to resist combined sinusoidal vibration, in dry heat or cold.

The purpose of testing is to investigate to what extent the optical, thermal, chemical and electrical performance characteristics of the specimen are affected by combined sinusoidal vibration, in dry heat or cold.

2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this part of ISO 9022. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this part of ISO 9022 are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 9022-1:1994¹⁾, *Optics and optical instruments — Environmental test methods — Part 1: Definitions, extent of testing.*

ISO 9022-2:1994, *Optics and optical instruments — Environmental test methods — Part 2: Cold, heat, humidity.*

ISO 9022-3:1994¹⁾, *Optics and optical instruments — Environmental test methods — Part 3: Mechanical stress.*

IEC 68-2-47:1982 *Environmental testing — Part 2: Tests — Mounting of components, equipment and other articles for dynamic tests including shock (Ea), bump (Eb), vibration (Fc and Fd) and steady-state acceleration (Ga) and guidance.*

IEC 68-2-50:1983 *Environmental testing — Part 2: Tests — Tests Z|AFc: Combined cold|vibration (sinusoidal) tests for both heat-dissipating and non-heat-dissipating specimens.*

IEC 68-2-51:1983 *Environmental testing — Part 2: Tests — Tests Z|BFc: Combined dry heat|vibration (sinusoidal) tests for both heat-dissipating and non-heat-dissipating specimens.*

IEC 68-2-53:1984 *Environmental testing — Part 2: Tests — Guidance to tests Z|AFc and Z|BFc: Combined temperature (cold and dry heat) and vibration (sinusoidal) tests.*

3 General information and test conditions

Exposure of the specimen to combined stress conditions renders the test much more severe than separate exposure to any one of the environmental conditions alone.

1) To be published.

The values of temperature specified in tables 1 and 3 are selected from ISO 9022-2, conditioning methods 10 and 11.

The tests shall be conducted in accordance with the requirements of ISO 9022-3.

The fixture for the specimen shall meet the requirements of IEC 68-2-47 and shall be thermally insulated, if appropriate.

In the event that the specimen is mounted on shock absorbers, time shall be allowed for temperature stabilization of the absorber elements.

4 Conditioning

The required exposure time shall not commence until all parts of the specimen have reached a temperature within 3 K of the test chamber temperature.

For heat-dissipating specimens, the period of exposure shall not begin until the temperature of the specimens changes not more than 1 K within 1 h at the stabilized test chamber temperature.

The last hour of the temperature soaking time shall be considered to be the first hour of the exposure period.

4.1 Conditioning method 61: Combined sinusoidal vibration, dry heat

See tables 1 and 2.

4.2 Conditioning method 62: Combined sinusoidal vibration, cold

See tables 3 and 4.

Table 1 — Degrees of severity for conditioning method 61: Combined sinusoidal vibration, dry heat applying frequency sweep

Degree of severity		01	02	03	04	05	06	07	08	09	10	11	12	13
Test chamber temperature	°C	40 ± 2			55 ± 2					63 ± 2				
Relative humidity	%	< 40												
Displacement	mm	0,035	0,15	0,075	0,15	0,15	0,15	0,35	0,35	0,15	0,15	0,35	0,35	1,0
Acceleration	m/s ²	4,9	19,6	9,8	19,6	—	19,6	49	49	19,6	19,6	49	49	—
	g multiples	0,5	2	1	2	—	2	5	5	2	2	5	5	—
Number of frequency cycles ¹⁾ for each axis, for frequency bands	10 Hz to 55 Hz	—	—	—	—	5	—	—	—	—	—	—	—	20
	10 Hz to 150 Hz	—	20	—	—	—	—	5	—	20	—	5	—	—
	10 Hz to 500 Hz	2	—	—	10	—	—	—	—	—	—	—	—	—
	10 Hz to 2 000 Hz	—	—	2	—	—	10	—	10	—	10	—	10	—
State of operation		0 or 1 or 2												

1) The sweep rate for the specified number of frequency cycles shall be one octave per minute.

Table 2 — Degrees of severity for conditioning method 61: Combined sinusoidal vibration, dry heat applying characteristic frequencies

Degree of severity		21	22	23	24	25	26	27	28	29	30	31	32	33	34	35	36	37	38	39	40	41	42	43	44	45	46
Degree of severity as in table 1		01	01	02	02	03	03	04	04	05	05	06	06	07	07	08	08	09	09	10	10	11	11	12	12	13	13
Exposure time of characteristic frequency	min	10	30	10	30	10	30	10	30	10	30	10	30	10	30	10	30	10	30	10	30	10	30	10	30	10	30
State of operation		0 or 1 or 2																									

Table 3 — Degrees of severity for conditioning method 62: Combined sinusoidal vibration, cold, applying frequency sweep

Degree of severity	01	02	03	04	05	06	07	08	09	10	11	12	13	14	15	16	17	
Test chamber temperature °C	- 10 ± 3		- 20 ± 3		- 25 ± 3				- 35 ± 3			- 55 ± 3			- 65 ± 3			
Relative humidity %	< 40																	
Displacement mm	0,035	0,15	0,035	0,15	0,075	0,15	0,15	0,35	0,075	0,15	0,15	0,35	0,15	0,15	0,35	1,0	0,15	
Acceleration	m/s ²	4,9	19,6	4,9	19,6	9,8	19,6	—	49	9,8	19,6	19,6	49	19,6	19,6	49	—	19,6
	g multiples	0,5	2	0,5	2	1	2	—	5	1	2	2	5	2	2	5	—	2
Number of frequency cycle ¹⁾ for each axis, for frequency band	10 Hz to 55 Hz	—	—	—	—	—	—	5	—	—	—	—	—	—	—	—	20	—
	10 Hz to 150 Hz	—	20	—	20	—	20	—	5	—	—	—	—	20	—	—	—	—
	10 Hz to 500 Hz	2	—	2	—	—	—	—	—	—	10	—	—	—	—	—	—	—
	10 Hz to 2 000 Hz	—	—	—	—	2	—	—	—	2	—	10	10	—	—	10	—	10
State of operation	0 or 1 or 2																	
1) The sweep rate for the specified number of frequency cycles shall be one octave per minute.																		

Table 4 — Degrees of severity for conditioning method 62: Combined sinusoidal vibration, cold, applying characteristic frequencies

Degree of severity	21	22	23	24	25	26	27	28	29	30	31	32	33	34	35	36	37
Degree of severity as in table 3	01	01	02	02	03	03	04	04	05	05	06	06	07	07	08	08	09
Exposure time at characteristic frequency min	10	30	10	30	10	30	10	30	10	30	10	30	10	30	10	30	10
State of operation	0 or 1 or 2																
Degree of severity	38	39	40	41	42	43	44	45	46	47	48	49	50	51	52	53	54
Degree of severity as in table 3	09	10	10	11	11	12	12	13	13	14	14	15	15	16	16	17	17
Exposure time at characteristic frequency min	30	10	30	10	30	10	30	10	30	10	30	10	30	10	30	10	30
State of operation	0 or 1 or 2																