



INTERNATIONAL STANDARD ISO 9873:1998
TECHNICAL CORRIGENDUM 1

Published 2000-06-15

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Dental hand instruments — Reusable mirrors and handles

TECHNICAL CORRIGENDUM 1

Instruments dentaires à main — Miroirs et manches de miroirs réutilisables

RECTIFICATIF TECHNIQUE 1

Technical Corrigendum 1 to International Standard ISO 9873:1998 was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 4, *Dental instruments*.

Page 7, subclause 5.8.1

Add the phrase "...with the casing." at the end of the second sentence.

Add the phrase "...with the side of the slot..." to the third sentence so that it reads:

"The depth of the slot shall be sufficient to ensure that the mirror stem does not come into contact with the side of the slot during the test."

Page 8, subclause 5.8.2

Delete the entire paragraph and replace it with the following text:

"Screw the casing and stem assembly C into the handle and clamp firmly in position using components A and B. Apply a moment of force of 1,25 N·m at the weld, in both directions, at 90° to the bottom of the mirror casing (see figure 4)."