



INTERNATIONAL STANDARD ISO 9951:1993
TECHNICAL CORRIGENDUM 1

Published 1994-11-15

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Measurement of gas flow in closed conduits — Turbine meters

TECHNICAL CORRIGENDUM 1

Mesure de débit de gaz dans les conduites fermées — Compteurs à turbine

RECTIFICATIF TECHNIQUE 1

Technical corrigendum 1 to International Standard ISO 9951:1993 was prepared by Technical Committee ISO/TC 30, *Measurement of fluid flow in closed conduits*, Subcommittee SC 10, *Turbine meters*.

Page 12

Subclause E.4.2

In the third line, delete "two elbows of diameter equal to D_1 ." and replace with "two elbows of radius of curvature equal to D_1 ".

UDC 532.575.57:681.121.4

Ref. No. ISO 9951:1993/Cor.1:1994(E)

Descriptors: gas flow, pipe flow, flow measurement, flowmeters, turbine meters, specifications, performance evaluation, dimensions, tests, pressure tests, calibration, inspection, marking.

© ISO 1994

Printed in Switzerland