



**INTERNATIONAL STANDARD ISO/IEEE 11073-20601:2016
TECHNICAL CORRIGENDUM 1**

Published 2016-06-15



INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION
INTERNATIONAL ELECTROTECHNICAL COMMISSION • МЕЖДУНАРОДНАЯ ЭЛЕКТРОТЕХНИЧЕСКАЯ КОМИССИЯ • COMMISSION ÉLECTROTECHNIQUE INTERNATIONALE

Health informatics — Personal health device communication
Part 20601:
Application profile — Optimized exchange protocol

TECHNICAL CORRIGENDUM 1

*Informatique de santé — Communication entre dispositifs de santé personnels —
Partie 20601: Profil d'application — Protocole d'échange optimisé*

RECTIFICATIF TECHNIQUE 1

Technical Corrigendum 1 to ISO/IEEE 11073-20601 was prepared by the IEEE 11073 Standards Committee of the IEEE Engineering in Medicine and Biology Society (as IEEE Std 11073-20601-2014/Cor 1:2015). It was adopted by Technical Committee ISO/TC 215, *Health informatics*, in parallel with its approval by the ISO member bodies, under the “fast-track procedure” defined in the Partner Standards Development Organization cooperation agreement between ISO and IEEE. IEEE is responsible for the maintenance of this document with participation and input from ISO member bodies.

STANDARDSISO.COM : Click to view the full PDF of ISO/IEEE 11073-20601:2016/Cor 1:2016

Health informatics—Personal health device communication

Part 20601: Application profile— Optimized Exchange Protocol

Corrigendum 1

Sponsor

IEEE 11073™ Standards Committee
of the
IEEE Engineering in Medicine and Biology Society

Approved 3 September 2015

IEEE-SA Standards Board

STANDARDSISO.COM : Click to view the full PDF of ISO/IEEE 11073-20601:2016/Cor 1:2016

Abstract: Within the context of the ISO/IEEE 11073 family of standards for device communication, this standard defines a common framework for making an abstract model of personal health data available in transport-independent transfer syntax required to establish logical connections between systems and to provide presentation capabilities and services needed to perform communication tasks. The protocol is optimized to personal health usage requirements and leverages commonly used methods and tools wherever possible.

This corrigendum removes the ambiguities and corrects the wrong nomenclature codes and qualifier status that have been identified in IEEE Std 11073-20601-2014 to improve implementation of the standard in an interoperable fashion.

Keywords: IEEE 11073™, IEEE 11073-20601™, medical device communication, personal health devices

The Institute of Electrical and Electronics Engineers, Inc.
3 Park Avenue, New York, NY 10016-5997, USA

Copyright © 2015 by The Institute of Electrical and Electronics Engineers, Inc.
All rights reserved. Published 27 October 2015. Printed in the United States of America.

IEEE is a registered trademark in the U.S. Patent & Trademark Office, owned by The Institute of Electrical and Electronics Engineers, Incorporated.

PDF: 978-0-7381-9995-5 STD20427
Print: 978-0-7381-9996-2 STDPD20427

IEEE prohibits discrimination, harassment, and bullying.

For more information, visit <http://www.ieee.org/web/aboutus/whatis/policies/p9-26.html>.

No part of this publication may be reproduced in any form, in an electronic retrieval system or otherwise, without the prior written permission of the publisher.

Important Notices and Disclaimers Concerning IEEE Standards Documents

IEEE documents are made available for use subject to important notices and legal disclaimers. These notices and disclaimers, or a reference to this page, appear in all standards and may be found under the heading “Important Notice” or “Important Notices and Disclaimers Concerning IEEE Standards Documents.”

Notice and Disclaimer of Liability Concerning the Use of IEEE Standards Documents

IEEE Standards documents (standards, recommended practices, and guides), both full-use and trial-use, are developed within IEEE Societies and the Standards Coordinating Committees of the IEEE Standards Association (“IEEE-SA”) Standards Board. IEEE (“the Institute”) develops its standards through a consensus development process, approved by the American National Standards Institute (“ANSI”), which brings together volunteers representing varied viewpoints and interests to achieve the final product. Volunteers are not necessarily members of the Institute and participate without compensation from IEEE. While IEEE administers the process and establishes rules to promote fairness in the consensus development process, IEEE does not independently evaluate, test, or verify the accuracy of any of the information or the soundness of any judgments contained in its standards.

IEEE does not warrant or represent the accuracy or content of the material contained in its standards, and expressly disclaims all warranties (express, implied and statutory) not included in this or any other document relating to the standard, including, but not limited to, the warranties of: merchantability; fitness for a particular purpose; non-infringement; and quality, accuracy, effectiveness, currency, or completeness of material. In addition, IEEE disclaims any and all conditions relating to: results; and workmanlike effort. IEEE standards documents are supplied “AS IS” and “WITH ALL FAULTS.”

Use of an IEEE standard is wholly voluntary. The existence of an IEEE standard does not imply that there are no other ways to produce, test, measure, purchase, market, or provide other goods and services related to the scope of the IEEE standard. Furthermore, the viewpoint expressed at the time a standard is approved and issued is subject to change brought about through developments in the state of the art and comments received from users of the standard.

In publishing and making its standards available, IEEE is not suggesting or rendering professional or other services for, or on behalf of, any person or entity nor is IEEE undertaking to perform any duty owed by any other person or entity to another. Any person utilizing any IEEE Standards document, should rely upon his or her own independent judgment in the exercise of reasonable care in any given circumstances or, as appropriate, seek the advice of a competent professional in determining the appropriateness of a given IEEE standard.

IN NO EVENT SHALL IEEE BE LIABLE FOR ANY DIRECT, INDIRECT, INCIDENTAL, SPECIAL, EXEMPLARY, OR CONSEQUENTIAL DAMAGES (INCLUDING, BUT NOT LIMITED TO: PROCUREMENT OF SUBSTITUTE GOODS OR SERVICES; LOSS OF USE, DATA, OR PROFITS; OR BUSINESS INTERRUPTION) HOWEVER CAUSED AND ON ANY THEORY OF LIABILITY, WHETHER IN CONTRACT, STRICT LIABILITY, OR TORT (INCLUDING NEGLIGENCE OR OTHERWISE) ARISING IN ANY WAY OUT OF THE PUBLICATION, USE OF, OR RELIANCE UPON ANY STANDARD, EVEN IF ADVISED OF THE POSSIBILITY OF SUCH DAMAGE AND REGARDLESS OF WHETHER SUCH DAMAGE WAS FORESEEABLE.

Translations

The IEEE consensus development process involves the review of documents in English only. In the event that an IEEE standard is translated, only the English version published by IEEE should be considered the approved IEEE standard.

Official statements

A statement, written or oral, that is not processed in accordance with the IEEE-SA Standards Board Operations Manual shall not be considered or inferred to be the official position of IEEE or any of its committees and shall not be considered to be, or be relied upon as, a formal position of IEEE. At lectures, symposia, seminars, or educational courses, an individual presenting information on IEEE standards shall make it clear that his or her views should be considered the personal views of that individual rather than the formal position of IEEE.

Comments on standards

Comments for revision of IEEE Standards documents are welcome from any interested party, regardless of membership affiliation with IEEE. However, IEEE does not provide consulting information or advice pertaining to IEEE Standards documents. Suggestions for changes in documents should be in the form of a proposed change of text, together with appropriate supporting comments. Since IEEE standards represent a consensus of concerned interests, it is important that any responses to comments and questions also receive the concurrence of a balance of interests. For this reason, IEEE and the members of its societies and Standards Coordinating Committees are not able to provide an instant response to comments or questions except in those cases where the matter has previously been addressed. For the same reason, IEEE does not respond to interpretation requests. Any person who would like to participate in revisions to an IEEE standard is welcome to join the relevant IEEE working group.

Comments on standards should be submitted to the following address:

Secretary, IEEE-SA Standards Board
445 Hoes Lane
Piscataway, NJ 08854 USA

Laws and regulations

Users of IEEE Standards documents should consult all applicable laws and regulations. Compliance with the provisions of any IEEE Standards document does not imply compliance to any applicable regulatory requirements. Implementers of the standard are responsible for observing or referring to the applicable regulatory requirements. IEEE does not, by the publication of its standards, intend to urge action that is not in compliance with applicable laws, and these documents may not be construed as doing so.

Copyrights

IEEE draft and approved standards are copyrighted by IEEE under U.S. and international copyright laws. They are made available by IEEE and are adopted for a wide variety of both public and private uses. These include both use, by reference, in laws and regulations, and use in private self-regulation, standardization, and the promotion of engineering practices and methods. By making these documents available for use and adoption by public authorities and private users, IEEE does not waive any rights in copyright to the documents.

Photocopies

Subject to payment of the appropriate fee, IEEE will grant users a limited, non-exclusive license to photocopy portions of any individual standard for company or organizational internal use or individual, non-commercial use only. To arrange for payment of licensing fees, please contact Copyright Clearance Center, Customer Service, 222 Rosewood Drive, Danvers, MA 01923 USA; +1 978 750 8400. Permission to photocopy portions of any individual standard for educational classroom use can also be obtained through the Copyright Clearance Center.

Updating of IEEE Standards documents

Users of IEEE Standards documents should be aware that these documents may be superseded at any time by the issuance of new editions or may be amended from time to time through the issuance of amendments, corrigenda, or errata. An official IEEE document at any point in time consists of the current edition of the document together with any amendments, corrigenda, or errata then in effect.

Every IEEE standard is subjected to review at least every ten years. When a document is more than ten years old and has not undergone a revision process, it is reasonable to conclude that its contents, although still of some value, do not wholly reflect the present state of the art. Users are cautioned to check to determine that they have the latest edition of any IEEE standard.

In order to determine whether a given document is the current edition and whether it has been amended through the issuance of amendments, corrigenda, or errata, visit the IEEE-SA Website at <http://ieeexplore.ieee.org/xpl/standards.jsp> or contact IEEE at the address listed previously. For more information about the IEEE-SA or IEEE's standards development process, visit the IEEE-SA Website at <http://standards.ieee.org>.

Errata

Errata, if any, for all IEEE standards can be accessed on the IEEE-SA Website at the following URL: <http://standards.ieee.org/findstds/errata/index.html>. Users are encouraged to check this URL for errata periodically.

Patents

Attention is called to the possibility that implementation of this standard may require use of subject matter covered by patent rights. By publication of this standard, no position is taken by the IEEE with respect to the existence or validity of any patent rights in connection therewith. If a patent holder or patent applicant has filed a statement of assurance via an Accepted Letter of Assurance, then the statement is listed on the IEEE-SA Website at <http://standards.ieee.org/about/sasb/patcom/patents.html>. Letters of Assurance may indicate whether the Submitter is willing or unwilling to grant licenses under patent rights without compensation or under reasonable rates, with reasonable terms and conditions that are demonstrably free of any unfair discrimination to applicants desiring to obtain such licenses.

Essential Patent Claims may exist for which a Letter of Assurance has not been received. The IEEE is not responsible for identifying Essential Patent Claims for which a license may be required, for conducting inquiries into the legal validity or scope of Patents Claims, or determining whether any licensing terms or conditions provided in connection with submission of a Letter of Assurance, if any, or in any licensing agreements are reasonable or non-discriminatory. Users of this standard are expressly advised that determination of the validity of any patent rights, and the risk of infringement of such rights, is entirely their own responsibility. Further information may be obtained from the IEEE Standards Association.

Participants

At the time this IEEE Standard was completed, the Personal Health Devices Working Group had the following membership:

Daidi Zhong, *Chair*
Michael J. Kirwan, *Chair*

Karsten Aalders
 Charles R. Abbruscato
 Nabil Abujbara
 Maher Abuzaid
 James Agnew
 Haidar Ahmad
 Manfred Aigner
 Jorge Alberola
 Murtaza Ali
 Rolf Ambuehl
 David Aparisi
 Lawrence Arne
 Diego B. Arquillo
 Serafin Arroyo
 Muhammad Asim
 Merat Bagha
 Doug Baird
 David Baker
 Anindya Bakshi
 Ananth Balasubramanian
 Sunlee Bang
 M. Jonathan Barkley
 Gilberto Barrón
 David Bean
 John Bell
 Rudy Belliardi
 Daniel Bernstein
 George A. Bertos
 Chris Biernacki
 Ola Björnsne
 Thomas Blackadar
 Marc Blanchet
 Thomas Bluethner
 Douglas P. Bogia
 Xavier Boniface
 Shannon Boucousis
 Julius Broma
 Lyle G. Bullock, Jr.
 Bernard Burg
 Chris Burns
 Anthony Butt
 Jeremy Byford-Rew
 Satya Calloji
 Carole C. Carey
 Craig Carlson
 Santiago Carot-Nemesio
 Randy W. Carroll
 Simon Carter
 Seungchul Chae
 Rahul Chauhan
 James Cheng
 Peggy Chien

David Chiu
 Chia-Chin Chong
 Saeed A. Choudhary
 Jinhan Chung
 Malcolm Clarke
 John A. Cogan
 John T. Collins
 Cory Condek
 Todd H. Cooper
 David Cornejo
 Douglas Coup
 Nigel Cox
 Hans Crommenacker
 Tomio Crosley
 David Culp
 Allen Curtis
 Ndifor Cyril Fru
 Jesús Daniel Trigo
 Eyal Dassau
 David Davenport
 Russell Davis
 Sushil K. Deka
 Ciro de la Vega
 Pedro de las-Heras-Quiros
 Jim DelloStritto
 Matthew d'Entremont
 Lane Desborough
 Kent Dicks
 Hyoungho Do
 Xiaolian Duan
 Brian Dubreuil
 Sourav Dutta
 Jakob Ehrensvarð
 Fredrik Einberg
 Roger M. Ellingson
 Michihiro Enokida
 Javier Escayola Calvo
 Mark Estes
 Leonardo Estevez
 Roger Feeley
 Bosco T. Fernandes
 Christoph Fischer
 Morten Flintrup
 Joseph W. Forler
 Russell Foster
 Eric Freudenthal
 Matthias Frohner
 Ken Fuchs
 Jing Gao
 Xuemei Gao
 Marcus Garbe
 John Garguilo

Rick Geimer
 Igor Gejdos
 Ferenc Gerbovics
 Nicolae Goga
 Julian Goldman
 Raul Gonzalez Gomez
 Chris Gough
 Channa Gowda
 Charles M. Gropper
 Amit Gupta
 Jeff Guttmacher
 Rasmus Haahr
 Christian Habermann
 Michael Hagerty
 Jerry Hahn
 Robert Hall
 Nathaniel Hamming
 Rickey L. Hampton
 Sten Hanke
 Jordan Hartmann
 Kai Hassing
 Marc Daniel Haunschild
 Wolfgang Heck
 Nathaniel Heintzman
 Charles Henderson
 Jun-Ho Her
 Takashi Hibino
 Timothy L. Hirou
 Allen Hobbs
 Alex Holland
 Arto Holopainen
 Kris Holtzclaw
 Robert Hoy
 Frank Hsu
 Anne Huang
 Sen-Der Huang
 Zhiqiang Huang
 Ron Huby
 David Hughes
 Robert D. Hughes
 Jiyoung Huh
 Hugh Hunter
 Hitoshi Ikeda
 Yutaka Ikeda
 Philip O. Isaacson
 Atsushi Ito
 Michael Jaffe
 Praduman Jain
 Wei Jin
 Danny Jochelson
 Chris Johnson
 Phaneeth Junga

Akiyoshi Kabe	Piotr Murawski	Sid Shaw
Steve Kahle	Soundharya Nagasubramanian	Frank Shen
Tomio Kamioka	Jae-Wook Nah	Liquan Shen
Kei Kariya	Alex Neefus	Bozhi Shi
Andy Kaschl	Trong-Nghia Nguyen-Dobinsky	Min Shih
Junzo Kashihara	Michael E. Nidd	Mazen Shihabi
Kohichi Kashiwagi	Tetsu Nishimura	Redmond Shouldice
Ralph Kent	Jim Niswander	Sternly K. Simon
Laurie M. Kermes	Hiroaki Niwamoto	Marjorie Skubic
Ikuo Keshi	Thomas Norgall	Robert Smith
Junhyung Kim	Anand Noubade	Ivan Soh
Minho Kim	Yoshiteru Nozoe	Motoki Sone
Min-Joon Kim	Abraham Ofek	Emily Sopensky
Taekon Kim	Brett Olive	Rajagopalan Srinivasan
Tetsuya Kimura	Begonya Otal	Andreas Staubert
Alfred Kloos	Charles Palmer	Nicholas Steblay
Jeongmee Koh	Bud Panjwani	Beth Stephen
Jean-Marc Koller	Carl Pantiskas	Lars Steubesand
John Koon	Harry P. Pappas	John (Ivo) Stivoric
Patty Krantz	Mikey Paradis	Raymond A. Strickland
Raymond Krasinski	Hanna Park	Chandrasekaran Subramaniam
Alexander Kraus	Jong-Tae Park	Hermanni Suominen
Ramesh Krishna	Myungeun Park	Lee Surprenant
Geoffrey Kruse	Soojun Park	Ravi Swami
Falko Kuester	Phillip E. Pash	Ray Sweidan
Rafael Lajara	TongBi Pei	Jin Tan
Pierre Landau	Lucian Pestritu	Haruyuyuki Tatsumi
Jaechul Lee	Soren Petersen	John W. Thomas
JongMuk Lee	James Petisce	Jonas Tirén
Kyong Ho Lee	Peter Piction	Alexandra Todiruta
Rami Lee	Michael Pliskin	James Tomcik
Sungkee Lee	Jeff Price	Janet Traub
Woojae Lee	Harald Prinzhorn	Gary Tschautscher
Yonghee Lee	John Quinlan	Masato Tsuchid
Joe Lenart	Arif Rahman	Ken Tubman
Kathryn A. Lesh	Tanzilur Rahman	Yoshihiro Uchida
Qiong Li	Steve Ray	Sunil Unadkat
Ying Li	Phillip Raymond	Fabio Urbani
Patrick Lichter	Tim Reilly	Philipp Urbauer
Jisoon Lim	Barry Reinhold	Laura Vanzago
Joon-Ho Lim	Brian Reinhold	Alpo Värri
John Lin	Melvin I. Reynolds	Dalimar Velez
Wei-Jung Lo	John G. Rhoads	Naveen Verma
Charles Lowe	Jeffrey S. Robbins	Rudi Voon
Don Ludolph	Moskowitz Robert	Isobel Walker
Christian Luszick	Timothy Robertson	David Wang
Bob MacWilliams	David Rosales	Jerry P. Wang
Srikanth Madhurbotheswaran	Bill Saltzstein	Yao Wang
Miriam L. Makhlouf	Benedikt Salzbrunn	Yi Wang
Romain Marmot	Giovanna Sannino	Steve Warren
Sandra Martinez	Jose A. Santos-Cadenas	Fujio Watanabe
Miguel Martínez de Espronceda	Stefan Sauermann	Toru Watsuji
Cámara	John Sawyer	Mike Weng
Peter Mayhew	Guillaume Schatz	Kathleen Wible
Jim McCain	Alois Schloegl	Paul Williamson
László Meleg	Paul S. Schluter	Jan Wittenber
Alexander Mense	Lars Schmitt	Jia-Rong Wu
Ethan Metsger	Mark G. Schnell	Will Wykeham
Jinsei Miyazaki	Richard A. Schrenker	Ariton Xhafa
Erik Moll	Antonio Scorpiniti	Yaxi Yan
Darr Moore	Kwang Seok Seo	Ricky Yang
Carsten Mueglitz	Riccardo Serafin	Melanie S. Yeung

Done-Sik Yoo
Jianchao Zeng

Jason Zhang
Zhiqiang Zhang
Thomas Zhao

Miha Zoubek
Szymon Zyskoter

The following members of the individual balloting committee voted on this standard. Balloters may have voted for approval, disapproval, or abstention.

Thomas Blackadar
Susan Burgess
Craig Carlson
Keith Chow
Sourav Dutta
Christoph Fischer
Randall Groves
Werner Hoelzl

Noriyuki Ikeuchi
Atsushi Ito
Piotr Karocki
Michael Kirwan
H. Moll
Bartien Sayogo
Stefan Schlichting

Lars Schmitt
Janek Schumann
Eugene Stoudenmire
Raymond Strickland
Walter Struppler
Jan Wittenber
Oren Yuen
Daidi Zhong

When the IEEE-SA Standards Board approved this standard on 3 September 2015, it had the following membership:

John Kulick, *Chair*
Jon Walter Rosdahl, *Vice Chair*
Richard H. Hulett, *Past Chair*
Konstantinos Karachalios, *Secretary*

Masayuki Ariyoshi
Ted Burse
Stephen Dukes
Jean-Philippe Faure
J. Travis Griffith
Gary Hoffman
Michael Janezic

Joseph L. Koepfinger*
David J. Law
Hung Ling
Andrew Myles
T. W. Olsen
Glenn Parsons
Ronald C. Petersen
Annette D. Reilly

Stephen J. Shellhammer
Adrian P. Stephens
Yatin Trivedi
Phillip Winston
Don Wright
Yu Yuan
Daidi Zhong

*Member Emeritus

STANDARDSISO.COM : Click to view the full PDF of ISO/IEEE 11073-20601:2016/Cor 1:2016

Introduction

This introduction is not part of IEEE Std 11073-20601™-2014/Cor 1-2015, Health informatics—Personal health device communication—Part 20601: Application profile—Optimized Exchange Protocol—Corrigendum 1.

ISO and IEEE 11073 standards enable communication between medical devices and external computer systems. This standard and corresponding IEEE 11073-104xx standards address a need for a simplified and optimized communication approach for personal health devices, which may or may not be regulated devices. These standards align with, and draw upon, the existing clinically focused standards to provide easy management of data from either a clinical or personal health device.

This document addresses a need for an openly defined, independent standard for converting the collected information into an interoperable transmission format so the information can be exchanged between agents and managers.

This standard removes the ambiguities and corrects the wrong nomenclature codes and qualifier status that have been identified in IEEE Std 11073-20601-2014 to improve implementation of the standard in an interoperable fashion.

STANDARDSISO.COM : Click to view the full PDF of ISO/IEEE 11073-20601:2016/Cor 1:2016

Contents

6. Personal health device DIM	2
6.3 Personal health object class definitions	2
7. Personal health device service model	4
7.4 Specific application of object access EVENT REPORT services for personal health devices.....	4
8. Communication model	5
8.7 Associating procedure	5
8.9 Operating procedure	5
Annex A (normative) ASN.1 definitions	7
A.4 ACTION-method-related data types	7
A.11 Data types for new object attributes and object services	7
Annex I (normative) Nomenclature codes.....	9

STANDARDSISO.COM : Click to view the full PDF of ISO/IEEE 11073-20601:2016/Cor 1:2016

Health informatics—Personal health device communication

Part 20601: Application profile— Optimized Exchange Protocol

Corrigendum 1

IMPORTANT NOTICE: *IEEE Standards documents are not intended to ensure safety, security, health, or environmental protection, or ensure against interference with or from other devices or networks. Implementers of IEEE Standards documents are responsible for determining and complying with all appropriate safety, security, environmental, health, and interference protection practices and all applicable laws and regulations.*

This IEEE document is made available for use subject to important notices and legal disclaimers. These notices and disclaimers appear in all publications containing this document and may be found under the heading “Important Notice” or “Important Notices and Disclaimers Concerning IEEE Documents.” They can also be obtained on request from IEEE or viewed at <http://standards.ieee.org/IPR/disclaimers.html>.

NOTE—The editing instructions contained in this corrigendum define how to merge the material contained therein into the existing base standard and its amendments to form the comprehensive standard.

The editing instructions are shown in ***bold italic***. Four editing instructions are used: change, delete, insert, and replace. ***Change*** is used to make corrections in existing text or tables. The editing instruction specifies the location of the change and describes what is being changed by using ~~strike through~~ (to remove old material) and underscore (to add new material). ***Delete*** removes existing material. ***Insert*** adds new material without disturbing the existing material. Insertions may require renumbering. If so, renumbering instructions are given in the editing instruction. ***Replace*** is used to make changes in figures or equations by removing the existing figure or equation and replacing it with a new one. Editing instructions, change markings, and this NOTE will not be carried over into future editions because the changes will be incorporated into the base standard.

6. Personal health device DIM

6.3 Personal health object class definitions

Change Source-Handle-Reference and Source-Handle-Reference-List rows in Table 6 as shown:

Table 6—Metric Attributes

Attribute name	Attribute ID	Attribute type	Remark	Qualifiers
Source-Handle-Reference	MDC_ATTR_SOURCE_HANDLE_REF	HANDLE	This attribute establishes a relation of this object instance to a source object (e.g., pulse references sourcing SpO2). This attribute is used whenever it is required to model an explicit relation between object instances to define dependencies. The usage of this attribute is defined by device specializations. A metric object may contain one of Source-Handle-Reference or Source-Handle-Reference-List, but not both. <u>If a measurement reports a Source-Handle-Reference or Source-Handle-Reference-List attribute, the measurement from the object(s) that it references shall be sent by the agent to the manager must have been sent prior to the sending of this measurement.</u>	Optional Dynamic
Source-Handle-Reference-List	MDC_ATTR_SOURCE_HANDLE_REF_LIST	HANDLE List	This attribute establishes a relation of this object instance to more than one source objects (e.g., body mass index (BMI) references sourcing Height and Weight). This attribute is used whenever it is required to model an explicit relation between object instances to define dependencies. The usage of this attribute is defined by device specializations. A metric object may contain one of Source-Handle-Reference or Source-Handle-Reference-List, but not both. <u>If a measurement reports a Source-Handle-Reference or Source-Handle-Reference-List attribute, the measurement from the object(s) that it references shall be sent by the agent to the manager must have been sent prior to the sending of this measurement.</u>	Optional Dynamic

6.3.7 PM-store class

6.3.7.3 PM-store class attributes

Delete the following text:

The attributes Handle and PM-Store-Capab are part of the agent configuration; therefore, the manager knows the corresponding attribute values after the Configuring procedure.

6.3.7.4 PM-store object methods

Change the first paragraph after Table 11 as shown:

If an agent supports the PM-store class, the support of the ~~Get-Segment-Info or Get-Segment-Id-List~~ ~~methods is mandatory~~, and support of the Trig-Segment-Data-Xfer method is mandatory. Support for the Clear-Segments and Get-Segment-Id-List method is optional and is indicated in the PM-Store-Capab attribute.

Delete text in Get-Segment-Info as shown:

— **Get-Segment-Info:**

This method allows the manager to retrieve PM-segment attributes of one or more PM-segments, with the exception of the Fixed-Segment-Data attribute, which contains the actual stored data and is retrieved by using the Trig-Segment-Data-Xfer method. In particular, the Get-Segment-Info method allows the manager to retrieve the attributes and their data contents from the PM-segment object instances identified by the SegmSelection parameter.

The agent shall support the all-segments choice in the SegmSelection action-info-args of the Get-Segment-Info method. The agent may support the segm-id-list and/or abs-time-range and/or bo-time-range choice in the SegmSelection action-info-args of the Get-Segment-Info method. In this case the agent shall set the pmsc-segm-id-list-select and/or pmsc-abs-time-select flag in the PM-Store-Capab attribute. If the manager sends the Get-Segment-Info method with the choice that the agent does not support, the agent shall reply with an unsupported-choice error (roer).

For PM-segment info returned by time, the segments are selected using the same mechanism as described under Clear-Segments.

If the manager supports sending the Get-Segment-Info method, the manager shall support at least the choice all-segments in the SegmSelection action-info-args of the Get-Segment-Info method. The manager may support additional choices.

If a standard configuration contains any PM-Store object, the manager must send Get-Segment-info or Get-Segment-Id-List at the beginning of accessing any PM-Store object.

If no PM-Segment matches the selection criteria in the SegmSelection action-info-args so that no PM-segments are found by the action, then this is not an error, a normal response is sent, and the segment info list will just be empty.

If the choice of SegmSelection in the Get-Segment-Info method is segm-id-list and the segm-id-list is empty, then the response shall be a segment-info-list that is empty.

~~If the agent supports the Get-Segment-Info method, the agent shall set the pmsc-get-segm-info-sup flag in the PM-Store-Capab attribute.~~

Insert text in Get-Segment-Id-List as shown:

— **Get-Segment-Id-List**

This method allows the manager to retrieve a list of the instance numbers of all the PM-segments of a PM-store. In particular, the Get-Segm-Id-List method allows the manager to then retrieve the attributes of selected PM-segment object instances and their data contents without needing to retrieve information of all PM-segments. This also allows the manager to retrieve multiple PM-segments as a series of requests.

If a standard configuration contains any PM-Store object, the manager must send Get-Segment-Info or Get-Segment-Id-List at the beginning of accessing any PM-Store object.

If the agent supports the Get-Segment-Id-List method, the agent shall set the pmsc-get-segm-id-list-sup flag in the PM-Store-Capab attribute.

If the agent supports the Get-Segment-Id-List method, the agent shall also support the Get-Segment-Info method with the segm-id-list choice in the SegmSelection action-info-args.

7. Personal health device service model

7.4 Specific application of object access EVENT REPORT services for personal health devices

7.4.3 Configuration event report

7.4.3.2 Agent device configuration

Change the second paragraph as shown:

The MDS object is not considered part of the configuration. A manager reassociating with an agent offering the same Dev-Configuration-Id cannot expect the MDS attribute values to be the same; for example, an agent may clear the mds-time-mgr-set-time ~~manager-set-time~~ bit as its clock has already been set.

STANDARDSISO.COM : Click to view the full PDF of ISO/IEEE 11073-20601:2016/Cor 1:2016