
**Biological evaluation of medical
devices —**

Part 19:
**Physico-chemical, morphological and
topographical characterization of
materials**

Évaluation biologique des dispositifs médicaux —

*Partie 19: Caractérisations physicochimique, morphologique et
topographique des matériaux*

STANDARDSISO.COM : Click to view the full PDF of ISO/TS 10993-19:2006



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

STANDARDSISO.COM : Click to view the full PDF of ISO/TS 10993-19:2006

© ISO 2006

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Symbols and abbreviated terms	2
5 General principles	2
6 Characterization procedure	3
6.1 General	3
6.2 Qualitative information	4
6.3 Material equivalence	4
6.4 Quantitative information	4
6.5 Quantitative assessment	4
7 Characterization parameters and methods	5
8 Reporting of data obtained	8
Annex A (informative) Principles for judging material equivalency	9
Annex B (informative) Nanoparticles — Special consideration in judging material equivalency and biological evaluation	10
Bibliography	11

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

In other circumstances, particularly when there is an urgent market requirement for such documents, a technical committee may decide to publish other types of normative document:

- an ISO Publicly Available Specification (ISO/PAS) represents an agreement between technical experts in an ISO working group and is accepted for publication if it is approved by more than 50 % of the members of the parent committee casting a vote;
- an ISO Technical Specification (ISO/TS) represents an agreement between the members of a technical committee and is accepted for publication if it is approved by 2/3 of the members of the committee casting a vote.

An ISO/PAS or ISO/TS is reviewed after three years in order to decide whether it will be confirmed for a further three years, revised to become an International Standard, or withdrawn. If the ISO/PAS or ISO/TS is confirmed, it is reviewed again after a further three years, at which time it must either be transformed into an International Standard or be withdrawn.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO/TS 10993-19 was prepared by Technical Committee ISO/TC 194, *Biological evaluation of medical devices*.

ISO 10993 consists of the following parts, under the general title *Biological evaluation of medical devices*:

- *Part 1: Evaluation and testing*
- *Part 2: Animal welfare requirements*
- *Part 3: Tests for genotoxicity, carcinogenicity and reproductive toxicity*
- *Part 4: Selection of tests for interactions with blood*
- *Part 5: Tests for in vitro cytotoxicity*
- *Part 6: Tests for local effects after implantation*
- *Part 7: Ethylene oxide sterilization residuals*
- *Part 9: Framework for identification and quantification of potential degradation products*

- *Part 10: Tests for irritation and delayed-type hypersensitivity*
- *Part 11: Tests for systemic toxicity*
- *Part 12: Sample preparation and reference materials*
- *Part 13: Identification and quantification of degradation products from polymeric medical devices*
- *Part 14: Identification and quantification of degradation products from ceramics*
- *Part 15: Identification and quantification of degradation products from metals and alloys*
- *Part 16: Toxicokinetic study design for degradation products and leachables*
- *Part 17: Establishment of allowable limits for leachable substances*
- *Part 18: Chemical characterization of materials*
- *Part 19: Physico-chemical, morphological and topographical characterization of materials*
- *Part 20: Principles and methods for immunotoxicology testing of medical devices*

STANDARDSISO.COM : Click to view the full PDF of ISO/TS 10993-19:2006

Introduction

ISO 14971 points out that a toxicological risk analysis should take account of the nature of the materials.

ISO 10993-1 provides a framework for a structured programme of assessment for the evaluation of biological safety. ISO 10993-1:2003, Clause 3, states that in the selection of materials to be used for device manufacture, the first consideration should be fitness for purpose. This should have regard to the characteristics and properties of the material, which include chemical, toxicological, physical, electrical, morphological and mechanical properties. This information is necessary prior to any biological evaluation. ISO 10993-1:2003, 7.2 notes that the continuing acceptability of a biological evaluation is an aspect of a quality management system.

The identification and evaluation of the physico-chemical, morphological and topographical properties of materials used in a finished medical device are important in determining the biological evaluation of that device and its materials. Such information can be used in:

- a) assessing the overall biological evaluation of a medical device (ISO 10993);
- b) screening of potential new materials and/or processes for suitability in a medical device for a proposed clinical application.

The compositional characteristics of the materials of manufacture are mainly under the control of the suppliers of these materials. However, other characteristics are chiefly influenced by the requirements to be met by the finished medical device as well as the production processes used by the medical device manufacturer.

STANDARDSISO.COM : Click to view the full PDF of ISO/TS 10993-19:2006

Biological evaluation of medical devices —

Part 19:

Physico-chemical, morphological and topographical characterization of materials

1 Scope

This Technical Specification provides a compilation of parameters and test methods that can be useful for the identification and evaluation of the physico-chemical, morphological and topographical (PMT) properties of materials in finished medical devices. Such an assessment is limited to those properties that are relevant to biological evaluation and the medical device's intended use (clinical application and duration of use) even if such properties overlap with clinical effectiveness.

This part of ISO 10993 does not address the identification or quantification of degradation products, which are covered in ISO 10993-9, ISO 10993-13, ISO 10993-14 and ISO 10993-15. Chemical characterization of materials is covered by ISO 10993-18.

The ISO 10993 series of International Standards is not applicable when the material or device does not contact the body directly or indirectly (see ISO 10993-1:2003, 4.2).

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 10993-1:2003, *Biological evaluation of medical devices — Part 1: Evaluation and testing*

ISO 10993-18, *Biological evaluation of medical devices — Part 18: Chemical characterization of materials*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 10993-1, ISO 10993-18 and the following apply.

3.1

physico-chemical

relating to the physical chemistry (of materials)

3.2

morphological

relating to the shape, contours and microstructural organization (of materials)

3.3

topographical

relating to the features of the surface (of materials)

4 Symbols and abbreviated terms

The following abbreviations are used throughout the document.

- NP: nanoparticle
- PMT: physico-chemical, morphological and topographical

The abbreviations listed in Table 1 are used in Clause 7.

Table 1 — Methodology abbreviations

Abbreviation	Analytical method
AES	Auger Electron Spectroscopy including scanning tunneling auger
AFM/SPM	Atomic Force Microscopy/Scanning Probe Microscopy including topographical roughness and phase contrast
BET	Brunauer-Emmett-Teller, a porosity measurement methodology
CLSM	Confocal Laser Scanning Microscopy
DMTA	Dynamic Mechanical Thermal Analysis
DSC	Differential Scanning Calorimetry
EPMA	Electron Probe Microanalyser
ESC	Equilibrium Solvent Content
EWC	Equilibrium Water Content
EDX-SEM	Energy Dispersive X-ray Analysis — Scanning Electron Microscopy
FTIR	Fourier Transform Infra Red (spectroscopy) including microscopy, imaging and diffuse reflectance
FTIR-ATR	Fourier Transform Infra Red (spectroscopy) — Attenuated Total Reflectance (multiple internal reflectance)
IR	Infra Red (spectroscopy)
OM	Optical Microscopy including polarized light and phase contrast microscopy
QCM	Quartz Crystal Microbalance (or other microbalance techniques)
SEM/TEM	Scanning Electron Microscopy/Transmission Electron Microscopy
SPR	Surface Plasmon Resonance
TOF/SIMS	Time of Flight — Secondary Ionization Mass Spectroscopy
TMA	Thermal Mechanical Analyser
XPS/ESCA	X-ray Photoelectron Spectroscopy/ Electron Spectroscopy for Chemical Analysis

5 General principles

Consideration of the PMT characterization of the materials from which a medical device is made, like chemical characterization of materials (addressed in ISO 10993-18), is a necessary step in assessing the biological safety and clinical effectiveness of the device. Both types of characterization are also important in judging equivalence of:

- a proposed material to a clinically established material,
- or
- a prototype device to a final device.

The relationship of PMT characteristics of materials used in devices, to their biocompatibility and clinical effectiveness, is still a developing area. However there are several examples of where these relationships are becoming better understood, as listed below.

- 1) The use of certain PMT characteristics of porous materials as surfaces on orthopaedic implants can encourage tissue in-growth at the surface of the implant and thus result in better integration with the surrounding tissue.
- 2) The use of material scaffolds and meshes, with certain PMT characteristics, as implants into injured soft or hard tissue can facilitate the beneficial infiltration of certain types of cells which aid the healing process (Dexter et al. [50]).
- 3) The PMT characteristics of the surfaces of materials used as catheters have a major influence on the adherence of bacteria and proteins to the inner and outer surfaces, which in turn influences the risk of infections and blockages.
- 4) Alterations to the micro-topography of surfaces, e.g. producing microgrooves or other defined patterns, has been shown to influence the adhesion and direction of movement of certain types of cells on that surface (Alaerts et al. [46]; Dewez et al. [49]).
- 5) For certain medical devices e.g. orthopaedic implants and vascular prostheses, mechanical properties may influence biological responses such as tissue re-modelling.

NOTE The shape and geometric form of medical devices and their components are known to affect the biological response, e.g. aspect ratio, thickness, form with relation to blood flow. Information on specific devices may be found in the applicable product standards.

This technical specification provides a range of examples of PMT characterization parameters and methods which may be usefully applied in the PMT characterization of materials utilized in medical devices.

Medical device manufacturers should select relevant parameters and methods and justify their selection. Manufacturers should document the level of characterization performed on their medical device and its component materials, appropriate to its clinical application.

The extent of characterization should reflect the nature and duration of the clinical exposure and may be useful for risk assessment of the biological safety of the device. The PMT characterization should also reflect the materials used and their physical form(s), e.g. solid, liquid, gel, polymer, metal, ceramic, composite or biologically sourced material. Characterization generally requires the close collaboration of material scientists, analytical scientists and risk assessors.

6 Characterization procedure

6.1 General

The analytical methods should be selected to give the required information for the evaluation. Prior to new method development, existing standards, monographs, scientific articles or other relevant scientific documents should be consulted to check for existing appropriate test methods. Methods from the literature may need to be adapted and validated before use. If suitable methods cannot be identified, appropriate new methods should be developed.

The analytical methods used should be validated, justified and reported in line with Clauses 7 and 8. The validation of an analytical method is the process by which it is established that the performance characteristics of the method meet the requirements for the intended analytical applications. Analytical methods should be validated as appropriate with respect to the following justified analytical characteristics: accuracy, precision, specificity, limit of detection, limit of quantification, linearity, range, ruggedness, robustness and system suitability.

At each step of the characterization procedure, a decision should be made on the adequacy of the data obtained, as a basis for the risk analysis. This procedure should consider each of the materials as they appear in the finished device.

NOTE The supplier may be a useful source of appropriate analytical methods. In the absence of any initial data on material properties, a literature study is recommended to assist in the selection of the most appropriate methods of analysis for the material concerned.

6.2 Qualitative information

Describe the material/device and its intended purpose. A documented, qualitative description of the PMT characteristics of the finished device is recommended, including the characteristics of each material used in the device (see 3.2 and Clause 4 of ISO 10993-1: 2003) (see Annex A). The level of qualitative data provided should reflect the category of medical device in terms of degree of invasiveness and clinical exposure duration, as well as the nature of the materials present.

The qualitative description should, where applicable, include details of batch or lot, supplier and material specification for each material.

Medical device manufacturers should obtain qualitative and quantitative material characterization information, relevant to the final product. Such information may be obtained from the supplier of the starting material, the literature or additional testing. The PMT characteristics of materials should either be in accordance with applicable materials standards or should be specified by the manufacturer. It is important to obtain as much information as possible at this early stage to be able to gain a thorough understanding of the hazards (potential risks) and potential benefits arising from the properties of the material, and to develop an initial assessment of the fitness for the intended purpose. This assessment will be further refined as additional information is gained during the product development process.

6.3 Material equivalence

As a part of material suitability assessment, a comparison of these data should be made to determine whether this material is equivalent to that utilized in a device with the same clinical exposure/use and having had the same manufacturing and sterilization processes applied, e.g. established safe and effective use of materials in a product to be used on intact skin. Annex A gives further guidance on judging material equivalency and Annex B gives information on the special case of material used as nanoparticles (≤ 100 nm in any one dimension).

6.4 Quantitative information

Where qualitative material characterization data alone have not provided sufficient data for a material suitability assessment to be completed, quantitative material characterization data should be established, documented and subjected to assessment of suitability and risk.

6.5 Quantitative assessment

Sufficient quantitative characterization information should be obtained in order to permit an assessment of the fitness of all of the materials in a finished device for their intended purpose as part of the overall biological evaluation of the medical device. This quantitative characterization information can be usefully compared with data for materials and/or finished medical devices clinically established as being safe and effective for the intended use. The characterization information may also be usefully compared to those materials/products found not to have the required characteristics for this use. This overall evaluation is outside the scope of this part of ISO 10993 and will combine information gained from many other parts of the ISO 10993 series of International Standards and will utilize ISO 14971.

7 Characterization parameters and methods

Clause 6 indicates the generation of qualitative and quantitative PMT characterization data for use in the suitability/risk assessment. Table 2 summarizes examples of parameters for material characterization and examples of methods which can be used to provide qualitative and/or quantitative data for these. Relevant standards and/or references are given for the parameter section, e.g. topography. Not all of the parameters or the associated methods may be applicable to all types of material. The characterization parameters used should be selected appropriate to the material or finished medical device. Due to the diversity of medical devices, it is recognized that not all of the parameters identified for a material will be relevant for all/some device uses. As noted in 6.2 the extent of characterization which should be considered is determined by the invasiveness and duration of clinical exposure in the intended use.

The analyst and material scientist in consultation with the manufacturer's assessor of the material fitness for use (risk assessor) should determine which parameters are relevant to the assessment of a material or medical device. Characterization data should be considered for all of the parameters considered relevant by the manufacturer's risk assessor.

NOTE For natural macromolecules, it is essential that the source organism (species) and breed/strain be clearly identified as a first step. The ISO 22442 series of standards covers the safe utilisation of non-human tissues and their derivatives in the manufacture of medical devices. EN 455-3 covers the assessment of risks associated with protein residues in natural rubber latex.

Natural macromolecules utilized in medical devices include but are not limited to proteins, glycoproteins, polysaccharides and ceramics. Examples include gelatine, collagen, elastin, fibrin, albumin, alginate, cellulose, heparin, chitosan, processed bone, coral and natural rubber. These materials may have been processed, purified and modified to different extents. Pharmacopoeia monographs exist for many of these materials and several ASTM F04 standards also cover the characterization of these materials.

Table 2 — Examples of parameters and test methodologies for characterization of materials including polymers, metals, alloys, ceramics and natural macromolecules

Examples of parameters to be analysed	Example methods (not comprehensive or exclusive)	Qualitative	Quantitative	Standard or reference
Porosity				
Classical	OM Gas adsorption (BET) Mercury porosimetry Helium pycnometry	× — — ×	— × × —	ASTM F1854-01 [39] ISO 18754 [22] ISO 18757 [23]
Connectivity	SEM AFM	— ×	× —	
Scaffolds	SEM AFM	— ×	× —	
Morphology				
Crystallinity	X-Ray Diffraction OM DSC SEM/TEM AFM	× × × × ×	× — × — —	ASTM F665 [35] ASTM F754 [38] ASTM F2081 [42] ASTM F2183 [44] Hasegawa and Hashimoto [55] Kajiyama et al. [58] Kajiyama et al. [59]
Amorphous	DMTA AFM	× ×	× —	
Multiple phases	OM AFM TEM	× × ×	× — —	
Hard/soft surfaces	OM AFM/SPM TEM Ultrasound	× × × ×	× × — —	

Table 2 (continued)

Examples of parameters to be analysed	Example methods (not comprehensive or exclusive)	Qualitative	Quantitative	Standard or reference
Surface energy/charge				
Hydrophobic	Wettability (contact angle)	×	×	EN 828 [27] Collier et al. [48] Dewez et al. [49] Dexter et al. [50] Ebara and Okahata [53] Ikada [56] Jenney and Anderson [57] Kishida et al. [60] MacDonald et al. [65] Niihura et al. [69] Quirk et al. [71] Senshu et al. [72] Senshu et al. [74] Tamada et al. [76] Wagner et al. [77] Weber et al. [79]
Hydrophilic	Wettability (contact angle)	×	×	
Protein adsorption	QCM or SPR	×	×	
	CLSM	×	×	
	Biochemical analysis	×	×	
	RIA	×	×	
Protein repulsion	QCM or SPR	×	×	
	CLSM	×	×	
	Biochemical analysis	×	×	
	Radioimmunoassay	×	×	
Cell attachment — Human general — Human blood cells — Human specific cells — Bacteria general — Bacteria specific class	OM	×	×	
	QCM	×	×	
	CLSM	×	×	
Cell repulsion — Human general — Human blood cells — Human specific cells — Bacteria general — Bacteria specific class	OM	×	×	
	QCM	×	×	
	CLSM	×	×	
Abrasion resistance				
Stability of treated surface	IR	×	—	ASTM D968 [31] ASTM D1044 [32] ASTM D1894 [33] ASTM D4060 [34] ASTM F732 [36] ASTM F735 [37] ASTM F1978 [41]
Surface friction	Volume loss, strain gauge	×	×	
	Coefficient of friction	×	×	
	AFM/SPM	×	×	

STANDARDSISO.COM : Click to view the full PDF of ISO/TS 10993-19:2006

Table 2 (continued)

Examples of parameters to be analysed	Example methods (not comprehensive or exclusive)	Qualitative	Quantitative	Standard or reference
Topography				
Surface chemical mapping	XPS/ESCA TOF-SIMS FTIR/ATR FTIR-Microscopy FTIR-Imaging EDX-SEM Raman EPMA	x x x x x x x x	x — — — — — — x	ISO 3274 [1] ISO 4287 [2] ISO 4288 [3] ISO 5436-1 [4] ISO 5436-2 [5] ISO 11562 [11] ISO 12179 [12] ISO 13565-1 [15] ISO 13565-2 [16] ISO 13565-3 [17]
Roughness				
— smooth — pitted — grooved — irregular terrain (“hills, valleys”)	SEM AFM/SPM Tribology Profilometry	x x x x	— x — —	ISO 18754 [22] ISO 18757 [23] EN 623-4 [25] Alaerts et al. [46] Ikada [56] Kajiyama et al. [58] Kumaki et al. [63] Senshu et al. [72] Senshu et al. [73] Senshu et al. [74]
Particles				
Size Size distribution 3D shape	OM Laser diffraction Image analysis Filters (sieves) SEM	x x x x x	x x x — —	ASTM F1877 [40] ISO 13319 [13] ISO 13320-1 [14] ISO/TS 13762 [18] ISO 17853 [21] EN 725-5 [26]
Shape and Form				
	SEM AFM/SPM OM	x x x	x x x	
Swelling				
Water absorption Solvent absorption Shape change Surface crazing Weight gain	EWC ESC Image analysis OM SEM TMA Microbalance	x x x x x x x	x x x x x x x	ISO 17190-5 [20] Moskala and Jones [66]

8 Reporting of data obtained

Test reports should clearly state the purpose of the characterization that has been performed and, where appropriate, should include the following:

- a) description and details of material or finished medical device;
- b) characterization methods;
- c) qualitative data generated;
- d) quantitative data generated.

For areas for which no standards exist, the qualitative and quantitative PMT data generated will be collected and documented for information purposes. These data may be used to assist in the development of appropriate future standards.

STANDARDSISO.COM : Click to view the full PDF of ISO/TS 10993-19:2006

Annex A (informative)

Principles for judging material equivalency

In 6.3 characterization data is used in risk assessment to judge equivalency of a proposed material to an existing clinically established material or finished medical device for the same type of clinical exposure. The key principle applied in making this judgement is that the proposed material or finished medical device has equivalent properties in terms of biological safety and clinical performance to that of the clinically established material or finished medical device. The following list contains examples where such equivalence is met and is provided as guidance to assist the assessment process as outlined in Clauses 5 and 6.

- a) The proposed material or finished medical device meets an existing standard for its intended use and duration of contact and invasiveness.
- b) The proposed material or finished medical device is already established in a more invasive comparable exposure than the less invasive proposed application.
- c) The proposed material or finished medical device has properties very close to a clinically accepted material or finished medical device.

Annex B (informative)

Nanoparticles — Special consideration in judging material equivalency and biological evaluation

Where a material is composed of nanoparticles (NPs, ≤ 100 nm diameter) it should not be assumed that the material has equivalent physico-chemical or biological exposure/effects to that composed of larger particles, e.g. in the μm range or larger. NPs currently in commercial use have been made from, e.g. transition metals, silicon, carbon (single-wall carbon nanotubes; fullerenes) and metal oxides (zinc oxide and titanium dioxide). NPs should be assessed on a case-by-case basis taking into account the route and duration of exposure and interactions with fluids, cells and tissues. NPs are known to have a tendency to aggregate.

Although very few toxicology studies of manufactured NPs exist, those that do, indicate that NPs have unique physico-chemical properties and pulmonary toxicity at high doses which are probably related to the large surface area:mass ratio (Donaldson et al. [51]; Dreher [52]; Everitt and Bermudez [54]; Lam et al. [64]; Warheit et al. [78]). Lung inflammation to carbon black NPs increased in direct proportion to the total particle surface area instilled and additional effects have been seen in the presence of transition metal ions on the surface of the NPs (Brown et al. [47]; Wilson et al. [80]). Large species difference in lung toxicity susceptibility to inhaled titanium dioxide NPs has been observed, even in rodents, with rat being most sensitive, mouse intermediate and hamster insensitive (Everitt and Bermudez [54]). NPs may have the ability to redistribute from their site of deposition, e.g. following inhalation exposure, NPs have been reported to travel via the nasal nerves to the brain and from the lung into the blood (Kreuter et al. [61]; Kreyling et al. [62]; Nemmar et al. [67]; Oberdorster et al. [70]). Lung instillation of polystyrene NPs had an effect on blood coagulability in the hamster and formation of clots was increased when the surface of the NPs were aminated, i.e. made more reactive (Nemmar et al. [68]). NPs of materials of otherwise low toxicity and low solubility may generate free radicals at their surface which can result in an effect on calcium signalling which can lead to inflammation (Stone et al. [75]). For exposure to NPs, the dose of NPs may be better described in terms of total surface area and/or number of particles/kg body weight rather than mg NPs/kg body weight.

Bibliography

- [1] ISO 3274:1996, *Geometrical Product Specifications (GPS) — Surface texture: Profile method — Nominal characteristics of contact (stylus) instruments*
- [2] ISO 4287:1997, *Geometrical Product Specifications (GPS) — Surface texture: Profile method — Terms, definitions and surface texture parameters*
- [3] ISO 4288:1996, *Geometrical Product Specifications (GPS) — Surface texture: Profile method — Rules and procedures for the assessment of surface texture*
- [4] ISO 5436-1:2000, *Geometrical Product Specifications (GPS) — Surface texture: Profile method; Measurement standards — Part 1: Material measures*
- [5] ISO 5436-2:2001, *Geometrical Product Specifications (GPS) — Surface texture: Profile method; Measurement standards — Part 2: Software measurement standards*
- [6] ISO 5832-1:1997, *Implants for surgery — Metallic materials — Part 1: Wrought stainless steel*
- [7] ISO 10993-9:1999, *Biological evaluation of medical devices — Part 9: Framework for identification and quantification of potential degradation products*
- [8] ISO 10993-13:1998, *Biological evaluation of medical devices — Part 13: Identification and quantification of degradation products from polymeric medical devices*
- [9] ISO 10993-14:2001, *Biological evaluation of medical devices — Part 14: Identification and quantification of degradation products from ceramics*
- [10] ISO 10993-15:2000, *Biological evaluation of medical devices — Part 15: Identification and quantification of degradation products from metals and alloys*
- [11] ISO 11562:1996, *Geometrical Product Specifications (GPS) — Surface texture: Profile method — Metrological characteristics of phase correct filters*
- [12] ISO 12179:2000, *Geometrical Product Specifications (GPS) — Surface texture: Profil method — Calibration of contact (stylus) instruments*
- [13] ISO 13319:2000, *Determination of particle size distributions — Electrical sensing zone method*
- [14] ISO 13320-1:1999, *Particle size analysis — Laser diffraction methods — Part 1: General principles*
- [15] ISO 13565-1:1996, *Geometrical Product Specifications (GPS) — Surface texture: Profile method; Surface having stratified functional properties — Part 1: Filtering and general measurement conditions*
- [16] ISO 13565-2:1996, *Geometrical Product Specifications (GPS) — Surface texture: Profile method; Surfaces having stratified functional properties — Part 2: Height characterization using the linear material ratio curve*
- [17] ISO 13565-3:1998, *Geometrical Product Specifications (GPS) — Surface texture: Profile method; Surfaces having stratified functional properties — Part 3: Height characterization using the material probability curve*
- [18] ISO/TS 13762:2001, *Particle size analysis — Small angle X-ray scattering method*
- [19] ISO 14971:2000, *Medical devices — Application of risk management to medical devices*