

**AEROSPACE  
MATERIAL  
SPECIFICATION**



**AMS 7295/7C**

Issued APR 1978  
Revised APR 1993  
Noncurrent JUL 1998  
Cancelled FEB 2006  
Superseding AMS 7295/7B

Radiographic Film  
3-30

**CANCELLATION NOTICE**

This specification has been declared "CANCELLED" by the Aerospace Materials Division, SAE, as of February, 2006. By this action, this document will remain listed in the Numerical Section of the Index of Aerospace Material Specifications indicating that it has been "CANCELLED".

Cancelled specifications are available from SAE.

SAENORM.COM : Click to view the full PDF of AMS 7295-7C

SAE Technical Standards Board Rules provide that: "This report is published by SAE to advance the state of technical and engineering sciences. The use of this report is entirely voluntary, and its applicability and suitability for any particular use, including any patent infringement arising therefrom, is the sole responsibility of the user."

SAE reviews each technical report at least every five years at which time it may be reaffirmed, revised, or cancelled. SAE invites your written comments and suggestions.

Copyright © 2006 SAE International

All rights reserved. No part of this publication may be reproduced, stored in a retrieval system or transmitted, in any form or by any means, electronic, mechanical, photocopying, recording, or otherwise, without the prior written permission of SAE.

**TO PLACE A DOCUMENT ORDER:** Tel: 877-606-7323 (inside USA and Canada)  
Tel: 724-776-4970 (outside USA)  
Fax: 724-776-0790  
Email: [custsvc@sae.org](mailto:custsvc@sae.org)  
**SAE WEB ADDRESS:** <http://www.sae.org>

**1. SCOPE:****1.1 Form:**

This specification covers one type of low-contrast, coarse-grain radiographic film in the form of cut sheets or rolls.

**1.2 Application:**

This film will provide a radiographic quality level of 2-2T sensitivity when properly utilized (See AMS 7295).

**1.3 Classification:**

Low contrast with gradient of 3.0 to 3.9 and exhibiting resolution of 20 to 39 lines per millimeter.

**2. APPLICABLE DOCUMENTS:**

See AMS 7295.

**3. TECHNICAL REQUIREMENTS:****3.1 Basic Specification:**

The complete requirements for film described herein and its procurement shall consist of this document and the latest issue of the basic specification, AMS 7295.

3.1.1 Definition: Properties shall be as defined in AMS 7295.

**3.2 Properties of Radiographic Film:**

3.2.1 Thickness: Shall be 0.007 inch  $\pm$  0.001 (0.18 mm  $\pm$  0.03).

3.2.2 Net Optical Density: Shall be not greater than 0.30 H&D units.

3.2.3 Gradient or Contrast: Shall be 3.0 to 3.9 at 1.5 to 3.5 H&D units.

3.2.4 Resolution: Shall be coarse, showing 20 to 39 lines per millimeter.

**4. QUALITY ASSURANCE PROVISIONS:**

See AMS 7295.

**5. PREPARATION FOR DELIVERY:**

See AMS 7295.