



AEROSPACE STANDARD

AS 1224

Society of Automotive Engineers, Inc.
TWO PENNSYLVANIA PLAZA, NEW YORK, N. Y. 10001

Issued 9-15-71
Revised

CONTINUOUS FLOW GENERAL AVIATION OXYGEN MASKS

1. SCOPE

- 1.1 **Purpose:** This standard defines the minimum requirement for the design, construction and performance of continuous flow oxygen masks for crew and passengers of general aviation civil aircraft.
- 1.2 **Types:** This standard includes the following types of continuous flow oxygen masks.
- Open port dilution rebreathing masks.
 - Valved or restrictive phase dilution rebreathing masks.
 - Valved or restrictive phase dilution reservoir masks.
 - Open port or restrictive dilution mask without rebreathing or reservoir bag.
- 1.3 **Description:** The masks shall be of an oronasal type covering the nose and mouth, utilizing a continuous supply of oxygen and consisting of the facepiece, valving, mask suspension device, rebreathing or reservoir bag (except Type "d"), supply tube, and including connector and flow indicator (when used).
- 1.4 **Definition:** The mask types are defined as follows:
- Open port dilution rebreathing mask - This is a type of mask incorporating a rebreather bag into which exhaled gases, high in oxygen content from the first portion of the previous exhalation, are introduced to be inspired again upon the next inspiration. Dilution of oxygen flowing into the mask is accomplished by fixed nonvariable orifices incorporated in the body of the mask or system, allowing dilution by introduction of ambient air during inspiration. Valving is not present between the mask facepiece and rebreathing bag.
 - Valved or restrictive phase dilution rebreathing mask - This type of mask may be either of the following:
 - (1) A mask utilizing a rebreather bag into which a constant flow of oxygen is introduced. A check valve between the mask and ambient air is provided so that the ambient air will not be admitted before the rebreather bag has been depleted. Valving is not present between the mask facepiece and the rebreather bag, or
 - (2) A mask utilizing a rebreathing bag but incorporating a restrictive sponge or other means which admits dilution air when subjected to a significant decrease in intra-mask pressure. Valving is not present between the mask facepiece and rebreather bag.

SAE Technical Board rules provide that: "All technical reports, including standards approved and practices recommended, are advisory only. Their use by anyone engaged in industry or trade is entirely voluntary. There is no agreement to adhere to any SAE standard or recommended practice, and no commitment to conform to or be guided by any technical report. In formulating and approving technical reports, the Board and its Committees will not investigate or consider patents which may apply to the subject matter. Prospective users of the report are responsible for protecting themselves against infringement of patents."

- c. Valved or restrictive phase dilution reservoir mask - This type of mask provides the most efficient physiological use of constant flow aircraft oxygen and may be one of the following:
- (1) A mask utilizing a reservoir bag incorporating a check valve between the mask facepiece and reservoir bag to prevent introduction of exhaled gases into the reservoir bag and to assure 100% oxygen in the reservoir. Dilution is accomplished at the later phases of inspiration by a loaded ambient air valve which introduces ambient air following depletion of the 100% oxygen content of the reservoir bag, or
 - (2) A phase dilution mask utilizing a reservoir bag with valve between the bag and the mask. A porous restrictive dilution port to provide inhalation of ambient air at a slower rate than oxygen from the reservoir, furthermore permitting the exhaled air to leave the mask.
- d. Open port or restrictive dilution mask without rebreathing or reservoir bag - This mask is defined as a mask which incorporates dilution ports or restrictive dilution such as by the use of open cell foam construction or one or more air inlet valves. This mask is not recommended for efficient physiological use of the aircraft oxygen, commensurate with requirements for pilot safety or survival in hypoxic environments at altitude.

2. MATERIALS

- 2.1 **General:** Materials shall be of type, grade and quality which experience and/or tests have shown to be suitable for the purpose intended. Materials shall not be used which contaminate oxygen or are adversely affected by continuous service with oxygen. Materials shall have at least flash resistant properties as specified in the Federal Aviation Regulations 23.853, or 3.1.3.1 of National Aerospace Standard (NAS) 1179.
- 2.1.1 **Facepiece:** The facepiece shall be free of objectionable odors. Materials in contact with the skin shall be selected to be as non-irritating, non-allergenic, soft and compliant to the facial configuration as practical.
- 2.1.2 **Cleaning and Sterilizing:** The mask shall be made of materials which will permit cleaning and sterilization without adverse effects and without disassembly. The method of cleaning and sterilizing shall be recommended by the manufacturer.
- 2.1.3 **Elastomer Components:** A tag or leaflet describing elastomeric components and a suggested method for inspection and detection of any deterioration in these components which may adversely affect the performance of the mask shall be attached to or included in the packaged mask prior to delivery to the user.
- 2.1.4 **Fungus:** Components shall be fungus-proofed by selection of parts and materials that are non-nutrient to fungus, or by treatment of the parts and materials prior to their use.
- 2.1.5 **Dissimilar Metals:** Unless suitably protected against electrolytic corrosion, dissimilar metals shall not be used in intimate contact with each other.

3. WORKMANSHIP

- 3.1 **General:** The mask shall be fabricated and finished in accordance with the highest grade practice in the manufacture of this type of equipment. The finished mask and all internal parts shall be clean throughout and free of fins, burrs, scale, oils, materials or any other conditions which might adversely affect the safe operation of the mask.

- 3.2 Finish: All materials which are not inherently corrosion-resistant shall be finished with a protective treatment or coating to minimize the effect of exposure to the environmental conditions which may be encountered in the service for which the equipment is intended. The protective treatment or coating shall not chip, flake, or powder, or otherwise contaminate the mask.
- 3.3 Strength: The rebreather bag, reservoir bag, oxygen supply tube, and facepiece shall be assembled in such a manner that when supported at the inlet end of the supply inlet, the assembly shall support a static load of 20 lb, applied at the facepiece for a period of three seconds after having been subjected to the environmental conditions as specified in 3.4.1.1.

The rebreather or reservoir bag shall be easily removed and replaced, but should be so designed that it may not be inadvertently detached while in use. All components of the mask assembly shall be resistant to tears, snags, embrittlement and cracking, and all other defects leading to malfunction of the article, which might be caused by normal handling during the service life.

- 3.3.1 Rebreather or Reservoir Bag: The rebreather or reservoir shall be capable of withstanding two psi pressure for three seconds without failure or damage.

3.3.2 Facepiece Assembly:

- 3.3.2.1 Fit: The facepiece shall be of sufficient resilience, size, and shape to conform readily to facial contours using no more pressure than supplied by the mask suspension device. Mask design should consider extremes of Nasion-Menton, Bizygomatic, Bigonial, and Nasion-Supramentale distances and other applicable anthropometric information required to provide an adequate fit. Suggested sources of information include:

Adults: Human Body Size in Military Aircraft and Personal Equipment, AAF Tech. Report 5501, June 10, 1946.

Anthropometry of Flying Personnel, WADC Tech. Report 52-321, USAF, Sept. 1954, Hertzberg, et. al.

Anthropometry of WAF Basic Trainees, WADC Tech. Report 53-12, July 1953, Danish, et. al.

Annotated Bibliography of Applied Physical Anthropology in Human Engineering, WADC Tech. Report 56-30, May 1958, edited by H. T. E. Hertzberg.

Children: Selected Facial Measurements of Children for Oxygen Mask Design, J. W. Young, FAA, Office of Aviation Medicine, Civil Aeromedical Institute, Report AM 66-9, April 1966, Okla. City, Okla.

- 3.3.2.2 Volume: The chamber formed between the face and the mask shall be so configured as to provide proper fit and adequate comfort with minimal dead-space which could contribute to carbon dioxide retention.
- 3.3.2.3 Resilience: The main body of the mask shall be resilient enough to minimize deformation due to incorrect handling, which might compromise operating performance.
- 3.3.2.4 Valve Design: Valves, when used, shall be designed to offer the following maximum resistance:

Oxygen inflow valve: 0.8 in. H₂O at 30 lpm.
2.0 in. H₂O at 70 lpm.

Exhalation valve: 2.0 in. H₂O at 85 lpm.

- 3.3.2.5 Attachment of the Mask Suspension Device: Provisions shall be made on the facepiece for attaching a mask suspension device.
- 3.3.3 Mask Suspension Device: A simple mask suspension device shall be provided for holding the mask on the user's face. A simple means of adjusting the tightness of the mask to the face shall be provided.
- 3.3.4 Gas Bag (when used): The gas bag, when used, shall be attached to the facepiece and shall be made of flexible material which will collapse when not in use. Strength of the attachment shall conform to para. 3.3. The gas bag(s) shall have the following minimum and maximum capacities when inflated to 0.5 inches of water positive pressure.

Type	Min. Vol. (cc.)	Max. Vol. (cc.)
Rebreather	800	1000
Reservoir	1000	1500

- 3.3.5 Tubing: The oxygen supply tube which is to be used for connecting the mask assembly to the oxygen supply source shall be lightweight and adequate for its intended use. The tubing shall be fully flexible within a temperature range of -29 C (-20 F) to 71 C (160 F). The tubing shall be capable of being bent to a 1.0 in. radius during stowage with less than 5% restriction to flow at 30 lpm. NTPD upon being removed from stowage.

3.4 Performance Requirement:

3.4.1 Environmental Conditions:

- 3.4.1.1 Temperature and Humidity: The assembly shall be capable of being stowed at temperatures of 71 C (160 F) for 120 hr and -55 C (-67 F) for 2 hr, at relative humidities varying from 5% to 95%, without affecting subsequent performance after return to normal temperature (72 F).
- 3.4.1.2 Condensation: Operation of the equipment shall not be adversely affected by moisture accumulated during use.
- 3.4.1.3 Vibration: The mask assembly shall be capable of performing adequately after exposure to one million cycles, or 8 hr, of vibration at the resonant frequency or frequencies as determined by a scan of vibration frequencies as plotted in Figure 1, NAS 1179, Feb. 1961. At each resonant frequency detected, the mask shall be exposed to one million cycles or 8 hr vibration at this frequency and amplitude, as shown in Figure 1, with the mask mounted: (a) with the direction of the vibration in line with the principal axis of the mask, and (b) in a direction perpendicular to the principal axis of the mask.
- If no resonances are found within the frequency range of Figure 1, vibrate the mask at 100 cycles per second and the corresponding amplitude of Figure 1, for one million cycles in each direction.
- 3.4.1.4 Leakage Tests: Gas and supply tube leakage - with all openings to the mask facepiece sealed, apply a pressure of at least 2 in. H₂O to the supply tube inlet. The gas bag, including connections to the facepiece and supply tubing, shall not exhibit a leakage in excess of 10 cc. per minute when pressurized to the above value.

4. MASK PERFORMANCE

4.1 Use: It is the intent of this standard that the mask assembly, as used by passengers and crew members of non-pressurized and pressurized general aviation aircraft, shall be capable of maintaining during inspiration a mean tracheal oxygen partial pressure of no less than 100 mm Hg when engaged in flights not to exceed 35,000 feet. Constant flow passenger masks which provide reliable concentrations approaching 100% oxygen shall be provided passengers of pressurized aircraft engaged in flights above 35,000 ft and not to exceed 41,000 ft for which a potential exists for pressurization failure and emergency cabin decompression to the flight altitude of the aircraft. These mask designs shall be capable of maintaining during inspiration a mean tracheal oxygen partial pressure of no less than 83.8 mm Hg. It is recommended that constant flow oxygen masks as detailed in this standard not be used by those individuals in control of pressurized or non-pressurized aircraft above an altitude of 35,000 feet. Oxygen requirements and masks for crew members controlling aircraft at higher altitudes are detailed in AIR 825, AIR 1069, and AS 452A.

4.1.1 Altitude Limitations: It is recommended that constant flow masks of the following types be limited to the following altitudes depending upon their use in pressurized or non-pressurized aircraft and whether they are to be worn by the passengers or by individuals in control of the aircraft.

- a. Open port dilution rebreathing mask (passengers & crew).
 - 1. Pressurized ----- 27,000 feet
 - 2. Non-Pressurized ----- 25,000 feet
- b. Valved or restrictive phase dilution rebreathing masks (passengers & crew).
 - 1. Pressurized ----- 35,000 feet
 - 2. Non-Pressurized ----- 30,000 feet
- c. Valved or restrictive phase dilution reservoir masks (passenger).
 - 1. Pressurized ----- 41,000 feet
 - 2. Non-Pressurized ----- 35,000 feet
 (crew)
 - 1. Pressurized ----- 35,000 feet
 - 2. Non-Pressurized ----- 35,000 feet
- d. Open port or restrictive phase dilution mask without a rebreather or reservoir bag (passenger & crew).
 - 1. Pressurized ----- 20,000 feet
 - 2. Non-Pressurized ----- 15,000 feet

4.1.2 Performance Test "A": The performance of the mask shall be established by first determining a typical fit leakage value on representative randomly selected subjects, and then testing the mask on a breathing machine which can consistently produce the minute and tidal volumes specified. The typical fit leakage tests shall be carried out as specified in para. 4.1.7.1 of NAS 1179. In lieu of the leakage test method prescribed in NAS 1179, 4.1.7.1.1, the nitrogen dilution method (which is considered to be a more definitive leakage test) as detailed in the final report of FAA Contract No. FAA-885, Feb. 8, 1962, may be used.

4.1.2.1 Breathing Machine Tests With Typical Fit Leakage: When using a breathing machine for test purposes, hermetically seal the facepiece of the test mask against the face of the breathing machine and incorporate a mechanism to provide the typical fit leakage. A suitable gas sampling orifice shall be provided in the portion of the breathing machine representing the trachea. The breathing machine shall have a simulated tracheal dead space of approximately 150 cc. and shall insure a thorough mixing of the gases in the simulated lung. The respiratory rate and tidal volume shall be sufficiently adjustable to meet requirements of this standard. The simulated respiratory flow shall essentially follow a pattern such as occurs in human respiration. An outline of a typical dynamic testing system for demand regulators which may be modified for constant flow system testing is detailed in ARP 1109.